

(19) World Intellectual Property Organization
International Bureau(43) International Publication Date
7 March 2002 (07.03.2002)

PCT

(10) International Publication Number
WO 02/17780 A1(51) International Patent Classification⁷: A61B 5/00Charles; 61 Mira Loma, Orinda, CA 94563 (US).
SWEDLOW, David; 2165 Canyon Oak Lane, Danville, CA 94506 (US).

(22) International Filing Date: 31 August 2001 (31.08.2001)

(74) Agents: CAHILL, Steven, J. et al.; Two Embarcadero Center, 8th Floor, San Francisco, CA 94111-3834 (US).

(25) Filing Language: English

(81) Designated States (national): AU, CA, JP.

(26) Publication Language: English

(84) Designated States (regional): European patent (AT, BE, CH, CY, DE, DK, ES, FI, FR, GB, GR, IE, IT, LU, MC, NL, PT, SE, TR).

(30) Priority Data:


60/229,616 31 August 2000 (31.08.2000) US
09/943,805 30 August 2001 (30.08.2001) US**Published:**

- with international search report
- before the expiration of the time limit for amending the claims and to be republished in the event of receipt of amendments

(71) Applicant: NELLCOR PURITAN BENNETT INC.
[US/US]; 4280 Hacienda Drive, Pleasanton, CA 94588 (US).

For two-letter codes and other abbreviations, refer to the "Guidance Notes on Codes and Abbreviations" appearing at the beginning of each regular issue of the PCT Gazette.

(54) Title: OXIMETER SENSOR WITH DIGITAL MEMORY ENCODING SENSOR DATA



WO 02/17780 A1

(57) Abstract: The present invention provides a memory chip for use in an oximeter sensor, or an associated adapter or connector circuit. The memory chip allows the storing of different data to provide enhanced capabilities for the oximeter sensor. In addition to providing unique data to store in such a memory, the invention describes unique uses of data stored in such a memory. The data stored in the memory chip may include information relating to use of the oximeter sensor. For example, the memory chip may encode a sensor model identification that can be displayed on a display screen when the sensor is connected to an oximeter monitor. The memory may also encode a range of operating parameters such as light levels over which the sensor can function or a maximum drive current. The operating parameters are read and interpreted by a controller circuit to control the pulse oximetry system.

OXIMETER SENSOR WITH DIGITAL MEMORY ENCODING SENSOR DATA

CROSS-REFERENCES TO RELATED APPLICATIONS

5 [01] This application claims priority to U.S. Provisional Patent Application No. 60/229,616, filed August 31, 2001, which is hereby incorporated by reference herein.

BACKGROUND OF THE INVENTION

10 [02] The present invention relates to oximetry sensors and, in particular, pulse oximetry sensors which include coded information relating to characteristics of the sensor.

15 [03] Pulse oximetry is typically used to measure various blood flow characteristics including, but not limited to, the blood-oxygen saturation of hemoglobin in arterial blood, the volume of individual blood pulsations supplying the tissue, and the rate of blood pulsations corresponding to each heartbeat of a patient. Measurement of these characteristics has been accomplished by use of a non-invasive sensor which passes light through a portion of the patient's tissue where blood perfuses the tissue, and photoelectrically senses the absorption of light in such tissue. The amount of light absorbed is then used to calculate the amount of blood constituent being measured.

20 [04] The light passed through the tissue is selected to be of one or more wavelengths that are absorbed by the blood in an amount representative of the amount of the blood constituent present in the blood. The amount of transmitted light passed through the tissue will vary in accordance with the changing amount of blood constituent in the tissue and the related light absorption. For measuring blood oxygen level, such sensors have been 25 provided with light sources and photodetectors that are adapted to operate at two different wavelengths, in accordance with known techniques for measuring blood oxygen saturation.

30 [05] An encoding mechanism is shown in U.S. Patent No. 4,700,708, the disclosure of which is incorporated herein by reference. This mechanism relates to an optical oximeter probe which uses a pair of light emitting diodes (LEDs) to direct light through blood-perfused tissue, with a detector picking up light which has not been absorbed by the tissue. The operation depends upon knowing the wavelength of the LEDs. Since the wavelength of LEDs can vary, a coding resistor is placed in the probe with the value of the

resistor corresponding to the actual wavelength of at least one of the LEDs. When the oximeter instrument is turned on, it first applies a current to the coding resistor and measures the voltage to determine the value of the resistor and thus the value of the wavelength of the LED in the probe.

5 [06] U.S. Patent 5,259,381 recognizes that the coded value of the wavelength of the red LED provided by a coding resistor may be inaccurate, since the actual wavelength can vary with temperature. Accordingly, this patent teaches including a temperature sensor in the oximeter probe to measure the actual temperature. With the actual temperature, and the coded wavelength value, a look-up table can be consulted to determine 10 the actual LED wavelength for that temperature.

15 [07] Another method of storing coded information regarding the characteristics of the LEDs is shown in U.S. Patent No. 4,942,877 assigned to Minolta. This patent discloses using an EPROM memory to store digital information, which can be provided in parallel or serially from the sensor probe to the remote oximeter. The memory is described as storing coefficients for the saturation equation, wavelength, subwavelength (where 2 peaks for LED), half-width of wavelength spectrum emitted by LED, intensity of LEDS or ratio, and on time of LEDS (written by the processor).

20 [08] Other examples of coding probe characteristics exist in other areas. Multiple calibration values are sometimes required, with this making the circuitry more complex or requiring many leads. In Patent No. 4,446,715, assigned to Camino Laboratories, Inc., a number of resistors are used to provide coded information regarding the characteristics of a pressure transducer. Patent No. 3,790,910 discloses another pressure transducer with a ROM storing characteristics of the individual transducer. Patent No. 4,303,984 shows another probe with digital characterization information stored in a PROM, which is read 25 serially using a shift register.

30 [09] Typically, the coding element is mounted in the probe itself. For instance, U.S. Patent No. 4,621,643 shows the coding resistor mounted in the probe element itself. In addition, U.S. Patent No. 5,246,003 shows the coding resistor being formed with a printed conductive material on the probe itself.

[10] In some devices, an electrical connector coupled by a cable to a device attached to a patient may include a coding element. For example, U.S. Patent No. 3,720,199 shows an intra-aortic balloon catheter with a connector between the catheter and a console. The connector includes a resistor with a value chosen to reflect the volumetric displacement of the particular balloon. U.S. Patent No. 4,684,245 discloses a fiberoptic catheter with a

module between the fiberoptic and electrical wires connected to a processor. The module converts the light signals into electrical signals, and includes a memory storing calibration signals so the module and catheter can be disconnected from the processor and used with a different processor without requiring a recalibration.

5 [11] Patent No. 5,645,059 teaches using a modulated signal to provide the coded data to a remote analyzer. Patent No. 5,429,129 shows using a voltage regulator to produce a specific voltage value in response to an attempt to read by the analyzer.

10 [12] Hewlett-Packard Company Patent No. 5,058,588 teaches an oximeter sensor with an encoding element that could be resistor, ROM, or customized integrated circuit. The encoding element encodes the type of sensor (in particular, type indicating area of placement on body - finger, ear, foot, arm; also, the type of sensor can indicate transmission/reflection type, or adult/neonate {indicating correction to be performed on theoretical oxygen saturation, allow switching between physiological limits such as minimum/maximum pulse rates for adults/neonates}; the maximum driving current may be 15 adapted according to type of sensor, and contact of sensor with tissue can be tested by means of an attenuation measurement if sensor type is known).

20 [13] Nellcor Patent No. 5,645,059, the disclosure of which is hereby incorporated herein by reference, teaches coding information in sensor memory used to provide pulse modulated signal, to indicate the type of sensor (finger, nose), the wavelength of a second LED, the number of LEDs, the numerical correction terms to the standard curves, and an identifier of the manufacturer.

25 [14] A number of catheter patents also discuss encoding information in the catheter. Sentron Patent No. 4,858,615 teaches encoding the type of sensor, type number, serial number, date of production, safe use life of the sensor, correction data for non-linearity, pressure sensitivity, offset, and temperature sensitivity.

30 [15] Interflo Medical Published PCT Application No. PCT/US92/08263, Publication No. WO 93/06776 teaches encoding patient specific data, size, manufacture date, batch number, sterilization date, expiration date, transducer number and type, manufacturer's name and address, thermistor heating element resistance, filament efficiency, program segments or patient historical data., format version for the calibration data, trademark information, catheter unique serial number, ship date, other date and time information, security code to identify manufacturer, thermal mass, filament composition, coefficient of resistance, layout byte, checksum, copyright, number of seconds since a certain date, patient

weight, patient height, timestamp of 1st CO data point, and a count of all CO data points in EEPROM.

[16] Duxel-Ohmeda of Boulder, Colorado markets an oximeter sensor product that encodes data into resistor values representing pointers to a lookup table 5 containing coefficients (as in U.S. Patent No. 4,700,708) as well as indicating a range of LED drive current to use with the sensor. The LEDs are driven with a higher or lower drive currents depending upon the value of the resistor in a particular sensor.

[17] Honeywell Patent No. 4,303,984 (expires 12-14-99) describes a 10 memory which stores characterization information, such as linearization information for a pressure sensor. Alnor Instrument Patent No. 5,162,725 describes storing both calibration and ID information in a sensor memory. Siemens Patent No. 5,016,198 describes a coding memory in a sensor with data for defining sensor's characteristic curve. McBean Patent No. 5,365,462 describes a date code in a sensor memory. Honeywell Patent No. 4,734,873 describes a pressure sensor with a PROM storing coefficients for a polynomial. Robert 15 Bosch Patent No. 4,845,649 describes a PROM in a sensor storing correcting data.

[18] McBean Patent No. 5,371,128 relates to EEPROM in sensor with 20 sensor type code and calibration data. McBean Patent No. 5,347,476 describes an accuracy code. Otax Patent No. 5,528,519 shows a PROM in a connector for oximeter.

[19] Square D Company Patent No. 5,070,732 shows calibration data in a 25 sensor memory. Baxter Patent No. 5,720,293 talks about different calibration information for a catheter, including a security code (encryption is discussed), serial number, model number, ID data such as calibration, manufacture, sterilization and ship date or other date and time information, a software program segment, security code for identifying whether sensor made by same manufacturer as monitor manufacturer, filament or transducer resistance, heat transfer coefficient, thermal mass, filament composition and coefficient of resistance, layout byte, copyright notice, checksum, random data bytes. Porsche Patent No. 5,008,843 describes a sensor with EEPROM ID and characteristics data.

BRIEF SUMMARY OF THE INVENTION

[20] The present invention provides a memory chip for use in an oximeter 30 sensor, or an associated adapter or connector circuit. The memory chip allows the storing of different data to provide enhanced capabilities for the oximeter sensor. In addition to providing unique data to store in such a memory, the invention describes unique uses of data stored in such a memory. The data stored in the memory chip includes information relating

to use of the oximeter sensor. For example, the memory chip may encode a sensor model identification that can be displayed on a display screen when the sensor is connected to an oximeter monitor. The memory may also encode a range of operating parameters such as light levels over which the sensor can function or a maximum drive current. The operating parameters are read by a controller circuit which uses the data read from the memory chip to control the functioning of the pulse oximetry system.

BRIEF DESCRIPTION OF THE DRAWINGS

[21] FIG. 1 is a block diagram of a pulse oximeter system in accordance 10 with the present invention.

DETAILED DESCRIPTION OF THE INVENTION

[22] FIG. 1 is a block diagram of a pulse oximeter system incorporating a calibration memory element 56 according to the invention. In one embodiment, memory 15 element 56 is a two-lead semiconductor digital memory chip. The calibration element is part of the sensor 50 which also includes red and infrared LEDs 52 as in the prior art, along with a detector 54. If desired, LEDs 52 may be replaced with other light emitting elements such as lasers.

[23] The oximeter includes read circuit 60, drive circuit 66, look-up tables 20 62 and 63, controller 64, amplifier 72, filter 74, and analog-to-digital converter 76. Read circuit 60 is provided for reading multiple coded values across the two leads 51, 53 connected to calibration element 56. One value is provided to a look-up table 62 to determine appropriate wavelength dependent coefficients for the oxygen saturation calculation, as in the prior art. The other value(s) are then provided to another look up table(s) 63 which provides 25 input (e.g., coefficients) to other calculations performed by controller 64. These additional calculations may enhance the performance and/or safety of the system. Controller 64 provides signals to a drive circuit 66, to control the amount of drive current provided to LEDs 52.

[24] As in the prior art, detector 54 is connected through an amplifier 72 30 and a filter 74 to an A/D converter 76. This forms a feedback path used by controller 64 to adjust the drive current to optimize the intensity range of the signal received. For proper operation the signal must be within the analog range of the circuits employed. The signal should also be well within the range of A/D converter 76 (e.g., one rule that may be applied is to adjust LED drives and amplifier gains so that both red and IR signals fall between 40%

and 80% of full scale reading of converter 76). This requires correct and independent settings for both the red and infrared LEDs. The current invention allows for another feedback path which may alter the LED settings based on other sensor characteristics contained in the coding of the calibration element 56, which is discussed in further detail below.

5 [25] Memory 56 may, for example, be implemented as a random access memory (RAM), a FLASH memory, a programmable read only memory (PROM), an electrically erasable PROM, a similar programmable and/or erasable memory, any kind of erasable memory, a write once memory, or other memory technologies capable of write operations. Various types of data useful to a pulse oximetry system can be stored in memory 10 56. For example, data indicating a sensor model identification code corresponding to a particular sensor model can be encoded in memory 56. Also, an action can be encoded into memory element 56 indicating an action to be performing by the oximeter monitor in response to reading the sensor model identification code.

15 [26] For example, an identification code in the form of text indicating the specific model of sensor can be digitally encoded into memory 56 and read by the oximeter monitor when the sensor is connected to the oximeter. An action indicating that the sensor model text is to be displayed by the oximeter monitor on a display screen can also be encoded in memory 56. The identification code can be displayed in human readable form on a display screen connected to the pulse oximeter monitor. The identification code allows the oximeter 20 instrument to display a text string indicating what sensor model is being used, e.g. "Nellcor OXISENSOR II D-25," "Adult Digit Sensor," or "Agilent N-25."

25 [27] Alternately, display text for a plurality of specific models of pulse oximeter sensors can be stored in a lookup table coupled in parallel with lookup tables 62 and 63 in the pulse oximeter monitor. The pulse oximeter monitor reads a sensor code from memory 56 when the sensor 50 is connected to the oximeter. The sensor identification code stored in memory 56 is used to locate display text stored in a lookup table that corresponds to a specific sensor model. The oximeter can display the display text for the specific sensor model on a display screen for viewing.

30 [28] The present invention eliminates the need for printing a model name and number on the sensor itself. Even when model names and numbers are printed on a sensor, the text may become illegible after several uses. Displaying text that corresponds to a specific sensor model can be highly useful for users of pulse oximetry sensors. For example, it may be important to identify a sensor model so that instructions relating to a particular sensor model in the manufacturer's handbook can be identified. In addition, it may be

necessavy to identify a sensor model name or identification number when corresponding with the manufacturer.

[29] Digitally encoded data indicating a sensor model type in memory 56 or in a lookup table may be used to determine whether a sensor model is compatible with a particular pulse oximeter monitor. For example, memory 56 may contain a code indicating a sensor model type that is read by controller 64. Memory 56 can also encode an action indicating that controller 64 is to compare the code from memory 56 with a list of codes in a lookup table (or other oximeter monitor memory device) to determine if the sensor is compatible. If controller 64 successfully matches the code read from the sensor, the display text indicating the sensor model type is displayed on the display screen. If controller 64 does not recognize the code, an error message may be displayed on the display screen indicating that the oximeter monitor does not recognize the sensor, and the oximeter may refuse to operate until the sensor is replaced.

[30] A code can be stored in the sensor memory 56 identifying the sensor manufacturer. An action indicating a use for the code by the oximeter can also be stored in memory 56. The code is read by controller 64 and is used for the purpose indicated by the action. The action may, for example, indicate that the code in memory 56 is to be used to indicate operability with oximeter monitors of other manufacturers. Controller 64 can recognize certain codes as indicating compatible oximeter sensors. If the oximeter monitor does not recognize the code, then controller 64 can display an error message on a display screen indicating that the sensor is not compatible, and/or controller 64 can shut down circuitry in the oximeter monitor that senses signals from the sensor until the sensor is replaced with a compatible sensor.

[31] Other information may also be encoded into memory 56, read by the oximeter monitor and displayed for user reference. For example, language codes or country codes can be stored in memory 56, read, and displayed to the user. The user can select a language or country code so that messages are displayed such as error messages are displayed in the selected language or a language corresponding to the selected country. Messages may also be encoded into memory 56. For example, safety messages relating to the proper use of the sensor can be encoded in memory 56 and displayed on a display screen in human-readable form.

[32] It is often desirable to upgrade the algorithms that are used by the oximeter to determine blood oxygen saturation levels, pulse rates, pulse amplitude, blood pressure, and other patient data as technology progresses and the operating parameters (such

as filter coefficients) are refined. Because oximeter sensors are typically much less expensive to replace than oximeter monitor instruments, it is desirable to encode data corresponding to the updated algorithms in the sensors rather than in the oximeter monitors.

[33] One method for performing these updates is by encoding revisions to 5 the algorithms used for calculating the patient parameters in memory within the oximeter monitor, while encoding updated software code or tuning coefficients in sensor memory 56. The updated code or coefficients correspond to updated algorithms that are read by the oximeter monitor so that the updated algorithms can be applied to the standard algorithms preprogrammed into the oximeter. For example, a line of software code in an algorithm used 10 by the oximeter monitor can be replaced by a updated line of code stored in memory 56.

[34] Controller 64 can read the updated code or coefficients from memory 56 and apply the updated algorithms to signals received from detector 54 to determine 15 accurate blood oxygen saturation levels, pulse rates, pulse amplitudes, perfusion data, blood pressure, and other patient data. The updated algorithms can also be used to allow only supported features to be used. In the preferred embodiment, once updated, the new code or 20 coefficients become permanently stored in the oximeter monitor, along with a new algorithm revision number, and are utilized for all future sensor use until later updated.

[35] Encoding a sensor model identification code could also be used to 25 accommodate sensor-specific operating parameters such as LED drive currents or “sensor off” characteristics (as an alternative to programming the value of drive current or “off” characteristics themselves). Under normal operating conditions, photosignals coming from the sensor LEDs generally fall within a certain range. When a sensor is removed from a patient, or falls off on it’s own, the photosignal usually changes. This is particularly true for the reusable clip-style sensor, because in their normal disconnected state, the LEDs shine directly onto the photodetector unimpeded by, for example, tissue! By programming a 30 “threshold photocurrent” into memory chip 56, reliable detection of a “sensor is off the patient” condition can be accomplished. In this example, exceeding a certain detected threshold light level is a sure sign the sensor isn’t on a finger or other opposed site.

[36] For certain other sensors, a low light level may be indicative of the 35 sensor being off. An adhesive sensor, for example, lays flat when in it’s natural state - little LED light may reach the detector. Encoding an expected range of light levels for the specific model of sensor being used into memory 56 allows enhanced detection of when the sensor is improperly placed or has been removed. When controller 64 senses that the light level output detector by photodetector 54 has fallen below or exceeded the expected range of light levels

encoded into memory 56, the oximeter monitor can display an “sensor off” message on a display screen indicating to the medical personnel that the sensor is not in an operable position and that valid data cannot be detected (i.e., a sensor off condition). The oximeter monitor can also emit an alarm signal until the light level detected by photodetector 54 reaches the expected range.

[37] If desired, expected ranges of light levels (or other parameters such as pulse size) that are specific to a particular patient may be encoded and saved into memory 56 by the clinical through the oximeter. The oximeter compares the expected range for the parameters encoded into memory 56 with data received from the photodetector to determine a sensor off condition each time the sensor is used until the range data is overwritten with new data. This is advantageous because light levels, pulse sizes, and other parameters detected by the photodetector can vary significantly from patient to patient.

[38] Existing pulse oximeter sensors determine whether a sensor is off the patient, or not in good contact, by using a number of metrics. Those metrics include pulse size, pulse variability, IR/Red correlation, light level variability, pulse shape, and pulse regularity. Not only the light level, but any of these other values could vary depending on the type of sensor, the characteristics of an individual patient, and the location on the body where the sensor is to be applied. Thus, sensor memory 56 could encode information about the expected variation in any of these metrics for the particular sensor type or model or a particular patient, for use in determining if a sensor is off from any combination of these or other metrics as an indication that the sensor is off the patient.

[39] For example, pulses could be typically weaker on the forehead compared to the finger. Memory device 56 of an oximeter sensor designed for use on the forehead of a patient can be encoded with a range of pulse sizes as well as a range of light levels that are expected from that particular oximeter sensor model. If desired, memory 56 can encode a range of numbers based upon light level and pulse size (and other parameters). For example, memory 56 can encode a range of numbers representing the expected range of pulse size times light level received from detector 54 for a specific sensor model.

[40] Controller 64 reads and decodes the pulse size, light level range, and other data encoded in memory 56. Controller 64 then compares the expected pulse size and light level range data with the information received from detector 54. When the pulse size and/or light level data received from detector 54 exceeds or falls below the expected range data encoded in memory 56, the oximeter monitor displays an output message, e.g., a warning of a poor signal, on the display screen indicating that the sensor is not operable or

emits an alarm signal. Further details of a Method and Circuit for Indicating Quality and Accuracy of Physiological Measurements are discussed in U.S. Patent Application No. 09/545,170, filed April 6, 2000 to Porges, et al., which is incorporated by reference herein in its entirety.

5 [41] Running LEDs 52 at a high drive current results in more light output from the LEDs, thus improving the signal-to-noise ratio of the blood oxygen saturation signal from detector 54, but comes at a cost of causing additional heat dissipation (i.e., the LEDs run “hotter”). As current flows through the sensor LEDs, the LED emits heat (i.e., the LED power = LED drive current times the voltage drop across the LED). The majority of the 10 energy output by the LEDs is dissipated as heat, and the smaller portion of the energy output by the LEDs is emitted as light. This heat typically causes the temperature of the skin under the sensor to rise by an amount that depends on the heat dissipation properties of the sensor. Current safety regulations and guidelines limit the temperature of the skin contacting portions of the sensor to remain at or below 41 °C. Sensors that do a poor job of directing the 15 heat away from the skin contacting surface, should use a lower LED drive current. Sensors with good thermal management can utilize higher drive currents without risk to the patient.

20 [42] Accordingly, by encoding the maximum safe LED drive current into the sensor itself, the oximeter instrument can utilize the highest safe drive current for the sensor being used to attain the greatest amount of LED light without risk of injury. The maximum safe drive current allowed to achieve a skin temperature at or below a maximum level can be determined in advance through testing for a given oximeter sensor model. That 25 maximum drive current can be encoded into memory 56 and read by controller 64 when the sensor is connected to the oximeter monitor. Controller 64 then communicates with drive circuit 66 to drive LEDs 52 at or near the maximum drive current value read from memory 56, but to prevent circuit 66 from driving LEDs 52 with a current that exceeds the maximum drive current.

30 [43] While the present invention has been described herein with reference to particular embodiments thereof, a latitude of modification, various changes and substitutions are intended in the foregoing disclosure, and it will be appreciated that in some instances some features of the invention will be employed without a corresponding use of other features without departing from the scope of the invention as set forth. Therefore, many modifications may be made to adapt a particular situation or material to the teachings of the invention without departing from the essential scope and spirit of the present invention. It is intended that the invention not be limited to the particular embodiments disclosed, but

that the invention will include all embodiments and equivalents falling within the scope of the claims.

WHAT IS CLAIMED IS:

1. An oximeter monitor system comprising:
a sensor; and a monitor,
said sensor including a memory device for storing data identifying an ID code associated with a sensor model,
said monitor reading said memory device, decoding said ID code, and selecting operating parameters based on said sensor model.
2. The oximeter monitor system of claim 1 wherein said operating parameters are a maximum safe drive current for a light emitter in said sensor.
3. The oximeter monitor system of claim 1 wherein said operating parameters are an expected range for signals received from a light detector in said sensor.
4. The oximeter monitor system of claim 3 wherein signals from said light detector being outside the expected range indicate a sensor-off condition.
5. An oximeter sensor comprising:
a light emitting element;
a light detecting element; and
a memory device for storing digital data, said digital data comprising an identification code that identifies a sensor model of said oximeter sensor and an action to be performed by an oximeter monitor coupled to the oximeter sensor in response to reading the identification code.
6. The oximeter sensor of claim 5 wherein said action comprises said oximeter monitor displaying the sensor model of said oximeter sensor by decoding said identification code.
7. An oximeter sensor comprising:
a light emitting element;
a light detecting element; and
a memory device for storing digital data, said digital data comprising a maximum safe drive current that can be applied to said light emitting element.

1 8. The oximeter sensor of claim 7 wherein an oximeter monitor coupled
2 to said oximeter sensor drives said light emitting element to said maximum safe drive current.

1 9. An oximeter sensor comprising:
2 a light emitting element;
3 a light detecting element; and
4 a memory device for storing digital data, said digital data comprising updated
5 code or coefficients for use by an oximeter monitor coupled to said oximeter sensor to
6 calculate data encoded in signals received from said light detecting element.

1 10. An oximeter sensor comprising:
2 a light emitting element;
3 a light detecting element; and
4 a memory device for storing digital data, said digital data comprising a
5 displayable message.

1 11. The oximeter sensor of claim 10 wherein the digital data further
2 comprises codes corresponding to a plurality of languages, wherein said message is
3 displayable in a selected language.

1 12. An oximeter sensor comprising:
2 a light emitting element;
3 a light detecting element; and
4 a memory device for storing digital data, said digital data comprising a code
5 identifying a manufacturer of said oximeter sensor and an action to be performed by an
6 oximeter monitor coupled to the oximeter sensor in response to reading the code.

1 13. A method for storing data in an oximeter sensor, the method
2 comprising:
3 emitting light from a light emitting element;
4 detecting light from the light emitting element using a photodetector; and
5 storing digitally encoded data in a memory in the sensor, the digitally encoded
6 data comprising an identification code that identifies a sensor model of said oximeter sensor
7 in said memory and an action to be performed by an oximeter monitor coupled to the
8 oximeter sensor in response to reading the identification code;

9 reading said identification code and said action using said oximeter monitor;
10 and
11 performing said action using said oximeter monitor.

1 14. The method of claim 13 wherein performing said action comprises:
2 displaying the sensor model of said oximeter sensor by decoding said
3 identification code.

1 15. A method for storing data in an oximeter sensor, the method
2 comprising:
3 emitting light from a light emitting element;
4 detecting light from the light emitting element using a photodetector; and
5 storing digitally encoded data in a memory in the sensor, the digitally encoded
6 data comprising a maximum safe drive current that can be applied to said light emitting
7 element in said memory.

1 16. The method of claim 15 further comprising:
2 driving said light emitting element to said maximum safe drive current.

1 17. A method for storing data in an oximeter sensor, the method
2 comprising:
3 emitting light from a light emitting element;
4 detecting light from the light emitting element using a photodetector; and
5 storing digitally encoded data in a memory in the sensor, the digitally encoded
6 data comprising a code identifying a manufacturer of said oximeter sensor in said memory
7 and an action to be performed by an oximeter monitor coupled to the oximeter sensor in
8 response to reading said code;
9 reading said identification code and said action using said oximeter monitor;
10 and
11 performing said action using said oximeter monitor.

1 18. The method of claim 17 wherein performing said action comprises:
2 displaying the manufacturer of said oximeter sensor by decoding said
3 identification code.

1 19. A method for storing data in an oximeter sensor, the method
2 comprising:
3 emitting light from a light emitting element;
4 detecting light from the light emitting element using a photodetector; and
5 storing digitally encoded data in a memory in the sensor, the digitally encoded
6 data comprising updated code or coefficients for use by an oximeter monitor coupled to said
7 oximeter sensor to calculate data from signals received from said photodetector.




FIG. 1

INTERNATIONAL SEARCH REPORT

Int'l Application No
PCT/US 01/27417A. CLASSIFICATION OF SUBJECT MATTER
IPC 7 A61B5/00

According to International Patent Classification (IPC) or to both national classification and IPC

B. FIELDS SEARCHED

Minimum documentation searched (classification system followed by classification symbols)

IPC 7 A61B

Documentation searched other than minimum documentation to the extent that such documents are included in the fields searched

Electronic data base consulted during the international search (name of data base and, where practical, search terms used)

EPO-Internal

C. DOCUMENTS CONSIDERED TO BE RELEVANT

Category °	Citation of document, with indication, where appropriate, of the relevant passages	Relevant to claim No.
X	US 6 044 283 A (FEIN ET AL.) 28 March 2000 (2000-03-28) column 2, line 49 -column 5, line 8 column 9, line 13 - line 31; figures 1-9 ----	1,2,5, 13-15,19
Y	US 5 443 801 A (LANGFORD) 22 August 1995 (1995-08-22) column 7, line 3 - line 23 column 23, line 53 -column 24, line 33; figures 1-18 ----	6,9,10, 12,17,18
Y	US 5 645 059 A (FEIN ET AL.) 8 July 1997 (1997-07-08) cited in the application the whole document ----	1,5,6,9, 12,13, 17,19 -/-

 Further documents are listed in the continuation of box C. Patent family members are listed in annex.

° Special categories of cited documents :

- "A" document defining the general state of the art which is not considered to be of particular relevance
- "E" earlier document but published on or after the international filing date
- "L" document which may throw doubts on priority claim(s) or which is cited to establish the publication date of another citation or other special reason (as specified)
- "O" document referring to an oral disclosure, use, exhibition or other means
- "P" document published prior to the international filing date but later than the priority date claimed

- "T" later document published after the international filing date or priority date and not in conflict with the application but cited to understand the principle or theory underlying the invention
- "X" document of particular relevance; the claimed invention cannot be considered novel or cannot be considered to involve an inventive step when the document is taken alone
- "Y" document of particular relevance; the claimed invention cannot be considered to involve an inventive step when the document is combined with one or more other such documents, such combination being obvious to a person skilled in the art.
- "&" document member of the same patent family

Date of the actual completion of the international search

Date of mailing of the international search report

31 January 2002

07/02/2002

Name and mailing address of the ISA
European Patent Office, P.B. 5818 Patentlaan 2
NL - 2280 HV Rijswijk
Tel. (+31-70) 340-2040, Tx. 31 651 epo nl,
Fax: (+31-70) 340-3016

Authorized officer

Hunt, B

INTERNATIONAL SEARCH REPORT

International Application No

PCT/US 01/27417

C.(Continuation) DOCUMENTS CONSIDERED TO BE RELEVANT

Category	Citation of document, with indication, where appropriate, of the relevant passages	Relevant to claim No.
Y	US 5 800 350 A (COPPLESON ET AL.) 1 September 1998 (1998-09-01) column 13, line 33 - line 54 column 14, line 49 - line 53; figures 1-18 ----	1,5,6,9, 12,13, 17,19
A	US 5 720 293 A (QUINN ET AL.) 24 February 1998 (1998-02-24) column 11, line 40 -column 13, line 31; figures 1-5 ----	1,5,7,9, 10,12, 13,15, 17,19
A	US 5 830 121 A (ENOMOTO ET AL.) 3 November 1998 (1998-11-03) the whole document ----	1,5,7,9, 10,12, 13,15, 17,19
X	US 5 987 343 A (KINAST) 16 November 1999 (1999-11-16) the whole document -----	1,5,9,12

INTERNATIONAL SEARCH REPORT

Information on patent family members

International Application No

PCT/US 01/27417

Patent document cited in search report	Publication date		Patent family member(s)	Publication date
US 6044283	A 28-03-2000	US	5779630 A	14-07-1998
		US	5645059 A	08-07-1997
		AU	690612 B2	30-04-1998
		AU	1211695 A	03-07-1995
		CA	2179022 A1	22-06-1995
		DE	69411947 D1	27-08-1998
		DE	69411947 T2	22-04-1999
		EP	0734221 A1	02-10-1996
		JP	10500026 T	06-01-1998
		WO	9516387 A1	22-06-1995
US 5443801	A 22-08-1995	US	5245845 A	21-09-1993
		US	5184633 A	09-02-1993
		AT	193664 T	15-06-2000
		AU	682870 B2	23-10-1997
		AU	6834794 A	12-12-1994
		CA	2162697 A1	24-11-1994
		DE	69424867 D1	13-07-2000
		DE	69424867 T2	15-02-2001
		EP	0700319 A1	13-03-1996
		JP	9501845 T	25-02-1997
		WO	9426432 A1	24-11-1994
		AU	2781492 A	03-05-1993
		CA	2120628 A1	15-04-1993
		EP	0642393 A1	15-03-1995
		JP	7500428 T	12-01-1995
		WO	9306948 A1	15-04-1993
		US	5207237 A	04-05-1993
US 5645059	A 08-07-1997	AU	690612 B2	30-04-1998
		AU	1211695 A	03-07-1995
		CA	2179022 A1	22-06-1995
		DE	69411947 D1	27-08-1998
		DE	69411947 T2	22-04-1999
		EP	0734221 A1	02-10-1996
		JP	10500026 T	06-01-1998
		US	6044283 A	28-03-2000
		WO	9516387 A1	22-06-1995
		US	5779630 A	14-07-1998
US 5800350	A 01-09-1998	AT	168872 T	15-08-1998
		AU	693112 B2	25-06-1998
		AU	7746194 A	25-05-1995
		CA	2134423 A1	02-05-1995
		CN	1106246 A	09-08-1995
		DE	69412030 D1	03-09-1998
		DE	69412030 T2	11-03-1999
		EP	0650694 A1	03-05-1995
		ES	2122175 T3	16-12-1998
		HK	1013943 A1	12-05-2000
		ZA	9408393 A	26-06-1995
US 5720293	A 24-02-1998	CA	2120532 A1	15-04-1993
		DE	69225803 D1	09-07-1998
		DE	69225803 T2	02-09-1999
		EP	0606356 A1	20-07-1994
		JP	6511172 T	15-12-1994
		WO	9306776 A1	15-04-1993

INTERNATIONAL SEARCH REPORT

Information on patent family members

Inte al Application No

PCT/US 01/27417

Patent document cited in search report	Publication date	Patent family member(s)		Publication date
US 5720293	A	US	5553622 A	10-09-1996
		US	5588438 A	31-12-1996
		US	5807269 A	15-09-1998
		US	5857976 A	12-01-1999
		US	5755670 A	26-05-1998
US 5830121	A	03-11-1998	JP 7171089 A	11-07-1995
			JP 7171090 A	11-07-1995
			US 5967969 A	19-10-1999
US 5987343	A	16-11-1999	NONE	

专利名称(译)	血氧计传感器与数字存储器编码传感器数据		
公开(公告)号	EP1315447A1	公开(公告)日	2003-06-04
申请号	EP2001968451	申请日	2001-08-31
[标]申请(专利权)人(译)	内尔科尔普里坦贝内特公司		
申请(专利权)人(译)	NELLCOR PURITAN BENNETT INCORPORATED		
当前申请(专利权)人(译)	NELLCOR PURITAN BENNETT INCORPORATED		
[标]发明人	FEIN MICHAEL E DI MANNHEIMER PAUL D MERCHANT ADNAN PORGES CHARLES SWEDLOW DAVID		
发明人	FEIN, MICHAEL, E. DI MANNHEIMER, PAUL, D. MERCHANT, ADNAN PORGES, CHARLES SWEDLOW, DAVID		
IPC分类号	A61B5/0245 A61B5/00 A61B5/022 A61B5/026 A61B5/145 A61B5/1455 G01D18/00		
CPC分类号	G01D18/00 A61B5/14552 A61B5/1495 A61B2562/085		
优先权	09/943805 2001-08-30 US 60/229616 2000-08-31 US		
其他公开文献	EP1315447B1		
外部链接	Espacenet		

摘要(译)

本发明提供一种用于血氧计传感器或相关适配器或连接器电路的存储器芯片。存储器芯片允许存储不同的数据以为血氧计传感器提供增强的能力。除了提供存储在这种存储器中的唯一数据之外，本发明还描述了存储在这种存储器中的数据的独特用途。存储在存储器芯片中的数据可以包括与血氧计传感器的使用有关的信息。例如，存储器芯片可以编码传感器模型标识，当传感器连接到血氧计监视器时，该传感器模型标识可以显示在显示屏上。存储器还可以编码一系列操作参数，例如传感器可以起作用的光水平或最大驱动电流。由控制器电路读取和解释操作参数以控制脉搏血氧测定系统。