

(11) **EP 1 384 433 A1** 

(12)

# **EUROPEAN PATENT APPLICATION**

(43) Date of publication:

28.01.2004 Bulletin 2004/05

(51) Int Cl.7: **A61B 5/00** 

(21) Application number: 03012263.4

(22) Date of filing: 11.06.2003

(84) Designated Contracting States:

AT BE BG CH CY CZ DE DK EE ES FI FR GB GR HU IE IT LI LU MC NL PT RO SE SI SK TR

Designated Extension States:

**AL LT LV MK** 

(71) Applicant: St. Jude Medical AB 175 84 Järfälla (SE)

(72) Inventor: Järverud, Karin 169 71 Solna (SE)


(30) Priority: 22.07.2002 SE 0202290

# (54) Apparatus for early detection of Ischemic Heart Disease

(57) A monitor for early detection of an ischemic heart disease of a patient comprises an impedance measurement means (1) including electrode means for measuring an intracardiac impedance and generating a corresponding impedance signal. A notch detector (3) is connected to the impedance measurement means for

detecting the occurrence of a notch in the impedance signal coincident with the entry of blood into the ventricle. A pattern recognition means (4) is provided to compare the measured post-notch impedance curve with a stored predetermined reference impedance curve template to detect an ischemic heart disease from the result of the comparison.





EP 1 384 433 A1

# Technical Field

**[0001]** The present invention relates to a monitor for early detection of an ischemic heart disease of a patient comprising an impedance measurement means including electrode means for measuring an intracardiac impedance and generating a corresponding impedance signal, and a notch detector connected to said impedance measurement means detecting the occurrence of a notch in said impedance signal coincident with the entry of blood into the ventricle.

1

# Background

[0002] A monitor of the above defined kind is described in WO 02/43587.

[0003] Congestive heart failure (CHF) is a fast growing health problem that mostly affects only adults. In this condition the heart is unable to pump enough blood to meet the need of the body's organs. Among the most common causes of CHF can be mentioned coronary artery disease, causing myocardial ischemia, myocardial infarction and cardiomyopathy. During ischemia the cardiac relaxation, i.e. diastolic, is changed or disturbed because the cardiac muscle is stiffened. A disturbed diastolic phase or diastolic failure is a very early kind of congestive heart failure, such that at this early state it might not even appear as symptoms to the patient. Detection of these early signs of disturbed relaxation patterns is described in the above mentioned WO 02/43587.

[0004] Early detection of ischemic heart disease (IHD) is required since that will give opportunities to prevent life threatening complications. By using a parameter such as diastolic intracardiac impedance IHD will be detected at a much earlier stage and myocardiac consequences such myocardial infarction (MI) and CHF may be prevented.

[0005] IHD prevents normal cardiac function because oxygen supply to the cardiac muscle is too low. In the acute stage of global cardiac ischemia (after 1-2s), relaxation failure occurs in the diastole (filling phase) as the myocardial tissue stiffens from lack of oxygen, followed by contraction failure (after 10s), increased filling pressures (after 20s), changes in ECG (after 25 s) and angina (30 s). Even if global ischemia is removed after a few seconds and oxygen supply is fully recovered, the disturbed relaxation may remain for hours or days. The diastole portion of the cardiac cycle seems to be the most sensitive parameter for IHD since it is effected first and residue effects remain for a long period after oxygen supply is restored. Cardiac tissue stiffness caused by IHD prevents adequate passive ventricular filling. The heart will try to compensate for this decreased filling by a largely increased atrial contraction. The atrial contribution to the ventricular filling in relation to passive filling

will thus increase and this is often seen as an inverted Early filling/Atrial contribution (E/A) quotient in echocardiographical measurements of mitral blood flow. Figure 1 shows schematically the early filling E and the atrial contribution A of mitral value blood flow for a normal healthy heart and an abnormal heart suffering from IHD. [0006] When IHD progresses it may lead to MI, CHF and/or the patient's death. In fact early detection of IHD can serve as an early marker for CHF risk factor.

[0007] Detection of the start of the diastolic phase by a "notch" detection is described in the above mentioned PCT/SE01/02615, cf. also Brian R. Pickett et. al., The American Journal of Cardiology, Vol. 71, May 1, 1993, " Usefulness of the Impedance Cardiogram to Reflect Left Ventricular Diastolic Function", which describes a study of the correlation between a dip in a non-invasively measured impedance and Doppler measurements for diastolic studies showing the appearance of a notch in the impedance signal corresponding to early diastole of a cardiac cycle.

**[0008]** The intracardiac impedance reflects the blood volume in the heart chambers, and the purpose of the present invention is to provide a monitor for early detection of an ischemic heart decease as described above by studying the measured intracardiac impedance curve during a selected portion of the cardiac cycle related to said notch.

#### Disclosure of the invention

**[0009]** This purpose is obtained by a monitor of the type defined in the introductory portion of the description and having the characterizing features of claim 1.

**[0010]** The above discussed increased stiffness of the heart tissue caused by IHD is reflected in the shape of the post notch portion of the measured impedance curve and this phenomena is thus utilized in the monitor according to the invention for early detection of a ischemic heart decease.

[0011] Generally the impedance curve has a plateau after the notch, which for IHD patients is followed by a dominating decrease coinciding with atrial contraction, indicating that the atria of these IHD patients are trying to compensate for the reduced passive ventricular filling. To detect this decrease, according to an advantageous embodiment of the monitor according to the invention, a differentiating means is provided to calculate the time derivative, dZ/dt, of the impedance curve, and the pattern of recognition means is adapted to compare the shape of the time derivative in the post notch portion of the impedance curve with a predetermined reference impedance derivative template.

**[0012]** According to other advantageous embodiments of the monitor according to the invention a calculation means is provided to calculate, for each cardiac cycle, an ischemic heart decease index, ZIHD-index, defined by

40

# ZIHD-index = $\Delta Z_1/\Delta Z_2$

where  $\Delta Z_1$  denotes, for each cardiac cycle, the difference between a maximum value of the measured impedance and the impedance value measured in a plateau occurring in the impedance curve after the notch, and  $\Delta Z_2$  the difference between the impedance value in the plateau and a minimum value of the measured impedance in an impedance decrease following the plateau, for detecting an ischemic heart disease from the calculated ZIHD-index. The ZIHD-index indicates the ratio between the passive filling and the contribution from the atrium to the ventricle filling. With the ZIHDindex the progression and degree of IHD can be monitored and the degree of cardiac compensation for the IHD can be calculated. The more the atrium is compensating for reduced passive ventricular filling the more advanced is the IHD. Atrial compensation is feasible only to a certain point and the more IHD progresses the more enlarged the atrium will become and the risk of atrial defibrillation AF is increased. By monitoring the ZI-HD-index development of AF problems can be avoided. Alternatively to the monitoring of the ZIHD-index a first comparison means can be provided to compare the absolute value  $|\Delta Z_1|$  of  $\Delta Z_1$  and/or the absolute value  $|\Delta Z_2|$ of  $\Delta Z_2$  with a predetermine  $\Delta Z_1$  threshold value and/or a predetermined  $\Delta Z_2$  threshold value respectively.  $\Delta Z_2$ is a measure of the function of the atrium and by monitoring  $\Delta Z_2$  it is possible to observe how the disease is advancing.

[0013] The post-notch plateau is not always a horizontal plateau and some patients have an overshoot after the notch. It is therefore advantageous, according to still another embodiment of the monitor according to the invention, to provide a first averaging means to determine an average value of the measured impedance curve in a predetermined time window after the notch for use as a plateau impedance value. The delay of said window after the notch is selectable and the length of the time window is preferably in the range of 100-150 msec.

[0014] According to yet other advantageous embodiments of the monitor according to the invention a loop creator is connected to the impedance measurement means and to the differentiating means to receive the impedance signal and the first time derivative dZ/dt of the impedance signal for plotting impedance values against related time derivative values to form a loop for each cardiac cycle, and a third comparison means is connected to the loop creator for comparing the loop with a predetermined loop template to detect an ischemic heart disease from the result of the comparison. The third comparison means is preferably adapted to compare the shape of the loop in that part of the loop, which corresponds to the post-notch portion of the impedance curve, with corresponding part of the loop template. By analyzing such a loop very clear indications can be obtained of deviations indicating an ischemic heart dis-

[0015] According to still other advantageous embodiments of the monitor according to the invention a second averaging means is connected to the impedance measuring means for determining an average impedance signal of impedance signals measured during a predetermined number of cardiac cycles, and the pattern recognition means is connected to the second averaging means to compare the average value of the post-notch impedance curves with the reference impedance curve template. A third averaging means can be connected to the differentiating means for determining an average time derivative of impedance signals measured during a predetermined number of cardiac cycles, and the pattern recognition means is connected to the third averaging means to compare the average time derivative of the post-notch portion of the measured impedance curves with corresponding part of the reference impedance derivative template. The loop creator is preferably connected to the second and third averaging means to receive the averaging impedance signal and average time derivative to form a corresponding average loop, and the third comparison means is adapted to compare the average loop with the predetermined loop template. By using average quantities in this way in the detection of an ischemic heart disease the reliability in the detection can be still improved.

#### Brief description of the drawings

[0016] To explain the invention in greater details an exemplifying embodiment of the monitor according to the invention with the described with reference to figure 5 on the drawings, on which figure 1 illustrates the ratio between the early passive filling and the atrial contribution to the mitral valve blood flow for a normal heart and an abnormal heart suffering from an ischemic heart disease, figure 2 shows results of impedance measurements formed in a humans heart, figure 3 shows average value and standard deviation for a plurality of impedance measurements, figure 4 illustrates use of the first time derivative of the impedance curve for detection of an ischemic heart disease, and figure 5 shows in the form of the block diagram an embodiment chosen as an example of the monitor according to the invention.

### Description of a preferred embodiment

[0017] Figure 2 shows results of right ventricular impedance measurements in four humans, viz. A) a 73 year old male, suffering from IHD and having a normal ejection fraction EF of 70% (patient number 14), B) a 67 year old male, suffering from IHD and having a reduced EF of 35% (patient number 6), C) a 24 year old male, having no IHD and reduced EF of 42% (patient number 20), and D) a healthy 27 year old female (patient number 19).

**[0018]** AC impedance measurements were performed using measuring catheters of type USCI Baid EP. Current was injected between a tip electrode positioned in right ventricular apex and a ring separated 10 mm from the tip. The resulting voltage response was measured between the same two electrodes.

**[0019]** In figure 2 the average value and the standard deviation for the measured impedance Z, the corresponding quantities for the first time derivative of the impedance curves dZ/dt, as well as a normalized ECG as a function of time.

**[0020]** These measurements can be used to discriminate IHD from non-IHD by using pattern recognition by comparing the post-notch impedance curve or its time derivative to "normal" impedance templates or "normal" first time derivative templates.

**[0021]** From figure 2 appears that for humans suffering from IHD the measured impedance curve Z exhibits, after the notch indicated in the figure by a vertical dashed bold line, a plateau followed by the steep decrease. The impedance curve Z for a non-IHD human has no such plateau and decrease, cf. diagrams C) and D) in figure 2.

**[0022]** In the corresponding dZ/dt curves IHD manifests itself as a distinct minimum in the post-notch region, encircled in diagram A) and B) in figure 2. Such a distinct minimum cannot be traced in the dZ/dt curves from non-IHD humans.

**[0023]** The measured Z and dZ/dt curves can be used for creating loops as described in the above mentioned PCT/SE01/02615 and IHD be detected by analyzing that part of the resulting loops which correspond to the post-notch portion of the impedance curves.

**[0024]** Figure 3 shows the average impedance curve and the standard deviation from measurements on the plurality of cardiac cycles for A) a IHD patient and B) a healthy patient.

**[0025]** It appears from figure 3 that patients having IHD, independent of EF status, exhibit a plateau after the notch in the impedance curve Z followed by a dominating, distinct decrease coinciding with atrial contraction. The position of the notch is marked in the diagrams of figure 3 by a dashed, vertical, bold line. As discussed above this decrease corresponds to the compensation by the atrium for the reduced passive filling of the ventricle of these patients.

**[0026]** The impedance curve Z for non-IHD patients may also have a plateau like portion after the impedance notch and atrial contraction may also be seen as a decrease in the impedance curve, but not as clear and dominating as for IHD patients.

**[0027]** Discriminating IHD patients from non-IHD patients can be accomplished by calculating a ZIHD-index =  $\Delta Z_1/\Delta Z_2$ .

[0028] As appears from figure 3

$$\Delta Z_1 = Z_{\text{max}} - Z_{\text{plateau}}$$

where  $Z_{max}$  denotes the maximum value of the impedance curve in the notch region and  $Z_{plateau}$  the impedance value of the post-notch plateau, and

$$\Delta Z_2 = Z_{\text{plateu}} - Z_{\text{min}}$$

where  $Z_{\text{min}}$  denotes minimum value in the decrease following the plateau.

**[0029]** This index ZIHD is smaller for IHD patients, diagram A), than for non-IHD patients, diagram B).

**[0030]** The maximum value  $Z_{max}$  appears in figure 3 before the notch, however sometimes a maximum value of the impedance after the notch can be used. For some patients it can be suitable to use as maximum value  $Z_{max}$  an average value of a maximum value occurring before the notch and a maximum value occurring after the notch, cf. figure 2A).

[0031] The plateau value Z<sub>plateau</sub> can preferably be determined as an average value of the measured impedance curve in a predetermined time window after the notch. The length of this window can suitably be in the range of 100-150 msec and the location of the time window can be delayed more or less after the notch depending on the appearance of the impedance curve in the vicinity of the notch (occurrence of possible overshoots, etc.). The time from the occurrence of the notch to the steep decrease in the measured Z curve is depending on the heart rate, the length of the plateau decreasing with increasing heart rate. This must also be taken into consideration when determining the length of the time window.

[0032] Figure 4 shows the dZ/dt curves and the normalized ECG:s of figure 2. From this figure it appears that by using a suitably selected discriminating threshold level, indicated by a horisontal dashed bold line after the notch in the dZ/dt-curves, the large down-slope coinciding with atrial activity for IHD patients can be correctly detected. Thus for humans A) and B) the dZ/dt curves decrease below the marked threshold value indicating IHD, whereas for humans C) and D) the dZ/dt curves never reach this threshold, indicating non-IHD.

**[0033]** Figure 5 shows a preferred embodiment of the monitor according to the invention in the form of a block diagram. The measured impedance Z, block 1, is supplied to a notch detector 3, of the kind described in PCT/SE01/02615. The notch detector 3 detects notch existence and timing in the measured impedance curve. The pattern recogniser 4 detects the shape of the post-notch impedance curve and compares it with a stored predetermined reference or normal impedance curve template.

**[0034]** A differentiating means is provided to calculate the first time derivative dZ/dt, block 2, which is also supplied to the pattern recogniser 4 for comparison of the shape of the dZ/dt curve in the post-notch portion with a predetermined reference impedance derivative template.

20

25

35

40

45

[0035] dZ/dt is also supplied to a dZ/dt marker detector 5 for detecting the large negative dZ/dt-values of IHD patients coinciding with atrial activity after detection of a notch as described above, eg. by comparing dZ/dt with a predetermined threshold value.

[0036] The ZIHD-index calculator 6 is calculating the ZIHD-index =  $\Delta Z_1/\Delta Z_2$  for use for distinguishing IHD patients from non-IHD patients as described above. With the aid of the ZIHD-index the IHD progression or degree can be monitored. Progressing IHD results in increasing atrial compensation for reduced passive ventricular filling. However, this compensation is only possible to a certain point. With still progressing IHD the atrium will become enlarged and the risk of atrial defibrillation is obvious. By monitoring the ZIHD-index future atrial defibrillation problems can be avoided.

[0037] The embodiment shown in figure 5 also comprises a monitor and diagnostics unit 7 for collecting, analyzing and storing measuring data. Sometimes the notch does not appear clearly and might be difficult to detect. Information from previous measurements on the patient in question can in this case be used for locating the decrease in the impedance curve corresponding to atrial activity.

#### **Claims**

- 1. A monitor for early detection of an ischemic heart disease of a patient comprising an impedance measurement means (1) including electrode means for measuring an intracardiac impedance and generating a corresponding impedance signal, and a notch detector (3) connected to said impedance measurement means for detecting the occurrence of a notch in said impedance signal coincident with the entry of blood into the ventricle, characterized in that a pattern recognition means (4) is provided to compare the measured post-notch impedance curve with a stored predetermined reference impedance curve template to detect an ischemic heart disease from the result of the comparison.
- The monitor according to claim 1, characterized in that said pattern recognition means (4) is adapted to compare the shape of the post-notch portion of the impedance curve with said impedance curve template.
- 3. The monitor according to claims 1 or 2, **characterized in that** a differentiating means (2) is provided to calculate the time derivative, dZ/dt, of the impedance curve, and **in that** said pattern recognition means (4, 5) is adapted to compare the shape of said time derivative in the post-notch portion of the impedance curve with a predetermined reference impedance derivative template.

4. The monitor according to any of the preceding claims, characterized in that a calculation means is provided to calculate, for each cardiac cycle, an ischemic heart disease index, ZIHD-index, defined by


# ZIHD-index = $\Delta Z_1/\Delta Z_2$


where  $\Delta Z_1$  denotes, for each cardiac cycle, the difference between a maximum value of the measured impedance and the impedance value measured in a plateau occurring in the impedance curve after said notch, and  $\Delta Z_2$  the difference between the impedance value in said plateau and a minimum value of the measured impedance in an impedance decrease following said plateau, for detecting an ischemic heart disease from the calculated ZIHD-index.


- 5. The monitor according to the claim 4, characterized in that a first comparison means is provided to compare the absolute value  $|\Delta Z_1|$  of  $\Delta Z_1$  and/or the absolute value  $|\Delta Z_2|$  of  $\Delta Z_2$  with a predetermined  $\Delta Z_1$  threshold value and/or a predetermined  $\Delta Z_2$  threshold value respectively.
- 6. The monitor according to claims 4 or 5, characterized in that said maximum value of the measured impedance is a maximum value located before the notch in the cardiac cycle.
- 7. The monitor according to the claims 4 or 5, characterized in that said maximum value of the measured impedance is a maximum value located after the notch in the cardiac cycle.
- 8. The monitor according to the claims 4 or 5, characterized in that said maximum value of the measured impedance is an average value of a maximum value located before the notch in the cardiac cycle and a maximum value located after the notch.
- 9. The monitor according to any of the claims 4 8, characterized in that a first averaging means is provided to determine an average value of the measured impedance curve in a predetermined time window after the notch for use as a plateau impedance value.
- 10. The monitor according to the claim 9, characterized in that the delay of time window after the notch is selectable.
- 11. The monitor according to the claims 9 or 10, characterized in that the length of said time window is in the range of 100-150 msec.


- 12. The monitor according to any of the claims 3-11, characterized in that said differentiating means (2) is adapted to determine the slope of said impedance decrease following the plateau in the impedance curve from said time derivative, dZ/dt, of the impedance signal, and in that a second comparison means (5) is provided to compare said slope with a predetermined slope threshold level.
- 13. The monitor according to any of the claims 3 12, characterized in that a loop creator is connected to said impedance measurement means (1) and to said differentiating means (2) to receive said impedance signal and said first time derivative dZ/dt of the impedance signal for plotting impedance values against related time derivative values to form a loop for each cardiac cycle, and in that a third comparison means is connected to said loop creator for comparing said loop with a predetermined loop template to detect an ischemic heart disease from the result of the comparison.
- 14. The monitor according to claim 13, characterized in that said third comparison means is adapted to compare the shape of the loop in that part of the loop, which corresponds to the post-notch portion of the impedance curve, with corresponding part of said loop template.
- 15. The monitor according to any of the preceding claims, **characterized in that** a second averaging means is connected to said impedance measuring means (1) for determining an average impedance signal of impedance signals measured during a predetermined number of cardiac cycles, and **in that** said pattern recognition means (4) is connected to said second averaging means to compare the average value of the post-notch impedance curves with said reference impedance curve template.
- 16. The monitor according to any of the claims 3 15, characterized in that a third averaging means is connected to said differentiating means (2) for determining an average time derivative of impedance signals measured during a predetermined number of cardiac cycles, and in that said pattern recognition means is connected to said third averaging means to compare said average time derivative (4, 5) of the post-notch portion of the measured impedance curves with corresponding part of said reference impedance derivative template.
- 17. The monitor according to claims 15 and 16, characterized in that said loop creator is connected to said second and third averaging means to receive said average impedance signal and said average time derivative to form a corresponding average loop, and in that said third comparison means is


- adapted to compare said average loop with said predetermined loop template.
- 18. The monitor according to any of the claims 5 17, characterized in that said first, second and third comparison means are realised by one and the same comparison means.
- 19. The monitor according to any of the claims 9-18, characterized in that said first, second and third averaging means are realised by one and the same averaging means.
- 20. The monitor according to any of the preceding claims, characterized in that said electrode means comprises a bipolar ventricular electrode and in that said impedance measuring means (1) is adapted to measure the impedance between electrode tip and ring.
- 21. The monitor according to any of the claims 1 19, said monitor being implantable, characterized in that said electrode means comprises an unipolar ventricular electrode and in that said impedance measuring means (1) is adapted to measure the impedance between electrode and monitor casing.
- 22. A heart stimulator comprising a pulse generator for generating stimulation pulses for delivery to a patient's heart by means of electrode means adapted for location in one or both of the lower chambers of the heart, **characterized by** a monitor according to any one of the preceding claims.
























# **EUROPEAN SEARCH REPORT**

Application Number EP 03 01 2263

	DOCUMENTS CONSID				
Category	Citation of document with in of relevant passa	ndication, where appropriate, ges		levant claim	CLASSIFICATION OF THE APPLICATION (Int.CI.7)
Y,D	MEDICAL AB (SE) 6 * page 2, paragraph	RVERUD KARIN ;ST JUDE June 2002 (2002-06-06) 1 1 * 1 6 - page 4, paragraph			A61B5/00
Y,D	IMPEDANCE CARDIOGRAVENTRICULAR DIASTOL	IC FUNCTION" F CARDIOLOGY, CAHNERS FTON, MA,, US, FTON, MA,, US, FTON, Pages FTON, Pages FTON, PAGES	1,2		
A	KUBICEK W G ET AL: IMPEDANCE CARDIOGRA APPLICATIONS" BIOMEDICAL ENGINEER PRESS, LONDON,, GB, vol. 9, no. 9, Sept pages 410-416, XPOO ISSN: 0006-2898 * whole document *	RING, UNITED TRADE tember 1974 (1974-09),	1-2	1	TECHNICAL FIELDS SEARCHED (Int.CI.7) A61N A61B
A	EP 0 607 511 B (SIE 27 July 1994 (1994- * whole document *		1-2	1	
A	US 5 282 840 A (HUD 1 February 1994 (19 * whole document *		1-2	1	
A	US 2002/002389 A1 ( 3 January 2002 (200 * whole document *	BRADLEY KERRY ET AL) 2-01-03)	1-2	1	
	The present search report has t	peen drawn up for all claims			
	Place of search	Date of completion of the search	<del>'</del>		Examiner
	MUNICH	22 October 2003		Edw	ard, V
X : parti Y : parti docu A : techi O : non-	TEGORY OF CITED DOCUMENTS cularly relevant if taken alone cularly relevant if combined with another ment of the same category nological background written disclosure mediate document	L : document cited	ocument, ate I in the app for other	but publis olication reasons	hed on, or

EPO FORM 1503 03.82 (P04C01)

# ANNEX TO THE EUROPEAN SEARCH REPORT ON EUROPEAN PATENT APPLICATION NO.

EP 03 01 2263

This annex lists the patent family members relating to the patent documents cited in the above-mentioned European search report. The members are as contained in the European Patent Office EDP file on The European Patent Office is in no way liable for these particulars which are merely given for the purpose of information.

22-10-2003

Patent do cited in sea		Publication date		Patent fam member(s		Publication date
WO 024358	7 A	06-06-2002	EP WO	1339320 0243587		03-09-2003 06-06-2002
EP 060751	1 B	27-07-1994	DE DE EP JP US	69330820 69330820 0607511 7000367 5427112	T2 A2 A	31-10-2001 25-04-2002 27-07-1994 06-01-1995 27-06-1995
US 528284	Э А	01-02-1994	AU CA DE DE EP JP WO	3797893 2102906 69302875 69302875 0586664 6508289 9318821	A1 D1 T2 A1 T	21-10-1993 27-09-1993 04-07-1996 16-01-1997 16-03-1994 22-09-1994 30-09-1993
US 2002002	2389 A1	03-01-2002	AU WO	6310801 0187410		26-11-2001 22-11-2001
			•• <i>•</i>			

FORM P0459

For more details about this annex : see Official Journal of the European Patent Office, No. 12/82



专利名称(译)	用于早期检测缺血性心脏病的装置						
公开(公告)号	EP1384433A1	公开(公告)日	2004-01-28				
申请号	EP2003012263	申请日	2003-06-11				
申请(专利权)人(译)	ST.犹达医疗用品AB						
当前申请(专利权)人(译)	ST.犹达医疗用品AB						
[标]发明人	JARVERUD KARIN						
发明人	JÄRVERUD, KARIN						
IPC分类号	A61B5/053 A61N1/365 A61B5/00						
CPC分类号	A61N1/36521 A61B5/0535 A61B5/	7239 Y10S128/92 Y10S128/92	3				
优先权	0202290 2002-07-22 SE						
其他公开文献	EP1384433B1						
外部链接	Espacenet						

# 摘要(译)

用于早期检测患者的缺血性心脏病的监测器包括阻抗测量装置(1),其包括用于测量心内阻抗并产生相应的阻抗信号的电极装置。陷波检测器(3)连接到阻抗测量装置,用于检测阻抗信号中的陷波的发生,该陷波与血液进入心室的情况一致。提供模式识别装置(4)以将测量的切口后阻抗曲线与存储的预定参考阻抗曲线模板进行比较,以从比较结果中检测缺血性心脏病。



