

(19)



(11)

**EP 1 799 099 B1**

(12)

**EUROPEAN PATENT SPECIFICATION**

(45) Date of publication and mention of the grant of the patent:

**09.05.2018 Bulletin 2018/19**

(51) Int Cl.:

**A61B 5/00 (2006.01)**

(21) Application number: **05803610.4**

(86) International application number:

**PCT/US2005/034999**

(22) Date of filing: **27.09.2005**

(87) International publication number:

**WO 2006/039394 (13.04.2006 Gazette 2006/15)**

**(54) PATIENT VISUAL INSTRUCTION TECHNIQUES FOR SYNCHRONIZING BREATHING WITH A MEDICAL PROCEDURE**

VISUELLE INSTRUKTIONSTECHNIKEN FÜR PATIENTEN ZUR SYNCHRONISIERUNG DER ATMUNG MIT EINEM MEDIZINISCHEN VERFAHREN

TECHNIQUES D'INSTRUCTIONS VISUELLES DE PATIENT PERMETTANT DE SYNCHRONISER LA RESPIRATION AVEC UNE PROCÉDURE MEDICALE

(84) Designated Contracting States:  
**AT BE BG CH CY CZ DE DK EE ES FI FR GB GR HU IE IS IT LI LT LU LV MC NL PL PT RO SE SI SK TR**

(72) Inventor: **MOSTAFAVI, Hassan**  
**Los Altos, CA 94024 (US)**

(30) Priority: **30.09.2004 US 957009**

(74) Representative: **Foster, Mark Charles et al**  
**Mathisen & Macara LLP**  
**Communications House**  
**South Street**  
**Staines-upon-Thames**  
**Middlesex, TW18 4PR (GB)**

(43) Date of publication of application:  
**27.06.2007 Bulletin 2007/26**

(73) Proprietors:  
• **Varian Medical Systems, Inc.**  
**Palo Alto, CA 94304 (US)**  
• **MOSTAFAVI, Hassan**  
**Los Altos, CA 94024 (US)**

(56) References cited:  
**WO-A-03/003796**      **WO-A1-2004/112903**  
**JP-A- 2000 325 339**      **JP-A- 2002 282 325**  
**US-A1- 2003 188 757**      **US-A1- 2004 082 853**

**EP 1 799 099 B1**

Note: Within nine months of the publication of the mention of the grant of the European patent in the European Patent Bulletin, any person may give notice to the European Patent Office of opposition to that patent, in accordance with the Implementing Regulations. Notice of opposition shall not be deemed to have been filed until the opposition fee has been paid. (Art. 99(1) European Patent Convention).

## Description

### Field of the Invention

[0001] This invention relates generally to systems and methods for prompting patient, and more specifically, to systems and methods for prompting patient to control patient movement.

### Background of the Invention

[0002] Computed tomography is an imaging technique that has been widely used in the medical field. In a procedure for computed tomography, an x-ray source and a detector apparatus are positioned on opposite sides of a portion of a patient under examination. The x-ray source generates and directs a x-ray beam towards the patient, while the detector apparatus measures the x-ray absorption at a plurality of transmission paths defined by the x-ray beam during the process. The detector apparatus produces a voltage proportional to the intensity of incident x-rays, and the voltage is read and digitized for subsequent processing in a computer. By taking a plurality of readings from multiple angles around the patient, relatively massive amounts of data are thus accumulated. The accumulated data are then analyzed and processed for reconstruction of a matrix (visual or otherwise), which constitutes a depiction of a density function of a volume of the bodily region being examined. By considering one or more sections in the volume, a skilled diagnostician can often diagnose various bodily ailments such as tumors, blood clots, etc.

[0003] Computed tomography has found its principal application to examination of bodily structures or the like which are in a relatively stationary condition. However, currently available computed tomographic apparatus may not be able to generate tomographic images with sufficient quality or accuracy due to physiological movement of a patient. For example, beating of a human heart and breathing have been known to cause degradation of quality in CT images. U.S. Patent No. 3,952,201 issued to Hounsfield, describes a system that compensates for a blurring of a radiograph due to heart motions by collecting image data of a patient at different angles while monitoring the patient's heart. The image data and the heart motion data are then correlated in a computer, and image data that was obtained when the heart's motion exceeded a threshold level is rejected. However, such method does not generate desirable CT images for a portion of a cardiac cycle when the heart's motion exceeds the threshold level.

[0004] Degradation of quality of CT images due to patient's breathing is more difficult to address than that associated with heart motion. Patients' breathing poses a unique problem to CT imaging that is different from heart motion. This is because the pattern and the period of a patient's breathing cycle is generally less consistent when compared to those of the patient's cardiac cycle.

As such, while a particular phase of a cardiac cycle may be predicted with sufficient accuracy, a particular phase of a breathing cycle may not be as easily predicted or determined.

5 [0005] Furthermore, there has been an increased desire to visualize organ motion by viewing a sequence of CT images as a movie sequence. However, collecting a large quantity of CT image data sufficient for forming a video while considering breathing motion may take a much longer time. This may cause a patient who is confined within a gantry opening to feel uncomfortable and subject the patient to excessive radiation.

10 [0006] To reduce durations of CT image acquisition procedures, patient prompting techniques have been used to control patients' breathing during CT procedures. Such techniques involve sending an audio signal to instruct a patient to inhale, exhale, or hold breath. However, successful implementation of such techniques requires the audio signal be provided to the patient in a timely manner. If the signal is provided too early, the patient may perform a required Physiological movement earlier than expected. On the other hand, if the signal is provided too late, the patient may not be able to perform the required physiological movement in time to meet the procedure's requirement. Also, different patients may have different reaction times - i.e., some patients may react faster in response to the audio signal, while others may react slower in response to the audio signal. As such, existing techniques for prompting a patient are difficult to implement, and may not provide satisfactory results due to different patients' reaction time.

20 [0007] For the foregoing, it would be desirable to have an improved method and system for prompting a patient in a medical procedure.

25 [0008] WO 03/003796 discloses a patient visual prompt that informs a patient a desired breathing level. The visual prompt allows a patient to see his/her own breathing level, but does not indicate the future time at which the desired breathing level is desired to be accomplished, nor does the visual prompt provide an indication to the patient of when such future time will arrive.

30 [0009] JP 2000 325339 discloses a voice control circuit for controlling a voice generating circuit. The voice generating circuit is configured to store a voice signal for a message for guidance in stopping of breathing, and generate a voice signal according to a control signal of the voice control circuit.

35 [0010] US 2003 188757 A1 discloses a respiration control apparatus having an inhalation circuit and an exhalation circuit for controlling a breathing of a patient.

### SUMMARY OF THE INVENTION

40 [0011] The present invention provides a method of prompting a patient, a computer program product and a system as defined in the independent claims. Optional features of the invention are defined in the dependent claims.

**[0012]** In accordance with some embodiments of the invention, a method of prompting a patient includes informing a patient a relationship between a result of an activity being performed by the patient, and a first target result to be achieved by the activity.

**[0013]** In accordance with other embodiments of the invention, a computer program product that includes a medium is provided. The medium includes a set of instructions, an execution of which causes a process to be performed, the process comprising informing a patient a relationship between a result of an activity being performed by the patient, and a first target result to be achieved by the activity.

**[0014]** In accordance with other embodiments of the invention, a system for prompting a patient includes means for informing a patient a relationship between a result of an activity being performed by the patient, and a first target result to be achieved by the activity.

**[0015]** In accordance with other embodiments of the invention, a user interface for prompting a patient includes a screen displaying graphics for informing a patient a relationship between a result of an activity being performed by the patient, and a first target result to be achieved by the activity.

**[0016]** In accordance with other embodiments of the invention, a method of prompting a patient includes informing a patient a relationship between a time and a target result to be achieved by an activity.

**[0017]** In accordance with other embodiments of the invention, a computer program product that includes a medium is provided. The medium includes a set of instructions, an execution of which causes a process to be performed, the process comprising informing a patient a relationship between a time and a target result to be achieved by an activity.

**[0018]** In accordance with other embodiments of the invention, a system for prompting a patient includes means for informing a patient a relationship between a time and a target result to be achieved by an activity.

**[0019]** In accordance with other embodiments of the invention, a user interface for prompting a patient includes a screen displaying graphics for informing a patient a relationship between a time and a target result to be achieved by an activity.

**[0020]** In accordance with other embodiments of the invention, a method of prompting a patient includes informing a patient a relationship between a position of a portion of the patient and a first target position for the portion.

**[0021]** In accordance with other embodiments of the invention, a computer program product that includes a medium is provided. The medium includes a set of instructions, an execution of which causes a process to be performed, the process comprising informing a patient a relationship between a position of a portion of the patient and a first target position for the portion.

**[0022]** In accordance with other embodiments of the invention, a system for prompting a patient includes

means for informing a patient a relationship between a position of a portion of the patient and a first target position for the portion.

**[0023]** In accordance with other embodiments of the invention, a user interface for prompting a patient includes a screen displaying graphics for informing a patient a relationship between a position of a portion of the patient and a first target position for the portion.

**[0024]** Other aspects and features of the invention will be evident from reading the following detailed description of the preferred embodiments, which are intended to illustrate, not limit, the invention.

## BRIEF DESCRIPTION OF THE DRAWINGS

**[0025]** The drawings illustrate the design and utility of preferred embodiments of the present invention, in which similar elements are referred to by common reference numerals. In order to better appreciate how advantages and objects of the present invention are obtained, a more particular description of the present invention briefly described above will be rendered by reference to specific embodiments thereof, which are illustrated in the accompanying drawings. Understanding that these drawings depict only typical embodiments of the invention and are not therefore to be considered limiting of its scope, the invention will be described and explained with additional specificity and detail through the use of the accompanying drawings in which:

**FIG. 1** illustrates a computed tomography system in which embodiments of the invention can be implemented;

**FIG. 2A-2C** illustrates examples of breathing waveforms relative to gantry rotational angles;

**FIG. 3A-3C** illustrate a user interface in accordance with some embodiments of the invention;

**FIG. 4** illustrates a user interface having a curvilinear bar in accordance with other embodiments of the invention;

**FIG. 5** illustrates a user interface in accordance with other embodiments of the invention, showing the interface displaying graphics that represent prescribed inhale and exhale levels of a breathing;

**FIG. 6** illustrates a user interface in accordance with other embodiments of the invention, showing the interface displaying a curve that represents a target breathing motion;

**FIG. 7** illustrates a user interface in accordance with other embodiments of the invention; and

**FIG. 8** is a block diagram that illustrates an embodiment of a computer system upon which embodiments of the invention may be implemented.

## DETAILED DESCRIPTION OF THE PREFERRED EMBODIMENTS

**[0026]** Various embodiments of the present invention

are described hereinafter with reference to the figures. It should be noted that the figures are not drawn to scale and elements of similar structures or functions are represented by like reference numerals throughout the figures. It should also be noted that the figures are only intended to facilitate the description of specific embodiments of the invention. They are not intended as an exhaustive description of the invention or as a limitation on the scope of the invention. Also, an aspect described in conjunction with a particular embodiment of the present invention is not necessarily limited to that embodiment and can be practiced in any other embodiments of the present invention. In addition, a means for performing a function may be used to perform another function, and a function may be performed by one or more means.

#### Computed Tomography Image Acquisition System

**[0027]** Referring now to the drawings, in which similar or corresponding parts are identified with the same reference numeral, **FIG. 1** illustrates a computed tomography image acquisition system 10, in which embodiments of the present invention can be employed. The system 10 includes a gantry 12 having an opening (or bore) 13, a patient support 14 for supporting a patient 16, and a control system 18 for controlling an operation of the gantry 12. In other embodiments, instead of the ring configuration shown, the gantry 12 can have other configurations. For example, the gantry 12 can have a C-arm configuration, such as that used in the cone beam CT machines manufactured by Varian Medical Systems, Inc., in Palo Alto, California. The system 10 also includes an x-ray source 20 that projects a beam (which can be a cone beam, a fan beam, or the like) of x-rays towards a detector 24 on an opposite side of the gantry 12 while the patient 16 is positioned at least partially between the x-ray source 20 and the detector 24. The detector 24 has a plurality of sensor elements configured for sensing a x-ray that passes through the patient 16. Each sensor element generates an electrical signal representative of an intensity of the x-ray beam as it passes through the patient 16.

**[0028]** In the illustrated embodiment, the control system 18 includes a processor 54, such as a computer processor, coupled to a gantry rotation control 40 and a patient prompting device 100. The control system 18 may also include a monitor 56 for displaying data and an input device 58, such as a keyboard or a mouse, for inputting data. During a scan to acquire x-ray projection data (i.e., CT image data), the gantry 12 rotates about the patient 16. The rotation of the gantry 12 and the operation of the x-ray source 20 are controlled by the gantry rotation control 40, which provides power and timing signals to the x-ray source 20 and controls a rotational speed and position of the gantry 12 based on signals received from the processor 54. Although the control 40 is shown as a separate component from the gantry 12 and the processor 54, in alternative embodiments, the control 40 can

be a part of the gantry 12 or the processor 54. The processor 54 is configured to send prompting signals to the patient prompting device 100 in a prescribed manner (e.g., in synchronization with a rotation of the gantry 12).

**[0029]** The patient prompting device 100 is configured to provide visual signals to the patient 16 during a procedure, thereby instructing the patient 16 to perform certain task(s). The patient prompting device 100 includes a screen 101, an image source 104, and a structure 106 to which the screen 101 and the image source 104 are coupled. The structure 106 has a low profile, thereby allowing the screen 101 be placed within the opening 13 of the gantry. The screen 101 is preferably made from a non-metallic material and does not include circuitry for preventing interference with a radiation field. In the illustrated embodiments, the screen 101 has a mirror surface, and the image source 104 includes a flat panel screen (or a monitor screen). During use, the image source 104 receives image data from the processor 54 and displays graphics in response thereto. The graphics is reflected by the mirror surface, and the patient 16 can see the reflected graphics by looking towards the mirror surface 102. In other embodiments, the screen 101 can include a non-mirror (e.g., a non-reflective) surface. In such cases, instead of the image source 104 being a flat panel or a screen, the image source 104 includes an image projector that projects image onto the surface. Also, in other embodiments, the image source 104 can include fiber optics for transmitting image signals to a viewing surface. In such case, the screen 101 can be a component of a glasses or goggles, with the viewing surface being an inside face of the glasses or goggles. Other types of image source can also be used in alternative embodiments. For example, in further embodiments, the screen 101 can itself be a LCD screen (an image source) provided that its electronics do not significantly interfere with a radiation field. In such cases, the patient prompting device 100 does not include the image source 104. The above described embodiments of the patient prompting device 100 has been described in U.S. Patent Application entitled, "Patient Multimedia Display", filed concurrently with this application. Other display mechanism can also be used as long as it provides a viewing surface from which the patient 16 can receive image signals. In the illustrated embodiments, the image/graphics as shown in the screen 101 provides visual signal to control the patient's breathing (e.g., by instructing the patient 16 to hold breath, to inhale, and/or to exhale) while the gantry 12 rotates around the patient 16 to collect image data, thereby ensuring that CT image data that correspond to a prescribed phase of a breathing cycle are obtained. However, in other embodiments, the image/graphics as shown in the screen 101 can be configured to instruct the patient 16 to perform other task(s).

**[0030]** The system 10 also includes a position monitoring system 200 for monitoring a position of a portion of the patient 16, and producing a position signal in response thereto. The position signal is transmitted to the

processor 54, which causes the image source 104 to generate a visual indicator representing a position of the patient portion. In the illustrated embodiments, the position monitoring system 200 includes a marker block 202 placed on the patient 16, and an optical device 204 coupled to the processor 54. The optical device 204 may be a camera or other imaging devices, and is configured to sense an image of the marker block 202. The marker block 202 preferably comprises a reflective or retro-reflective material that can reflect light, whether in the visible or invisible wavelengths. The marker block 202 has a rectangular shape with multiple retro-reflective elements 203 located on its surface. Alternatively, the marker block 202 can have a different shape, such as a hemispherical shape, or a disk shape, as long as the size, spacing, and position of the reference locations are configured such that the optical device 204 can view and generate an image that accurately shows the positioning of the marker block 202.

**[0031]** When using the position monitoring system 200, one or more marker block 202 is placed on or secured to the patient 16, and the optical device 204 is used to sense the marker block(s) 202. The optical device 204 produces a set of image coordinates for the marker elements on the marker block(s) 202. The position and distance of the marker elements located on the marker block(s) 202 is known relative to other marker elements on the same respective marker block(s) 202. By comparing the position and distance between the marker elements on a recorded image frame with a reference position and image stored for the position monitoring system 200, the absolute position and orientation of the marker block(s) 202 can be determined with a high degree of accuracy. This, in turn, provides an accurate position and orientation estimation for the patient position or the patient body position upon which the marker block(s) 202 is attached. Such and similar physiological gating systems have been described in U.S. Patent No. 6,621,889, and U.S. Patent Application Serial No. 09/893,122, filed June 26, 2001.

**[0032]** It should be noted that other types of patient position monitoring system can also be used. For example, in alternative embodiments, instead of using an optical system, another device, such as a spirometer, strain-gauge, a laser sensor, a pressure sensor, or any of other devices known in the art of position or activity state monitoring, can be used to sense and/or measure a patient movement or an activity state during a procedure.

#### Patient Prompting Interface

**[0033]** During a CT image acquisition procedure, as the radiation source 20 is rotated about the patient 16 to generate CT image data at a plurality of gantry rotational angles, the patient prompting system 100 is used to guide the patient 16 to control his/her breathing such that one or more breathing state of the patient 16 can be synchronized with the rotation of the gantry 12, thereby allowing

desired image quality to be obtained at prescribed gantry angles. Particularly, the processor 54 is configured (e.g., programmed) to cause the image source 104 to provide visual signals for prompting the patient 16 during a procedure based on signals/data received from the optical device 204 and a rotational position of the gantry 12 (or the radiation source 20).

**[0034]** FIGS. 2A-2C illustrate examples of the patient's breathing waveforms as the gantry 12 rotates around the patient's 16 to collect image data. As the gantry 12 rotates 360° around the patient 16 in a first direction during a first rotation, the patient 16 undergoes a plurality of breathing cycles (the amplitude of which is represented by curve 360). When the patient's inhale level satisfies a prescribed criteria (e.g., lies between a minimum prescribed inhale level and a maximum prescribed inhale level, represented by lines 350, 352, respectively), the processor 54 then causes the radiation source 20 to activate to deliver a radiation beam and generate CT image data. Similarly, when the patient's inhale level falls outside the boundaries 350, 352, the processor 54 then causes the radiation source 20 to deactivate and cease generating CT image data. Alternatively, the radiation source 20 can continue to transmit radiation at a plurality of gantry rotational angles even when the patient's inhale level falls outside the prescribed range. In such cases, the processor 54 maintains a record of the CT image data that are collected, and chooses only desired CT image data (e.g., CT image data that correspond to a desired breathing level) for reconstruction of tomography image(s).

**[0035]** As shown in FIG. 2A, during a first gantry rotation, CT image data are collected at gantry angles that are between 50°-120°, 210°-275°, and 325°-360° (represented by clear portions 372a, 372b, 372c, respectively, of a bar 370), with CT image data desired to be collected at remaining gantry angles represented by hatched portions 374a, 374b, 374c of the bar 370.

**[0036]** During a next gantry rotation, the gantry 12 rotates (e.g., in an opposite direction), and additional CT image data are collected at gantry angles that correspond to the patient's inhale level satisfying the prescribed criteria (e.g., inhale level lying between prescribed levels 350, 352). As shown in FIG. 2B, during the second gantry rotation, with the aid of the patient prompting device 100, the patient 16 attempts to cause additional CT image data be collected by trying to match his/her breathing with the portions 374a-374c. As a result, additional CT image data are collected at gantry angles at which image data (or desirable image data) were not previously obtained in the first gantry rotation. In the illustrated example, CT image data are collected at gantry angles that are between 0°-50°, 140°-210°, and 260°-330°, corresponding to portions 376a, 376b, 376c shown in the graph, with CT image data desired to be collected at remaining gantry angles that are represented by hatched portions 378a, 378b, 378c.

**[0037]** In some embodiments, it may be desirable to

have CT image data collected from one gantry rotation overlap CT image data collected from another gantry rotation. For example, although image data have been collected between gantry rotational angles 210°-275° (corresponding to portion 372b) during the first gantry rotation, and image data have been collected between gantry rotational angles 140°-210° (corresponding to portion 376b) during the second gantry rotation, it may be desirable to have additional image data collected at or adjacent the seam between the two portions 372b, 376b. As such, in the illustrated example, hatched portion 378a is shown to correspond to a portion of a gantry rotation at which overlapping image data are desired. Similar is true for portion 378a and part of portion 378b. In alternative embodiments, CT image data collected from one gantry rotation is not required to overlap CT image data collected from another gantry rotation.

**[0038]** During a next gantry rotation, the gantry 12 rotates (e.g., in an opposite direction from that of the last rotation), and additional CT image data are collected at gantry angles that correspond to the patient's inhale level satisfying the prescribed criteria. As shown in **FIG. 2C**, during the third gantry rotation, with the aid of the patient prompting device 100, the patient 16 attempts to cause additional CT image data be collected by trying to match his/her breathing with the portions 378a-378c. As a result, additional CT image data are collected at gantry angles at which image data (or desirable image data) were not obtained in the first and the second gantry rotations. In the illustrated example, CT image data are collected at portions 380a, 380b, 380c of the gantry rotation. When all desired image data have been collected, the CT image data acquisition procedure is then terminated. Although the CT image data acquisition procedure has been described with reference to the gantry 12 making three gantry rotations in the illustrated example, it should be understood that a CT image data acquisition procedure may require more or less than three gantry rotations.

**[0039]** As mentioned previously, the patient prompting device 100 assists or guides the patient 16 in controlling his/her breathing during a CT image data acquisition procedure. **FIG. 3A-3C** illustrate a user interface 300 for prompting the patient 16 in accordance with some embodiments of the invention. The user interface 300 includes a first indicator (displayed object) 302 for indicating a state/result of a patient activity, and a second indicator (displayed object) 304 for indicating a target state/target result of the patient activity. The first and the second indicators 302, 304 are displayed in the screen 101 of the patient prompting device 100. The first indicator 302 includes a bar, the position of which relative to the screen 101 represents a breathing level of the patient 16. In the illustrated embodiments, the optical device 204 generates an image of the marker block 202, and transmits image signals to the processor 54. The processor 54 analyzes the image signals to determine the position and/or orientation of the marker block 202, and output a

signal to cause the image source 104 to display the first indicator 302 at a position in the screen 101 that is indicative of, or associated with, the determined position and/or orientation of the marker block 202. As such, as the patient 16 breaths, the marker block 202 moves in response to the patient's breathing, and the processor 54 in turn causes the first indicator 302 to move in correspondence with the patient's breathing. In the illustrated embodiments, when the patient 16 inhales, the first indicator 302 moves in a first direction 312, and when the patient 16 exhales, the first indicator 302 moves in a second direction 314.

**[0040]** As shown in **FIG. 3A**, the second indicator 304 includes a bar 306 having a first line 308, and a second line 310. The first line 308 represents a prescribed minimum inhale level of a breathing, and the second line 310 represents a prescribed maximum inhale level of a breathing. The bar 306 also has a length 318, which represents at least a portion of a range of gantry angles at which CT image data is desired to be collected. In some embodiments, the length of the bar 306 represents an entire range of gantry angles at which CT image data are desired to be collected. Alternatively, the bar 306 as displayed in the screen 101 represents different portion(s) of an entire range of gantry angles at which CT image data are desired to be collected as the bar 306 moves across the screen 101. During use, the processor 54 receives an instantaneous gantry angle read-out, and generate the bar 306 (or at least portion(s) thereof) based on the instantaneous gantry angle. The bar 306 moves in a direction indicated by arrow 311, and in synchronization with a rotation of the gantry 12 (or the radiation source 20). The patient 16 is instructed (either before or during a radiation procedure) to move the first indicator 302 in between the first and the second lines 308, 310 by inhaling, and keep the first indicator 302 in between the first and the second lines 308, 310 by breath-holding.

**[0041]** When the first indicator 302 has been positioned by the patient 16 in between the first and the second lines 308, 310, the processor 54 sends a "ON" signal to activate the radiation source 20, thereby creating CT image data. The radiation source 20 is activated at a plurality of gantry rotational angles as the gantry 12 is rotated about the patient 16 as long as the first indicator 302 is between the first and the second lines 308, 310 (representing a desired breathing state for collecting CT image data). When the patient 16 no longer holds his/her breath within the prescribed range inhale levels, the patient 16 will exhale, causing the first indicator 302 to move out of the prescribed range of inhale levels. In such cases, the processor 54 sends a "OFF" signal to deactivate the radiation source 20 to cease transmitting radiation. Such feature is advantageous because it minimizes an amount of radiation delivered to the patient 16. However, in alternative embodiments, the radiation source 20 can continue to transmit radiation at a plurality of gantry rotational angles even when the first indicator 302 has moved out of the prescribed range of inhale levels. In such cases,

the processor 54 maintains a record of the CT image data that are collected, and chooses only the desired image data (e.g., image data that correspond to desired breathing level) for reconstruction of tomography image(s).

**[0042]** As **FIG. 3B** shows, after the patient 16 has moved the first indicator 302 in between the first and the second lines 308, 310, a portion 320 of the bar 306 where the first indicator 302 intercepts the bar 306 changes color (or appears to be removed if the changed color is similar to a background color). The portion 320 represents part of the prescribed range of gantry angles at which CT image data has been collected, while a remaining portion(s) 322 of the bar 306 represents the part(s) of the prescribed range of gantry angles at which CT image data are still desired to be collected. Since the bar 306 is moving in the direction 311 in synchronization with the gantry 12 rotation, more portion of the bar 306 will change color (or the portion 320 will grow) as long as the patient 16 controls his/her breathing to maintain the first indicator 302 between the first and the second lines 308, 310.

**[0043]** When the patient 16 exhales, the first indicator 302 moves out of the location between the first and the second lines 308, 310 in response thereto, and the bar 306 ceases changing. However, the bar 306 continues to move in the direction 311 in correspondence with a rotation of the gantry 12. The patient 16 can repeat the step of inhaling to move the first indicator 302 between the first and the second lines 308, 310 again to cause another portion of the bar 306 to change color (or to be removed) - thereby causing CT image data to be collected at additional gantry rotational angles. For example, during a CT image acquisition procedure, the patient 16 can inhale and exhale for a plurality of times, resulting in a plurality of portions 320a, 320b of the bar 306 changing color (or being removed), with the remaining portions 322a, 322b of the bar 306 representing the parts of the prescribed range of gantry angles at which CT image data are still desired to be collected (**FIG. 3C**). In some embodiments, the processor 54 is configured to make adjustment to the bar 306 such that additional (or overlapping) CT image data can be collected (e.g., at a range of gantry angles that are adjacent to that represented by the interface between the portion 322a and the portion 320b). For example, the processor 54 can be configured to change at least a part of a removed portion 320 back to a hatched portion 322.

**[0044]** In the illustrated embodiments, the uncovered gantry angle intervals resulting from the patient's breathing being non-compliance defines the configuration of the bar 306 (e.g., spacing and length of the remaining portions (or indicator boxes) 322) for the next gantry rotation. As remaining portion(s) of the bar 306 approach the first indicator 302 a second time (corresponding to a second gantry rotation), the patient 16 can repeat the breathing step(s) in an attempt to cause additional CT image data be collected at remaining gantry angles at

which CT image data were not previously collected. The gantry 12 continues to make additional rotation, and the user interface 300 continues to display remaining portions of the bar 306 (e.g., based on an instantaneous gantry angle read-out), until the patient 16 has caused the entire bar 306 to change color (or the entire bar 306 to be removed), representing the condition that CT image data at all prescribed gantry rotational angles have been collected.

**[0045]** Providing the first indicator 302 is advantageous because it informs the patient 16 a relationship between a result of an activity being performed by the patient, and a first target result desired to be achieved by the activity, thereby allowing the patient 16 to gauge himself/herself while performing the activity. For example, the patient 16 can breath harder or less based on a position of the first indicator 302 as observed by the patient 16. In addition, providing the second indicator 304 is also advantageous because it informs the patient 16 a relationship between a time and a target result desired to be achieved by an activity, thereby allowing the patient 16 to decide when to perform a certain task based on his/her own awareness of a target result desired to be accomplished. The combination of these two features allows a patient to control his/her breathing such that CT image data at prescribed range(s) of gantry angles can be collected efficiently.

**[0046]** It should be noted that the user interface 300 should not be limited by the example described previously, and that the user interface 300 can have other configurations in alternative embodiments. **FIG. 4** illustrates a variation of the user interface 300. In such cases, instead of having a straight profile, the bar 306 has a profile that resembles a portion of a circle. The arc angle of the bar 306 corresponds to a prescribed range of gantry angles at which CT image data are desired to be collected. During use, the bar (in the form of an arc) rotates about a center 390 in a direction and speed that corresponds to a rotation of the gantry 12 (or the radiation source 20), and the patient 16 is instructed (either before or during a radiation procedure) to move the first indicator 302 in the first direction 312 (e.g., by inhaling) such that the first indicator 302 intercepts the bar 306, as similarly described previously. Initially, the bar 306 has a first color (represented by hatch in the example). As the patient 16 successfully causes the first indicator 302 to intercept the bar 306, portion(s) 320 of the bar 306 changes color (or is removed), indicating to the patient 16 that the CT image data at the corresponding gantry angles have been collected. The patient 16 continues to control his/her breathing to cause the first indicator 302 to intercept remaining portion(s) 322 as the bar 306 rotates about the center 390 (either continuously in one direction, or in a back-and-forth manner), until CT image data at all prescribed gantry angles have been collected. In alternative embodiments, instead of the bar 306 being an arc, the bar 306 can be a complete circle.

**[0047]** In the above embodiments, the user interface

300 is configured to guide the patient's breathing by prescribing a range of inhale levels. In alternative embodiments, the user interface 300 can be configured to guide the patient's breathing by prescribing a range of exhale levels. In such cases, the user interface 300 includes lines that represent a minimum exhale level and a maximum exhale level (that are similar to the minimum and maximum inhale levels 308, 310 discussed previously).

**[0048]** In further embodiments, the user interface 300 can be configured to guide a patient's breathing by prescribing both ranges of inhale and exhale levels. **FIG. 5** illustrates a user interface 400 for prompting a patient in accordance with other embodiments of the invention. The user interface 400 includes a first indicator (displayed object) 402 for indicating a state/result of a patient activity, and a plurality of indicators (displayed object) 404 representing target state/ target result of the patient activity. The indicators 402, 404 are displayed in the screen 101 of the patient prompting device 100. The first indicator 402 includes a bar, the position of which relative to the screen 101 represents a breathing level of the patient 16. In the illustrated embodiments, the optical device 204 generates an image of the marker block 202, and transmits image signals to the processor 54. The processor 54 analyzes the image signals to determine the position and/or orientation of the marker block 202, and output a signal to cause the image source 104 to display the first indicator 302 at a position in the screen 101 that is indicative of, or associated with, the determined position and/or orientation of the marker block 202. As such, as the patient 16 breaths, the marker block 202 moves in response to the patient's breathing, and the processor 54 in turn causes the first indicator 302 to move in correspondence with the patient's breathing.

**[0049]** As shown in **FIG. 5**, the first target 404a represents a prescribed inhale level desired to be accomplished by the patient 16. The target 404a has bottom and top sides that correspond to a minimum inhale level 420 and a maximum inhale level 422, respectively. Similar is true for the target 404c. The second target 404b represents a prescribed exhale level desired to be accomplished by the patient 16. The target 404b has top and bottom sides that correspond to a minimum exhale level 424 and a maximum exhale level 426, respectively. In some embodiments, the distances (e.g., distances, 430, 432) between successive targets 404 can be adjusted, depending on a particular need of a procedure. Also, in other embodiments, a length of a target can be adjusted, e.g., made longer to indicate that a breath-hold is desired. In some embodiments, the prescribed inhale level, the prescribed exhale level, and the distances 430, 432 are user specific, and can be determined during a training session. Such will make the patient more comfortable because the interface is prescribing a manner of breathing in which the patient is accustomed to performing.

**[0050]** During use of the interface 400, the targets 404 move in a direction indicated by arrow 411, and in syn-

chronization with a rotation of the gantry 12 (or the radiation source 20). The patient 16 is instructed (either before or during a radiation procedure) to move the first indicator 402 to intercept the targets 404, e.g., one after the other, by inhaling and exhaling, as the targets 404 moves across the screen 101, thereby controlling substantially all phases of the patient's breathing cycle. Such result is particularly desirable in the case in which sets of image data are desired to be collected for different prescribed phases of a breathing cycle, which requires patient breathing period to be as constant as possible and the phase of the periodic breathing motion to be synchronized with the gantry angle at any given time. In the illustrated embodiments, the processor 54 is configured to activate the radiation source 20 to generate image data when the first indicator 402 intercepts a target (one of the targets 404), and deactivate the radiation source 20 when the first indicator 402 misses a target. In some embodiments, a physician can prescribe a number N of phases into which a breathing cycle is divided. In such cases, the processor 54 is configured to provide appropriate visual signals via the patient prompting interface 400, and to provide timing signals to the gantry control 40 such that sets of image data for each of the N prescribed phases of a breathing cycle can be collected.

**[0051]** In other embodiments, instead of collecting CT image data for each of the N phases of a breathing cycle, CT image data are collected only for a desired phase range of a breathing cycle. For example, a breathing cycle may be divided into  $N = 3$  intervals, but CT image data may be desired to be collected for only the first 1/3 of a breathing cycle. In such cases, the patient prompting interface 400 can accordingly be configured to provide appropriate visual signals to the patient 16 such that the patient 16 can control his/her breathing at the right time to enable all desired CT image data be collected for the first 1/3 of the breathing cycle. Moreover, in some cases, fewer rotations can be used by foregoing having data for the entire 360° coverage for the phase(s) of interest and instead having data for such phase(s) from a sufficient number of angles to produce an image of sufficient quality for the purpose at hand.

**[0052]** If CT image data for all prescribed phases of a breathing cycle have been collected, then the CT image data collection process is terminated, and CT images are reconstructed using the collected CT image data. The reconstruction of CT images may be performed using a technique or method known in the art. In one embodiment, the processor 54 sorts the CT image data according to the phase segment of the respiratory cycle at which it was acquired, and synchronizes the collected CT image data such that data corresponding to a given phase of a respiratory cycle are combined to reconstruct an image for that phase. When CT images for all phases of a respiratory cycle have been reconstructed, the CT images may be displayed in a sequence to form a video.

**[0053]** It should be noted that the configuration of the user interface 400 is not limited by the examples dis-

cussed previously, and that the user interface 400 can have other configurations in alternative embodiments. For example, instead of targets 404 that are in the form of a block, in alternative embodiments, the user interface 400 can display a curve 450 that represents desired breathing waveforms to be accomplished (FIG. 6). In such cases, the curve 450 is configured to move across the screen 101 in the direction 411, and the patient 16 is instructed (either before or during a radiation procedure) to move the first indicator 402, e.g., by inhaling and exhaling, such that the first indicator 402 stays on the curve 450 as much as possible, as the curve 450 moves across the screen 101. In the illustrated example, the curve 450 has a shape that resembles a sine wave, but can have other shapes, such as a customized shape, in other embodiments. For example, the shape of the curve 450 can be specific to each patient and can be obtained in a training session. In the illustrated embodiments, the processor 54 sends a "ON" signal to activate the radiation source 20 to emit radiation for generating CT image data when the first indicator 402 intercepts the curve 450, and a "OFF" signal to deactivate the radiation source 20 when the first indicator 402 does not intercept the curve 450. In some embodiments, after image data at a desired gantry rotational angle have been collected for a prescribed phase of a breathing cycle, the processor 504 can change a color of (or remove) the corresponding portion of the curve 450 to indicate that image data have been collected for the corresponding gantry angle. The user interface 400 continues to display the curve 450 across the screen 101, and the patient 16 continues to operate the first indicator 402 by breathing, until CT image data for all prescribed gantry rotational angles and for all prescribed phases of a breathing cycle are collected.

**[0054] FIG. 7** shows a variation of the user interface 400, which includes a curve 500 having a first side 502 and a second side 506. In such cases, the curve 500 is configured to move across the screen 101 in the direction 411, and the patient 16 is instructed (either before or during a radiation procedure) to move the first indicator 402, e.g., by inhaling and exhaling, such that the first indicator 402 stays in between the first and the second sides 502, 506 as the curve 500 moves across the screen 101. The sides/boundaries 502, 506 are separated by a distance 504, which represents a range of prescribed breathing levels at which image data will be collected. The distance 504 can be uniform along a length of the curve 500, or alternatively, can vary along a length of the curve 500. In the illustrated example, the curve 500 has a shape that resembles a sine wave, but can have other shapes, such as a customized shape, in other embodiments. In the illustrated embodiments, the processor 54 sends a "ON" signal to activate the radiation source 20 to emit radiation for generating CT image data when the first indicator 402 is within the boundaries 502, 506 of the curve 500, and a "OFF" signal to deactivate the radiation source 20 when the first indicator 402 is outside the boundaries 502, 506. In some embodiments, after

CT image data at a desired gantry rotational angle have been collected for a prescribed phase of a breathing cycle, the processor 504 can change a color of (or remove) the corresponding portion of the curve 500 to indicate that image data have been collected. The user interface 400 continues to display the curve 500 across the screen 101, and the patient 16 continues to operate the first indicator 402 by breathing, until CT image data for all prescribed gantry rotational angles and for all prescribed phases of a breathing cycle are collected.

**[0055]** Although several examples of a patient prompting interface have been described, it should be noted that the scope of the invention should not be so limited. In alternative embodiments, the user interfaces 300, 400 can have other configurations for informing a patient a relationship between a result of an activity being performed by the patient, and a first target result desired to be achieved by the activity and/or a relationship between a time and the target result desired to be achieved by an activity. For example, in alternative embodiments, instead of a bar (e.g., the bar 306), the processor 54 can be configured to display an object having a different form for prompting the patient 16. As used in this specification, the term "object" can include any item, such as a graph, a text, a number, a message, a symbol, a line, a bar, an object having a geometric or customized shape, etc. In other embodiments, instead of the object being displayed, an "object" can be one or more lights emitted by an optical device. Also, instead of each indicator being an object, in other embodiments, two or more indicators can be represented by a single object, or alternatively, an indicator can be represented by two or more objects. In addition, in other embodiments, the processor 54 can be configured to cause CT image data be collected when a patient's inhale (or exhale) level has reached a minimum prescribed level, regardless of how much the patient's inhale (or exhale) level exceed the minimum prescribed level. In such cases, the user interface does not have an indicator for representing a maximum inhale (or exhale) level.

**[0056]** In further embodiments, the patient prompting device 100 can further include a speaker. In such cases, the processor 54 can be configured to cause the speaker to emit audio signal(s) for prompting the patient 16. Also, in any of the embodiments described herein, the processor 54 can be configured to provide a visual signal or an audio signal when a desired task has been performed by the patient 16. For example, the visual signal can be a score, and the audio signal can be a game sound, thereby providing a game interface for prompting the patient 16, which may make the image acquisition procedure more fun and engaging for the patient 16.

#### Computer System Architecture

**[0057] FIG. 8** is a block diagram that illustrates an embodiment of a computer system 800 upon which an embodiment of the invention may be implemented. Compu-

ter system 800 includes a bus 802 or other communication mechanism for communicating information, and a processor 804 coupled with the bus 802 for processing information. The processor 804 may be an example of the processor 54, or alternatively, an example of a component of the processor 54, of FIG. 1. The computer system 800 also includes a main memory 806, such as a random access memory (RAM) or other dynamic storage device, coupled to the bus 802 for storing information and instructions to be executed by the processor 804. The main memory 806 also may be used for storing temporary variables or other intermediate information during execution of instructions to be executed by the processor 804. The computer system 800 further includes a read only memory (ROM) 808 or other static storage device coupled to the bus 802 for storing static information and instructions for the processor 804. A data storage device 810, such as a magnetic disk or optical disk, is provided and coupled to the bus 802 for storing information and instructions.

**[0058]** The computer system 800 may be coupled via the bus 802 to a display 87, such as a cathode ray tube (CRT), for displaying information to a user. An input device 814, including alphanumeric and other keys, is coupled to the bus 802 for communicating information and command selections to processor 804. Another type of user input device is cursor control 816, such as a mouse, a trackball, or cursor direction keys for communicating direction information and command selections to processor 804 and for controlling cursor movement on display 87. This input device typically has two degrees of freedom in two axes, a first axis (e.g., x) and a second axis (e.g., y), that allows the device to specify positions in a plane.

**[0059]** Embodiments of the invention are related to the use of computer system 800 for generating visual signals for prompting a patient. According to one embodiment of the invention, such use is provided by computer system 800 in response to processor 804 executing one or more sequences of one or more instructions contained in the main memory 806. Such instructions may be read into the main memory 806 from another computer-readable medium, such as storage device 810. Execution of the sequences of instructions contained in the main memory 806 causes the processor 804 to perform the process steps described herein. One or more processors in a multi-processing arrangement may also be employed to execute the sequences of instructions contained in the main memory 806. In alternative embodiments, hard-wired circuitry may be used in place of or in combination with software instructions to implement the invention. Thus, embodiments of the invention are not limited to any specific combination of hardware circuitry and software.

**[0060]** The term "computer-readable medium" as used herein refers to any medium that participates in providing instructions to the processor 804 for execution. Such a medium may take many forms, including but not limited to, non-volatile media, volatile media, and transmission media. Non-volatile media includes, for example, optical

or magnetic disks, such as the storage device 810. Volatile media includes dynamic memory, such as the main memory 806. Transmission media includes coaxial cables, copper wire and fiber optics, including the wires that comprise the bus 802. Transmission media can also take the form of acoustic or light waves, such as those generated during radio wave and infrared data communications.

**[0061]** Common forms of computer-readable media include, for example, a floppy disk, a flexible disk, hard disk, magnetic tape, or any other magnetic medium, a CD-ROM, any other optical medium, punch cards, paper tape, any other physical medium with patterns of holes, a RAM, a PROM, and EPROM, a FLASH-EPROM, any other memory chip or cartridge, a carrier wave as described hereinafter, or any other medium from which a computer can read.

**[0062]** Various forms of computer-readable media may be involved in carrying one or more sequences of one or more instructions to the processor 804 for execution. For example, the instructions may initially be carried on a magnetic disk of a remote computer. The remote computer can load the instructions into its dynamic memory and send the instructions over a telephone line using a modem. A modem local to the computer system 800 can receive the data on the telephone line and use an infrared transmitter to convert the data to an infrared signal. An infrared detector coupled to the bus 802 can receive the data carried in the infrared signal and place the data on the bus 802. The bus 802 carries the data to the main memory 806, from which the processor 804 retrieves and executes the instructions. The instructions received by the main memory 806 may optionally be stored on the storage device 810 either before or after execution by the processor 804.

**[0063]** The computer system 800 also includes a communication interface 818 coupled to the bus 802. The communication interface 818 provides a two-way data communication coupling to a network link 820 that is connected to a local network 822. For example, the communication interface 818 may be an integrated services digital network (ISDN) card or a modem to provide a data communication connection to a corresponding type of telephone line. As another example, the communication interface 818 may be a local area network (LAN) card to provide a data communication connection to a compatible LAN. Wireless links may also be implemented. In any such implementation, the communication interface 818 sends and receives electrical, electromagnetic or optical signals that carry data streams representing various types of information.

**[0064]** The network link 820 typically provides data communication through one or more networks to other devices. For example, the network link 820 may provide a connection through local network 822 to a host computer 824 or to medical equipment 826 such as a radiation beam source or a switch operatively coupled to a radiation beam source. The data streams transported over the

network link 820 can comprise electrical, electromagnetic or optical signals. The signals through the various networks and the signals on the network link 820 and through the communication interface 818, which carry data to and from the computer system 800, are exemplary forms of carrier waves transporting the information. The computer system 800 can send messages and receive data, including program code, through the network(s), the network link 820, and the communication interface 818.

**[0065]** Although the embodiments of the systems and methods have been described with reference to generating CT images while passively controlling a patient's breathing in a synchronized manner, it should be understood that the systems and methods may also be implemented to control other physiological motions in different medical (which may or may not involve a radiation source) or non-medical procedures. In addition, although the gantry 12 has been described as making a 360° rotation around the patient 16 during CT image data acquisition, such needs not be the case. For example, if a full cone detector is used, the system 10 may acquire data while the gantry 12 rotates 180° plus the angle of the beam pattern. Other angles of rotation may also be used, depending on the particular system being employed and the particular need of an application. For example, in some procedures, it may be desirable to obtain image data (which may, for example, be incomplete for the purpose of performing 3D reconstruction) at certain prescribed of gantry angle(s). In such cases, the processor 54 can be configured to generate visual signals (that correspond to the prescribed range(s) of gantry angles) on the screen 101 for prompting the patient 16 accordingly. Also, instead of rotating the gantry 12 in alternating opposite directions in successive rotations, in alternative embodiments, the gantry 12 can be configured to rotate in a same direction in successive rotations. Furthermore, instead of a CT procedure, the above described user interfaces or similar user interfaces can be implemented in a laminar tomography procedure, a MRI procedure, a PET procedure, or other imaging procedures, in which a plurality of image data is desired to be generated. As such, the gantry 12 should not be limited to a rotatable structure as described previously, and could have other configurations, such as a sliding or translating configuration, depending on the particular application or procedure.

**[0066]** Although particular embodiments of the present inventions have been shown and described, it will be understood that it is not intended to limit the present inventions to the preferred embodiments, and it will be obvious to those skilled in the art that various changes and modifications may be made without departing from the scope of the present inventions. For example, the operations performed by the processor 54 can be performed by any combination of hardware and software within the scope of the invention, and should not be limited to particular embodiments comprising a particular definition of "processor". Also, the term "image" as used in this specification

includes image data that may be stored in a circuitry or a computer-readable medium, and should not be limited to image data that is displayed visually. The specification and drawings are, accordingly, to be regarded in an illustrative rather than restrictive sense. The present inventions are intended to cover alternatives, modifications, and equivalents, which may be included within the scope of the present inventions as defined by the claims.

## Claims

1. A method of prompting a patient (16), comprising:
  - informing a patient (16) of a relationship between a target result to be achieved by an activity, and time and a result of the activity being performed by the patient, wherein the patient (16) is informed of the relationship between the result of the activity and the target result by presenting a relationship between a position of a portion of the patient and a target position (420, 422, 424, 426) for the portion, wherein the informing comprises: displaying a first object (302, 402) that represents the position of the portion and displaying a second object (304) that represents the target position (420, 422, 424, 426), wherein the target position (420, 422, 424, 426) varies in the time, and the method further comprises informing the patient of the relationship between the time and the target position (420, 422, 424, 426), and wherein the informing the patient of the relationship between the time and the target position (420, 422, 424, 426) comprises displaying the second object (304) in a screen (101), the second object (304) having a plurality of portions (322a, 322b, 404a, 404b, 404c), each of the portions (322a, 322b, 404a, 404b, 404c) representing a portion of a gantry rotation during which image data is to be collected.
2. The method of claim 1, wherein the target result comprises a prescribed inhale level of a breathing, a prescribed exhale level of a breathing, or a prescribed duration of a breath-hold.
3. The method of claim 1, wherein the activity comprises breathing.
4. The method of any one preceding claim, further comprising moving the second object, wherein the moving the second object is performed in synchronization with a movement of a medical device.
5. The method of claim 4, wherein the medical device

comprises a radiation source.

6. The method of claim 5, wherein the one or more portions of the second object have a length representing a range of positions of the radiation source at which image data is to be obtained using the radiation source. 5
7. The method of claim 1, further comprising providing an audio signal when the target result is achieved by the activity. 10
8. The method of claim 1, further comprising providing a visual signal when the target result is achieved by the activity. 15
9. The method of claim 1, wherein the portion comprises a body part that is moveable in association with a breathing of the patient (16). 20
10. The method of claim 1, further comprising moving the second object in synchronization with a rotation of a radiation source.
11. A computer program product that includes a medium, the medium comprises a set of instructions, an execution of which causes any of the methods of claims 1 to 10 to be performed. 25
12. A system for performing any of the methods of claims 1 to 10 comprising a processor (54) configured to inform the patient (16) of the relationship between the target result to be achieved by the activity and time and a result of the activity being performed by the patient by generating the graphics, and the screen for displaying the graphics. 30 35

#### Patentansprüche

1. Verfahren zum Auffordern eines Patienten (16), umfassend:

Informieren eines Patienten (16) über eine Beziehung zwischen einem Zielergebnis, das durch eine Aktivität und Zeit erreicht werden soll, und einem Ergebnis der Aktivität, die durch den Patienten durchgeführt wird, wobei der Patient (16) über die Beziehung zwischen dem Ergebnis der Aktivität und dem Zielergebnis durch Darstellen einer Beziehung zwischen einer Position eines Abschnitts des Patienten und einer Zielposition (420, 422, 424, 426) des Abschnitts informiert wird, wobei das Informieren Folgendes umfasst: Anzeigen eines ersten Objekts (302, 402), welches die Position des Abschnitts darstellt, und Anzeigen eines zweiten Objekts (304), welches die

Zielposition (420, 422, 424, 426) darstellt, wobei die Zielposition (420, 422, 424, 426) in der Zeit variiert, und das Verfahren fernere Informieren des Patienten über die Beziehung zwischen der Zeit und der Zielposition (420, 422, 424, 426) umfasst, und wobei das Informieren des Patienten über die Beziehung zwischen der Zeit und der Zielposition (420, 422, 424, 426) Anzeigen des zweiten Objekts (304) auf einem Bildschirm (101) umfasst, wobei das zweite Objekt (304) eine Vielzahl von Abschnitten (322a, 322b, 404a, 404b, 404c) aufweist, wobei jeder der Abschnitte (322a, 322b, 404a, 404b, 404c) einen Abschnitt einer Gantry-Rotation darstellt, während welchem Bilddaten gesammelt werden sollen.

2. Verfahren nach Anspruch 1, wobei das Zielergebnis ein vorgeschriebenes Einatmungsniveau der Atmung, ein vorgeschriebenes Ausatmungsniveau der Atmung oder eine vorgeschriebene Dauer des Luftanhaltens umfasst.
3. Verfahren nach Anspruch 1, wobei die Aktivität Atmen umfasst.
4. Verfahren nach einem der vorhergehenden Ansprüche, ferner umfassend Bewegen des zweiten Objekts, wobei das Bewegen des zweiten Objekts synchron mit einer Bewegung einer medizinischen Vorrichtung durchgeführt wird.
5. Verfahren nach Anspruch 4, wobei die medizinische Vorrichtung eine Strahlenquelle umfasst.
6. Verfahren nach Anspruch 5, wobei der eine oder die mehreren Abschnitte des zweiten Objekts eine Länge aufweisen, die einen Bereich von Positionen der Strahlenquelle darstellt, bei welchem Bilddaten unter Verwendung der Datenquelle gesammelt werden sollen.
7. Verfahren nach Anspruch 1, ferner umfassend Bereitstellen eines Audiosignals, wenn das Zielergebnis durch die Aktivität erreicht wird.
8. Verfahren nach Anspruch 1, ferner umfassend Bereitstellen eines visuellen Signals, wenn das Zielergebnis durch die Aktivität erreicht wird.
9. Verfahren nach Anspruch 1, wobei der Abschnitt einen Körperteil umfasst, der in Assoziation mit einer Atmung des Patienten (16) bewegt werden kann.
10. Verfahren nach Anspruch 1, ferner umfassend Bewegen des zweiten Objekts synchron mit einer Strahlung der Strahlenquellen.

11. Computerprogrammprodukt, welches ein Medium beinhaltet, wobei das Medium einen Satz von Instruktionen umfasst, dessen Ausführung die Durchführung eines der Verfahren nach den Ansprüchen 1 bis 10 veranlasst.
12. System zum Durchführen eines der Verfahren nach den Ansprüchen 1 bis 10, umfassend einen Prozessor (54), der konfiguriert ist, um den Patienten (16) über die Beziehung zwischen dem Zielergebnis, das durch die Aktivität und Zeit erreicht werden soll, und einem Ergebnis der Aktivität, die durch den Patienten durchgeführt werden soll, durch Erzeugen der Grafiken und den Bildschirm zum Anzeigen der Grafiken zu informieren.

### Revendications

1. Procédé de sollicitation d'un patient (16) comprenant :
- le fait d'informer un patient (16) d'un rapport entre un résultat cible à atteindre par une activité et le temps et un résultat de l'activité étant réalisée par le patient, dans lequel le patient (16) est informé du rapport entre le résultat de l'activité et le résultat cible en présentant un rapport entre une position d'une partie du patient et une position cible pour la partie (420, 422, 424, 426) pour la partie, dans lequel le fait d'informer comprend : l'affichage d'un premier objet (302, 402) qui représente la position de la partie et l'affichage d'un second objet (304) qui représente la position cible (420, 422, 424, 426), dans lequel la position cible (420, 422, 424, 426) varie dans le temps et le procédé comprend en outre le fait d'informer le patient du rapport entre le temps et la position cible (420, 422, 424, 426), et dans lequel le fait d'informer le patient du rapport entre le temps et la position cible (420, 422, 424, 426) comprend l'affichage du second objet (304) dans un écran (101), le second objet (304) ayant une pluralité de parties (322a, 322b, 404a, 404b, 404c), chacune des parties (322a, 322b, 404a, 404b, 404c) représentant une partie d'une rotation de portique au cours de laquelle des données d'image doivent être collectées.
2. Procédé selon la revendication 1, dans lequel le résultat cible comprend un niveau d'inspiration prescrit d'une respiration, un niveau d'expiration prescrit d'une respiration ou une durée prescrite d'une respiration retenue.
3. Procédé selon la revendication 1, dans lequel l'activité comprend la respiration.
4. Procédé selon une quelconque revendication précédente, comprenant en outre le déplacement du second objet, dans lequel le déplacement du second objet est réalisé en synchronisation avec le déplacement d'un dispositif médical.
5. Procédé selon la revendication 4, dans lequel le dispositif médical comprend une source de rayonnement.
6. Procédé selon la revendication 5, dans lequel les une ou plusieurs parties du second objet ont une longueur représentant une plage de positions de la source de rayonnement dans laquelle des données d'image doivent être obtenues en utilisant la source de rayonnement.
7. Procédé selon la revendication 1, comprenant en outre la fourniture d'un signal audio lorsque le résultat cible est atteint par l'activité.
8. Procédé selon la revendication 1, comprenant en outre la fourniture d'un signal visuel lorsque le résultat cible est atteint par l'activité.
9. Procédé selon la revendication 1, dans lequel la partie comprend une partie de corps qui peut être déplacée en association avec une respiration du patient (16).
10. Procédé selon la revendication 1, comprenant en outre le déplacement du second objet en synchronisation avec la rotation d'une source de rayonnement.
11. Produit de programme informatique qui comprend un support, le support comprenant un ensemble d'instructions dont l'exécution entraîne la réalisation de l'un quelconque des procédés selon les revendications 1 à 10.
12. Système de réalisation de l'un quelconque des procédés selon les revendications 1 à 10 comprenant un processeur (54) configuré pour informer le patient (16) du rapport entre le résultat cible à atteindre par l'activité et le temps et un résultat de l'activité étant réalisée par le patient en générant les graphiques, et l'écran pour afficher les graphiques.




FIG. 1




FIG. 2A




FIG. 2B




FIG. 2C




FIG. 3A




FIG. 3B




FIG. 3C




FIG. 4




FIG. 5




FIG. 6




FIG. 7




Fig. 8

**REFERENCES CITED IN THE DESCRIPTION**

*This list of references cited by the applicant is for the reader's convenience only. It does not form part of the European patent document. Even though great care has been taken in compiling the references, errors or omissions cannot be excluded and the EPO disclaims all liability in this regard.*

**Patent documents cited in the description**

- US 3952201 A, Hounsfield [0003]
- WO 03003796 A [0008]
- JP 2000325339 A [0009]
- US 2003188757 A1 [0010]
- US 6621889 B [0031]
- US 89312201 A [0031]

专利名称(译)	用于使呼吸与医疗程序同步的患者视觉指导技术		
公开(公告)号	<a href="#">EP1799099A2</a>	公开(公告)日	2007-06-27
申请号	EP2005803610	申请日	2005-09-27
[标]申请(专利权)人(译)	創新醫療系統技術公司 MOSTAFAVI HASSAN		
申请(专利权)人(译)	瓦里安医疗系统TECHNOLOGIES , INC. Mostafavi哈桑		
当前申请(专利权)人(译)	瓦里安医疗系统TECHNOLOGIES , INC. Mostafavi哈桑		
[标]发明人	VARIAN MEDICAL SYST TECH INC MOSTAFAVI HASSAN		
发明人	VARIAN MEDICAL SYSTEMS TECHNOLOGIES, INC. MOSTAFAVI, HASSAN		
IPC分类号	A61B5/00 G01R33/28 A61B5/055 A61B5/11 A61B6/00 A61B6/03 A61B19/00 A61N5/10 G01B11/03 G01R33/48 G01T1/00 G01T1/161		
CPC分类号	A61B5/055 A61B5/1127 A61B5/1135 A61B5/7285 A61B6/463 A61B6/541 A61B90/39 A61B2090/371 A61N5/1064 A61N5/1065 A61N5/1069 A61N2005/1059		
优先权	10/957009 2004-09-30 US		
其他公开文献	EP1799099B1 EP1799099A4		
外部链接	<a href="#">Espacenet</a>		

#### 摘要(译)

提示患者的方法包括通知患者患者正在进行的活动的结果与由该活动实现的第一目标结果之间的关系。提示患者的方法包括通知患者时间和活动要达到的目标结果之间的关系。一种提示患者的方法包括通知患者患者的一部分的位置与该部分的第一目标位置之间的关系。