



US008175670B2

(12) **United States Patent**  
**Baker, Jr. et al.**

(10) **Patent No.:** **US 8,175,670 B2**  
(45) **Date of Patent:** **May 8, 2012**

- (54) **PULSE OXIMETRY SIGNAL CORRECTION USING NEAR INFRARED ABSORPTION BY WATER** 4,364,008 A 12/1982 Jacques  
4,711,244 A 12/1987 Kuzara  
4,714,341 A 12/1987 Hamaguri et al.  
4,723,554 A 2/1988 Oman et al.  
4,805,623 A 2/1989 Jobsis  
4,850,365 A 7/1989 Rosenthal  
4,860,753 A 8/1989 Amerena  
4,883,055 A 11/1989 Merrick  
4,907,594 A 3/1990 Muz  
4,911,167 A 3/1990 Corenman et al.  
4,936,679 A 6/1990 Mersch  
4,972,331 A 11/1990 Chance  
5,057,695 A 10/1991 Hirao et al.  
5,086,781 A 2/1992 Bookspan  
5,111,817 A 5/1992 Clark et al.  
5,119,815 A 6/1992 Chance  
5,122,974 A 6/1992 Chance  
5,146,091 A 9/1992 Knudson
- (75) Inventors: **Clark R. Baker, Jr.**, Newman, CA (US);  
**Edward Karst**, South Pasadena, CA (US); **Carine Hoarau**, Lafayette, CA (US)
- (73) Assignee: **Nellcor Puritan Bennett LLC**, Boulder, CO (US)
- (\* ) Notice: Subject to any disclaimer, the term of this patent is extended or adjusted under 35 U.S.C. 154(b) by 1630 days.

(Continued)

(21) Appl. No.: **11/521,960**

(22) Filed: **Sep. 15, 2006**

(65) **Prior Publication Data**

US 2007/0106137 A1 May 10, 2007

**Related U.S. Application Data**

(63) Continuation-in-part of application No. 10/797,475, filed on Mar. 9, 2004, now Pat. No. 7,277,741.

(51) **Int. Cl.**

**A61B 5/1455** (2006.01)

**A61B 5/02** (2006.01)

(52) **U.S. Cl.** ..... **600/336; 600/324; 600/502**

(58) **Field of Classification Search** ..... **600/310, 600/322, 323, 336**

See application file for complete search history.

(56) **References Cited**

**U.S. PATENT DOCUMENTS**

- 3,638,640 A 2/1972 Shaw
- 3,998,550 A 12/1976 Konishi et al.
- 4,066,068 A 1/1978 Nilsson et al.

**FOREIGN PATENT DOCUMENTS**

CA 2353007 A1 6/2000

(Continued)

**OTHER PUBLICATIONS**

Wheeler, Owen H., "Near Infrared Spectra of Organic Compounds," Department of Chemistry, College of Agriculture and Mechanic Arts, University of Puerto Rico (Mar. 1929).


(Continued)

*Primary Examiner* — Eric Winakur

(57) **ABSTRACT**

A method and system for measuring a physiological parameter, comprising collecting a first absorbance at a first wavelength, chosen to be primarily absorbed by water; collecting a second absorbance at a second wavelength, chosen to be primarily absorbed by hemoglobin; and combining the first signal and the second signal to generate a combined plethysmograph signal which is proportionate lower in noise caused by motion-related interference.

**39 Claims, 9 Drawing Sheets**



U.S. PATENT DOCUMENTS							
5,167,230	A	12/1992	Chance	6,591,122	B2	7/2003	Schmitt
5,224,478	A	7/1993	Sakai et al.	6,592,574	B1	7/2003	Shimmick et al.
5,277,181	A	1/1994	Mendelson et al.	6,594,513	B1	7/2003	Jobsis et al.
5,279,295	A	1/1994	Martens et al.	6,600,946	B1	7/2003	Rice
5,282,467	A	2/1994	Piantadosi et al.	6,606,509	B2	8/2003	Schmitt
5,297,548	A	3/1994	Pologe	6,615,064	B1	9/2003	Aldrich
5,337,745	A	8/1994	Benaron	6,622,095	B2	9/2003	Kobayashi et al.
5,337,937	A	8/1994	Blank et al.	6,635,491	B1	10/2003	Khalil et al.
5,348,004	A	9/1994	Hollub	6,636,759	B2	10/2003	Robinson
5,355,880	A	10/1994	Thomas et al.	6,643,543	B2	11/2003	Takehara et al.
5,372,136	A	12/1994	Steuer et al.	6,654,620	B2	11/2003	Wu et al.
5,377,674	A	1/1995	Kuestner	6,658,277	B2	12/2003	Wasserman
5,385,143	A	1/1995	Aoyagi	6,662,030	B2	12/2003	Khalil et al.
5,431,170	A*	7/1995	Mathews ..... 600/323	6,668,181	B2	12/2003	Wenzel et al.
5,482,036	A	1/1996	Diab et al.	6,671,526	B1	12/2003	Aoyagi et al.
5,499,627	A	3/1996	Steuer et al.	6,671,528	B2	12/2003	Steuer et al.
5,553,614	A	9/1996	Chance	6,675,029	B2	1/2004	Monfre et al.
5,564,417	A	10/1996	Chance	6,678,543	B2	1/2004	Diab et al.
5,575,285	A	11/1996	Takanashi et al.	6,687,519	B2	2/2004	Steuer et al.
5,615,689	A	4/1997	Kotler	6,690,958	B1	2/2004	Walker et al.
5,630,413	A	5/1997	Thomas et al.	6,708,048	B1	3/2004	Chance
5,645,059	A	7/1997	Fein et al.	6,711,424	B1	3/2004	Fine et al.
5,645,060	A	7/1997	Yorkey	6,711,425	B1	3/2004	Reuss
5,687,721	A	11/1997	Kuhls	6,777,240	B2	8/2004	Hazen et al.
5,692,503	A	12/1997	Kuenstner	6,785,568	B2	8/2004	Chance
5,701,902	A	12/1997	Vari et al.	6,801,797	B2	10/2004	Mannheimer et al.
5,720,284	A	2/1998	Aoyagi et al.	6,801,799	B2	10/2004	Mendelson
5,735,284	A	4/1998	Tsoglin et al.	6,849,046	B1	2/2005	Eyal-Bickels
5,747,789	A	5/1998	Godik	6,873,865	B2	3/2005	Steuer et al.
5,755,672	A	5/1998	Arai et al.	6,949,081	B1	9/2005	Chance
5,758,644	A	6/1998	Diab et al.	6,950,699	B1	9/2005	Manwaring et al.
5,769,785	A	6/1998	Diab et al.	6,961,598	B2	11/2005	Diab
5,779,631	A	7/1998	Chance	7,006,856	B2	2/2006	Baker, Jr. et al.
5,788,643	A	8/1998	Feldman	7,027,849	B2	4/2006	Al-Ali
5,803,908	A	9/1998	Steuer et al.	7,039,538	B2	5/2006	Baker, Jr.
5,827,181	A	10/1998	Dias et al.	7,206,621	B2	4/2007	Aoyagi
5,830,139	A	11/1998	Abreu	7,209,774	B2	4/2007	Baker, Jr.
5,833,602	A	11/1998	Osemwota	7,215,984	B2	5/2007	Diab et al.
5,842,981	A	12/1998	Larsen et al.	7,215,986	B2	5/2007	Diab et al.
5,853,364	A	12/1998	Baker, Jr. et al.	7,236,811	B2	6/2007	Schmitt
5,860,919	A	1/1999	Kiani-Azarbayjany et al.	7,239,902	B2	7/2007	Schmitt et al.
5,873,821	A	2/1999	Chance et al.	7,277,741	B2	10/2007	Debreczeny et al.
5,906,582	A	5/1999	Kondo et al.	7,328,053	B1	2/2008	Diab et al.
5,995,856	A	11/1999	Mannheimer et al.	7,376,453	B1	5/2008	Diab et al.
6,011,986	A	1/2000	Diab et al.	7,383,070	B2	6/2008	Diab et al.
6,064,898	A	5/2000	Aldrich	7,392,075	B2	6/2008	Baker, Jr.
6,120,460	A	9/2000	Abreu	2001/0005773	A1	6/2001	Larsen et al.
6,125,297	A	9/2000	Siconolfi	2001/0020122	A1	9/2001	Steuer et al.
6,134,460	A	10/2000	Chance	2001/0039376	A1	11/2001	Steuer et al.
6,149,591	A	11/2000	Henderson et al.	2001/0044700	A1	11/2001	Kobayashi et al.
6,157,850	A	12/2000	Diab et al.	2002/0026106	A1	2/2002	Khalil et al.
6,163,715	A	12/2000	Larsen et al.	2002/0035318	A1	3/2002	Mannheimer et al.
6,178,342	B1	1/2001	Thompson et al.	2002/0038079	A1	3/2002	Steuer et al.
6,181,958	B1	1/2001	Steuer et al.	2002/0042558	A1	4/2002	Mendelson
6,222,189	B1	4/2001	Misner et al.	2002/0049389	A1	4/2002	Abreu
6,230,035	B1	5/2001	Aoyagi et al.	2002/0062071	A1	5/2002	Diab et al.
6,246,894	B1	6/2001	Steuer et al.	2002/0111748	A1	8/2002	Kobayashi et al.
6,266,546	B1	7/2001	Steuer et al.	2002/0133068	A1	9/2002	Huiku
6,280,396	B1	8/2001	Clark et al.	2002/0161287	A1	10/2002	Schmitt
6,312,393	B1	11/2001	Abreu	2002/0161290	A1	10/2002	Chance
6,336,044	B1	1/2002	Ghiassi et al.	2002/0165439	A1	11/2002	Schmitt
6,370,426	B1	4/2002	Campbell et al.	2002/0198443	A1	12/2002	Ting
6,397,091	B2	5/2002	Diab et al.	2003/0023140	A1	1/2003	Chance
6,400,971	B1	6/2002	Finarov et al.	2003/0055324	A1	3/2003	Wasserman
6,402,690	B1	6/2002	Rhee et al.	2003/0060693	A1	3/2003	Monfre et al.
6,415,236	B2	7/2002	Kobayashi et al.	2003/0139687	A1	7/2003	Abreu
6,438,399	B1	8/2002	Kurth	2003/0144584	A1	7/2003	Mendelson
6,442,408	B1	8/2002	Wenzel et al.	2003/0220548	A1	11/2003	Schmitt
6,466,807	B1	10/2002	Dobson et al.	2003/0220576	A1	11/2003	Diab
6,487,439	B1	11/2002	Skladnev et al.	2004/0010188	A1	1/2004	Wasserman
6,488,677	B1	12/2002	Bowman et al.	2004/0054270	A1	3/2004	Pewzner et al.
6,501,974	B2	12/2002	Huiku	2004/0087846	A1	5/2004	Wasserman
6,501,975	B2	12/2002	Diab et al.	2004/0107065	A1	6/2004	Al-Ali
6,512,936	B1	1/2003	Monfre et al.	2004/0127779	A1	7/2004	Steuer et al.
6,526,301	B2	2/2003	Larsen et al.	2004/0171920	A1	9/2004	Mannheimer et al.
6,544,193	B2	4/2003	Abreu	2004/0176670	A1	9/2004	Takamura et al.
6,546,267	B1	4/2003	Sugiura et al.	2004/0176671	A1	9/2004	Fine et al.
6,549,795	B1	4/2003	Chance	2004/0230106	A1	11/2004	Schmitt et al.
				2005/0080323	A1	4/2005	Kato

2005/0101850	A1	5/2005	Parker
2005/0107676	A1	5/2005	Acosta et al.
2005/0113656	A1	5/2005	Chance
2005/0168722	A1	8/2005	Forstner et al.
2005/0192488	A1	9/2005	Bryenton et al.
2005/0203357	A1	9/2005	Debreczeny et al.
2005/0209517	A1	9/2005	Diab et al.
2005/0267346	A1	12/2005	Faber et al.
2006/0009688	A1	1/2006	Lamego et al.
2006/0015021	A1	1/2006	Cheng
2006/0020181	A1	1/2006	Schmitt
2006/0030763	A1	2/2006	Mannheimer et al.
2006/0030764	A1	2/2006	Porges et al.
2006/0052680	A1	3/2006	Diab
2006/0058683	A1	3/2006	Chance
2006/0084864	A1	4/2006	Schmitt et al.
2006/0200014	A1	9/2006	Fine et al.
2006/0200016	A1	9/2006	Diab et al.
2006/0217609	A1	9/2006	Diab et al.
2006/0253016	A1	11/2006	Baker, Jr. et al.
2006/0287587	A1	12/2006	Yarita
2006/0287588	A1	12/2006	Yarita
2007/0106137	A1	5/2007	Baker, Jr. et al.
2007/0118027	A1	5/2007	Baker, Jr. et al.
2007/0129614	A1	6/2007	Schmitt et al.
2007/0225581	A1	9/2007	Diab et al.
2007/0249918	A1	10/2007	Diab et al.
2007/0291832	A1	12/2007	Diab et al.
2008/0004514	A1	1/2008	Diab et al.
2008/0009690	A1	1/2008	Debreczeny et al.
2008/0033266	A1	2/2008	Diab et al.
2008/0036752	A1	2/2008	Diab et al.
2008/0045823	A1	2/2008	Diab et al.
2008/0255436	A1	10/2008	Baker

## FOREIGN PATENT DOCUMENTS

DE	19855521	A1	6/2000
DE	102 13 692	A1	10/2003
EP	1135184	A1	6/2000
EP	1184663	A2	3/2002
FR	2710517		4/1995
JP	04-040940		2/1992
JP	5-212016		8/1993
JP	5-329163		12/1993
JP	11-244266		9/1999
JP	25095606		4/2005
WO	WO 92/20273		11/1992
WO	WO 94/03102		2/1994
WO	WO 97/49330		12/1997
WO	WO 98/34097		8/1998
WO	WO 00/32262	A1	6/2000
WO	WO 00/71025	A1	11/2000
WO	WO 93/13706	A2	1/2001
WO	WO 01/16577	A1	3/2001
WO	WO 01/45553	A1	6/2001

## OTHER PUBLICATIONS

Pace, Nello, et al., "Studies on Body Composition: III. The body water and chemically combined nitrogen content in relation to fat content," Naval Medical Research Institute, Bethesda, Maryland (Jan. 11, 1945).

Mitchell, H. M., et al., "The Chemical Composition of the Adult Human Body and Its bearing on the Biochemistry of Growth," Division of Animal Nutrition, Departments of Physiology and Animal Husbandry, University of Illinois, pp. 625-637 (Feb. 1945).

Schloerb, Paul R., et al., "The Measurement of Total Body Water in the Human Subject by Deuterium Oxide Dilution," *Surgical Research Laboratories of the Peter Bent Brigham Hospital, and the Department of Surgery and the Biophysical Laboratory of the Harvard Medical School*, pp. 1296-1310 (Mar. 20, 1950).

Forbes, R.M., et al., "The Composition of the Adult Human Body as Determined by Chemical Analysis," Division of Animal Nutrition, and the Department of Anatomy, University of Illinois, Jan. 19, 1953.

Buijs, K., et al., "Near-Infrared Studies of the Structure of Water. I. Pure Water," *The Journal of Chemical Physics*, vol. 39, No. 8, pp. 2035-2041 (Oct. 15, 1963).

Choppin, G.R., et al., "Near-Infrared Studies of the Structure of Water. II. Ionic Solution," *The Journal of Chemical Physics*, vol. 39, No. 8, pp. 2042-2050 (Oct. 15, 1963).

Goldstein, R., et al., "The Near-Infrared Absorption of Liquid Water at Temperatures Between 27 and 209° C.," *J. Quant. Spectrosc. Radiat Transfer.*, vol. 4, pp. 441-451 (1964).

Ben-Gera, I., et al., "Influence of Fat Concentration on the Absorption Spectrum of Milk in the Near-Infrared Region," *Israel J. Agric. Res.*, Vol. 18, No. 3, pp. 117-124 (Jul. 1968).

Houseman, R.A., et al., "The measurement of total body water in living pigs by deuterium oxide dilution and its relation to body composition," *Br. J. Nutr.*, vol. 30, pp. 149-156 (1973).

Krikorian, S. Edward, et al., "The identification and origin of N-H overtone and combination bands in the near-infrared spectra of simple primary and secondary amides," *Spectrochimica Acta*, vol. 29A, pp. 1233-1246 (1973).

Lesser, G.T., et al., "Body water compartments with human aging using fat-free mass as the reference standard," *Am J. Physiol Regul Integr Comp Physiol.*, vol. 236, pp. 215-220 (1979).

Sheng, Hwai-Ping, et al., "A review of body composition studies with emphasis on total body water and fat," *The American Journal of Clinical Nutrition*, vol. 32., pp. 630-647 (Mar. 1979).

Martens, H., et al., "Unscrambling Multivariate Data from Mixtures: I: Fat, water and protein determination in meat by near-infrared reflectance spectroscopy, II: soy protein and collagen determination in meat products from amino acid data," *Meat Res. Workers, Proc. European Meeting*, pp. 146-149 (1980).

Fomon, Samuel J., et al., "Body composition of reference children from birth to age 10 years," *The American Journal of clinical Nutrition*, vol. 35, pp. 1169-1175, (May 1982).

Lanza, Elaine, "Determination of Moisture, Protein, Fat, and Calories in Raw Pork and Beef by near Infrared Spectroscopy," *Journal of Food Science*, vol. 48, pp. 471-474 (1983).

Shields, R. G., Jr., et al., "Efficacy of Deuterium Oxide to Estimate Body Composition of Growing Swine"; *Journal of Animal Science*, vol. 57, No. 1, pp. 66-73, (1983).

Wolfgang, Armeth, "Multivariate Infrared and near-infrared Spectroscopy: rapid analysis of protein, fat and water in meat," *Food Res and Data Analysis, Proc from IUOST Symp*, Oslo, Norway, pp. 239-251 (1983).

Cohn, S.H., et al., "Assessment of cellular mass and lean body mass by noninvasive nuclear techniques," *J. Lab Clin Med.*, vol. 105, pp. 305-311 (1985).

Hannon, John P., et al., "Splenic red cell sequestration and blood volume measurements in conscious pigs," *Am J. Physiol.*, vol. 248, pp. R293-R301 (1985).

Potts, R.O., et al., "A Noninvasive, In Vivo Technique to Quantitatively measure Water Concentration of the Stratum Corneum Using Attenuated Total-Reflectance Infrared Spectroscopy," *Arch. Dermatol Res.*, vol. 277, pp. 489-495 (1985).

Cox, Patrick, et al., "Variations in Lipids in Different Layers of Porcine Epidermis," *J. Invest Dermatol.*, vol. 87, pp. 741-744 (1986).

Valdes, E. V., et al., "Determination of Crude Protein and Fat in Carcass and Breast Muscle Samples of Poultry by Near Infrared Reflectance Spectroscopy," *Poultry Science*, vol. 65, pp. 485-490 (1986).

Hedberg, Christopher L., et al., "The Time Course of Lipid Biosynthesis in Pig Epidermis," *J. Invest Dermatol.*, vol. 91, pp. 169-174 (1988).

Hedberg, Christopher L., et al., "The nonpolar Lipids of Pig Epidermis," *J. Invest Dermatol.*, vol. 90, pp. 225-229 (1988).

Trapp, Scott A., et al., "An improved spectrophotometric bromide assay for the estimation of extracellular water volume," *Clinica Chimica Acta.*, vol. 181, pp. 207-212, (1989).

Bommannan, D., et al., "Examination of Stratum Corneum Barrier Function In Vivo by Infrared Spectroscopy," *J. Invest Dermatol*, vol. 95, pp. 403-408 (1990).

Hannon, John P., et al., "Normal pPhysiological Values for Conscious Pigs Used in Biomedical Research," *Laboratory Animal Science*, vol. 40, No. 3, May 1990.

Mak, Vivien H.W., et al., "Oleic Acid Concentration and Effect in Human Stratum Corneum: Non-Invasive determination by Attenu-

- ated Total Reflectance Infrared Spectroscopy In Vivo," *Journal of Controlled Release*, vol. 12, pp. 67-75 (1990).
- Edwardson, P. et al., "The Use of FT-IR for the Determination of Stratum Corneum Hydration in Vitro and in Vivo," *J. of Pharmaceutical & Biomed. Analysis*, vol. 9, Nos. 10-12, pp. 1089-1094, 1991.
- Drummer, C., et al., "Effects of an acute saline infusion on fluid and electrolyte metabolism in humans," *Am J. Physiol.*, vol. 262, pp. F744-F754 (1992).
- Horber, F.F., et al., "Impact of hydration status on body composition as measured by dual energy X-ray absorptiometry in normal volunteers and patients on haemodialysis," *The British Journal of Radiology*, vol. 65, pp. 895-900 (1992).
- Schmitt et al., *Proc. SPIE*, "Measurement of blood hematocrit by dual-wavelength near-IP photoplethysmography," 1641:150-161 (1992).
- Diaz-Carrillo, E., et al., "Near infrared calibrations for goat's milk components; protein, total casein,  $\alpha_s$ -,  $\beta$ - and k-caseins, fat and lactose," *J. near Infrared Spectrosc.*, vol. 1, pp. 141-146 (1993).
- Martin, K., "Direct Measurement of Moisture in Skin by NIR spectroscopy," *J. Soc. Cosmet. Chem.*, 44:249-261 (1993).
- Richard, Stephanie, et al., "Characterization of the Skin In Vivo by High Resolution Magnetic Resonance Imaging: Water Behavior and Age-Related Effects," *The Journal of Investigative Dermatology*, vol. 100, No. 5, pp. 705-709 (May 1993).
- Thompson et al., "Can bioelectrical impedance be used to measure total body water in dialysis patients?," *Physiol. Meas.*, 14:455-461 (1993).
- Bewig, Karen M., et al., "Discriminant Analysis of Vegetable Oils by Near-Infrared Reflectance Spectroscopy," *JAOCs*, vol. 71, No. 2, pp. 195-200 (Feb. 1994).
- Kamishikiryō-Yamashita, Hiromi, et al., "Protein Content in Milk by Near-Infrared Spectroscopy," *Journal of Food Science*, vol. 59, No. 2, pp. 313-315 (1994).
- Matcher, S. J., et al., "Absolute quantification of deoxyhaemoglobin concentration in tissue near infrared spectroscopy," *Phys. Med. Biol.*, vol. 39, pp. 1295-1312 (1994).
- Simanonok, Karl E., et al., "A Comprehensive Guyton Model Analysis of Physiologic Responses to Preadapting the Blood volume as a Countermeasure to Fluid Shifts," *J. Clin Pharmacol.*, vol. 34, pp. 440-453 (1994).
- Steven, Alasdair C., et al., "Protein composition of cornified cell envelopes of epidermal keratinocytes," *Journal of Cell Science*, vol. 107, pp. 693-700 (1994).
- Takeo, T. et al., "Skin Hydration State Estimation Using a Fiber-Optic Refractometer," *Applied Optics*, vol. 33, No. 19, Jul. 1994, p. 4267-72.
- Warren, Joan L., et al., "The burden and Outcomes Associates with Dehydration among US Elderly, 1991," *American Journal of Public Health*, vol. 84, No. 8, pp. 1265-1269 (Aug. 1994).
- Åneman, Anders, et al., "Splanchnic and Renal Sympathetic Activity in Relation to Hemodynamics During Isoflurane Administration in Pigs," *Anesth Analg.*, vol. 80, pp. 135-142, (1995).
- Kisch, Hille, et al., "Accuracy and reproducibility of the measurement of actively circulating blood volume with an integrated fiberoptic monitoring system," *Critical Care Medicine*, vol. 23, No. 5, pp. 885-893 (1995).
- Isaksson, Tomas, et al., "Non-Destructive Determination of Fat, Moisture and Protein in Salmon Fillets by Use of Near-Infrared Diffuse Spectroscopy," *J. Sci Food Agric.*, vol. 69, pp. 95-100 (1995).
- Quiniou, N., et al., "Prediction of Tissular Body Composition from Protein and Lipid Deposition in Growing Pigs," *J. Anim. Sci.*, vol. 73, pp. 1567-1575, (1995).
- Avis, N.J., et al.; "In vitro multifrequency electrical impedance measurements and modeling of the cervix in late pregnancy," *Physiological Measurement*, vol. 17, pp. A97-A103, 1996.
- Gniadecka, M., et al., "Assessment of dermal water by high-frequency ultrasound: comparative studies with nuclear magnetic resonance," *British Journal of Dermatology*, vol. 135, pp. 218-224, (1996).
- Finn, Patrick J., et al., "Progressive cellular dehydration and proteolysis in critically ill patients," *The Lancet*, vol. 347, pp. 654-646 (Mar. 9, 1996).
- Johnson et al., "Monitoring of Extracellular and Total Body Water during Hemodialysis Using Multifrequency Bio-Electrical Impedance Analysis," *Kidney and Blood Pressure Research*, 19:94-99 (1996).
- Kotler, D.P., et al.; "Prediction of body cell mass, fat-free mass, and total body water with bioelectrical impedance analysis: effects of race, sex, and disease," *Am J. Clin. Nutr.* 64(suppl):489S-97S (1996).
- Kumar, Gitesh, et al., "Non-Invasive Optical Assessment of Tissue Hydration," *International conference on Biomedical Engineering*, Jun. 3-5, 1996, Hong Kong, pp. C2-C5.
- Schmitt et al., *Proc. SPIE*, "Optimum wavelengths for measurement of blood hemoglobin content and tissue hydration by NIR spectrophotometry," 2678:442-453 (1996).
- De Fijter, W.M., et al., "Assessment of total body water ad lean body mass from anthropometry, Watson formula, creatinine kinetics, and body electrical impedance compared with antipyrine kinetics and peritoneal dialysis patients," *Nephrol Dial Transplant*, vol. 12, pp. 151-156 (1997).
- Johansen, Lars Bo, et al., "Hemodilution, central blood volume, and renal responses after an isotonic saline infusion in humans," *Am J. Physiol.*, vol. 272, pp. R549-R556 (1997).
- Visser, Marjolein, et al., "Density of fat-free body mass: relationship with race, age, and level of body fatness," *Am J. Physiol.*, vol. 272, pp. E781-E787, (1997).
- Alanen, Esko, et al., "Measurement of dielectric properties of subcutaneous fat with open-ended coaxial sensors," *Phys. Med. Biol.*, vol. 43, pp. 475-485 (1998).
- Alanen, Esko, et al., "Variational Formulation of Open-Ended Coaxial line in Contact with Layered Biological Medium," *IEEE Transactions on Biomedical Engineering*, vol. 45, No. 10, pp. 1241-1248 (Oct. 1998).
- Bonadonna, Riccardo C., et al., "Role of Tissue-Specific Blood Flow and Tissue Recruitment in Insulin-Mediated Glucose Uptake of Human Skeletal Muscle," *Circulation*, vol. 98, pp. 234-241, (1998).
- Bracco, David, et al., "Bedside determination of fluid accumulation after cardiac surgery using segmental bioelectrical impedance," *Crit Care Med*, vol. 26, No. 6, pp. 1065-1070 (1998).
- Gniadecka, Monika, et al., "Water and Protein Structure in Photoaged and Chronically Aged Skin," *J. Invest Dermatol.*, vol. 111, pp. 1129-1133 (1998).
- Gniadecka, Monika, et al., "Structure of Water, Proteins, and Lipids in Intact Human Skin, Hair, and Nail," *J. Invest Dermatol.*, vol. 110, pp. 393-398 (1998).
- Gow, Kenneth W., et al., "Effect of crystalloid administration on oxygen extraction in endotoxemic pigs," *J. Appl. Physiol.*, vol. 85, No. 5, pp. 1667-1675 (1998).
- Husby, P., et al., "Midazolam-fentanyl-isoflurane anaesthesia is suitable for haemodynamic and fluid balance studies in pigs," *Laboratory Animals*, vol. 32, pp. 316-323 (1998).
- Mitchell, A. D., et al., "Composition Analysis of Pork Carcasses by Dual-Energy X-Ray Absorptiometry," *J. Anim. Sci.*, vol. 76, pp. 2104-2114 (1998).
- Mahan, D. C., et al., "Essential and Nonessential Amino Acid Composition of Pigs from Birth to 145 Kilograms of Body Weight, and Comparison to Other Studies," *J. Anim. Sci.*, vol. 76, pp. 513-521, (1998).
- Martin, Kathleen, "In Vivo Measurements of Water in Skin by Near-Infrared Reflectance," *Applied Spectroscopy*, vol. 52, No. 7, 1998, pp. 1001-1007.
- Schou, Henning, et al., "Uncompensated Blood Loss is not Tolerated During Acute Normovolemic Hemodilution in Anesthetized Pigs," *Anesth Analg.*, vol. 87, pp. 786-794 (1998).
- Stranc, M.F., et al., "Assessment of tissue viability using near-infrared spectroscopy," *British Journal of Plastic Surgery*, vol. 51, pp. 210-217, (1998).
- Thomas, B. J., et al., "Bioimpedance Spectrometry in the Determination of Body Water Compartments: Accuracy and Clinical Significance," *Appl. Radiat. Isot.*, vol. 49, No. 5/6, pp. 447-455, (1998).
- Wilhelm, K.P., "Possible Pitfalls in Hydration Measurements," *Skin Bioengineering Techniques and Applications in Dermatology and Cosmetology*, vol. 26, pp. 223-234 (1998).
- Vrhovski, Bernadette, et al., "Biochemistry of tropoelastin," *Eur. J. Biochem.*, vol. 258, pp. 1-18 (1998).

- Alanen, Esko, et al., "Penetration of electromagnetic fields of an open-ended coaxial probe between 1 MHz and 1 GHz in dielectric skin measurements," *Phys. Med. Biol.*, vol. 44, pp. N169-N176 (1999).
- Dickens, Brian, et al., "Estimation of Concentration and Bonding Environment of Water Dissolved in Common Solvents Using Near Infrared Absorptivity," *J. Res. Natl. Inst. Stand. Technol.*, vol. 104, No. 2, pp. 173-183 (Mar.-Apr. 1999).
- Fornetti, Willa C., et al., "Reliability and validity of body composition measures in female athletes," *Journal of Applied Physiology*, vol. 87, pp. 1114-1122, (1999).
- Fusch, Christoph, et al., "Neonatal Body COMposition: Dual-Energy X-Ray Absorptiometry, Magnetic Resonance Imaging, and Three-Dimensional Chemical Shift Imaging versus Chemical Analysis in Piglets," *Pediatric Research*, vol. 46, No. 4, pp. 465-473 (1999).
- Gudivaka, R., et al., "Single- and multifrequency models for bioelectrical impedance analysis of body water compartments," *J. Appl. Physiol.*, vol. 87, No. 3, pp. 1087-1096 (1999).
- Jennings, Graham, et al., "The Use of infrared Spectrophotometry for Measuring Body Water Spaces," vol. 45, No. 7, pp. 1077-1081 (1999).
- Kalantar-Zadeh, Kamyar, et al., "Near infra-red interactance for nutritional assessment of dialysis patients," *Nephrol Dial Transplant*, vol. 14, pp. 169-175 (1999).
- Kayser-Jones, Jeanie, et al., "Factors Contributing to Dehydration in Nursing Homes: Inadequate Staffing and Lack of Professional Supervision," *J. Am Geriatr. Soc.*, vol. 47, pp. 1187-1194 (1999).
- Lange, Neale R., et al., "The measurement of lung water," *Critical Care*, vol. 3, pp. R19-R24 (1999).
- Marken Lichtenbelt, Wouter D. Van, et al., "Increased extracellular water compartment, relative to intracellular water compartment, after weight reduction," *Journal of Applied Physiology*, vol. 87, pp. 294-298 (1999).
- Rennie, Michael J., "PERSPECTIVES—Teasing out the truth about collagen," *Journal of Physiology*, vol. 521, p. 1 (1999).
- Sowa et al., "Near-infrared spectroscopic assessment of tissue hydration following surgery," *Journal of Surgical Research*, 86:62-69 (1999).
- Wagner, J.R., et al., "Analysis of Body Composition Changes of Swine During Growth and Development," *J. Anim. Sci.*, vol. 77, pp. 1442-1466 (1999).
- Wang, Zimian, et al., "Hydration of fat-free body mass: new physiological modeling approach," *Am. J. Physiol.*, vol. 276, pp. E995-E1003 (1999).
- Wang, Zimian, et al., "Hydration of fat-free body mass: review and critique of a classic body-composition constant," *Am J. Clin. Nutr.*, vol. 69, pp. 833-841 (1999).
- Ward, L., et al., "Multiple frequency bioelectrical impedance analysis: a cross-validation study of the inductor circuit and Cole models," *Physiol. Meas.*, vol. 20, pp. 333-347 (1999).
- Wells, Jonathan CK, et al., "Four-component model of body composition in children: density and hydration of fat-free mass and comparison with simpler models," *Am J. Clin. Nutr.*, vol. 69, pp. 904-912 (1999).
- Butte, Nancy F., et al., "Body Composition during the First 2 Years of Life; An Updated Reference," *Pediatric Research*, vol. 47, No. 5, pp. 578-585 (2000).
- Feigenbaum, Matthew S., et al., "Contracted Plasma and Blood volume in Chronic Heart Failure," *J Am Coll. Cardiol.*, vol. 35, No. 1, pp. 51-55 (Jan. 2000).
- Kays, Sandra E., et al., "Predicting protein content by near infrared reflectance spectroscopy in diverse cereal food products," *J. Near Infrared Spectrosc.*, vol. 8, pp. 35-43 (2000).
- Lucassen, G., et al., "Water Content and Water Profiles in Skin Measured by FTIR and Raman Spectroscopy," *Proc. SPIE*, vol. 4162, pp. 39-45 (2000).
- Plank, L. D., et al., "Similarity of Changes in Body Composition in Intensive Care Patients following Severe Sepsis or Major Blunt Injury," *Annals New York Academy of Sciences*, pp. 592-602 (2000).
- Ritz, P., et al., "Body Water Spaces and Cellular Hydration during Healthy Aging," *Annals New York Academy of Sciences*, pp. 474-483 (2000).
- Schoeller, Dale, "Bioelectrical Impedance Analysis—What does it measure?" *Annals New York Academy of Sciences*, pp. 159-162 (2000).
- Starcher, Barry C., "Lung Elastin and Matrix," *Chest*, vol. 117, No. 5, pp. 229S-234S, May 2000 Supplement.
- Young, A.E.R., et al., "Behaviour of near-infrared light in the adult human head: implications of clinical near-infrared spectroscopy," *British Journal of Anaesthesia*, vol. 84, No. 1, pp. 38-42 (2000).
- Zembrzski, Cora, "Nutrition and Hydration," Best Practices in Nursing Care to Older Adults, The Hartford Institute for Geriatric Nursing, vol. 2, No. 2, Sep. 2000, 2 pages.
- Attas, Michael, et al., "Visualization of cutaneous hemoglobin oxygenation and skin hydration using near-infrared spectroscopic imaging," *Skin Research and Technology*, vol. 7, pp. 238-245, (2001).
- Bray, George A., et al., "Evaluation of body fat in fatter and leaner 10-y-old African American and white children: the Baton Rouge Children's Study," *Am J. Clin Nutr.*, vol. 73, pp. 687-702 (2001).
- Campbell, Wayne W., et al., "The Recommended Dietary Allowance for Protein May Not Be Adequate for Older People to Maintain Skeletal Muscle," *Journal of Gerontology*, vol. 56A, No. 6, pp. M373-M-380 (2001).
- Divert, Victor E., "Body Thermal State Influence on Local Skin Thermosensitivity," *International Journal of Circumpolar Health*, vol. 60, pp. 305-311 (2001).
- Du, Y., et al., "Optical properties of porcine skin dermis between 900 nm and 1500 nm," *Phys. Med. Biol.*, vol. 46, pp. 167-181 (2001).
- Endo, Yutaka, et al., "Water drinking causes a biphasic change in blood composition in humans," *Pflügers Arch—Eur J. Physiol.*, vol. 442, pp. 362-368 (2001).
- Garulet, Marta, et al., "Site-specific differences in the fatty acid composition of abdominal adipose tissue in an obese population from a Mediterranean area: relation with dietary fatty acids, plasma lipid profile, serum insulin, and central obesity," *Am J. Clin. Nutr.*, vol. 74, pp. 585-591 (2001).
- Haga, Henning A., et al., "Electroencephalographic and cardiovascular indicators of nociception during isoflurane anaesthesia in pigs," *Veterinary Anaesthesia and Analgesia*, vol. 28, pp. 126-131 (2001).
- Kalantar-Zadeh, Kamyar, et al., "Near infra-red interactance for Longitudinal Assessment of Nutrition in Dialysis Patients," *Journal of Renal Nutrition*, vol. 11, No. 1, pp. 23-31 (Jan. 2001).
- Kamba, Masayuki, et al., "Proton magnetic resonance spectroscopy for assessment of human body composition," *Am J. Clin. Nutr.*, vol. 73, pp. 172-176 (2001).
- Lever, M., et al., "Some ways of looking at compensatory kosmotropes and different water environments," *Comparative Biochemistry and Physiology*, vol. 130, Part A, pp. 471-486, (2001).
- Mingrone, G., et al., "Unreliable use of standard muscle hydration value in obesity," *Am J. Physiol Endocrinol Metab.*, vol. 280, pp. E365-E371, (2001).
- Sašic, Slobodan, et al., "Short-Wave Near-Infrared Spectroscopy of Biological Fluids. 1. Quantitative Analysis of Fat, Protein, and Lactose in Raw Milk by Partial Least-Squares Regression and Band Assignment," *Anal. Chem.*, vol. 73, pp. 64-71 (2001).
- Schnickel, A.P., et al., "Evaluation of alternative measures of pork carcass composition," *J. Anim. Sci.*, vol. 79, pp. 1093-1119, (2001).
- Sowa et al., "Near infrared spectroscopic assessment of hemodynamic changes in the early post-burn period," *Burns*, 27(3):241-9 (2001).
- Troy, Tamara L., et al., "Optical properties of human skin in the near infrared wavelength range of 1000 to 2200nm," *Journal of Biomedical Optics*, vol. 6, No. 2, pp. 167-176 (Apr. 2001).
- Tsukahara, K., et al., "Dermal fluid translocation is an important determinant of the diurnal variation in human skin thickness," *British Journal of Dermatology*, vol. 145, pp. 590-596 (2001).
- Vescovi, Jason D., et al., "Evaluation of the BOD POD for estimating percentage body fat in a heterogeneous group of adult humans," *Eur J. Appl. Physiol.*, vol. 85, pp. 326-332 (2001).
- Wang, Zimian, et al., "Magnitude and variation of ratio of total body potassium to fat-free mass: a cellular level modeling study," *Am J. Physiol. Endocrinol. Metab.*, vol. 281, pp. E1-E7, (2001).
- Watson, Walter, "Hydration of fat-free body mass: new physiological modeling approach," *Am J. Physiol. Endocrinol. Metab.*, Letters to the Editor, vol. 278, pp. E752-E753 (2001).

- Attas, E. Michael, et al., "Near-IR Spectroscopic Imaging for Skin Hydration: The Long and the Short of It," *Biopolymers*, vol. 67, No. 2, pp. 96-106 (2002).
- Attas, M. et al., "Long-Wavelength Near-Infrared Spectroscopic Imaging for In-Vivo Skin Hydration Measurements," *Vibrational spectroscopy* (Feb. 28, 2002), vol. 28, No. 1, p. 37-43.
- Blank, T.B., et al., "Clinical Results from a Non-Invasive Blood Glucose Monitor," *Photonics West 2002 Meeting*, San Jose, California, Jan. 19-25, 2002 (25 pages).
- Chamney, Paul W., et al., "A new technique for establishing dry weight in hemodialysis patients via whole body bioimpedance," *Kidney International*, vol. 61, pp. 2250-2258 (2002).
- Drobin, Dan, et al., "Kinetics of Isotonic and Hypertonic Plasma volume Expanders," *Anesthesiology*, vol. 96, No. 6, pp. 1371-1380 (Jun. 2002).
- Endo, Yutaka, et al., "Changes in Blood Pressure and Muscle Sympathetic Nerve Activity during Water Drinking in Humans," *Japanese Journal of Physiology*, vol. 52, pp. 421-427 (2002).
- Haga, Henning A., et al., "Motor responses to stimulation during isoflurane anaesthesia in pigs," *Veterinary Anaesthesia and Analgesia*, vol. 29, pp. 69-75 (2002).
- Klaus, Stephan, et al., "Assessment of fluid balance by measurement of skin tissue thickness during clinical anaesthesia," *Clin. Physiol. & Func. Im.*, vol. 22, pp. 197-201 (2002).
- Meglinski, Igor V., et al., "Quantitative assessment of skin layers absorption and skin reflectance spectra simulation in the visible and near-infrared spectral regions," *Physiol. Meas.*, vol. 23, pp. 741-753, (2002).
- Perez-de-Sá, Valéria, et al., "Mild Hypothermia Has Minimal Effects on the Tolerance to Severe Progressive Normovolemic Anemia in Swine," *Anesthesiology*, Vol. 97, pp. 1189-1197 (2002).
- Ponec, Maria, et al., "Characterization of Reconstructed Skin Models," *Skin Pharmacol Appl Skin Physiol.*, vol. 15, Supplement 1, pp. 4-17, (2002).
- Querleux, B., et al., "Anatomy and physiology of subcutaneous adipose tissue by in vivo magnetic resonance imaging and spectroscopy: Relationships with sex and presence of cellulite," *Skin Research and Technology*, vol. 8, pp. 118-124 (2002).
- Van Bommel, Jasper, et al., "Intestinal and Cerebral Oxygenation during Severe Isovolemic Hemodilution and Subsequent Hyperoxic Ventilation in a Pig Model," *Anesthesiology*, vol. 97, No. 3, pp. 660-670 (Sep. 2002).
- Wong, William W., et al., "Evaluating body fat in girls and female adolescents: advantages and disadvantages of dual-energy X-ray absorptiometry," *Am J. Clin Nutr.*, vol. 76, pp. 384-389 (2002).
- Baković, Darija, et al., "Spleen volume and blood flow response to repeated breath-hold apneas," *J. Appl. Physiol.*, vol. 95, pp. 1460-1466 (2003).
- Bartok, Cynthia, et al., "Measurement of nutritional status in simulated microgravity by bioelectrical impedance spectroscopy," *J. Appl. Physiol.*, vol. 95, pp. 225-232 (2003).
- Bouwstra, Joke A., et al., "Water Distribution and Related Morphology in Human Stratum Corneum at Different Hydration Levels," *J. Invest Dermatol.*, vol. 150, pp. 750-758 (2003).
- Butte, Nancy F., et al., "Composition of gestational weight gain impacts maternal fat retention and infant birth weight," *Am J. Obstet Gynecol.*, vol. 189, pp. 1423-1432 (2003).
- Cloonan, Clifford C., "Don't Just Do Something, Stand There!: To Teach of not to Teach, That is the Question—Intravenous Fluid Resuscitation Training for Combat Lifesavers," *The Journal of TRAUMA, Injury, Infection, and Critical Care*, vol. 54, No. 5, pp. S20-S25 (May Supplement 2003).
- Cook, Lynda S., "IV Fluid Resuscitation," *Journal of Infusion Nursing*, vol. 26, No. 5, pp. 296-303 (Sep./Oct. 2003).
- Dey, D.K., et al., "Body composition estimated by bioelectric impedance in the Swedish elderly. Development of population-based prediction equation and reference values of fat-free mass and body fat for 70- and 75-y olds," *European Journal of Clinical Nutrition*, vol. 57, pp. 909-916 (2003).
- Farstad, M., et al., "Fluid extravasation during cardiopulmonary bypass in piglets—effects of hypothermia and different cooling protocols," *Acta Anaesthesiol Scand.*, vol. 47, pp. 397-406 (2003).
- Grandjean et al., "Hydration: issues for the 21<sup>st</sup> century", *Nutrition Reviews*, 61(8):261-271 (2003).
- Heise, H.M., et al., "Reflectance spectroscopy can quantify cutaneous haemoglobin oxygenation by oxygen uptake from the atmosphere after epidermal barrier disruption," *Skin Research and Technology*, vol. 9, pp. 295-298 (2003).
- Kasemsumran, Sumaporn, et al., "Simultaneous determination of human serum albumin,  $\gamma$ -globulin, and glucose in a phosphate buffer solution by near-infrared spectroscopy with moving window partial least-squares regression," *Analyst*, vol. 128, pp. 1471-1477 (2003).
- Kemming, G.I., et al., "Hyperoxic ventilation at the critical haematocrit," *Resuscitation*, vol. 56, pp. 289-297 (2003).
- Kurita, T., et al., "Comparison of isoflurane and propofol-fentanyl anaesthesia in a swine model of asphyxia," *British Journal of Anaesthesia*, vol. 91, No. 6, pp. 871-877 (2003).
- Laaksonen, DE, et al., "Changes in abdominal subcutaneous fat water content with rapid weight loss and long-term weight maintenance in abdominally obese men and women," *International Journal of Obesity*, vol. 27, pp. 677-683 (2003).
- Mao, Jinshu, et al., "Study of Novel Chitosan-gelatin artificial skin in vitro," *J. Miomed Mater Res.*, vol. 64, Part A, pp. 301-308 (2003).
- Mauran, P., et al., "Renal and hormonal responses to isotonic saline infusion after 3 days' dead-down tilt vs. supine and seated positions," *Acta Physiol. Scand.*, vol. 177, pp. 167-176, (2003).
- McHugh, Gerard, "Letter—Passive leg elevation and head-down tilt: effects and duration of changes," *Critical Care*, vol. 7, No. 3, p. 246 (Jun. 2003).
- Meglinski, I.V., et al., "Computer simulation of the skin reflectance spectra," *Computer Methods and Programs in Biomedicine*, vol. 70, pp. 179-186, (2003).
- Mendelsohn, Richard, et al., "Infrared microspectroscopic imaging maps the spatial distribution of exogenous molecules in skin," *Journal of Biomedical Optics*, vol. 8, No. 2, pp. 185-190 (Apr. 2003).
- Mentes, Janet C., et al., "Reducing Hydration=Linked events in Nursing Home Residents," *Clinical Nursing Research*, vol. 12, No. 3, pp. 210-225 (Aug. 2003).
- Merritt, Sean, et al., "Coregistration of diffuse optical spectroscopy and magnetic resonance imaging in a rat tumor model," *Applied Optics*, vol. 42, No. 16, pp. 2951-2959 (Jun. 2003).
- Parker, Lisa, et al., "Validity of Six Field and Laboratory Methods for Measurement of Body Composition in Boys," *Obesity Research*, vol. 11, No. 7, pp. 852-858 (Jul. 2003).
- Petäjä L., et al., "Dielectric constant of skin and subcutaneous fat to assess fluid changes after cardiac surgery", *Physiological Measurement*, 24: 3383-390, 2003.
- Rhodes, Andrew, et al., "Book Report—Haemodynamic monitoring in critically ill patients," *Critical Care*, vol. 8, p. 203 (2004).
- Richardson, Andrew D., et al., "Multivariate analyses of visible/near infrared (VIS/NIR) absorbance spectra reveal underlying spectral differences among dried, ground conifer needle samples from different growth environments," *New Phytologist*, vol. 161, pp. 291-301 (2003).
- Ritz, Patrick, "Chronic Cellular Dehydration in the Aged Patient," *Journal of Gerontology*, vol. 56A, No. 6, pp. M349-M352 (2001).
- Robinson, Martin P., et al., "A novel method of studying total body water content using a resonant cavity: experiments and numerical simulation," *Phys. Med. Biol.*, vol. 48, pp. 113-125, (2003).
- Sergi, Giuseppe, et al., "Changes in Fluid Compartments and Body Composition in Obese Women after Weight Loss Induced by Gastric Banding," *Ann. Nutr. Metab.*, vol. 47., pp. 152-157 (2003).
- Wang, Zimian, et al., "Magnitude and variation of fat-free mass density: a cellular level body composition modeling study," *Am J. Physiol. Endocrinol. Metab.*, vol. 284, pp. E267-E273 (2003).
- Windberger, U, et al., "Whole blood viscosity, plasma viscosity and erythrocyte aggregation in nine mammalian species; reference values and comparison of data," *Exp., Physiol.*, vol. 88, No. 3, pp. 431-440 (2003).
- Wolf, Martin, et al., "Absolute Frequency-Domain pulse Oximetry of the Brain: Methodology and Measurements," *Oxygen Transport to Tissue XXXIV*, Chapter 7, Dunn and Swartz, Kluwer Academic/Plenum Publishers, pp. 61-73 (2003).

- Ackland, G.L., et al., "Assessment of preoperative fluid depletion using bioimpedance analysis," *British Journal of Anaesthesia*, vol. 92, No. 1, pp. 134-136 (2004).
- Arimoto et al., "Non-contact skin moisture measurement based on near-infrared spectroscopy," *Applied Spectroscopy*, 58(12):1439-1445 (2004).
- Davidhiz, R., et al., "A review of the literature on how important water is to the world's elderly population," *International Nursing Review*, vol. 51, pp. 159-166 (2004).
- Dullenkopf, A., et al., "Non-invasive monitoring of haemoglobin concentration in paediatric surgical patients using near-infrared spectroscopy," *Anaesthesia*, vol. 59, pp. 453-458 (2004).
- Finlay, Jarod C., et al., "Hemoglobin oxygen saturations in phantoms and in vivo from measurements of steady-state diffuse reflectance at a single, short source-detector separation," *Medical Physics*, vol. 31, No. 7, pp. 1949-1959 (Jul. 2004).
- Hendriks, F.M., et al., "Influence of hydration and experimental length scale on the mechanical response of human skin in vivo, using optical coherence tomography," *Skin Research and Technology*, vol. 10, pp. 231-241 (2004).
- Hieda, I., et al., "Basic characteristics of the radio imaging method for biomedical application," *Medical Engineering & Physics*, vol. 26, pp. 431-437 (2004).
- Ikizler, T. Alp, et al., "Urea space and total body water measurements by stable isotopes in patients with acute renal failure," *Kidney International*, vol. 65, pp. 725-732 (2004).
- Isenring, E., et al., "Evaluation of foot-to-foot bioelectrical impedance analysis for the prediction of total body water in oncology outpatients receiving radiotherapy," *European Journal of Clinical Nutrition*, vol. 58, pp. 46-51 (2004).
- Jacobi, Ute, et al., "In vivo determination of skin surface topography using an optical 3D device," *Skin Research and Technology*, vol. 10, pp. 207-214 (2004).
- Kao, Bunsho, et al., "Evaluation of Cryogen Spray Cooling Exposure on In Vitro Model Human Skin," *Lasers in Surgery and Medicine*, vol. 34, pp. 146-154 (2004).
- Kyle, Urusula G., et al., Bioelectrical impedance analysis—part II: utilization in clinical practice, *Clinical Nutrition*, vol. 23, pp. 1430-1453 (2004).
- Lof, Marie, et al., "Hydration of fat-free mass in healthy women with special reference to the effect of pregnancy," *Am J. Clin. Nutr.*, vol. 80, pp. 960-965 (2004).
- Lowrie, Edmund G., "Urea space and body water," *Kidney Intl.*, vol. 66, No. 2, p. 868, Aug. 2004.
- Mirrashed, F., et al., "Pilot study of dermal and subcutaneous fat structures by MRI in individuals who differ in gender, BMI, and cellulite grading," *Skin Research and Technology*, vol. 10, pp. 161-168 (2004).
- Mirrashed, Fakhreh, et al., "In vivo morphological characterization of skin by MRI micro-imaging methods," *Skin Research and Technology*, vol. 10, pp. 149-160, (2004).
- Notingher, Ioan, et al., "Mid-infrared in vivo depth-profiling of topical chemicals on skin," *Skin Research and Technology*, vol. 10, pp. 113-121, (2004).
- Nouveau-Richard, S., et al., "In vivo epidermal thickness measurement: ultrasound vs. confocal imaging," *Skin Research and Technology*, vol. 10, pp. 136-140, (2004).
- Nuutinen, J., et al., "Validation of a new dielectric device to assess changes of tissue water in skin and subcutaneous fat," *Physiol. Meas.*, vol. 25, pp. 447-454, (2004).
- St-Onge, Marie-Pierre, et al., "Dual-energy X-ray absorptiometry lean soft tissue hydration: independent contributions of intra- and extracellular water," *Am J. Physiol. Endocrinol Metab.*, vol. 287, pp. E842-E847, Jul. 6, 2004.
- Schou, A. J., et al., "Methodological aspects of high-frequency ultrasound of skin in children," *Skin Research and Technology*, vol. 10, pp. 200-206, (2004).
- Stone, Darren A., et al., "Total body water measurements using resonant cavity perturbation techniques," *Phys. Med. Biol.*, vol. 49, pp. 1773-1788, (2004).
- Takiwaki, Hirotsugu, et al., "Analysis of the absorbance spectra of skin lesions as a helpful tool for detection of major pathophysiological changes," *Skin Research and Technology*, vol. 10, pp. 130-135 (2004).
- Van Kemenade, Patricia M., et al., "Do somatic forces play a role in the uptake of water by human skin?," *Skin Research and Technology*, vol. 10, pp. 109-112 (2004).
- Wang, Zimian, et al., "Body cell mass: model development and validation at the cellular level of body composition," *Am J. Physiol. Endocrinol. Metab.*, vol. 286, pp. E123-E128 (2004).
- Arimoto, Hidenobu, et al., "Depth profile of diffuse reflectance near-infrared spectroscopy for measurement of water content in skin," *Skin Research and Technology*, vol. 11, pp. 27-35 (2005).
- Burmeister, J.J., et al., "Spectroscopic considerations for noninvasive blood glucose measurements with near infrared spectroscopy," *LEOS Newsletter*, vol. 12, No. 2, 1998, <http://www.ieee.org/oroanizations/pubs/newsletters/leos/apr98/infrared.htm> (last accessed, Nov. 30, 2005).
- Haroun, D., et al., "Composition of the fat-free mass in obese and nonobese children: matched case—control analyses," *International Journal of Obesity*, vol. 29, pp. 29-36 (2005).
- Ivorra, Antoni, et al., "Bioimpedance dispersion width as a parameter to monitor living tissues," *Physiol. Meas.*, vol. 26, pp. S165-S173 (2005).
- Remote ICU Monitoring, *U.S. News & World Report*, pp. 45-61 (Aug. 1, 2005).
- Sarkar, Shubho R., et al., "Assessment of Body Composition in Long-Term Hemodialysis Patients: Rationale and Methodology," *Journal of Renal Nutrition*, vol. 15, No. 1, pp. 152-158 (Jan. 2005).
- Youcef-Toumi K., et al., "Noninvasive blood glucose analysis using near infrared absorption spectroscopy," MIT Home Automation and Healthcare Consortium, Progress Report No. 2-5, <http://darbelofflab.mit.edu/ProgressReports/HomeAutomation/Report2-5/Chapter04.pdf> (last accessed, Nov. 30, 2005).
- Garcia-Olmo, J., et al., "Advantages and disadvantages of multiple linear regression and partial least squares regression equations for the prediction of fatty acids," pp. 253-258 (undated).
- Wang, Zimian, et al., "Cellular-Level Body Composition Model—A New Approach to Studying Fat-free Mass Hydration," *Annals New York Academy of Sciences*, pp. 306-311 (undated).
- Lee, Jason C.S., et al., "Measurement of Percent Carboxyhemoglobin with Pulse-Oximetry Technique," *IEEE Engineering in Medicine & Biology Society 10<sup>th</sup> Annual International Conference*, CH2566-88, vol. 88, pp. 1781-1782 (1988).
- Lee, Jason C.S., et al., "Simultaneous Measurement of Percent Carboxyhemoglobin and Functional Oxygen Saturation," *IEEE Engineering in Medicine and Biology Society*, CH2770-6, vol. 89, pp. 1092-1093. (1989).
- Bongard, Frederic S., et al., "Continuous Dual Oximetry in Surgical critical care—Indications and Limitations," *Annals of Surgery*, vol. 216, No. 1, pp. 60-68 (1992).

\* cited by examiner




FIG. 1




FIG. 2




FIG. 3




FIG. 4




FIG. 5




FIG. 6




FIG. 7




FIG. 8




FIG. 9

# PULSE OXIMETRY SIGNAL CORRECTION USING NEAR INFRARED ABSORPTION BY WATER

## CROSS REFERENCE TO RELATED APPLICATIONS

This application is a continuation-in-part of the U.S. patent Ser. No. 10/797,475, entitled "PULSE OXIMETRY MOTION ARTIFACT REJECTION USING NEAR INFRARED ABSORPTION BY WATER", filed Mar. 9, 2004 now U.S. Pat. No. 7,277,741, which is herein incorporated by reference in its entirety.

## BACKGROUND OF THE INVENTION

### 1. Field of the Invention

The present invention relates generally to the processing of signals obtained from a medical diagnostic apparatus, such a pulse oximeter, using near infrared spectroscopy, to remove artifact or noise effects from the signal representative of a physiological parameter of interest.

### 2. Description of the Related Art

This section is intended to introduce the reader to various aspects of art that may be related to various aspects of the present invention, which are described and/or claimed below. This discussion is believed to be helpful in providing the reader with background information to facilitate a better understanding of the various aspects of the present invention. Accordingly, it should be understood that these statements are to be read in this light, and not as admissions of prior art.

A typical pulse oximeter measures two physiological parameters, percent oxygen saturation of arterial blood hemoglobin ( $SpO_2$  or sat) and pulse rate. Oxygen saturation can be estimated using various techniques. In one common technique, the photocurrent generated by the photo-detector is conditioned and processed to determine the ratio of modulation ratios (ratio of ratios) of the red to infrared signals. This modulation ratio has been observed to correlate well to arterial oxygen saturation. The pulse oximeters and sensors are empirically calibrated by measuring the modulation ratio over a range of in vivo measured arterial oxygen saturations ( $SaO_2$ ) on a set of patients, healthy volunteers, or animals. The observed correlation is used in an inverse manner to estimate blood oxygen saturation ( $SpO_2$ ) based on the measured value of modulation ratios of a patient. Most pulse oximeters extract the plethysmographic signal having first determined saturation or pulse rate, both of which are susceptible to interference.

In general, pulse oximetry takes advantage of the fact that in live human tissue, hemoglobin is a strong absorber of light between the wavelengths of 500 and 1100 nm. The pulsation of arterial blood through tissue is readily measurable, using light absorption by hemoglobin in this wavelength range. A graph of the arterial pulsation waveform as a function of time is referred to as an optical plethysmograph. The amplitude of the plethysmographic waveform varies as a function of the wavelength of the light used to measure it, as determined by the absorption properties of the blood pulsing through the arteries. By combining plethysmographic measurements at two different wavelength regions, where oxy- and deoxy-hemoglobin have different absorption coefficients, the oxygen saturation of arterial blood can be estimated. Typical wavelengths employed in commercial pulse oximeters are 660 and 890 nm.

It is known that rapid motion or application of pressure to a tissue site can have the effect of changing the optical prop-

erties being measured at or near that site. The amplitude of the optical signal changes associated with such events, known as motion artifacts, can easily be larger than that due to the arterial pulse. In practice, this can lead to inaccurate estimation of the percent oxygen saturation by pulse oximetry. Various techniques for addressing and removing undesired signal effects, including motion artifacts are known. As used herein, noise refers to signal portions that are undesired or are not directly related to changes in optical properties that are related to the arterial pulse, and which may include motion artifact. The optical signal through the tissue can be degraded by both noise and motion artifact. One source of noise is ambient light which reaches the light detector. Another source of noise is electromagnetic coupling from other electronic instruments. Motion of the patient also introduces noise and affects the signal. For example, the contact between the detector and the skin, or the emitter and the skin, can be temporarily disrupted when motion causes either to move away from the skin. In addition, since blood is a fluid, it responds differently than the surrounding tissue to inertial effects, thus resulting in momentary changes in volume at the point near which the oximeter probe is attached.

Motion artifact can degrade a pulse oximetry signal relied upon by a health care provider, without the provider's awareness. This is especially true if the monitoring of the patient is remote, the motion is too small to be observed, or the health care provider is watching the instrument or other parts of the patient, and not the sensor site. There are various known techniques for addressing the effects of noise and/or motion artifacts.

For example, U.S. Pat. No. 4,714,341 discloses a method for combining three wavelengths to detect the presence of motion. The wavelengths are used two at a time to separately compute the oxygen saturation percentage. When the oxygen saturation values computed using different wavelength combinations are in poor agreement, this is assumed to be caused by motion artifact, and the value is discarded. A disadvantage of this approach is that the agreement or lack thereof between the saturation values may or may not be due to motion artifact. In addition, this approach does not identify or remove the effects of motion artifact, but instead discards values that appear suspect

Another approach involves the filtering of pulse oximetry signals. However, filtering methods require assumptions about the properties of the artifact that do not always hold in practice. In addition, this approach does not measure the motion-induced signal.

U.S. Pat. No. 5,482,036 provides another approach, and describes a signal processing method for artifact reduction that functions when the artifact-related signal is associated with blood that is at a lower oxygen saturation than the arterial blood. Such a method relies on the generation of an artificial noise signal, which is combined with the physiological parameter to reduce the effect of the unknown noise signal. This approach for reducing the effects of artifact, without separately measuring the motion signal, is based on assumptions about the effect of motion on the plethysmographic signal. Assumptions may or may not be true, and many assumptions are invalid

Each of the known techniques for compensating for motion artifact has its own limitations and drawbacks. It is therefore desirable that a pulse oximetry system be designed which more effectively and accurately reports blood-oxygen levels during periods of motion. While many have attempted to isolate the effects of undesired signal portions, such as motion-induced artifacts, by making potentially invalid assumptions or by rejecting suspect estimates of desired sig-

nal values, there still remains a need for a deterministic identification, determination and measurement of artifact signals, to enable an accurate measurement of the desired signal values in the presence of undesired signal portions.

### SUMMARY

Certain aspects commensurate in scope with the originally claimed invention are set forth below. It should be understood that these aspects are presented merely to provide the reader with a brief summary of certain forms of the invention might take and that these aspects are not intended to limit the scope of the invention. Indeed, the invention may encompass a variety of aspects that may not be set forth below.

By measuring the artifact signal, the present technique allows motion artifacts to be separated from the plethysmographic signal without the limiting assumptions of prior known techniques. The present technique provides methods for measuring the motion signal associated with changes in tissue optical properties and using the measurement to compensate plethysmographic measurements made at other wavelengths.

In one embodiment, the present technique provides a method of determining a physiological parameter, including measuring an absorbance at a wavelength chosen to be primarily absorbed by water, and measuring an absorbance at a wavelength chosen to be primarily absorbed by hemoglobin. A ratio-of-ratios is calculated between these absorbances, and the ratio-of-ratios is used to identify motion noise.

In another embodiment, there is provided a system for the minimization of motion noise artifacts in pulse oximetry. This system uses a pulse oximeter monitor configured to analyze an absorbance signal that is primarily reflective of motion noise, an absorbance signal chosen to be primarily absorbed by hemoglobin, and another absorbance signal chosen to be primarily absorbed by hemoglobin. In another aspect, the monitor may be configured to use the first and second absorbances to obtain a metric that identifies the presence of motion noise.

In another embodiment, there is provided one or more tangible, machine readable media, containing code which controls the measurement of an absorbance at a wavelength chosen to be primarily absorbed by water and the measurement of an absorbance at a wavelength chosen to be primarily absorbed by hemoglobin. Under the control of this code, a ratio-of-ratios is calculated between these absorbances, and the ratio-of-ratios is used to identify motion noise.

### BRIEF DESCRIPTION OF THE DRAWINGS

Advantages of the invention may become apparent upon reading the following detailed description and upon reference to the drawings in which:

FIG. 1 is a block diagram of an exemplary oximeter, in accordance with aspects of the present technique.

FIG. 2 is a graph of the plethysmographic amplitude measured on the human ear as a function of wavelength.

FIG. 3 is a graph of absorption spectra of the principal components in human blood, scaled to typical physiological concentration.

FIG. 4 is a graph of absorption spectra of the principal components in human skin, scaled to typical physiological concentration.

FIG. 5 is a graph of absorption spectra of the principal components in human skin, scaled to equal volume-fraction concentration.

FIG. 6 is a graph of plethysmographs measured on a human ear at 4 different wavelengths of approximately 920, 1050, 1180 and 1300 nm respectively.

FIG. 7 is a graph of an exemplary plethysmographic artifact reduction by combining measurements at 2 near infrared wavelengths.

FIG. 8 is a flowchart of one approach to using a third wavelength to compensate for motion artifacts, in accordance with aspects of the present technique.

FIG. 9 is a graph of the oximetry results showing the error compensation using the recombination technique.

### DETAILED DESCRIPTION OF SPECIFIC EMBODIMENTS

One or more specific embodiments of the present invention will be described below. In an effort to provide a concise description of these embodiments, not all features of an actual implementation are described in the specification. It should be appreciated that in the development of any such actual implementation, as in any engineering or design project, numerous implementation-specific decisions must be made to achieve the developers' specific goals, such as compliance with system-related and business-related constraints, which may vary from one implementation to another. Moreover, it should be appreciated that such a development effort might be complex and time consuming, but would nevertheless be a routine undertaking of design, fabrication, and manufacture for those of ordinary skill having the benefit of this disclosure.

By measuring the artifact signal, the present technique allows motion artifact to be separated from the plethysmographic signal without the limiting assumptions of prior known techniques. The present technique provides methods for measuring the motion signal associated with changes in tissue optical properties and using the measurement to compensate plethysmographic measurements made at other wavelengths.

FIG. 1 is a block diagram of an exemplary pulse oximeter that may be configured to implement the embodiments of the present technique. The embodiments of the present technique can be a data processing algorithm that is executed by the microprocessor 122, described below. Light from light source 110 passes into patient tissue 112, and is scattered and detected by photodetector 114. A sensor 100 containing the light source and photodetector may also contain an encoder 116 which provides signals indicative of the wavelength of light source 110 to allow the oximeter to select appropriate calibration coefficients for calculating oxygen saturation. Encoder 116 may, for instance, be a resistor.

Sensor 100 is connected to a pulse oximeter 120. The oximeter includes a microprocessor 122 connected to an internal bus 124. Also connected to the bus are a RAM memory 126 and a display 128. A time processing unit (TPU) 130 provides timing control signals to light drive circuitry 132 which controls when light source 110 is illuminated, and if multiple light sources are used, the multiplexed timing for the different light sources. TPU 130 also controls the gating-in of signals from photodetector 114 through an amplifier 133 and a switching circuit 134. These signals are sampled at the proper time, depending upon which of multiple light sources is illuminated, if multiple light sources are used. The received signal is passed through an amplifier 136, a low pass filter 138, and an analog-to-digital converter 140. The digital data is then stored in a queued serial module (QSM) 142, for later downloading to RAM 126 as QSM 142 fills up. In one

embodiment, there may be multiple parallel paths of separate amplifiers, filters and A/D converters for multiple light wavelengths or spectra received.

Based on the value of the received signals corresponding to the light received by photodetector 114, microprocessor 122 will calculate the oxygen saturation using various algorithms. These algorithms require coefficients, which may be empirically determined, corresponding to, for example, the wavelengths of light used. These are stored in a ROM 146. In one embodiment of a two-wavelength system, the particular set of coefficients chosen for any pair of wavelength spectra is determined by the value indicated by encoder 116 corresponding to a particular light source in a particular sensor 100. In one embodiment, multiple resistor values may be assigned to select different sets of coefficients. In another embodiment, the same resistors are used to select from among the coefficients appropriate for an infrared source paired with either a near red source or far red source. The selection between whether the near red or far red set will be chosen can be selected with a control input from control inputs 154. Control inputs 154 may be, for instance, a switch on the pulse oximeter, a keyboard, or a port providing instructions from a remote host computer. Furthermore, any number of methods or algorithms may be used to determine a patient's pulse rate, oxygen saturation or any other desired physiological parameter. For example, the estimation of oxygen saturation using modulation ratios is described in U.S. Pat. No. 5,853,364, entitled "METHOD AND APPARATUS FOR ESTIMATING PHYSIOLOGICAL PARAMETERS USING MODEL-BASED ADAPTIVE FILTERING," issued Dec. 29, 1998, and U.S. Pat. No. 4,911,167, entitled "METHOD AND APPARATUS FOR DETECTING OPTICAL PULSES," issued Mar. 27, 1990, both of which are incorporated herein by reference in their entirety. Furthermore, the relationship between oxygen saturation and modulation ratio is further described in U.S. Pat. No. 5,645,059, entitled "MEDICAL SENSOR WITH MODULATED ENCODING SCHEME," issued Jul. 8, 1997 and incorporated herein by reference in its entirety.

Having described an exemplary pulse oximeter above, the methods for reducing noise, including motion artifact effects in the received signals, according to embodiments of the present technique, are described below.

FIG. 2 is a plot of the average plethysmographic amplitude as a function of wavelength measured through the earlobe of 36 subjects, and normalized to measurements at a wavelength of approximately 900 nm. Measurements, such as those shown in FIG. 2, reveal that the amplitude of the photoplethysmographic waveform diminishes as a function of wavelength between approximately 900 and 1300 nm, having a minimum value at approximately 1285 nm. The inventors herein have discovered that at wavelengths beyond approximately 900-920 nm, water, which is at much higher concentrations than hemoglobin, also becomes a major light absorber in tissue. FIG. 3 is a graph of some of the light absorbing components found in blood at typical concentrations, in units of absorbance in  $\text{cm}^{-1}$  vs. wavelength in nm. FIG. 3 shows that at approximately 1300 nm, blood should have only about 20% as much total absorbance as at 900 nm, with water being the dominant absorber. This theoretical model is in rough agreement with the pooled data shown in FIG. 2, where average plethysmographic amplitude was about  $\frac{1}{3}$  as much at 1300 nm as at 900 nm.

FIG. 4 is a graph of absorption spectra ( $\text{cm}^{-1}$ ) of the principal components in human skin, scaled to typical physiological concentration, as a function of wavelength in nm. This figure shows that the absorbance due to water has a peak

value at approximately 1180 nm, and that similar peaks are present for protein at slightly above 1150 and for lipids at approximately 1200 nm.

FIG. 5 is a graph of absorption spectra of the principal components in human skin, scaled to equal volume-fraction concentration. This figure shows that at approximately 1185 nm, the volume-fraction scaled absorbance for water, lipids and proteins are approximately equal.

While not being limited to any particular theory, the present inventors have, particularly in plethysmographic data from reflectance sensors, noted a weaker effect of water than would be theoretically predicted from absorption spectra. One potential reason for this effect lies in the fact that hemoglobin is largely confined to the blood vessels, whereas water is present at high concentrations both in the blood vessels and in the surrounding tissue. As a result, the pulse-induced expansion of arterial vessels through a tissue bed results in a localized increase in hemoglobin concentration, but only a small net change in water concentration. To the extent that the water concentration in the blood is equal to the water concentration in tissue, the change in light absorption by water is expected to approach zero.

The embodiments of the present technique exploit the finding that in spectral regions where hemoglobin absorbs weakly and water absorbs strongly, the plethysmograph is more sensitive to motion-related events that perturb tissue than arterial pulsation, compared with spectral regions where hemoglobin is a strong absorber and water is a weak absorber.

The weak magnitude of the plethysmograph in regions of strong water absorption is exploited to enable the separation of arterial-pulse-related signal from a motion artifact signal. By measuring the optical plethysmograph at a wavelength where water is the dominant absorber, the change in tissue optical properties associated with motion or pressure can be measured, with little interference from the underlying arterial pulsation. Plethysmographs at four near-infrared wavelengths measured through a human ear undergoing occasional motion are shown in FIG. 6, in absorbance units vs. scaled time (i.e., time per point is 43 ms). At approximately 920 nm, where hemoglobin absorption is strong and water absorption is weak, the plethysmograph contains regular arterial pulsations that are interrupted occasionally by motion-related events. As the wavelength is increased to approximately 1300 nm, where water is the predominant absorber, the arterial pulsations diminish and the measured signal becomes largely due to the motion-related events.

By combining the plethysmograph measured in a spectral region where water is the dominant absorber with a plethysmograph measured where blood is a major absorber, the motion-related signal can be selectively removed. FIG. 7 shows the plethysmograph of a human ear measured at approximately 920 nm, and the result of subtracting a portion of the plethysmograph measured at approximately 1180 nm from that measured at 920 nm. In particular, FIG. 7 shows the plethysmograph of a human ear measured at 920 nm, and the result of subtracting approximately 60% of the plethysmograph measured at approximately 1180 nm from that measured at approximately 920 nm. For different wavelength combinations, other multipliers are used based on the ratios of the absorbance of water as compared to that of oxy-hemoglobin or based on empirical determination(s).

By applying the same analysis to a diverse pool of 36 patients measured in a hospital setting, an average signal to noise increase of a factor of 2 of the plethysmograph at 910 nm was observed. By allowing the multiplier for the 1180 nm plethysmograph to vary between subjects, higher signal to noise improvements are achieved.

## Theoretical Model

The derivation below and the alternative description that follows demonstrate mechanisms by which the effect of motion-induced changes in optical scattering on a plethysmograph measured at one wavelength can be compensated by plethysmographic measurement at a second wavelength. These are provided as examples of techniques for reducing motion-induced optical changes, but are not the only mechanisms by which the present technique may function, and thus are not meant to limit the embodiments of the present technique.

A starting point for the analysis is the diffusion theory of light transport in tissue (for example, see "Diffusion Theory of Light Transport", Willem M. Star, in Optical-Thermal Response of Laser-Irradiated Tissue, edited by Ashley J. Welch and Martin J. C. van Gemert, Plenum Press, New York, 1995, pgs. 131-206). In the case where the transport-corrected scattering coefficient,  $\mu'_s$ , is much larger than the absorption coefficient,  $\mu_a$ , the diffuse intensity of light,  $I(\lambda)$ , measured at wavelength,  $\lambda$ , by a detector positioned a distance,  $l$ , away from a light source, can be described as follows (for example, see "Measurement of Blood Hematocrit by Dual-Wavelength Near-IR Photoplethysmography", Schmitt, J. M.; Guan-Xiong, G.; Miller, J., SPIE, Vol. 1641, 1992, pgs. 150-161):

$$I(\lambda) \propto \exp(-l\sqrt{3\mu_a(\lambda)\mu_s(\lambda)}) \quad (\text{eqn. 1})$$

For small changes in the absorption coefficient, such as those caused by arterial pulsation, the resulting change in intensity can be described by the derivative of intensity with respect to the absorption coefficient:

$$\frac{dI(\lambda)}{d\mu_a(\lambda)} = \frac{AC(\lambda)}{DC(\lambda)} = -l\sqrt{\frac{3\mu'_s(\lambda)}{4\mu_a(\lambda)}} \Delta V^{art} \mu_a^{art}(\lambda) \quad (\text{eqn. 2})$$

where  $\Delta V^{art}$  is the fractional volume change due to arterial pulsation,  $\mu_a^{art}$  is the absorption coefficient of the arterial blood under measurement,  $AC(\lambda)$  refers to the time varying portion of the optical signal and  $DC(\lambda)$  refers to the average or non-time varying portion of the optical signal.

The arterial oxygen saturation,  $SpO_2$ , is estimated if the AC-DC ratio described by equation 2 is measured at two wavelengths,  $\lambda_1$  and  $\lambda_2$ , that are chosen so that oxy- and deoxy-hemoglobin are readily differentiated (e.g.,  $\lambda_1 \sim$  approximately 660 nm,  $\lambda_2 \sim$  approximately 910 nm):

$$R = \frac{\frac{AC(\lambda_1)}{DC(\lambda_1)}}{\frac{AC(\lambda_2)}{DC(\lambda_2)}} = \Omega_{12} \frac{\mu_a^{art}(\lambda_1)}{\mu_a^{art}(\lambda_2)} \quad (\text{eqn. 3a})$$

where:

$$\Omega_{12} = \sqrt{\frac{\mu'_s(\lambda_1)\mu_a(\lambda_2)}{\mu'_s(\lambda_2)\mu_a(\lambda_1)}} \quad (\text{eqn. 3b})$$

from which:

$$SpO_2 = \frac{\mu_a^{HHb}(\lambda_1) - R\Omega_{12}^{-1}\mu_a^{HHb}(\lambda_2)}{R\Omega_{12}^{-1}(\mu_a^{O_2Hb}(\lambda_2) - \mu_a^{HHb}(\lambda_2)) + \mu_a^{HHb}(\lambda_1) - \mu_a^{O_2Hb}(\lambda_1)} \quad (\text{eqn. 3c})$$

where  $\mu_a^{HHb}$  and  $\mu_a^{O_2Hb}$  are the respective absorption coefficients for deoxy- and oxy-hemoglobin in arterial blood, and  $R$  is the ratio of the AC to DC ratios.

The effect of small changes in the scattering coefficient, such as may be brought about by compression of tissue or motion artifact, are as set forth below by eqn. 4:

$$\frac{dI(\lambda)}{d\mu'_s(\lambda)} = \frac{AC(\lambda)}{DC(\lambda)} = -l\sqrt{\frac{3\mu_a(\lambda)}{4\mu'_s(\lambda)}} \Delta\mu'_s(\lambda) \quad (\text{eqn. 4})$$

By measuring the AC-DC ratio at a third wavelength,  $\lambda_3$ , chosen so that the absorption due to hemoglobin is weak but the absorption due to water is strong, the effect of the motion-induced scattering change are removed from the AC-DC measurement at  $\lambda_2$  by subtracting the scaled AC-DC measurement at  $\lambda_3$ . The resulting motion-corrected plethysmograph,  $P$ , can be expressed as:

$$P = \frac{AC(\lambda_2)}{DC(\lambda_2)} - \frac{AC(\lambda_3)}{DC(\lambda_3)} \Omega_{23}^{-1} \quad (\text{eqn. 5a})$$

where:

$$\Omega_{23} = \sqrt{\frac{\mu'_s(\lambda_2)\mu_a(\lambda_3)}{\mu'_s(\lambda_3)\mu_a(\lambda_2)}} \quad (\text{eqn. 5b})$$

When the effects of arterial pulsation (equation 2) and motion artifact (equation 4) are additive, equation 5 is expanded as follows:

$$P = -l\sqrt{\frac{3\mu'_s(\lambda_2)}{4\mu_a(\lambda_2)}} \Delta V^{art} \mu_a^{art}(\lambda_2) - l\sqrt{\frac{3\mu_a(\lambda_2)}{4\mu'_s(\lambda_2)}} \Delta\mu'_s(\lambda_2) + \Omega_{23}^{-1} \left[ l\sqrt{\frac{3\mu'_s(\lambda_3)}{4\mu_a(\lambda_3)}} \Delta\mu'_s(\lambda_3) + l\sqrt{\frac{3\mu_a(\lambda_3)}{4\mu'_s(\lambda_3)}} \Delta\mu'_s(\lambda_3) \right] \quad (\text{eqn. 6})$$

When water absorption dominates the absorption of light by tissue at  $\lambda_3$ , and the water concentration in the arteries and surrounding tissue is nearly equal,  $\Delta\mu_a(\lambda_3)$  is approximately zero, and equation 6 simplifies to:

$$P = -l\sqrt{\frac{3\mu'_s(\lambda_2)}{4\mu_a(\lambda_2)}} \Delta V^{art} \mu_a^{art}(\lambda_2) \quad (\text{eqn. 7})$$

Equation 7 depends only on the effect of arterial pulsation at  $\lambda_2$ ; the effect of the motion artifact has been removed. In a similar manner the plethysmograph measured at  $\lambda_3$  may be used to remove the motion effects from the plethysmograph measured at  $\lambda_1$ . The corrected plethysmographs measured at

$\lambda_1$  and  $\lambda_2$  may then be combined and used to estimate oxygen saturation, as described, for example, by equation 3.

Several wavelengths in the range between approximately 900 and 1300 nm and more specifically in the range between approximately 1150 and 1350 nm have been tested and found effective at reducing motion-artifact from plethysmographs measured at approximately 910 nm. Wavelengths at the longer wavelength side of this range have the advantage of weaker absorbance of hemoglobin compared to that of water (for example, see FIGS. 3 and 4). However, wavelengths at the shorter end of this range have the advantage of reduced variation with changing tissue composition. As can be seen in FIG. 5, where the major components of tissue have been normalized to equal volume fraction, water, lipid, and non-hemoglobin protein all have approximately equal absorbance at approximately 1185 nm. Therefore the absorbance of tissue at approximately 1185 nm will vary little with changes in the relative concentration of these principal components.

It is known that the detection of light beyond approximately 1100 nm cannot readily be accomplished with the silicon (Si) detectors that are commonly employed in commercial oximeters. For example, the detector used to collect the data displayed in FIGS. 2-7 employed Indium Gallium Arsenide (InGaAs) as the photosensitive material. The most common type of InGaAs detectors are sensitive to light between approximately 800 and 1700 nm. Therefore, in a pulse oximeter designed in accordance with the embodiments of the present technique, with the conventional wavelengths of 660 and 890 nm, in addition to a new light source that emits at wavelengths that are absorbed strongly by water (such as approximately 1180 nm or approximately between 900-1400 nm), an additional detector(s) is used. One such scheme employs two detectors, one Si and one InGaAs, placed side-by-side. An alternative arrangement uses a collinear ("sandwich") detector containing separate Si and InGaAs layers, such as those commercially available, for example, from the Hamamatsu corporation. Yet another alternate arrangement uses two Si detectors placed symmetrically on either side of an InGaAs detector. Alternately, a germanium detector (Ge) is used as a substitute for the InGaAs detector.

An Implementation of Motion Noise Reduction Technique

A practical technique by which the effect of motion-induced changes in optical scattering on a plethysmograph measured at one wavelength may be reduced by plethysmographic measurement at a second wavelength is described below. This technique and the derivation above should be considered examples, and are not the only mechanisms by which the present technique may function. They are not meant to limit the embodiments of the present technique.

In one example, three wavelengths of light are used: a red wavelength at 660 nm, a near infrared (NIR) wavelength at 890 nm, and a NIR wavelength at 1300 nm. The first two wavelengths are both chosen to be primarily absorbed by hemoglobin, and the third wavelength is chosen to be primarily absorbed by water. After these wavelengths of light from the light source 110 (See FIG. 1) are passed through the tissue, the light is collected by a photodetector 114 (See FIG. 1) generating plethysmographs at each frequency.

Turning now to FIG. 8, in an exemplary embodiment, the red plethysmograph 200, the NIR plethysmograph 202, and the NIR (water) plethysmograph 204 are pre-processed (Block 210) prior to use. In this step, the waveforms are converted to a natural logarithm, and may be filtered to reduce noise, such as with a bandpass filter. The preprocessed plethysmographs 218 are then mathematically combined (Block 220) to identify periods of high motion noise and to generate plethysmographs with reduced motion noise.

In one such embodiment, the preprocessed NIR plethysmograph 214 and the preprocessed NIR (water) plethysmograph 216 are used to identify periods of high and/or low motion noise. This is performed by calculating a ratio-of-ratios,  $R_{1300,890}$ , between the absorbances at the NIR wavelength (890 nm) and the water wavelength (1300 nm) (See Eqn. 3a above for an example). The value of this ratio is less than 1.0 for periods when there are little or no motion artifacts, ranging from around 0.2 to 0.7 for most subjects. In one embodiment, a default value of 0.4 may be selected for initial use by an algorithm as described herein. In an exemplary embodiment,  $R_{1300,890}$  is calculated on two periods: once using three seconds of data for rapid detection of motion artifacts, and once using fifteen seconds of data for use in adjusting the combined weights, as discussed further below.

In one embodiment, a three step process is used to generate plethysmographs with reduced motion noise. The first step is to subtract fractions (F) of the preprocessed NIR (water) plethysmograph 216 (Preprocessed<sub>1300</sub>) from the preprocessed red plethysmograph 212 (Preprocessed<sub>660</sub>), and the preprocessed NIR plethysmograph 214 (Preprocessed<sub>890</sub>), to generate corrected waveforms:

$$\text{Corrected}_{890} = \text{Preprocessed}_{890} - F_{1300,890} * \text{Preprocessed}_{1300} \quad (\text{eqn. 8})$$

$$\text{Corrected}_{660} = \text{Preprocessed}_{660} - F_{1300,660} * \text{Preprocessed}_{1300} \quad (\text{eqn. 9})$$

In such an embodiment, the second step is to rescale the corrected waveforms to preserve the ratio-of-ratios 222 ( $R_{660,890}$ ) between the absorbance signals at the red (660 nm) and NIR (890 nm) wavelengths, so that the coefficients in eqn. 3b will not need to change. This is performed by estimating the fractions ( $C_{890}$  and  $C_{660}$ ) of the arterial pulse that were cancelled in Corrected<sub>890</sub> and Corrected<sub>660</sub>:

$$\text{Rescaled}_{890} = \text{Corrected}_{890} / (1.0 - C_{890}) \quad (\text{eqn. 10})$$

$$\text{Rescaled}_{660} = \text{Corrected}_{660} / (1.0 - C_{660}) \quad (\text{eqn. 11})$$

where:

$$C_{890} = R_{1300,890} * F_{1300,890} \quad (\text{eqn. 12})$$

$$C_{660} = R_{1300,890} * F_{1300,660} \quad (\text{eqn. 13})$$

$R_{660,890}$  222 may be supplied by the two wavelength oximetry algorithm 230, which calculates this value for determination of the oxygen saturation. An alternative method for rescaling Corrected<sub>660</sub> is to add a percentage of Corrected<sub>890</sub> to maintain a constant value for  $R_{660,890}$ :

$$\text{Rescaled}_{660} = \text{Corrected}_{660} + (F_{1300,660} / F_{1300,890}) * C_{890} * \text{Rescaled}_{890} \quad (\text{eqn. 14})$$

In this embodiment, the third step in generating plethysmographs with reduced motion noise is to adjust the fractions,  $F_{1300,660}$  and  $F_{1300,890}$ , of the NIR (water) plethysmograph 216 (Preprocessed<sub>1300</sub>) subtracted from the other two waveforms. Mathematical techniques may be selected that minimize the power, standard deviation, or amplitude of the resulting waveforms. Alternatively, techniques may be chosen that minimize the skewness of the derivative of the rescaled waveforms, or enhance some other recognized metric, or combination of metrics, of signal quality. The techniques for adjusting the fractions,  $F_{1300,660}$  and  $F_{1300,890}$  may be selected based on their efficacy in reducing saturation or pulse rate errors in representative sets of oximetry data that include motion artifact.

An example of one technique for calculating  $F_{1300,660}$  and  $F_{1300,890}$  is to use the summations given below:

$$F_{1300,890} = \frac{\sum \text{Preprocessed}_{d_{890,t}} \text{Preprocessed}_{d_{1300,t}}}{\sum \text{Preprocessed}_{d_{1300,t}^2}} - \frac{\sum R_{1300,890}(1.0 - C_{890}) \text{Rescaled}_{d_{890,t}^2}}{\sum \text{Preprocessed}_{d_{1300,t}^2}} \quad (\text{eqn. 15})$$

$$F_{1300,660} = \frac{\sum \text{Preprocessed}_{d_{660,t}} \text{Preprocessed}_{d_{1300,t}}}{\sum \text{Preprocessed}_{d_{1300,t}^2}} - \frac{\sum R_{1300,660}(1.0 - C_{660}) \text{Rescaled}_{d_{660,t}^2}}{\sum \text{Preprocessed}_{d_{1300,t}^2}} \quad (\text{eqn. 16})$$

These summations may be adequately represented by the approximations shown below:

$$F_{1300,890} = \frac{\sum \text{Preprocessed}_{d_{890,t}} \text{Preprocessed}_{d_{1300,t}}}{\sum \text{Preprocessed}_{d_{1300,t}^2}} - \epsilon \quad (\text{eqn. 17})$$

$$F_{1300,660} = \frac{\sum \text{Preprocessed}_{d_{660,t}} \text{Preprocessed}_{d_{1300,t}}}{\sum \text{Preprocessed}_{d_{1300,t}^2}} - \epsilon \quad (\text{eqn. 18})$$

In one implementation using these summations, a value of 0.03 has been found to work well for  $\epsilon$ . Alternatively, these summations may be approximated with infinite impulse response (IIR) filters. The values for  $F_{1300,660}$  and  $F_{1300,890}$  typically range from 0.6-0.9, and, in one embodiment, may be limited to range between 0.5-1.0 with a default value of 0.7. As will be understood by those skilled in the art, these constants may vary due to factors such as wavelength selection or sensor site or geometry.

As shown in FIG. 8, the plethysmographs which have been adjusted to reduce the noise motion artifacts,  $\text{Rescaled}_{890}$  226 and  $\text{Rescaled}_{660}$  224, are then used in a two wavelength algorithm 230 to calculate a value for oxygen saturation 232 and pulse rate 234. In one embodiment, the two wavelength algorithm 230 may be similar to that described in U.S. Pat. No. 5,853,364, but without the preprocessing that has already been done in block 210 of FIG. 8.

The improvements afforded by this technique are illustrated in the graph shown in FIG. 9. For this test, an oxygen sensor was attached to a test subject's ear lobe, which is highly susceptible to motion artifacts. As a control, another sensor was attached to a digit on the test subject. This second sensor was connected to an oximeter using a standard two wavelength algorithm. In the graph, the oxygen saturation 300 calculated from the preprocessed red plethysmograph 212 (See FIG. 8) and the preprocessed NIR plethysmograph 214 (See FIG. 8), using a standard two wavelength algorithm, showed a significant drop during periods of motion, such as nodding or shaking of the head. A control value 304 was calculated from the sensor located on the digit and remained steady. In contrast to the oxygen saturation 300 calculated from the uncombined preprocessed plethysmographs 218 (See FIG. 8), the oxygen saturation 302 calculated from the combined plethysmographs 228 (See FIG. 8) closely tracked the control. This is further illustrated by the % modulation curves at the bottom of the graph in FIG. 9. Prior to correction, the % modulation signal 306 shows the motion noise to be far larger than the % modulation signal 308 after the technique above is used.

In a larger test, a test group of 10 subjects using the standard two wavelength algorithm showed a pooled root-mean-square-difference (RMSD) in oxygen saturation of 4.55%, with some periods of 25% errors, between the moving sensor and a non-moving control. In contrast, the same data processed by the three wavelength algorithm discussed above showed a RMSD of 2.61% for the pooled subjects.

The values calculated in the algorithm detailed above may be used in a number of ways to display more accurate information to the user, while minimizing the load on the processor. For example, turning back to FIG. 8, during periods of very low motion artifacts, such as where  $R_{1300,890} < 0.85$ , the calculation above may be deactivated and the oxygen saturation 232 and pulse rate 234 calculated using the uncombined data from the preprocessed plethysmographs 218. Conversely, in such an embodiment, during periods of high motion artifacts, the calculation may remain active or may be activated and the oxygen saturation 232 and pulse rate 234 calculated using the combined data from the preprocessed plethysmographs 218. Alternatively, the value for  $R_{1300,890}$  could be used to gradually interpolate between values calculated from the preprocessed plethysmographs 218 and the combined plethysmographs 228. This technique would be useful in cases where the NIR (water) plethysmograph 204 was weaker, perhaps due to small pulse amplitude or a thick sensor site. In this case, the NIR (water) plethysmograph 204 would have a poor signal-to-noise ratio, and using the combined plethysmographs 228 only during periods of high motion artifacts would provide the most accurate information.

Additional useful modifications could take advantage of the extra signals provided by the technique. For example, additional preprocessing filters may be implemented prior to the calculation of the adjusted waveforms. In another example, various algorithms in the oximeter, such as sensor off detection, may continue to use the preprocessed plethysmographs 218, while the oxygen saturation and pulse rate calculation use the combined plethysmographs 228.

In addition, an alternative wavelength selection to the above-described augmentation to conventional pulse oximetry is an all-NIR pulse oximeter. An example of an all NIR oximeter is an oximeter employing light sources emitting at approximately 940, 1040, and 1180 nm used in conjunction with a single InGaAs detector. In addition to the advantage of requiring only one detector, the all-NIR implementation has advantages associated with the optical properties of tissue. The accuracy of measurements made using pulse oximetry depends, in part, on the extent to which the paths traveled by the different colors of light are the same. The mean path length and penetration depth of light at a particular wavelength traveling through tissue is strongly affected by the absorption and scattering coefficients of tissue at that wavelength. In conventional pulse oximetry, in order to achieve the same mean path length and penetration depth at two wavelengths, the scattering and absorption coefficients at the two wavelengths need to be matched. The scattering of light by tissue decreases rapidly as a function of wavelength, with the result that the scattering properties of tissue at approximately 940, 1040, and 1180 nm will be more closely matched than the scattering properties of tissue at a combination of both visible and NIR wavelengths such as approximately 660, 890, and 1180 nm, for reasons discussed below. The absorption properties of oxy- and deoxy-hemoglobin are such that at high oxygen saturation values the net (i.e., combined effects of oxy and deoxy) absorption coefficient due to hemoglobin will be matched reasonably well at 660 nm and 940 nm. However, as oxygen saturation values decrease, the high

absorption coefficient of deoxy-hemoglobin at approximately 660 nm will result in an increasingly strong mismatch between the net absorption coefficient of hemoglobin at approximately 660 and approximately 940 nm. The net absorption coefficients of hemoglobin at approximately 940 and approximately 1040 nm, will be more closely matched than at approximately 660 and approximately 940 nm, over the full range of measurable oxygen saturation values.

The choice of the wavelength used to measure the motion-artifact signal depends partially on the need for matching the optical path length to that of the signals to be corrected. Beyond approximately 950 nm, the absorption coefficient of water, protein, and non-hemoglobin protein, in addition to that of hemoglobin needs to be considered in order to achieve close matching of path lengths. Although about 1300 nm is a currently preferred wavelength for measuring the motion-artifact signal, other alternative wavelength values are also effective, for example, wavelengths between approximately 1050 and 1400 nm and between approximately 1500 and 1850 nm.

The embodiments of the present technique may be practiced by placing the optical components directly at the tissue interface, or alternatively, by transporting the light to and from the tissue with fiber optics. The former implementation has the advantage of more efficient delivery and collection of the light, whereas the latter implementation has the advantages of being less costly. The less costly solution is enabled by the fact that when employing fiber optic delivery, the light sources and detectors can reside in the monitor as opposed to the sensor, and considering that such components may be more expensive than the fiber, this will result in a less expensive device.

As will be understood by those skilled in the art, other equivalent or alternative methods for the measurement of motion artifact signal associated with changes in tissue optical properties, and using the measurement to compensate plethysmographic measurements made at other wavelengths, according to the embodiments of the present technique can be envisioned without departing from the essential characteristics thereof. For example, a combination of visible and NIR or an all NIR wavelength combination may be used to make the measurements. Moreover, individuals skilled in the art of near-infrared spectroscopy would recognize that additional terms can be added to the algorithms used herein to incorporate reflectance measurements made at additional wavelengths and thus improve accuracy further. Also, light sources or light emission optics other than LED's including and not limited to incandescent light and narrowband light sources appropriately tuned to the desired wavelengths and associated light detection optics may be placed near the tissue location or may be positioned within a remote unit; and which deliver light to and receive light from the tissue location via optical fibers. Additionally, sensor arrangements functioning in a back-scattering or a reflection mode to make optical measurements of reflectances, as well as other embodiments, such as those working in a forward-scattering or a transmission mode may be used to make these measurements.

While the invention may be susceptible to various modifications and alternative forms, specific embodiments have been shown by way of example in the drawings and have been described in detail herein. However, it should be understood that the invention is not intended to be limited to the particular forms disclosed. Rather, the invention is to cover all modifications, equivalents, and alternatives falling within the spirit and scope of the invention as defined by the following appended claims.

What is claimed is:

1. A method of determining a physiological parameter, comprising:
  - obtaining a first absorbance at a first wavelength, wherein the first wavelength is chosen to be primarily absorbed by water;
  - obtaining a second absorbance at a second wavelength, wherein the second wavelength is chosen to be primarily absorbed by hemoglobin;
  - estimating a first ratio-of-ratios between the first absorbance and the second absorbance; and using the first ratio-of-ratios to identify motion noise.
2. The method of claim 1, comprising
  - obtaining a third absorbance at a third wavelength, wherein the third wavelength is chosen to be primarily absorbed by hemoglobin;
  - calculating a corrected second absorbance by subtracting a first fraction of the first absorbance from the second absorbance; and
  - calculating a corrected third absorbance by subtracting a second fraction of the first absorbance from the third absorbance.
3. The method of claim 2, comprising:
  - rescaling the corrected second absorbance to obtain a rescaled second absorbance; and
  - rescaling the corrected third absorbance to obtain a rescaled third absorbance,
 wherein the rescaling of the corrected second absorbance and the corrected third absorbance maintains a ratio-of-ratios between the rescaled second absorbance and the rescaled third absorbance that is independent of changes in the subtracted first and second fractions of the first absorbance.
4. The method of claim 3, further comprising:
  - calculating a first oxygen saturation value using a two wavelength algorithm based on the rescaled second absorbance and the rescaled third absorbance; and
  - calculating a second oxygen saturation value using a two wavelength algorithm based on the second absorbance and the third absorbance.
5. The method of claim 4, further comprising:
  - displaying the first oxygen saturation value during periods of high motion noise;
  - displaying the second oxygen saturation value during periods of low motion noise;
  - calculating an intermediate oxygen saturation value using the first and second oxygen saturation values; and
  - displaying the intermediate oxygen saturation value during periods of intermediate motion noise.
6. The method of claim 3, further comprising:
  - calculating an oxygen saturation value using a two wavelength algorithm based on the rescaled second absorbance and the rescaled third absorbance.
7. The method of claim 3, further comprising:
  - calculating a first pulse rate value using the rescaled second absorbance; and
  - calculating a second pulse rate value using the second absorbance.
8. The method of claim 7, further comprising:
  - displaying the first pulse rate value during periods of high motion noise;
  - displaying the second pulse rate value during periods of low motion noise;
  - calculating an intermediate pulse rate value using the first and second pulse rate values; and
  - displaying the intermediate pulse rate value during periods of intermediate motion noise.

15

9. The method of claim 3, further comprising:

calculating a pulse value using the corrected second absorbance or the rescaled second absorbance.

10. The method of claim 2, further comprising using an algorithm to calculate the value of the first fraction and the value of the second fraction.

11. The method of claim 10, wherein the algorithm minimizes the standard deviation of a resulting waveform.

12. The method of claim 10, wherein the algorithm minimizes the power of a resulting waveform.

13. The method of claim 10, wherein the algorithm minimizes the skewness of a resulting waveform.

14. The method of claim 10, wherein the algorithm is chosen to reduce errors in the calculation of oxygen saturation and pulse rate.

15. The method of claim 10, wherein the algorithm comprises an infinite impulse response (IIR) filter.

16. A system for signal correction in pulse oximetry, the system comprising:

a pulse oximeter monitor, wherein the pulse oximeter monitor is configured to analyze a first absorbance signal substantially corresponding to motion noise, to analyze a second absorbance signal chosen to be primarily absorbed by hemoglobin, to analyze a third absorbance signal chosen to be primarily absorbed by hemoglobin, to adjust the second absorbance signal to compensate for noise by subtracting a first fraction of the first absorbance signal from the second absorbance signal to obtain a corrected second absorbance signal, and to adjust the third absorbance signal to compensate for noise by subtracting a second fraction of the first absorbance signal from the third absorbance signal to obtain a corrected third absorbance signal, wherein the pulse oximeter monitor is further configured to rescale the corrected second absorbance signal to obtain a rescaled second absorbance signal, and to rescale the corrected third absorbance signal to obtain a rescaled third absorbance signal, wherein the rescaling maintains a ratio-of-ratios between the rescaled second absorbance signal and the rescaled third absorbance signal that is independent of changes in the subtracted first and section fractions of the first absorbance.

17. The system of claim 16, further comprising a pulse oximetry sensor, wherein the sensor comprises optical emitters and detectors configured to emit and detect light at a first wavelength chosen to be primarily absorbed by water, to emit and detect light at a second wavelength chosen to be primarily absorbed by hemoglobin, and to emit and detect light at a third wavelength chosen to be primarily absorbed by hemoglobin, wherein the first absorbance signal analyzed by the pulse oximeter monitor is based on the first wavelength emitted and detected by the pulse oximetry sensor.

18. The system of claim 16, wherein the pulse oximeter monitor is configured to combine the first absorbance signal and the second absorbance signal to obtain a metric that identifies the presence of motion noise.

19. The system of claim 16, wherein the pulse oximeter monitor is configured to calculate a first oxygen saturation value using the rescaled second absorbance signal and the rescaled third absorbance signal, and to calculate a second oxygen saturation value using the first absorbance signal and the second absorbance signal.

20. The system of claim 19, wherein the pulse oximeter monitor is configured to display the first oxygen saturation value during periods of high motion noise, to display the second oxygen saturation value during periods of low motion noise, to calculate an intermediate oxygen saturation value

16

using the first oxygen saturation value and the second oxygen saturation value, and to display the intermediate oxygen saturation value during periods of intermediate motion noise.

21. The system of claim 16, wherein the pulse oximeter monitor is configured to calculate an oxygen saturation value using the rescaled second absorbance signal and the rescaled third absorbance signal.

22. The system of claim 16, wherein the pulse oximeter monitor is configured to calculate a first pulse rate value from the rescaled second absorbance signal and the rescaled third absorbance signal, and to calculate a second pulse rate value from the second absorbance signal and the third absorbance signal.

23. The system of claim 22, wherein the pulse oximeter monitor is configured to display the first pulse rate value during periods of high motion noise, to display the second pulse rate value during periods of low motion noise, to calculate an intermediate pulse rate value using the first pulse rate value and the second pulse rate value, and to display the intermediate pulse rate value during periods of intermediate motion noise.

24. The system of claim 16, wherein the pulse oximeter monitor is configured to calculate a pulse rate value from the rescaled or corrected second and third absorbance signals.

25. One or more tangible, machine readable media, comprising code executable to perform the acts of:

obtaining a first absorbance at a first wavelength, wherein the first wavelength is chosen to be primarily absorbed by water;

obtaining a second absorbance at a second wavelength, wherein the second wavelength is chosen to be primarily absorbed by hemoglobin;

estimating a first ratio-of-ratios between the first absorbance and the second absorbance; and

using the first ratio-of-ratios to identify motion noise.

26. The one or more tangible, machine readable media of claim 25, further comprising code executable to perform the acts of:

obtaining a third absorbance at a third wavelength, wherein the third wavelength is chosen to be primarily absorbed by hemoglobin;

calculating a corrected second absorbance by subtracting a first fraction of the first absorbance from the second absorbance; and

calculating a corrected third absorbance by subtracting a second fraction of the first absorbance from the third absorbance.

27. The one or more tangible, machine readable media of claim 26, further comprising code executable to perform the acts of:

rescaling the corrected second absorbance to obtain a rescaled second absorbance; and

rescaling the corrected third absorbance to obtain a rescaled third absorbance,

wherein the rescaling of the corrected second absorbance and the corrected third absorbance maintains a ratio-of-ratios between the rescaled second absorbance and the rescaled third absorbance that is independent of changes in the subtracted first and section fractions of the first absorbance.

28. The one or more tangible, machine readable media of claim 27, further comprising code executable to perform the acts of:

calculating a first oxygen saturation value using a two wavelength algorithm based on the rescaled second absorbance and the rescaled third absorbance; and

calculating a second oxygen saturation value using a two wavelength algorithm based on the second absorbance and the third absorbance.

29. The one or more tangible, machine readable media of claim 28, further comprising code executable to perform the acts of:

displaying the first oxygen saturation value during periods of high motion noise;

displaying the second oxygen saturation value during periods of low motion noise;

calculating an intermediate oxygen saturation value using the first oxygen saturation value and the second oxygen saturation value; and

displaying the intermediate oxygen saturation value during periods of intermediate motion noise.

30. The one or more tangible, machine readable media of claim 27, further comprising code executable to perform the acts of:

calculating a first pulse rate value using the rescaled second absorbance; and

calculating a second pulse rate value using the second absorbance.

31. The one or more tangible, machine readable media of claim 30, further comprising code executable to perform the acts of:

displaying the first pulse rate value during periods of high motion noise;

displaying the second pulse rate value during periods of low motion noise;

calculating an intermediate pulse rate value using the first pulse rate value and the second pulse rate value; and

displaying the intermediate pulse rate value during periods of intermediate motion noise.

32. The one or more tangible, machine readable media of claim 27, further comprising code executable to perform the acts of:

calculating a pulse rate value using the rescaled second absorbance or the corrected second absorbance.

33. The one or more tangible, machine readable media of claim 26, further comprising code executable to perform the acts of:

Calculating an oxygen saturation value using a two wavelength algorithm based on the rescaled second absorbance and the rescaled third absorbance.

34. The one or more tangible, machine readable media of claim 26, further comprising code executable to perform the act of using an algorithm to calculate the value of the first fraction and the value of the second fraction.

35. The one or more tangible, machine readable media of claim 34, wherein the algorithm minimizes the standard deviation of a resulting waveform.

36. The one or more tangible, machine readable media of claim 34, wherein the algorithm minimizes the power of a resulting waveform.

37. The one or more tangible, machine readable media of claim 34, wherein the algorithm minimizes the skewness of a resulting waveform.

38. The one or more tangible, machine readable media of claim 34, wherein the algorithm is chosen to reduce errors in the calculation of oxygen saturation and pulse rate.

39. The one or more tangible, machine readable media of claim 34, wherein the algorithm is an infinite impulse response (IIR) filter.

\* \* \* \* \*

专利名称(译)	利用水的近红外吸收进行脉搏血氧饱和度信号校正		
公开(公告)号	<a href="#">US8175670</a>	公开(公告)日	2012-05-08
申请号	US11/521960	申请日	2006-09-15
[标]申请(专利权)人(译)	BAKER JR CLARK - [R] 喀斯特爱德华 HOARAU CARINE		
申请(专利权)人(译)	BAKER CLARK - [R JR] 喀斯特爱德华 HOARAU CARINE		
当前申请(专利权)人(译)	COVIDIEN LP		
[标]发明人	BAKER JR CLARK R KARST EDWARD HOARAU CARINE		
发明人	BAKER, JR., CLARK R. KARST, EDWARD HOARAU, CARINE		
IPC分类号	A61B5/1455 A61B5/02 A61B5/00 A61B5/024		
CPC分类号	A61B5/14551 A61B5/7207 A61B5/02416		
其他公开文献	US20070106137A1		
外部链接	<a href="#">Espacenet</a> <a href="#">USPTO</a>		

摘要(译)

一种用于测量生理参数的方法和系统，包括收集第一波长的第一吸光度，选择为主要被水吸收；收集第二波长的第二吸光度，选择主要被血红蛋白吸收；并且组合第一信号和第二信号以产生组合的体积描记器信号，其与由运动相关的干扰引起的噪声成比例地降低。

