



US 20080183092A1

(19) **United States**  
(12) **Patent Application Publication**  
Smith et al.

(10) **Pub. No.: US 2008/0183092 A1**  
(43) **Pub. Date: Jul. 31, 2008**

(54) **APPARATUS FOR MONITORING FETAL HEART-BEAT**

(30) **Foreign Application Priority Data**

(75) Inventors: **Mark John Smith**, Worcestershire (GB); **Richard William Penney**, Worcestershire (GB)

Oct. 3, 2001 (GB) ..... 0123772.6  
Sep. 27, 2002 (GB) ..... PCT/GB02/04410

Correspondence Address:  
**MCDONNELL BOEHNEN HULBERT & BERGHOFF LLP**  
**300 S. WACKER DRIVE, 32ND FLOOR**  
**CHICAGO, IL 60606**

**Publication Classification**

(51) **Int. Cl.**  
*A61B 5/0444* (2006.01)  
(52) **U.S. Cl.** ..... **600/511**

(57) **ABSTRACT**

Apparatus for monitoring a fetal heart-beat is able to extract one or more fetal electrocardiograms (fECGs) from a composite signal detected at the abdomen of a pregnant woman. The apparatus includes specific low-noise components: a plurality of low-noise electrodes (1-3, R) for placement on the abdomen during pregnancy and a low-noise signal recording and processing means (34). Screened leads are also used, as required. Signals indicative of voltages developed between each abdominal electrode (1-3) and a reference electrode (R) are recorded in a plurality of signal channels. Data within each channel is digitised and processed in order to generate a plurality of separated source signals, at least one of which relates to the fECG of a single fetus. Single-channel fECGs may be reconstructed using the source signals identified as belonging to the same fetus.


(73) Assignee: **QINETIQ LIMITED**



(21) Appl. No.: **12/026,267**

(22) Filed: **Feb. 5, 2008**

**Related U.S. Application Data**

(63) Continuation of application No. 10/490,611, filed on Mar. 24, 2004, filed as application No. PCT/GB02/04410 on Sep. 27, 2002.








Figure 3




Figure 4




Figure 5a




Figure 5b




Figure 6




Figure 7







Figure 8

# QinetiQ Fetal ECG

SCUA v1.53:  
 Name GST52 Date: xx/xx/xxxx  
 Singleton Gestation 24 weeks  
 No. sensors 12 Length of dataset 60 s  
 Data used: 60 s - From 0 s to 60 s  
 Passband from 5 Hz to 150 Hz  
 Data quality Very good  
 Separation Good

## Comments

Mean heartrate: 143 bpm  
 Max heartrate: 150 bpm  
 Min heartrate: 114 bpm



Rhythm strip trace 1: from 16 secs to 22 ses.



Rhythm strip trace 2: from 16 secs to 22 ses.



Figure 9

## APPARATUS FOR MONITORING FETAL HEART-BEAT

**[0001]** This is a continuation of U.S. application Ser. No. 10/490,611, filed on Mar. 24, 2004, which is a 371 application of PCT/GB02/04410, filed on Sep. 27, 2002, which claims the benefit of the filing date of United Kingdom 0123772.6 filed on Oct. 3, 2001, the specifications of each of which are incorporated herein by reference.

### BACKGROUND OF THE INVENTION

**[0002]** (1) Field of the Invention

**[0003]** This invention relates to the field of medical electrocardiographs, in particular to those adapted to monitor non-invasively the heart-beat of an unborn fetus.

**[0004]** (2) Description of the Art

**[0005]** The electrocardiogram (ECG) is a key tool in the diagnosis of heart disease and abnormalities in both children and adults. The heart-beat is instigated and controlled by electrical conduction through the heart. In a healthy human a certain characteristic sequence of electrical impulses, which is repeatedly cycled, controls the heart-beat. If voltage sensors are placed on a patient's chest then the electrical activity, and its variation from beat to beat, can be detected and displayed. This is the basis behind the ECG. Alternative detection tools such as Magnetic Resonance Imaging (MRI) and Ultrasound may complement the ECG in providing an understanding of the physiology of the heart, but it is the ECG that importantly indicates the detail of the cardiac rhythm.

**[0006]** The unique detection capability provided by the ECG gives it a very important role to play in the diagnosis and management of abnormal cardiac rhythms, and accordingly it is widely used in hospitals throughout the world. An ECG can, for example, help with the diagnosis of the causes of chest pain and breathlessness and is crucial to the proper use of thrombolysis in treating myocardial infarction. ECG equipment is moreover generally cheaper, more portable and simpler to use than apparatus used for alternative monitoring techniques such as MRI and ultrasound. It therefore does not require such highly trained personnel to operate it and ECG readings can be taken over extended periods (e.g. 24 hours) even whilst the subject is ambulatory.

**[0007]** It is anticipated that the ability to acquire routinely a fetal ECG (fECG) would have at least a similar clinical value to the unborn child as the ECG currently has to children and adults. In addition it is expected to contribute to the early detection and monitoring of ischaemia and heart arrhythmias and abnormalities that might lead to premature death or long-term damage. Unfortunately, as in any non-invasive fetal observation technique, the situation is complicated by the need to extract weaker fetal information from a composite signal containing data relating to both the fetus and the mother. It is significant that subtle cardiovascular indications of dysfunction may be present well before the 20<sup>th</sup> week of pregnancy, but these cannot be detected until further into the gestation period by currently-used techniques.

**[0008]** The problem in multiple pregnancies is further compounded not only by the need to acquire at least two fetal indications from the composite signal, but also by the fact that accurate detection of cardiac development is of far greater importance to this high risk group. Multiple pregnancies face substantially increased risks of perinatal mortality and morbidity compared with singletons and it is therefore envisaged

that antenatal acquisition of their fECGs will have correspondingly greater clinical application. For example monozygous twins face a risk of 3.6% of congenital heart disease. Twin-Twin Transfusion Syndrome (TTTS) is a complication unique to monochorionic twins in which one fetus (the recipient) receives too much blood through the shared placenta at the expense of the other (the donor). Cardiac overload is a typical feature exhibited by the recipient fetus, which would be readily detectable by ECG. TTTS complicates 15% of monochorionic twins and accounts for 17% of all perinatal mortality in twins. In addition, growth restriction, which complicates 5% of singleton pregnancies, affects 25% of dichorionic and 42% of monochorionic twins. In comparison with traditional methods of fetal monitoring, it is envisaged that fECG will offer earlier detection of ischaemia or cardiac compromise, and will provide useful information on the physiological responses to divergent blood volume loads found in TTTS.

**[0009]** Current fetal monitoring tools such as Doppler ultrasound and cardiotocography (CTG) lack sensitivity and specificity. Diagnostic use of CTG appears to have no significant effect on perinatal mortality or morbidity in high-risk pregnancies. Indeed in the Cochrane database there is a trend towards increased perinatal mortality (odds ratio 2.85, 95% confidence interval 0.99 to 7.12) in those assessed by CTG. Although the use of Doppler ultrasound in high-risk pregnancies appears to improve a number of obstetric care outcomes and appears promising in helping to reduce perinatal deaths, it has not been shown to be of benefit in low-risk populations.

**[0010]** The limitations of these techniques, combined with the potential benefits that are offered by fECG has prompted the more recent development of both invasive and non-invasive techniques to record fECGs.

**[0011]** Invasive techniques involve the direct attachment of an electrode to the scalp of the baby during labour. The use of scalp electrodes has however increased risks of perinatal infection and so, despite a demonstrable reduction in birth asphyxia, is limited in many countries, including the UK. Such techniques do serve however to indicate the importance of acquiring a detailed fECG recording. In *The Lancet*, Vol 358, pp 534-8 (2001), Amer-Wahlin et al showed that monitoring the ST segment of a fetal heart pulse during labour provided a useful diagnostic indicator of hypoxia. There is therefore a demonstrable need for the ability to obtain reliably this level of fECG detail non-invasively.

**[0012]** A variety of non-invasive techniques have been tested. It was shown as long ago as 1906 that the electrical activity of the heart of the unborn fetus can be detected non-invasively at the surface of the maternal abdomen. Electrodes can therefore be used to detect a composite signal containing information relating to the cardiac activity of both the mother and fetus. However the amplitude of the maternal signal is typically around 100  $\mu$ V or more at the abdomen, whereas that of the fetus is of only 10-20  $\mu$ V, and may even be less depending on the position of the electrodes on the abdomen and on the presentation and gestation of the fetus. Clearly, to provide useful information, it is necessary to separate the detailed fECG from the composite signal in which it is significantly masked by both the maternal signal and noise.

**[0013]** Many attempts have been made to obtain a meaningful fECG signal from the composite signal detected at the maternal abdomen. D. Callaerts in his PhD thesis "Signal separation methods based on Singular Value Decomposition and their application to real-time extraction of the fetal elec-

trocadiogram from cutaneous recordings”, Katholieke Universiteit Leuven, December 1989 describes one method of processing data collected using dedicated hardware equipment. This technique however requires electrodes to be placed along three mutually orthogonal axes intersecting at the location of the fetal heart, in addition to those required to be placed on the mother’s chest. This not only means that electrodes must be placed on the mother’s back as well as her abdomen, adding to her discomfort, but it also requires an a priori knowledge of the position of the fetus e.g. acquired by use of ultrasound.

**[0014]** De Lathauwer et al. in “Fetal electrocardiogram by Blind Source Subspace Separation” IEEE Trans. Biomed. Eng. 47(5) 567-572 (2000), apply an improved algorithm to data previously recorded by Callaerts and co-workers. Callaerts reported analysis of the composite data using an algorithm known as Singular Value Decomposition (SVD). De Lathauwer demonstrated that Blind Signal Separation (BSS) algorithms based on Independent Component Analysis (ICA) can be more effective in separating out the fetal signal in those cases for which the earlier SVD technique was found to work. Potential insights on or improvements to the recording equipment or its configuration were not addressed.

**[0015]** Spencer et al. in “Antenatal abdominal fetal electrocardiogram recording—preliminary results of a compact and portable monitor” (Abstract) XVI FIGO World Congress of Gynaecology and Obstetrics, Washington (2000) report results for fetal separation again using purpose-built equipment. However the quality of the separated signals is very poor, with the interval between successive beats in the fECG being discernable in only 59% of cases. Given the level of noise, it is highly unlikely that this equipment could be adapted for effective monitoring of detailed fECGs in all but singleton pregnancies. Indeed the emphasis of such devices is not upon acquisition of detail at all, but only on the determination of the fetal heart rate. This is a far simpler measurement to extract. Based on what is known about ECGs taken in children and adults it is expected that more detailed examination of the cardiac rhythm will yield further important diagnostic parameters.

**[0016]** In summary therefore currently-available non-invasive techniques for obtaining the fetal electrocardiogram are generally only of limited effect, even in singleton pregnancies, and/or require a complicated arrangement of electrodes for the recording, making it impossible to be taken by all but highly trained staff. There is a perceived need to provide apparatus for the recording of fetal ECGs from electrodes on the maternal abdomen which is capable of reliably extracting the detailed fetal signal and which admits of more straightforward application than the aforementioned prior art apparatus. It is an object of this invention to provide such apparatus. In particular, it is an object of this invention to provide fECG equipment which is capable of extracting fECGs during multiple pregnancies; this being one of the high-risk groups most likely to gain from the kind of information that a detailed fECG might have to offer.

#### SUMMARY OF THE INVENTION

**[0017]** Accordingly the present invention provides apparatus for the detection of a fetal electrocardiogram (fECG), the apparatus comprising: a plurality of low-noise electrodes for external skin placement during pregnancy, each electrode being connectable to low-noise signal recording and processing means wherein, when in use, the signal recording and

processing means is arranged: to record signals indicative of voltages developed between pairs of electrodes (for example, wherein one electrode is used as a common reference) in a plurality of signal channels; to process digitised data within the plurality of signal channels in order to generate at least one source signal which relates to the fECG of a single fetus; and on identification of the at least one source signal relating to the single-fetus fECG, to reconstruct, for at least one signal channel, that component of the digitised data within the channel which is attributable to that fetus, and therefore corresponds to a single-channel fECG. Preferably, when processing the digitised data, the signal recording and processing means is arranged to generate a plurality of separated source signals, at least one of which relates to the fECG of a single fetus

**[0018]** This invention is capable of extracting improved fECGs as compared to the prior art. It makes use of improved hardware for significant reduction of unwanted noise in the raw input data and appropriate digital signal processing for separation of the desired fetal signal. Once the electrical noise is reduced in this way (in one or more aspects), the contributions due to maternal muscle noise are seen to be surprisingly high. Once this has been noticed however, the invention provides the advantage of being able to observe objectively the relaxation state of the mother. Previously, in practical application of fetal monitoring equipment, it had been almost impossible to judge, even by the mother, a degree of relaxation below a certain level. Nor was it possible, with all other contributing noise factors, to realise the importance of providing such relaxation. Now, the noise on each signal channel can be observed prior to recording and recording is only begun when the mother is sufficiently relaxed. Accordingly, use of this invention will be of enormous benefit to operators of fECG equipment in judging when best to make a recording. Once the quality of recording is improved, more information can be extracted from the fECG, including respective fECGs in multiple pregnancies. In addition the likelihood of making a successful recording when the fECG signal is weak is greatly increased. Not only is the signal weak at early gestation, as might be expected, but also in the 27 to about 32 week gestation period. This phenomenon is thought to be due to a non-conducting layer forming around the fetus during this period. In any event the effect on fetal ECGs is well-known and discussed in detail by Oostendorp in his PhD thesis “Modelling the fetal ECG”, University of Nijmegen, January 1989. This equipment is consequently better able to extract fECGs at early gestation and in the problem area at around 27 to 32 weeks gestation than prior art fECG monitors.

**[0019]** Preferably the electrodes may be placed on the skin such that skin impedance at each electrode is less than 5 k $\Omega$  and, ideally, 2 k $\Omega$ .

**[0020]** It is preferred that the signal recording and processing means comprises electronic components for processing analogue voltage signals developed between pairs of electrodes to provide digital signals in the plurality of signal channels and data processing means to process the digital signals.

**[0021]** The electronic components most preferably comprise, for each abdominal electrode, a low-noise differential amplifier for amplifying the difference between each electrode’s voltage signal and a signal derived from the voltage developed at the reference electrode, an anti-aliasing low-pass filter and an analogue to digital converter. The analogue to digital converter is preferably a simultaneous multi-chan-

nel A/D converter which enables simultaneous sampling (and therefore synchronous digitisation) of the signals from each electrode.

**[0022]** These electronic components may be located in a multichannel lead box, remote from the patient, and connectable to the electrodes by, for each electrode, respective screened leads. This multichannel lead box may be one that is suitable for use in taking electroencephalography (EEG) scans. Additionally, a pre-amplifier may be located adjacent each electrode. This amplifies each signal to some extent before it is transmitted along the leads, decreasing relative losses and making it more robust to the effects of noise.

**[0023]** Alternatively, the lead box may be located near to a patient and in communication with the data processing means over a wireless link. This gives the mother considerably more freedom of movement while connected for a scan.

**[0024]** Consideration of the noise level is important, but its criticality depends on the stage of gestation and also whether or not there is more than one fetus present. All the low-noise factors listed above contribute to the possibility of providing input signal channels in practice with a noise component of less than 10  $\mu\text{V}$  and, ideally, less than 3  $\mu\text{V}$ . Prior art devices, without such noise-reducing features, have only ever been able to obtain fECGs for singletons, after about 20 weeks of gestation. By way of contrast, prototype apparatus built in accordance with this invention has demonstrated extraction of fECGs for triplets at 20 weeks gestation, twins at 18 weeks and singletons at 15 weeks.

**[0025]** The low-noise electrodes preferably comprise a plurality of abdominal electrodes for placement on the skin in the abdominal area and a common reference electrode, and the signal recording and processing means is arranged to record signals indicative of voltages developed between each abdominal electrode and the reference electrode.

**[0026]** They preferably further include a low-noise electrode which is connected to earth. The taking of voltage readings with reference to a common electrode is referred to as a unipolar configuration. Prior art techniques have always used a bipolar configuration. That is, measurements are taken between multiple pairs of electrodes. The electrode pairs in the bipolar configuration are placed sufficiently close that it may be assumed that the same noise contribution is made to each channel. By taking a difference signal therefore, the noise is reduced. Of course this relies on the signal being different in closely-spaced electrodes. A unipolar configuration is inherently more sensitive to the signal itself but consideration of the noise of the system has actively discouraged any attempt to make use of unipolar readings. One skilled in the art has hitherto believed that, without the noise reduction offered by a bipolar configuration, it would be impossible to extract a fetal ECG from a composite signal. It is only with development of apparatus in accordance with the present invention that this long-held belief has been proved erroneous.

**[0027]** A unipolar configuration offers advantages to both data processing and, rather unexpectedly, to noise reduction. Most prior art processing techniques to date have been based on the use of SVD or Principal Component Analysis to achieve source separation. Both these techniques require bipolar readings to be input, which, in turn, places constraints on the geometry of the electrodes used to gather the data. Data collected using a unipolar configuration is however open to analysis by further separation algorithms, for example, ICA. Thus there is an increased flexibility afforded the signal pro-

cessing by the present invention, with the added advantage that equipment may be operated successfully by a less expert user.

**[0028]** Unipolar signal channels allow an observer to view directly both the signal and noise on the channel. It is therefore far easier to recognise the noise component and to observe its level. If it can be made to fall, the reduction is readily observed. Thus, in a practical application of the apparatus according to this aspect of the invention, given the fact that maternal muscle noise makes a significant contribution to overall system noise, it is easier to see if the mother is relaxed prior to taking any measurements. In this way, recording quality can be considerably improved and significant information extracted from fetal ECGs, even in multiple pregnancies and at early gestation.

**[0029]** Moreover, unipolar channels, as provided in the present apparatus, are digitised and then available for processing electronically by whatever means is required. This allows software to be used to replicate a bipolar system by calculating differences between respective pairs of unipolar electrode channels. This also enables the same hardware to be used, for example, to measure a conventional ECG.

**[0030]** If more than one source signal relates to the same fetal fECG, the signal processing means may be arranged, on identification of the more than one source signal relating to the single fetus to combine these source signals with appropriate weighting in the reconstruction of the single-channel fECG. In this way the natural variation of the fECG waveform morphology over the maternal abdomen may be observed in a similar way to that in which the ordinary ECG waveform morphology is seen to vary between different chest leads.

**[0031]** In detecting fetal electrocardiograms in a multiple pregnancy, the processing means is preferably arranged, on identification of the at least one source signal relating to each fetus, to reconstruct, for at least one abdominal electrode, components of the digitised data within the corresponding signal channel which is attributable to each fetus. Components of digitised data may be reconstructed for all signal channels, thereby enabling construction of an abdominal surface intensity map of signal strength from each fetus. This may be used to provide an indication of fetal position and so ensure that a given diagnosis is assigned to the correct fetus.

**[0032]** The processing means is preferably arranged to generate the separated source signals by ICA. The capabilities of the ICA algorithm are more readily exploited when unipolar, rather than bipolar, readings are taken. If only unipolar readings are processed, there is no need for any special attention to be paid to electrode arrangement or to fetal presentation. The system is therefore more versatile and routine usage more practical. The standard operator of the fECG apparatus in accordance with this embodiment of the invention may be a person less highly trained than a doctor such as a midwife. Use of unipolar readings is facilitated by the care taken in the reduction of noise and, in particular, muscle noise.

**[0033]** The leads are preferably electrically screened cables, which ideally are kept close together to reduce noise from varying magnetic fields. The electrodes may be self-adhesive and preferably able to resolve signals over a bandwidth which includes 0.5 to 200 Hz. Preferably, the signal channels are arranged to contain a visible noise component on a display apparatus of less than 10  $\mu\text{V}$  and, ideally, less than 3  $\mu\text{V}$ . Subsequent digital signal processing and signal separation is arranged thereafter to reduce the noise level in separated fECG signals to values considerably lower than this and

certainly sufficient to identify detail in the fetal P and T waves, which are generally around 1  $\mu\text{V}$  in amplitude.

**[0034]** In a third aspect, this invention provides a method of recording a fetal electrocardiogram (fECG) comprising the steps of:

**[0035]** (a) attaching a plurality of abdominal electrodes to skin in an abdominal area during pregnancy;

**[0036]** (b) attaching a common reference electrode to the skin;

**[0037]** (c) measuring skin impedance at each electrode and, if the skin impedance is greater than 5 k $\Omega$ , reattaching that electrode;

**[0038]** (d) observing time-varying electrical skin potentials developed between each abdominal electrode and the reference electrode through a plurality of signal channels;

**[0039]** (e) if the noise on any raw data composite signal channel is greater than 10  $\mu\text{V}$ , reducing environmental noise contributory factors, including that due to muscle noise;

**[0040]** (f) collecting time-varying electrical skin potentials through the plurality of signal channels;

**[0041]** (g) processing data within the plurality of signal channels to generate a plurality of separated source signals;

**[0042]** (h) identifying within the plurality of source signals those which relate to the fECG of one among one or more fetuses; and

**[0043]** (i) reconstructing, for at least one of the plurality of signal channels respective components of the data within that channel which are attributable to each of the one or more fetuses, thereby generating a single-channel fECG for each individual fetus.

**[0044]** A further aspect of this invention provides a computer system configured to extract one or more fetal electrocardiograms from composite signals detected at electrodes distributed over a pregnant woman's abdomen, the system comprising:

filtering means for filtering digitised composite signals, each signal corresponding to a difference between a voltage developed at a signal electrode and that developed at a reference electrode, in order to remove unwanted frequency components;

processing means arranged to generate a plurality of separated source signals from the corresponding plurality of filtered composite signals, wherein each composite data signal is assumed to be a linear mixture of the unknown source signals;

selection means for identifying those source signal(s) which correspond to a single-fetus ECG;

reconstruction means arranged, for each single-fetus ECG identified, to reconstruct, for at least one signal channel, that component of the filtered composite signal within the channel which is attributable to that fetus and therefore corresponds to a single-channel fECG; and

display means arranged to display each fECG.

**[0045]** In a further aspect the present invention also provides a computer-readable medium embodying instructions for execution by a processor, the instructions relating to the extraction of one or more fetal electrocardiograms from composite signals detected at electrodes distributed over a pregnant woman's abdomen, wherein the computer-readable medium comprises:

program code for filtering multi-channel digitised composite signals, each signal corresponding to a difference between a

voltage developed at a signal electrode and that developed at a reference electrode, in order to remove unwanted frequency components;

program code for generating a plurality of separated source signals from the corresponding plurality of filtered composite signals, wherein each composite data signal is assumed to be a linear mixture of the unknown source signals;

program code for identifying, either automatically or by prompting for a user input, those source signal(s) which correspond to a single-fetus ECG;

program code for reconstructing, for each single-fetus ECG identified, and for at least one signal channel, that component of the filtered composite signal within the channel which is attributable to that fetus and therefore corresponds to a single-channel fECG; and

program code for displaying at least one reconstructed fECG to a user.

#### DESCRIPTION OF THE FIGURES

**[0046]** Embodiments of the invention will now be described by way of example only and with reference to the accompanying drawings.

**[0047]** FIG. 1 is an example of a single-channel ECG reading taken from an adult.

**[0048]** FIG. 2 is a schematic expanded portion of one cycle of the ECG illustrated in FIG. 1 showing conventional labelling of specific features of the heart-beat waveform.

**[0049]** FIG. 3 is a schematic illustration of apparatus suitable for recording fECGs in accordance with the present invention.

**[0050]** FIG. 4 is a schematic illustration of an arrangement of electrodes on the maternal abdomen suitable for recording fECG signals in accordance with this invention.

**[0051]** FIG. 5a is a schematic illustration of apparatus suitable for recording fECGs in accordance with a second embodiment of the present invention.

**[0052]** FIG. 5b is a schematic illustration of apparatus suitable for recording fECGs in accordance with a third embodiment of the present invention.

**[0053]** FIG. 6 is an exemplary illustration of the composite signal recorded during a singleton pregnancy on the 12 channels of the electrodes of FIG. 4.

**[0054]** FIG. 7 illustrates an example of data obtained in performing signal analysis on the 12 channels of composite signal of FIG. 5.

**[0055]** FIG. 8 shows an example of a parameterised average fECG taken using the apparatus of this invention.

**[0056]** FIG. 9 is an illustration of a display generated by a prototype apparatus of the invention for interpretation by an operator.

#### DESCRIPTION OF SEVERAL EMBODIMENTS

**[0057]** With reference to FIG. 1, a typical single-channel adult ECG trace 20 can be seen to comprise a series of regular pulses (22a, 22b, 22c) of amplitude  $\sim 0.8$  mV, one such pulse being produced roughly every 0.75 s; that is, around 80 per minute. Each pulse corresponds to a single heart-beat.

**[0058]** FIG. 2 shows a schematic view of an averaged pulse 22 such as that seen in the trace 20 shown in FIG. 1. This is referred to as the underlying cardiac complex which repeats with each beat of the heart. The general waveform of the complex 22 has various features which are known to provide important diagnostic information. These include (in order of

appearance during the beat) P, Q, R, S and T waves. The QRS complex corresponds to the principal, powerful beat of the heart. Using this notation the position of the onset of the P and Q waves, and of the termination of the S and T waves in particular can be determined and labelled as indicated in the figure. In this way relative timings and durations of different parts of the cardiac complex are routinely obtained. These can be compared with known equivalents in a healthy heart and diagnosis aided. Parameters with particularly important known diagnostic properties in ordinary ECGs are, for example, PR and QT intervals (periods between onset of P wave and onset of Q wave and between onset of Q and termination of T wave respectively), QRS duration and relative heights of the PQ and ST segments (approximately flat readings between wave features) compared with the isoelectric line.

[0059] Extraction of a detailed fECG would provide the ability to display the instantaneous peak-to-peak heart rate in a rhythm strip showing the P and T waves and, in addition, the ability to examine and characterise the detail in the underlying waveform by measuring quantities such as the PR and QT intervals and QRS duration, etc. From comparison with the benefits of ordinary ECGs, it is expected that the former ability would enable the diagnosis of cardiac arrhythmias and anomalies such as atrial or ventricular ectopic beats or heart block. The latter would enable diagnosis of more subtle conditions not manifested directly in the heart-rate such as long-QT syndrome.

[0060] It is to be noted that extraction of the detailed fetal ECG from a maternal abdominal signal is clearly a more demanding problem than finding merely the fetal heart-rate after suppressing the maternal QRS peaks. It is necessary to detect the fetal P and T waves in order that a detailed characterisation may be performed.

[0061] FIG. 3 is an illustration of apparatus suitable for implementing this invention, indicated generally by 30. The apparatus 30 comprises a number of electrodes (G, Rs, Rl, 1, 2, 3, . . .) suitable for attaching to the mother's skin and monitoring voltage signals generated thereon. The electrodes G, Rs, Rl, 1, 2, 3, . . . are connected via respective screened leads (32a, b, c, d, . . .) to a lead box 34. At the lead box 34 the signals are amplified and converted to digital readings ready for recording and processing by a computer 36. For convenience only six electrodes G, Rs, Rl, 1, 2, 3 are illustrated in this Figure, but in this specific embodiment of the invention there are, in fact, 15 as will become apparent later. Processing electronics within the lead box 34 for the signal electrodes (1, 2, 3, . . .) are shown inset in FIG. 3. These comprise, for each abdominal electrode (1, 2, 3, . . .), a low-noise differential amplifier 37 and an anti-aliasing low-pass filter 38 as well as a common (to all electrode channels) simultaneous multichannel analogue to digital (A/D) converter 39.

[0062] Somewhat surprisingly it has been found that a commercially available Electroencephalography (EEG) system has proved suitable for adaptation for acquisition and display of raw input composite fECG data readings. Accordingly the computer 36 is that from a portable EEG system (SYS98-Port24-CL) supplied by Micromed Electronics UK Ltd (Woking, Surrey) and which therefore comprises a battery-powered laptop computer running System '98 EEG recording and analysis software (SYS-98) under Microsoft Windows NT operating system. The SYS-98 software provides a convenient interface from the A/D outputs to display apparatus (screen, not shown) and to a data storage medium (hard disk).

Bespoke software is also run on this computer 36, this software being specifically designed to enable reading of recorded (by EEG-specific software) data, separation and processing of the fetal contribution(s) and for display of the fECG and parameters derived from it (such as fetal heart-rate, PR, QRS, QT intervals, etc.). Details of the processing carried out by this bespoke software will be explained later. The type of computer 36 is clearly not critical however, all that is required is that it has sufficient processing capacity for running the recording, processing and display software and sufficient memory for storing the recorded data, processed results and the display itself. Preferably the computer should be portable. Not only does this provide for ease of transfer to patients, but portable computers may be run on batteries and so, in this way, the computer 36 may be isolated from the mains supply.

[0063] The lead box 34 and the computer 36, including the computer display screen and recording and display software for the raw composite data (as opposed to processed data which is specific to this application), as well as their connecting lead are all part of the portable EEG system. The leads 32a, b, c, d, e and their connectors to the lead box 34 and to the electrodes G, Rs, Rl, 1, 2, 3, etc. are purpose-built for use in this invention. It is to be noted that a commercial EEG machine proved convenient for use in constructing a prototype apparatus. It is envisaged that bespoke equipment will ultimately prove more suitable for implementing this invention.

[0064] The electrodes G, Rs, Rl, 1, 2, 3, etc. are commercially available, disposable, self-adhesive neurology electrodes (type 710 01-K) manufactured by Neuroline®. The principal preferences for the electrodes G, R, 1, 2, 3, 4 are that they are low-noise and of a type that is readily attached to a patient in such a way as to result in an impedance at the skin of less than 2 kΩ. Moreover they must be of sufficient number to allow effective signal separation by the processing software. Each electrode G, R, 1, 2, 3, 4 with its respective screened lead 32a, b, c, d contributes a single, separate, channel of data to a multichannel recording.

[0065] Note that the 710 01-K electrodes sold commercially have a 10 cm length of ordinary (unscreened) cable attached to them. This type was selected as the attached cable length is the shortest available. It is preferred that this length of wire is nearer 1 cm, or that the electrodes are attached directly to the screened leads 32a, b, c, d, e as this would reduce electrical noise further. Disposable electrodes with shorter leads specific to fECG could be trivially made to the same design.

[0066] The screened leads 32a, b, c, d are made from 0.9 mm coaxial screened cable of a type suitable for biomedical applications. They should be screened sufficiently to reduce the noise level during fECG recordings to less than 3 μV. Such cables make simple, convenient connections from the disposable electrodes G, Rs, Rl, 1, 2, 3 to the lead box 34. Connection is made to the lead box 34 by means of a metal D-type connector (not shown) with its body connected to ground, which arrangement provides electrical screening.

[0067] The precise number of electrodes G, Rs, Rl 1, 2, 3 and respective leads 32a, b, c, d is not important to signal separation, although it does determine the number of distinct sources which will be obtained by the analysis. Electrodes are to be distinguished however by function. That is, the system includes one earth electrode G and two common reference electrodes Rs, Rl and a number of electrodes 1, 2, 3 for

attachment to the mother's abdomen. As a rough guide, eight or more abdominal electrodes **1, 2, 3** are generally sufficient to provide adequate abdominal coverage and to permit signal separation into sufficient distinct sources. For example, two or three sources generally result from the maternal heart and typically two per fetus. Additional electrodes allow the separation of unwanted artefacts such as those associated with maternal breathing, unwanted electrical interference, etc. Larger numbers of abdominal electrodes **1, 2, 3** may be used, subject only to limits of practicality such as time needed to apply them, comfort and convenience of the mother and limitations of the processing and display systems. Only one common reference electrode Rs, Rl is used at a time. One of them Rl is connected to a longer screened lead **32f** than the other electrodes G, Rs, **1, 2, 3**. Thus either a long-Rl or short-Rs leaded reference electrode can be placed on the mother, whichever is the more able to reach a conveniently-chosen reference attachment. For example, if the convenient attachment is the mother's ankle, some distance away from the abdomen, the long-leaded electrode Rl is used.

[0068] FIG. 4 is an illustration of one possible arrangement of electrodes **1, 2, 3, G** around the mother's abdomen. In this example, the embodiment comprises twelve abdominal electrodes (**1-12**), the earth electrode G and the common reference electrode Rs or Rl (not shown in this figure) which are all attached to the mother's skin **40**. Placement is indicated in the Figure by shaded circles indexed with reference numerals of the corresponding electrodes. The common reference electrode Rl is attached to the mother's ankle (not shown), the remainder to her abdominal area. Alternatively, the common reference electrode Rs is attached to the mother's abdomen adjacent to G and the remainder also to her abdominal area. Electrode positions **1-12** are shown linked by network lines **42** which indicate that an approximately hexagonal arrangement of electrodes is ideally employed for even abdominal coverage. This is not however critical: the degree of separation achieved is not critically dependent on exact electrode location.

[0069] In order to achieve good separation the abdominal electrodes **1-12** should not be placed too close together and should involve a wide coverage of the abdomen. Typically, a regularly spaced arrangement of 12 electrodes, results in an electrode separation of about 10 cm. A practical placement, as shown in FIG. 6, includes coverage from one side of the abdomen to the other and from the pubic hairline to the likely upper limit of the uterus. This latter can be judged from gestation or by following a standard configuration which is sufficient for the maximum height of the uterus (fundal height) which occurs late in pregnancy. It is a feature of this invention that suitable placement can readily be achieved by, for example, a midwife.

[0070] The common reference electrode Rs, Rl is selected to be of appropriate length for placement at a conveniently-reached point on the mother's body. In some instances the ankle may be appropriate as this is far from the abdomen and the signal that the remaining electrodes **1-12** are detecting. That is, neither signal nor noise will appear artificially reduced when measuring a unipolar voltage difference between abdominal electrode **1-12** and reference R.

[0071] On the other hand it has been found that placement of the reference electrode Rs on the abdomen has advantages in reducing the amount of noise seen on the screen in the raw composite data. That is, use of an abdominal reference allows use of the short connecting lead **32e**. The disadvantage of

using a distant location for the reference electrode R is that it necessitates using the long lead **32f**. This creates a larger conducting loop, which leads to higher magnetic induction and greater scope for electrical noise to enter the system.

[0072] It may therefore prove appropriate to use either reference electrode Rs, Rl, depending on the situation. Both options are therefore made available in this embodiment of the invention. It is also possible to employ a combination of reference electrodes Rs, Rl and leads **32e, 32f** (including additional electrodes and leads as required), such as is used in conventional ECG. Whatever combination is used, the fact remains that all leads **32a, b, c, d, e, f** should be as close to the skin and to each other as possible in order to reduce electrical and magnetic noise through magnetic flux linkage of loops formed by the combination of mother and leads.

[0073] The earth electrode is placed in position G, close to the mother's navel. Again, an alternative site close to the abdominal area may be chosen.

[0074] In preparation for attachment and recording, the mother will ideally lie comfortably on a bed with the lead box **34** close by, but touching neither the patient nor the bed frame. She should be allowed to relax for a few minutes to help reduce involuntary muscle activity.

[0075] The screened leads **32a, b, c, d** connect the electrodes **1-12, G, R** to the lead box **34**. An outer braided mesh layer of the coaxial cable comprising each screened lead **32a, b, c, d** is connected to isolated ground at the lead box **34** and to the metal case of the D-type connector. The earth electrode G is also connected to isolated ground at the lead box **34**. This provides a return bias current path to the mother's body for common mode interference which will not be passed by the amplifier **38**.

[0076] Voltage signals arising from cardiac activity and other sources are picked up by the electrodes **1-12, R** attached to the skin **40**. The signals are then communicated to the lead box **34** via the screened leads **32a, b, c, d**. The lead box **34** is the SAM 25R "headbox" of the Micromed Electronics EEG system. The advantage of an EEG headbox as opposed to an ECG lead box, is that the former has superior electronics (accordingly less noisy electronics) and an increased number of input channels available for use. The input channels are, importantly, configured for unipolar use.

[0077] The particular lead box **34** used has a number of possible connections (more, in fact, than are required in order to implement the present invention). There are 21 unipolar input channels and 5 ground connections in addition to the common reference connections. Further specifications of the SAM 25R lead box of relevance to a prototype apparatus built to implement this invention are: touch-proof safety connections, 512 Hz sampling, passband from 0.5-256 Hz, low-pass anti-aliasing filter with cut-off frequency at 1 kHz and 12-bit resolution covering a voltage range of  $\pm 2$  mV. Clearly in connecting the electrode arrangement shown in FIG. 4 to the lead box **34** only 12 of the 21 unipolar channels are used. Additional abdominal electrodes may therefore be used if required. This may be a useful facility in special cases (for example triplets or higher multiple pregnancies) in which there is particular concern for the health of the fetus or fetuses.

[0078] The multiple channel inputs to the lead box **34** are used in a unipolar configuration. That is, voltage readings are taken between each abdominal electrode **1-12** and the common reference electrode R (whichever Rs, Rl is selected).

This is to be compared with prior art ECG devices which have attempted to solve the problem of system noise by taking bipolar readings.

**[0079]** A conventional ECG reading is taken between electrodes arranged on a patient's chest and a specifically configured reference formed from leads located on the patient's wrists and ankles. In this way six unipolar readings are available for processing. Adaptation of conventional ECG equipment to measure fECGs by positioning ECG electrodes on the mother's abdomen encounters two fundamental problems. The first is that the noise level within the equipment itself is too high. The second is that only six channels are available. Although in some instances, for example a singleton pregnancy some way into gestation, this may be sufficient to separate the fetal signal, there may be insufficient coverage for more complicated situations requiring data separation using an ICA algorithm. For a "blind" approach, that is one in which no knowledge of fetal position or presentation is assumed a priori, it is recommended that more than six electrodes, and more preferably eight, are used to achieve appropriate coverage.

**[0080]** The bipolar procedure was adopted by Callaerts when attempting to derive fECG details from measurements taken at the mother's abdomen. Use of such pairs of electrodes and the SVD algorithm renders the measurements highly dependent on the geometric arrangement of electrodes and orientation of the fetus in the uterus.

**[0081]** The unipolar configuration has become a practical proposition in the present invention by the careful reduction of electrical noise. This is why use has been made of the low noise electrical electrodes, screened leads and electronics of the EEG headbox. This therefore overturns the long-held belief that such an arrangement would be incapable of overcoming the noise problem.

**[0082]** Within the lead box **34**, analogue voltage signals from each abdominal electrode (**1-12**) are fed to one input of a respective differential amplifier **37** and the voltage signal from the reference electrode (R) is fed to the other. Each differential amplifier **37** therefore outputs an amplified signal proportional to the difference between the voltage developed at the associated abdominal electrode (**1-12**) and that developed at the reference electrode (R): a unipolar voltage. The resulting amplified signals are filtered by respective anti-aliasing low-pass filters **38** and digitised by the simultaneous multi-channel A/D converter **39**. The advantage of using a multi-channel A/D converter **39** is that simultaneous sampling can be arranged on all channels (**1-12**). These digitised signals are then passed to the computer **36** for signal processing.

**[0083]** It is to be noted that although a unipolar configuration has advantages, a bipolar configuration is by no means precluded. One may be replicated simply by taking differences between digitised unipolar channel outputs, if such a bipolar configuration is required.

**[0084]** In setting up the equipment for making ECG and fECG recordings it is important to reduce ambient and system noise. The following procedure has been found to produce sufficiently low-noise readings:

**[0085]** i). The mother's skin **40** where each electrode **1-12**, G, R is to be placed is lightly excoriated using standard abrasive preparation tape (e.g. "Skinprep", manufactured by 3M) and then cleaned with an alcohol- or water-based swab.

**[0086]** ii). Each electrode **1-12**, G, R, a 2 cm self-adhesive pad, is attached to the skin with light finger pressure, and arranged such that the short trailing wire points towards the earth electrode G.

**[0087]** iii). Each electrode **1-12**, G, R is then connected to the corresponding screened lead **32a, b, c, d**.

**[0088]** iv). The screened leads **32a, b, c, d** are connected to the lead box **34** via the D-type screened connector (not shown).

**[0089]** v). The recording system is switched on, using battery power only—i.e. isolated from the mains supply.

**[0090]** vi). The skin impedance at each electrode is measured and any electrode having a skin impedance greater than about 2 k $\Omega$  is reapplied.

**[0091]** vii). Individual screened leads are gathered together and, in addition, held as close to the mother's skin **40** as possible in order to minimise magnetic pickup.

**[0092]** viii). The recording system (electrodes **1-12**, G, R, leads **32a, b, c, d** and lead box **34**) is set to display real-time signals from abdominal electrodes **1-12**. On screen traces corresponding to outputs from all electrode channels are displayed simultaneously.

**[0093]** ix). Possible sources of electrical interference (such as mains leads in the room) are disconnected if possible.

**[0094]** x). The mother is asked to relax as much as possible and her posture is adjusted (for example, using pillows under her legs, ankles, etc.) until the noise level of all the traces is less than 10  $\mu$ V and preferably as low as possible.

**[0095]** xi). Once the operator is satisfied as to the quality of the traces, recording is started. The computer **36** records composite raw data represented in the traces and any display settings and saves them to, for example, its hard disk.

**[0096]** These various steps contribute to lowering the noise level as far as possible. Another surprising observation has been made following the procedure listed at step x). This is, that maternal muscle noise is a major contributing factor to the noise of the system. Once steps such as ensuring that the mother is sufficiently relaxed are taken, the overall noise can be made to fall significantly. This, plus the electrical noise reduction described above, has been found in this embodiment of the invention to reduce noise sufficiently to allow use of unipolar channel inputs. It has previously been unappreciated that muscle noise made such a significant contribution.

**[0097]** Once sufficient data has been collected, recording is stopped, the screened leads disconnected from the lead box **34** using the D-type connector and the electrodes **1-12**, G, R removed from the mother.

**[0098]** Using prototype apparatus, it has been found that each recording takes approximately 15 minutes, including application and removal of the sensors.

**[0099]** There are two major advantages of the hardware described in relation to this embodiment of the invention. The first is that commercial-off-the-shelf, portable, battery-powered equipment is used. This makes the system relatively inexpensive and also mobile. Recordings can be taken at home, by a hospital bed, etc wherever convenient. The second advantage is that the same equipment is capable of performing neo-natal and adult ECGs. Alternative processing of the signals will be required: that is, six channel unipolar inputs will have to be processed in order to generate a conventional 12-lead ECG trace and additional reference connections (ECG requires that the reference used is an average signal from a set of reference voltages measured at standard locations), for example to the subject's wrists and ankles, will be

required. However this is a straightforward matter, and the approach to take will be readily apparent to one skilled in the art. The ability to take neo-natal measurements is important because it enables a better comparison between fECG and neo-natal ECG to be made. Equipment differences may blur the comparison if traces are obtained using separate pieces of apparatus. Prior art fECG monitors are made from specialised hardware which is not capable of being adapted to take conventional ECGs.

[0100] As mentioned previously, the SAM 25R EEG "headbox" 34 is described in relation to this embodiment as it was the most appropriate for use in a prototype. That is, it was readily adaptable to perform the functions required and so avoided the need to build bespoke equipment at this stage of development. It is of course to be expected that improved performance can be obtained with a purpose-built lead box 34. An improved design would include features which match the performance of the lead box 34 more closely with the fECG data. In particular, the SAM 25R lead box 34 has a non-ideal low-pass filter at its input and an amplifier which is overly noisy. It is anticipated that the filter should be redesigned with a passband that rejects frequencies greater than around 200 Hz, as opposed to the ~1 kHz limit of the EEG headbox. This feature would reject more unwanted noise and provide improved anti-aliasing. Amplifiers with noise less than 0.1  $\mu\text{V}$  are available and should preferably be used (the EEG box amplifier has a noise level of 0.16  $\mu\text{V}$ ). Such a redesign would improve acquisition of P and T waves in the fetal heart-beat complex, which can be of just 1  $\mu\text{V}$  or less in amplitude. Additional connections, to allow for the additional reference limb connection for conventional ECG, should also be provided.

[0101] Alternative embodiments of apparatus suitable for implementing this invention are shown in FIGS. 5a and 5b. In both these Figures, components of the system(s) common to those shown in FIG. 3 are like referenced.

[0102] Considering FIG. 5a first, this embodiment 44, as with the earlier prototype 30, comprises a number of electrodes (G, Rs, Rl, 1, 2, 3, . . .) suitable for attaching to the mother's skin and monitoring voltages signals generated thereon. In this embodiment however, each electrode (G, Rs, Rl, 1, 2, 3, . . .) is connected first to its own pre-amplifier 46 (illustrated schematically inset 47) and from there, via respective screened leads (32a, b, c, d, . . .), to the lead box 34. The detail of electronics suitable for pre-amplification will be apparent to one skilled in the art. As before, the lead box 34 contains the differential amplifiers 37 and low-pass filters 38 for each electrode channel and the multi-channel A/D converter 39. The computer 36 performs data processing on the digitised output from the lead box 34.

[0103] In the FIG. 3 embodiment 30, the lead box 34 houses the electronics responsible for performing all processing functions for the different input channels (via leads 32a, b, c, d, . . .). These functions include: amplification, low-pass anti-alias filtering, high-pass filtering, digitisation and optical isolation. In this present embodiment 44, individual amplifiers 47 are disposed adjacent respective electrodes (G, Rs, Rl, 1, 2, 3, . . .) and each 47 therefore provides a stage of pre-processing on its individual channel. In this way, the signals propagating along the leads 32a, b, c, d, . . . have already been amplified to some extent and are therefore much larger and more robust to electric and magnetic noise sources.

As a consequence electric and magnetic noise in the leads is far less significant than for the embodiment 30 shown in FIG. 3.

[0104] An electric guard potential may be applied to the earth shielding on the screened leads 32a, b, c, d, . . . Methods of implementing this will be apparent to one skilled in the art. The guard potential has the effect of reducing lead capacitance and minimising mismatch between input capacitances. This increases the common mode noise component of the detected signal that is rejected by the differential amplifier 37. Although the guard potential may be similar to the signal voltages of interest, the earth shielding must be driven from a low impedance source. For example, from a voltage follower driven by the signal of interest.

[0105] FIG. 5b illustrates a further embodiment 48 of the present invention. This embodiment 48 also comprises a number of electrodes (G, Rs, Rl, 1, 2, 3, . . .) suitable for attaching to the mother's skin and monitoring voltage signals generated thereon, with each electrode (G, Rs, Rl, 1, 2, 3, . . .) being connected to its own dedicated pre-amplifier 46. In this embodiment 48 the lead box 34, although housing the same electronics as described in relation to previous embodiments 30, 44, is also connected to a transmitter 49a. A corresponding receiver 49b is connected to the computer 36. Pre-amplification is again performed at the electrodes, as for the embodiment 44 shown in FIG. 5a. The lead box 34 is however placed close to or on the patient (for example, using a belt) and its output is transmitted to the computer 36 over a wireless (for example, infra red) link 49a,b. Transmitting amplified data in this way enables the screened leads 32a, b, c, d, . . . to be far shorter than previously: they only need reach the closely-positioned lead box 34. This further reduces the amount of noise and signal loss arising from the leads 32a, b, c, d, . . . In addition, the lack of long trailing leads and their physical connection to the computer 36 enables the mother to move around more freely, without leads or electrodes having to be disconnected, potentially allowing her to relax more readily when a recording is to be taken. Finally, this embodiment also offers the potential for readings to be made whilst the mother is ambulatory, if sufficient relaxation can be induced.

[0106] FIG. 6 is an illustration 50 of 5 s worth of raw composite data traces taken using the equipment described in FIG. 3 attached as shown in FIG. 4. Twelve traces 52a, b, . . . , l are generated corresponding to the twelve abdominal electrodes 1-12 of FIGS. 3 and 4. In all traces both maternal 54 and fetal 56 heart-beats are visible. For example, trace 52k exhibits a number of fetal heart-beats . . . 56a, 56b, 56c, 56d, 56e . . . , although one of these 56c is masked by the far stronger maternal signal 54a. However it can readily be seen that information relating to the detail of the fECG in particular is not at all apparent.

[0107] Referring once again to the embodiment 30 shown in FIG. 3, the computer 36 receives, from the lead box 34, digital data relating to the traces 52a, b, c . . . l for signal processing. In this embodiment of the invention the digitised signals are filtered (this time in software) in order to remove further unwanted frequency components. The filters used consist of a high-pass infinite impulse-response (IIR) filter of 6 filter taps, and a low-pass finite impulse-response (FIR) filter of 9 filter taps. The high-pass filter is designed using an IIR Butterworth filter with a passband of 2 Hz, stop-band of 0.1 Hz and stop-band attenuation of 120 dB, resulting in a 3 dB point of 1 Hz and a passband ripple of 0.01 dB. The

low-pass filter is designed using a Blackman window with a band-edge at 150 Hz. Filtering is implemented using a zero phase forward and reverse digital filtering technique. This band-pass filtering reduces baseline wander to acceptable levels and also removes high frequency interference which lies outside the frequency range of interest.

**[0108]** The filtered signals are then subject to a Blind Signal Separation (BSS) technique based on Independent Component Analysis (ICA), I. J. Clarke "Direct Exploitation of non-Gaussianity as a Discriminant", EUSIPCO '98, Rhodes, Greece, 8-11 September, 1998. ICA is a powerful statistical and computational technique for revealing hidden factors that underlie sets of random variables, measurements or signals. In this situation therefore it is used to analyse the twelve signal traces **52a-1** obtained from the abdominal electrodes **1-12**. ICA defines a model for observed composite data variables  $\underline{x}_i$  based on the assumption that each is a linear or nonlinear mixture of some unknown latent sources  $\underline{s}_j$ . The mixing system is also unknown and the sources are assumed mutually independent and non-Gaussian. In cases for which the composite data variables are provided as a set of parallel signals or time series, and no prior knowledge about the signals, sensors or method of propagation, etc. is employed, the term Blind Source Separation is used to characterise the problem.

**[0109]** Thus, in this instance, the electrodes **52a-1** will be considered indexed by the subscript  $i$ , each of the  $i=1, \dots, n$ , electrodes producing a sensor output  $\underline{x}_i$ . Each sensor output  $\underline{x}_i$  has been digitised by the lead box **34** and so comprises  $m$  time samples of recorded data. The ICA algorithm takes the  $m \times n$  matrix  $X$  of sensor outputs and generates a mixing matrix  $M$  and a set of  $n$  independent sources  $\underline{s}_j$  such that each sensor output  $\underline{x}_i$  can be written as a different linear combination of the sources  $\underline{s}_j$  i.e.:

$$\underline{x}_i = \sum_{j=1}^n m_{ij} \underline{s}_j \text{ or simply } X = SM \quad (1)$$

where  $X$  is a matrix whose columns are the  $n$  sensor outputs  $\underline{x}_i$  and  $S$  is the  $m \times n$  matrix whose columns are the set of  $n$  independent sources  $\underline{s}_j$ . In this way the composite data  $X$  is separated into different independent sources of interest  $\underline{s}_j$ . The various sources will comprise the fetal ECG, maternal ECG and also some separated unwanted noise sources. In multiple pregnancies there will, of course, be more than one fetal ECG. Signals of interest relating to selected sources may be separated and examined individually in isolation.

**[0110]** This model assumes that the sources  $\underline{s}_j$  are point sources, which is clearly not the case for a physiological source such as the heart which is of finite extent. In this situation it is an artefact of the calculation that multiple, separated sources are found. By using abdominal sensors alone the number of sources found per heart varies from about one to three but depends on factors such as proximity of the electrode to the source, fetal presentation and details of electrical conduction to the surface. This then is the origin of the requirement of two electrodes per fetal heart which was referred to previously.

**[0111]** A beneficial additional consequence is the ability to observe the variation in the structural morphology of the fetal heart-beat over the maternal abdomen. In a standard ECG it is

well known that the detail (morphology) of the measured heart-beat waveform varies over the chest i.e. depending on which electrode the signal is observed. For example, the P wave is known to appear biphasic, rather than the peak shown in FIG. 2, if observed at certain chest positions. This variation in heart-beat morphology may also be observed in the fECG over the mother's abdomen using the equipment described herein. This provides a confirmation to a doctor that the technique is functioning as intended, as well as perhaps indicating fetal presentation and position and may ultimately prove to be an additional diagnostic tool.

**[0112]** ICA is a well known analytical technique and it is not necessary to expand on the presentation given herein. Further details are described in "Independent Component Analysis—theory and applications" by T-W. Lee, published by Kluwer Academic, Boston (1998). Many developments of the basic ICA technique are also well known and it is anticipated that these may well be applied beneficially to the signal processing described herein. In particular, signal separation should be achievable in real time, and run as the data is recorded. Other improved algorithms, tailored to the specific application, should enable application of the fECG technique described herein to situations in which continuous monitoring is required. Such monitoring is known to be important to a number of areas, including long-term assessment of heart-rate variability, identification of intermittent cardiac arrhythmias and anomalies and the use of fECGs during labour.

**[0113]** FIG. 7 illustrates stages involved and results obtained in processing data collected from a singleton pregnancy using the apparatus described herein. The Figure is divided into three columns **60**, **62**, **66**. The first column **60** comprises 12 channels of input data, one channel collected at each electrode, the number of the electrode being indicated at the left hand side of each trace. Accordingly, this column contains a portion of the information shown in FIG. 6. A second column **62** illustrates the 12 separated sources found by ICA. Each source is annotated at its left hand side with the percentage of total energy found in that source, and reference will be made to this value when referring to individual sources. To the right of each source is a source-selection button **64**. A third column **66** comprises modified data generated using only those sources from the second column **62** which are of interest.

**[0114]** Examination of the separated sources **62** readily indicates that two sources have been found for the maternal heart-beat: the two strongest sources **73**, **20** show pulses of the expected frequency. Similarly two weaker sources **3.4**, **0.2** exhibit readily noticeable pulses at a typical fetal heart-beat frequency and are aligned with each other. The remainder of the sources typically correspond to noise sources such as maternal breathing, muscle noise, mains and other electrical interference, etc.

**[0115]** In this example it is assumed that the fetal ECG is of interest. In order to extract this, each electrode trace **1-12** in the first column should be reconstructed using only the fetal heart sources **3.4**, **0.2**. That is, each electrode trace **1-12** is first modelled as a mixture of sources **73-0.05**, and then reconstructed using only those sources, and associated coefficients, of interest ( $\underline{x}_i(m) = \sum_j m_{ij} \underline{s}_j$ ,  $j$  being restricted to the index associated with particular separated sources). This gives rise to twelve modified data traces **m1-m12**, as illustrated in the third column **66**. In these traces **m1-m12**, the fetal heart-beat is readily apparent and the maternal and other noise sources have been suppressed.

[0116] In this embodiment of the invention, selection of the required sources is made by means of the source-selection button 64 which is displayed on the computer 36 next to each separated source 73-0.05. Each button may be toggled between a “no” indicator, meaning discard the source and a “yes” indicator, meaning to make use of it. Selection may be made by the operator. FIG. 6 illustrates that the fetal sources 3.4, 0.2 are selected by means of the source-selection button 64. It is clear however that this selection process can readily be automated, thus enabling a fetal trace to be displayed without the need for operator intervention.

[0117] From this analysis it can be seen that, although noise reduction is important in enabling a fECG to be extracted from the raw composite data, an apparently noisy signal may still prove tractable. This is because, as can be seen, separable noise will be isolated by the ICA technique and can be discarded. The problem is that separable noise is not often distinguishable from unseparable noise in the composite signal. Consequently, all possible noise is minimised in constructing and operating apparatus in accordance with this invention. This gives the best chance for extraction of the fECG, although clearly it may still be extractable from an apparently noisy signal under certain circumstances.

[0118] FIG. 8 illustrates a representative average 70 of the underlying waveform of the modified data m1-m12 shown in the third column 66 of FIG. 7. This data is extracted in accordance with standard methods of averaging the signals in the area around the peaks in a particular trace. The P wave 72 and T wave 74 are clearly seen in addition to the QRS complex. Positions of P and Q wave onsets and S and T wave terminations are also marked. From these details, information 76 concerning certain diagnostically important intervals has been extracted. The QT interval is a particularly important diagnostic parameter but it depends on determination of the end of the T wave. No prior art example of fECG has managed to quantify this parameter. It is clearly displayed as 263 ms in the display of FIG. 8.

[0119] In the present embodiment of the invention determination of the positions of wave onsets and terminations and performance of consequent calculations of intervals are carried out semi-automatically. That is, cross-hairs are displayable on the display screen and can be electronically dragged and dropped by an operator. Once marker positions have been set in this way, the diagnostic intervals are automatically calculated by the computer 36 and added to the display and to an automatically generated patient record. It is a straightforward matter for one skilled in the art to provide for fully automatic characterisation and measurement of fetal waveforms.

[0120] Underlying waveforms 70 such as that shown in FIG. 8 may be generated for each individual electrode 1-12 channel and displayed superimposed in an outline of the abdominal surface. This abdominal surface map therefore shows the average fetal waveform at each electrode position 1-12 and this can be used by a doctor to assess fetal health and development. The waveform 70 shows the expected morphological variation, which is expected to be of assistance in clinical diagnosis.

[0121] Abdominal surface intensity maps are readily generated from the separated sources by shading a map of the geometrical locations of the sensors according to the strength of the coefficient  $m_{ij}^2$  of the desired source  $j$  (or combination of sources) at each sensor  $i$ .

[0122] Brighter areas are used to indicate higher levels of signal strength, and the different fetal positions can be distinguished in a multiple pregnancy.

[0123] In using this equipment, waveforms 70 may be extracted for each fetal heart-beat. Not only is the waveform 70 a useful diagnostic parameter itself, but it may also be used for “gating” another imaging tool such as ultrasound. The location of the peak in the activity of the fetal heart-beat is recorded. This location is fixed on the cardiac cycle and is obviously useful in determining the instantaneous heart-rate of the fetus. It can also be used however to send a pulse at the appropriate time to a second piece of monitoring or imaging equipment such as one based on Doppler ultrasound. Imaging can be improved by making use of information about the position of the cardiac cycle. For example, 3D images of the fetal heart may be reconstructed from time-aligned 2D ultrasound images. Timing information from the fECG peak locations can therefore be used to reduce motion artefacts.

[0124] Signal separation does not have to be carried out using ICA, but this technique is very much preferred. SVD, another commonly-used analysis tool in signal processing, requires, for fECG separation, careful consideration of the geometrical relationship between the abdominal electrodes. With the reduction in noise provided by the hardware of this invention, it is envisaged that less sophisticated signal processing techniques may also yield acceptable results. For example, if an additional electrode is placed on the mother's heart, the maternal signal can be subtracted from the composite signals simply by removing components which correlate with the additional electrode signal. Of course the noise contributions would not be separated out, but if these are sufficiently low, useable results may be obtained. It is a feature of this invention that alternative or further improved signal separation techniques may be incorporated primarily with software changes and without substantial alteration to the equipment.

[0125] Use of a computer 36 to display the processed data clearly offers extreme flexibility in which of the range of extractable parameters is displayed. An example display is shown in FIG. 9.

[0126] It is clear that a variety of parameters may be set up for display and output. Examples of apparatus capabilities which it is envisaged will prove useful are:

[0127] i). Display of patient data, recording and processing details for patient and hospital records.

[0128] ii). Display of the multichannel abdominal input data (unipolar or chosen bipolar configuration).

[0129] iii). Means for manual or automatic selection of the source or sources of interest e.g. fetal, maternal, or raw data channel in the case of the use of the equipment for performing conventional ECG.

[0130] iv). Ability to perform a projection of any or all of the channels of data onto the subspace spanned by the selected sources i.e. to eliminate the contributions to all the data channels from sources other than the source or sources of interest, and display the results.

[0131] v). Detection of the positions of the QRS peaks to be used as fiducial markers for the time-sequenced averaging of the cardiac waveform. A number of different peak detection algorithms can be employed including the use of a simple threshold.

[0132] vi). Plotting and reporting of heart rate over the recording interval together with statistical parameters concerning the heart rate and its variability. These parameters

include: maximum, minimum and mean heart rate, presence of gross changes in heart rate, maximum variation in heart rate (maximum-minimum heart rate), relative variation in heart rate (maximum variation divided by mean heart rate), coefficient of variation (standard deviation divided by mean heart rate), etc.

[0133] vii). Fitting of data window around the marker for averaging purposes.

[0134] viii). Averaging of the data windows time-aligned to the marker to produce an average waveform for the chosen data channel.

[0135] ix). Parameterisation of the average waveform by manual or automatic labelling of the positions of such features as the P-wave onset, the Q-wave onset, the S-wave termination and T-wave termination in order that the PR interval, QRS duration and QT interval and any other parameters of interest may readily be determined.

[0136] x). Automatic generation and annotation of parameters such as PR and QT intervals and QRS duration from the semi-automatic measurements of P-wave and Q-wave onsets and S and T-wave termination on a screen displaying the average fetal waveform.

[0137] xi). Option to display a rhythm strip of a standard length on the screen or of particular section of interest.

[0138] xii). Option to display a full set of rhythm strips such as one minute on one page.

[0139] xiii). Option to enter and display patient details on screen.

[0140] xiv). Automatic generation of a patient record containing all the parameters of interest along with the average fECG waveform, the fetal heart rate over the interval of data of interest and the chosen rhythm strip which might include particular abnormalities or features of interest.

[0141] xv). Automatic appending of results to a database of previously generated patient data.

[0142] xvi). Abdominal surface map of features of interest such as source strength over the abdomen, or average fECG waveform at the positions of the electrodes.

[0143] xvii). Ability to zoom in or out of the display in order to focus on fine detail in any view for detailed analysis of features such as the structure of the heart-beat, heart rate or waveform.

[0144] In addition to monitoring the fECG during pregnancy, the apparatus described herein may also be used during labour. This may appear somewhat at odds with the requirement for the mother to be relaxed, but very useful measurements can be extracted at certain times, such as the time following a contraction. Whilst maternal muscle noise and uterine contractions will add to the background noise level, there are at least two mitigating factors that make the problem less intractable. First, high-risk pregnancies are often delivered under epidural anaesthetic. Under these conditions the mother's movements are limited. Secondly, it is the variation in the fetal heart activity in response to (i.e. soon after) a uterine contraction that is of particular clinical interest. Accordingly, it is not essential (although it may be desirable) for monitoring to continue throughout contractions. To this end it is envisaged that the signal processing technique may be varied so as to make use of information gained between contractions in order to follow the fetal heart activity through the next contraction.

[0145] It has been shown by invasive fECG using a fetal scalp electrode that analysis of the ST segment of the heart-beat in particular has diagnostic value during labour for the

identification of hypoxia and fetal distress. Use of apparatus according to this invention in the manner described above will enable this information to be obtained non-invasively. The display should also be altered in order to present information useful during labour such as, for example, a continuous plot of the fetal heart activity alongside the maternal heart activity and uterine contractions in real time. Real-time signal separation may be achieved, for example, by consecutive processing of a number of overlapping blocks of data. The outputs from this processing can then be aligned to produce a continuous separated fetal signal using correlation of the output of consecutive blocks of data.

1. Apparatus for the detection of a fetal electrocardiogram (fECG), the apparatus comprising:

a plurality of low-noise electrodes for external skin placement during pregnancy, each electrode being connectable to low-noise signal recording and processing means wherein, when in use, the signal recording and processing means is arranged:

to record signals indicative of voltages developed between pairs of electrodes in a plurality of signal channels;

to process digitised data within the plurality of signal channels in order to generate at least one source signal which relates to the fECG of a single fetus; and

on identification of the at least one source signal relating to the single-fetus fECG, to reconstruct, for at least one signal channel, that component of the digitised data within the channel which is attributable to that fetus, and therefore corresponds to a single-channel fECG.

2. Apparatus according to claim 1 wherein the signal recording and processing means is arranged, when processing the digitised data, to generate a plurality of separated source signals, at least one of which relates to the fECG of a single fetus.

3. Apparatus according to claim 1 wherein the electrodes are placeable on the skin such that skin impedance at each electrode is less than 5 k $\Omega$ .

4. Apparatus according to claim 3 wherein the electrodes are placeable such that the skin impedance at each electrode is less than 2 k $\Omega$ .

5. Apparatus according to claim 1 wherein the electrodes comprise a plurality of abdominal electrodes for placement on the skin in the abdominal area and a common reference electrode, and the signal recording and processing means is arranged to record signals indicative of voltages developed between each abdominal electrode and the reference electrode.

6. Apparatus according to claim 5 wherein the apparatus also includes a low-noise electrode which is connected to earth.

7. Apparatus according to claim 5 wherein the signal recording and processing means comprises electronic components for processing analogue voltage signals developed between pairs of electrodes to provide digital signals in the plurality of signal channels and data processing means to process the digital signals.

8. Apparatus according to claim 7 wherein the electronic components comprise a low-noise differential amplifier for amplifying the difference between each electrode's voltage signal and a signal derived from the voltage developed at the reference electrode, an anti-aliasing low-pass filter and an analogue to digital converter.

9. Apparatus according to claim 8 wherein the electronic components are located in a multichannel lead box, remote

from a patient, and connectable to the electrodes by, for each electrode, respective screened leads.

10. Apparatus according to claim 9 wherein the multichannel lead box 34 is suitable for use in taking electroencephalography (EEG) scans.

11. Apparatus according to claim 9 wherein the electronic components additionally include, for each electrode, a pre-amplifier located adjacent that electrode and connectable to the remote lead box by respective screened leads.

12. Apparatus according to claim 8 wherein the electronic components are located in a lead box adjacent to a patient and, for the pre-amplifiers, in the vicinity of the electrodes wherein the lead box and data processing means are in communication via a wireless link.

13. Apparatus according to claim 1 wherein the signal channels exhibit a noise component within raw composite data of less than 10  $\mu$ V.

14. Apparatus according to claim 13 wherein the signal channels exhibit a noise component of less than 5  $\mu$ V and ideally less than 3  $\mu$ V.

15. Apparatus according to claim 1 wherein more than one source signal relates to the same fetal fECG and the signal recording and processing means is arranged, on identification of the more than one source signal relating to the single fetus to combine these source signals with appropriate weighting in the reconstruction of the single-channel fECG.

16. Apparatus according to claim 15 for the detection of fetal electrocardiograms in a multiple pregnancy, wherein the signal recording and processing means is arranged, on identification of the at least one source signal relating to each fetus, to reconstruct, for at least one abdominal electrode, components of the digitised data within the corresponding signal channel which is attributable to each fetus.

17. Apparatus according to claim 16 wherein the signal recording and processing means is arranged to reconstruct, for each abdominal electrode, components of the digitised data within each corresponding signal channel which is attributable to each fetus and thereby to construct an abdominal surface intensity map of signal strength from each fetus.

18. Apparatus according to claim 1 wherein signal recording and processing means is arranged to generate the separated source signals by Independent Component Analysis (ICA).

19. Apparatus according to claim 1 wherein the electrodes are self-adhesive.

20. A method of recording a fetal electrocardiogram (fECG) comprising the steps of:

- (a) attaching a plurality of abdominal electrodes to skin in an abdominal area during pregnancy;
- (b) attaching a common reference electrode to the skin;
- (c) measuring skin impedance at each electrode and, if the skin impedance is greater than 5 k, reattaching that electrode;
- (d) observing electrical skin potentials developed between each abdominal electrode and the reference electrode through a plurality of signal channels;
- (e) if the noise on any raw data composite signal channel is greater than 10  $\mu$ V, reducing environmental noise contributory factors, including that due to muscle noise;
- (f) collecting electrical skin potentials through the plurality of signal channels;
- (g) processing data within the plurality of signal channels to generate a plurality of separated source signals;

(h) identifying within the plurality of source signals those which relate to the fECG of one among one or more fetuses; and

(i) reconstructing, for at least one of the plurality of signal channels respective components of the data within that channel which are attributable to each of the one or more fetuses, thereby generating a single-channel fECG for each individual fetus.

21. A method according to claim 20 wherein at step (c), the electrode is reattached if the skin impedance is greater than 2 k.

22. A method according to claim 20 wherein at step (e) measures are taken to reduce environmental noise contributory factors if the noise within any channel is greater than 5  $\mu$ V.

23. A method according to claim 20 wherein it includes the additional step of characterising the fECG for each fetus by examination of fetal heart-beat waveform and extraction of diagnostic indicators such as PR, QRS and QT intervals.

24. A computer system configured to extract one or more fetal electrocardiograms from composite signals detected at electrodes distributed over a pregnant woman's abdomen, the system comprising:

filtering means for filtering digitised composite signals, each signal corresponding to a difference between a voltage developed at a signal electrode and that developed at a reference electrode, in order to remove unwanted frequency components;

processing means arranged to generate a plurality of separated source signals from the corresponding plurality of filtered composite signals, wherein each composite data signal is assumed to be a linear mixture of the unknown source signals ( $X=SM$ );

selection means for identifying those source signal(s) which correspond to a single-fetus ECG;

reconstruction means arranged, for each single-fetus ECG identified, to reconstruct, for at least one signal channel, that component of the filtered composite signal within the channel which is attributable to that fetus and therefore corresponds to a single-channel fECG; and

display means arranged to display each fECG.

25. A computer-readable medium embodying instructions for execution by a processor, the instructions relating to the extraction of one or more fetal electrocardiograms from composite signals detected at electrodes distributed over a pregnant woman's abdomen, wherein the computer-readable medium comprises:

program code for filtering multi-channel digitised composite signals, each signal corresponding to a difference between a voltage developed at a signal electrode and that developed at a reference electrode, in order to remove unwanted frequency components;

program code for generating a plurality of separated source signals from the corresponding plurality of filtered composite signals, wherein each composite data signal is assumed to be a linear mixture of the unknown source signals ( $X=SM$ );

program code for identifying, either automatically or by prompting for a user input, those source signal(s) which correspond to a single-fetus ECG;

program code for reconstructing, for each single-fetus ECG identified, and for at least one signal channel, that component of the filtered composite signal within the

channel which is attributable to that fetus and therefore corresponds to a single-channel fECG; and program code for displaying at least one reconstructed fECG to a user.

**26.** Apparatus for the detection of a fetal electrocardiogram (fECG), the apparatus comprising:

- (a) a plurality of low-noise electrodes for abdominal skin placement during pregnancy, the electrodes being in number sufficient to enable at least eight signal sources to be monitored; and
- (b) low-noise signal processing apparatus connectable to the electrodes and of sufficient sensitivity to detect signals of magnitude comparable with electroencephalography signals, the signal processing apparatus being for:
  - i. detecting signals developed by the electrodes; and
  - ii. processing electrode signals to obtain data and using a Blind Source Separation technique to process the data and distinguish independent sources and thereby to derive at least one source signal which relates to a single fetus fECG.

**27.** A method of recording a fetal electrocardiogram (fECG) comprising the steps of:

- (a) during pregnancy, attaching a plurality of low-noise electrodes to a patient's skin providing coverage from one side of the patient's abdomen to the other and from pubic hairline to upper limit of uterus, the electrodes being in number sufficient to enable at least eight signal sources to be monitored;
- (b) reattaching any and all electrodes which do not meet a predetermined skin impedance criterion;
- (c) if noise associated with any electrode does not meet a predetermined noise criterion, reducing at least one of electrical interference, magnetic pickup and patient muscle noise until the noise criterion is met;
- (d) detecting signals developed by the electrodes and processing such signals to obtain data using low-noise signal processing apparatus of sufficient sensitivity to detect signals of magnitude comparable with electroencephalography signals; and
- (e) processing the data with a Blind Source Separation technique to distinguish independent sources and thereby to derive at least one source signal relating to a single fetus fECG.

\* \* \* \* \*

专利名称(译)	监测胎儿心跳的装置		
公开(公告)号	<a href="#">US20080183092A1</a>	公开(公告)日	2008-07-31
申请号	US12/026267	申请日	2008-02-05
[标]申请(专利权)人(译)	秦内蒂克有限公司		
申请(专利权)人(译)	奎奈蒂克有限公司		
当前申请(专利权)人(译)	奎奈蒂克有限公司		
[标]发明人	SMITH MARK JOHN PENNEY RICHARD WILLIAM		
发明人	SMITH, MARK JOHN PENNEY, RICHARD WILLIAM		
IPC分类号	A61B5/0444 A61B5/05 A61B5/00 A61B5/0402 A61B5/0408 A61B5/0428 A61B5/0472 A61B5/0476 A61B5/0478 A61B5/0492		
CPC分类号	A61B5/0006 A61B5/4362 A61B5/0444 A61B5/0011		
优先权	2001023772 2001-10-03 GB PCT/GB2002/004410 2002-09-27 WO		
外部链接	<a href="#">Espacenet</a> <a href="#">USPTO</a>		

摘要(译)

用于监测胎儿心跳的装置能够从在孕妇腹部检测到的复合信号中提取一个或多个胎儿心电图 (fECG)。该装置包括特定的低噪声部件：用于在怀孕期间放置在腹部上的多个低噪声电极 (1-3, R) 和低噪声信号记录和处理装置 (34)。根据需要, 还使用屏蔽引线。指示在每个腹部电极 (1-3) 和参考电极 (R) 之间产生的电压的信号被记录在多个信号通道中。对每个通道内的数据进行数字化和处理, 以产生多个分离的源信号, 其中至少一个源信号与单个胎儿的fECG有关。可以使用被识别为属于同一胎儿的源信号来重建单通道fECG。

