



US008463343B2

(12) **United States Patent**
Kuhn et al.

(10) **Patent No.:** **US 8,463,343 B2**
(45) **Date of Patent:** ***Jun. 11, 2013**

(54) **OPTICAL SENSOR FOR MEDICAL DEVICE**

(56) **References Cited**

(75) Inventors: **Jonathan L. Kuhn**, Ham Lake, MN (US); **Thomas A. Anderson**, New Hope, MN (US); **Can Cinbis**, Shoreview, MN (US); **Jeffrey M. Jelen**, New Hope, MN (US); **Timothy Davis**, Coon Rapids, MN (US); **James K. Carney**, Brooklyn Park, MN (US)

U.S. PATENT DOCUMENTS

3,985,449 A	10/1976	Patrin
4,100,562 A	7/1978	Sugawara et al.
4,202,339 A	5/1980	Wirtzfeld et al.
4,467,807 A	8/1984	Bornzin
4,730,389 A	3/1988	Baudino et al.
5,010,381 A	4/1991	Shiba
5,144,381 A	9/1992	Furuyama et al.

(Continued)

FOREIGN PATENT DOCUMENTS

WO	0025862 A1	5/2000
WO	2007/122375 A2	11/2007

(Continued)

OTHER PUBLICATIONS

Reply to Written Opinion dated Oct. 19, 2010, from international application No. PCT/US2010/032096, filed Nov. 28, 2011, 15 pp.

(Continued)

Primary Examiner — Rochelle-Ann J Blackman

(74) *Attorney, Agent, or Firm* — Evans M. Mburu; Stephen W. Bauer

(57) **ABSTRACT**

An optical sensor for a medical device includes a fixed lens spacing between emit and receive modules to achieve target sensor sensitivity, while varying other sensor parameters in order to increase signal amplitude without increasing power demand. An optical sensor connected to a housing of a medical device includes a circuit board, an opto-electronic component, a wall, a lens, and a ferrule. The circuit board is arranged within the housing. The opto-electronic component is mounted on a surface of the circuit board. The wall protrudes from the surface of the circuit board and surrounds the opto-electronic component. The lens is offset from the surface of the circuit board. The ferrule is connected to the housing, the lens and the wall. An inner surface of the wall mates with an outer surface of the ferrule.

28 Claims, 12 Drawing Sheets

(73) Assignee: **Medtronic, Inc.**, Minneapolis, MN (US)

(*) Notice: Subject to any disclaimer, the term of this patent is extended or adjusted under 35 U.S.C. 154(b) by 420 days.

This patent is subject to a terminal disclaimer.

(21) Appl. No.: **12/765,396**

(22) Filed: **Apr. 22, 2010**

(65) **Prior Publication Data**

US 2011/0190609 A1 Aug. 4, 2011


Related U.S. Application Data

(60) Provisional application No. 61/299,705, filed on Jan. 29, 2010.

(51) **Int. Cl.**
A61B 5/00 (2006.01)

(52) **U.S. Cl.**
USPC **600/316; 600/325; 600/333; 600/334; 600/339**

(58) **Field of Classification Search**
USPC **600/316, 310, 322-327, 333, 339, 600/341; 438/65; 257/85, E31.001**
See application file for complete search history.



U.S. PATENT DOCUMENTS

5,556,421 A 9/1996 Prutchi et al.
 5,902,326 A 5/1999 Lessar et al.
 6,125,290 A 9/2000 Miesel
 6,144,866 A 11/2000 Miesel et al.
 6,198,952 B1 3/2001 Miesel
 6,356,774 B1 3/2002 Bernstein et al.
 6,409,675 B1 6/2002 Turcott
 6,491,639 B1 12/2002 Turcott
 6,661,167 B2 12/2003 Eliashovich et al.
 6,711,423 B2 3/2004 Colvin
 7,013,178 B2 3/2006 Reinke et al.
 7,152,977 B2 12/2006 Ruda et al.
 7,167,309 B2 1/2007 Saxena et al.
 7,415,298 B2 8/2008 Casciani et al.
 2004/0176669 A1 9/2004 Colvin, Jr.
 2004/0197267 A1 10/2004 Black et al.
 2004/0220629 A1 11/2004 Kamath et al.
 2004/0246744 A1 12/2004 Krupa et al.
 2005/0006651 A1 1/2005 LeBoeuf et al.
 2006/0255353 A1 11/2006 Taskar et al.
 2009/0076353 A1 3/2009 Carpenter et al.
 2009/0156905 A1 6/2009 Ries et al.
 2009/0156912 A1 6/2009 Kuhn et al.
 2009/0326346 A1 12/2009 Kracker et al.
 2010/0185262 A1 7/2010 Kuhn et al.

FOREIGN PATENT DOCUMENTS

WO 2008118042 A1 10/2008
 WO 2009/076063 A1 6/2009
 WO 2009137426 A1 11/2009

OTHER PUBLICATIONS

Reply to Written Opinion dated Oct. 8, 2010, from international application No. PCT/US2010/032091, filed Nov. 28, 2011, 18 pp.
 Reply to Written Opinion dated Oct. 15, 2010, from international application No. PCT/US2010/032080, filed Nov. 28, 2011, 11 pp.

Thomas Jensen et al., "Independent Component Analysis Applied to Pulse Oximetry in the Estimation of the Arterial Oxygen Saturation (S_pO_2)— a Comparative Study", 31st Annual International Conference of the IEEE EMBS, Minneapolis, Minnesota Sep. 2-6, 2009 (pp. 4039-4044).

Sune Duun et al., "A Novel Ring Shaped Photodiode for Reflectance Pulse Oximetry in Wireless Applications", IEEE Sensors Conference, 2007 (pp. 596-599).

U.S. Appl. No. 61/146,469, filed Jan. 22, 2009 entitled "Co-Location of Emitters and Detectors and Method of Operation".

U.S. Appl. No. 12/765,345, filed Apr. 22, 2010 entitled "Optical Sensor for Medical Device".

U.S. Appl. No. 12/765,434, filed Apr. 22, 2010 entitled "Optical Sensor for Medical Device".

Written Opinion dated Apr. 10, 2012, from international application No. PCT/US2010/032091, 11 pp.

Written Opinion dated Apr. 10, 2012, from international application No. PCT/US2010/032096, 10 pp.

Written Opinion dated Apr. 10, 2012, from international application No. PCT/US2010/032080, 10 pp.

Mendelson et al. "Design and Evaluation of a New Reflectance Pulse Oximeter Sensor," Medical Instrumentation, Philadelphia, PA, vol. 22, No. 4, Aug. 1, 1988, pp. 167-173.

International Search Report and Written Opinion of international application No. PCT/US2010/032080, dated Oct. 15, 2010, 13 pp.

Kuhn, et al., "Optical Sensor for Medical Device", filed Apr. 22, 2010, U.S. Appl. No. 12/765,345, 37 pages.

Kuhn, et al., "Optical Sensor for Medical Device", filed Apr. 22, 2010, U.S. Appl. No. 12/765,434, 36 pages.

International Preliminary Report on Patentability from international application No. PCT/US2010/032091, dated Jun. 15, 2012, 28 pp.

International Preliminary Report on Patentability from international application No. PCT/US2010/032096, dated Jun. 15, 2012, 23 pp.

International Preliminary Report on Patentability from international application No. PCT/US2010/032080, dated Jun. 15, 2012, 19 pp.




FIG. 9




FIG. 10




FIG. 11




FIG. 12

OPTICAL SENSOR FOR MEDICAL DEVICE

This application claims the benefit of U.S. Provisional Application No. 61/299,705, filed Jan. 29, 2010, the entire content of which is incorporated herein by this reference.

TECHNICAL FIELD

This disclosure relates to medical devices, and, more particularly, to optical sensors employed in such devices.

BACKGROUND

A variety of implantable medical devices for delivering a therapy and/or monitoring a physiological condition have been clinically implanted or proposed for clinical implantation in patients. Implantable medical devices may deliver electrical stimulation or drug therapy to, and/or monitor conditions associated with, the heart, muscle, nerve, brain, stomach or other organs or tissue, as examples. Some implantable medical devices may employ one or more elongated electrical leads carrying stimulation electrodes, sense electrodes, and/or other sensors. Implantable medical leads may be configured to allow electrodes or other sensors to be positioned at desired locations for delivery of stimulation or sensing. For example, electrodes or sensors may be carried at a distal portion of a lead. A proximal portion of the lead may be coupled to an implantable medical device housing, which may contain circuitry such as stimulation generation and/or sensing circuitry. Other implantable medical devices may employ one or more catheters through which the devices deliver a therapeutic fluid to a target site within a patient.

Implantable medical devices may include one or more physiological sensors, which may be used in conjunction with the device to provide signals related to various physiological conditions from which a patient state or the need for a therapy can be assessed. Examples of such implantable medical devices include heart monitors, pacemakers, implantable cardioverter defibrillators (ICDs), myostimulators, neurostimulators, therapeutic fluid delivery devices, insulin pumps, and glucose monitors.

Optical sensors may be employed in implantable medical devices as physiological sensors configured to detect changes in light modulation by a body fluid or tissue volume caused by a change in a physiological condition in the body fluid or tissue. Such optical sensors can be used, for example, for detecting changes in metabolite levels in the blood, such as oxygen saturation levels or glucose level, or changes in tissue perfusion. Monitoring such physiological conditions provides useful diagnostic measures, and can be used in managing therapies for treating a medical condition. For example, a decrease in blood oxygen saturation or in tissue perfusion may be associated with insufficient cardiac output or respiratory function. Thus monitoring such signals allows an implantable medical device to respond to a decrease in oxygen saturation or tissue perfusion, for example by delivering electrical stimulation therapies to the heart to restore a normal hemodynamic function.

SUMMARY

In general, examples disclosed herein are directed to an optical sensor for a medical device that includes a fixed lens spacing between emit and receive modules to achieve target sensor sensitivity, while varying other sensor parameters in order to increase signal amplitude without increasing power demand.

In one example, a medical device includes a housing and an optical sensor module connected to the housing. The optical sensor module includes a circuit board, an opto-electronic component, a wall, a lens, and a ferrule. The circuit board is arranged within the housing. The opto-electronic component is mounted on a surface of the circuit board. The wall protrudes from the surface of the circuit board and surrounds the opto-electronic component. The lens is offset from the surface of the circuit board. The ferrule is connected to the housing, the lens and the wall. An inner surface of the wall mates with an outer surface of the ferrule.

In another example, an optical sensor connected to a housing of a medical device includes a circuit board, an opto-electronic component, a wall, a lens, and a ferrule. The circuit board is arranged within the housing. The opto-electronic component is mounted on a surface of the circuit board. The wall protrudes from the surface of the circuit board and surrounds the opto-electronic component. The lens is offset from the surface of the circuit board. The ferrule is connected to the housing, the lens and the wall. An inner surface of the wall mates with an outer surface of the ferrule.

In another example, A method of constructing an optical sensor connected to a housing of a medical device, the method includes connecting a lens to a ferrule, connecting the ferrule to the housing, mounting an opto-electronic component on a surface of a circuit board, connecting a wall to the circuit board to protrude from the surface of the circuit board and surround the opto-electronic component, and connecting the wall to the ferrule such that an inner surface of the wall mates with an outer surface of the ferrule.

The details of one or more examples are set forth in the accompanying drawings and the description below. Other features, objects, and advantages of the disclosure will be apparent from the description and drawings, and from the claims.

BRIEF DESCRIPTION OF DRAWINGS

FIG. 1 is a conceptual drawing illustrating an example system that includes an implantable medical device (IMD) coupled to implantable medical leads.

FIG. 2 is a conceptual drawing illustrating the example IMD of FIG. 1 including a number of optical sensor modules.

FIG. 3 is a functional block diagram illustrating an example configuration of the IMD of FIG. 1.

FIG. 4 is a functional block diagram illustrating an example configuration of an external programmer that facilitates user communication with the IMD.

FIG. 5 is a block diagram illustrating an example system that includes an external device, such as a server, and one or more computing devices that are coupled to the IMD and programmer shown in FIG. 1 via a network.

FIGS. 6A and 6B are conceptual drawings respectively illustrating a top view and a cross-sectional view of an example optical sensor.

FIGS. 7A and 7B are plan and section views, respectively, of a ferrule of an optical sensor module including a flange masking the module lens.

FIG. 8 is a partial section view of one module of the optical sensor of FIGS. 6A and 6B.

FIG. 9 is a conceptual drawing illustrating an example medical lead including an optical sensor.

FIGS. 10-12 are flow diagrams illustrating several example methods of constructing an optical sensor according to this disclosure connected to a housing of an IMD.

DETAILED DESCRIPTION

FIG. 1 is a conceptual diagram illustrating an example system 10 that may be used for sensing of physiological

parameters of patient **14** and/or to provide therapy to heart **12** of patient **14**. Therapy system **10** includes IMD **16**, which is coupled to leads **18**, **20**, and **22**, and programmer **24**. IMD **16** may be, for example, an implantable pacemaker, cardioverter, and/or defibrillator that provides electrical signals to heart **12** via electrodes coupled to one or more of leads **18**, **20**, and **22**. Patient **14** is ordinarily, but not necessarily, a human patient.

Leads **18**, **20**, **22** extend into the heart **12** of patient **16** to sense electrical activity of heart **12** and/or deliver electrical stimulation to heart **12**. In the example shown in FIG. **1**, right ventricular (RV) lead **18** extends through one or more veins (not shown), the superior vena cava (not shown), and right atrium **26**, and into right ventricle **28**. Left ventricular (LV) coronary sinus lead **20** extends through one or more veins, the vena cava, right atrium **26**, and into the coronary sinus **30** to a region adjacent to the free wall of left ventricle **32** of heart **12**. Right atrial (RA) lead **22** extends through one or more veins and the vena cava, and into the right atrium **26** of heart **12**.

System **10** may, in some examples, additionally or alternatively include one or more leads or lead segments (not shown in FIG. **1**) that deploy one or more electrodes within the vena cava or other vein. These electrodes may allow alternative electrical sensing configurations that may provide improved or supplemental sensing in some patients. Furthermore, in some examples, therapy system **10** may include temporary or permanent epicardial or subcutaneous leads, instead of or in addition to leads **18**, **20** and **22**. Such leads may be used for one or more of cardiac sensing, pacing, or cardioversion/defibrillation.

IMD **16** may sense electrical signals attendant to the depolarization and repolarization of heart **12** via electrodes (not shown in FIG. **1**) coupled to at least one of the leads **18**, **20**, **22**. In some examples, IMD **16** provides pacing pulses to heart **12** based on the electrical signals sensed within heart **12**. The configurations of electrodes used by IMD **16** for sensing and pacing may be unipolar or bipolar. IMD **16** may detect arrhythmia of heart **12**, such as tachycardia or fibrillation of ventricles **28** and **32**, and may also provide defibrillation therapy and/or cardioversion therapy via electrodes located on at least one of the leads **18**, **20**, **22**. In some examples, IMD **16** may be programmed to deliver a progression of therapies, e.g., pulses with increasing energy levels, until a fibrillation of heart **12** is stopped. IMD **16** detects fibrillation employing any of a number of known fibrillation detection techniques.

In addition to employing electrodes connected to one of the leads **18**, **20**, **22** and/or the housing of IMD **16** to sense the physiological state of patient **14**, IMD **16** includes at least one optical sensor that may be configured to detect changes in light modulation by a body fluid or tissue volume of the patient caused by a change in a physiological condition in the body fluid or tissue. For example, the optical sensor(s) of IMD **16** may be employed to detect changes in metabolite levels in the blood of patient **14** including, e.g., oxygen saturation levels or glucose level, or changes in tissue perfusion. IMD **16** may employ the optical sensor(s) to monitor physiological conditions of patient **14** for diagnostic purposes and/or to manage therapy delivered to the patient. For example, a decrease in blood oxygen saturation or in tissue perfusion may be associated with insufficient cardiac output or respiratory function. By monitoring these physiological parameters IMD **16** may, for example, respond to a decrease in oxygen saturation or tissue perfusion by delivering electrical stimulation via one or more of leads **18**, **20**, and **22** to heart **12** of patient **14** to restore normal hemodynamic function.

Programmer **24** shown in FIG. **1** may be a handheld computing device, computer workstation, or networked comput-

ing device. Programmer **24** may include a user interface that receives input from a user. The user interface may include, for example, a keypad and a display, which may for example, be a cathode ray tube (CRT) display, a liquid crystal display (LCD), an organic or standard light emitting diode (OLED/LED) display. The keypad may take the form of an alphanumeric keypad or a reduced set of keys associated with particular functions. Programmer **24** can additionally or alternatively include a peripheral pointing device, such as a mouse, via which a user may interact with the user interface. In some examples, a display of programmer **24** may include a touch screen display, and a user may interact with programmer **24** via the display. In some examples, the user may also interact with programmer **24** remotely via a networked computing device.

A user, such as a physician, technician, surgeon, electrophysiologist, or other clinician, may interact with programmer **24** to communicate with IMD **16**. For example, the user may interact with programmer **24** to retrieve physiological or diagnostic information from IMD **16** that was obtained based on signals received from the optical sensors of the device. A user may also interact with programmer **24** to program IMD **16**, e.g., to select values for operational parameters of the IMD.

For example, the user may use programmer **24** to retrieve information from IMD **16** regarding the rhythm of heart **12**, trends therein over time, or arrhythmic episodes. As another example, the user may use programmer **24** to retrieve information from IMD **16** regarding other sensed physiological parameters of patient **14**, such as intracardiac or intravascular pressure, activity, posture, respiration, or thoracic impedance. The user may also employ programmer **24** to retrieve information from IMD **16** regarding physiological parameters of patient **14** detected by the optical sensor(s) of the device including, e.g., blood oxygen saturation and tissue perfusion. As another example, the user may use programmer **24** to retrieve information from IMD **16** regarding the performance or integrity of IMD **16** or other components of system **10**, such as leads **18**, **20** and **22**, or a power source of IMD **16**. In some examples, this information may be presented to the user as an alert.

The user may use programmer **24** to program a therapy progression, select electrodes used to deliver defibrillation pulses, select waveforms for the defibrillation pulse, or select or configure a fibrillation detection algorithm for IMD **16**. The user may also use programmer **24** to program aspects of other therapies provided by IMD **16**, such as cardioversion or pacing therapies. IMD **16** and programmer **24** may communicate via wireless communication using any number of known techniques.

FIG. **2** is a conceptual diagram illustrating IMD **16** including optical sensors **42** and **44**, each of which include emit module **46** and receive module **48** laterally offset from one another. While, generally speaking, emit module **46** of sensor **42** emits optical signals to receive module **48** of sensor **42** and emit module **46** of sensor **44** emits optical signals to receive module **48** of sensor **44**, examples according to this disclosure include emit module **46** of sensor **42** emitting optical signals to receive module **48** of sensor **44**, as well as other emit and receive module combinations of sensors **42** and **44**. Furthermore, although IMD **16** of FIG. **2** includes two optical sensors **42**, **44**, other example devices may include fewer or more optical sensors. Finally, as described in greater detail with reference to FIG. **9**, optical sensors employed in therapy system **10** of FIG. **1**, or other medical systems, may be positioned in locations other than connected to IMD **16**. For example, optical sensors **42**, **44** may be connected to one or

more of leads **18**, **20**, or **22**. In another example, optical sensors may be located in a separate device that communicates via telemetry with IMD **16** or communicates via telemetry or wire with Programmer **24**.

Referring again to FIG. **2**, each of emit and receive module **46**, **48** include an opto-electronic component (not shown) arranged within a cavity in housing **17** of IMD **16**. The cavity is generally defined by a circuit board (not shown), a side wall (not shown) protruding from the circuit board and surrounding the opto-electronic component, a lens offset from the circuit board, and a ferrule (**50** or **52** for emit and receive modules **46**, **48** respectively) connected to housing **17** of IMD **16**, the lens and the side wall (not shown). As used herein, the term “opto-electronic component” refers to any electrical circuit component capable of emitting light in response to an applied voltage or current or emitting current or changing an inherent physical property in response to exposure to light, including, e.g., light emitting diodes (LEDs), vertical cavity surface emitting lasers (VCSELs), photoresistors or light dependent resistors, photodiodes, phototransistors, photovoltaic cells, or charge-coupled devices.

In order to detect changes in a physiological condition of patient **14**, the opto-electronic component of one of the emit modules **46** emits an optical signal out of transparent lens **54** into adjacent body fluid or tissue volume of the patient. The optical signal from emit module **46** interacts with the adjacent body fluid or tissue volume of patient **14**, after which a portion of the optical signal modulated by the body fluid or tissue volume of patient **14** is reflected off of the fluid or volume through lens **56** of one of receive modules **48**. The optical signal received by receive module **48** is detected by the opto-electronic component of the receive module. The modulation of the optical signal received by receive module **48** from emit module **46** may be interpreted by IMD **16** to detect a physiological condition of patient **14**, e.g. to detect blood oxygen saturation or tissue perfusion of the patient. Employing sensors **42**, **44**, IMD **16** may, for example, respond to decreases in oxygen saturation or tissue perfusion by delivering electrical stimulation via one or more of leads **18**, **20**, and **22** to heart **12** of patient **14** to restore normal hemodynamic function. Optical sensors **42**, **44** may be operated continuously at a relatively low duty cycle, or the sensors may be triggered by other sensor signals including, e.g., electrocardiogram (ECG) information for patient **14** produced by electrodes connected to one or more of leads **18**, **20**, **22** and/or housing **17** of IMD **16** that signals an episode of interest.

Generally speaking, increasing distance between emit and receive lenses **54**, **56** of emit and receive modules **46**, **48** respectively increases sensitivity, thereby increasing the granularity of physiological change each of sensors **42**, **44** is capable of detecting and improving the limits within which signals from the emit module are detectable by receive module. Although the precise relationship between lens distance and sensor sensitivity may not be known, empirical data gathered from a study of an implantable medical device including optical sensors similar to those described in this disclosure showed that 10 millimeter center-to-center lens spacing provides substantially improved sensitivity over 7 millimeters in sensors with emit and receive modules including approximately 2.5 millimeter diameter windows. However, the amplitude of the optical signal that reaches the opto-electronic component of the receive module decreases exponentially as lens spacing is increased. Although signal amplitude may be increased by increasing power to the emit module of the optical sensor, such increased power demand

disadvantageously decreases device longevity and increases detection times by decreasing the thermal stability of the sensor.

In view of the foregoing relationships between sensor configuration and performance, the examples described in this disclosure are directed to techniques for improving optical sensor signal amplitude for a target sensor sensitivity. Lens spacing is one of the most significant controlling parameters for optical signal body fluid or tissue interaction, i.e. optical sensor sensitivity. As such, the disclosed techniques fix lens spacing to achieve target sensor sensitivity, while varying other sensor parameters in order to increase signal amplitude without increasing power demand. In particular, the disclosed examples select the size, D_E and D_R of emit and receive lenses **54**, **56** respectively in FIG. **2**, and the offset between each of lenses **54**, **56**, and the respective opto-electronic components of each of emit and receive modules **46**, **48** to increase the amplitude of the signal that is received by the receive module of an optical sensor from the emit module. Additionally, in some examples, the arrangement of the opto-electronic component within the receive module is improved by masking the receive module lens with an opaque member to create a masked lens leading edge that is aligned with a leading edge of the opto-electronic component and that acts to effectively increase the edge-to-edge lens spacing between the emit and receive modules.

Masking receive lens **56** may improve the sensitivity of the light receiving opto-electronic component versus lens spacing, e.g. edge-to-edge distance. Without the mask, due to the exponential decay of light versus distance from the light emitting opto-electronic component to the light receiving opto-electronic component, a disproportionately large amount of the signal produced by the light receiving opto-electronic component comes from the “leading edge” of receive lens **56** nearest to emit lens **54**. As you move away from this leading edge or point on receive lens **56**, the signal quality is improved, but the amplitude decays exponentially. By masking lens **56** of receive module **48** the light emitted by emit module **46** and received by the receive module is more uniformly distributed across the light receiving opto-electronic component of the receive module.

In this disclosure, spacing between emit and receive lenses **54**, **56** is generally expressed as edge-to-edge lens distance, E , as shown in FIG. **2**. However, lens spacing may, in other examples, be expressed as other distances including, e.g., center-to-center lens spacing in optical sensors employing circular windows.

The configuration of therapy system **10** illustrated in FIGS. **1** and **2** is merely one example. In other examples, a therapy system may include epicardial leads and/or patch electrodes instead of or in addition to the transvenous leads **18**, **20**, **22** illustrated in FIG. **1**. Further, IMD **16** need not be implanted within patient **14**. In examples in which IMD **16** is not implanted in patient **14**, IMD **16** may deliver defibrillation pulses and other therapies to heart **12** via percutaneous leads that extend through the skin of patient **14** to a variety of positions within or outside of heart **12**. As noted above, in some examples, IMD **16** may be implanted along with a separate sensor device that includes optical sensors **42**, **44** communicatively connected to the IMD. Such a separate sensor device may be either implanted at the same or a different site from IMD **16** or may be external to patient **14**.

In addition, in other examples, a therapy system may include any suitable number of leads coupled to IMD **16**, and each of the leads may extend to any location within or proximate to heart **12**. For example, other examples of therapy systems may include three transvenous leads located as illus-

trated in FIGS. 1 and 2, and an additional lead located within or proximate to left atrium 36. Other example therapy systems may include a single lead that extends from IMD 16 into right atrium 26 or right ventricle 28, or two leads that extend into a respective one of the right ventricle 26 and right atrium 26. Additionally, in examples in which IMD 16 or a separate sensor device including sensors 42, 44 is employed for integrated diagnostic applications, there may be no external leads or other therapeutic components.

Although optical sensing techniques according to this disclosure are described in the context of devices including cardiac stimulation and/or sensing leads, the examples disclosed herein may also be employed in other types of implantable medical devices including, e.g., myostimulators, neurostimulators, therapeutic fluid delivery devices, insulin pumps, and glucose monitors.

FIG. 3 is a functional block diagram illustrating an example configuration of IMD 16. In the illustrated example, IMD 16 includes a processor 80, memory 82, signal generator 84, electrical sensing module 86, optical sensor module 87, telemetry module 88, and power source 90. Memory 82 includes computer-readable instructions that, when executed by processor 80, cause IMD 16 and processor 80 to perform various functions attributed to IMD 16 and processor 80 herein. Memory 82 may include any volatile, non-volatile, magnetic, optical, or electrical media, such as a random access memory (RAM), read-only memory (ROM), non-volatile RAM (NVRAM), electrically-erasable programmable ROM (EEPROM), flash memory, or any other digital or analog media.

Processor 80 may include any one or more of a microprocessor, a controller, a digital signal processor (DSP), an application specific integrated circuit (ASIC), a field-programmable gate array (FPGA), or equivalent discrete or analog logic circuitry. In some examples, processor 80 may include multiple components, such as any combination of one or more microprocessors, one or more controllers, one or more DSPs, one or more ASICs, or one or more FPGAs, as well as other discrete or integrated logic circuitry. The functions attributed to processor 80 herein may be embodied as software, firmware, hardware or any combination thereof.

Processor 80 controls signal generator 84 to deliver stimulation therapy to heart 12 according to one or more of therapy programs, which may be stored in memory 82. For example, processor 80 may control signal generator 84 to deliver electrical pulses with the amplitudes, pulse widths, frequency, or electrode polarities specified by a therapy program stored in memory 82.

Signal generator 84 is electrically coupled to electrodes via conductors of the respective lead 18, 20, 22, or, in the case of a housing electrode connected to housing 17 of IMD 16, via an electrical conductor disposed within the housing. Signal generator 84 is configured to generate and deliver electrical stimulation therapy to heart 12. For example, signal generator 84 may deliver defibrillation shocks as therapy to heart 12 via at least two electrodes connected to one or more of leads 18, 20, and 22. Signal generator 84 may also deliver pacing pulses via, e.g. ring or helical electrodes coupled to leads 18, 20, and 22. In some examples, signal generator 84 delivers pacing, cardioversion, or defibrillation stimulation in the form of electrical pulses. In other examples, signal generator may deliver one or more of these types of stimulation in the form of other signals, such as sine waves, square waves, or other substantially continuous time signals.

Signal generator 84 may include a switch module, and processor 80 may use the switch module to select, e.g., via a data/address bus, which of the available electrodes are used to

deliver defibrillation pulses or pacing pulses. The switch module may include a switch array, switch matrix, multiplexer, or any other type of switching device suitable to selectively couple stimulation energy to selected electrodes.

Electrical sensing module 86 monitors signals from at least one electrode connected to one of leads 18, 20, or 22 in order to monitor electrical activity of heart 12 of patient 14. Electrical sensing module 86 may also include a switch module to select which of the available electrodes are used to sense the heart activity. In some examples, processor 80 selects the electrodes that function as sense electrodes, or the sensing configuration, via the switch module within electrical sensing module 86, e.g., by providing signals via a data/address bus.

In some examples, electrical sensing module 86 includes multiple detection channels, each of which comprise an amplifier. Each sensing channel detects electrical activity in respective chambers of heart 12, and may be configured to detect either R-waves or P-waves. In some examples, electrical sensing module 86 or processor 80 includes an analog-to-digital converter for digitizing the signal received from a sensing channel for electrogram signal processing by processor 80. In response to the signals from processor 80, the switch module within electrical sensing module 86 couples the outputs from the selected electrodes to one of the detection channels or the analog-to-digital converter.

Processor 80 may maintain escape interval counters using electrical sensing module 86. Signal generator 84 may include pacer output circuits that are coupled, e.g., selectively by a switching module, to any combination of electrodes connected to leads 18, 20, and 22 appropriate for delivery of a bipolar or unipolar pacing pulse to one or more of the chambers of heart 12. Processor 80 may control signal generator 84 to deliver a pacing pulse to a chamber upon expiration of an escape interval. Processor 80 may reset the escape interval counters upon the generation of pacing pulses by stimulation generator 84, or detection of an intrinsic depolarization in a chamber, and thereby control the basic timing of cardiac pacing functions. The escape interval counters may include P-P, V-V, RV-LV, A-V, A-RV, or A-LV interval counters, as examples. The value of the count present in the escape interval counters when reset by sensed R-waves and P-waves may be used by processor 80 to measure the durations of R-R intervals, P-P intervals, P-R intervals and R-P intervals. Processor 80 may use the count in the interval counters to detect a tachyarrhythmia event, such as an atrial or ventricular fibrillation or tachycardia and/or detect a heart rate, such as an atrial rate or ventricular rate.

Processor 80 may also derive other physiological parameters from signals sensed via electrical sensing module 86. For example, processor 80 may establish one or more indicators of ejection fraction and/or heart failure status from electrical signals sensed via electrical sensing module 86. In particular, impedance signals may be used to determine flow or pressure, which may indicate ejection fraction and/or heart failure status.

IMD 16 also includes optical sensor module 87, which is coupled to optical sensors 42 and 44. Optical sensor module 87 may include a variety of analog and/or digital circuitry configured to drive and process signals of optical sensors 42 and 44. Generally speaking, optical sensor module 87 may include an application specific integrated circuit (ASIC), an electronic circuit, a processor (shared, dedicated, or group) and memory that execute one or more software or firmware programs, a combinational logic circuit, or other suitable components that provide the described functionality. In some examples, optical sensor module 87 includes a sensor driver circuit and sensor processor circuit. Conductor elements 88

extending from optical sensors **42, 44**, which may correspond to any of the sensor examples described herein or variations thereof, provide connection to the sensor driver circuit and sensor processor circuit of optical sensor module **87** via any necessary connector elements including, e.g., feedthroughs. The sensor driver circuit provides operational power for optical sensors **42, 44** from power source **90** and controls the timing of sensor operation. Sensor processor circuit receives optical sensor signal output and processes the signal output to estimate a change in a physiological condition, such as blood oxygen saturation, glucose saturation, tissue perfusion or any other physiological condition causing alterations in light modulation by the measurement body fluid or tissue volume.

In other examples, the functions attributed to optical sensor module **87** herein may be executed by other components of IMD **16** including, e.g., processor **80** and memory **82**.

In some examples, IMD **16** may also include one or more sensors in addition to optical sensors **42, 44** and separate from the electrodes connected to leads **18, 20, and 22**. Processor **80** may monitor a variety of physiological parameters of patient **14** that are relevant to delivering efficacious therapy to the patient via signals generated by such additional sensors. Example sensors include an intracardiac or intravascular pressure sensor, as well as an accelerometer or other sensor capable of detecting heart or blood sounds or vibrations, or patient motion, activity, or posture. Optical sensors **42, 44**, as well as any other sensors employed in conjunction with IMD **16**, may also be arranged in a device separate from the IMD, which may, in some examples include processor, memory, and/or telemetry modules for processing and communicating signals generated by the sensors.

Telemetry module **88** includes any suitable hardware, firmware, software or any combination thereof for communicating with another device, such as programmer **24** (FIG. 1). Under the control of processor **80**, telemetry module **88** may receive downlink telemetry from and send uplink telemetry to programmer **24** with the aid of an antenna, which may be internal and/or external. Processor **80** may provide the data to be uplinked to programmer **24** and the control signals for the telemetry circuit within telemetry module **88**, e.g., via an address/data bus. In some examples, telemetry module **88** may provide received data to processor **80** via a multiplexer.

In some examples, processor **80** may transmit atrial and ventricular heart signals (e.g., electrocardiogram signals) produced by atrial and ventricular sense amp circuits within electrical sensing module **86** to programmer **24**. Programmer **24** may interrogate IMD **16** to receive the heart signals. Processor **80** may store heart signals within memory **82**, and retrieve stored heart signals from memory **82**. Processor **80** may also generate and store marker codes indicative of different cardiac events that electrical sensing module **86** detects, and transmit the marker codes to programmer **24**. In some examples, processor **80** may also retrieve data based on optical signals generated by optical sensor module **87** and transmit the data to programmer **24** in order to, e.g., display changes in a physiological condition of patient **14** including, e.g., blood oxygen saturation, glucose saturation, tissue perfusion or any other physiological condition causing alterations in light modulation via body fluid or tissue volume interaction.

The various components of IMD **16** are coupled to power source **90**, which may include a rechargeable or non-rechargeable battery. A non-rechargeable battery may be capable of holding a charge for several years, while a rechargeable battery may be inductively charged from an external device, e.g., on a daily or weekly basis.

FIG. 4 is a functional block diagram illustrating an example configuration of programmer **24**. As shown in FIG. 4, programmer **24** may include a processor **110**, memory **112**, user interface **114**, telemetry module **116**, and power source **118**. Programmer **24** may be a dedicated hardware device with dedicated software for programming of IMD **16**. Alternatively, programmer **24** may be an off-the-shelf computing device running an application that enables programmer **24** to program IMD **16**.

A user may use programmer **24** to select therapy programs (e.g., sets of stimulation parameters), generate new therapy programs, modify therapy programs through individual or global adjustments or transmit the new programs to a medical device, such as IMD **16** (FIG. 1). The clinician may interact with programmer **24** via user interface **114**, which may include a display to present a graphical user interface to a user, and a keypad or another mechanism for receiving input from a user.

Processor **110** can take the form of one or more microprocessors, DSPs, ASICs, FPGAs, programmable logic circuitry, or the like, and the functions attributed to processor **110** herein may be embodied as hardware, firmware, software or any combination thereof. Memory **112** may store instructions that cause processor **110** to provide the functionality ascribed to programmer **24** herein, and information used by processor **110** to provide the functionality ascribed to programmer **24** herein. Memory **112** may include any fixed or removable magnetic, optical, or electrical media, such as RAM, ROM, CD-ROM, hard or floppy magnetic disks, EEPROM, or the like. Memory **112** may also include a removable memory portion that may be used to provide memory updates or increases in memory capacities. A removable memory may also allow patient data to be easily transferred to another computing device, or to be removed before programmer **24** is used to program therapy for another patient.

Programmer **24** may communicate wirelessly with IMD **16**, such as using RF communication or proximal inductive interaction. This wireless communication is possible through the use of telemetry module **116**, which may be coupled to an internal antenna or an external antenna. An external antenna that is coupled to programmer **24** may correspond to the programming head that may be placed over heart **12**, as described above with reference to FIG. 1. Telemetry module **116** may be similar to telemetry module **88** of IMD **16** (FIG. 3).

Telemetry module **116** may also be configured to communicate with another computing device via wireless communication techniques, or direct communication through a wired connection. Examples of local wireless communication techniques that may be employed to facilitate communication between programmer **24** and another computing device include RF communication according to the 802.11 or Bluetooth specification sets, infrared communication, e.g., according to the IrDA standard, or other standard or proprietary telemetry protocols. In this manner, other external devices may be capable of communicating with programmer **24** without needing to establish a secure wireless connection. An additional computing device in communication with programmer **24** may be a networked device such as a server capable of processing information retrieved from IMD **16**.

Power source **118** delivers operating power to the components of programmer **24**. Power source **118** may include a battery and a power generation circuit to produce the operating power. In some examples, the battery may be rechargeable to allow extended operation. Recharging may be accomplished by electrically coupling power source **118** to a cradle or plug that is connected to an alternating current (AC) outlet.

In addition or alternatively, recharging may be accomplished through proximal inductive interaction between an external charger and an inductive charging coil within programmer 24. In other examples, traditional batteries (e.g., nickel cadmium or lithium ion batteries) may be used. In addition, programmer 24 may be directly coupled to an alternating current outlet to power programmer 24. Power source 118 may include circuitry to monitor power remaining within a battery. In this manner, user interface 114 may provide a current battery level indicator or low battery level indicator when the battery needs to be replaced or recharged. In some cases, power source 118 may be capable of estimating the remaining time of operation using the current battery.

FIG. 5 is a block diagram illustrating an example system that includes an external device, such as a server 124, and one or more computing devices 130A-130N, that are coupled to the IMD 16 and programmer 24 shown in FIG. 1 via a network 122. In this example, IMD 16 may use its telemetry module 88 to communicate with programmer 24 via a first wireless connection, and to communication with an access point 120 via a second wireless connection. In the example of FIG. 5, access point 120, programmer 24, server 124, and computing devices 130A-130N are interconnected, and able to communicate with each other, through network 122. In some cases, one or more of access point 120, programmer 24, server 124, and computing devices 130A-130N may be coupled to network 122 through one or more wireless connections. IMD 16, programmer 24, server 124, and computing devices 130A-130N may each comprise one or more processors, such as one or more microprocessors, DSPs, ASICs, FPGAs, programmable logic circuitry, or the like, that may perform various functions and operations, such as those described herein.

Access point 120 may comprise a device that connects to network 122 via any of a variety of connections, such as telephone dial-up, digital subscriber line (DSL), or cable modem connections. In other examples, access point 120 may be coupled to network 122 through different forms of connections, including wired or wireless connections. In some examples, access point 120 may be co-located with patient 14 and may comprise one or more programming units and/or computing devices (e.g., one or more monitoring units) that may perform various functions and operations described herein. For example, access point 120 may include a home-monitoring unit that is co-located with patient 14 and that may monitor the activity of IMD 16.

In some examples, server 124 or computing devices 130 may perform any of the various functions or operations described herein. As shown in FIG. 5, server 124 may include an input/output device 126 and processors 128, similar to programmer 24. A user may interact with server 124 via input/output device 126, similar to programmer 24. In addition, processors 128 may perform any calculations, data processing, communication relay, or any other task required to treat or monitor patient 14.

Network 122 may comprise a local area network, wide area network, or global network, such as the Internet. In some cases, programmer 24 or server 124 may assemble sensing integrity information in web pages or other documents for viewing by and trained professionals, such as clinicians, via viewing terminals associated with computing devices 130A-130N. The system of FIG. 5 may be implemented, in some aspects, with general network technology and functionality similar to that provided by the Medtronic CareLink® Network developed by Medtronic, Inc., of Minneapolis, Minn.

FIGS. 6A and 6B are schematic illustrations of optical sensor 42 including emit module 46 and receive module 48

offset laterally from the emit module. Optical sensor 42 illustrated in FIGS. 6A and 6B may correspond to either of the optical sensors illustrated in FIGS. 2 and 3. FIG. 6A is a plan view of optical sensor 42 including emit module 46 and receive module 48 offset laterally from the emit module. FIG. 6B is a section view of emit module 46 and receive module 48 cut along section line A-A of FIG. 6A.

Emit module 46 includes circuit board 150, side wall 152, light emitting opto-electronic component 154, lens 54, and ferrule 50. Side wall 152 protrudes from circuit board 150 and surrounds light emitting opto-electronic component 154. Lens 54 is offset from circuit board 150. Ferrule 50 is connected to housing 17 of IMD 16, lens 54, and side wall 152.

Side wall 152 surrounds light emitting opto-electronic component 154 to prevent scattering of light and promote transmission of light through lens 54 toward adjacent body fluid or tissue volume 140. Body fluid or tissue volume 140 may correspond to any bodily fluid, such as blood, or body tissue, such as skeletal muscle, neural tissue, myocardium, or the like. Side wall 152 may be formed from a rigid, light-insulating material, including, e.g., a liquid crystal polymer or a metal. In some examples, side wall 152 may be formed from other light-insulating materials including, e.g., any polymer material formed as a molded component, and, in other examples, wall 152 may be formed of a non-rigid material. Side wall 152 is securely coupled to circuit board 150 by, e.g., applying a coating as a hard, die coat dam to hold the wall to the board. In order to increase transmission of light through lens 54 toward adjacent body fluid or tissue volume 140, circuit board 150 and/or side wall 152 may be formed from or coated with a highly reflective material.

Light emitting opto-electronic component 154 may be any electrical circuit component capable of emitting light in response to an applied voltage or current, including, e.g., LEDs and VCSELs. Light emitting opto-electronic component 154 is mounted on circuit board 150 to enable the necessary connections for applying a voltage or current to cause device 154 to emit light. Light emitting component 154 may emit light corresponding to different wavelengths or colors. In one example, in which the sensor including emit module 46 is used for sensing blood oxygen saturation, light emitting opto-electronic component 154 may emit red light (660 nanometer wavelength) or infrared light (890 nanometer wavelength). In another example, in which the sensor including emit module 46 is used for sensing tissue perfusion, light emitting component 154 may emit near-infrared light (800 nanometer wavelength).

In some examples, light emitting opto-electronic component 154 may include multiple light emitting components, e.g. a number of LEDs to emit light of multiple colors and wavelengths. In one such example, one LED of light emitting opto-electronic component 154 may emit red light (660 nanometer wavelength) and another LED of light emitting opto-electronic component 154 may emit infrared light (890 nanometer wavelength). The number of LEDs or other opto-electronic light sources and corresponding light emission wavelengths employed in device 154 may be selected according to the requirements of a particular application and may also depend on the physiological condition being monitored.

Receive module 48 includes circuit board 160, side wall 162, light receiving opto-electronic component 164, lens 56, and ferrule 52. Side wall 162 protrudes from circuit board 160 and surrounds light receiving opto-electronic component 164. Lens 56 is offset from circuit board 160. Ferrule 52 is connected to housing 17 of IMD 16, lens 56, and side wall 162. Light receiving opto-electronic component 164 may be any electrical circuit component capable of generating cur-




FIG. 1

rent or changing a physical property in response to exposure to light, including, e.g., photoresistors or light dependent resistors, photodiodes, phototransistors, photovoltaic cells, or charge-coupled devices.

Side wall **162** surrounds light receiving opto-electronic component **164** to promote transmission of light from adjacent body fluid or tissue volume **140** through lens **56** toward the light receiving component of receive module **48**. Side wall **162** may be formed from a rigid, light-insulating material, including, e.g., a liquid crystal polymer or a metal. In some examples, side wall **162** may be formed from other light-insulating materials including, e.g., any polymer material formed as a molded component, and, in other examples, wall **162** may be formed of a non-rigid material. Side wall **162** is securely coupled to circuit board **160** by, e.g., applying a coating as a hard, die coat dam to hold the wall to the board. In order to increase transmission of light through lens **56** toward light receiving opto-electronic component **164**, circuit board **160** and/or side wall **162** may be formed from or coated with a highly reflective material. Additionally, in some examples, side wall **162** of receive module **48** may share a common portion with side wall **152** of emit module **46**.

Light receiving opto-electronic component **164** may be any electrical circuit component capable of generating current or changing a physical property in response to exposure to light, including, e.g., photoresistors or light dependent resistors, photodiodes, phototransistors, photovoltaic cells, or charge-coupled devices. In some examples, light receive component **164** may include more than one light receiving component. In one example, light receiving opto-electronic component **164** is embodied as an LED formed from a direct band-gap semiconductor that emits narrow spectrum light when electrically biased in the forward direction of the p-n junction. Instead of biasing the LED of light receiving component **164** to emit light, the LED is biased to generate current upon exposure to light, allowing it to function as a light detector. However, in other examples, light receiving opto-electronic component **164** is embodied as another type of light detector including, e.g. a photodiode.

Although example emit and receive modules **46**, **48** of FIGS. 6A and 6B include separate circuit boards **150**, **160**, respectively, in other examples, a single circuit board may be employed that spans both emit and receive modules.

The opto-electronic components of light emitting component **154** and light receiving component **164** may be selected such that one light-detecting component of device **164** is sensitive to the same color of light emitted by one light-emitting component of device **154**. In one example, in which a sensor including emit and receive modules **46**, **48** is used for sensing blood oxygen saturation, as described above, light emitting opto-electronic component **154** is configured to emit and light receiving opto-electronic component **164** is configured to detect red light (660 nanometer wavelength) or light emitting component **154** is configured to emit and light receiving component **164** is configured to detect infrared light (890 nanometer wavelength). In another example, in which a sensor is used for sensing tissue perfusion, light emitting opto-electronic component **154** is configured to emit and light receiving opto-electronic component **164** is configured to detect near-infrared light (800 nanometer wavelength).

In some examples, one or both of emit module **46** and receive module **48** may include index matching material **168** filling some or all of the cavity of each of the modules in which light emitting opto-electronic component **154** and light receiving opto-electronic component **164** are respectively arranged. Index matching material **168** may be configured with a high refractive index, also referred to herein as "index

of refraction", for optically coupling light emitting opto-electronic component **154** and light receiving opto-electronic component **164** with lenses **54** and **56** respectively. Employing index matching material **168** to optically couple light emitting opto-electronic component **154** with lens **54** may act to reduce the reflection of emitted light as light leaves opto-electronic component **154**, as compared to the amount of light reflected when opto-electronic component **154** is interfaced with air. Similarly, employing index matching material **168** to optically couple light receiving opto-electronic component **164** with lens **56** may act to reduce reflections of received light at lens **56** as compared to reflections that would otherwise occur when lens **56** is interfaced with air.

Employing index matching material **168** to reduce light reflections at the interface between the cavity of emit module **46** and lens **54** and the interface between lens **56** and the cavity of receive module **48** may generally increase the optical efficiency of the modules. Example materials and their respective index of refraction values that may be employed as index matching material **168** are disclosed in U.S. application Ser. No. 11/955,056, entitled "IMPLANTABLE OPTICAL SENSOR AND METHOD FOR MANUFACTURE," to Kuhn et al., filed Dec. 12, 2007 and published as U.S. Publication No. 2009/0156912 on Jun. 18, 2009, which is incorporated in its entirety herein by this reference. Index matching material **168** may substantially fill the cavity of each of emit and receive modules **46**, **48** in which light emitting opto-electronic component **154** and light receiving opto-electronic component **164** are respectively arranged, as illustrated in FIG. 6B. In other examples, however, the cavity of one or both of emit and receive modules **46**, **48** may be only partially filled with index matching material **168** or may be filled with multiple materials, at least one of which is an index matching material.

In order to detect changes in a physiological condition of patient **14**, light emitting opto-electronic component **154** emits an optical signal through transparent lens **54** into adjacent body fluid or tissue volume **140**. The optical signal from light emitting opto-electronic component **154** interacts with adjacent body fluid or tissue volume **140** of patient **14**, after which a portion of the optical signal modulated by the body fluid or tissue volume of the patient is reflected off of the fluid or volume through lens **56** of receive module **48**. In some examples, lens **56** may include a thin film or other coating that acts to filter light signals at colors and/or wavelengths of interest. The optical signal received by receive module **48** is detected by light receiving opto-electronic component **164**. The modulation of the optical signal received by light receiving opto-electronic component **164** from light emitting opto-electronic component **154** may be interpreted by optical sensor module **87** and/or processor **80** of IMD **16** to detect a physiological condition of patient **14**, e.g. to detect blood oxygen saturation or tissue perfusion of the patient, in response to which IMD **16** may, e.g., deliver electrical stimulation therapy or modify electrical stimulation therapy already being delivered via one or more of leads **18**, **20**, and **22** to heart **12** of patient **14** to restore normal hemodynamic function.

The examples described in this disclosure are directed to techniques for improving optical sensor signal amplitude for a target sensor sensitivity, i.e. for a target edge-to-edge lens spacing. In the examples disclosed, lens spacing is set to a value to achieve target sensor sensitivity, and other optical sensor parameters are varied in order to increase signal amplitude without increasing power demand. In some examples, the size, D_E and D_R of emit and receive lenses **54**, **56** respectively, and the offset between each of lenses **54**, **56**, and emit

and light emitting and receiving devices **154**, **164**, respectively, e.g. offset distance **O** shown in FIG. **6B**, are selected to increase the amplitude of the signal that is received by light receiving opto-electronic component **164** of receive module **48** from light emitting opto-electronic component **154** of emit module **46**.

With reference to the size of lenses **54**, **56** of the example of FIGS. **6A** and **6B**, optical efficiency generally increases with decreasing lens diameter for constant edge-to-edge lens spacing, holding emit and receive lens diameters, D_E , D_R , respectively, equal, and coating circuit boards **150**, **160** and side walls **152**, **162** with a reflective material. The particular lens diameter selected for emit and receive lenses **54**, **56** may be bounded by upper and lower limits. An upper limit may define, for example, a lens diameter above which optical efficiency is only nominally affected by the size of lenses **54**, **56** in either or both of emit and receive modules **46**, **48**, respectively. In some examples, the diameter, D_E , of emit lens **54**, and the diameter, D_R , of receive lens **56** may be less than approximately 6 millimeters (0.24 inches). In some examples, there may also be a lower limit on the diameter, D_E , of emit lens **54**, or the diameter, D_R , of receive lens **56**. For example, generally speaking, in order to best communicate the optical signal emitted by light emitting opto-electronic component **154** to body fluid or tissue volume **140**, the diameter, D_E , of emit lens **54** should be greater than an area of the light signal emitted by light emitting opto-electronic component **154**. In examples in which light emitting opto-electronic component **154** includes, e.g., an LED, the light signal generated by the LED is essentially a point source that radiates light out in all directions from a point on the component. In examples in which light emitting opto-electronic component **154** includes a VCSEL or other collimated light source, on the other hand, the light signal is generally in the form of a coherent beam. In one example, the diameter, D_E , of emit lens **54** may be greater than in a range from approximately 5 to approximately 10 times the area of the light signal emitted by light emitting opto-electronic component **154**. The light emitting area of light emitting opto-electronic component may be, e.g., the area of the projected beam at the outer surface of lens **54**.

Optical efficiency of emit and receive modules **46**, **48** may be further improved by decreasing emit lens **54** diameter, D_E , while maintaining or increasing receive lens **56** diameter, D_R . Increasing optical efficiency by decreasing emit lens diameter may be caused by an increased concentration of light transmitted by light emitting opto-electronic component **154** through lens **54** to body fluid or tissue volume **140**. However, decreasing receive lens **56** diameter, D_R , may act to decrease the total light received by light receiving opto-electronic component **164**. Therefore, although, in some configurations where emit and receive lens diameters are equal, the impact of decreased reception may be overcome by the benefit of increased transmission, optical efficiency of emit and receive modules **46**, **48** may be further improved by forming the emit module with emit lens **54** diameter, D_E , that is less than receive lens **56** diameter, D_R .

Although lens size has been described with reference to the diameter of disc shaped lenses **54** and **56** of the examples of FIGS. **2**, **3** and **6**, in other examples according to this disclosure the lenses employed in emit and receive modules of an optical sensor may have different shapes and their sizes may therefore be characterized by dimensions other than a diameter. In one example, instead of having the generally circular configuration of lenses **54** and **56**, emit and/or receive modules of an optical sensor may have generally rectilinear

lenses. In such examples, lens size may be characterized by, e.g., the length or the width of the lens.

Another parameter that may affect the optical efficiency of emit and receive modules **46**, **48** is the offset **O** between one of light emitting and receiving devices **154**, **164** and respective one of lenses **54**, **56**. Decreasing the opto-electronic component-lens offsets O_E and O_R of both emit and receive modules **46**, **48**, respectively, may generally act to increase optical efficiency. In some examples, the opto-electronic component-lens offsets O_E and O_R of both emit and receive modules **46**, **48**, respectively, may be selected based on the diameter of emit and receive lenses **54**, **56**, respectively.

As with lens diameter, the particular opto-electronic component-lens offsets O_E and O_R of emit and receive modules **46**, **48**, respectively, may be bounded by upper and lower limits. An upper limit may include, for example, an opto-electronic component-lens offset that is approximately equal to 1 times the diameter of the lens. For example, the distance that light emitting opto-electronic component **154** is offset from lens **54**, i.e. O_E shown in FIG. **6B**, may be approximately equal to the diameter, D_E , of the emit lens. A lower limit on the opto-electronic component-lens offsets O_E and O_R for emit module **46** and receive module **48** may be approximately equal to zero such that the opto-electronic component abuts the module lens. For example, the distance that light emitting opto-electronic component **154** is offset from lens **54**, i.e. O_E shown in FIG. **6B**, may be approximately equal to zero such that the light emitting opto-electronic component abuts the emit lens.

In addition to lens size and opto-electronic component-lens offset, in some examples, the arrangement of light receiving opto-electronic component **164** within receive module **48** and the effective spacing of lenses **54**, **56** are improved by masking receive module lens **56** with flange **170** to create masked lens edge **56a** that is aligned with light receiving opto-electronic component leading edge **164a** and that acts to effectively increase the edge-to-edge lens spacing between emit and receive modules **46**, **48**. In the example of FIGS. **6A** and **6B**, flange **170** protrudes from ferrule **52** of receive module **48**. FIGS. **7A** and **7B** are plan and section views, respectively, which illustrate in greater detail ferrule **52** including flange **170** masking lens **56**. The section view of FIG. **7B** is cut along section line B-B of FIG. **7A**. Although the example of FIGS. **6A-7A** illustrate flange **170** that creates a substantially linear masked lens edge **56a** of lens **56**, in other examples, the opaque member masking the lens of receive module **48** may be curved or otherwise contoured in a non-linear configuration.

As described above, generally speaking, sensor sensitivity may be controlled by varying the spacing between lenses **54**, **56** of emit and receive modules **46**, **48** respectively. Varying lens spacing, and, in particular, increasing lens spacing increases sensor sensitivity. However, as lens spacing is increased, the optical signal reaching light receiving opto-electronic component **164** decreases exponentially. The distribution of light over receive lens **56** also decreases exponentially with distance from the emit lens **54**. Therefore, for a given lens edge-to-edge distance, the optimal position of light receiving opto-electronic component **164** is to be as close to the leading edge in the direction of light travel between emit module **46** and receive module **48** of receive lens **56** as possible.

Lens spacing may generally be defined by edge-to-edge spacing between lenses **54**, **56**, which is illustrated in FIG. **6A** as E_1 . Sensor lenses, including, e.g., lenses **54** and **56** of the example of FIGS. **6A** and **6B**, are commonly produced in a disc shape. Opto-electronic components, on the other hand,

including, e.g., emitting and receiving opto-electronic components **154**, **164** of the example of FIGS. **6A** and **6B**, are commonly produced in rectilinear shapes. As a result, rectilinear light emitting opto-electronic component **154** and light receiving opto-electronic component **164** cannot be arranged with their respective leading edges **154a**, **164a** in the direction of light travel from emit module **46** to receive module **48** aligned with the edges of disc shaped lenses **54** and **56** respectively.

The misalignment of opto-electronic leading edge and lens edge may have a greater performance impact on receive module **48** than emit module **46**. In some examples according to this disclosure, therefore, flange **170** may be employed to mask lens **56** to create masked lens edge **56a** that is aligned with light receiving opto-electronic component leading edge **164a**. In examples employing flange **170** to mask lens **56**, edge-to-edge lens spacing may be defined by E_2 , instead of E_1 . Therefore, masking lens **56** to create masked lens edge **56a** aligned with light receiving opto-electronic component leading edge **164a** also acts to effectively increase the edge-to-edge lens spacing between emit and receive modules **46**, **48** by a distance A .

In some examples, lens **56** may be masked by a different opaque member than flange **170** protruding from ferrule **52** shown in the example of FIGS. **6A-7B**. In one example, lens **56** may be masked by a flange protruding from another component of receive module **48** including, e.g., side wall **162**. In another example, a separate component arranged, e.g. within the cavity of receive module **48** may be employed to mask lens **56**. Additionally, in one example, lens **56** may be masked by an opaque film. As shown in FIGS. **6A** and **6B**, light emitting module **46** and light receiving module **48** may include transparent polymeric seals **156**, **166**, respectively, which may be formed over lenses **54**, **56** and ferrules **50**, **52**, respectively. Seals **156**, **166** may be formed, for example, from silicone rubber, polyurethane, or another optically transparent and biocompatible material. Seals **156**, **166** may function to protect the joints between lenses **54**, **56** and ferrules **50**, **52** from the corrosive effects of bodily fluids and provide a smooth, convex surface that reduces the susceptibility of light emitting and receiving modules **46**, **48** to blood clot formation and excessive tissue encapsulation over lenses **54**, **56**. In some examples, seals employed to cover one or both of lenses **54**, **56** may include rougher surface textures, as well as shapes other than the convex shape of example seals **156**, **166** shown in FIGS. **6A** and **6B**. In examples employing seals **156**, **166**, e.g. the example of FIGS. **6A** and **6B**, an opaque film, which may include a coating or an etched region, may be arranged over seal **166** to mask lens **56** arranged below seal **166**. In another example, however, the opaque film may coat or be etched into the inner or outer surface of lens **56**. Improving lens spacing and opto-electronic component placement may, in some examples, also be accomplished by, e.g., employing a rectilinear lens. However, such lens shapes may be less reliable and more expensive to manufacture and/or assemble.

In one example employing an opaque film to mask lens **56** of receive module **48**, the film may include a coating of a refractory or non-refractory material applied to the inner surface or outer surface of the lens. Refractory materials may be employed, e.g., in examples in which the film that masks lens **56** is arranged on an outer surface of the lens or is applied over seal **166**. Example materials that may be used for an opaque film to mask lens **56** of receive module **48** include niobium, refractory, and gold, non-refractory.

Housing **17** of IMD **16** may be formed, in some examples, from titanium, stainless steel, ceramic, glass, or a rigid polymer. In one example, housing **17** and ferrules **50**, **52** are each

formed from titanium and the ferrules are welded within openings formed in housing **17** to maintain hermeticity of light emitting and receiving modules **46**, **48** and IMD **16**. Lenses **54**, **56** of light emitting and receiving modules **46**, **48** respectively may be formed from, e.g., sapphire and are hermetically sealed in openings formed in housing **17** of IMD **16** using ferrules **50**, **52**, respectively. In some examples, lenses **54**, **56** and ferrules **50**, **52** may be bonded in a number of ways including using, e.g., gold braze joints or a polymer adhesive. The appropriate bond between lenses **54**, **56** and ferrules **50**, **52** may depend on the material from which the ferrules are formed and other manufacturing processes used in fabricating light emitting and receiving modules **46**, **48**.

As illustrated in example emit and receive modules **46**, **48** of FIGS. **6A** and **6B**, inner surfaces **152a**, **162a** of sidewalls **152**, **162** mate with respective outer surfaces **50a**, **52a** of ferrules **50**, **52**. In different examples of emit and receive modules **46**, **48**, inner surfaces **152a**, **162a** of sidewalls **152**, **162** may be joined with respective outer surfaces **50a**, **52a** of ferrules **50**, **52** in a variety of ways. In one example, side walls **152**, **162** may be welded to ferrules **50**, **52**, respectively. In another example, side walls **152**, **162** and respective ferrules **50**, **52** may be joined with braze joints or an adhesive. In one example, side walls **152**, **162** and respective ferrules **50**, **52** may be sized such that inner surfaces **152a**, **162a** of the sidewalls form an interference fit with respective outer surfaces **50a**, **52a** of the ferrules.

In examples in which circuit boards **150**, **160** are merged to include a single board that spans both emit and receive modules **46**, **48**, one side wall, e.g. **162** may be mated to one ferrule **52** with a relatively small dimensional tolerance, while the other side wall, e.g. side wall **152** is mated to the other ferrule **50a** with a larger dimensional tolerance. Such an arrangement in which the dimensional tolerances of the respective joints between side walls **152**, **162** and ferrules **50**, **52** may act to allow tolerance stack-ups. To provide additional flexibility in the fabrication of emit and receive modules **46**, **48**, the joint between side wall and ferrule, e.g. side wall **152** and ferrule **50** may include a slot, e.g. in one of inner sidewall surface **152a** or outer ferrule surface **50a** to allow additional relief when assembly the components of the modules.

One example in which the inner surface of an optical sensor module side wall mates with the outer surface of a ferrule is illustrated in FIG. **8**. FIG. **8** is a partial section view of receive module **48** of FIG. **6B** illustrating in greater detail the junction between housing **17**, side wall **162**, and ferrule **52**. To more clearly illustrate the features of the example of FIG. **8**, ferrule **52** has been exploded away from side wall **162** and housing **17**. In FIG. **8**, inner surface **162a** of side wall **162** includes protrusion **162b** received by concavity **52b** in outer surface **52a** of ferrule **52**. Protrusion **162b** may, in some examples, include a number of protrusions distributed around inner surface **162a** of side wall **162**. In examples in which protrusion **162b** includes a number of protrusions distributed around inner surface **162a** of side wall **162**, concavity **52b** may include a corresponding number of dimples distributed around outer surface **52a** of ferrule **52** such that each protrusion in the inner surface of the side wall is received by a respective one of the dimples in the outer surface of the ferrule. In other examples, however, protrusion **162b** may be formed as a substantially continuous rib around all or most of inner surface **162a** of side wall **162**, in which case, concavity **52b** may be formed as a corresponding continuous channel in outer surface **52a** of ferrule **52** to receive the rib protruding from the inner surface of the side wall.

Although protrusion **162b** and concavity **52b** in side wall **162** and ferrule **52** respectively are illustrated in the example

of FIG. 8 as formed in a generally arcuate shape, in other examples, the side wall protrusion and corresponding ferrule concavity may have other shapes including, e.g., a generally rectilinear shape. Additionally, although inner surface 162a of side wall 162 is shown as including protrusion 162b and outer surface 52a of ferrule 52 is shown as including concavity 52b in the example of FIG. 8, in other examples, the inner surface of the side wall may include a concavity that is configured to receive a protrusion in the outer surface of the ferrule.

Although the foregoing examples have been described with reference to an optical sensor employed in IMD 16, optical sensors according to this disclosure may also be employed in other types of implantable devices. For example, FIG. 9 is a conceptual drawing illustrating example medical lead 200 including body 202, optical sensor 204, conductors 206 and 208, connector assembly 210, and fixation member 212. Lead 200 includes elongated body 202 extending between proximal end 202a and distal end 202b. Optical sensor 204 is positioned along lead body 202, in some examples, near distal end 202b of the lead body. Sensor 204 includes at least two modules, e.g. emit module 214 and receive module 216, which modules respectively include windows 218 and 220 through which emitted light and scattered light travels from/to light emitting opto-electronic component 222 and light receiving opto-electronic component 224, respectively, of sensor 204.

Lead body 202 carries separately insulated conductor pairs 206 and 208 between proximal connector assembly 210 and sensor 204. Conductor pair 206 is provided for carrying drive signals from proximal connector assembly 210 to light emitting opto-electronic component 218 via integrated circuitry in emit module 214. Conductor pair 208 is provided for carrying current generated by light receiving opto-electronic component 224 included in receive module 216 to proximal connector assembly 210. Connector assembly 210 may be coupled to an implantable medical device to thereby couple optical sensor 204 to a sensor module included in the medical device, e.g. optical sensor module 87 of IMD 16 shown in FIG. 3. The sensor module may include sensor driver circuitry and signal processing circuitry (not shown in FIG. 9).

Lead 200 is shown having a distal fixation member 212 for anchoring the position of distal end 202b of lead body 202 at a targeted implant location, e.g. within patient 14. In some examples, fixation member 212 may serve as an electrode and be coupled to an insulated conductor extending to proximal connector assembly 210. Additionally, in some examples lead 200 may include other sensors and/or electrodes including, e.g. electrical stimulation electrodes. As such, it is recognized that the particular configurations of lead body 202, conductors carried by the lead body and proximal connector assembly 210 will depend on the particular configuration of electrodes and sensors carried by lead 200. In some examples, lead 200 may also include an open lumen, for example for use in delivering a fluid agent or passing a guide wire.

FIGS. 10-12 are flow diagrams illustrating several example methods of constructing an optical sensor according to this disclosure connected to a housing of an IMD. The methods of FIGS. 10-12 will be described with reference to the example sensor of FIGS. 6A and 6B connected to housing 17 of IMD. However, the methods may be applied to optical sensors connected to other medical devices including, e.g., lead 200 of FIG. 9.

FIG. 10 is a flow diagram of an example method of constructing an example of optical sensor 42 of FIGS. 6A and 6B. The method of FIG. 10 includes providing emit module 46 and receive module 48 offset laterally from the emit module

(250). In the example of FIG. 10, each of emit and receive modules 46, 48 includes a cavity within which one of light emitting opto-electronic component 154 or light receiving opto-electronic component 164 is arranged. One of emit or receive lenses 54, 56 generally defines one end of the cavity offset from the one of light emitting opto-electronic component 154 or light receiving opto-electronic component 164. The method of FIG. 10 also includes selecting a size of lens 54 of emit module 46 to be less than a size of lens 56 of receive module 48 (252).

FIG. 11 is a flow diagram of another example method of constructing another example of optical sensor 42 of FIGS. 6A and 6B. Light receiving opto-electronic component 164 including first edge 164a is arranged within a cavity of receive module 48 (270). One end of the cavity of receive module 48 offset from light receiving opto-electronic component 164 is enclosed with lens 56 (272) and a portion of lens 56 is covered with an opaque member including, e.g. flange 170 to form masked lens edge 56a substantially aligned with first edge 164a of light receiving opto-electronic component 164 (274).

FIG. 12 is a flow diagram of another example method of constructing another example of optical sensor 42 of FIGS. 6A and 6B. For convenience, the method of FIG. 12 is described with reference to the construction of emit module 46. However, the method may also be applied to the construction of receive module 48. The method of FIG. 12 includes connecting lens 56 to ferrule 50 (280), e.g., by brazing the lens to the ferrule. Ferrule 50 is connected to housing 17 of IMD 16 (282), e.g., by welding the ferrule to the housing. Light emitting opto-electronic component 154 is mounted on a surface of circuit board 150 (284). Wall 152 is connected to circuit board 150 to protrude from the surface of the circuit board and surround light emitting opto-electronic component 154 (286). Finally, wall 152 is connected to ferrule 50 such that inner surface 152a of the wall mates with outer surface 50a of the ferrule (288).

The techniques disclosed herein may provide several advantages for optical sensors employed in implantable medical devices. In particular, the examples described in this disclosure are directed to techniques for improving optical sensor signal amplitude for a target sensor sensitivity without increasing power demand. In some examples, sensor signal amplitude is increased by decreasing lens size and/or an offset between opto-electronic component and lens for one or both of an emit module and receive module of an optical sensor. Additionally, in some examples, the arrangement of the opto-electronic component within the receive module is improved by masking the receive module lens with an opaque member to create a masked lens leading edge that is aligned with a leading edge of the opto-electronic component and that acts to effectively increase the edge-to-edge lens spacing between the emit and receive modules.

Various examples have been described. These and other examples are within the scope of the invention defined by the following claims.

The invention claimed is:

1. A medical device comprising:

a housing; and

an optical sensor module connected to the housing and comprising:

a circuit board arranged within the housing;

an opto-electronic component mounted on a surface of the circuit board;

a wall protruding from the surface of the circuit board and surrounding the opto-electronic component;

a lens offset from the surface of the circuit board;

21

an opaque member partially covering the lens to form a masked lens edge substantially aligned with a first edge of the opto-electronic component; and
 a ferrule connected to the housing, the lens and the wall, wherein an inner surface of the wall mates with an outer surface of the ferrule.

2. The device of claim 1, wherein the wall and the ferrule are sized such that the inner surface of the wall and the outer surface of the ferrule form an interference fit.

3. The device of claim 1, wherein a joint between the inner surface of the wall and the outer surface of the ferrule comprises at least one of a weld, a braze, or an adhesive joint.

4. The device of claim 1, wherein the outer surface of the ferrule comprises one of a protrusion or a concavity and the inner surface of the wall comprises the other of the concavity or the protrusion, the protrusion being configured to be received by the concavity.

5. The device of claim 4, wherein at least one of the protrusion or the concavity comprises at least one of an arcuate or rectilinear shape.

6. The device of claim 1, wherein the opto-electronic component comprises at least one of a light emitting diode, a vertical cavity surface emitting laser, a photoresistor, a light dependent resistor, a photodiode, a phototransistor, a photo-voltaic cell, or a charge-coupled device.

7. The device of claim 1, wherein an inner surface of the wall and a surface of the circuit board to which the opto-electronic component is mounted each comprise a reflective material.

8. The device of claim 1, wherein a space defined by the circuit board, the wall, the lens, and the ferrule is at least partially filled with an index matching material.

9. The device of claim 1 further comprising a transparent seal arranged over an outer surface of the lens.

10. The device of claim 9, wherein the seal comprises a biocompatible polymeric material.

11. The device of claim 9, wherein the seal comprises at least one of silicone or polyurethane.

12. The device of claim 1, wherein the medical device comprises at least one of a pulse generator or a medical lead.

13. The device of claim 1, wherein the optical sensor module comprises at least one of a light emitting module or a light receiving module.

14. The device of claim 1 further comprising an optical sensor comprising the optical sensor module, wherein the optical sensor comprises at least one of a blood oxygen saturation, a blood glucose level, or a tissue perfusion sensor.

15. An optical sensor connected to a housing of a medical device, the sensor comprising:
 a circuit board arranged within the housing;
 an opto-electronic component mounted on a surface of the circuit board;
 a wall protruding from the surface of the circuit board and surrounding the opto-electronic component;
 a lens offset from the surface of the circuit board;
 an opaque member partially covering the lens to form a masked lens edge substantially aligned with a first edge of the opto-electronic component; and

22

a ferrule connected to the housing, the lens and the wall, wherein an inner surface of the wall mates with an outer surface of the ferrule.

16. The sensor of claim 15, wherein the wall and the ferrule are sized such that the inner surface of the wall and the outer surface of the ferrule form an interference fit.

17. The sensor of claim 15, wherein a joint between the inner surface of the wall and the outer surface of the ferrule comprises at least one of a weld, a braze, or an adhesive joint.

18. The sensor of claim 15, wherein the outer surface of the ferrule comprises one of a protrusion or a concavity and the inner surface of the wall comprises the other of the concavity or the protrusion, the protrusion being configured to be received by the concavity.

19. The sensor of claim 18, wherein at least one of the protrusion or the concavity comprises at least one of an arcuate or rectilinear shape.

20. The sensor of claim 15, wherein the opto-electronic component comprises at least one of a light emitting diode, a vertical cavity surface emitting laser, a photoresistor, a light dependent resistor, a photodiode, a phototransistor, a photo-voltaic cell, or a charge-coupled device.

21. The sensor of claim 15, wherein an inner surface of the wall and a surface of the circuit board to which the opto-electronic component is mounted each comprise a reflective material.

22. The sensor of claim 15, wherein a space defined by the circuit board, the wall, the lens, and the ferrule is at least partially filled with an index matching material.

23. The sensor of claim 15 further comprising a transparent seal arranged over an outer surface of the lens.

24. The sensor of claim 23, wherein the seal comprises a biocompatible polymeric material.

25. The sensor of claim 23, wherein the seal comprises at least one of silicone or polyurethane.

26. The sensor of claim 15, wherein the medical device comprises at least one of a pulse generator or a medical lead.

27. The sensor of claim 15, wherein the optical sensor comprises at least one of a blood oxygen saturation, a blood glucose level, or a tissue perfusion sensor.

28. A method of constructing an optical sensor connected to a housing of a medical device, the method comprising:
 connecting a lens to a ferrule;
 connecting the ferrule to the housing;
 mounting an opto-electronic component on a surface of a circuit board;
 connecting a wall to the circuit board to protrude from the surface of the circuit board and surround the opto-electronic component;
 masking a portion of the lens with an opaque member to form a masked lens edge substantially aligned with a first edge of the opto-electronic component; and
 connecting the wall to the ferrule such that an inner surface of the wall mates with an outer surface of the ferrule.

* * * * *




FIG. 2




FIG. 3




FIG. 4




FIG. 5




FIG. 6A




FIG. 6B




FIG. 7A




FIG. 7B




FIG. 8

专利名称(译)	医疗设备用光学传感器		
公开(公告)号	US8463343	公开(公告)日	2013-06-11
申请号	US12/765396	申请日	2010-04-22
[标]申请(专利权)人(译)	美敦力公司		
申请(专利权)人(译)	美敦力公司, INC.		
当前申请(专利权)人(译)	美敦力公司, INC.		
[标]发明人	KUHN JONATHAN L ANDERSON THOMAS A CINBIS CAN JELEN JEFFREY M DAVIS TIMOTHY CARNEY JAMES K		
发明人	KUHN, JONATHAN L. ANDERSON, THOMAS A. CINBIS, CAN JELEN, JEFFREY M. DAVIS, TIMOTHY CARNEY, JAMES K.		
IPC分类号	A61B5/00		
CPC分类号	A61B5/14503 A61B5/14532 A61B5/1455 A61N1/36557 A61N1/3702 A61B5/6846 Y10T29/49002		
代理人(译)	宝华, STEPHEN W.		
优先权	61/299705 2010-01-29 US		
其他公开文献	US20110190609A1		
外部链接	Espacenet USPTO		

摘要(译)

用于医疗设备的光学传感器包括发射和接收模块之间的固定透镜间隔，以实现目标传感器灵敏度，同时改变其他传感器参数以便在不增加功率需求的情况下增加信号幅度。连接到医疗装置的壳体的光学传感器包括电路板，光电元件，壁，透镜和套圈。电路板布置在壳体内部。光电元件安装在电路板的表面上。壁从电路板的表面突出并围绕光电元件。透镜偏离电路板的表面。套圈连接到壳体，透镜和壁。壁的内表面与套圈的外表面配合。

