



(19) **United States**

(12) **Patent Application Publication** (10) **Pub. No.: US 2003/0114762 A1**

Balas et al.

(43) **Pub. Date: Jun. 19, 2003**

(54) **METHOD AND SYSTEM FOR CHARACTERIZATION AND MAPPING OF TISSUE LESIONS**

(52) **U.S. Cl. 600/476**

(76) **Inventors: Konstantinos Balas, Heraklion (GR); Demetrios Pelecoudas, Patras (GR)**

(57) **ABSTRACT**

Correspondence Address:
THE FIRM OF KARL F ROSS
5676 RIVERDALE AVENUE
PO BOX 900
RIVERDALE (BRONX), NY 10471-0900 (US)

The present invention provides a method and an apparatus for the in vivo, non-invasive, early detection of alterations and mapping of the grade of these alterations, caused in the biochemical and/or in the functional characteristics of epithelial tissues during the development of tissue atypias, dysplasias, neoplasias and cancers. The method is based, at least in part, on the simultaneous measurement of the spatial, temporal and spectral alterations in the characteristics of the light that is re-emitted from the tissue under examination, as a result of a combined tissue excitation with light and special chemical agents. The topical or systematic administration of these agents result in an evanescent contrast enhancement between normal and abnormal areas of tissue. The apparatus enables the capturing of temporally successive imaging in one or more spectral bands simultaneously. Based on the measured data, the characteristic curves that express the agent-tissue interaction kinetics, as well as numerical parameters derived from these data, are determined in any spatial point of the examined area. Mapping and characterization of the lesion, are based on these parameters.

(21) **Appl. No.: 10/240,367**

(22) **PCT Filed: Mar. 28, 2001**


(86) **PCT No.: PCT/GR01/00017**

(30) **Foreign Application Priority Data**


Mar. 28, 2000 (GR) 20000-100102
Dec. 15, 2000 (US) 09/739089


Publication Classification

(51) **Int. Cl.⁷ A61B 6/00**



KI





KI




Figure 1




Figure 2




Figure 3




Figure 4




Figure 5




Figure 6




Figure 7

METHOD AND SYSTEM FOR CHARACTERIZATION AND MAPPING OF TISSUE LESIONS

RELATED APPLICATIONS

[0001] This application claims priority to Greek National Application Ser. No. 20000100102, filed on Mar. 28, 2000 and to United States National application Ser. No. 09/739089 filed on Dec. 15, 2000.

FIELD OF THE INVENTION

[0002] The present invention is directed to a method and apparatus for the in vivo, non-invasive detection and mapping of the biochemical and/or functional pathologic alterations of human tissues.

BACKGROUND OF THE INVENTION

[0003] Cancer precursor signs are the so-called pre-cancerous states, which are often curable if they are detected at an early stage. If left untreated, the pre-cancerous state can develop into invasive cancer, which can subsequently metastasize. At this stage, the possibilities of successful therapy are dramatically diminished. Consequently, the early detection and the objective identification of the severity of the pre-cancerous state are of crucial importance.

[0004] Conventional methods that utilize optical instruments are very limited in their ability to detect cancerous and pre-cancerous tissue lesions. This is due to the fact that the structural and metabolic changes, which take place during the development of the disease, do not significantly and specifically alter the spectral characteristics of the pathological tissue.

[0005] In order to obtain a more accurate diagnosis, biopsy samples are obtained from suspicious areas, which are submitted for histological examination. However, biopsies pose several problems, such as a) a risk for sampling errors associated with the visual limitations in detecting and localizing suspicious areas; b) a biopsy can alter the natural history of the intraepithelial lesion; c) mapping and monitoring of the lesion require multiple tissue sampling, which is subjected to several risks and limitations; and d) the diagnostic procedure performed with biopsy sampling and histologic evaluation is qualitative, subjective, time consuming, costly and labor intensive.

[0006] In recent years, a few methods and systems have been developed to overcome the disadvantages of the conventional diagnostic procedures. These methods can be classified into two categories: a) methods which are based on the spectral analysis of tissues in vivo, in an attempt to improve the diagnostic information, and b) methods which are based on the chemical excitation of tissues with the aid of special agents, which can interact with pathologic tissue and alter its optical characteristics selectively, thus enhancing the contrast between lesion and healthy tissue.

[0007] In the first case, the experimental use of spectroscopic techniques has been motivated by the ability of these techniques to detect alterations in the biochemical and/or the structural characteristics of tissue as the disease progresses. In particular, fluorescence spectroscopy has been extensively used in various tissues. With the aid of a light source (usually laser) of short wave length (blue-ultraviolet range),

the tissue is first excited. Next, the intensity of the fluorescent light emitted by the tissue as a function of the wavelength of the light is measured.

[0008] Garfield and Glassman in U.S. Pat. Nos. 5,450,857 and Ramanajum et al. in 5,421,339 have presented a method based on the use of fluorescence spectroscopy for the diagnosis of cancerous and pre-cancerous lesions of the cervix. The main disadvantage of fluorescence spectroscopy is that the existing biochemical modifications associated with the progress of the disease are not manifested in a direct way as modifications in the measured fluorescence spectra. The fluorescence spectra contain limited diagnostic information for two basic reasons: a) Tissues contain non-fluorescent chromophores, such as hemoglobin. Absorption by such chromophores of the emitted light from fluorophores can result in artificial dips and peaks in the fluorescence spectra. In other words the spectra carry convoluted information for several components and therefore it is difficult to assess alterations in tissue features of diagnostic importance; and b) The spectra are broad because a large number of tissue components are optically excited and contribute to the measured optical signal. As a result, the spectra do not carry specific information of the pathologic alterations and thus they are of limited diagnostic value. In short, the aforementioned fluorescent technique suffers from low sensitivity and specificity in the detection and classification of tissue lesions.

[0009] Aiming to enhance the sensitivity and specificity of the preceding method, Ramanujan et al. in the Patent No. WO 98/24369 have presented a method based on the use of neural networks for the analysis of the spectral data. This method is based on the training of a computing system with a large number of spectral patterns, which have been taken from normal and from pathologic tissues. The spectrum that is measured each time is compared with the stored spectral data, facilitating in this way the identification of the tissue pathology.

[0010] R. R. Kortun et al, in U.S. Pat. No. 5,697,373, seeking to improve the quality of the measured diagnostic information, have presented a method based on the combination of fluorescence spectroscopy and Raman scattering. The latter has the ability of providing more analytical information; however, Raman spectroscopy requires complex instrumentation and ideal experimental conditions, which substantially hinders the clinical use thereof.

[0011] It is generally known that tissues are characterized by the lack of spatial homogeneity. Consequently the spectral analysis of distributed spatial points is insufficient for the characterization of their status.

[0012] Dombrowski in U.S. Pat. No. 5,424,543, describes a multi-wavelength, imaging system, capable of capturing tissue images in several spectral bands. With the aid of such a system it is possible in general to map characteristics of diagnostic importance based on their particular spectral characteristics. However, due to the insignificance of the spectral differences between normal and pathologic tissue, which is in general the case, inspection in narrow spectral bands does not allow the highlighting of these characteristics and even more so, the identification and staging of the pathologic area.

[0013] D. R. Sandison et al., in U.S. Pat. No. 5,920,399, describe an imaging system, developed for the in vivo

investigation of cells, which combines multi-band imaging and light excitation of the tissue. The system also employs a dual fiber optic bundle for transmitting light from the source to the tissue, and then from the tissue to an optical detector. These bundles are placed in contact with the tissue, and various wavelengths of excitation and imaging are combined in attempt to enhance the spectral differentiation between normal and pathologic tissue.

[0014] In U.S. Pat. No. 5,921,926, J. R. Delfyett et al. have presented a method for the diagnosis of diseases of the cervix, which is based on the combination of Spectral Interferometry and Optical Coherence Tomography (OCT). This system combines three-dimensional imaging and spectral analysis of the tissue.

[0015] Moreover, several improved versions of colposcopes have been presented, (D. R. Craine et al., U.S. Pat. Nos. 5,791,346 and K. L. Blaiz 5,989,184) in most of which, electronic imaging systems have been integrated for image capturing, analysis of tissue images, including the quantitative assessment of lesion's size. For the enhancement of the optical differentiation between normal and pathologic tissue, special agents are used in various fields of biomedical diagnostics, which are administered topically or systematically. Such agents include acetic acid solution, toluidine blue, and various photosensitizers (porphyrines) (S. Anderson Engels, C. Klinteberg, K. Svanberg, S. Svanberg, *In vivo* fluorescence imaging for tissue diagnostics, *Phys Med. Biol.* 42 (1997) 815-24). The selective staining of the pathologic tissue arises from the property of these agents to interact with the altered metabolic and structural characteristics of the pathologic area. This interaction enhances progressively and reversibly the differences in the spectral characteristics of reflection and/or fluorescence between normal and pathologic tissue. Despite the fact that the selective staining of the pathologic tissue is a dynamic phenomenon, in clinical practice the intensity and the extent of the staining are assessed qualitatively and statically. Furthermore, in several cases of early pathologic conditions, the phenomenon of temporary staining after administering the agent, is short-lasting and thus the examiner is not able to detect the alterations and even more so, to assess their intensity and extent. In other cases, the staining of the tissue progresses very slowly, resulting in patient discomfort and the creation of problems for the examiner in assessing the intensity and extent of the alterations, since they are continuously changing. The above have as direct consequence the downgrading of the diagnostic value of these diagnostic procedures. Thus, their usefulness is limited to facilitating the localization of suspected areas for obtaining biopsy samples.

[0016] Summarizing the above, the following conclusions are drawn:

[0017] a) Various conventional light dispersion spectroscopic techniques (fluorescence, elastic, non-elastic scattering, etc.) have been proposed and experimentally used for the *in vivo* detection of alterations in the structural characteristics of pathologic tissue. The main disadvantage of these techniques is that they provide point information, which is inadequate for the analysis of the spatially non-homogenous tissue. Multi-band imaging has the potential to solve this problem by providing spectral information, of lesser resolution as a rule, in any spatial point of the area under examination.

These imaging and non-imaging techniques, however, provide information of limited diagnostic value because the structural tissue alterations, which accompany the development of the disease, are not manifested as significant and characteristic alterations in the measured spectra. Consequently, the captured spectral information cannot be directly correlated with the tissue pathology, a fact that limits the clinical usefulness of these techniques.

[0018] b) The conventional (non-spectral) imaging techniques provide the capability of mapping characteristics of diagnostic importance in two or three dimensions. They are basically used for measuring morphological characteristics and as clinical documentation tools.

[0019] c) The diagnostic methods that are based on the selective staining of pathologic tissue with special agents allow the enhancement of the optical contrast between normal and pathologic tissue. Nevertheless they provide limited information for the *in vivo* identification and staging of the disease.

[0020] The selective interaction of pathologic tissue with the agents, which enhance the optical contrast with healthy tissue, is a dynamic phenomenon. It is therefore reasonable to suggest that the measurement and analysis of kinetic properties could provide important information for the *in vivo* detection, identification and staging of tissue lesions. In a previous publication, in which one of the inventors is a co-author, (C. Balas, A. Dimoka, E. Orfanoudaki, E. Koumandakis, "In vivo assessment of acetic acid-cervical tissue interaction using quantitative imaging of back-scattered light: Its potential use for the *in vivo* cervical cancer detection grading and mapping", *SPIE-Optical Biopsies and Microscopic Techniques*, Vol. 3568 pp. 31-37, (1998)), measurements of the alterations in the characteristics of the back-scattered light as a function of wave-length and time are presented. These alterations occur in the cervix by the topical administration of acetic acid solution. In this particular case, a general-purpose multi-spectral imaging system built around a tunable liquid crystal monochromator was used for measuring the variations in intensity of the back-scattered light as a function of time and wavelength at selected spatial points. It was found that the lineshapes of curves of intensity of back-scattered light versus time provide advanced information for the direct identification and staging of tissue neoplasias. Unpublished results of the same research team indicate that similar results can also be obtained with other agents, which have the property of enhancing the optical contrast between normal and pathologic tissue. Nevertheless, the experimental method employed in the published paper is characterized by quite a few disadvantages, such as: The imaging monochromator requires time for changing the imaging wavelength and as a consequence it is inappropriate for multispectral imaging and analysis of dynamic phenomena. It does not constitute a method for the mapping of the grade of the tissue lesions, as the presented curves illustrate the temporal alterations of intensity of the back-scattered light in selected points. The lack of data modeling and parametric analysis of kinetics data in any spatial point of the area of interest restricts the usefulness of the method in experimental studies and hinders its clinical implementation. The optics used for the imaging of the area of interest is of general purpose and does not

comply with the special technical requirements for the clinical implementation of the method. Clinical implementation of the presented system is also hindered by the fact that it does not integrate appropriate means for ensuring the stability of the relative position between the tissue surface and image capturing module during the snapshot imaging procedure. This is very important since small movements of the patient (i.e. breathing) are always present during the examination procedure. If, after the application of the agent, micro-movements occur while an image is being recorded, then the spatial features of the captured images may not be accurate. This may substantially reduce the accuracy of the calculation of the curves in any spatial point that express the kinetics of marker-tissue interaction.

SUMMARY OF THE INVENTION

[0021] The present invention provides a method for monitoring the effects of a pathology-differentiating agent on a tissue sample. The method includes applying a pathology differentiating agent, e.g., acetic acid, on a tissue sample and measuring a spectral property, such as an emission spectrum, of the tissue sample over time, thereby monitoring the effects of a pathology differentiating agent on a tissue sample. The tissue may be a sample from: the cervix of the uterus, the vagina, the skin, the uterus, the gastrointestinal track or the respiratory track. Without intending to be limited by theory, it is believed that the pathology-differentiating agent induces transient alterations in the light scattering properties of the tissue, e.g., the abnormal epithelium.

[0022] In another aspect, the present invention features a method for the *in vivo* diagnosis of a tissue abnormality, e.g., a tissue atypia, a tissue dysplasia, a tissue neoplasia (such as a cervical intraepithelial neoplasia, CINI, CINII, CINIII) condylomas or cancer, in a subject. The method includes applying a pathology differentiating agent, e.g., an acetic acid solution or a combination of solutions selected from a plurality of acidic and basic solutions, to a tissue. The method further includes exposing the tissue in the subject to optical radiation, and monitoring the intensity of light emitted from the tissue over time, thereby diagnosing a tissue abnormality in a subject. The optical radiation may be broad band optical radiation, preferably polarized optical radiation.

[0023] The non-invasive methods of the present invention are useful for *in vivo* early detection of tissue abnormalities/alterations. The methods are also useful for mapping the grade of abnormalities/alterations in epithelial tissues during the development of tissue atypias, dysplasias, neoplasias and cancers.

[0024] In one embodiment, the tissue area of interest is illuminated with a broad band optical radiation and contacted with a pathology differentiating agent, e.g., an agent or a combination of agents which interact with pathologic tissue areas characterized by an altered biochemical composition and/or cellular functionality and provoke a transient alteration in the characteristics of the light that is re-emitted from the tissue. The light that is re-emitted from the tissue may be in the form of reflection, diffuse scattering, fluorescence or combinations or subcombinations thereof. The intensity of the light emitted from the tissue may be measured, e.g., simultaneously, in every spatial point of the tissue area of interest, at a given time point or over time (e.g.,

for the duration of agent-tissue interaction). A diagnosis may be made based on the quantitative assessment of the spatial distribution of alterations in the characteristics of the light re-emitted from the tissue at given time points before and after the optical and chemical excitation of the tissue. The diagnosis may also be made based on the spatial distribution of parameters calculated from kinetics curves obtained from the light re-emitted from the tissue. These curves are simultaneously measured in every spatial point of the area under examination during the optical and chemical excitation of the tissue.

[0025] In one embodiment of the invention, the step of tissue illumination comprises exposing the tissue area under analysis to optical radiation of narrower spectral width than the spectral width of the light emitted by the illumination source. In another embodiment, the step of measuring the intensity of light comprises measuring the intensity of the re-emitted light in a spectral band, the spectral width of which is narrower than the spectral width of the detector's sensitivity. In yet another embodiment, the step of measuring the intensity of light comprises measuring simultaneously the intensity of the re-emitted light in a plurality of spectral bands, the spectral widths of which are narrower than the spectral width of the detector's sensitivity.

[0026] In yet another aspect, the present invention features an apparatus for the *in vivo*, non-invasive early detection of tissue abnormalities/alterations and mapping of the grade of these tissue abnormalities/alterations caused in the biochemical and/or in the functional characteristics of epithelial tissues, during the development of tissue atypias, dysplasias, neoplasias and cancers. The apparatus includes optics for collecting the light re-emitted by the area under analysis, selecting magnification and focusing the image of the area. The apparatus may also include optical imaging detector(s), means for the modulation, transfer, display and capturing of the image of the tissue area of interest. In addition, the apparatus can include a computer, which has data storage, processing and analysis means, a monitor for displaying images, curves and numerical data, optics for the optical multiplication of the image of the tissue area of interest, and a light source for illuminating the area of interest. The apparatus may also include optical filters for selecting the spectral band of imaging and illumination, means for transmitting light and illuminating the area of interest, control electronics, and optionally, software for the analysis and processing of data. The software can help with the tissue image capturing and storing in specific time points and for a plurality of time points, before and after administration of the pathology-differentiating agent.

[0027] Using the foregoing apparatus, an image or a series of images may be created which express the spatial distribution of the characteristics of the kinetics of the induced alterations in the tissue's optical characteristics, before and after the administration of the agent. Pixel values of the image correspond to the spatial distribution of the alterations in the intensity of the light emitted from the tissue at given times, before and after the optical and chemical excitation of tissue. The spatial distribution of parameters may be associated with pixel gray values as a function of time. The foregoing function may be calculated from the measured and stored images and for each row of pixels with the same spatial coordinates.

[0028] In one embodiment, the step of optical filtering the imaging detector comprises an optical filter that is placed in the optical path of the rays that form the image of the tissue, for the recording of temporally successive images in a selected spectral band, the spectral width of which is narrower than the spectral width of the detector's sensitivity.

[0029] In yet another embodiment, the image multiplication optics includes light beam splitting optics that creates two identical images of the area of interest. The images are recorded by two imaging detectors, in front of which optical filters are placed. The filters are capable of transmitting light having a spectral width that is shorter than the spectral width of the detector's sensitivity, so that two groups of temporally successive images of the same tissue area are recorded simultaneously, each one corresponding to a different spectral band.

[0030] In another embodiment, the image multiplication optics include more than one beam splitter for the creation of multiple identical images of the area of interest. The images are recorded by multiple imaging detectors, in front of which optical filters are placed. The filters have different transmission characteristics and are capable of transmitting light of spectral width shorter than the spectral width of the detector's sensitivity. Thus, multiple groups of temporally successive images of the same tissue area are recorded simultaneously, each one corresponding to a different spectral band.

[0031] In a further embodiment, the image multiplication optics comprise one beam splitter for the creation of multiple identical images of the area of interest, which are recorded by multiple imaging detectors, in front of which optical filters are placed with, preferably, different transmission characteristics and capable of transmitting light of spectral width shorter than the spectral width of the detector's sensitivity, so that multiple groups of temporally successive images of the same tissue area are recorded simultaneously, each one corresponding to a different spectral band.

[0032] In yet a further embodiment, the image multiplication optics include one beam splitter for the creation of multiple identical images of the area of interest, which are recorded in different sub-areas of the same detector. Optical filters having different transmission characteristics are placed in the path of the split beams. The filters are capable of transmitting light of spectral width shorter than the spectral width of the detector's sensitivity. Multiple groups of temporally successive images of the same tissue area are recorded simultaneously in the different areas of the detector, each one corresponding to a different spectral band.

[0033] In another embodiment, the step of filtering the light source comprises an optical filter, which is placed in the optical path of an illumination light beam, and transmits light of spectral width shorter than the spectral width of sensitivity of the detector used.

[0034] In a further embodiment, the step of filtering the light source includes providing a plurality of optical filters and a mechanism for selecting the filter that is disposed in the path of the illumination light, thus enabling the tuning of the center wavelength and the spectral width of the light illuminating the tissue.

[0035] In another embodiment, the mapping of the grade of the alterations associated with the biochemical and/or

functional characteristics of the tissue area of interest is based on the pixel values of one image from the group of the recorded temporally successive images of the tissue area of interest.

[0036] In a further embodiment, this mapping is based on the pixel values belonging to a plurality of images, which are members of the group of the recorded temporally successive images of the tissue area of interest.

[0037] In another embodiment, this mapping is based on numerical data derived from the pixel values belonging to a plurality of images, which are members of the group of the recorded temporally successive images of the tissue area of interest.

[0038] In a further embodiment, a pseudo-color scale, which represents with different colors the different pixel values of the image or of the images used for the mapping of abnormal tissue areas, is used for the visualization of the mapping.

[0039] In one embodiment, the image or images are used for the in vivo detection, and identification of the borders of epithelial lesions.

[0040] In another embodiment, the pixel values of the image or of the images, which are determined for the mapping of the grade of alterations in biochemical and/or functional characteristics of tissue, are used as diagnostic indices for the in vivo identification and staging of epithelial lesions.

[0041] In yet another embodiment, the image or the images can be superimposed on the color or black and white image of the same area of tissue under examination displayed on the monitor. Abnormal tissue areas are highlighted and their borders are demarcated, facilitating the selection of a representative area for taking a biopsy sample, the selective surgical removal of the abnormal area and the evaluation of the accuracy in selecting and removing the appropriate section of the tissue.

[0042] In a further embodiment, the image or the images which are determined for the mapping of the grade of alterations in biochemical and/or functional characteristics of tissue are used for the evaluation of the effectiveness of various therapeutic modalities such as radiotherapy, nuclear medicine treatments, pharmacological therapy, and chemotherapy.

[0043] In another embodiment, the optics for collecting the light re-emitted by the area under analysis includes optomechanical components employed in microscopes used in clinical diagnostic examinations, surgical microscopes, colposcopes and endoscopes.

[0044] In one embodiment of the invention directed to colposcopy applications, the apparatus may comprise a speculum, an articulated arm onto which the optical head is attached. The optical head includes a refractive objective lens, focusing optics, a mechanism for selecting the magnification, an eyepiece, a mount for attaching a camera, and an illuminator. The speculum is attached so that the central longitudinal axis of the speculum is perpendicular to the central area of the objective lens. Thus, when the speculum is inserted into the vagina and fixed in it, the relative position of the image-capturing optics and of the tissue area of

interest remain unaltered, regardless of micro-movements of the cervix, which are taking place during the examination of the female subject.

[0045] In a further embodiment, the apparatus may further comprise an atomizer for delivering the agent. The atomizer is attached to the articulated arm-optical head of the apparatus and in front of the vaginal opening, where the spraying of the tissue may be controlled and synchronized with a temporally successive image capturing procedure with the aid of electronic control means.

[0046] In another embodiment of the apparatus of the invention, the image capturing detector means and image display means include a camera system. The camera system has a detector with a spatial resolution greater than 1000×1000 pixels and a monitor of

BRIEF DESCRIPTION OF THE DRAWINGS

[0047] FIG. 1 is a schematic representation of the present method's basic principle. at least 17 inches/43.18 cm (diagonal), so that high magnification is ensured together with a large field of view while the image quality is maintained.

[0048] In a further embodiment directed to microscopes used in clinical diagnostic examinations surgical microscopes and colposcopes, a system includes an articulated arm onto which the optical head is attached. The optical head includes an objective lens, focusing optics, a mechanism for selecting the magnification, an eyepiece, a mount for attaching a camera, an illuminator and two linear polarizers. One linear polarizer is disposed in the optical path of the illuminating light beam and the other in the optical path of the rays, that form the image of the tissue. The polarization planes of these polarizers may be rotated. When the planes are perpendicular to each other, the contribution of the tissue's surface reflection to the formed image is eliminated.

[0049] In another embodiment directed to endoscopy, an endoscope may include optical means for transferring light from the light source to the tissue surface. The optical means may also allow the collection and transferring of rays along substantially the same axis. The optical means also allow the focusing of the rays that form the image of the tissue. The endoscope may also include two linear polarizers. One linear polarizer is disposed in the optical path of the illuminating light beam and the other in the optical path of the rays that form the image of the tissue. The polarization planes of these polarizers may be rotated. When the planes are perpendicular to each other, the contribution of the tissue's surface reflection to the formed image is eliminated.

[0050] In another embodiment, microscopes used in clinical diagnostic examinations, surgical microscopes and colposcopes may include a reflective objective lens that replaces a retroactive lens. The reflective objective lens is contracted so that a second reflection mirror is disposed in the central part of its optical front aperture. In the rear, non-reflective part of this mirror, illumination means are attached from which light is emitted toward the object. With or without illumination zooming and focusing optics, the central ray or the emitted light cone is coaxial with the central ray of the light-beam that enters the imaging lens. With the aid of illumination zooming and focusing optics, which may be adjusted simultaneously and automatically

with the mechanism for varying the magnification the optical imaging system, the illuminated area and the field-of-view of the imaging system can vary simultaneously and proportionally. Any decrease in image brightness caused by increasing the magnification is compensated with the simultaneously zooming and focusing of the illumination beam.

[0051] Other features and advantages of the invention will be apparent from the following detailed description and claims.

BRIEF DESCRIPTION OF THE DRAWINGS

[0052] FIG. 1 is a schematic representation of the present method's basic principle.

[0053] In FIG. 1, the tissue (T), is sprayed using an atomizer (A), which contains the agent, e.g., acetic acid. At the same time, the tissue is illuminated with a source that emits light having a frequency within a specific spectral band that depends on the optical characteristics of both the agent and the tissue. The characteristics of the light emitted from the source can be controlled by choosing particular sources (LS), and optical filters (OFS). Sources of light for illuminating the tissue include light emitting diodes, and lasers.

[0054] For imaging the area of interest, light collection optics (L) may be used, which focus the image onto a two-dimensional optical detector (D). The output signal of the latter is amplified, modulated and digitized with the aid of appropriate electronics (EIS) and finally the image is displayed on a monitor (M) and stored in the data-storing means of a personal computer (PC). Between tissue (T) and detector (D), optical filters (OFI) can be interposed. The filter can be interposed for tissue (T) imaging in selected spectral bands, at which the maximum contrast is obtained between areas that are subjected to different grade of alterations in their optical characteristics after administering the appropriate agent.

[0055] Before administration of the latter, images can be obtained and used as references. After the agent has been administered, the detector (D) helps to capture images of the tissue, in successive time instances, which are then stored in the computer's data-storage means. The measuring rate is proportional to the rate at which the tissue's optical characteristics are altered, following the administration of the agent.

[0056] As used herein, an optical property, P, is a property that arises from the interaction of electromagnetic waves and a material sample, e.g., a tissue, such as a tissue within a subject. For example, the property can be the intensity of light after it interacts with matter, as manifested by an absorption, emission, or Raman spectrum. A dynamic optical property is a property that is obtained from a time-dependent optical property, $P(t)$, and is determined from the measurement of $P(t)$ at more than one time. For example, a dynamical optical property can be a relaxation time, or a time integral of $P(t)$.

[0057] In FIG. 1, images of the same tissue area are schematically illustrated, which have been stored successively before and after administering the agent (STI). In these images, the black areas represent tissue areas that do not alter their optical characteristics (NAT), while the gray-white tones represent areas that alter their optical charac-

teristics (AT), following the administration of the agent. The simultaneous capture of the intensity of the light re-emitted from every spatial point of the tissue area under analysis and in predetermined time instances, allows the calculation of the kinetics of the induced alterations.

[0058] In FIG. 1, two curves are illustrated: pixel value at position xy (Pv_{xy}), versus time t . The curve ATC corresponds to an area where agent administration induced alterations (AT) in the tissue's optical characteristics. The curve (NATC) corresponds to an area where no alteration took place (NAT).

[0059] Each pixel, (x,y) , can be associated with a pixel value, such as intensity I , which generally depends on time. For example, at time t_i and pixel (x,y) , the pixel value can be denoted by $PV_{xy}(t_i)$. One useful dynamical spectral property, which can be obtained by measuring pixel value versus time at a particular pixel (x,y) , is the relaxation time $t_{rel}(x,y)$. Letting the maximum of a Pv_{xy} versus time curve be denoted by A , then $t_{rel}(x,y)$ satisfies $PV_{xy}(t_{rel})=A/e$, where e is the base of the natural logarithm. For example, if the pixel value versus time curve can be approximated by an exponential with relaxation rate r , $PV_{xy}(t)=A \exp(-rt)$, where $r>0$, then $t_{rel}(x,y)=1/r$.

[0060] The calculation of these parameters (P) at every spatial point of the area under analysis allows kinetic information (KI) to be obtained, with pixel values that are correlated with these parameters. These values can be represented with a scale of pseudocolors (P_{min} , P_{max}), the spatial distribution of which allows for immediate optical evaluation of the intensity and extent of the induced alterations. Depending on the correlation degree between the intensity and the extent of the induced alterations with the pathology and the stage of the tissue lesion, the measured quantitative data and the derived parameters allow the mapping, the characterization and the border-lining of the lesion. The pseudocolor image of the phenomenon's kinetics (KI), which expresses the spatial distribution of one or more parameters, can be superimposed (after being calculated) on the tissue image, which is displayed in real-time on the monitor. Using the superimposed image as a guide facilitates the identification of the lesion's boundaries, for successful surgical removal of the entire lesion, or for locating suspicious areas to obtain a biopsy sample(s). Furthermore, based on the correlation of the phenomenon's kinetics with the pathology of the tissue, the measured quantitative data and the parameters that derive from them can provide quantitative clinical indices for the in vivo staging of the lesion or of sub-areas of the latter.

[0061] In some cases it is necessary to capture the kinetics of the phenomenon in more than one spectral band. This can help in the in vivo determination of illumination and/or imaging spectral bands at which the maximum diagnostic signal is obtained. Furthermore, the simultaneous imaging in more than one spectral band can assist in minimizing the contribution of the unwanted endogenous scattering, fluorescence and reflection of the tissue, to the optical signal measured by the detector. The measured optical signal comprises the optical signal generated by the marker-tissue interaction and the light emitted from the endogenous components of the tissue. In many cases, the recorded response of the components of the tissue constitutes noise since it occludes the generated optical signal, which carries the

diagnostic information. Therefore, separation of these signals, based on their particular spectral characteristics, results in the maximization of the signal-to-noise ratio and consequently in the improvement of the obtained diagnostic information.

[0062] FIG. 2 illustrates a method for measuring in two spectral bands simultaneously and in any spatial point of the area under analysis, the kinetics of the alterations in the characteristics of the light emitted from the tissue, before and the after the administration of the contrast enhancing agent. The light emitted from the tissue is collected and focused by the optical imaging module (L) and allowed to pass through a beam splitting (BSP) optical element. Thus, two identical images of the tissue (T) are generated, which can be captured by two detectors (D1, D2). In front of the detector, appropriate optical filters (Of_{λ_1}), (Of_{λ_2}) can be placed, so that images with different spectral characteristics are captured. Besides beam splitters, optical filters, dichroic mirrors, etc., can also be used for splitting the image of the object. The detectors (D1), (D2) are synchronized so that they capture simultaneously the corresponding spectral images of the tissue (Ti_{λ_1}), (Ti_{λ_2}) and in successive time-intervals, which are stored in the computer's data storage means. Generalizing, multiple spectral images can be captured simultaneously by combining multiple splitting elements, filters and sources.

[0063] FIG. 3 illustrates another method for capturing in different spectral bands simultaneously and in any spatial point of the area under analysis, the kinetics of the alterations in the characteristics of the light emitted from the tissue, before and the after the administration of the contrast enhancing agent. With the aid of a special prism (MIP) and imaging optics, it is possible to form multiple copies of the same image onto the surface of the same detector (D). Various optical filters (Of_{λ_1}), (Of_{λ_2}), (Of_{λ_3}), and (Of_{λ_4}), can be interposed along the length of the optical path of the rays that form the copies of the object's image, so that the multiple images correspond to different spectral areas.

[0064] For the clinical use of the methods of the invention, the different implementations of imaging described above can be integrated to conventional optical imaging diagnostic devices. Such devices are the various medical microscopes, colposcopes and endoscopes, which are routinely used for the in vivo diagnostic inspection of tissues. Imaging of internal tissues of the human body requires in most cases the illumination and imaging rays to travel along the same optical path, through the cavities of the body. As a result, in the common optical diagnostic devices the tissue's surface reflection contributes substantially to the formed image. This limits the imaging information for the subsurface characteristics, which is in general of great diagnostic importance. This problem becomes especially serious in epithelial tissues such as the cervix, larynx, and oral cavity, which are covered by fluids such as mucus and saliva. Surface reflection also obstructs the detection and the measurement of the alterations in the tissue's optical properties, induced after the administration of agents, which enhance the optical contrast between normal and pathologic tissue. More specifically, when an agent alters selectively the scattering characteristics of the pathologic tissue, the strong surface reflection that takes place in both pathologic (agent responsive) and normal (agent non responsive) tissue areas, occludes the diagnostic signal that originates from the

interaction of the agent with the subsurface features of the tissue. In other words, surface reflection constitutes optical noise in the diagnostic signal degrading substantially the perceived contrast between agent responsive and agent non-responsive tissue areas.

[0065] For accurate diagnoses using the aforementioned imaging devices, appropriate optics can be used to eliminate noise arising from surface reflection. **FIG. 4** illustrates a schematic diagram of a medical microscope that includes a light source (LS), a magnification selection mechanism (MS), an eyepiece (EP) and a mount for attaching the image capturing module (CA), (detector(s), readout electronics etc). To eliminate surface reflection, a pair of linear polarizers is employed. Light from the source passes through a linear polarizer (LPO) with the resulting linearly polarized light (LS) then impinging on the tissue. The surface reflected light (TS) has the same polarization plane as the incident light (Fresnel reflection). By placing another linear polarizer (IPO), oriented at a right angle with respect to the first, in the path of the light emitted from the tissue, the contribution of the surface reflected light is eliminated. The light that is not surface reflected enters the tissue, where due to multiple scattering, light polarization is randomized. Thus, a portion of the re-emitted light passes through the imaging polarization optics, carrying improved information for the subsurface features.

[0066] **FIG. 5** illustrates an endoscope that includes an eyepiece (EP), which can be adapted to an electronic imaging system, and optical fibers or crystals for the transmission of both illumination and image rays. The endoscope also includes a first linear polarizer (LPO), disposed in the optical path of the illumination rays (LE), and a second polarizer (IPO), oriented at right angles to the first, disposed in the path of the light emitted by the tissue (II). The polarizer (LPO) can be disposed as shown in the figure, or, alternatively, where the light enters the endoscope (IL). In the latter case, the endoscope has to be constructed using polarization preserving crystals or fiber optics for transferring the light. If polarization preserving light transmission media are used, then the polarizers for the imaging rays can be disposed in their path, in front or in back of the eyepiece (EP).

[0067] A problem for the effective clinical implementation of the method described above involves the micro-movements of the patient, which are present during the snapshot imaging of the same tissue area. This problem is eliminated when the patient is under anesthesia (open surgery). In most cases, however, the movements of the tissue relative to the image capturing module, occurring during the successive image capturing time-course, result in image pixels, with the same image coordinates, which do not correspond to exactly the same spatial point x,y of the tissue area under examination. This problem is typically encountered in colposcopy. A method for eliminating the influence to the measured temporal data of the relative movements between tissue and image capturing module is presented below.

[0068] A colposcopic apparatus, illustrated in **FIG. 6**, includes an articulated arm (AA), onto which the optical head (OH) is affixed. The head (OH) includes a light source (LS), an objective lens (OBJ), an eyepiece (EP) and optics for selecting the magnification (MS). The image-capturing module is attached to the optical head (OH), through an opto-mechanical adapter. A speculum (KD), which is used to open-up the vaginal canal for the visualization of the cervix, is connected mechanically to the optical head (OH), so that its longitudinal symmetry axis (LA) is perpendicular to the

central area of the objective lens (OBJ). The speculum enters the vagina and its blades are opened up compressing the side walls of the vagina. The speculum (KD), being mechanically connected with the optical head (OH), transfers any micro-movement of the patient to the optical head (OH), which, being mounted on an articulated arm (AA), follows these movements. Thus the relative position between tissue and optical head remains almost constant.

[0069] An important issue that must also be addressed for the successful clinical implementation of the diagnostic method described herein is the synchronization of the application of the pathology differentiating agent with the initiation of the snapshot imaging procedure. **FIG. 6**, illustrates an atomizer (A) attached to the optical head of the microscope. The unit (MIC) is comprised of electronics for controlling the agent sprayer and it can incorporate also the container for storing the agent. When the unit (MIC) receives the proper command from the computer, it sprays a predetermined amount of the agent onto the tissue surface, while the same or another command initiates the snapshot image capturing procedure.

[0070] The diagnostic examination of non-directly accessible tissues located in cavities of the human body (ear, cervix, oral cavity, esophagus, colon, stomach) is performed with the aid of common clinical microscopes. In these devices, the illumination-imaging rays are near co-axial. More specifically, the line perpendicular to the exit point of light into the air, and the line perpendicular to the objective lens, form an angle of a few degrees. As a result, these microscopes operate at a specific distance from the subject (working distance), where the illuminated tissue area coincides with the field-of-view of the imaging system. These microscopes are found to be inappropriate in cases where tissue imaging through human body cavities of small diameter and at short working distances is required. These technical limitations hinder the successful clinical implementation of the method described herein. As discussed above, elimination of surface reflection results in a substantial improvement of the diagnostic information obtained from the quantitative assessment of marker-tissue interaction kinetics. If a common clinical microscope is employed as the optical imaging module, then as a result of the above-mentioned illumination-imaging geometry, multiple reflections occur in the walls of the cavity before the light reaches the tissue under analysis. Multiple reflections are more numerous in colposcopy because of the highly reflective blades of the speculum, which is inserted into the vagina to facilitate the inspection of the cervix.

[0071] If the illuminator of the imaging apparatus emits linearly polarized light, the multiple reflections randomize the polarization plane of the incident light. As discussed above, if the light impinging on the tissue is not linearly polarized, then the elimination of the contribution from the surface reflection to the image can not be effective.

[0072] **FIG. 7** illustrates an optical imaging apparatus that includes a light source located at the central part of its front-aperture. With this arrangement, the central ray of the emitted light cone is coaxial with the central ray of the light beam that enters the imaging apparatus. This enables illumination rays to directly reach the tissue surface under examination before multiple reflections occur with the wall of the cavity or speculum. A reflective-objective lens is used, which includes a first reflection (1RM) and a second reflection (2RM) mirror. A light source (LS) is disposed at the rear of the second reflection mirror (2RM), together with, if

required, optics for light beam manipulation such as zooming and focusing (SO). The reflective-objective lens (RO), by replacing the common refractive-objective used in conventional microscopes, provides imaging capability in cavities of small diameter with the freedom of choosing the working distance. The zooming and focusing optics of the light beam can be adjusted simultaneously with the mechanism for varying the magnification of the optical imaging system so that the illumination area and the field-of-view of the imaging system vary simultaneously and proportionally. Thus, image brightness is preserved regardless of the magnification level of the lens. The imaging-illumination geometry embodied in this optical imaging apparatus, along with the light beam manipulation options, helps to eliminate the surface reflection contribution to the image and consequently helps to efficiently implement the method described herein.

[0073] Equivalents

[0074] Those skilled in the art will recognize, or be able to ascertain using no more than routine experimentation, many equivalents to the specific embodiments of the invention described herein. Such equivalents are intended to be encompassed by the following claims.

What is claimed:

1. Use of a differentiating agent in a method for monitoring the effects of said pathology differentiating agent on a tissue, comprising:

administering a pathology differentiating agent to said tissue,

exposing said tissue to broad band, continuous wave optical radiation; and

measuring a dynamic optical property of said tissue, after said tissue has interacted with said pathology differentiating agent, wherein said dynamic optical property is obtained from at least two values of a time-dependent optical property, thereby monitoring the effects of a pathology differentiating agent on a tissue.

2. Use of a differentiating agent in a method for monitoring the effects of said pathology differentiating agent on a tissue, comprising:

administering a pathology differentiating agent to said tissue;

exposing said tissue to optical radiation; and

measuring a non-fluorescent dynamic optical property of said tissue, after said tissue has interacted with said pathology differentiating agent, wherein said dynamic optical property is obtained from at least two values of a time-dependent optical property, thereby monitoring the effects of a pathology differentiating agent on a tissue.

3. Use of a differentiating agent in a method for monitoring the effects of said pathology differentiating agent on a tissue, the method comprising:

administering a pathology differentiating agent to said tissue;

exposing said tissue to optical radiation; and

while exposing said tissue to optical radiation, measuring a dynamic optical property of said tissue, after said tissue been interacted with said pathology differentiat-

ing agent, wherein said dynamic optical property is obtained from at least two values of a time-dependent optical property,

thereby monitoring the effects of a pathology differentiating agent on a tissue.

4. Use of a differentiating agent in a method for monitoring and mapping the effects of said pathology differentiating agent on a tissue, comprising:

administering a pathology differentiating agent to said tissue;

exposing said tissue to broad band, continuous wave optical radiation; and

after said tissue has interacted with said pathology differentiating agent, measuring a dynamic optical property at several locations of the tissue to produce a spatial map of the tissue, wherein said dynamic optical property is obtained from at least two values of a time-dependent optical property,

thereby monitoring and mapping the effects of a pathology differentiating agent on a tissue.

5. The use of any one of claims 1, 2, 3, or 4, further comprising measuring an optical property of said tissue before said tissue has interacted with said pathology differentiating agent.

6. The use of any one of claims 1, 2, 3, or 4, wherein said pathology differentiating agent is selected from the group consisting of an acidic solution and a basic solution.

7. The use of any one of claims 1, 2, 3, or 4, wherein said tissue is a tissue from the cervix of the uterus.

8. The use of any one of claims 1, 2, 3, or 4, wherein said tissue is a vagina tissue.

9. The use of any one of claims 1, 2, 3, or 4, wherein said tissue is a uterus tissue.

10. The use of any one of claims 1, 2, 3, or 4, wherein said tissue is selected from the group consisting of a skin tissue, an oral cavity tissue, a gastrointestinal track tissue, and a respiratory track tissue.

11. The use of a differentiating agent in a method for the in vivo diagnosis and mapping of a tissue abnormality in a subject, comprising

administering a pathology differentiating agent to a subject;

exposing a tissue in said subject to optical radiation; and

while exposing said tissue to optical radiation, monitoring the dynamic optical property of said tissue after said tissue has interacted with said pathology differentiating agent, wherein said dynamic optical property is obtained from at least two values of a time-dependent optical property,

thereby diagnosing and mapping a tissue abnormality in a subject.

12. The use of claim 11, wherein said optical radiation is broad band optical radiation.

13. The use of claim 11, wherein said optical radiation is polarized optical radiation.

14. The use of claim 11, wherein said tissue abnormality is selected from the group consisting of a tissue atypia, a tissue dysplasia, a tissue neoplasia, condylomas and cancer.

15. The use of claim 11, wherein said tissue abnormality is neoplasia of any grade.

16. The use of claim 11, wherein said tissue abnormality is a cervical intraepithelial neoplasia.

17. The use of claim 11, wherein said pathology differentiating agent is selected from the group consisting of an acidic solution and a basic solution.

18. The use of claim 11, wherein said tissue is a sample from the cervix of the uterus.

19. The use of claim 11, wherein said tissue is a vaginal tissue.

20. The use of claim 11, wherein said tissue is a uterine tissue.

21. The use of claim 11, wherein said tissue is selected from the group consisting of a skin tissue, an oral cavity tissue a gastrointestinal track tissue, and a respiratory track tissue.

22. A system for diagnosing and mapping a tissue abnormality, the system comprising a detector for measuring a difference in a dynamic optical property between a normal tissue and an abnormal tissue, said difference being caused by applying a pathology differentiating agent to the tissue.

23. The system of claim 22, further comprising an applicator for applying a pathology differentiating agent to the tissue.

24. The system of claim 23, wherein the applicator is an atomizer.

25. The system of claim 22, wherein the pathology differentiating agent is capable of altering the reflection characteristics of the abnormal tissue.

26. The system of claim 22, wherein the pathology differentiating agent is selected from the group consisting of an acidic solution and a basic solution.

27. The system of claim 22, wherein the detector is suitable for obtaining an emission spectrum for measuring the dynamic optical property of the tissue.

28. The system of claim 22, wherein the detector is suitable for measuring light emitted by the tissue for measuring the dynamic optical property of the tissue.

29. The system of claim 28, further comprising a beam splitter for splitting the light emitted by the tissue into a plurality of beams.

30. The system of claim 29, further comprising a plurality of optical filters, each of the filters transmitting over a specific spectral range, for filtering each of the plurality of beams, wherein the dynamic optical property of the tissue can be measured for each of the plurality of filtered beams.

31. The system of claim 22, further comprising a light source for shining light on the tissue to obtain the difference in the dynamic optical property.

32. The system of claim 22, wherein the light source includes at least one of a laser and a light-emitting diode.

33. The system of claim 22, wherein the light shone on the tissue is linearly polarized in a first direction.

34. The system of claim 33, further comprising a linear polarizer for polarizing light emitted from the tissue, in response to shining light thereon, in a direction substantially perpendicular to the first direction, thereby reducing a contribution of light reaching the detector arising from tissue surface reflection.

35. The system of claim 31, wherein the light source and the detector are included on an optical head.

36. The system of claim 35, further comprising a probing device connected to the optical head, said probing device coupled with the tissue so that relative motion between the tissue and the optical head is substantially eliminated.

37. The system of claim 36, wherein the tissue includes vaginal tissue, and the probing device includes a speculum for insertion into a vagina.

38. The system of claim 36, wherein the tissue includes gastrointestinal tissue or respiratory tissue, and the probing device includes an endoscope for insertion into a gastrointestinal track or a respiratory track, respectively.

39. The system of claim 36, further comprising a reflective objective lens so that the light source emits a cone of light having a symmetry axis that is substantially coaxial with a longitudinal symmetry axis of the objective lens to substantially reduce multiple reflections of light by at least one of the probing device and the tissue.

40. An apparatus for the in vivo diagnosis and mapping of a tissue abnormality comprising

a. Optics for collecting light re-emitted by the tissue under analysis for magnification and focusing an image of the tissue;

b. Optical imaging detector;

c. Means for modulating, transferring, displaying and capturing of the image of the tissue;

d. Computer which includes data storage, processing and analysis means;

e. Monitor for displaying images, curves and numerical data;

f. Optics for the optical multiplication of the image of the tissue;

g. Light source consisting of broad band, continuous wave optical radiation for illuminating the tissue;

h. Optics filters or selecting a spectral band of imaging and illumination;

i. Means for transmitting light and illuminating the tissue;

j. Control electronics; and

k. Software for analyzing and processing of data, and for capturing, registering and storing of tissue images obtained at specific spectral bands and time instances, before and after administering an agent or a combination of agents that alter the optical properties of atypical and pathological tissue areas, wherein dynamic, optical properties and their spatial map is obtained from the stored images.

41. The apparatus of claim 40, wherein a pseudo-color scale, which represents different values of the dynamic optical properties with different colors, is used for the visualization of tissue areas with abnormalities of different grade.

42. The apparatus of claim 40 or 41, wherein the image is used for the selective surgical removal of the atypical and pathological tissue area.

43. The apparatus of claim 40 or 41, wherein the image is used as a guide for taking a biopsy sample.

44. The apparatus if claim 40 or 41, wherein the image is used for the in vivo detection, mapping, and determination of the borders of epithelial lesions.

45. The apparatus or claim 40 or 41, wherein the parameters determined as a function of dynamic optical properties are used as diagnostic indices for the in vivo identification, grading of epithelial lesions and for the mapping of tissue areas with epithelial lesions of different grade.

46. A system for diagnosing and mapping a tissue abnormality, the system comprising:

- a detector for measuring a difference in a dynamic optical property between a normal tissue and an abnormal tissue, said difference being caused by applying a pathology differentiating agent to the tissue;
- a light source for shining light on the tissue, wherein the light source and the detector are included on an optical head;
- a probing device connected to the optical head, said probing device coupled with the tissue so that relative motion between the tissue and the optical head is substantially eliminated; and
- an applicator for applying a pathology differentiating agent to the tissue.

47. The system of claim 46, wherein the tissue includes vaginal tissue, and the probing device includes a speculum for insertion into a vagina.

48. The system of claim 46, wherein the tissue includes gastrointestinal tissue or respiratory tissue, and the probing

device includes an endoscope for insertion into a gastrointestinal track or a respiratory track, respectively.

49. The system of claim 46, further comprising

a beam splitter for splitting the light emitted by the tissue into a plurality of beams;

a plurality of optical filters, each of the filters transmitting over a specific spectral range, for filtering each of the plurality of beams, wherein the dynamic optical property of the tissue can be measured for each of the plurality of filtered beams.

50. The system of claim 46, wherein the light shone on the tissue is linearly polarized in a first direction.

51. The system of claim 50, further comprising a linear polarizer for polarizing light emitted from the tissue, in response to shining light thereon, in a direction substantially perpendicular to the first direction, thereby reducing a contribution of light reaching the detector arising from tissue surface reflection.

* * * * *

专利名称(译)	用于表征和绘制组织病变的方法和系统		
公开(公告)号	US20030114762A1	公开(公告)日	2003-06-19
申请号	US10/240367	申请日	2001-03-28
[标]申请(专利权)人(译)	巴拉斯KONSTANTINOS PELECOUDAS德米特里		
申请(专利权)人(译)	巴拉斯KONSTANTINOS PELECOUDAS德米特里		
[标]发明人	BALAS KONSTANTINOS PELECOUDAS DEMETRIOS		
发明人	BALAS, KONSTANTINOS PELECOUDAS, DEMETRIOS		
IPC分类号	G01N21/64 A61B1/00 A61B1/267 A61B1/273 A61B1/303 A61B1/307 A61B1/31 A61B5/00 A61B10/00 G01N21/27 G01N21/65 G01N21/75 G01N33/15 G01N33/48 A61B6/00		
CPC分类号	A61B1/303 A61B5/0059 A61B5/0075 A61B1/0646 A61B5/0088 A61B1/00186 A61B5/0084 A61B1/043 A61B5/0071 G01N21/21 G01N21/31 G01N21/6486 G01N2021/6417 G01N2021/6423		
优先权	20000100102 2000-03-28 GR 09/739089 2000-12-15 US		
其他公开文献	US7515952		
外部链接	Espacenet USPTO		

摘要(译)

本发明提供了用于体内，非侵入性，早期检测改变的方法和装置，以及在组织发育期间在上皮组织的生物化学和/或功能特征中引起的这些改变的等级的作图。异型，发育不良，瘤形成和癌症。该方法至少部分地基于同时测量从被检查组织重新发射的光的特征中的空间，时间和光谱变化，这是由于组合的组织激发与光和特殊化学剂。这些药剂的局部或系统给药导致正常和异常组织区域之间的渐逝对比度增强。该装置能够同时捕获一个或多个光谱带中的时间上连续的成像。基于测量数据，在检查区域的任何空间点确定表示药剂 - 组织相互作用动力学的特征曲线，以及从这些数据导出的数值参数。病变的定位和表征基于这些参数。

