



US008430817B1

(12) **United States Patent**
Al-Ali et al.

(10) **Patent No.:** **US 8,430,817 B1**
(45) **Date of Patent:** **Apr. 30, 2013**

(54) **SYSTEM FOR DETERMINING CONFIDENCE
IN RESPIRATORY RATE MEASUREMENTS**

(75) Inventors: **Ammar Al-Ali**, San Juan Capistrano,
CA (US); **Bilal Muhsin**, San Clemente,
CA (US); **Michael O'Reilly**, Dana Point,
CA (US); **James P. Welch**, Mission
Viejo, CA (US)

(73) Assignee: **Masimo Corporation**, Irvine, CA (US)

(*) Notice: Subject to any disclaimer, the term of this
patent is extended or adjusted under 35
U.S.C. 154(b) by 60 days.

(21) Appl. No.: **12/905,530**

(22) Filed: **Oct. 15, 2010**

Related U.S. Application Data

(60) Provisional application No. 61/366,866, filed on Jul.
22, 2010, provisional application No. 61/261,199,
filed on Nov. 13, 2009, provisional application No.
61/252,086, filed on Oct. 15, 2009.

(51) **Int. Cl.**
A61B 5/00 (2006.01)

(52) **U.S. Cl.**
USPC **600/301; 600/300; 600/323; 600/529;**
128/920

(58) **Field of Classification Search** **600/300–301,**
600/309–310, 320, 323, 372–384, 481, 529,
600/534–536, 586

See application file for complete search history.

(56) **References Cited**

U.S. PATENT DOCUMENTS

3,682,161 A 8/1972 Alibert
4,127,749 A 11/1978 Atoji et al.
4,326,143 A 4/1982 Guth et al.
4,537,200 A 8/1985 Widrow

4,884,809 A 12/1989 Rowan
4,960,128 A 10/1990 Gordon et al.
4,964,408 A 10/1990 Hink et al.
5,033,032 A 7/1991 Houghtaling
5,041,187 A 8/1991 Hink et al.
5,069,213 A 12/1991 Polczynski
5,163,438 A 11/1992 Gordon et al.
5,319,355 A 6/1994 Russek
5,337,744 A 8/1994 Branigan

(Continued)

FOREIGN PATENT DOCUMENTS

CA 2262236 4/2008
EP 0716628 12/1998

(Continued)

OTHER PUBLICATIONS

Imhoff, M. et al; "Alarm Algorithms in Critical Care Monitoring";
Anesth Analg 2006;102:1525-37.*

(Continued)

Primary Examiner — Bill Thomson


Assistant Examiner — Marie Archer

(74) *Attorney, Agent, or Firm* — Knobbe Martens Olson &
Bear LLP

(57) **ABSTRACT**

This disclosure describes, among other features, systems and methods for using multiple physiological parameter inputs to determine multiparameter confidence in respiratory rate measurements. For example, a patient monitoring system can programmatically determine multiparameter confidence in respiratory rate measurements obtained from an acoustic sensor based at least partly on inputs obtained from other non-acoustic sensors or monitors. The patient monitoring system can output a multiparameter confidence indication reflective of the programmatically-determined multiparameter confidence. The multiparameter confidence indication can assist a clinician in determining whether or how to treat a patient based on the patient's respiratory rate.

12 Claims, 22 Drawing Sheets



U.S. PATENT DOCUMENTS					
5,341,805 A	8/1994	Stavridi et al.	6,280,213 B1	8/2001	Tobler et al.
D353,195 S	12/1994	Savage et al.	6,285,896 B1	9/2001	Tobler et al.
D353,196 S	12/1994	Savage et al.	6,301,493 B1	10/2001	Marro et al.
5,377,676 A	1/1995	Vari et al.	6,317,627 B1	11/2001	Ennen et al.
D359,546 S	6/1995	Savage et al.	6,321,100 B1	11/2001	Parker
5,431,170 A	7/1995	Mathews	6,325,761 B1	12/2001	Jay
D361,840 S	8/1995	Savage et al.	6,334,065 B1	12/2001	Al-Ali et al.
D362,063 S	9/1995	Savage et al.	6,343,224 B1	1/2002	Parker
5,448,996 A	9/1995	Bellin et al.	6,349,228 B1	2/2002	Kiani et al.
5,452,717 A	9/1995	Branigan et al.	6,360,114 B1	3/2002	Diab et al.
D363,120 S	10/1995	Savage et al.	6,368,283 B1	4/2002	Xu et al.
5,456,252 A	10/1995	Vari et al.	6,371,921 B1	4/2002	Caro et al.
5,479,934 A	1/1996	Imran	6,377,829 B1	4/2002	Al-Ali
5,482,036 A	1/1996	Diab et al.	6,388,240 B2	5/2002	Schulz et al.
5,490,505 A	2/1996	Diab et al.	6,397,091 B2	5/2002	Diab et al.
5,494,043 A	2/1996	O'Sullivan et al.	6,430,437 B1	8/2002	Marro
5,533,511 A	7/1996	Kaspari et al.	6,430,525 B1	8/2002	Weber et al.
5,534,851 A	7/1996	Russek	6,463,311 B1	10/2002	Diab
5,561,275 A	10/1996	Savage et al.	6,470,199 B1	10/2002	Kopotic et al.
5,562,002 A	10/1996	Lalin	6,486,588 B2	11/2002	Doron et al.
5,590,649 A	1/1997	Caro et al.	6,501,975 B2	12/2002	Diab et al.
5,602,924 A	2/1997	Durand et al.	6,505,059 B1	1/2003	Kollias et al.
5,632,272 A	5/1997	Diab et al.	6,515,273 B2	2/2003	Al-Ali
5,638,816 A	6/1997	Kiani-Azarbayjany et al.	6,517,497 B2	2/2003	Rymut et al.
5,638,818 A	6/1997	Diab et al.	6,519,487 B1	2/2003	Parker
5,645,440 A	7/1997	Tobler et al.	6,525,386 B1	2/2003	Mills et al.
5,685,299 A	11/1997	Diab et al.	6,526,300 B1	2/2003	Kiani et al.
5,724,983 A	3/1998	Selker et al.	6,541,756 B2	4/2003	Schulz et al.
D393,830 S	4/1998	Tobler et al.	6,542,764 B1	4/2003	Al-Ali et al.
5,743,262 A	4/1998	Lepper, Jr. et al.	6,580,086 B1	6/2003	Schulz et al.
5,758,644 A	6/1998	Diab et al.	6,584,336 B1	6/2003	Ali et al.
5,760,910 A	6/1998	Lepper, Jr. et al.	6,595,316 B2	7/2003	Cybulski et al.
5,769,785 A	6/1998	Diab et al.	6,597,932 B2	7/2003	Tian et al.
5,782,757 A	7/1998	Diab et al.	6,597,933 B2	7/2003	Kiani et al.
5,785,659 A	7/1998	Caro et al.	6,606,511 B1	8/2003	Ali et al.
5,791,347 A	8/1998	Flaherty et al.	6,632,181 B2	10/2003	Flaherty et al.
5,810,734 A	9/1998	Caro et al.	6,639,668 B1	10/2003	Trepagnier
5,819,007 A	10/1998	Elghazzawi	6,640,116 B2	10/2003	Diab
5,823,950 A	10/1998	Diab et al.	6,643,530 B2	11/2003	Diab et al.
5,830,131 A	11/1998	Caro et al.	6,650,917 B2	11/2003	Diab et al.
5,833,618 A	11/1998	Caro et al.	6,654,624 B2	11/2003	Diab et al.
5,860,919 A	1/1999	Kiani-Azarbayjany et al.	6,658,276 B2	12/2003	Kiani et al.
5,865,736 A	2/1999	Baker, Jr. et al.	6,661,161 B1	12/2003	Lanzo et al.
5,890,929 A	4/1999	Mills et al.	6,671,531 B2	12/2003	Al-Ali et al.
5,904,654 A	5/1999	Wohlmann et al.	6,678,543 B2	1/2004	Diab et al.
5,919,134 A	7/1999	Diab	6,684,090 B2	1/2004	Ali et al.
5,934,925 A	8/1999	Tobler et al.	6,684,091 B2	1/2004	Parker
5,940,182 A	8/1999	Lepper, Jr. et al.	6,697,656 B1	2/2004	Al-Ali
5,995,855 A	11/1999	Kiani et al.	6,697,657 B1	2/2004	Shehada et al.
5,997,343 A	12/1999	Mills et al.	6,697,658 B2	2/2004	Al-Ali
6,002,952 A	12/1999	Diab et al.	RE38,476 E	3/2004	Diab et al.
6,011,986 A	1/2000	Diab et al.	6,699,194 B1	3/2004	Diab et al.
6,027,452 A	2/2000	Flaherty et al.	6,714,804 B2	3/2004	Al-Ali et al.
6,036,642 A	3/2000	Diab et al.	RE38,492 E	4/2004	Diab et al.
6,045,509 A	4/2000	Caro et al.	6,721,582 B2	4/2004	Trepagnier et al.
6,067,462 A	5/2000	Diab et al.	6,721,585 B1	4/2004	Parker
6,081,735 A	6/2000	Diab et al.	6,725,075 B2	4/2004	Al-Ali
6,088,607 A	7/2000	Diab et al.	6,728,560 B2	4/2004	Kollias et al.
6,110,522 A	8/2000	Lepper, Jr. et al.	6,735,459 B2	5/2004	Parker
6,124,597 A	9/2000	Shehada et al.	6,745,060 B2	6/2004	Diab et al.
6,128,521 A	10/2000	Marro et al.	6,754,516 B2 *	6/2004	Mannheimer 600/323
6,129,675 A	10/2000	Jay	6,760,607 B2	7/2004	Al-Ali
6,144,868 A	11/2000	Parker	6,770,028 B1	8/2004	Ali et al.
6,151,516 A	11/2000	Kiani-Azarbayjany et al.	6,771,994 B2	8/2004	Kiani et al.
6,152,754 A	11/2000	Gerhardt et al.	6,792,300 B1	9/2004	Diab et al.
6,157,850 A	12/2000	Diab et al.	6,813,511 B2	11/2004	Diab et al.
6,165,005 A	12/2000	Mills et al.	6,816,741 B2	11/2004	Diab
6,184,521 B1	2/2001	Coffin, IV et al.	6,822,564 B2	11/2004	Al-Ali
6,206,830 B1	3/2001	Diab et al.	6,826,419 B2	11/2004	Diab et al.
6,229,856 B1	5/2001	Diab et al.	6,830,711 B2	12/2004	Mills et al.
6,232,609 B1	5/2001	Snyder et al.	6,850,787 B2	2/2005	Weber et al.
6,236,872 B1	5/2001	Diab et al.	6,850,788 B2	2/2005	Al-Ali
6,241,683 B1	6/2001	Macklem et al.	6,852,083 B2	2/2005	Caro et al.
6,248,083 B1	6/2001	Smith et al.	6,861,639 B2	3/2005	Al-Ali
6,253,097 B1	6/2001	Aronow et al.	6,898,452 B2	5/2005	Al-Ali et al.
6,256,523 B1	7/2001	Diab et al.	6,920,345 B2	7/2005	Al-Ali et al.
6,263,222 B1	7/2001	Diab et al.	6,931,268 B1	8/2005	Kiani-Azarbayjany et al.
6,278,522 B1	8/2001	Lepper, Jr. et al.	6,934,570 B2	8/2005	Kiani et al.
			6,939,305 B2	9/2005	Flaherty et al.

US 8,430,817 B1

Page 3

6,943,348	B1	9/2005	Coffin IV	7,500,950	B2	3/2009	Al-Ali et al.
6,950,687	B2	9/2005	Al-Ali	7,509,154	B2	3/2009	Diab et al.
6,961,598	B2	11/2005	Diab	7,509,494	B2	3/2009	Al-Ali
6,970,792	B1	11/2005	Diab	7,510,849	B2	3/2009	Schurman et al.
6,979,812	B2	12/2005	Al-Ali	7,526,328	B2	4/2009	Diab et al.
6,985,764	B2	1/2006	Mason et al.	7,530,942	B1	5/2009	Diab
6,993,371	B2	1/2006	Kiani et al.	7,530,949	B2	5/2009	Al Ali et al.
6,996,427	B2	2/2006	Ali et al.	7,530,955	B2	5/2009	Diab et al.
6,999,904	B2	2/2006	Weber et al.	7,563,110	B2	7/2009	Al-Ali et al.
7,003,338	B2	2/2006	Weber et al.	7,596,398	B2	9/2009	Al-Ali et al.
7,003,339	B2	2/2006	Diab et al.	7,618,375	B2	11/2009	Flaherty
7,015,451	B2	3/2006	Dalke et al.	D606,659	S	12/2009	Kiani et al.
7,024,233	B2	4/2006	Ali et al.	7,647,083	B2	1/2010	Al-Ali et al.
7,027,849	B2	4/2006	Al-Ali	D609,193	S	2/2010	Al-Ali et al.
7,030,749	B2	4/2006	Al-Ali	D614,305	S	4/2010	Al-Ali et al.
7,039,449	B2	5/2006	Al-Ali	7,690,378	B1 *	4/2010	Turcott 128/204.23
7,041,060	B2	5/2006	Flaherty et al.	RE41,317	E	5/2010	Parker
7,044,918	B2	5/2006	Diab	7,729,733	B2	6/2010	Al-Ali et al.
7,067,893	B2	6/2006	Mills et al.	7,734,320	B2	6/2010	Al-Ali
7,096,052	B2	8/2006	Mason et al.	7,761,127	B2	7/2010	Al-Ali et al.
7,096,054	B2	8/2006	Abdul-Hafiz et al.	7,761,128	B2	7/2010	Al-Ali et al.
7,096,060	B2	8/2006	Arand et al.	7,764,982	B2	7/2010	Dalke et al.
7,132,641	B2	11/2006	Schulz et al.	D621,516	S	8/2010	Kiani et al.
7,142,901	B2	11/2006	Kiani et al.	7,791,155	B2	9/2010	Diab
7,149,561	B2	12/2006	Diab	7,801,581	B2	9/2010	Diab
7,186,966	B2	3/2007	Al-Ali	7,822,452	B2	10/2010	Schurman et al.
7,190,261	B2	3/2007	Al-Ali	RE41,912	E	11/2010	Parker
7,215,984	B2	5/2007	Diab	7,844,313	B2	11/2010	Kiani et al.
7,215,986	B2	5/2007	Diab	7,844,314	B2	11/2010	Al-Ali
7,221,971	B2	5/2007	Diab	7,844,315	B2	11/2010	Al-Ali
7,225,006	B2	5/2007	Al-Ali et al.	7,865,222	B2	1/2011	Weber et al.
7,225,007	B2	5/2007	Al-Ali	7,873,497	B2	1/2011	Weber et al.
RE39,672	E	6/2007	Shehada et al.	7,880,606	B2	2/2011	Al-Ali
7,239,905	B2	7/2007	Kiani-Azarbayjany et al.	7,880,626	B2	2/2011	Al-Ali et al.
7,245,953	B1	7/2007	Parker	7,891,355	B2	2/2011	Al-Ali et al.
7,254,429	B2	8/2007	Schurman et al.	7,894,868	B2	2/2011	Al-Ali et al.
7,254,431	B2	8/2007	Al-Ali	7,899,507	B2	3/2011	Al-Ali et al.
7,254,433	B2	8/2007	Diab et al.	7,899,518	B2	3/2011	Trepagnier et al.
7,254,434	B2	8/2007	Schulz et al.	7,904,132	B2	3/2011	Weber et al.
7,272,425	B2	9/2007	Al-Ali	7,909,772	B2	3/2011	Popov et al.
7,274,955	B2	9/2007	Kiani et al.	7,910,875	B2	3/2011	Al-Ali
D554,263	S	10/2007	Al-Ali	7,919,713	B2	4/2011	Al-Ali et al.
7,280,858	B2	10/2007	Al-Ali et al.	7,937,128	B2	5/2011	Al-Ali
7,289,835	B2	10/2007	Mansfield et al.	7,937,129	B2	5/2011	Mason et al.
7,292,883	B2	11/2007	De Felice et al.	7,937,130	B2	5/2011	Diab et al.
7,295,866	B2	11/2007	Al-Ali	7,941,199	B2	5/2011	Kiani
7,328,053	B1	2/2008	Diab et al.	7,951,086	B2	5/2011	Flaherty et al.
7,332,784	B2	2/2008	Mills et al.	7,957,780	B2	6/2011	Lamego et al.
7,340,287	B2	3/2008	Mason et al.	7,962,188	B2	6/2011	Kiani et al.
7,341,559	B2	3/2008	Schulz et al.	7,962,190	B1	6/2011	Diab et al.
7,343,186	B2	3/2008	Lamego et al.	7,976,472	B2	7/2011	Kiani
D566,282	S	4/2008	Al-Ali et al.	7,988,637	B2	8/2011	Diab
7,355,512	B1	4/2008	Al-Ali	7,990,382	B2	8/2011	Kiani
7,356,365	B2	4/2008	Schurman	7,991,446	B2	8/2011	Ali et al.
7,371,981	B2	5/2008	Abdul-Hafiz	8,000,761	B2	8/2011	Al-Ali
7,373,193	B2	5/2008	Al-Ali et al.	8,008,088	B2	8/2011	Bellott et al.
7,373,194	B2	5/2008	Weber et al.	RE42,753	E	9/2011	Kiani-Azarbayjany et al.
7,376,453	B1	5/2008	Diab et al.	8,019,400	B2	9/2011	Diab et al.
7,377,794	B2	5/2008	Al Ali et al.	8,028,701	B2	10/2011	Al-Ali et al.
7,377,899	B2	5/2008	Weber et al.	8,029,765	B2	10/2011	Bellott et al.
7,383,070	B2	6/2008	Diab et al.	8,036,728	B2	10/2011	Diab et al.
7,398,115	B2	7/2008	Lynn	8,046,040	B2	10/2011	Ali et al.
7,415,297	B2	8/2008	Al-Ali et al.	8,046,041	B2	10/2011	Diab et al.
7,428,432	B2	9/2008	Ali et al.	8,046,042	B2	10/2011	Diab et al.
7,438,683	B2	10/2008	Al-Ali et al.	8,048,040	B2	11/2011	Kiani
7,440,787	B2	10/2008	Diab	8,050,728	B2	11/2011	Al-Ali et al.
7,454,240	B2	11/2008	Diab et al.	RE43,169	E	2/2012	Parker
7,467,002	B2	12/2008	Weber et al.	8,118,620	B2	2/2012	Al-Ali et al.
7,469,157	B2	12/2008	Diab et al.	8,126,528	B2	2/2012	Diab et al.
7,471,969	B2	12/2008	Diab et al.	8,128,572	B2	3/2012	Diab et al.
7,471,971	B2	12/2008	Diab et al.	8,130,105	B2	3/2012	Al-Ali et al.
7,483,729	B2	1/2009	Al-Ali et al.	8,145,287	B2	3/2012	Diab et al.
7,483,730	B2	1/2009	Diab et al.	8,150,487	B2	4/2012	Diab et al.
7,489,958	B2	2/2009	Diab et al.	8,175,672	B2	5/2012	Parker
7,496,391	B2	2/2009	Diab et al.	8,180,420	B2	5/2012	Diab et al.
7,496,393	B2	2/2009	Diab et al.	8,182,443	B1	5/2012	Kiani
D587,657	S	3/2009	Al-Ali et al.	8,185,180	B2	5/2012	Diab et al.
7,499,741	B2	3/2009	Diab et al.	8,190,223	B2	5/2012	Al-Ali et al.
7,499,835	B2	3/2009	Weber et al.	8,190,227	B2	5/2012	Diab et al.

8,203,438	B2	6/2012	Kiani et al.	
8,224,411	B2	7/2012	Al-Ali et al.	
8,228,181	B2	7/2012	Al-Ali	
8,229,533	B2	7/2012	Diab et al.	
8,255,026	B1	8/2012	Al-Ali	
8,255,028	B2	8/2012	Al-Ali et al.	
8,260,577	B2	9/2012	Weber et al.	
8,274,360	B2	9/2012	Sampath et al.	
2002/0193670	A1	12/2002	Garfield et al.	
2003/0015368	A1	1/2003	Cybulski et al.	
2003/0158466	A1	8/2003	Lynn et al.	
2004/0133087	A1	7/2004	Ali et al.	
2004/0158162	A1	8/2004	Narimatsu	
2006/0047215	A1	3/2006	Newman et al.	
2006/0149144	A1	7/2006	Lynn et al.	
2006/0155206	A1	7/2006	Lynn	
2006/0155207	A1	7/2006	Lynn et al.	
2006/0161071	A1	7/2006	Lynn et al.	
2006/0189880	A1	8/2006	Lynn et al.	
2006/0195041	A1	8/2006	Lynn et al.	
2006/0235324	A1	10/2006	Lynn	
2006/0238333	A1	10/2006	Welch et al.	
2007/0093721	A1	4/2007	Lynn et al.	
2007/0129647	A1	6/2007	Lynn	
2007/0149860	A1	6/2007	Lynn et al.	
2007/0185397	A1	8/2007	Govari et al.	
2007/0282212	A1	12/2007	Sierra et al.	
2008/0039735	A1	2/2008	Hickerson	
2008/0076972	A1	3/2008	Dorogusker et al.	
2009/0018429	A1	1/2009	Saliga et al.	
2009/0093687	A1	4/2009	Telfort et al.	
2009/0187065	A1	7/2009	Basinger	
2009/0299157	A1	12/2009	Telfort et al.	
2010/0274099	A1	10/2010	Telfort et al.	
2010/0298661	A1 *	11/2010	McCombie et al.	600/301
2010/0324377	A1 *	12/2010	Woehrle	600/300
2011/0009710	A1 *	1/2011	Kroeger et al.	600/300
2011/0118573	A1 *	5/2011	McKenna	600/323

FOREIGN PATENT DOCUMENTS

EP	0659058	1/1999
EP	1207536	5/2002
GB	2358546	11/1999
JP	6214898	1/1987
JP	01-309872	6/1998
JP	10-155755	6/1998
JP	2001-50713	5/1999
JP	2003-329719	11/2003
WO	WO 94/05207	3/1994
WO	WO 94/13207	6/1994
WO	WO 95/29632	11/1995
WO	WO 99/53277	10/1999
WO	WO 00/10462	3/2000
WO	WO 01/34033	5/2001
WO	WO 01/78059	10/2001
WO	WO 01/97691	12/2001
WO	WO 02/03042	1/2002
WO	WO 03/058646	7/2003
WO	WO 03/087737	10/2003

WO	WO 2004/000111	12/2003
WO	WO 2004/004411	1/2004
WO	WO 2005/096931	10/2005
WO	WO 2005/099562	10/2005
WO	WO 2008/017246	2/2008
WO	WO 2008/148172	12/2008
WO	WO 2009093159	A1 * 7/2009
WO	WO 2009/137524	11/2009

OTHER PUBLICATIONS

Gorges, M. et al; "Improving Alarm Performance in the Medical Intensive Care Unit Using Delays and Clinical Context"; Technology, Computing, and Simulation; vol. 108, No. 5, May 2009; p. 1546-1552.*

Chambrin, M-C.; "Alarms in the intensive care unit: how can the number of false alarms be reduced?"; Critical Care Aug. 2001, vol. 5 No. 4; p. 1-5.*

Watt, R. C.; "Alarms and Anesthesia. Challenges in the design of Intelligent systems for Patient Monitoring"; IEEE Engineering in Medicine and biology; Dec. 1993, p. 34-41.*

Sierra et al., Monitoring Respiratory Rate Based on Tracheal Sounds. First Experiences, Proceedings of the 26th Annual Int'l Conf. of the IEEE EMBS (Sep. 2004), 317-320.

International Search Report and Written Opinion issued in application No. PCT/US2010/052756 on Feb. 6, 2012.

U.S. Appl. No. 12/904,775, filed Oct. 14, 2010, Fechter et al.

U.S. Appl. No. 12/904,789, filed Oct. 14, 2010, Telfort, Valery et al.

U.S. Appl. No. 12/904,823, filed Oct. 14, 2010, Al-Ali et al.

U.S. Appl. No. 12/904,836, filed Oct. 14, 2010, Al-Ali, Ammar.

U.S. Appl. No. 12/904,890, filed Oct. 14, 2010, Telfort et al.

U.S. Appl. No. 12/904,907, filed Oct. 14, 2010, Telfort et al.

U.S. Appl. No. 12/904,931, filed Oct. 14, 2010, Telfort et al.

U.S. Appl. No. 12/904,938, filed Oct. 14, 2010, Telfort et al.

U.S. Appl. No. 12/905,036, filed Oct. 14, 2010, Kiani et al.

U.S. Appl. No. 12/905,384, filed Oct. 15, 2010, Al-Ali et al.

U.S. Appl. No. 12/905,449, filed Oct. 15, 2010, Al-Ali et al.

U.S. Appl. No. 12/905,489, filed Oct. 15, 2010, Weber et al.

U.S. Appl. No. 12/960,325, filed Dec. 3, 2010, Al-Ali, Ammar et al.

Analog Devices, 12-Bit Serial Input Multiplying D/A Converter, Product Data Sheet, 2000.

International Search Report & Written Opinion, PCT Application PCT/US2010/052758, Feb. 10, 2011; 12 pages.

International Search Report & Written Opinion, PCT Application PCT/US2010/058981, Feb. 17, 2011; 11 pages.



International Search Report, PCT Application PCT/CA2003/000536, Dec. 11, 2003; 2 pages.

International Search Report, PCT Application PCT/US2009/069287, Mar. 30, 2010; 7 pages.

Welch Allyn, ECG ASIC, Product Data Sheet, 2001.

Eldor et al., "A device for monitoring ventilation during anaesthesia; the paratracheal audible respiratory monitor", Canadian Journal of Anaesthesia, 1990, vol. 9, No. 1, p. 95-98.

* cited by examiner






FIG. 1C




FIG. 2




FIG. 3A




FIG. 3B




FIG. 3C




FIG. 4




FIG. 5




FIG. 6A




FIG. 6B-1

FIG. 6B-1

FIG. 6B-2

FIG. 6B




FIG. 6B-2




FIG. 6C




FIG. 7




FIG. 8A




FIG. 8B




FIG. 8C




FIG. 8D




FIG. 9




FIG. 10




FIG. 11



FIG. 12

FIG. 13




FIG. 14




FIG. 15




FIG. 16




FIG. 17

SYSTEM FOR DETERMINING CONFIDENCE IN RESPIRATORY RATE MEASUREMENTS

RELATED APPLICATIONS

This application claims priority from U.S. Provisional Patent Application No. 61/252,086 filed Oct. 15, 2009, entitled "Pulse Oximetry System for Determining Confidence in Respiratory Rate Measurements," from U.S. Provisional Patent Application No. 61/261,199, filed Nov. 13, 2009, entitled "Pulse Oximetry System with Adjustable Alarm Delay," and from U.S. Provisional Patent Application No. 61/366,866, filed Jul. 22, 2010, entitled "Pulse Oximetry System for Determining Confidence in Respiratory Rate Measurements," the disclosures of which are hereby incorporated by reference in their entirety.

BACKGROUND

Hospitals, nursing homes, and other patient care facilities typically include patient monitoring devices at one or more bedsides in the facility. Patient monitoring devices generally include sensors, processing equipment, and displays for obtaining and analyzing a patient's physiological parameters. Physiological parameters include, for example, blood pressure, respiratory rate, oxygen saturation (SpO_2) level, other blood constitutions and combinations of constitutions, and pulse, among others. Clinicians, including doctors, nurses, and certain other caregiver personnel use the physiological parameters obtained from the patient to diagnose illnesses and to prescribe treatments. Clinicians can also use the physiological parameters to monitor a patient during various clinical situations to determine whether to increase the level of care given to the patient. Various patient monitoring devices are commercially available from Masimo Corporation ("Masimo") of Irvine, Calif.

During and after surgery and in other care situations, respiratory rate is a frequently monitored physiological parameter of a patient. Respiratory rate can be indicated as the number of breaths a person takes within a certain amount of time, such as breaths per minute. For example, a clinician (such as a nurse, doctor, or the like) can use respiratory rate measurements to determine whether a patient is experiencing respiratory distress and/or dysfunction.

SUMMARY OF DISCLOSURE

A system for determining multiparameter confidence in a respiratory rate measurement from a medical patient, the system comprising: an optical sensor comprising: a light emitter configured to impinge light on body tissue of a living patient, the body tissue comprising pulsating blood, and a detector responsive to the light after attenuation by the body tissue, wherein the detector is configured to generate a photoplethysmographic signal indicative of a physiological characteristic of the living patient; an ECG sensor configured to obtain an electrical signal from the living patient; an acoustic sensor, the acoustic sensor configured to obtain an acoustic respiratory signal from the living patient; and a processor configured to: derive a first respiratory rate measurement from the acoustic respiratory signal, derive a second respiratory rate measurement from one or both of the photoplethysmographic signal and the electrical signal, and use the second respiratory rate measurement to calculate a confidence in the first respiratory rate measurement.

A system for determining confidence in a respiratory rate measurement from a medical patient, the system comprising:

a first physiological sensor configured to obtain a physiological signal from a patient, the first physiological sensor comprising at least one of the following: an ECG sensor, a bio-impedance sensor, and a capnography sensor; an acoustic sensor configured to obtain an acoustic respiratory signal from the living patient; and a processor configured to: obtain a first respiratory rate measurement from the physiological signal, obtain a second respiratory rate measurement from the acoustic respiratory signal, and calculate a confidence in the first respiratory rate measurement responsive to the first and second respiratory rate measurements.

A method of analyzing respiratory rate monitoring parameters to determine confidence in a measured respiratory rate, the method comprising: obtaining a first respiratory measurement from a first physiological device, the first physiological device comprising an acoustic sensor; obtaining a second respiratory measurement from a second physiological device; determining a third respiratory rate measurement based at least in part on the first and second respiratory rate measurements; and outputting the third respiratory rate measurement.

For purposes of summarizing the disclosure, certain aspects, advantages and novel features of the inventions have been described herein. It is to be understood that not necessarily all such advantages can be achieved in accordance with any particular embodiment of the inventions disclosed herein. Thus, the inventions disclosed herein can be embodied or carried out in a manner that achieves or optimizes one advantage or group of advantages as taught herein without necessarily achieving other advantages as can be taught or suggested herein.

BRIEF DESCRIPTION OF THE DRAWINGS

Throughout the drawings, reference numbers can be re-used to indicate correspondence between referenced elements. The drawings are provided to illustrate embodiments of the inventions described herein and not to limit the scope thereof.

FIGS. 1A-B are block diagrams illustrating physiological monitoring systems in accordance with embodiments of the disclosure;

FIG. 1C is a top perspective view illustrating portions of a sensor assembly in accordance with an embodiment of the disclosure;

FIG. 2 illustrates a block diagram of an embodiment of a multiparameter patient monitoring system that includes an acoustic respiratory monitoring (ARM) sensor and an optical sensor.

FIG. 3A illustrates an embodiment of an envelope of a photoplethysmograph waveform.

FIG. 3B schematically illustrates an example calculation of pulse wave transit time from two physiological signal inputs.

FIG. 3C illustrates example power spectrum plots of pulse wave transit time variability and heart rate variability for determining respiratory rate measurements.

FIG. 4 illustrates an embodiment of the multiparameter patient monitoring system of FIG. 2 coupled to a patient.

FIG. 5 illustrates a block diagram of an embodiment of a multiparameter patient monitoring system.

FIGS. 6A through 6C illustrate block diagrams of embodiments of respiratory rate measurement calculation systems.

FIG. 7 illustrates an example multiparameter physiological monitor.

FIGS. 8A through 8D illustrate example multiparameter physiological monitor displays.

FIG. 9 illustrates an embodiment of a patient monitoring process in which a user can specify a delay time for an alarm to be triggered.

FIG. 10 illustrates an embodiment of a multiparameter patient monitoring process that allows for dynamic modification of an alarm delay.

FIGS. 11 through 17 illustrate embodiments of parameter confidence displays.

DETAILED DESCRIPTION

Acoustic sensors, including piezoelectric acoustic sensors, can be used to measure breath sounds and other biological sounds of a patient. Breath sounds obtained from an acoustic sensor placed on the neck, chest, and/or other suitable location can be processed by a patient monitor to derive one or more physiological parameters of a patient, including respiratory rate. Respiratory rate can also be determined from other physiological signals (for example, an ECG signal, a plethysmographic signal, a bioimpedance signal, and/or the like) obtained using other sensors and/or instruments.

Respiratory rate measurements derived from a single sensor or sensor type can be less accurate at times due to noise, sensor limitations, body movement, and/or other reasons. Accordingly, improved multiparameter respiratory rate measurements can be obtained by jointly processing multiple physiological signals from multiple sensors and/or sensor types. Alternately, or in addition, respiratory rate measurements derived from a physiological signal obtained from one type of sensor can be used to continuously or periodically refine or assess confidence in the respiratory rate measurements derived from a physiological signal obtained from another type of sensor. For example, in certain embodiments, respiratory rate measurements derived from a plethysmographic signal obtained by an optical sensor can be used to improve or determine confidence in the respiratory rate derived from an acoustic signal obtained by an acoustic sensor.

This disclosure describes, among other features, systems and methods for using multiple physiological signals to improve respiratory or other physiological parameter measurements reflective of a patient's condition and/or to determine confidence in these physiological parameter measurements. In certain embodiments, a patient monitoring system comprises one or more physiological sensors applied to a living patient and a processor to monitor the physiological signals received from physiological sensors. The physiological sensors can include, for example, acoustic sensors for acquiring breath and/or heart sounds, electrodes for acquiring ECG and/or bioimpedance signals, and noninvasive optical sensors to perform pulse oximetry and related noninvasive analysis of blood constituents.

In particular, in certain embodiments, a patient monitoring system can programmatically determine multiparameter confidence in respiratory rate measurements obtained from an acoustic sensor based at least partly on inputs obtained from other non-acoustic sensors or monitors. The patient monitoring system can output a multiparameter confidence indication reflective of the programmatically-determined multiparameter confidence. The multiparameter confidence indication can assist a clinician in determining whether or how to treat a patient based on the patient's respiratory rate.

In certain embodiments, the patient monitoring system can determine the multiparameter confidence at least in part by receiving signals from multiple physiological parameter monitoring devices that are reflective of respiratory rate. For example, a multiparameter patient monitoring unit can

receive a signal reflective of a respiratory rate from both an acoustic sensor and an optical sensor. Respiratory rate measurements can be extracted and/or derived from each of the physiological parameter signals. The respiratory rate measurement from the acoustic sensor can be compared with the respiratory rate measurement derived from the optical sensor signal. Based at least partly on this comparison, a determination of multiparameter confidence in the acoustically-derived respiratory rate can be made. A visual or audible indicator corresponding to this multiparameter confidence determination, including possibly an alarm, can be output for presentation to a clinician.

Additionally, in certain embodiments, the respiratory rate measurement output to the clinician can be generated based at least partly on a combination of the multiple respiratory rate measurements. For example, a respiratory rate measurement derived from an optical sensor signal can be combined with a respiratory rate measurement derived from an acoustic sensor signal to produce an overall respiratory rate. The overall respiratory rate measurement can be output to the patient monitor display and/or can be output over a network to another device.

Moreover, in certain embodiments, the patient monitoring systems and methods disclosed herein can assess multiparameter confidence and/or determine respiratory rate based at least partly on signals received from other physiological parameter monitoring devices. For example, various measurements obtained from a capnograph, an electrocardiograph (ECG), a bioimpedance device, or from other monitoring devices or sensors can be used to assess multiparameter confidence in acoustic respiratory rate measurements and/or to determine an overall respiratory rate output.

For purposes of illustration, this disclosure is described primarily in the context of respiratory rate. However, the features described herein can be applied to other respiratory parameters, including, for example, inspiratory time, expiratory time, inspiratory to expiratory ratio, inspiratory flow, expiratory flow, tidal volume, minute volume, apnea duration, breath sounds (including, e.g., rales, rhonchi, or stridor), changes in breath sounds, and the like. Moreover, the features described herein can also be applied to other physiological parameters and/or vital signs. For example, outputs from multiple monitoring devices or sensors (e.g., an optical sensor and an ECG monitor) can be used to assess multiparameter confidence in heart rate measurements, among other parameters.

Referring to the drawings, FIGS. 1A through 1C illustrate example patient monitoring systems, sensors, and cables that can be used to derive a respiratory rate measurement from a patient. FIGS. 2 through 8 illustrate multiparameter respiratory rate embodiments. The embodiments of FIGS. 2 through 8 can be implemented at least in part using the systems and sensors described in FIGS. 1A through 1C.

Turning to FIG. 1A, an embodiment of a physiological monitoring system 10 is shown. In the physiological monitoring system 10, a medical patient 12 is monitored using one or more sensor assemblies 13, each of which transmits a signal over a cable 15 or other communication link or medium to a physiological monitor 17. The physiological monitor 17 includes a processor 19 and, optionally, a display 11. The one or more sensors 13 include sensing elements such as, for example, acoustic piezoelectric devices, electrical ECG leads, optical sensors, or the like. The sensors 13 can generate respective signals by measuring a physiological parameter of the patient 12. The signals are then processed by one or more processors 19. The one or more processors 19 then communicate the processed signal to the display 11. In an embodi-

ment, the display 11 is incorporated in the physiological monitor 17. In another embodiment, the display 11 is separate from the physiological monitor 17. In one embodiment, the monitoring system 10 is a portable monitoring system. In another embodiment, the monitoring system 10 is a pod, without a display, that is adapted to provide physiological parameter data to a display.

For clarity, a single block is used to illustrate the one or more sensors 13 shown in FIG. 1A. It should be understood that the sensor 13 shown is intended to represent one or more sensors. In an embodiment, the one or more sensors 13 include a single sensor of one of the types described below. In another embodiment, the one or more sensors 13 include one or more acoustic sensors. In still another embodiment, the one or more sensors 13 include one or more acoustic sensors and one or more ECG sensors, optical sensors, bioimpedance sensors, capnography sensors, and the like. In each of the foregoing embodiments, additional sensors of different types are also optionally included. Other combinations of numbers and types of sensors are also suitable for use with the physiological monitoring system 10.

In some embodiments of the system shown in FIG. 1A, all of the hardware used to receive and process signals from the sensors are housed within the same housing. In other embodiments, some of the hardware used to receive and process signals is housed within a separate housing. In addition, the physiological monitor 17 of certain embodiments includes hardware, software, or both hardware and software, whether in one housing or multiple housings, used to receive and process the signals transmitted by the sensors 13.

As shown in FIG. 1B, the acoustic sensor assembly 13 can include a cable 25. The cable 25 can include three conductors within an electrical shielding. One conductor 26 can provide power to a physiological sensor 13, one conductor 28 can provide a ground signal from the physiological monitor 17, and one conductor 28 can transmit signals from the sensor 13 to the physiological monitor 17. For multiple sensors 13, one or possibly more cables 13 can be provided.

In some embodiments, the ground signal is an earth ground, but in other embodiments, the ground signal is a patient ground, sometimes referred to as a patient reference, a patient reference signal, a return, or a patient return. In some embodiments, the cable 25 carries two conductors within an electrical shielding layer, and the shielding layer acts as the ground conductor. Electrical interfaces 23 in the cable 25 can enable the cable to electrically connect to electrical interfaces 21 in a connector 20 of the physiological monitor 17. In another embodiment, the sensor assembly 13 and the physiological monitor 17 communicate wirelessly.

FIG. 1C illustrates an embodiment of a sensor system 100 including a sensor assembly 101 and a monitor cable 111 suitable for use with any of the physiological monitors shown in FIGS. 1A and 1B. The sensor assembly 101 includes a sensor 115, a cable assembly 117, and a connector 105. The sensor 115, in one embodiment, includes a sensor subassembly 102 and an attachment subassembly 104. The cable assembly 117 of one embodiment includes a sensor 107 and a patient anchor 103. A sensor connector subassembly 105 is connected to the sensor cable 107.

The sensor connector subassembly 105 can be removably attached to an instrument cable 111 via an instrument cable connector 109. The instrument cable 111 can be attached to a cable hub 120, which includes a port 121 for receiving a connector 112 of the instrument cable 111 and a second port 123 for receiving another cable. In certain embodiments, the second port 123 can receive a cable connected to an optical sensor or other sensor. In addition, the cable hub 120 could

include additional ports in other embodiments for receiving additional cables. The hub includes a cable 122 which terminates in a connector 124 adapted to connect to a physiological monitor (not shown).

The sensor connector subassembly 105 and connector 109 can be configured to allow the sensor connector 105 to be straightforwardly and efficiently joined with and detached from the connector 109. Embodiments of connectors having connection mechanisms that can be used for the connectors 105, 109 are described in U.S. patent application Ser. No. 12/248,856 (hereinafter referred to as "the '856 Application"), filed on Oct. 9, 2008, which is incorporated in its entirety by reference herein. For example, the sensor connector 105 could include a mating feature (not shown) which mates with a corresponding feature (not shown) on the connector 109. The mating feature can include a protrusion which engages in a snap fit with a recess on the connector 109. In certain embodiments, the sensor connector 105 can be detached via one hand operation, for example. Examples of connection mechanisms can be found specifically in paragraphs [0042], [0050], [0051], [0061]-[0068] and [0079], and with respect to FIGS. 8A-F, 13A-E, 19A-F, 23A-D and 24A-C of the '856 Application, for example.

The sensor connector subassembly 105 and connector 109 can reduce the amount of unshielded area in and generally provide enhanced shielding of the electrical connection between the sensor and monitor in certain embodiments. Examples of such shielding mechanisms are disclosed in the '856 Application in paragraphs [0043]-[0053], [0060] and with respect to FIGS. 9A-C, 11A-E, 13A-E, 14A-B, 15A-C, and 16A-E, for example.

In an embodiment, the acoustic sensor assembly 101 includes a sensing element, such as, for example, a piezoelectric device or other acoustic sensing device. The sensing element can generate a voltage that is responsive to vibrations generated by the patient, and the sensor can include circuitry to transmit the voltage generated by the sensing element to a processor for processing. In an embodiment, the acoustic sensor assembly 101 includes circuitry for detecting and transmitting information related to biological sounds to a physiological monitor. These biological sounds can include heart, breathing, and/or digestive system sounds, in addition to many other physiological phenomena. The acoustic sensor 115 in certain embodiments is a biological sound sensor, such as the sensors described herein. In some embodiments, the biological sound sensor is one of the sensors such as those described in the '883 Application. In other embodiments, the acoustic sensor 115 is a biological sound sensor such as those described in U.S. Pat. No. 6,661,161, which is incorporated by reference herein in its entirety. Other embodiments include other suitable acoustic sensors.

The attachment sub-assembly 104 includes first and second elongate portions 106, 108. The first and second elongate portions 106, 108 can include patient adhesive (e.g., in some embodiments, tape, glue, a suction device, etc.). The adhesive on the elongate portions 106, 108 can be used to secure the sensor subassembly 102 to a patient's skin. One or more elongate members 110 included in the first and/or second elongate portions 106, 108 can beneficially bias the sensor subassembly 102 in tension against the patient's skin and reduce stress on the connection between the patient adhesive and the skin. A removable backing can be provided with the patient adhesive to protect the adhesive surface prior to affixing to a patient's skin.

The sensor cable 107 can be electrically coupled to the sensor subassembly 102 via a printed circuit board ("PCB") (not shown) in the sensor subassembly 102. Through this

contact, electrical signals are communicated from the multiparameter sensor subassembly to the physiological monitor through the sensor cable **107** and the cable **111**.

In various embodiments, not all of the components illustrated in FIG. **1C** are included in the sensor system **100**. For example, in various embodiments, one or more of the patient anchor **103** and the attachment subassembly **104** are not included. In one embodiment, for example, a bandage or tape is used instead of the attachment subassembly **104** to attach the sensor subassembly **102** to the measurement site. Moreover, such bandages or tapes can be a variety of different shapes including generally elongate, circular and oval, for example. In addition, the cable hub **120** need not be included in certain embodiments. For example, multiple cables from different sensors could connect to a monitor directly without using the cable hub **120**.

Additional information relating to acoustic sensors compatible with embodiments described herein, including other embodiments of interfaces with the physiological monitor, are included in U.S. patent application Ser. No. 12/044,883, filed Mar. 7, 2008, entitled "Systems and Methods for Determining a Physiological Condition Using an Acoustic Monitor," (hereinafter referred to as "the '883 Application"), the disclosure of which is hereby incorporated by reference in its entirety. An example of an acoustic sensor that can be used with the embodiments described herein is disclosed in U.S. Patent Application No. 61/252,076, filed Oct. 15, 2009, titled "Acoustic Sensor Assembly," the disclosure of which is hereby incorporated by reference in its entirety.

FIG. **2** illustrates an embodiment of a multiparameter patient monitoring system **200**, which can implement any of the features described above. The multiparameter patient monitoring system **200** includes a multiparameter patient monitor **205** that receives signals from multiple physiological parameter measurement devices. The multiparameter patient monitor **205** can use the multiple received signals to determine a confidence value for respiratory rate measurements derived from the signals. The confidence value can advantageously reflect a degree to which the respiratory rate measurements derived from the different signals correspond. In addition, in some embodiments, the multiparameter patient monitor **205** can generate one or more respiratory rate outputs based at least partly on the multiple received signals.

The patient monitor **205** can include any of the features of the physiological monitor **17** described above. The patient monitor **205** can include one or more processors, a display, memory, one or more input/output (I/O) devices (such as input control buttons, speakers, etc), a wireless transceiver, a power supply, and/or processing and filtration circuitry. In certain embodiments, the patient monitor **205** can communicate with external devices, such as processing devices, output devices, mass storage devices, and the like. The patient monitor **205** can communicate with the external devices via a wired and/or wireless connection. The external devices can include a central monitoring station (such as a nurses' monitoring station), a server, a laptop computer, a cell phone, a smart phone, a personal digital assistant, a kiosk, other patient monitors, or other clinician devices. The patient monitor **205** can send physiological data to the external devices.

In the depicted embodiment, the patient monitor **205** is in communication with an acoustic sensor **210** and an optical sensor **215**. The acoustic sensor **210** can be a piezoelectric sensor or the like that obtains physiological information reflective of one or more respiratory parameters of a patient, including respiratory rate, expiratory flow, tidal volume, minute volume, apnea duration, breath sounds, riles, rhonchi, stridor, and changes in breath sounds, such as decreased vol-

ume or change in airflow. In addition, in some cases the acoustic sensor **210** can measure other physiological sounds, such as heart rate (e.g., to help with probe-off detection). In certain embodiments, the acoustic sensor **210** can include any of the features described in U.S. Patent Application No. 61/252,076, filed Oct. 15, 2009, titled "Acoustic Sensor Assembly," the disclosure of which is hereby incorporated by reference in its entirety.

The optical sensor **215** can include a noninvasive optical sensor that obtains physiological information reflective of one or more blood parameters of the patient. These parameters can include one or more of the following: a photoplethysmograph, oxygen saturation (SpO₂), HbCO, HbMet, FaO₂, fractional oxygen, total hemoglobin (Hbt), other hemoglobin species, carbon monoxide, carbon dioxide, pulse rate, perfusion index, pleth variability index, and optionally others, including concentrations or actual analyte values of the same. The optical sensor **215** can include one or more emitters capable of irradiating a tissue site (such as a finger) with one or more wavelengths of light, such as red and/or infrared (IR) wavelengths. In one embodiment, the optical sensor **215** is a pulse oximetry sensor. While many optical sensors emit two wavelengths, certain of the features described herein can be implemented by a photoplethysmograph sensor that emits a single wavelength. Further, the optical sensor **215** need not emit red or infrared wavelengths in certain embodiments but can also emit other wavelengths. The optical sensor **215** can also include one or more detectors capable of detecting the light after attenuation by pulsatile blood and tissue at the measurement site. The one or more detectors can generate a signal responsive to the attenuated light, which can be provided to the patient monitor **205**.

The patient monitor **205** can receive signals indicative of one or more physiological parameters from the acoustic sensor **210** and from the optical sensor **215**. The patient monitor **205** can extract and/or derive respiratory rate measurements from signals provided by both the acoustic sensor **210** and the optical sensor **215**. The patient monitor **205** can also output one or more respiratory rate measurements for display based at least in part on the received signals. Example techniques for deriving respiratory rate from the optical sensor measurements are described below with respect to FIG. **3**.

In certain embodiments, the patient monitor **205** can use pulse oximetry respiratory rate measurements to determine a multiparameter confidence in the acoustic respiratory rate measurements. The multiparameter confidence can be a value that reflects a degree of correspondence between the respiratory rate measurements obtained from the two sensors **210**, **215**. A close correspondence (e.g., small difference) between the two respiratory rate measurements can cause the patient monitor **205** to assign a higher multiparameter confidence to the acoustic respiratory rate measurement. Conversely, a larger difference between the two measurements can result in a lower multiparameter confidence. In certain embodiments, the patient monitor **205** can instead or also use the difference in respiratory rate values to assign a multiparameter confidence to the pulse-oximetry-derived respiratory rate measurement.

More generally, any comparative metric can be used to determine the multiparameter confidence. The comparative metric can be a difference between the measurements of the two sensors **210**, **215** but need not be. Instead, in some embodiments, the comparative metric can be a ratio between the measurements from the sensors **210**, **215**, a percentage derived from such a ratio, or the like. Such a ratio or percentage might be more meaningful than an absolute difference in some situations. Similarly, the comparative metric can be a

normalization of the measurements from the two sensors **210**, **215**, such as the following quotient: (the acoustic respiratory rate—the oximeter respiratory rate)/(the acoustic respiratory rate) or the like. Other comparative metrics can also be used.

Additionally, the patient monitor **205** can use pulse oximetry respiratory rate measurements to refine or adjust the acoustic respiratory rate measurements in some implementations. For example, the respiratory rate measurements derived from the two sensors **210**, **215** can be combined to form an overall respiratory measurement. The patient monitor **205** can average the two measurements, for example. The combined respiratory rate measurement can be more accurate than a respiratory rate measurement from either sensor **210**, **215** alone.

The patient monitor **205** can output the respiratory rate measurement derived from either or both of the acoustic and optical sensors **210**, **215**. In addition, the patient monitor **205** can output a multiparameter confidence indicator that reflects the calculated multiparameter confidence. Examples of multiparameter confidence indicators are described in greater detail below.

In certain embodiments, a signal received from the optical sensor **215** can be analyzed to determine a respiratory rate measurement. As an illustration of such a signal, FIG. **3A** depicts an example photoplethysmograph (pleth) waveform **300** derived from an optical sensor. The pleth waveform **300** can be derived from the received signal by the patient monitor **205**. The pleth waveform **300** is plotted on an intensity axis **301** versus a time axis **302**. The pleth waveform **300** has multiple pulses **312**, each with a peak **314** and a valley **316** and extending over a time period **318**. A curve extending along the peaks **314** of the pleth waveform **300** represents an envelope **322** of the pleth waveform **300**.

In certain embodiments, a respiratory rate measurement can be determined from an analysis of the pleth waveform **300**. A respiratory rate measurement can be determined from the pleth waveform **300** in the time domain and/or in the frequency domain. In certain embodiments, a respiratory rate measurement can be determined from the modulation in the amplitude of the pleth waveform **300**. For example, the time-varying frequency of the envelope **322** can correspond to the respiratory rate of the patient. The frequency of the pleth envelope **322** can be determined from the inverse of the period **324** of the envelope **322**. The envelope **322** of the pleth waveform **300** can be detected by an envelope detector. The envelope can be identified using an analog envelope detector such as a diode-based envelope detector or a digital detector employing such techniques as a Hilbert transform, squaring and low-pass filtering, or the like.

The respiratory rate can also be determined from a frequency analysis of the pleth waveform **300**. A frequency spectrum of the pleth waveform **300** can be generated, for example, by performing a Fast Fourier Transform (FFT) or other mathematical transform of the pleth waveform **300**. The respiratory rate can be identified by a peak in the spectrum (e.g., which corresponds to the frequency of the pleth envelope **322**). In certain embodiments, the peak can be identified by identifying the highest peak in a range of typical respiratory rates of a human patient. This range can differ for different patients based on factors such as age, gender, comorbidity, and the like. A respiratory rate value can be derived from the frequency of the selected peak. Additional methods of determining respiratory rate from the pleth waveform **300** and/or an optical signal are also possible.

In certain embodiments, instead of or in addition to analyzing the pleth waveform **300** to obtain respiratory rate, the patient monitor **205** can obtain respiratory rate from variability

detected in oxygen saturation measurements obtained from the optical sensor **215**. Variations in the oxygen saturation can track or approximately track the patient's respiratory cycle (e.g., a cycle of recruitment and collapse of alveoli), as is described in greater detail in U.S. Application No. 61/222,087, filed Jun. 30, 2009, titled "Pulse Oximetry System for Adjusting Medical Ventilation," the disclosure of which is hereby incorporated by reference in its entirety. The magnitude of the time-domain variations in the oxygen saturation can reflect the degree of recruitment and collapse of alveoli in the respiratory cycle. In the frequency domain, a peak in a magnitude response of the SpO₂ variability within an expected respiratory rate range can be used to determine a respiratory rate measurement.

In certain embodiments, the patient monitor **205** can obtain a respiratory rate measurement from variability detected in a patient's heart rate. The heart rate can be derived from an ECG signal, a bioimpedance signal, an acoustic signal, a plethysmograph signal, and/or combinations of the same.

In one embodiment, an instantaneous heart rate can be derived by determining the interval between successive R waves of the ECG signal and then converting the interval to beats per minute (bpm). For example, the heart rate can be calculated as **60** divided by the R-R interval in seconds. In another embodiment, the instantaneous heart rate can be derived from successive peaks in the plethysmograph signal. For example, the instantaneous heart rate can be calculated as **60** divided by the interval in seconds between the two successive peaks.

Other techniques can be used to derive the heart rate. For instance, the heart rate can be determined by analyzing any successive landmark of an ECG or plethysmograph signal. Further, to improve noise immunity, the patient monitor **205** can use a more robust technique to measure the interval, such as autocorrelation of the ECG or plethysmograph waveform from one beat to the next. More generally, any technique for reliably measuring the period from one beat to the next can be used.

The instantaneous heart rate can be plotted over time to illustrate variability in the patient's heart rate. In certain embodiments, the variability in the patient's heart rate is reflective of the patient's respiratory rate. For example, analysis of the variability in the instantaneous heart rate in the frequency domain (for example, by taking the Fourier transform of the instantaneous heart rate signal in the time domain) can provide an indication of respiratory rate that can be used to assess confidence in a respiratory rate measurement derived from an acoustic sensor or another type of sensor.

In certain embodiments, the patient monitor **205** can also obtain a respiratory rate measurement by measuring arterial pulse wave propagation time from the heart to an extremity. This propagation time is typically used by blood pressure monitoring systems and can be estimated by detecting a time difference between points on an ECG waveform and a photoplethysmograph waveform. This estimated propagation time is sometimes referred to as pulse wave transit time (PWTT) or time difference of arrival (TDOA). Currently available blood pressure monitoring systems trigger an automatic occlusive cuff to take a blood pressure measurement based on detected changes in PWTT.

Variability in the PWTT can be modulated by respiration. Thus, in certain embodiments, the patient monitor **205** can calculate PWTT and determine the variability in PWTT measurements over time. The patient monitor **205** can derive respiratory rate values from the calculated variability. The

patient monitor **205** can use these values to improve the accuracy of or calculate confidence in acoustically-derived respiratory values.

As illustrated in FIG. 3B, in one embodiment, PWTT is determined as a time difference between a peak of an R-wave **335** of a QRS complex of an ECG signal **330** to the foot point **345** of a plethysmograph signal **340**. The R-wave **335** represents the first upward, or positive, deflection of the QRS complex and corresponds to the time of ventricular depolarization. The foot point **345** of the plethysmograph signal **340** can correspond to the time of earliest onset of arrival of the pulse at a location away from the heart (e.g., at a patient's finger). More generally, PWTT can be taken as a time interval from any feature of the ECG waveform to any feature of the pleth waveform. For example, PWTT can be taken as the interval between the Q or S points of the ECG waveform and a point such as the midpoint of the pleth waveform.

The PWTT calculation can be improved by accounting for a patient's pre-ejection period (PEP). The PEP can include the difference in time between initiation of ventricular contraction (e.g., as detected by an ECG) and ejection of blood from the ventricles into the aorta. The PEP can also be considered as an interval between the onset of the QRS complex (of an electrocardiogram) and cardiac ejection. PWTT compensated for PEP can more accurately represent the propagation time of the arterial pulse from the heart to an extremity. In order to determine the PEP, in one embodiment an acoustic sensor is coupled with the patient to detect a patient's heart sound. The time difference between a feature of the ECG signal and a feature of the heart sound (represented as a signal) can be an estimate of PEP. In another embodiment, a bioimpedance sensor can be used to estimate PEP by taking a time difference between features of ECG and bioimpedance sensor signals. The arterial PWTT can then be calculated by subtracting the PEP from the initial PWTT calculation obtained from the ECG and plethysmograph signals. The patient monitor **205** can employ any of the systems or methods for determining PWTT and PEP described in more detail in U.S. Provisional Application No. 61/366,862, titled "System for Triggering A Non-Invasive Blood Pressure Device," filed Jul. 22, 2010, the disclosure of which is hereby incorporated by reference in its entirety.

In yet other embodiments, the PWTT is determined from a landmark of a first plethysmograph signal to a landmark of a second plethysmograph signal. In some embodiments, the first plethysmograph signal is acquired from a sensor applied to a finger of a patient and the second plethysmograph signal is acquired from a sensor applied to a toe of a patient; however other sensor locations can be used as desired and/or required.

The analysis of the heart rate and/or PWTT variability can include correlation in the time, frequency, or other transform domains. In one embodiment of a frequency domain analysis, FIG. 3C illustrates power spectrums **350A**, **350B** of the PWTT variability and the heart rate variability of a patient being monitored with the patient monitor **205**. The power spectrums **350A**, **350B** plot power amplitude (having an expanded scale) versus frequency. In one embodiment, the respiratory rate measurement is determined from the power spectrums **350A**, **350B** by the highest spectral peak in the frequency range corresponding to the normal range of respiratory rates. The respiratory peak **355** of the power spectrums **350A**, **350B** is approximately 0.3 Hz, which corresponds to a respiratory rate of approximately 18 breaths per minute. This is an example frequency value that can vary for different patients or even for the same patient over time.

The respiratory rate measurement derived from the PWTT variability and the respiratory rate measurement from the

heart rate variability can be compared with each other and/or with other respiratory rate measurements to determine an overall respiratory rate measurement or to assess confidence in a respiratory rate measurement derived from another physiological signal, as described in further detail below.

In certain embodiments, the PWTT and/or heart rate variability data can be smoothed or otherwise filtered by various signal processing methods, such as moving average smoothing, sliding average smoothing, box smoothing, binomial (Gaussian) smoothing, polynomial smoothing, and/or the like, to improve the accuracy of, or confidence in, the respiratory rate measurements.

FIG. 4 illustrates an embodiment of a multiparameter patient monitoring system **400** coupled to a patient **401**. The multiparameter patient monitoring system **400** includes a patient monitor **405**, an acoustic sensor **410**, and an optical sensor **415**. The acoustic sensor **410** and the optical sensor **415** can obtain physiological signals from the patient **401** and transmit the signals to the patient monitor **405** through cables **403A**, **403B**.

As shown, the acoustic sensor **410** is attached to the skin of the patient **401** on the neck near the trachea. The acoustic sensor **410** can include adhesive elements (e.g., tape, glue, or the like) to secure the acoustic sensor **410** to the skin. The acoustic sensor **410** can additionally be secured to the patient using an anchor **408**, which can be affixed near a subclavian region of the patient **401** or at other regions. The anchor **408** can reduce stress on the connection between the acoustic sensor **410** and the skin during movement. Other placement locations for the acoustic sensor **410** and the patient anchor **408** are also possible, such as other parts of the neck, the chest, or the like.

The optical sensor **415** can be removably attached to the finger of the patient **401**. In other embodiments, the optical sensor **415** can be attached to a toe, foot, and/or ear of the patient **401**. The optical sensor **415** can include a reusable clip-type sensor, a disposable adhesive-type sensor, a combination sensor having reusable and disposable components, or the like. Moreover, the optical sensor **415** can also include mechanical structures, adhesive or other tape structures, Vello™ wraps or combination structures specialized for the type of patient, type of monitoring, type of monitor, or the like.

In certain embodiments, the various sensors and/or monitors can communicate with the patient monitor **405** wirelessly. The wireless communication can employ any of a variety of wireless technologies, such as Wi-Fi (802.11x), Bluetooth, cellular telephony, infrared, RFID, combinations of the same, and the like.

In certain embodiments, the multiparameter patient monitoring system **200** can include additional physiological parameter measurement devices. FIG. 5 illustrates an example of a multiparameter respiratory monitoring system **500** that includes multiple additional measurement devices. In particular, a patient monitor **505** receives inputs from an acoustic sensor **510**, an optical sensor **515**, an electrocardiograph (ECG) **520**, a capnograph **525**, a bioimpedance monitor **530**, and possibly other physiological monitors or sensors **535**.

In certain embodiments, the multiparameter patient monitor **505** derives respiratory rate measurements from signals received from each of the depicted physiological parameter measurement devices and/or sensors. In certain embodiments, the respiratory rate measurements derived from one or more of the optical sensor **515**, the ECG **520**, the capnograph **525**, and/or the bioimpedance monitor **530** can be compared with the respiratory rate measurement from the acoustic sen-

sor **510**. The monitor **505** can compare one or more of these measurements with the acoustically-derived measurement in order to derive a multiparameter confidence value reflecting a confidence in the acoustic respiratory rate measurement (or confidence in any other of the respiratory rate measurements).

In other embodiments, one or more of the respiratory rate measurements from the ECG **520**, the capnograph **525** and the bioimpedance monitor **530** can be combined with the respiratory rate measurements from the acoustic sensor **510** and/or the optical sensor **515** to generate a combined respiratory rate output. In certain embodiments, the combined respiratory rate output can have greater accuracy than the respiratory rate measurement obtained from any one of the devices shown.

The ECG **520** can monitor electrical signals generated by the cardiac system of a patient. The ECG **520** can include one or more sensors adapted to be attached to the skin of a patient, which can be used to detect electrical heart activity of the patient. The ECG **520** can determine any of a variety of electrical physiological parameters based upon electrical signals received from the one or more sensors, such as heart rate. In certain embodiments, the ECG **520** can generate an electrocardiogram waveform. The patient monitor **505** can compare one or more features of the waveform with an acoustically-derived respiratory rate measurement to determine multiparameter confidence in the acoustically-derived respiratory rate. For instance, the R-R time period of the ECG waveform, or the like can be correlated with respiratory rate in certain individuals. More generally, an envelope of the ECG waveform can include peaks that the patient monitor **505** can correlate in frequency with respiratory rate in certain situations.

The capnograph **525** can determine the carbon dioxide content in inspired and/or expired air from a patient. For example, the capnograph **525** can monitor the inhaled and/or exhaled concentration or partial pressure of carbon dioxide through a breathing mask or nasal cannula. In certain embodiments, the capnograph **525** can generate a capnogram responsive to the patient's breathing. The capnograph **525** can also identify end tidal carbon dioxide (EtCO_2) levels and/or other values. From the EtCO_2 values, the capnograph **525** can determine a respiratory rate of the patient. The capnograph **525** can provide this respiratory rate measurement to the patient monitor **505**, which can compare the respiratory rate with the acoustically-derived respiratory rate to determine multiparameter confidence.

The bioimpedance monitor **530** can determine electrical impedance or resistance in body tissue of a medical patient. For example, the bioimpedance monitor **530** can include two or more sensors or electrodes positioned on a patient so as to measure the bioelectrical impedance or resistance across the chest region. The measured bioelectrical impedance can vary as a result of the expansion of the chest due to breathing, and from this variance, a respiratory rate measurement can be derived. In certain embodiments, the bioimpedance monitor **530** is a Transthoracic Impedance Monitor or the like, having two or more electrodes that can optionally be combined with ECG electrodes. In other embodiments, the bioimpedance monitor **530** is an impedance tomograph, having many more electrodes that can also be used to form a spatial image of the impedance variation.

The respiratory rate measurement can be derived by the bioimpedance monitor **530**, or alternatively, the bioimpedance monitor **530** can provide impedance values with respect to time to the patient monitor **505**, which can derive the respiratory rate. The patient monitor **505** can also compare

the impedance-derived respiratory rate with the acoustically-derived respiratory rate to determine multiparameter confidence.

Additional sensors and/or monitors of different types can also be included. The other patient monitors **135** can include, for example, thermistor-based breathing sensors or pneumatic breathing belt sensors.

In certain embodiments, the electrocardiograph **520**, the capnograph/capnometer **525**, and the bioimpedance monitor **530** are standalone patient monitors that can provide filtered and/or processed signals to the patient monitor **505**. In other embodiments, the electrocardiograph **520**, the capnograph **525**, and the bioimpedance monitor **530** can be replaced with respective sensors, which each provide physiological data directly to the patient monitor **505**. In still other embodiments, the acoustic sensor **510** and the optical sensor **515** can be replaced with an acoustic respiratory monitor and a pulse oximeter, respectively. Thus, any combination of sensors and monitors can provide inputs to the patient monitor **505**, including any subset of the devices shown.

The patient monitor **505** can output for display the respiratory rate value derived from the acoustic sensor **550**. In addition, the patient monitor **505** can output respiratory rate values derived from any of the other devices shown.

In certain embodiments, the respiratory rate measurements derived from one or more of the sensors can be used for sequential hypothesis testing.

FIGS. **6A** through **6C** illustrate embodiments of systems **600A**, **600B**, and **600C** for determining multiparameter confidence of respiratory rate measurements and for outputting respiratory rate values. The systems **600A**, **600B**, and **600C** can be implemented by any of the patient monitors described herein, such as the patient monitors **205**, **405**, and **505**, or by the patient monitors described below. Each of the depicted blocks of the systems **600A**, **600B**, and **600C** can be implemented by hardware and/or software.

Referring to FIG. **6A**, the system **600A** receives signal inputs reflective of physiological parameters from an acoustic sensor and from an optical sensor, such as any of the sensors described above. The signal inputs can be received by respiratory rate determination blocks **640a**, **640b**, respectively. Each of the respiratory rate determination blocks **640** can determine a respiratory rate based at least in part on its respective signal input. For example, the respiratory rate determination block **640b** can determine respiratory rate of a patient from a time domain or frequency analysis of a photopleth input signal.

In certain embodiments, the respiratory rate determination block **640b** can be part of any of the patient monitors described above. Thus, for example, an optical sensor could provide the photopleth signal to the respiratory rate determination block **640b** of a patient monitor, which derives a respiratory rate. The respiratory rate determination block **640b** could instead be part of a pulse oximetry monitor. The pulse oximetry monitor could determine a respiratory rate measurement based at least in part on the photopleth signal. The pulse oximetry monitor could provide the calculated respiratory rate to the patient monitor (e.g., **205**, **405**, **505**, or the like).

For convenience, the acoustic respiratory rate measurement will be described using the shorthand RR_{AR} and the photopleth respiratory rate measurement will be described using the shorthand RR_{PO} . In the depicted embodiment, the RR_{AR} and the RR_{PO} measurements are provided to a respiratory rate analyzer **645**. The respiratory rate analyzer **645** can analyze the RR_{AR} and the RR_{PO} measurements to determine a multiparameter confidence in the RR_{AR} measurement. For example, the respiratory rate analyzer **645** can compare the

two measurements to determine a difference between the two measurements. The respiratory rate analyzer can derive a multiparameter confidence or multiparameter confidence value from this calculated difference. In certain embodiments, the greater the difference between the RR_{AR} and the RR_{PO} measurements, the lower is the multiparameter confidence determined for the RR_{AR} measurement. Conversely, in certain embodiments, the respiratory rate analyzer **645** can use the difference between the two measurements to assign a multiparameter confidence to the RR_{PO} measurement.

The respiratory rate analyzer **645** can output for display a multiparameter confidence indicator **660** responsive to the calculated multiparameter confidence along an output respiratory rate measurement (RR_{OUT} , described below). The multiparameter confidence indicator **660** can include a visual and/or audible indication in various embodiments.

Moreover, in certain embodiments, the respiratory rate analyzer **645** can generate the respiratory rate output RR_{OUT} based on a combination of the inputs RR_{AR} and RR_{PO} . For example, the respiratory rate analyzer **645** could average the two respiratory rate inputs. This average could be a weighted average or the like (see, e.g., FIG. 6C).

In another embodiment, the respiratory rate analyzer selects one of the respiratory rate inputs (RR_{AR} and RR_{PO}) to output as the respiratory rate output RR_{OUT} . The respiratory rate analyzer **645** could make this selection based at least partly on single parameter confidence values generated by each respiratory rate determination block **640a**, **640b**. These single parameter confidence values can reflect a quality of the signal received by each block **640a**, **640b**. Single parameter confidence values can be distinguished from multiparameter confidence values, in certain embodiments, in that single parameter confidence values can reflect confidence that a respiratory rate derived from a single parameter is accurate. In contrast, multiparameter confidence values can reflect respiratory rate accuracy as determined by an analysis of multiple parameters (e.g., photopleth and ECG).

For example, the respiratory rate determination block **640b** could determine single parameter confidence of the photopleth signal using techniques such as those described in U.S. Pat. No. 6,996,427, titled "Pulse Oximetry Data Confidence Indicator," filed Dec. 18, 2003, (the "'427 patent") the disclosure of which is hereby incorporated by reference in its entirety. Analogous techniques could be used by the respiratory rate determination block **640a** to determine single parameter confidence in the quality of the acoustic respiratory signal received.

The respiratory rate analyzer **645** could select either the RR_{AR} respiratory rate value or the RR_{PO} respiratory rate value to provide as the respiratory rate output RR_{OUT} based on, for example, which signal has a higher calculated signal quality. In another embodiment, the respiratory rate analyzer **645** could weight a combination of the two respiratory rate values based at least in part on the single parameter confidence values. In various embodiments, the respiratory rate analyzer **645** can also select the respiratory rate value to output based on patient-specific factors, such as age, gender, comorbidity, and the like. For instance, for some patients, one respiratory rate measurement derived from a particular parameter might be more reliable than other respiratory rate measurements derived from other parameters. Many other variations are also possible.

Although the respiratory rate analyzer **645** has been described as being able to average respiratory rate values or select respiratory rate values, the distinction between averaging and selecting can blur. Selecting, for instance, can be considered a subset of weighting where respiratory rate val-

ues selected are given a weight of "1" (or substantially 1) and respiratory rate values not selected are given a weight of "0" (or substantially 0).

FIG. 6B extends the embodiment shown in FIG. 6A to include additional parameter inputs. In FIG. 6B, the system **600B** receives an acoustic respiratory signal, a photopleth signal, an ECG signal, a capnograph signal, and a bioimpedance signal. Signal inputs from other types of sensors and/or monitors, or additional sensors of the types listed, can also be received. The respective signal inputs are received by respiratory rate determination blocks **640a**, **640b**, **640c**, **640d**, and **640e**. As described above, the respiratory rate determination blocks **640a**, **640b** can determine a respiratory rate measurement using any of the techniques described above and optionally a single parameter confidence value based at least in part on its respective signal input. Likewise, the respiratory rate determination blocks **640c**, **640d**, and **640e** can calculate respiratory rate measurements and optionally single parameter confidence values.

Signal inputs can also be used to determine respiratory rate measurements derived from heart rate variability and/or PWTT variability. As shown in FIG. 6B, the signal inputs (e.g., an acoustic respiratory signal, a photopleth signal, an ECG signal, a bioimpedance signal and/or other signals) are received by a heart rate determination block **670** and a PWTT determination block **675**. In other embodiments, more or fewer signal inputs can be received by the heart rate determination block **670** and/or the PWTT determination block **675**. The heart rate determination block **670** can derive the patient's heart rate from one or more of the signal inputs. The PWTT determination block **675** can determine the patient's PWTT from one or more of the signal inputs using any of the techniques described above with respect to FIG. 3B.

The respiratory rate determination blocks **640f** and **640g** can determine respiratory rate measurements based at least in part on an analysis of the heart rate variability and the PWTT variability, respectively, of the patient, using any of the techniques described above. For example, the respiratory rate determination blocks **640f** and **640g** can determine respiratory rate measurements from a frequency analysis of heart rate and/or PWTT signals over time. The respiratory rate measurements calculated by the respiratory rate determination blocks **640f** and **640g** can be provided to the respiratory rate analyzer **650** along with any of the respiratory rate measurements calculated by the respiratory rate determination blocks **640a**, **640b**, **640c**, **640d** and **640e**. The respiratory rate determination blocks **640f** and **640g** can also calculate single or multiple parameter confidence values.

The respiratory rate determination blocks **640** can be implemented in any of the patient monitors **205**, **405**, **505**, etc. described herein. Thus, for example, a patient monitor can receive sensor inputs from one or more of an acoustic sensor, an optical sensor, an ECG sensor or sensors, a capnometry sensor, and a bioimpedance sensor. Not all of the inputs shown need be received by a patient monitor; rather, a subset can be received by any patient monitor. From the inputs, the patient monitor implementing the respiratory rate determination blocks **640** can calculate individual respiratory rate measurements corresponding to each input, using any of the techniques described above. The patient monitor can further implement the respiratory rate determination blocks **640** by calculating single parameter confidence in each block in an analogous manner to that described in the '427 patent incorporated by reference above. In another embodiment, the respiratory rate calculation for certain of the parameters is performed in a separate monitor. For instance, a capnograph monitor can determine a respiratory rate of a patient and

provide this respiratory rate value to a respiratory rate analyzer **650** of the patient monitor.

The respiratory rate determination blocks **640** can provide respiratory rate values and optionally single parameter confidence values to a respiratory rate analyzer **650**. The respiratory rate analyzer **650** can operate in a similar manner to the respiratory rate analyzer **645** described above. For instance, the respiratory rate analyzer **650** can analyze one or more of the respiratory rate measurements to determine a multiparameter confidence in the RR_{ARM} measurement, using any of the techniques described above.

In one embodiment, the respiratory rate analyzer **650** determines multiparameter confidence by comparing the RR_{AR} measurement to one or more of the other respiratory rate measurements. The multiparameter confidence calculated by the respiratory rate analyzer **650** can reflect the differences between the measurements. For example, the respiratory rate analyzer **650** can average the differences to generate a multiparameter confidence value, use a weighted average of the differences to generate a multiparameter confidence value, can select the greatest difference as the multiparameter confidence value, can use any of the above to further derive a multiparameter confidence value (e.g., by looking up the difference value in a look-up table to obtain a corresponding multiparameter confidence value, or by multiplying the difference value by a scalar to obtain a multiparameter confidence value), or by a host of other techniques. Moreover, in certain embodiments, the respiratory rate analyzer **650** can analyze any subset of the respiratory rate measurements received to determine a multiparameter confidence in any given one of the respiratory rate measurements.

The respiratory rate analyzer **650** can output for display a multiparameter confidence indicator **660** responsive to the calculated multiparameter confidence along an output respiratory rate measurement (RR_{OUT} , described below). The multiparameter confidence indicator **660** can include a visual and/or audible indication in various embodiments. The multiparameter confidence indicator **660** can be output to a display **655** along with a respiratory rate output RR_{OUT} .

Moreover, like the respiratory rate analyzer **645** described above, the respiratory rate analyzer **650** can generate the respiratory rate output RR_{OUT} based on a combination or selection of any of the respiratory rate inputs received from the various sensors or monitors. For example, the respiratory rate output RR_{OUT} can be the acoustic respiratory rate (RR_{AR}), or a selected one of the other respiratory rate measurements. Or, the respiratory rate analyzer **650** could average, perform a weighted average (e.g., based on respective single parameter confidences), or otherwise combine the respiratory rate measurements to determine the respiratory rate output RR_{OUT} . In various embodiments, the respiratory rate analyzer **645** can also select and/or combine the respiratory rate values to determine an output based on patient-specific factors, such as age, gender, comorbidity, and the like. For instance, for some patients, one respiratory rate measurement derived from a particular parameter might be more reliable than other respiratory rate measurements derived from other parameters. Many other variations are also possible.

In other embodiments, the combiner/selector module **650** can compare the derived respiratory rate measurements to determine, which, if any, of the respiratory rate determination blocks **640** provided outliers. The combiner/selector module **650** could reject the outliers and combine (e.g., average) the outputs of the remaining respiratory rate determination blocks **640**.

In yet other embodiments, the combiner/selector module **650** could determine which of the outputs from the respira-

tory rate determination blocks **640** are close to each other (e.g., within a tolerance) and output a combination of those outputs. For example, if three of the five respiratory rate determination blocks **640** produce a similar output and two are outliers, the combiner/selector module **650** could average the three similar outputs or select one of the three outputs as the final RR_{OUT} measurement. Moreover, the combiner/selector module **650** can learn over time and can select the output derived from one of the sensors or monitors based on past performance. Many other configurations and extensions of the combiner/selector module **650** are possible.

In certain embodiments, the respiratory rate output measurements and/or the multiparameter confidence values can be output to an external device over a network, instead of, or in addition to, being output to the display **655**. For example, the output data can be output to a central monitoring station (such as a nurses' monitoring station), a server, a laptop computer, a cell phone, a smart phone, a personal digital assistant, other patient monitors, or other clinician devices, for example. In some embodiments, the patient monitor **505** can transmit data to an external device via a wireless network using a variety of wireless technologies, such as Wi-Fi (802.11x), Bluetooth, cellular telephony, infrared, RFID, combinations of the same, and the like.

FIG. 6C illustrates yet another embodiment of a system **600C** for calculating multiparameter confidence in respiratory rate measurements. In the system **600C**, acoustic and photopleth signal inputs are provided, as well as optionally any number of other signal inputs (such as any of the inputs described above). As above, respiratory rate determination blocks **640a**, **640b**, and so forth down to **640n** can receive these signal inputs. The respiratory rate determination blocks **640a**, **640b**, . . . , **640n** can calculate respiratory rate values based on the signal inputs, as well as associated internal confidence values. Each of the internal confidence values can reflect an individual respiratory rate block **640** algorithm's confidence in the respiratory rate measurements.

A respiratory rate analyzer **660** receives the respiratory rate and confidence measurements **642** calculated by the respiratory rate determination blocks **640**. The respiratory rate analyzer **660** can have some or all the features of the respiratory rate analyzers described above. In addition, the respiratory rate analyzer **660** can use the internal confidence values calculated by the respiratory rate blocks **640** to weight, select, or otherwise determine appropriate overall respiratory rate and confidence values **652**. The respiratory rate analyzer **660** outputs these values **652** to a display **655** or to some other device.

The respiratory rate analyzer **660** can use any of a variety of techniques to calculate the overall respiratory rate and confidence values **652**. Some example techniques are described herein. To illustrate, in one embodiment, the respiratory rate analyzer **660** can perform a weighted average of the respiratory rate values from each respiratory rate determination block **640**. The weights can be derived from, or can be, their respective confidence values.

More complex weighting schemes can also be devised. One example weighting algorithm can implement an adaptive algorithm for dynamically adjusting the weights applied to each respiratory rate value over time. The weights can be adapted based on minimizing some cost function, such as may be applied by a Kalman filter, for instance. More generally, any of a variety of adaptive algorithms may be used to adjust the weights. For example, the respiratory rate analyzer **660** can implement one or more of the following: a least mean squares algorithm (LMS), a least squares algorithm, a recursive least squares (RLS) algorithm, wavelet analysis, a joint

process estimator, an adaptive joint process estimator, a least-squares lattice joint process estimator, a least-squares lattice predictor, a correlation canceller, optimized or frequency domain implementations of any of the above, any other linear predictor, combinations of the same, and the like.

In another embodiment, the respiratory rate analyzer **660** can select the top N available sources having the highest confidence level, where N is an integer. For instance, the respiratory rate analyzer **660** can choose the output of N respiratory rate determination blocks **640** having confidence values that exceed a threshold. This threshold may be determined relative to the confidence values provided (e.g., via a ratio or the like) or can be an absolute threshold. The respiratory rate analyzer **660** can then perform a weighted average of the remaining values or select from these values, for example, based on confidence values.

Internal confidence of each respiratory rate determination block **640** can depend on a variety of factors, such as signal to noise ratio, irregularities in the data, probe-off conditions, and the like. A probe off condition, for instance, can result in a zero confidence value, a gradual taper down to zero confidence over time, or the like. Likewise, the confidence values can be derived from the signal to noise ratio for each respiratory rate determination block **640**.

FIG. 7 illustrates an example noninvasive multiparameter physiological monitor **700** that can implement any of the features described herein. An embodiment of the monitor **700** includes a display **701** showing data for multiple physiological parameters. For example, the display **701** can include a CRT or an LCD display including circuitry similar to that available on physiological monitors commercially available from Masimo Corporation of Irvine, Calif. sold under the name Radical™, and disclosed in U.S. Pat. Nos. 7,221,971; 7,215,986; 7,215,984 and 6,850,787, for example, the disclosures of which are hereby incorporated by reference in their entirety. However, many other display components can be used that are capable of displaying respiratory rate and other physiological parameter data along with the ability to display graphical data such as plethysmographs, respiratory waveforms, trend graphs or traces, and the like.

The depicted embodiment of the display **701** includes a measured value of respiratory rate **712** (in breaths per minute (bpm)) and a respiratory rate waveform graph **706**. In addition, other measured blood constituents shown include SpO₂ **702**, a pulse rate **704** in beats per minute (BPM), and a perfusion index **708**. Many other blood constituents or other physiological parameters can be measured and displayed by the multiparameter physiological monitor **700**, such as blood pressure, ECG readings, EtCO₂ values, bioimpedance values, and the like. In some embodiments, multiple respiratory rates, corresponding to the multiple input sensors and/or monitors, can be displayed.

FIGS. 8A through 8C illustrate example multiparameter physiological monitor displays **801A-801C** that output multiparameter confidence indicators **814**. The multiparameter confidence indicators **814** can be generated using any of the techniques described above.

Referring to FIG. 8A, an example display **801A** is shown that includes parameter data for respiratory rate, including a measured respiratory rate value **812** in breaths per minute (bpm) and a respiratory waveform graph **806**. The display **801A** also includes parameter data for SpO₂ **802** and pulse rate **804** in beats per minute (BPM). A respiratory rate multiparameter confidence indicator **814A** is also depicted. In the depicted embodiment, the multiparameter confidence indicator **814A** includes text that indicates that the current respiratory rate has a low multiparameter confidence level. The

multiparameter confidence indicator **814A** can function as a visual multiparameter confidence-based alarm by flashing, changing color, or the like when the multiparameter confidence is below a threshold level. The multiparameter confidence indicator can include symbols other than (or in addition to) text in certain embodiments. An audible multiparameter confidence-based alarm can alternatively, or additionally, be output through a speaker or other audio output device. A multiparameter confidence-based alarm can be generated as described in the '427 patent described above.

In certain embodiments, an alarm can be output when the monitored respiratory rate of the patient deviates beyond a patient-specific and/or patient-independent threshold. The utility and effectiveness of an alarm based on a respiratory rate measurement determined solely from an acoustic signal can be improved by joint processing of ancillary signals from multiple monitored physiological parameters, such as those described herein (e.g., electrical signals, photoplethysmographic signals, bioimpedance signals, and/or the like).

For example, respiratory rate measurements determined from the ancillary signals can be used to continuously or periodically refine or assess confidence in the respiratory rate measurements derived from the acoustic signal. If the multiparameter confidence in the acoustic respiratory rate measurement is low, the alarm can be suppressed, at least pending further consideration; however, if the multiparameter confidence in the acoustic respiratory rate measurement is sufficiently high, the alarm can be output without further consideration.

In other embodiments, the ancillary signals can be used to estimate the initial respiratory rate or timing information to assist an acoustic signal processing algorithm in capturing a respiratory component of the acoustic signal. The use of the ancillary signals from multiple parameters to assist in the capturing of the respiratory component of the acoustic signal can lead to increased confidence in the accuracy of the respiratory rate measurement, thereby increasing the accuracy, reliability, and effectiveness of the alarm based on the respiratory rate measurement from the acoustic signal.

The display **801B** of FIG. 8B includes the same parameter data as the display **801A**. However, the display **801B** includes a multiparameter confidence indicator **814B** that indicates the current multiparameter confidence level numerically, rather than textually (displayed as a percentage in the depicted embodiment). A present multiparameter confidence of 90% is shown by the multiparameter confidence indicator **814B**.

The display **800C** of FIG. 8C includes a respiratory rate trend graph **816**, which depicts respiratory rate measurements over a period of time. The display **801C** also depicts a multiparameter confidence indicator **814C** in the form of a bar graph below the trend graph **816**. The multiparameter confidence indicator **814C** includes bars that can correspond to occurrences of breaths of a patient. The bars can have a height that corresponds to a degree of multiparameter confidence in the respiratory rate measurements for any given breath. As the breaths change over time, the multiparameter confidence can also change over time, resulting in a changing multiparameter confidence indicator **814C**. The multiparameter confidence indicator **814C** can be generated using analogous techniques to those described in the '427 patent described above.

In certain embodiments, the bars can all be depicted with the same color and/or pattern or with varying colors and/or patterns depending on the multiparameter confidence level. For example, bars within a desired multiparameter confidence range can be displayed with a first color and/or pattern, bars within a tolerable multiparameter confidence range can be displayed with a second color and/or pattern, and bars

within a low multiparameter confidence range can be displayed with a third color and/or pattern. In certain embodiments, the bars can be replaced with pulses, lines, or other shapes.

FIG. 8D illustrates another example multiparameter physiological monitor **800** having a display **801D**. As shown, the physiological monitor **800** can be configured with a vertical display instead of a horizontal display. The display **801D** can include similar parameter data as shown in displays **801A-801C**. The multiparameter physiological monitor **800** includes a multiparameter confidence indicator **814D** that is positioned off the display **801D**. The multiparameter confidence indicator **814D** can include one or more light emitting diodes (LEDs) positioned adjacent to text, such as "LOW RR CONF" or "LOW SQ" or "LOW SIQT™," where SQ and SIQ stand for signal quality and signal intelligence quotient, respectively. The multiparameter confidence indicator **810D** can be activated to inform a caregiver that a measured value of the multiparameter confidence of the incoming signal is below a certain threshold, for instance. In certain embodiments, different colored LEDs can be used to represent different multiparameter confidence range levels, such as in the manner described above.

The example displays **801A-801D** in FIGS. 8A-8D are merely illustrative examples. Many other variations and combinations of multiparameter confidence indicators **814** are also possible in other implementations without departing from the spirit and/or scope of the disclosure.

Moreover, in certain embodiments, the features described in U.S. Pat. No. 6,129,675, filed Sep. 11, 1998 and issued Oct. 10, 2000 and in U.S. patent application Ser. No. 11/899,512, filed Sep. 6, 2007, titled "Devices and Methods for Measuring Pulsus Paradoxus," each of which is hereby incorporated by reference in its entirety, can be used in combination with the features described in the embodiments herein.

FIG. 9 illustrates an embodiment of a patient monitoring process **900** in which a user (e.g., a clinician) has the ability to specify a delay time for an alarm to be triggered. In one implementation, the patient monitoring process **900** is performed by any of the patient monitoring systems (e.g., systems **10**, **200**, **400**, **500**, **600a**, **600b**) and/or the patient monitors (e.g., monitors **205**, **405**, **505**, **700**, **800**) described above. More generally, the patient monitoring process **900** can be implemented by a machine having one or more processors. Advantageously, in certain embodiments, the patient monitoring process **900** provides a user-customizable alarm delay that can reduce nuisance alarms.

Currently available patient monitoring devices often generate alarms prematurely or generate alarms that may not correspond to a clinically significant event. For example, a monitoring device can generate an alarm even though the patient's physiological state or condition does not warrant attention. Instead of providing useful, actionable information, these "nuisance" alarms can result in unnecessary worry or stress of the patient and/or clinician and wasted time on the part of the clinician in responding to the nuisance alarms. The patient monitoring process **900** can advantageously reduce or suppress the number of nuisance alarms by providing an alarm delay period. The alarm delay period can advantageously be adjusted by a user.

The patient monitoring process **900** begins by receiving user input of an alarm delay time at block **902**. For example, a user such as a clinician can select a desired alarm delay by inputting the desired delay time into a physiological monitor via a user interface, a numerical keypad, or the like. The alarm delay time can correspond to a particular physiological parameter to be monitored. The physiological parameter can

include, for example, blood pressure, respiratory rate, oxygen saturation (SpO₂) level, other blood constitutions and combinations of constitutions, and pulse, among others. The input from the clinician can adjust a default alarm delay. For example, the default alarm delay time might be 15 seconds, and the clinician input can change the alarm delay time to 30 seconds.

At block **904**, the user-specified alarm delay time is stored in a memory device. At block **906**, the physiological parameter corresponding to the user-specified alarm delay time is monitored by a patient monitor of a patient monitoring system. At decision block **908**, it is determined whether a value of the monitored physiological parameter has remained past a threshold (e.g., above or below a threshold or thresholds) for the user-specified alarm delay time. If it is determined that the value of the monitored physiological parameter has passed a threshold for the time period of the user-specified alarm delay, an alarm is output at block **910**. If, however, it is determined that the value of the monitored physiological parameter has not remained past the threshold for the time period of the user-specified alarm delay, the patient monitoring process **900** loops back to block **906** to continue monitoring. In various implementations, the threshold can be set or adjusted by a user (e.g., a clinician) depending on patient-specific factors (e.g., age, gender, comorbidity, or the like).

The alarm can be provided as a visual and/or audible alarm. In one embodiment, the alarm is output by a patient monitor. In another embodiment, the patient monitor transmits the alarm to another device, such as a computer at a central nurses' station, a clinician's end user device (e.g., a pod, a pager), or the like, which can be located in a hospital or at a remote location. The patient monitor can transmit the alarm over a network, such as a LAN, a WAN, or the Internet.

As one example, a user can set an alarm delay time for a respiratory rate to be sixty seconds. In certain situations, a respiratory rate that is outside a threshold range of values for less than sixty seconds can be considered an apnea event. Accordingly, an alarm generated before the respiratory rate has remained outside the threshold range of values for a time period of more than sixty seconds may not be desirable or provide useful information for a clinician to act on. In one embodiment, the patient monitor can monitor the respiratory rate by receiving signals from an acoustic sensor, such as any of the acoustic sensors described herein. When the patient monitor determines that the respiratory rate has been outside a threshold range of values for at least sixty seconds, then an alarm can be output by the patient monitor.

In certain embodiments, an indication can be provided to a user (e.g., a clinician) regarding a current status of the alarm delay period. The indication can be audible and/or visual. In one embodiment, a confidence indicator can be altered or modified based on the alarm delay period. For example, the confidence indicator can be modified to reflect a "countdown" to the time of triggering of the alarm. If the confidence indicator is represented by an LED, for example, the LED can blink once the alarm delay has been initiated and can blink faster as the trigger time of the alarm grows closer. If the confidence indicator is represented by a bar graph, for example, the bars can be modified during the period from initiation of the alarm delay until the time of triggering of the alarm. A separate countdown timer that is not coupled with the confidence indicator could also be provided, which counts down seconds remaining in the alarm delay period.

FIG. 10 illustrates an embodiment of a multiparameter patient monitoring process **1000**. In the multiparameter patient monitoring process **1000**, an alarm delay time for a first physiological parameter can be modified dynamically

based on a measurement of a second physiological parameter. In one implementation, the patient monitoring process 1000 is performed by any of the patient monitoring systems (e.g., systems 10, 200, 400, 500, 600a, 600b) and/or the patient monitors (e.g., monitors 205, 405, 505, 700, 800) described above. More generally, the patient monitoring process 1000 can be implemented by a machine having one or more processors.

At block 1006, first and second physiological parameters are monitored by a multiparameter patient monitor. In one embodiment, respiratory rate and SpO₂ are the two monitored physiological parameters. In other embodiments, the first and second monitored physiological parameters can include, for example, blood pressure, respiratory rate, oxygen saturation (SpO₂) level, other blood constituents and combinations of constituents, and pulse, among others.

At decision block 1008, it is determined whether the current monitored value of the first physiological parameter has passed a threshold. If so, then at decision block 1010, it is determined whether an alarm delay time corresponding to the first parameter has been reached. The alarm delay time can be a default alarm delay time or a user-selected delay time, such as the user-selected delay time described above with respect to FIG. 9. If the first physiological parameter has not passed the threshold, the multiparameter patient monitoring process 1000 loops back to block 1006 to continue monitoring.

If it is determined at decision block 1010 that the user-specified alarm delay time has been reached, then an alarm is output at block 1012. If, however, it is determined that the alarm delay time has not been reached, then the multiparameter patient monitoring process 1000 proceeds to decision block 1014. The alarm can have similar features as described above and can be provided by, on, or to any of the devices described above.

At decision block 1014, it is determined whether a value of the second monitored physiological parameter has deviated from a previous value. If so, then the alarm delay time is dynamically modified at block 1016, and the process 1000 loops back to block 1006 to continue monitoring. If not, the process 1000 loops back to decision block 1006 to continue monitoring without changing the delay time.

In certain embodiments, a deviation from a previous value for the second monitored physiological parameter includes a reduction or increase in value. In other embodiments, a deviation from a previous value includes a deviation beyond a threshold or threshold range of acceptable values. The threshold or threshold range for the second monitored physiological parameter can be set or adjusted by a user (e.g., a clinician) depending on patient-specific factors (e.g., age, gender, comorbidity, or the like). The threshold range of values can be set to include any range of values.

In one embodiment, the degree of modification of the alarm delay can depend on the degree of deviation of the second monitored physiological parameter. In another embodiment, the degree of modification of the alarm delay can also depend on the value of the user-specified or default alarm delay time and/or the identity of the first physiological parameter being monitored.

The dynamic modification can be performed in a linear, step-wise, logarithmic, proportional, or any other fashion. For example, the change in the alarm delay corresponding to the first monitored physiological parameter can be proportional to the change or deviation in the second monitored physiological parameter. In another embodiment, a series of successive threshold ranges of values of the second physiological parameter can be provided, wherein each threshold range corresponds to a different amount of delay adjustment.

For example and not by way of limitation, the first physiological parameter can be respiratory rate and the second physiological parameter can be SpO₂. In one embodiment, a user-specified or default alarm delay time can be sixty seconds. If the respiratory rate is less than a given threshold for less than the alarm delay time, it can be determined whether the current SpO₂ level has deviated. Based at least partly on this deviation, the alarm delay time can be adjusted. For example, if the SpO₂ level has dropped, the alarm delay time can be reduced, for example, to 30 seconds, or to 15 seconds, or to another value. If the second monitored physiological parameter deviates too far beyond a threshold range, an alarm corresponding to the second monitored physiological parameter can also be triggered.

FIGS. 11 through 17 illustrate additional example embodiments of physiological parameter displays 1100-1700. These displays 1100-1700 can be implemented by any physiological monitor, including any of the monitors described herein. The displays 1100-1700 shown illustrate example techniques for depicting parameter values and associated confidence. The displays 1100-1700 can be used to depict single parameter (e.g., internal) confidence, multiparameter confidence, or both. The displays 1100-1700 can be implemented for respiratory rate or for any other physiological parameter, including, but not limited to, SpO₂, hemoglobin species (including total hemoglobin), pulse rate, glucose, or any of the other parameters described herein.

Referring initially to FIG. 11, the display 1100 includes an example parameter value scale 1102 and a plot area 1106. An indicator 1110 displayed in the plot area 1106 plots a parameter value 1108 together with associated confidence. In the depicted embodiment, the indicator 1110 is a normal or Gaussian density function (e.g., bell curve) that includes a peak 1112. The indicator 1110 can represent a current (or most recent) parameter value at the peak 1112 and the confidence associated with that parameter value.

The value of the parameter at the peak 1112 matches the parameter value scale 1102. Thus, for instance, the normal density function is centered at 8.0, and a superimposed value 1108 of "8.0" is superimposed on the indicator 1110, indicating a value of 8.0 for the measured parameter. If the parameter is respiratory rate, the 8.0 can correspond to breaths per minute. If the parameter were hemoglobin (SpHb), the value can be reported as a concentration in g/dL (grams per deciliter) or the like. Other parameters, such as glucose or SpO₂, can have different parameter value scales. The parameter value scale 1102 and/or the superimposed value 1108 are optional and may be omitted in certain embodiments. Likewise, vertical grid lines 1104 are shown but can be optional, and horizontal grid lines can also be provided.

The confidence is represented in certain embodiments by the characteristics of the indicator 1110 as a normal density function (or a variation thereof). The normal density function can be plotted using the Gaussian function or bell curve:

$$f(x) = \frac{1}{\sqrt{2\pi\sigma^2}} e^{-\frac{(x-\mu)^2}{2\sigma^2}}$$

where parameters μ and σ^2 are the mean and variance, respectively. Other related formulas can also be used. In one embodiment, the indicator 1110 can be plotted by assigning μ to be the parameter value and σ^2 (or σ , the standard deviation) to be the computed confidence value (internal, multiparameter, or a combination of the same). Then, the location on the parameter scale 1102 of the indicator 1110 can depend on the

value of the parameter (μ), and the width or dispersion of the indicator **1110** can depend on the confidence (σ^2 or σ). Thus, with higher confidence, the variance (σ^2) can be lower, and the curve of the indicator **1110** can be narrower. With lower confidence, the variance can be higher, and the curve of the indicator **1110** can be wider or more dispersed.

The parameter value and/or confidence can have values that are some linear combination of μ and σ^2 . For instance, the parameter value can be represented as $\alpha\mu$, where α is a real number. Likewise, the confidence value can be represented as $\beta\sigma^2$, where β is a real number.

Advantageously, in certain embodiments, the indicator **1110** provides an at-a-glance view of a parameter value and associated confidence. Because the confidence can be represented as the width of the indicator **1110**, the indicator **1110** can rapidly convey qualitative as well as quantitative information about confidence to a clinician. Most clinicians may be familiar with the normal density function or its associated distribution and may therefore readily associate the shape of the indicator **1110** with qualitative meaning regarding confidence.

Other features of the display **1100** include a phantom indicator **1120**, shown as dashed lines, that represents the previous-calculated parameter and confidence values. The phantom indicator **1120** can be used for the immediately previous values, or multiple phantom indicators **1120** can be used for multiple sets of previous values. A safety zone bar **1130** is also displayed. The safety zone bar **1130** includes three areas—a red zone **1132**, a yellow zone **1134**, and a green zone **1136**, representing unsafe, marginally safe, and safe parameter values, respectively. If the peak **1112** of the indicator **1110** is in the green zone **1132**, the value is represented as being safe, and so forth. The colors, including any colors discussed herein, may be outlines instead of solid colors. Further, the colors can be replaced with hatch marks, lines, dots, or any of a variety of other indications to represent different zones of safety.

Some example safety zone ranges for respiratory rate for an adult are as follows. The red, or danger zone **1132** can include about 5 breaths per minute (BPM) or less. A second red zone (see, e.g., FIG. **12**) might include about 30 BPM or more. The yellow, or marginally safe zone **1134**, can include about 6 BPM to about 10 BPM. A second yellow zone (see, e.g., FIG. **12**) can include about 24 BPM to about 30 BPM. The green zone **1136** can include about 11 BPM to about 23 BPM. These ranges are merely examples, however, and can vary considerably depending on, for instance, patient age, gender, comorbidity, medications, current activities (e.g., exercising or sitting), combinations of the same, and the like.

Although the normal density function has been used to illustrate confidence, other indicators in other embodiments can be illustrated using different probability density functions (such as binomial or Poisson functions). Further, the indicator need not be illustrated using a probability density function but can instead be illustrated using one or more boxes, circles, triangles, or other geometric shapes whose width, length, height, or other property changes with changing confidence (see, e.g., FIG. **16**). Further, the characteristics of the density function can depend on other factors in addition to or instead of confidence, such as patient comorbidities (other diseases can affect the confidence of the measurement), drugs taken by the patient (which can also affect the confidence), age, gender, combinations of the same, and the like. Further, the parameter value scale **1102** can change depending on the range of the parameter being considered, and a clinician can optionally zoom in or zoom out to a smaller or larger range.

Moreover, the safety ranges on the safety zone bar **1130** can depend on or otherwise be adjusted by a clinician based on patient comorbidity (e.g., hemoglobinopathy or thalassemia can affect the safe zones for hemoglobin), medications, age, gender, current activities or patient condition (such as donating blood, which can result in a higher start point of the green safety zone for hemoglobin), combinations of the same, and the like. The alternative implementations described with respect to FIG. **11**, as well as any of the other features of the display **1100**, can be used for any of the displays **1200-1700** described below as well.

A variant of the display **1100** is shown as the display **1200** in FIG. **12**. The display **1200** also shows an indicator **1210**, which can represent a normal density function as described above with respect to FIG. **11**. As such, the indicator **1210** can represent a parameter value at a peak of the indicator **1210** and a confidence value associated with the shape of the indicator. In this indicator **1210**, however, the indicator **1210** itself is colored to show vertical safety zones **1240**, **1242**, and **1244**. These safety zones can be similar to the safety zones described above with respect to the safety zone bar **1130** of FIG. **11**.

The safety zone **1240** can represent a green or safe zone, the safety zone **1242** can represent a yellow or marginal zone, and the zone **1244** can represent a red or danger zone. Although the indicator **1210** is centered on these zones, different values of the parameter represented by the indicator **1210** can shift the indicator **1210** closer toward one or more of the zones. Thus, the color of the indicator **1210** can change, for example, be entirely yellow, or entirely red, or entirely green, or some different combination of the same. Further, the display **1200** can be modified in some embodiments to add a safety zone bar like the safety zone bar **1130** of FIG. **11**. The colors of the safety zone bar can vertically match the colors above the bar shown in the indicator **1210** (see, e.g., FIG. **14**).

FIG. **13** depicts another embodiment of a display **1300**. Similar to the displays **1100**, **1200**, the display **1300** includes an indicator **1310** that uses normal density function features to represent a parameter value and confidence. However, the indicator **1310** includes horizontal safety zones **1340**, **1342**, **1344**. The horizontal zones can be similar to the safety zones described above and can include, for example, red, green, and yellow (or other) colors. In another embodiment (not shown), the indicator **1310** can be a single solid color corresponding to the safety zone of the peak of the indicator **1310**. The indicator **1310** can also include gradual instead of abrupt transitions between colors.

Referring to FIG. **14**, a display **1400** includes an indicator **1410** having the density function characteristics described above. In addition, the indicator **1410** is colored vertically with safety zones **1442** and **1444**, similar to the indicator **1210** above. In addition, a safety zone bar **1430** is also shown, which has colors that correspond vertically to the colors of the indicator **1410**. Thus, a zone **1440** can be green, the zones **1442** and **1452** (of the bar **1430**) can be yellow, and the zones **1444** and **1454** (of the bar **1430**) can be red, or the like.

FIG. **15** illustrates a display **1500** having an indicator **1510** without color. Instead, the indicator **1510**, which can include the features of the indicators described above, includes markings **1520** to reflect percentages of standard deviations of the normal density function. These standard deviations can correspond to confidence intervals. These markings **1520** include horizontal arrows, vertical lines, and associated percentage numbers to mark a first standard deviation (e.g., 68% confidence that the parameter lies within the interval marked by the arrow), a second standard deviation (e.g., 95% confidence that the parameter lies within the interval marked by the

arrow), and the third standard deviation (e.g., 99% confidence interval). Color or safety zones can be added to the indicator **1510** as in any of the other example indicators described herein.

FIG. **16** illustrates yet another display **1600** with an indicator **1600**. Unlike the indicators described above, the indicator **1610** is not a bell curve but instead a geometric arrangement of vertical bars **1612**. The vertical bars **1612** can approximate a bell curve, however. Horizontal bars may also be used similarly. The width of the vertical bars **1612** and/or the width of the indicator **1600** as a whole can represent the confidence of a parameter. Further, the parameter value can be represented by the center bar **1612** on a parameter value scale (not shown; see FIG. **11**). The narrower the indicator **1610**, the more confidence is represented, and the wider the indicator **1610**, the less confidence is represented, in one embodiment.

Referring to FIG. **17**, another embodiment of a display **1700** is shown. The display **1700** illustrates additional features that can be combined with any of the embodiments described above. The display **1700** includes two indicators **1710a**, **1710b**. Each of the indicators **1710** is asymmetrical instead of bell-curve shaped. On one side of the peak **1712a**, **1712b** for each indicator **1710a**, **1710b**, a portion **1714a**, **1714b** of the curve is wider or more dispersed than the other side, leading to the asymmetry.

In one embodiment, the indicator **1710** can be asymmetrical if the confidence measure indicates higher confidence on one side of the parameter value as opposed to the other. Asymmetric confidence can occur for some parameters due to bias. For instance, with hemoglobin, a bias for more positive values at lower values of hemoglobin may occur based on the levels of other blood constituents such as oxygen saturation (SpO_2) or carboxyhemoglobin (SpCO). Similarly, hemoglobin can have a bias for more negative values at higher values of hemoglobin based on levels of other blood constituents. Thus, for lower values of hemoglobin, the curve may be wider in the positive direction, and vice versa.

Positive asymmetry for lower parameter values is illustrated by the indicator **1710a**, while negative asymmetry for higher parameter values is illustrated by the indicator **1710b**. To illustrate both positive and negative asymmetry, two indicators **1710a**, **1710b** are depicted on the display **1700**. However, in one implementation, only one indicator **1710** is displayed. Multiple indicators are also possible, such as for multiple parameters on a single display. Moreover, the features described herein with respect to FIG. **17** can be extended to arbitrary geometric shapes. A triangle, for instance, can have one half that is wider than another half based on positive or negative bias in confidence levels.

Any of the displays **1100-1700** can be used to indicate the occurrence and/or severity of an alarm. For instance, the indicators described above can pulsate or flash when an alarm occurs, optionally in conjunction with an audible alarm. The seriousness of the alarm can depend at least partially on the measured confidence. Higher confidence (e.g., a narrow indicator) can result in a more urgent alarm, whereas less urgent alarms can result for less confident parameter values. This urgency can be displayed in a variety of ways, for example, by increasing the rate that the indicator flashes, increasing the frequency and/or pitch of an audible alarm, combinations of the same, and the like.

Conditional language used herein, such as, among others, "can," "could," "might," "may," "e.g.," and the like, unless specifically stated otherwise, or otherwise understood within the context as used, is generally intended to convey that certain embodiments include, while other embodiments do

not include, certain features, elements and/or states. Thus, such conditional language is not generally intended to imply that features, elements and/or states are in any way required for one or more embodiments or that one or more embodiments necessarily include logic for deciding, with or without author input or prompting, whether these features, elements and/or states are included or are to be performed in any particular embodiment.

Depending on the embodiment, certain acts, events, or functions of any of the methods described herein can be performed in a different sequence, can be added, merged, or left out all together (e.g., not all described acts or events are necessary for the practice of the method). Moreover, in certain embodiments, acts or events can be performed concurrently, e.g., through multi-threaded processing, interrupt processing, or multiple processors or processor cores, rather than sequentially.

The various illustrative logical blocks, modules, circuits, and algorithm steps described in connection with the embodiments disclosed herein can be implemented as electronic hardware, computer software, or combinations of both. To clearly illustrate this interchangeability of hardware and software, various illustrative components, blocks, modules, circuits, and steps have been described above generally in terms of their functionality. Whether such functionality is implemented as hardware or software depends upon the particular application and design constraints imposed on the overall system. The described functionality can be implemented in varying ways for each particular application, but such implementation decisions should not be interpreted as causing a departure from the scope of the disclosure.

The various illustrative logical blocks, modules, and circuits described in connection with the embodiments disclosed herein can be implemented or performed with a general purpose processor, a digital signal processor (DSP), an application specific integrated circuit (ASIC), a field programmable gate array (FPGA) or other programmable logic device, discrete gate or transistor logic, discrete hardware components, or any combination thereof designed to perform the functions described herein. A general purpose processor can be a microprocessor, but in the alternative, the processor can be any conventional processor, controller, microcontroller, or state machine. A processor can also be implemented as a combination of computing devices, e.g., a combination of a DSP and a microprocessor, a plurality of microprocessors, one or more microprocessors in conjunction with a DSP core, or any other such configuration.

The blocks of the methods and algorithms described in connection with the embodiments disclosed herein can be embodied directly in hardware, in a software module executed by a processor, or in a combination of the two. A software module can reside in RAM memory, flash memory, ROM memory, EPROM memory, EEPROM memory, registers, a hard disk, a removable disk, a CD-ROM, or any other form of computer-readable storage medium known in the art. An exemplary storage medium is coupled to a processor such that the processor can read information from, and write information to, the storage medium. In the alternative, the storage medium can be integral to the processor. The processor and the storage medium can reside in an ASIC. The ASIC can reside in a user terminal. In the alternative, the processor and the storage medium can reside as discrete components in a user terminal.

While the above detailed description has shown, described, and pointed out novel features as applied to various embodiments, it will be understood that various omissions, substitutions, and changes in the form and details of the devices or

algorithms illustrated can be made without departing from the spirit of the disclosure. As will be recognized, certain embodiments of the inventions described herein can be embodied within a form that does not provide all of the features and benefits set forth herein, as some features can be used or practiced separately from others. The scope of certain inventions disclosed herein is indicated by the appended claims rather than by the foregoing description. All changes which come within the meaning and range of equivalency of the claims are to be embraced within their scope.

What is claimed is:

1. A method of triggering an alarm based on information related to a physiological parameter, the method comprising: by a processor:
 - providing functionality for a clinician to specify an alarm delay time;
 - monitoring a first physiological parameter reflecting physiological information obtained from a physiological sensor coupled to a patient, the first physiological parameter comprising respiratory rate;
 - determining whether the first physiological parameter has satisfied a first alarm threshold for an amount of time corresponding to the alarm delay time;
 - triggering an alarm in response to said determining that a value of the first physiological parameter has satisfied the first alarm threshold
 - monitoring a second physiological parameter, the second physiological parameter comprising oxygen saturation;
 - determining whether the second physiological parameter has fallen below a second alarm threshold;
 - automatically reducing the alarm delay time in response to the second physiological parameter having fallen below the second alarm threshold; and
 - modifying a display of a confidence indicator based on the amount of time that the first physiological parameter has satisfied the first alarm threshold.
2. The method of claim 1, wherein said automatically reducing the alarm delay time comprises reducing the alarm delay time to a value that corresponds to an amount that the second physiological parameter has fallen below the second alarm threshold.
3. The method of claim 1, wherein said providing functionality for a clinician to specify the alarm delay time comprises providing a user interface comprising an option to adjust the alarm delay time.
4. A method of triggering an alarm based on information related to a physiological parameter, the method comprising:

by a processor:

- monitoring a first physiological parameter reflecting physiological information obtained from a physiological sensor coupled to a patient, the first physiological parameter comprising respiratory rate;
- automatically adjusting an alarm delay threshold for the first physiological parameter based on a value of a second physiological parameter, the second physiological parameter comprising oxygen saturation, wherein said adjusting comprises reducing the alarm delay threshold in response to the value of the second physiological parameter meeting a first threshold;
- determining whether the first physiological parameter has satisfied a second threshold for an amount of time corresponding to the alarm delay threshold; and
- triggering an alarm in response to said determining that a value of the first physiological parameter has satisfied the second threshold for the amount of time corresponding to the alarm delay threshold.
5. The method of claim 4, wherein an increase in the oxygen saturation causes said adjusting to increase the alarm delay time.
6. The method of claim 4, wherein said automatically adjusting the alarm delay time comprises reducing the alarm delay time to a value that corresponds to an amount that the second physiological parameter has fallen below the first threshold.
7. The method of claim 4, further comprising providing functionality for a user to adjust the alarm delay time.
8. The method of claim 4, further comprising outputting an indicator configured to reflect an amount of time which the first physiological parameter has satisfied the second threshold.
9. The method of claim 8, wherein the indicator is a confidence indicator.
10. The method of claim 8, wherein said outputting the indicator comprises modifying the indicator to reflect a countdown to the alarm delay threshold.
11. The method of claim 8, wherein said outputting the indicator comprises blinking the indicator to indicate that the value of the first physiological parameter has satisfied the second threshold prior to reaching the alarm delay time.
12. The method of claim 4, further comprising providing functionality for a clinician to initially specify the alarm delay time.

* * * * *

UNITED STATES PATENT AND TRADEMARK OFFICE
CERTIFICATE OF CORRECTION

PATENT NO. : 8,430,817 B1
APPLICATION NO. : 12/905530
DATED : April 30, 2013
INVENTOR(S) : Al-Ali et al.

Page 1 of 1

It is certified that error appears in the above-identified patent and that said Letters Patent is hereby corrected as shown below:

In column 2 (title page 2 item 56) at line 42, Under U.S. Patent Documents, change
“Kianl et al.” to --Kiani et al.--.

In column 2 (title page 4 item 56) at line 12, Under Other Publications, change
“Experieances,” to --Experiences,--.

In column 2 (title page 4 item 56) at line 39, Under Other Publications, change
“Sheete,” to --Sheets,--.

In the Drawings:

Sheet 10 of 22 (Fig. 6B-1) at line 6 (approx.), change “Photpleth” to --Photopleth--.

Sheet 12 of 22 (Fig. 6C) at line 6 (approx.), change “Photpleth” to --Photopleth--.

In the Specification:

In column 26 at line 3, change “thalacemia” to --thalassemia--.

In the Claims:

In column 29 at line 22, In Claim 1, after “satisfied” delete “an”.

Signed and Sealed this
Eighth Day of October, 2013



Teresa Stanek Rea
Deputy Director of the United States Patent and Trademark Office

专利名称(译)	用于确定呼吸速率测量的置信度的系统		
公开(公告)号	US8430817	公开(公告)日	2013-04-30
申请号	US12/905530	申请日	2010-10-15
[标]申请(专利权)人(译)	AL ALI AMMAR 穆赫辛BILAL 奥莱利MICHAEL 韦尔奇JAMES P		
申请(专利权)人(译)	AL-ALI , AMMAR 穆赫辛 , BILAL 奥赖利MICHAEL 韦尔奇JAMES P.		
当前申请(专利权)人(译)	Masimo公司		
[标]发明人	AL ALI AMMAR MUHSIN BILAL OREILLY MICHAEL WELCH JAMES P		
发明人	AL-ALI, AMMAR MUHSIN, BILAL O'REILLY, MICHAEL WELCH, JAMES P.		
IPC分类号	A61B5/00		
CPC分类号	A61B5/7221 A61B5/746 G06F19/3418 A61B5/14551 A61B5/0816 A61B5/7278 G16H40/60 G16H40/63 A61B5/0205 A61B5/021 A61B5/024 A61B5/02416 A61B5/0402 A61B5/044 A61B5/0803 A61B5/14542 A61B5/742 A61B7/003 G16H40/67		
助理审查员(译)	ARCHER , MARIE		
优先权	61/366866 2010-07-22 US 61/261199 2009-11-13 US 61/252086 2009-10-15 US		
外部链接	Espacenet USPTO		

摘要(译)

除了其他特征之外，本公开还描述了用于使用多个生理参数输入来确定呼吸速率测量中的多参数置信度的系统和方法。例如，患者监测系统可以至少部分地基于从其他非声学传感器或监测器获得的输入以编程方式确定从声学传感器获得的呼吸速率测量的多参数置信度。患者监测系统可以输出反映编程确定的多参数置信度的多参数置信度指示。多参数置信度指示可以帮助临床医生基于患者的呼吸率确定是否或如何治疗患者。

