



US010278648B2

(12) **United States Patent**
Al-Ali et al.

(10) **Patent No.:** **US 10,278,648 B2**
(45) **Date of Patent:** ***May 7, 2019**

(54) **AUTOMATED CCHD SCREENING AND DETECTION**

(71) Applicant: **MASIMO CORPORATION**, Irvine, CA (US)

(72) Inventors: **Ammar Al-Ali**, San Juan Capistrano, CA (US); **Bilal Muhsin**, San Clemente, CA (US)

(73) Assignee: **MASIMO CORPORATION**, Irvine, CA (US)

(*) Notice: Subject to any disclaimer, the term of this patent is extended or adjusted under 35 U.S.C. 154(b) by 94 days.

This patent is subject to a terminal disclaimer.

(21) Appl. No.: **15/279,134**

(22) Filed: **Sep. 28, 2016**

(65) **Prior Publication Data**

US 2017/0014084 A1 Jan. 19, 2017

Related U.S. Application Data

(60) Continuation of application No. 15/195,037, filed on Jun. 28, 2016, which is a division of application No. (Continued)

(51) **Int. Cl.**
A61B 5/00 (2006.01)
A61B 5/1455 (2006.01)
(Continued)

(52) **U.S. Cl.**
CPC **A61B 5/7275** (2013.01); **A61B 5/0205** (2013.01); **A61B 5/0208** (2013.01);
(Continued)

(58) **Field of Classification Search**
CPC . A61B 5/0205; A61B 5/0208; A61B 5/1455; A61B 5/14551; A61B 5/14552;
(Continued)

(56) **References Cited**

U.S. PATENT DOCUMENTS

4,869,253 A 9/1989 Craig et al.
4,960,128 A 10/1990 Gordon et al.
(Continued)

FOREIGN PATENT DOCUMENTS

WO WO 2013/103885 7/2013

OTHER PUBLICATIONS

Granelli A.D., Mellander et al., "Screening for duct-dependent congenital heart disease with pulse oximetry: A critical evaluation of strategies to maximize sensitivity", Acta Paediatrica, 2005; 94:1590-1596, http://www.masimo.com/pdf/Granelli_Article.pdf, 1 page downloaded and printed from the World Wide Web.

(Continued)


Primary Examiner — Eric F Winakur
Assistant Examiner — Chu Chuan Liu

(74) *Attorney, Agent, or Firm* — Knobbe Martens Olson & Bear LLP

(57) **ABSTRACT**

Automated critical congenital heart defect ("CCHD") screening systems and processes are described. A caregiver may be guided to use a single or dual sensor pulse oximeter to obtain pre- and post-ductal blood oxygenation measurements. A delta of the measurements indicates the possible existence or nonexistence of a CCHD. Errors in the measurements are reduced by a configurable measurement confidence threshold based on, for example, a perfusion index. Measurement data may be stored and retrieved from a remote data processing center for repeated screenings.

19 Claims, 16 Drawing Sheets



- Related U.S. Application Data**
- 13/733,782, filed on Jan. 3, 2013, now Pat. No. 9,392,945.
- (60) Provisional application No. 61/703,132, filed on Sep. 19, 2012, provisional application No. 61/583,143, filed on Jan. 4, 2012.
- (51) **Int. Cl.**
A61B 5/02 (2006.01)
A61B 5/0205 (2006.01)
A61B 5/0295 (2006.01)
A61B 5/024 (2006.01)
A61B 5/08 (2006.01)
- (52) **U.S. Cl.**
 CPC *A61B 5/0295* (2013.01); *A61B 5/14551* (2013.01); *A61B 5/14552* (2013.01); *A61B 5/684* (2013.01); *A61B 5/6825* (2013.01); *A61B 5/6829* (2013.01); *A61B 5/7221* (2013.01); *A61B 5/7278* (2013.01); *A61B 5/7282* (2013.01); *A61B 5/743* (2013.01); *A61B 5/024* (2013.01); *A61B 5/0816* (2013.01); *A61B 2503/04* (2013.01)
- (58) **Field of Classification Search**
 CPC *A61B 5/684*; *A61B 5/6825*; *A61B 5/6829*; *A61B 5/7221*; *A61B 5/7275*; *A61B 5/7278*; *A61B 5/743*
 See application file for complete search history.

(56) **References Cited**

U.S. PATENT DOCUMENTS

4,964,408 A	10/1990	Hink et al.	5,785,659 A	7/1998	Caro et al.
5,041,187 A	8/1991	Hink et al.	5,791,347 A	8/1998	Flaherty et al.
5,069,213 A	12/1991	Polczynski	5,810,734 A	9/1998	Caro et al.
5,078,714 A	1/1992	Katims	5,823,950 A	10/1998	Diab et al.
5,163,438 A	11/1992	Gordon et al.	5,830,131 A	11/1998	Caro et al.
5,319,355 A	6/1994	Russek	5,833,618 A	11/1998	Caro et al.
5,337,744 A	8/1994	Branigan	5,860,919 A	1/1999	Kiani-Azarbayjany et al.
5,341,805 A	8/1994	Stavridi et al.	5,890,929 A	4/1999	Mills et al.
D353,195 S	12/1994	Savage et al.	5,904,654 A	5/1999	Wohltmann et al.
D353,196 S	12/1994	Savage et al.	5,919,134 A	7/1999	Diab
5,377,676 A	1/1995	Vari et al.	5,934,277 A	8/1999	Mortz
5,398,682 A	3/1995	Lynn	5,934,925 A	8/1999	Tobler et al.
D359,546 S	6/1995	Savage et al.	5,940,182 A	8/1999	Lepper, Jr. et al.
5,431,170 A	7/1995	Mathews	5,987,343 A	11/1999	Kinast
D361,840 S	8/1995	Savage et al.	5,995,855 A	11/1999	Kiani et al.
D362,063 S	9/1995	Savage et al.	5,997,343 A	12/1999	Mills et al.
5,452,717 A	9/1995	Branigan et al.	6,002,952 A	12/1999	Diab et al.
D363,120 S	10/1995	Savage et al.	6,011,986 A	1/2000	Diab et al.
5,456,252 A	10/1995	Vari et al.	6,027,452 A	2/2000	Flaherty et al.
5,479,934 A	1/1996	Imran	6,036,642 A	3/2000	Diab et al.
5,482,036 A	1/1996	Diab et al.	6,045,509 A	4/2000	Caro et al.
5,490,505 A	2/1996	Diab et al.	6,067,462 A	5/2000	Diab et al.
5,494,043 A	2/1996	O'Sullivan et al.	6,081,735 A	6/2000	Diab et al.
5,533,511 A	7/1996	Kaspari et al.	6,083,172 A	7/2000	Baker, Jr. et al.
5,534,851 A	7/1996	Russek	6,088,607 A	7/2000	Diab et al.
5,561,275 A	10/1996	Savage et al.	6,110,522 A	8/2000	Lepper, Jr. et al.
5,562,002 A	10/1996	Lalin	6,124,597 A	9/2000	Shehada
5,590,649 A	1/1997	Caro et al.	6,128,521 A	10/2000	Marro et al.
5,602,924 A	2/1997	Durand et al.	6,129,675 A	10/2000	Jay
5,632,272 A	5/1997	Diab et al.	6,144,868 A	11/2000	Parker
5,638,816 A	6/1997	Kiani-Azarbayjany et al.	6,151,516 A	11/2000	Kiani-Azarbayjany et al.
5,638,818 A	6/1997	Diab et al.	6,152,754 A	11/2000	Gerhardt et al.
5,645,440 A	7/1997	Tobler et al.	6,157,850 A	12/2000	Diab et al.
5,685,299 A	11/1997	Diab et al.	6,165,005 A	12/2000	Mills et al.
D393,830 S	4/1998	Tobler et al.	6,184,521 B1	2/2001	Coffin, IV et al.
5,743,262 A	4/1998	Lepper, Jr. et al.	6,206,830 B1	3/2001	Diab et al.
5,758,644 A	6/1998	Diab et al.	6,229,856 B1	5/2001	Diab et al.
5,760,910 A	6/1998	Lepper, Jr. et al.	6,232,609 B1	5/2001	Snyder et al.
5,769,785 A	6/1998	Diab et al.	6,236,872 B1	5/2001	Diab et al.
5,782,757 A	7/1998	Diab et al.	6,241,683 B1	6/2001	Macklem et al.
			6,253,097 B1	6/2001	Aronow et al.
			6,256,523 B1	7/2001	Diab et al.
			6,263,222 B1	7/2001	Diab et al.
			6,278,522 B1	8/2001	Lepper, Jr. et al.
			6,280,213 B1	8/2001	Tobler et al.
			6,285,896 B1	9/2001	Tobler et al.
			6,301,493 B1	10/2001	Marro et al.
			6,308,089 B1	10/2001	von der Ruhr et al.
			6,317,627 B1	11/2001	Ennen et al.
			6,321,100 B1	11/2001	Parker
			6,321,110 B1	11/2001	Ito et al.
			6,325,761 B1	12/2001	Jay
			6,334,065 B1	12/2001	Al-Ali et al.
			6,343,224 B1	1/2002	Parker
			6,349,228 B1	2/2002	Kiani et al.
			6,360,114 B1	3/2002	Diab et al.
			6,368,283 B1	4/2002	Xu et al.
			6,371,921 B1	4/2002	Caro et al.
			6,377,829 B1	4/2002	Al-Ali
			6,388,240 B2	5/2002	Schulz et al.
			6,397,091 B2	5/2002	Diab et al.
			6,430,437 B1	8/2002	Marro
			6,430,525 B1	8/2002	Weber et al.
			6,463,311 B1	10/2002	Diab
			6,470,199 B1	10/2002	Kopotic et al.
			6,501,975 B2	12/2002	Diab et al.
			6,505,059 B1	1/2003	Kollias et al.
			6,515,273 B2	2/2003	Al-Ali
			6,519,487 B1	2/2003	Parker
			6,525,386 B1	2/2003	Mills et al.
			6,526,300 B1	2/2003	Kiani et al.
			6,541,756 B2	4/2003	Schulz et al.
			6,542,764 B1	4/2003	Al-Ali et al.
			6,580,086 B1	6/2003	Schulz et al.
			6,584,336 B1	6/2003	Ali et al.
			6,595,316 B2	7/2003	Cybulski et al.
			6,597,932 B2	7/2003	Tian et al.
			6,597,933 B2	7/2003	Kiani et al.
			6,606,511 B1	8/2003	Ali et al.

(56)

References Cited

U.S. PATENT DOCUMENTS

6,632,181	B2	10/2003	Flaherty et al.	RE39,672	E	6/2007	Shehada et al.
6,639,668	B1	10/2003	Trepagnier	7,239,905	B2	7/2007	Kiani-Azarbayjany et al.
6,640,116	B2	10/2003	Diab	7,245,953	B1	7/2007	Parker
6,643,530	B2	11/2003	Diab et al.	7,254,429	B2	8/2007	Schurman et al.
6,650,917	B2	11/2003	Diab et al.	7,254,431	B2	8/2007	Al-Ali
6,654,624	B2	11/2003	Diab et al.	7,254,433	B2	8/2007	Diab et al.
6,658,276	B2	12/2003	Kiani et al.	7,254,434	B2	8/2007	Schulz et al.
6,661,161	B1	12/2003	Lanzo et al.	7,272,425	B2	9/2007	Al-Ali
6,671,531	B2	12/2003	Al-Ali et al.	7,274,955	B2	9/2007	Kiani et al.
6,678,543	B2	1/2004	Diab et al.	D554,263	S	10/2007	Al-Ali
6,684,090	B2	1/2004	Ali et al.	7,280,858	B2	10/2007	Al-Ali et al.
6,684,091	B2	1/2004	Parker	7,289,835	B2	10/2007	Mansfield et al.
6,697,656	B1	2/2004	Al-Ali	7,292,883	B2	11/2007	De Felice et al.
6,697,657	B1	2/2004	Shehada et al.	7,295,866	B2	11/2007	Al-Ali
6,697,658	B2	2/2004	Al-Ali	7,328,053	B1	2/2008	Diab et al.
RE38,476	E	3/2004	Diab et al.	7,332,784	B2	2/2008	Mills et al.
6,699,194	B1	3/2004	Diab et al.	7,340,287	B2	3/2008	Mason et al.
6,714,804	B2	3/2004	Al-Ali et al.	7,341,559	B2	3/2008	Schulz et al.
RE38,492	E	4/2004	Diab et al.	7,343,186	B2	3/2008	Lamego et al.
6,721,582	B2	4/2004	Trepagnier et al.	D566,282	S	4/2008	Al-Ali et al.
6,721,585	B1	4/2004	Parker	7,355,512	B1	4/2008	Al-Ali
6,725,075	B2	4/2004	Al-Ali	7,356,365	B2	4/2008	Schurman
6,728,560	B2	4/2004	Kollias et al.	7,371,981	B2	5/2008	Abdul-Hafiz
6,735,459	B2	5/2004	Parker	7,373,193	B2	5/2008	Al-Ali et al.
6,745,060	B2	6/2004	Diab et al.	7,373,194	B2	5/2008	Weber et al.
6,760,607	B2	7/2004	Al-Ali	7,376,453	B1	5/2008	Diab et al.
6,770,028	B1	8/2004	Ali et al.	7,377,794	B2	5/2008	Al Ali et al.
6,771,994	B2	8/2004	Kiani et al.	7,377,899	B2	5/2008	Weber et al.
6,792,300	B1	9/2004	Diab et al.	7,383,070	B2	6/2008	Diab et al.
6,813,511	B2	11/2004	Diab et al.	7,415,297	B2	8/2008	Al-Ali et al.
6,816,741	B2	11/2004	Diab	7,428,432	B2	9/2008	Ali et al.
6,822,564	B2	11/2004	Al-Ali	7,438,683	B2	10/2008	Al-Ali et al.
6,826,419	B2	11/2004	Diab et al.	7,440,787	B2	10/2008	Diab
6,830,711	B2	12/2004	Mills et al.	7,454,240	B2	11/2008	Diab et al.
6,850,787	B2	2/2005	Weber et al.	7,467,002	B2	12/2008	Weber et al.
6,850,788	B2	2/2005	Al-Ali	7,469,157	B2	12/2008	Diab et al.
6,852,083	B2	2/2005	Caro et al.	7,471,969	B2	12/2008	Diab et al.
6,861,639	B2	3/2005	Al-Ali	7,471,971	B2	12/2008	Diab et al.
6,898,452	B2	5/2005	Al-Ali et al.	7,483,729	B2	1/2009	Al-Ali et al.
6,920,345	B2	7/2005	Al-Ali et al.	7,483,730	B2	1/2009	Diab et al.
6,931,268	B1	8/2005	Kiani-Azarbayjany et al.	7,489,958	B2	2/2009	Diab et al.
6,934,570	B2	8/2005	Kiani et al.	7,496,391	B2	2/2009	Diab et al.
6,939,305	B2	9/2005	Flaherty et al.	7,496,393	B2	2/2009	Diab et al.
6,943,348	B1	9/2005	Coffin, IV	D587,657	S	3/2009	Al-Ali et al.
6,950,687	B2	9/2005	Al-Ali	7,499,741	B2	3/2009	Diab et al.
6,961,598	B2	11/2005	Diab	7,499,835	B2	3/2009	Weber et al.
6,970,792	B1	11/2005	Diab	7,500,950	B2	3/2009	Al-Ali et al.
6,979,812	B2	12/2005	Al-Ali	7,509,154	B2	3/2009	Diab et al.
6,985,764	B2	1/2006	Mason et al.	7,509,494	B2	3/2009	Al-Ali
6,993,371	B2	1/2006	Kiani et al.	7,510,849	B2	3/2009	Schurman et al.
6,996,427	B2	2/2006	Ali et al.	7,526,328	B2	4/2009	Diab et al.
6,999,904	B2	2/2006	Weber et al.	7,530,942	B1	5/2009	Diab
7,003,338	B2	2/2006	Weber et al.	7,530,949	B2	5/2009	Al Ali et al.
7,003,339	B2	2/2006	Diab et al.	7,530,955	B2	5/2009	Diab et al.
7,015,451	B2	3/2006	Dalke et al.	7,563,110	B2	7/2009	Al-Ali et al.
7,024,233	B2	4/2006	Ali et al.	7,596,398	B2	9/2009	Al-Ali et al.
7,027,849	B2	4/2006	Al-Ali	7,618,375	B2	11/2009	Flaherty
7,030,749	B2	4/2006	Al-Ali	D606,659	S	12/2009	Kiani et al.
7,039,449	B2	5/2006	Al-Ali	7,647,083	B2	1/2010	Al-Ali et al.
7,041,060	B2	5/2006	Flaherty et al.	D609,193	S	2/2010	Al-Ali et al.
7,044,918	B2	5/2006	Diab	D614,305	S	4/2010	Al-Ali et al.
7,048,687	B1	5/2006	Reuss et al.	RE41,317	E	5/2010	Parker
7,067,893	B2	6/2006	Mills et al.	7,729,733	B2	6/2010	Al-Ali et al.
7,096,052	B2	8/2006	Mason et al.	7,734,320	B2	6/2010	Al-Ali
7,096,054	B2	8/2006	Abdul-Hafiz et al.	7,761,127	B2	7/2010	Al-Ali et al.
7,132,641	B2	11/2006	Schulz et al.	7,761,128	B2	7/2010	Al-Ali et al.
7,142,901	B2	11/2006	Kiani et al.	7,764,982	B2	7/2010	Dalke et al.
7,149,561	B2	12/2006	Diab	D621,516	S	8/2010	Kiani et al.
7,186,966	B2	3/2007	Al-Ali	7,791,155	B2	9/2010	Diab
7,190,261	B2	3/2007	Al-Ali	7,801,581	B2	9/2010	Diab
7,215,984	B2	5/2007	Diab	7,822,452	B2	10/2010	Schurman et al.
7,215,986	B2	5/2007	Diab	RE41,912	E	11/2010	Parker
7,221,971	B2	5/2007	Diab	7,844,313	B2	11/2010	Kiani et al.
7,225,006	B2	5/2007	Al-Ali et al.	7,844,314	B2	11/2010	Al-Ali
7,225,007	B2	5/2007	Al-Ali	7,844,315	B2	11/2010	Al-Ali
				7,865,222	B2	1/2011	Weber et al.
				7,873,497	B2	1/2011	Weber et al.
				7,880,606	B2	2/2011	Al-Ali
				7,880,626	B2	2/2011	Al-Ali et al.

(56)

References Cited

U.S. PATENT DOCUMENTS

7,891,355 B2	2/2011	Al-Ali et al.	8,374,665 B2	2/2013	Lamego
7,894,868 B2	2/2011	Al-Ali et al.	8,385,995 B2	2/2013	Al-Ali et al.
7,899,507 B2	3/2011	Al-Ali et al.	8,385,996 B2	2/2013	Smith et al.
7,899,518 B2	3/2011	Trepagnier et al.	8,388,353 B2	3/2013	Kiani et al.
7,904,132 B2	3/2011	Weber et al.	8,399,822 B2	3/2013	Al-Ali
7,909,772 B2	3/2011	Popov et al.	8,401,602 B2	3/2013	Kiani
7,910,875 B2	3/2011	Al-Ali	8,405,608 B2	3/2013	Al-Ali et al.
7,919,713 B2	4/2011	Al-Ali et al.	8,414,499 B2	4/2013	Al-Ali et al.
7,937,128 B2	5/2011	Al-Ali	8,418,524 B2	4/2013	Al-Ali
7,937,129 B2	5/2011	Mason et al.	8,423,106 B2	4/2013	Lamego et al.
7,937,130 B2	5/2011	Diab et al.	8,428,967 B2	4/2013	Olsen et al.
7,941,199 B2	5/2011	Kiani	8,430,817 B1	4/2013	Al-Ali et al.
7,942,824 B1	5/2011	Kayyali et al.	8,437,825 B2	5/2013	Dalvi et al.
7,951,086 B2	5/2011	Flaherty et al.	8,455,290 B2	6/2013	Siskavich
7,957,780 B2	6/2011	Lamego et al.	8,457,703 B2	6/2013	Al-Ali
7,962,188 B2	6/2011	Kiani et al.	8,457,707 B2	6/2013	Kiani
7,962,190 B1	6/2011	Diab et al.	8,463,349 B2	6/2013	Diab et al.
7,976,472 B2	7/2011	Kiani	8,466,286 B2	6/2013	Bellot et al.
7,988,637 B2	8/2011	Diab	8,471,713 B2	6/2013	Poeze et al.
7,990,382 B2	8/2011	Kiani	8,473,020 B2	6/2013	Kiani et al.
7,991,446 B2	8/2011	Ali et al.	8,483,787 B2	7/2013	Al-Ali et al.
8,000,761 B2	8/2011	Al-Ali	8,489,364 B2	7/2013	Weber et al.
8,008,088 B2	8/2011	Bellott et al.	8,498,684 B2	7/2013	Weber et al.
RE42,753 E	9/2011	Kiani-Azarbayjany et al.	8,504,128 B2	8/2013	Blank et al.
8,019,400 B2	9/2011	Diab et al.	8,509,867 B2	8/2013	Workman et al.
8,028,701 B2	10/2011	Al-Ali et al.	8,515,509 B2	8/2013	Bruinsma et al.
8,029,765 B2	10/2011	Bellott et al.	8,523,781 B2	9/2013	Al-Ali
8,036,727 B2	10/2011	Schurman et al.	8,529,301 B2	9/2013	Al-Ali et al.
8,036,728 B2	10/2011	Diab et al.	8,532,727 B2	9/2013	Ali et al.
8,046,040 B2	10/2011	Ali et al.	8,532,728 B2	9/2013	Diab et al.
8,046,041 B2	10/2011	Diab et al.	D692,145 S	10/2013	Al-Ali et al.
8,046,042 B2	10/2011	Diab et al.	8,547,209 B2	10/2013	Kiani et al.
8,048,040 B2	11/2011	Kiani	8,548,548 B2	10/2013	Al-Ali
8,050,728 B2	11/2011	Al-Ali et al.	8,548,549 B2	10/2013	Schurman et al.
RE43,169 E	2/2012	Parker	8,548,550 B2	10/2013	Al-Ali et al.
8,118,620 B2	2/2012	Al-Ali et al.	8,560,032 B2	10/2013	Al-Ali et al.
8,126,528 B2	2/2012	Diab et al.	8,560,034 B1	10/2013	Diab et al.
8,128,572 B2	3/2012	Diab et al.	8,570,167 B2	10/2013	Al-Ali
8,130,105 B2	3/2012	Al-Ali et al.	8,570,503 B2	10/2013	Vo et al.
8,145,287 B2	3/2012	Diab et al.	8,571,617 B2	10/2013	Reichgott et al.
8,150,487 B2	4/2012	Diab et al.	8,571,618 B1	10/2013	Lamego et al.
8,175,672 B2	5/2012	Parker	8,571,619 B2	10/2013	Al-Ali et al.
8,180,420 B2	5/2012	Diab et al.	8,577,431 B2	11/2013	Lamego et al.
8,182,443 B1	5/2012	Kiani	8,581,732 B2	11/2013	Al-Ali et al.
8,185,180 B2	5/2012	Diab et al.	8,584,345 B2	11/2013	Al-Ali et al.
8,190,223 B2	5/2012	Al-Ali et al.	8,588,880 B2	11/2013	Abdul-Hafiz et al.
8,190,227 B2	5/2012	Diab et al.	8,600,467 B2	12/2013	Al-Ali et al.
8,203,438 B2	6/2012	Kiani et al.	8,606,342 B2	12/2013	Diab
8,203,704 B2	6/2012	Merritt et al.	8,626,255 B2	1/2014	Al-Ali et al.
8,204,566 B2	6/2012	Schurman et al.	8,630,691 B2	1/2014	Lamego et al.
8,219,172 B2	7/2012	Schurman et al.	8,634,889 B2	1/2014	Al-Ali et al.
8,224,411 B2	7/2012	Al-Ali et al.	8,641,631 B2	2/2014	Sierra et al.
8,228,181 B2	7/2012	Al-Ali	8,652,060 B2	2/2014	Al-Ali
8,229,533 B2	7/2012	Diab et al.	8,663,107 B2	3/2014	Kiani
8,233,955 B2	7/2012	Al-Ali et al.	8,666,468 B1	3/2014	Al-Ali
8,244,325 B2	8/2012	Al-Ali et al.	8,667,967 B2	3/2014	Al-Ali et al.
8,255,026 B1	8/2012	Al-Ali	8,670,811 B2	3/2014	O'Reilly
8,255,027 B2	8/2012	Al-Ali et al.	8,670,814 B2	3/2014	Diab et al.
8,255,028 B2	8/2012	Al-Ali et al.	8,676,286 B2	3/2014	Weber et al.
8,260,577 B2	9/2012	Weber et al.	8,682,407 B2	3/2014	Al-Ali
8,265,723 B1	9/2012	McHale et al.	RE44,823 E	4/2014	Parker
8,274,360 B2	9/2012	Sampath et al.	RE44,875 E	4/2014	Kiani et al.
8,280,473 B2	10/2012	Al-Ali	8,690,799 B2	4/2014	Telfort et al.
8,301,217 B2	10/2012	Al-Ali et al.	8,700,112 B2	4/2014	Kiani
8,306,596 B2	11/2012	Schurman et al.	8,702,627 B2	4/2014	Telfort et al.
8,310,336 B2	11/2012	Muhsin et al.	8,706,179 B2	4/2014	Parker
8,315,683 B2	11/2012	Al-Ali et al.	8,712,494 B1	4/2014	MacNeish, III et al.
RE43,860 E	12/2012	Parker	8,715,206 B2	5/2014	Telfort et al.
8,337,403 B2	12/2012	Al-Ali et al.	8,718,735 B2	5/2014	Lamego et al.
8,346,330 B2	1/2013	Lamego	8,718,737 B2	5/2014	Diab et al.
8,353,842 B2	1/2013	Al-Ali et al.	8,718,738 B2	5/2014	Blank et al.
8,355,766 B2	1/2013	MacNeish, III et al.	8,720,249 B2	5/2014	Al-Ali
8,359,080 B2	1/2013	Diab et al.	8,721,541 B2	5/2014	Al-Ali et al.
8,364,223 B2	1/2013	Al-Ali et al.	8,721,542 B2	5/2014	Al-Ali et al.
8,364,226 B2	1/2013	Diab et al.	8,723,677 B1	5/2014	Kiani
			8,740,792 B1	6/2014	Kiani et al.
			8,754,776 B2	6/2014	Poeze et al.
			8,755,535 B2	6/2014	Telfort et al.
			8,755,856 B2	6/2014	Diab et al.

(56)

References Cited

U.S. PATENT DOCUMENTS

8,755,872	B1	6/2014	Marinow	9,192,329	B2	11/2015	Al-Ali
8,761,850	B2	6/2014	Lamego	9,192,351	B1	11/2015	Telfort et al.
8,764,671	B2	7/2014	Kiani	9,195,385	B2	11/2015	Al-Ali et al.
8,768,423	B2	7/2014	Shakespeare et al.	9,211,072	B2	12/2015	Kiani
8,771,204	B2	7/2014	Telfort et al.	9,211,095	B1	12/2015	Al-Ali
8,777,634	B2	7/2014	Kiani et al.	9,218,454	B2	12/2015	Kiani et al.
8,781,543	B2	7/2014	Diab et al.	9,226,696	B2	1/2016	Kiani
8,781,544	B2	7/2014	Al-Ali et al.	9,241,662	B2	1/2016	Al-Ali et al.
8,781,549	B2	7/2014	Al-Ali et al.	9,245,668	B1	1/2016	Vo et al.
8,788,003	B2	7/2014	Schurman et al.	9,259,185	B2	2/2016	Abdul-Hafiz et al.
8,790,268	B2	7/2014	Al-Ali	9,267,572	B2	2/2016	Barker et al.
8,801,613	B2	8/2014	Al-Ali et al.	9,277,880	B2	3/2016	Poeze et al.
8,821,397	B2	9/2014	Al-Ali et al.	9,289,167	B2	3/2016	Diab et al.
8,821,415	B2	9/2014	Al-Ali et al.	9,295,421	B2	3/2016	Kiani et al.
8,830,449	B1	9/2014	Lamego et al.	9,307,928	B1	4/2016	Al-Ali et al.
8,831,700	B2	9/2014	Schurman et al.	9,323,894	B2	4/2016	Kiani
8,840,549	B2	9/2014	Al-Ali et al.	D755,392	S	5/2016	Hwang et al.
8,847,740	B2	9/2014	Kiani et al.	9,326,712	B1	5/2016	Kiani
8,849,365	B2	9/2014	Smith et al.	9,333,316	B2	5/2016	Kiani
8,852,094	B2	10/2014	Al-Ali et al.	9,339,220	B2	5/2016	Lamego et al.
8,852,994	B2	10/2014	Wojtczuk et al.	9,341,565	B2	5/2016	Lamego et al.
8,868,147	B2	10/2014	Stippick et al.	9,351,673	B2	5/2016	Diab et al.
8,868,150	B2	10/2014	Al-Ali et al.	9,351,675	B2	5/2016	Al-Ali et al.
8,870,792	B2	10/2014	Al-Ali et al.	9,364,181	B2	6/2016	Kiani et al.
8,886,271	B2	11/2014	Kiani et al.	9,368,671	B2	6/2016	Wojtczuk et al.
8,888,539	B2	11/2014	Al-Ali et al.	9,370,325	B2	6/2016	Al-Ali et al.
8,888,708	B2	11/2014	Diab et al.	9,370,326	B2	6/2016	McHale et al.
8,892,180	B2	11/2014	Weber et al.	9,370,335	B2	6/2016	Al-Ali et al.
8,897,847	B2	11/2014	Al-Ali	9,375,185	B2	6/2016	Al-Ali et al.
8,909,310	B2	12/2014	Lamego et al.	9,386,953	B2	7/2016	Al-Ali
8,911,377	B2	12/2014	Al-Ali	9,386,961	B2	7/2016	Al-Ali et al.
8,912,909	B2	12/2014	Al-Ali et al.	9,392,945	B2	7/2016	Al-Ali et al.
8,920,317	B2	12/2014	Al-Ali et al.	9,397,448	B2	7/2016	Al-Ali et al.
8,921,699	B2	12/2014	Al-Ali et al.	9,408,542	B1	8/2016	Kinast et al.
8,922,382	B2	12/2014	Al-Ali et al.	9,436,645	B2	9/2016	Al-Ali et al.
8,929,964	B2	1/2015	Al-Ali et al.	9,445,759	B1	9/2016	Lamego et al.
8,942,777	B2	1/2015	Diab et al.	9,466,919	B2	10/2016	Kiani et al.
8,948,834	B2	2/2015	Diab et al.	9,474,474	B2	10/2016	Lamego et al.
8,948,835	B2	2/2015	Diab	9,480,422	B2	11/2016	Al-Ali
8,965,471	B2	2/2015	Lamego	9,480,435	B2	11/2016	Olsen
8,983,564	B2	3/2015	Al-Ali	9,492,110	B2	11/2016	Al-Ali et al.
8,989,831	B2	3/2015	Al-Ali et al.	9,510,779	B2	12/2016	Poeze et al.
8,996,085	B2	3/2015	Kiani et al.	9,517,024	B2	12/2016	Kiani et al.
8,998,809	B2	4/2015	Kiani	9,532,722	B2	1/2017	Lamego et al.
9,028,429	B2	5/2015	Telfort et al.	9,538,949	B2	1/2017	Al-Ali et al.
9,037,207	B2	5/2015	Al-Ali et al.	9,538,980	B2	1/2017	Telfort et al.
9,060,721	B2	6/2015	Reichgott et al.	9,549,696	B2	1/2017	Lamego et al.
9,066,666	B2	6/2015	Kiani	9,554,737	B2	1/2017	Schurman et al.
9,066,680	B1	6/2015	Al-Ali et al.	9,560,996	B2	2/2017	Kiani
9,072,474	B2	7/2015	Al-Ali et al.	9,560,998	B2	2/2017	Al-Ali et al.
9,078,560	B2	7/2015	Schurman et al.	9,566,019	B2	2/2017	Al-Ali et al.
9,084,569	B2	7/2015	Weber et al.	9,579,039	B2	2/2017	Jansen et al.
9,095,316	B2	8/2015	Welch et al.	9,591,975	B2	3/2017	Dalvi et al.
9,106,038	B2	8/2015	Telfort et al.	9,622,692	B2	4/2017	Lamego et al.
9,107,625	B2	8/2015	Telfort et al.	9,622,693	B2	4/2017	Diab
9,107,626	B2	8/2015	Al-Ali et al.	D788,312	S	5/2017	Al-Ali et al.
9,113,831	B2	8/2015	Al-Ali	9,636,055	B2	5/2017	Al-Ali et al.
9,113,832	B2	8/2015	Al-Ali	9,636,056	B2	5/2017	Al-Ali
9,119,595	B2	9/2015	Lamego	9,649,054	B2	5/2017	Lamego et al.
9,131,881	B2	9/2015	Diab et al.	9,662,052	B2	5/2017	Al-Ali et al.
9,131,882	B2	9/2015	Al-Ali et al.	9,668,679	B2	6/2017	Schurman et al.
9,131,883	B2	9/2015	Al-Ali	9,668,680	B2	6/2017	Bruinsma et al.
9,131,917	B2	9/2015	Telfort et al.	9,668,703	B2	6/2017	Al-Ali
9,138,180	B1	9/2015	Coverston et al.	9,675,286	B2	6/2017	Diab
9,138,182	B2	9/2015	Al-Ali et al.	9,687,160	B2	6/2017	Kiani
9,138,192	B2	9/2015	Weber et al.	9,693,719	B2	7/2017	Al-Ali et al.
9,142,117	B2	9/2015	Muhsin et al.	9,693,737	B2	7/2017	Al-Ali
9,153,112	B1	10/2015	Kiani et al.	9,697,928	B2	7/2017	Al-Ali et al.
9,153,121	B2	10/2015	Kiani et al.	9,717,425	B2	8/2017	Kiani et al.
9,161,696	B2	10/2015	Al-Ali et al.	9,717,458	B2	8/2017	Lamego et al.
9,161,713	B2	10/2015	Al-Ali et al.	9,724,016	B1	8/2017	Al-Ali et al.
9,167,995	B2	10/2015	Lamego et al.	9,724,024	B2	8/2017	Al-Ali
9,176,141	B2	11/2015	Al-Ali et al.	9,724,025	B1	8/2017	Kiani et al.
9,186,102	B2	11/2015	Bruinsma et al.	9,743,887	B2	8/2017	Al-Ali et al.
9,192,312	B2	11/2015	Al-Ali	9,749,232	B2	8/2017	Sampath et al.
				9,750,442	B2	9/2017	Olsen
				9,750,461	B1	9/2017	Telfort
				2004/0034293	A1	2/2004	Kimball
				2004/0097797	A1	5/2004	Porges et al.

(56)

References Cited

U.S. PATENT DOCUMENTS

2005/0165316	A1	7/2005	Lowery et al.	2014/0142401	A1	5/2014	Al-Ali et al.
2006/0161054	A1	7/2006	Reuss et al.	2014/0163344	A1	6/2014	Al-Ali
2007/0282478	A1	12/2007	Al-Ali et al.	2014/0163402	A1	6/2014	Lamego et al.
2008/0009754	A1	1/2008	Chang	2014/0166076	A1	6/2014	Kiani et al.
2008/0071155	A1*	3/2008	Kiani A61B 5/02416 600/324	2014/0171763	A1	6/2014	Diab
2008/0221464	A1	9/2008	Al-Ali	2014/0180038	A1	6/2014	Kiani
2009/0062664	A1	3/2009	Chang et al.	2014/0180154	A1	6/2014	Sierra et al.
2009/0247984	A1	10/2009	Lamego et al.	2014/0180160	A1	6/2014	Brown et al.
2009/0275813	A1	11/2009	Davis	2014/0187973	A1	7/2014	Brown et al.
2009/0275844	A1	11/2009	Al-Ali	2014/0194709	A1	7/2014	Al-Ali et al.
2010/0004518	A1	1/2010	Vo et al.	2014/0194711	A1	7/2014	Al-Ali
2010/0030040	A1	2/2010	Poeze et al.	2014/0194766	A1	7/2014	Al-Ali et al.
2011/0001605	A1	1/2011	Kiani et al.	2014/0206963	A1	7/2014	Al-Ali
2011/0082711	A1	4/2011	Poeze et al.	2014/0213864	A1	7/2014	Abdul-Hafiz et al.
2011/0087083	A1	4/2011	Poeze et al.	2014/0243627	A1	8/2014	Diab et al.
2011/0105854	A1	5/2011	Kiani et al.	2014/0266790	A1	9/2014	Al-Ali et al.
2011/0125060	A1	5/2011	Telfort et al.	2014/0275808	A1	9/2014	Poeze et al.
2011/0208015	A1	8/2011	Welch et al.	2014/0275835	A1	9/2014	Lamego et al.
2011/0213212	A1	9/2011	Al-Ali	2014/0275871	A1	9/2014	Lamego et al.
2011/0230733	A1	9/2011	Al-Ali	2014/0275872	A1	9/2014	Merritt et al.
2011/0237911	A1	9/2011	Lamego et al.	2014/0275881	A1	9/2014	Lamego et al.
2011/0237969	A1	9/2011	Eckerbom et al.	2014/0276115	A1	9/2014	Dalvi et al.
2011/0288383	A1	11/2011	Diab	2014/0288400	A1	9/2014	Diab et al.
2011/0301444	A1	12/2011	Al-Ali	2014/0303520	A1	10/2014	Telfort et al.
2012/0041316	A1	2/2012	Al-Ali et al.	2014/0316217	A1	10/2014	Purdon et al.
2012/0046557	A1	2/2012	Kiani	2014/0316218	A1	10/2014	Purdon et al.
2012/0059267	A1	3/2012	Lamego et al.	2014/0316228	A1	10/2014	Blank et al.
2012/0088984	A1	4/2012	Al-Ali et al.	2014/0323825	A1	10/2014	Al-Ali et al.
2012/0165629	A1	6/2012	Merritt et al.	2014/0323897	A1	10/2014	Brown et al.
2012/0179006	A1	7/2012	Jansen et al.	2014/0323898	A1	10/2014	Purdon et al.
2012/0209082	A1	8/2012	Al-Ali	2014/0330092	A1	11/2014	Al-Ali et al.
2012/0209084	A1	8/2012	Olsen et al.	2014/0330098	A1	11/2014	Merritt et al.
2012/0227739	A1	9/2012	Kiani	2014/0330099	A1	11/2014	Al-Ali et al.
2012/0283524	A1	11/2012	Kiani et al.	2014/0333440	A1	11/2014	Kiani
2012/0296178	A1	11/2012	Lamego et al.	2014/0336481	A1	11/2014	Shakespeare et al.
2012/0319816	A1	12/2012	Al-Ali	2014/0343436	A1	11/2014	Kiani
2012/0330112	A1	12/2012	Lamego et al.	2014/0357966	A1	12/2014	Al-Ali et al.
2013/0018241	A1	1/2013	Bezzarides et al.	2014/0371548	A1	12/2014	Al-Ali et al.
2013/0023775	A1	1/2013	Lamego et al.	2014/0371632	A1	12/2014	Al-Ali et al.
2013/0041591	A1	2/2013	Lamego	2014/0378784	A1	12/2014	Kiani et al.
2013/0045685	A1	2/2013	Kiani	2015/0005600	A1	1/2015	Blank et al.
2013/0046204	A1	2/2013	Lamego et al.	2015/0011907	A1	1/2015	Purdon et al.
2013/0060147	A1	3/2013	Welch et al.	2015/0012231	A1	1/2015	Poeze et al.
2013/0096405	A1	4/2013	Garfio	2015/0018650	A1	1/2015	Al-Ali et al.
2013/0096936	A1	4/2013	SamPATH et al.	2015/0025406	A1	1/2015	Al-Ali
2013/0197328	A1	8/2013	Diab et al.	2015/0032029	A1	1/2015	Al-Ali et al.
2013/0211214	A1	8/2013	Olsen	2015/0038859	A1	2/2015	Dalvi et al.
2013/0243021	A1	9/2013	Siskavich	2015/0045637	A1	2/2015	Dalvi
2013/0253334	A1	9/2013	Al-Ali et al.	2015/0051462	A1	2/2015	Olsen
2013/0262730	A1	10/2013	Al-Ali et al.	2015/0080754	A1	3/2015	Purdon et al.
2013/0296672	A1	11/2013	O'Neil et al.	2015/0087936	A1	3/2015	Al-Ali et al.
2013/0296713	A1	11/2013	Al-Ali et al.	2015/0094546	A1	4/2015	Al-Ali
2013/0317370	A1	11/2013	Dalvi et al.	2015/0097701	A1	4/2015	Al-Ali et al.
2013/0324808	A1	12/2013	Al-Ali et al.	2015/0099950	A1	4/2015	Al-Ali et al.
2013/0331660	A1	12/2013	Al-Ali et al.	2015/0099951	A1	4/2015	Al-Ali et al.
2013/0331670	A1	12/2013	Kiani	2015/0099955	A1	4/2015	Al-Ali et al.
2013/0338461	A1	12/2013	Lamego et al.	2015/0101844	A1	4/2015	Al-Ali et al.
2014/0012100	A1	1/2014	Al-Ali et al.	2015/0106121	A1	4/2015	Muhsin et al.
2014/0034353	A1	2/2014	Al-Ali et al.	2015/0112151	A1	4/2015	Muhsin et al.
2014/0051953	A1	2/2014	Lamego et al.	2015/0116076	A1	4/2015	Al-Ali et al.
2014/0058230	A1	2/2014	Abdul-Hafiz et al.	2015/0126830	A1	5/2015	Schurman et al.
2014/0066783	A1	3/2014	Kiani et al.	2015/0133755	A1	5/2015	Smith et al.
2014/0077956	A1	3/2014	SamPATH et al.	2015/0140863	A1	5/2015	Al-Ali et al.
2014/0081100	A1	3/2014	Muhsin et al.	2015/0141781	A1	5/2015	Weber et al.
2014/0081175	A1	3/2014	Telfort	2015/0165312	A1	6/2015	Kiani
2014/0094667	A1	4/2014	Schurman et al.	2015/0196237	A1	7/2015	Lamego
2014/0100434	A1	4/2014	Diab et al.	2015/0201874	A1	7/2015	Diab
2014/0114199	A1	4/2014	Lamego et al.	2015/0208966	A1	7/2015	Al-Ali
2014/0120564	A1	5/2014	Workman et al.	2015/0216459	A1	8/2015	Al-Ali et al.
2014/0121482	A1	5/2014	Merritt et al.	2015/0230755	A1	8/2015	Al-Ali et al.
2014/0121483	A1	5/2014	Kiani	2015/0238722	A1	8/2015	Al-Ali
2014/0127137	A1	5/2014	Bellott et al.	2015/0245773	A1	9/2015	Lamego et al.
2014/0129702	A1	5/2014	Lamego et al.	2015/0245794	A1	9/2015	Al-Ali
2014/0135588	A1	5/2014	Al-Ali et al.	2015/0257689	A1	9/2015	Al-Ali et al.
				2015/0272514	A1	10/2015	Kiani et al.
				2015/0351697	A1	12/2015	Weber et al.
				2015/0351704	A1	12/2015	Kiani et al.
				2015/0359429	A1	12/2015	Al-Ali et al.
				2015/0366472	A1	12/2015	Kiani

(56)

References Cited

U.S. PATENT DOCUMENTS

2015/0366507 A1 12/2015 Blank
 2015/0374298 A1 12/2015 Al-Ali et al.
 2015/0380875 A1 12/2015 Coverston et al.
 2016/0000362 A1 1/2016 Diab et al.
 2016/0007930 A1 1/2016 Weber et al.
 2016/0029932 A1 2/2016 Al-Ali
 2016/0029933 A1 2/2016 Al-Ali et al.
 2016/0045118 A1 2/2016 Kiani
 2016/0051205 A1 2/2016 Al-Ali et al.
 2016/0058338 A1 3/2016 Schurman et al.
 2016/0058347 A1 3/2016 Reichgott et al.
 2016/0066823 A1 3/2016 Kind et al.
 2016/0066824 A1 3/2016 Al-Ali et al.
 2016/0066879 A1 3/2016 Telfort et al.
 2016/0072429 A1 3/2016 Kiani et al.
 2016/0073967 A1 3/2016 Lamego et al.
 2016/0081552 A1 3/2016 Wojtczuk et al.
 2016/0095543 A1 4/2016 Telfort et al.
 2016/0095548 A1 4/2016 Al-Ali et al.
 2016/0103598 A1 4/2016 Al-Ali et al.
 2016/0113527 A1 4/2016 Al-Ali et al.
 2016/0143548 A1 5/2016 Al-Ali
 2016/0166182 A1 6/2016 Al-Ali et al.
 2016/0166183 A1 6/2016 Poeze et al.
 2016/0166188 A1 6/2016 Bruinsma et al.
 2016/0166210 A1 6/2016 Al-Ali
 2016/0192869 A1 7/2016 Kiani et al.
 2016/0196388 A1 7/2016 Lamego
 2016/0197436 A1 7/2016 Barker et al.
 2016/0213281 A1 7/2016 Eckerbom et al.
 2016/0228043 A1 8/2016 O'Neil et al.
 2016/0233632 A1 8/2016 Scruggs et al.
 2016/0234944 A1 8/2016 Schmidt et al.
 2016/0267804 A1 9/2016 Pemba et al.
 2016/0270735 A1 9/2016 Diab et al.
 2016/0274572 A1 9/2016 Littrell et al.
 2016/0283665 A1 9/2016 Sampath et al.
 2016/0287090 A1 10/2016 Al-Ali et al.
 2016/0287786 A1 10/2016 Kiani
 2016/0296169 A1 10/2016 McHale et al.
 2016/0310052 A1 10/2016 Al-Ali et al.
 2016/0314260 A1 10/2016 Kiani
 2016/0324486 A1 11/2016 Al-Ali et al.
 2016/0324488 A1 11/2016 Olsen
 2016/0327984 A1 11/2016 Al-Ali et al.
 2016/0328528 A1 11/2016 Al-Ali et al.
 2016/0331332 A1 11/2016 Al-Ali
 2016/0367173 A1 12/2016 Dalvi et al.
 2017/0000394 A1 1/2017 Al-Ali et al.
 2017/0007134 A1 1/2017 Al-Ali et al.
 2017/0007190 A1 1/2017 Al-Ali et al.
 2017/0007198 A1 1/2017 Al-Ali et al.
 2017/0014083 A1 1/2017 Diab et al.
 2017/0021099 A1 1/2017 Al-Ali et al.
 2017/0027456 A1 2/2017 Kinast et al.
 2017/0042488 A1 2/2017 Muhsin
 2017/0055847 A1 3/2017 Kiani et al.
 2017/0055851 A1 3/2017 Al-Ali

2017/0055882 A1 3/2017 Al-Ali et al.
 2017/0055887 A1 3/2017 Al-Ali
 2017/0055896 A1 3/2017 Al-Ali et al.
 2017/0079594 A1 3/2017 Telfort et al.
 2017/0086723 A1 3/2017 Al-Ali et al.
 2017/0143281 A1 5/2017 Olsen
 2017/0147774 A1 5/2017 Kiani
 2017/0156620 A1 6/2017 Al-Ali et al.
 2017/0173632 A1 6/2017 Al-Ali
 2017/0187146 A1 6/2017 Kiani et al.
 2017/0188919 A1 7/2017 Al-Ali et al.
 2017/0196464 A1 7/2017 Jansen et al.
 2017/0196470 A1 7/2017 Lamego et al.
 2017/0202490 A1 7/2017 Al-Ali et al.
 2017/0224216 A1 8/2017 Al-Ali
 2017/0224231 A1 8/2017 Al-Ali
 2017/0224233 A1 8/2017 Al-Ali
 2017/0224262 A1 8/2017 Al-Ali
 2017/0231537 A1 8/2017 Al-Ali
 2017/0245790 A1 8/2017 Al-Ali et al.
 2017/0251974 A1 9/2017 Shreim et al.
 2017/0251975 A1 9/2017 Shreim et al.
 2017/0258403 A1 9/2017 Abdul-Hafiz et al.
 2017/0340293 A1 11/2017 Al-Ali et al.
 2017/0360310 A1 12/2017 Kiani et al.

OTHER PUBLICATIONS

Koppel, Robert I et al., "Effective of Pulse Oximetry Screening for Congenital Heart Disease in Asymptomatic Newborns", *Pediatrics*, 2003, vol. 111, No. 3, 451-455.

Hoke et al., "Oxygen saturation as a screening test for critical congenital heart disease: a preliminary study", *Pediatr Cardiol* 23:403-409, 2002.

PCT International Search Report and Written Opinion, dated Apr. 17, 2013, re: PCT Application No. PCT/US2013/020377, application dated Jan. 4, 2013, in 15 pages.

PCT Preliminary Report of Patentability, dated Jul. 8, 2014, for application No. PCT/US2013/020377 in 8 pages.

U.S. Appl. No. 13/907,638, Congenital Heart Disease Monitor, filed May 31, 2013.

U.S. Appl. No. 15/195,037, Automated CCHD Screening and Detection, filed Jun. 28, 2016.

U.S. Appl. No. 15/681,256, Automated Condition Screening and Detection, filed Aug. 18, 2017.

U.S. Appl. No. 15/634,502, Congenital Heart Disease Monitor, filed Jun. 27, 2017.

"Newborn Screening for CCHD, Answers and Resources for Primary Care Pediatricians", American Academy of Pediatrics, <http://www.aap.org/en-us/advocacy-and-policy/aap-health-initiatives/>, printed on Jun. 6, 2017, in 8 pages.

"Tools and Technology to Start Your CCHD Screening Program", Masimo Corporation, 2010-2011, in 16 pages.

"User Interface Elements", <http://www.usabilit.gov/how-to-and-tools/methods/user-interface-elements.html>, printed on Jun. 19, 2017, in 7 pages.

* cited by examiner




FIG. 1




FIG. 2
(PRIOR ART)




FIG. 3A




FIG. 3B





FIG. 3C






FIG. 5




FIG. 6




FIG. 7




FIG. 8




FIG. 9




FIG. 10A




FIG. 10B




FIG. 11A





FIG. 11B

Step 1: Pre-Ductal Instruction

FIG. 12A


1200 ↗



Step 2: Pre-Ductal Instruction

FIG. 12B


1220 ↗



Step 2: Sensor Error

FIG. 12C


1240 ↗



Step 3: Post-Ductal Instruction

FIG. 13A


1300 ↗



Step 4: Post-Ductal Measure

FIG. 13B

1320 ↗



Step 4: Sensor Error

FIG. 13C

1340 ↗




FIG. 14A

Step 5: CCHD Screen Results: Pass

1400

1405

step 5: screening results		
first screen	05-13-2012	05:42am
👶 pre-ductal ⓘ	96% SpO ₂	1.7PI
👶 post-ductal ⓘ	95% SpO ₂	1.8PI
Δ SpO ₂ difference ⓘ	1%	

negative screen: "pass" result
screen complete

close

FIG. 14B

Step 5: CCHD Screen Results: Conditional Fail (1)

1420

1425

step 5: screening results		
first screen	05-13-2012	05:42am
👶 pre-ductal ⓘ	96% SpO ₂	1.7PI
👶 post-ductal ⓘ	91% SpO ₂	1.8PI
Δ SpO ₂ difference ⓘ	5%	

positive screen: post-ductal SpO₂
less than 95%
perform second screen in one hour

close

FIG. 14C

Step 5: CCHD Screen Results: Conditional Fail (2)

1440

1445

step 5: screening results		
first screen	05-13-2012	06:42am
👶 pre-ductal ⓘ	96% SpO ₂	1.7PI
👶 post-ductal ⓘ	91% SpO ₂	1.8PI
Δ SpO ₂ difference ⓘ	5%	

positive screen: post-ductal SpO₂
less than 95%
perform third screen in one hour

close

FIG. 14D

Step 5: CCHD Screen Results: Fail (3)

1460

1465

step 5: screening results		
first screen	05-13-2012	07:42am
👶 pre-ductal ⓘ	96% SpO ₂	1.7PI
👶 post-ductal ⓘ	91% SpO ₂	1.8PI
Δ SpO ₂ difference ⓘ	5%	

positive screen: difference greater than 3%
rerer newborn for further medical evaluation

close




FIG. 15




FIG. 16

AUTOMATED CCHD SCREENING AND DETECTION

PRIORITY CLAIM AND RELATED APPLICATIONS

This application is a continuation of U.S. patent application Ser. No. 15/195,037, filed Jun. 28, 2016, and titled "AUTOMATED CCHD SCREENING AND DETECTION", which application is a divisional of U.S. patent application Ser. No. 13/733,782, filed Jan. 3, 2013, and titled "AUTOMATED CCHD SCREENING AND DETECTION", which application claims a priority benefit under 35 U.S.C. § 119 to the following U.S. Provisional Patent Applications:

Serial No.	Date	Title
61/583,143,	Jan. 4, 2012,	SYSTEMS AND METHODS AUTOMATING CCHD SCREENING AND DETECTION,
61/703,132,	Sep. 19, 2012,	SYSTEMS AND METHODS AUTOMATING CCHD SCREENING AND DETECTION.

Each of the foregoing disclosures is incorporated by reference herein in its entirety.

FIELD OF THE DISCLOSURE

The present application relates to the field of pulse oximetry. Specifically, the present application relates to the field of infant oxygen saturation monitoring and congenital heart defects screening.

BACKGROUND OF THE DISCLOSURE

Pulse oximetry screening can identify some critical congenital heart defects ("CCHDs"), which also are known collectively in some instances as critical congenital heart disease. CCHDs are structural heart defects that often are associated with hypoxia among infants during the newborn period. Infants with CCHDs are at risk for significant morbidity or mortality. There are several defects that could be considered CCHDs. However, in the context of newborn pulse oximetry screening at the time of preparation of the present application, the Centers for Disease Control and Prevention ("CDC") for the U.S. government classify seven (7) defects as CCHD: hypoplastic left heart syndrome, pulmonary atresia (with intact septum), tetralogy of Fallot, total anomalous pulmonary venous return, transposition of the great arteries, tricuspid atresia, and truncus arteriosus. According to the CDC, these seven CCHDs represent about seventeen to about thirty one percent (17-31%) of all congenital heart disease.

Patent ductus arteriosus ("PDA") is common in infants with several or more of the above seven (7) defects. In the developing fetus, the ductus arteriosus ("DA") **102** shown in FIG. **1** is the vascular connection between the pulmonary artery **106** and the aortic arch **104** that allows most of the blood from the right ventricle **110** to bypass the fetus' fluid-filled compressed lungs. During fetal development, this shunt protects the right ventricle **110** from pumping against the high resistance in the lungs **108**, which can lead to right ventricular failure if the DA **102** closes in-utero.

When the newborn takes its first breath, the lungs open and pulmonary vascular resistance decreases. In normal

newborns, the DA is substantially closed within twelve to twenty four (12-24) hours after birth, and is completely sealed after three (3) weeks.

In the case of PDA, high pressure oxygenated blood from the aorta **104** leaks or flows back into the pulmonary artery **112** and back to the lungs **108** with normal deoxygenated venous blood. The additional fluid returning to the lungs increases lung pressure to the point that the infant may have greater difficulty inflating the lungs. This uses more calories than normal and often interferes with feeding in infancy. Moreover, an open (patent) DA **102** alters the flow in the descending aorta **118**, which, as a result, changes the blood oxygen saturation in the feet.

Without screening, some newborns with CCHDs might be missed because the signs of CCHD might not be evident before an infant is discharged from the hospital after birth. Other heart defects might be considered secondary screening targets. Some of these heart defects can be just as severe as the primary screening targets and also require intervention soon after birth. These secondary targets include aortic arch atresia or hypoplasia, interrupted aortic arch, coarctation of the aorta, double-outlet right ventricle, Ebstein anomaly, pulmonary stenosis, atrioventricular septal defect, ventricular septal defect, and single ventricle defects (other than hypoplastic left heart syndrome and tricuspid atresia).

Current CDC recommendations focus on screening infants in the well-baby nursery and in intermediate care nurseries or other units in which discharge from the hospital is common during an infant's first week of life. At the time of preparation of the present application, the CDC promulgated a CCHD screening process **200** reproduced as FIG. **2**, directed toward oxygen saturation measurements, or percentages measured using, for example, a standard pulse oximeter.

According to the CDC's CCHD screening process **200** of FIG. **2**, a screen is considered positive (see box **228**) if (1) any oxygen saturation measurement is less than ninety percent (<90%) (in the initial screen or in repeat screens) (see boxes **206**, **214**, and **222**); (2) the oxygen saturation measurement is less than ninety five percent (<95%) in the right hand and foot on three measures (see boxes **208**, **216**, and **224**), each separated by one (1) hour (see boxes **204**, **212**, **220**); or (3) a greater than three percent (>3%) absolute difference exists in oxygen saturation measurements between the right hand and foot on three measures (see boxes **208**, **216**, and **224**), each separated by one (1) hour. Any screening that is greater than or equal to ninety five percent (≥95%) in the right hand or foot with a less than or equal to three percent (≤3%) absolute difference in oxygen saturation measurements between the right hand and foot is considered a negative screen and screening would end (see boxes **210**, **218**, **226**, and **230**).

The CDC recommends any infant receiving a positive screen receive a diagnostic echocardiogram, which would involve an echocardiogram within the hospital or birthing center, transport to another institution for the procedure, or use of telemedicine for remote evaluation. This can be expensive, disruptive, and possibly harmful to the infant. For example, at the time of preparation of the present application, an echocardiogram to verify an out-of-range (positive) screen could cost several hundred dollars.

Thus, false positives are to be avoided. The CDC believes that false positives are decreased if the infant is alert, and timing pulse oximetry screening around the time of newborn hearing screening improves efficiency.

Pulse oximetry screening may not detect all CCHDs, so it is possible for a baby with a negative screening result to still have a CCHD or other congenital heart defect.

SUMMARY OF THE DISCLOSURE

The CCHD screening process of FIG. 2 incorporates multiple measurement sites, such as, for example, a baseline site of the right hand and a secondary site of the feet or left hand. In some cases, a stereo pulse oximeter, such as the one disclosed in U.S. Pat. No. 6,334,065 (the '065 patent), incorporated by reference herein, could use two (2) or more of its associated sensors, one for each of multiple sites to accomplish the measurements. In fact, the '065 patent discusses the use of its sensor for multiple sites to determine indications of PDA, Persistent Pulmonary Hypertension in Neonates ("PPHN"), and Aortic Coarctation. However, very few if any institutions possess stereo pulse oximeters.

Rather, in most if not all circumstances, the CCHD screening process of FIG. 2 will be implemented on a single site pulse oximeter. The caregiver will apply the sensor to the first site, such as the baseline right hand, and take measurements. The caregiver will then remove the sensor from the first site, and transfer it to the second site, such as a foot, left hand, etc. and take measurements. As shown in FIG. 2, the CCHD screening process relies on the delta (or difference) between the two measurements.

Drawbacks may occur using the single sensor implementation. For example, there will be a time differential between the first baseline measurement and the second alternate site measurement when the caregiver changes sites. In infants, the parameters measurable with today's oximeters, including but not limited to oxygen saturation ("SpO₂"), vary within relatively short periods. This is exacerbated when infants are excited, crying, or otherwise agitated.

In the most straightforward scenario where the SpO₂ measurements are somewhat varying, the single sensor implementation of CCHD screening may determine, for example, an infant's SpO₂ during a valley or trough of varying SpO₂ values for the baseline measurement, and quite accidentally during a peak of varying SpO₂ values for the alternate measurement. Such time displaced measurements could appear anywhere on an infant's SpO₂ waveform. Thus, the differential between the baseline measurement and the alternate measurement, which is the key to determining positive or negative screenings under the CCHD screening process, could be subject to error. This is exacerbated as the CCHD screening process may require three (3) or more screenings before rendering a conclusion. Thus, a screening process may include measurements taken under different data conditions at each screen, and then again across the screens.

To overcome these and other drawbacks, the present disclosure includes systems and methods automating CCHD screening and detection. In an embodiment, a processor executes one or more processing modules to improve a likelihood that during a single sensor implementation of CCHD screening, the measurement values while time displaced, correspond to data conditions similar to one another. In addition, the processing module may determine the best sites for measurements.

In an embodiment, an oximeter or communicating monitor controls and tracks the implementation of the screening process, including instructions to caregivers on next steps. For example, a straightforward instruction may include "Attach Sensor to Right Hand," "Attach Sensor to Alternate Site," "Attach Sensor to Right Foot," "Attach Sensor to Left

Foot," "Calm Patient," "Adjust Sensor Positioning," or the like. The oximeter may also include a quality indicator providing information on the confidence in the screening measurements. A quality measure may be included for each measurement, for the entire screen, or the like. For example, the display may indicate "Positive Screen, 72% Confidence." In an embodiment, a minimum confidence threshold may be used to instruct a caregiver to repeat the measurements and/or restart the screening process. Moreover, the oximeter may produce an audio/visual alarm indicating time for a repeat screen, may accept patient information including a patient identifier, and the like.

In other embodiments, the oximeter may communicate with a host digital network or system to store or upload measurement data associated with a unique identifier to a remote processing center. That network or system may include multiple networks or systems. However, the oximeter may access previously stored information, such as, for example, earlier screening data stored at the remote network, to complete or increment the CCHD screening process. In an embodiment, a first network may be an institutional network such as a hospital data system, a cellular or other data system, or the like, wirelessly communicating with the oximeter or monitor. The system or systems eventually allow communication to a remote data server or processing center that stores the measurement information in a manner that provides for retrieval and appropriate association with newly acquired data.

For purposes of summarizing the disclosure, certain aspects, advantages and novel features thereof have been described herein. Of course, it is to be understood that not necessarily all such aspects, advantages or features will be embodied in any particular embodiment of the invention.

BRIEF DESCRIPTION OF THE DRAWINGS

The following drawings and the associated descriptions are provided to illustrate embodiments of the present disclosure and do not limit the scope of the claims.

FIG. 1 illustrates a diagram of a portion of a circulatory system including a heart having a patent or open ductus arteriosus.

FIG. 2 illustrates an embodiment of a CCHD screening process provided by the CDC.

FIG. 3A illustrates a simplified view of a pulse oximeter configured to perform a CCHD screening process, according to an embodiment of the disclosure.

FIG. 3B illustrates a simplified block diagram of a pulse oximeter configured to perform a CCHD screening process, according to an embodiment of the disclosure.

FIG. 3C illustrates an example CCHD screening process according to an embodiment of the present disclosure.

FIG. 4 illustrates a simplified perspective view and block diagram of a pulse oximeter configured to perform a CCHD screen process including dual sensors, according to an embodiment of present disclosure.

FIG. 5 illustrates a data diagram of a data sharing system, according to an embodiment of present disclosure.

FIG. 6 illustrates an example screen shot of a pulse oximeter monitor configured for measuring CCHD, according to an embodiment of the present disclosure.

FIG. 7 illustrates an example screen shot, including spot check settings, of a pulse oximeter monitor configured for measuring CCHD, according to an embodiment of the present disclosure.

FIGS. 8-9 illustrate example screen shots, including test results, of a pulse oximeter monitor configured for measuring CCHD, according to an embodiment of the present disclosure.

FIGS. 10A-B illustrate example screen shots, including alternative screening settings, of a pulse oximeter monitor configured for measuring CCHD, according to an embodiment of the present disclosure.

FIGS. 11A-B illustrate example screen shots, including additional settings, of a pulse oximeter monitor configured for measuring CCHD, according to an embodiment of the present disclosure.

FIGS. 12A-C, 13A-C, and 14A-D illustrate example screen shots, including instructions for single sensor operation, of a pulse oximeter monitor configured for measuring CCHD, according to an embodiment of the present disclosure.

FIG. 15 illustrates an example plot of pre- and post-ductal SpO₂ measurements in a CCHD screen process, according to an embodiment of the present disclosure.

FIG. 16 illustrates an example plot of pre-ductal perfusion index and confidence measurement in a CCHD screen process, according to an embodiment of the present disclosure.

DETAILED DESCRIPTION

The present disclosure generally relates to systems and methods automating critical congenital heart defects (“CCHDs”) screening and detection. In an embodiment, the CCHD screening process may be implemented on a single site pulse oximeter. A caregiver will apply the pulse oximeter sensor to the first site, such as the right hand, and take baseline measurements including, for example, blood oxygen saturation (“SpO₂”). The caregiver will then remove the pulse oximeter sensor from the first site, and transfer it to the second site, such as, for example, a foot or left hand, and take measurements. As shown in FIG. 2, and as described above, the CCHD screening process relies on the delta (or difference) between the two measurements. Then, a processor executes one or more processing modules to improve a likelihood that during a single sensor implementation of CCHD screening, the measurement values, while time displaced, correspond to data conditions similar to one another. In addition, the processing module may determine the best sites for measurements.

As used herein, the terms pulse oximeter, CCHD screening system, CCHD measurement device, and CCHD monitor may be used interchangeably.

FIG. 1 illustrates an infant heart 100 having patent ductus arteriosus (“PDA”), and lungs 108. The infant heart 100 includes a ductus arteriosus 102, an aorta 104, a pulmonary artery 106, a right ventricle 110, a main pulmonary artery 112, a left ventricle 114, an innominate artery 116, a descending aorta 118, and a subclavian artery 120. The ductus arteriosus 102 frequently fails to close in premature infants, allowing left-to-right shunting, i.e. oxygenated “red” blood flows from the aorta 104 to the now unobstructed pulmonary artery 106 and recirculates through the lungs 108. A persistent patent results in pulmonary hyperperfusion and an enlarged right ventricle 110, which leads to a variety of abnormal respiratory, cardiac and genitourinary symptoms. Current PDA diagnosis involves physical examination, chest x-ray, blood gas analysis, echocardiogram, or a combination of the above. For example, large PDAs may be associated with a soft, long, low-frequency murmur detectable with a stethoscope. As another example, two-

dimensional, color Doppler echocardiography may show a retrograde flow from the ductus arteriosus 102 into the main pulmonary artery 112. Once a problematic PDA is detected, closure can be effected medically with indomethacin or ibuprofen or surgically by ligation. Multiple doses of indomethacin are commonplace but can still result in patency, demanding ligation. A drawback to current diagnostic techniques is that clinical symptoms of a PDA can vary on an hourly basis, requiring extended and inherently intermittent testing.

In a single sensor implementation of noninvasive PDA determination or CCHD screen (as described below in referenced to FIGS. 3A-C), a sensor, such as a blood oxygenation sensor, may be placed on the right hand to determine a baseline of physiological data. For example, a pulse oximetry sensor at the right hand provides physiological data signals indicative of arterial oxygen saturation and a plethysmograph for blood circulating from the left ventricle 114 through the innominate artery 116, which supplies the right subclavian artery leading to the right arm. Because the innominate artery 116 is upstream from the PDA shunt at the ductus arteriosus 102, the oxygen saturation value and plethysmograph waveform obtained from the right hand are relatively unaffected by the shunt and serve as a baseline for comparison with readings from other tissue sites.

A sensor may then be placed on a foot to provide oxygen status for blood supplied from the descending aorta 118. The shunt at the ductus arteriosus 102 affects aortic flow. In particular, the shunt allows a transitory left-to-right flow during systole from the high pressure aorta 104 to the low pressure pulmonary artery 106 circulation. This left-to-right flow through the shunt at the ductus arteriosus 102 alters the flow in the descending aorta 118 and, as a result, affects the oxygen saturation value and plethysmograph waveform measured at the foot. The PDA condition, therefore, may be manifested as a normal plethysmograph with a characteristically narrow peak and well-defined dicrotic notch at the right-hand baseline site compared with a damped plethysmograph with a broadened peak and reduced or missing notch at the foot site. Further, the foot site waveform may be phase shifted from the baseline waveform. These plethysmograph differences are accompanied by comparable differences in arterial oxygen saturation values between the right-hand site and the foot site.

As an alternative, the sensor may be placed on the left hand to provide oxygen status for blood circulating from the left ventricle through the left subclavian artery 120 that supplies the left arm. Because the left subclavian artery 120 is nearer the shunt at the ductus arteriosus 102 than the further upstream innominate artery 116, it may experience some alteration in flow due to the shunt at the ductus arteriosus 102. The PDA condition, therefore, may also be manifested as an altered plethysmograph waveform at a left hand site as compared with the right hand baseline site, although likely to a lesser degree than with a foot site. Thus, the PDA condition, and thus a CCHD condition, can be detected and its treatment monitored from a delta in saturation (i.e., difference in SpO₂) values and plethysmograph morphology and phase comparisons between a right hand baseline sensor site and one or more other sites, such as the left hand or foot. One of ordinary skill will recognize that multiple site comparisons using an oximeter may also be used to detect other cardiac abnormalities that cause mixing of oxygenated and deoxygenated blood, such as a ventricular hole or a patent foramen. Further, abnormal mixing of oxygenated and deoxygenated blood may also be manifested

in physiological data measurements other than oxygen saturation provided by an advanced patient monitor or pulse oximeter.

FIG. 3A illustrates a simplified view of a pulse oximeter configured to perform a CCHD screening process with a single sensor (as described above), according to an embodiment of the disclosure. FIG. 3A includes a pulse oximeter 300, a display 302, and a sensor 304. In an embodiment, the sensor 304 may be noninvasively attached to the patient's finger. The sensor 304 may measure various blood analytes noninvasively using multi-stream spectroscopy. In an embodiment, the multi-stream spectroscopy may employ visible, infrared and near infrared wavelengths. The sensor 304 may be capable of noninvasively measuring blood analytes or percentages thereof (e.g., saturation) based on various combinations of features and components.

The sensor 304 may include photocommunicative components, such as an emitter, a detector, and other components. The emitter may include a plurality of sets of optical sources that, in an embodiment, are arranged together as a point source. The various optical sources may emit a sequence of optical radiation pulses at different wavelengths towards a measurement site, such as a patient's finger. Detectors may then detect optical radiation from the measurement site. The optical sources and optical radiation detectors may operate at any appropriate wavelength, including infrared, near infrared, visible light, and ultraviolet. In addition, the optical sources and optical radiation detectors may operate at any appropriate wavelength, and such modifications to the embodiments desirable to operate at any such wavelength will be apparent to those skilled in the art. In some embodiments, the sensor 304 may be any of a disposable, reusable, and/or resposable sensor. Generally, for CCHD measurements, a sensor configured for use with an infant is desirable. In some embodiments, this may include a finger, toe, or ear sensor. In an embodiment, the sensor 304 may also be a wrist-type sensor configured to surround the wrist or ankle of an infant.

The sensor 304 is coupled to the pulse oximeter 300 that processes and/or displays the sensor 304's output, on, for example, display 302. The sensor 304 may additionally be coupled to one or more monitors that process and/or display the sensor 304's output. As described below in reference to FIG. 3B, the pulse oximeter 300 may include various components, such as a sensor front end, a signal processor, and/or a display, among other things.

The sensor 304 may be integrated with a monitor (such as the pulse oximeter 300), for example, into a handheld unit including the sensor 304, a display and user controls. In other embodiments, the sensor 304 may communicate with one or more processing devices. The communication may be through wire(s), cable(s), flex circuit(s), wireless technologies, or other suitable analog or digital communication methodologies and devices to perform those methodologies. Many of the foregoing arrangements allow the sensor 304 to be attached to the measurement site while the device (such as the pulse oximeter 300) is attached elsewhere on a patient, such as the patient's arm, or placed at a location near the patient, such as a bed, shelf or table. The sensor 304 and/or pulse oximeter 300 may also provide outputs to a storage device or network interface.

FIG. 3B illustrates a simplified block diagram of a pulse oximeter configured to perform a CCHD screening process, as described above with reference to FIG. 3A. FIG. 3B includes the pulse oximeter 300, the sensor 304, and a communications link 322. The pulse oximeter 300 includes a front-end interface 324, a signal processor 326, a user

interface processor 330, the display 302, a storage 334, a network interface 336, and an optional memory 328. In an embodiment, the signal processor 326 includes processing logic that determines measurements, for example, SpO₂ and/or PI. The signal processor 326 may be implemented using one or more microprocessors or subprocessors (e.g., cores), digital signal processors, application specific integrated circuits (ASICs), field programmable gate arrays (FPGAs), combinations of the same, and the like.

The signal processor 326 may provide various signals through the front-end interface 324 and the communications link 322 that control the operation of the sensor 304. For example, the signal processor 326 may provide an emitter control signal to the sensor 304. Additionally, measurement data may be transmitted from the sensor 304 to the signal processor 326. As also shown, the optional memory 328 may be included in the front-end interface 324 and/or in the signal processor 326. This optional memory 328 may serve as a buffer or storage location for the front-end interface 324 and/or the signal processor 326, among other uses.

The user interface processor 330 may provide an output, for example, on the display 302, for presentation to a user of the pulse oximeter 300. The user interface processor 330 and/or the display 302 may be implemented as a touch-screen display, an LCD display, an organic LED display, or the like. In addition, the user interface processor 330 and/or display 302 may include a flip screen, a screen that can be moved from one side to another on the pulse oximeter 300, or may include an ability to reorient its display indicia responsive to user input or device orientation. In alternative embodiments, the pulse oximeter 300 may be provided without the display 302 and may simply provide an output signal to a separate display or system.

The storage 334 and the network interface 336 represent other optional output connections that can be included in the pulse oximeter 300. The storage 334 may include any computer-readable medium, such as a memory device, hard disk storage, EEPROM, flash drive, or the like. Various software and/or firmware applications can be stored in the storage 334, which may be executed by the signal processor 326 and/or another processor of the pulse oximeter 300. The network interface 336 may be a serial bus port (RS-232/RS-485), a Universal Serial Bus (USB) port, an Ethernet port, a wireless interface (e.g., WiFi such as any 802.1x interface, including an internal wireless card), or other suitable communication device(s) that allows the pulse oximeter 300 to communicate and share data with other devices (such as, for example, a remote data processing center as described below in reference to FIG. 5). The pulse oximeter 300 may also include various other components not shown, such as a microprocessor, graphics processor, and/or controller to output a user interface, to control data communications, to compute data trending, and/or to perform other operations.

Although not shown in the depicted embodiment, the pulse oximeter 300 may include various other components or may be configured in different ways. For example, the sensor 304 may measure additional advanced parameters. As described below, the pulse oximeter 300 may prompt the user to take specific actions through the display 302 to, for example, accomplish the CCHD screening process.

FIG. 3C illustrates an example CCHD screening process according to an embodiment of the present disclosure. This CCHD screening process may be accomplished with, for example, the pulse oximeter 300 of FIGS. 3A and 3B. At box 362, sensor data is gathered at a first site. For example, at this point the user/caregiver may be instructed to attach the pulse oximeter sensor (for example, the sensor 304) to the

patient's right hand for a pre-ductal measurement. At box 364, sensor data is gathered at a second site. For example, at this point the user/caregiver may be instructed to attach the pulse oximeter sensor (for example, the sensor 304) to the patient's foot or left hand for a post-ductal measurement. In each of boxes 362 and 364, the measurement data may be transmitted from the sensor 304, through the communications link 322 and front-end interface 324, to the signal processor 326.

At box 366, the gathered pre- and post-ductal measurement data is processed by, for example, the signal processor 326. The gathered data is processed to provide both screening results (for example, the pre- and post-ductal SpO₂ measurements and the delta between the two), but also to reduce errors in the measurements. As stated above, drawbacks arise when data from each time-displaced measurement (baseline and alternate, pre- and post-ductal) is not carefully selected. For example, normal infant SpO₂ values may drift up and down more than two percent (2%). If a baby holds its breath during onset of crying, values may drift more than twenty percent (20%) in a very short time. If by chance the baseline was taken at a coincidental peak in normal variation, and the alternate was taken during crying or even at a valley or trough of normal variation, such error may impact a CCHD screening process, which, as shown in FIG. 2, may change the result based on a three percent (3%) differential between baseline and alternate.

As described above, FIG. 2 depicts the Centers for Disease Control and Prevention's ("CDC's") recommended CCHD screening process 200. A screen is considered positive (see box 228) if (1) any oxygen saturation measurement is less than ninety percent (<90%) (in the initial screen or in repeat screens) (see boxes 206, 214, and 222); (2) the oxygen saturation measurement is less than ninety five percent (<95%) in the right hand and foot on three measures (see boxes 208, 216, and 224), each separated by one (1) hour (see boxes 204, 212, 220); or (3) a greater than three percent (>3%) absolute difference exists in oxygen saturation measurements between the right hand and foot on three measures (see boxes 208, 216, and 224), each separated by one (1) hour. Any screening that is greater than or equal to ninety five percent ($\geq 95\%$) in the right hand or foot with a less than or equal to three percent ($\leq 3\%$) absolute difference in oxygen saturation measurements between the right hand and foot is considered a negative screen and screening would end (see boxes 210, 218, 226, and 230).

Referring back to FIG. 3C and box 366, in an embodiment, a process or processes executing on one or more signal processors (such as signal processor 326) seek to reduce errors introduced by, for example, the time differential in measurements in the single sensor implementation of CCHD screening processes. In an embodiment, confidence thresholds may indicate when signal quality is sufficiently high to use measurement data. Confidence information may also be used to weight data as it is combined for comparisons. Confidence information may be assigned to windows of data and may be used to reduce an impact of a particular window on a measurement value or adjust the value itself. Such confidence information may be advantageously incorporated into a visual queue, such as a number or a graphic, indicating a confidence in the measurement or screening itself. An embodiment of a confidence measurement is displayed in FIG. 16. FIG. 16 shows an example plot 1600 of a pre-ductal PI graph 1605 and a pre-ductal confidence graph 1610, each versus time (in seconds). The pre-ductal PI graph 1605 will be described below. Pre-ductal confidence is indicated by pre-ductal confidence graph 1610, which varies over time

based on, for example, signal quality from the sensor 304 and/or PI measurements as shown in the pre-ductal PI graph 1605. In an embodiment, a sufficiently low confidence level (and/or threshold) results in a measurement error, and no result is reported to the user/caregiver. For example, a confidence threshold may be (on a scale of 0 to 1) 0.5, 0.6, 0.7, 0.8, 0.85, 0.9, 0.95, 0.96, 0.97, 0.98, 0.99, and/or 1, among other possibilities.

Referring back to box 366 in FIG. 3C, additional processing may be accomplished to ensure the measured data is reliable, and/or the user/caregiver may be notified of possible problems with the gathered data. Alternatively, the gathered data may be processed after it is gathered. For example, the sensor data at site one may be gathered and processed, followed by the gathering and processing of the sensor data of site two.

In an embodiment, indications of motion may advantageously cause an audio/visual message to be presented to a caregiver to calm the patient before measurements can be used. In an embodiment, a minimum wait time may ensure that actual stabilization of the data occurs, such as for example, about ten (10) to twenty (20) seconds or more may need to pass after an indication of cessation of motion in the patient.

In an embodiment, the features of the plethsmographic data or of the oxygen saturation data values can be analyzed to determine when to use measurement data. In the case of the plethsmographic data, determination of how well the waveform fits a model may guide confidence measurements or indicate signal noise. With saturation data, troughs and peaks, and their respective severities may be determined so that measurements for each site are chosen during similar or the same waveform feature, such as, for example, using measurements that correspond to peaks for each site. In an embodiment, the natural high and low cycles of oxygen saturation are used to correlate the measurements. For example, typical pre- and post-ductal SpO₂ measurements are illustrated in FIG. 15. FIG. 15 shows an example plot 1500, including a pre-ductal SpO₂ graph 1502 and a post-ductal SpO₂ graph 1504, each versus time (in seconds). The SpO₂ measurements vary cyclically over time. Thus, in order to compare measurements from two different sites taken at different times, the natural cycles of variance in the SpO₂ are used to correlate the measurement data. In the example of FIG. 15, a 60 second window of data is taken of both the pre- and post-ductal measurement. The data is time adjusted so that the natural cycles of the SpO₂ measurements occur at roughly the same time. A window or timeframe, for example, a 30 second window of time, is analyzed, indicated as 1510 and 1512, respectively, in FIG. 15. This window or timeframe is then used to determine the SpO₂ delta 1514. As a result of correlating the pre- and post-ductal SpO₂ measurements using the natural cycles, a more accurate delta can be determined.

Referring back to box 366 in FIG. 3C, in an embodiment, additional processing may be performed and/or intermediate data may be accessed to assist in determining which measurements to select for screening deltas. For example, often sophisticated filters are used to smooth saturation measurements. In an embodiment, pre-filtered results may provide more accurate or immediate information on measurement selection. Similarly, ratio or other information may be used in addition to or in place of focusing on oxygen saturation values.

In an embodiment, phases in the respiration cycle may be accounted for to select measurement data. For example, measurement values may correspond to only data during, for

example, the inspiration phase, or the like. In an embodiment, respiration or pulse rates may qualify or disqualify measurement data, based on, for example, rate stability or the like.

Other parameter information may also be used. For example, perfusion index (“PI”) information may provide indicators on when to select measurement data. In an embodiment, PI may vary for reasons unrelated to CCHD and therefore can be used in certain implementations, such as ranges that qualify or disqualify measurement data or the like. For example, the perfusion index may indicate that signal quality is sufficiently high to use measurement data. A PI measurement is also illustrated in FIG. 16 as pre-ductal PI graph 1605. Although both confidence and PI have been illustrated with respect to a pre-ductal measurement in FIG. 16, it is to be understood that a similar PI and confidence measure may also be determined for a post-ductal measurement.

In box 368 of FIG. 3C, the pulse oximeter 300 may optionally communicate with a remote data processing center, through, for example, the network interface 336. As described below in reference to FIG. 5, current measurement data may be transmitted to the remote data processing center, and/or previous measurement data may be accessed from the remote data processing center, among other things. Thus, the results of multiple measurements may be stored, retrieved, and/or compared as part of the CCHD screening process. Alternatively, measurement data may be stored at the pulse oximeter 300. In an embodiment, the data processing of box 366 takes place at the remote data processing center.

In box 370, the results of the CCHD screening are reported to the caregiver/patient/user. This may be accomplished, for example, by displaying the results on the display 302 as a number, color, and/or other symbol, and/or aurally in the form of, for example, an alarm.

An artisan will recognize from the disclosure herein a wide variety of indicators or combinations of indicators for determining when to select measurement data for use in CCHD screening processes. For example, segments or whole windows of data for the various parameters and indicators discussed in the foregoing may be combined to provide additional insight into measurement selection. Moreover, any or combinations of the foregoing may be used to adjust a particular measurement instead of seeking a different measurement.

FIG. 4 illustrates an embodiment of the CCHD screening system incorporating dual (or stereo) sensors. FIG. 4 includes a pulse oximeter 400, a signal processing device 406, a first sensor 408, and a second sensor 409. Further, the pulse oximeter 400 includes a display 402, a communication port 403, and a portable oximeter 404. Such a pulse oximeter is commercially available from Masimo Corporation of Irvine, Calif. Other CCHD systems may employ more than two sensors. In an embodiment, the pulse oximeter 400 includes a USB or other communication port 403 that connects to the signal processing device 406. In this embodiment, each of the dual sensors (first sensor 408 and second sensor 409) connects to the signal processing device 406. Either of the first sensor 408 or the second sensor 409, or both the first sensor 408 and the second sensor 409, may comprise any combination of disposable, reusable, and/or resposable sensors. As described above, the first sensor 408 and/or the second sensor 409 may be configured for use with an infant. Additionally, the two sensors are not required to be of the same type, but may be any combination of suitable sensors. For example, in some embodiments, the first sensor 408 and/or the second sensor 409 may include finger, toe, or

ear sensors, or any combination of finger, toe, or ear sensors. In another example, the first sensor 408 and/or the second sensor 409 may be a wrist-type sensor configured to surround the wrist or ankle of an infant. The signal processing device 406 provides stereo processing of sensor signals, such as those disclosed in the ’065 patent. In an embodiment, such processing may include separate oximeter calculators logically similar to having two independent oximeters, one for each sensor. The pulse oximeter 400 may include components similar to those of the pulse oximeter 300 of FIG. 3B. Specifically, the pulse oximeter 400 may include a signal processor, among other things. Processing of sensor data may take place in the pulse oximeter 400 and/or the signal processing device 406.

In this stereo embodiment of two (2) or more sensors, while there may not necessarily be time differentials between measurements, the data from each sensor may be of varying quality. Thus, many of the same procedures disclosed in the foregoing will apply. For example, at a particular time, the data from the baseline sensor may be clean with a high confidence while the data from the alternate sensor may have low quality from motion artifact, such as, for example, an infant kicking but not moving their right hand. Thus, the signal processing device 406 may use the foregoing processes to select measurement values from each sensor at different times. In such cases, determining which measurements should be used involves determinations similar to those used in the single sensor implementation. Alternatively, the signal processing device 406 may wait and select a time when both sensors produce usable measurement data, similar data conditions, or the like.

In an embodiment, the signal processing device 406 determines measurements for each of first sensor 408 and second sensor 409, and forwards measurement values to the pulse oximeter 400. The monitor 400 advantageously includes CCHD screening modules that guide a caregiver through the screening process. In other embodiments, the signal processing device 406 executes the screening and sends flags or messages to the display 402 and/or the portable oximeter 404 directing the display 402 and/or the portable oximeter 404 to display caregiver instructions and/or output results.

Although disclosed as the processing device 406 separate from the pulse oximeter 400, an artisan will recognize from the disclosure herein that the processing of the signal processing device 406 may be incorporated into the pulse oximeter 400 and/or the portable oximeter 404.

FIG. 5 illustrates a data diagram of a data sharing system, according to an embodiment of present disclosure. FIG. 5 includes a patient monitor 502, a remote data processing center 504, and a network 506. In an embodiment, the patient monitor 502 may include, for example, either the pulse oximeter 300 of FIG. 3A or the pulse oximeter 400 of FIG. 4. In an embodiment, the patient monitor 502 and the network 506 may communicate with one another, and may be located in a hospital, while both the patient monitor 502 and the network 506 may advantageously communicate with the remote data processing center 504. The network 506 may include, for example, Masimo’s Patient SafetyNet System, a hospital patient data system, and/or any other wired or wirelessly connected system. An artisan will recognize that communications among any of the patient monitor 502, the remote data processing center 504, and the network 506 may be through any appropriate wired and/or wireless data transmission, and may use any suitable communications protocol. For example, communication may be serial or parallel, through Universal Serial Bus (USB) (wired or

wireless), Ethernet, Bluetooth, Near Field Communications (NFC), radio frequency (RF), infrared, and/or WiFi (such as any 802.1x interface), among others as is known in the art.

In an embodiment, the remote data processing center **504** may store, for example, Patient ID's, device information (such as, for example, patient monitor **502** device information), sensor information, measurement data, screening data, and/or screening determinations for comparisons with later screening events. For example, the CCHD screening process of FIG. **2** includes multiple repeat screens. In this example, previous screening data for a particular patient, among other things, may be requested by the patient monitor **502**, when for example, a new screening is performed. Thus, the data from the two screening may be compared or used in some other way. Such communication may advantageously be wirelessly directed from the patient monitor **502** to the remote data processing center **504**, or through one or more intermediary networks (such as the network **506**). For example, the patient monitor **502** may include wireless communication to a hospital or other network which eventually communicates with the remote data processing center **504**.

In an embodiment, the pulse oximeter **300** may communicate with network **506** (or other host digital network or system) to store or upload measurement data associated with a unique identifier to remote data processing center **504**. The network **506** may include multiple networks or systems. The pulse oximeter **300** may access previously stored information, such as, for example, earlier screening data stored at the remote data processing center **504** of a remote network, to complete or increment the CCHD screening process. In an embodiment, a first network may be an institutional network such as a hospital data system, a cellular or other data system, or the like, wirelessly communicating with the pulse oximeter **300** or monitor. The system or systems eventually allow communication to a remote data processing center **504** or other processing center that stores the measurement information in a manner that provides for retrieval and appropriate association with newly acquired data.

As will be described in detail below, FIGS. **6-9**, **10A-B**, **11A-B**, **12A-C**, **13A-C**, and **14A-D** illustrate example screen shots and/or displays of a pulse oximeter monitor (such as the pulse oximeter **300** of FIG. **3A** or the pulse oximeter **400** of FIG. **4**) and/or other CCHD measurement device configured for measuring CCHD, according to embodiments of the present disclosure.

FIG. **6** illustrates a screen shot **600** of a pulse oximeter monitor configured for measuring CCHD. The screen shot **600** includes a first screen **602**, second screen **604**, and third screen **606** for measuring CCHD. These three screens correlate to the three separate measurement recommendations for measuring CCHD, as described in reference to FIG. **2**. A touch screen can be provided in the pulse oximeter monitor so that a caregiver need only touch each measurement (for example, first screen **602**, second screen **604**, and third screen **606**) in succession to begin the three different measurement processes.

FIG. **7** illustrates various spot check settings that are available on the pulse oximeter monitor and/or CCHD measurement device of the present disclosure. A SpO₂ delta **702** allows a caregiver to dictate a specific difference in SpO₂ between measurement sites that will indicate a CCHD problem. The value can range between 1 and 10 with a default value of 3. The low SpO₂ threshold **704** indicates a minimal passing SpO₂ value for a test result. The value can range between 85-100 with a default value of 95. The all mute **706** mutes all noise from the pulse oximeter monitor

and/or CCHD measurement device. The low PI threshold **708** indicates a minimum PI measurement value that may still indicate a valid test result. This value can range between 0.1 and 1 with a default value of 0.7.

FIG. **8** illustrates a display of a first test results page **800**. The first test results page **800** includes a first test results indication **801**. In the embodiment of first test results page **800**, a positive result was obtained, but it was at a saturation value of less than 95%. The first test results page **800** also includes the time of the test **802**, the pre-ductal results **804** and post-ductal results **810**, including the respective SpO₂ values **806**, **812** and PI values **808** and **814**, as well as the delta SpO₂ **816** illustrating the delta SpO₂ value **818**. The first test results page **800** also includes a next step indication **820**. For example, in the embodiment of FIG. **8**, the next step may indicate that a second screen test is to be performed in an hour. Alternatively, the next step may indicate that no further testing is required (in the event of, for example, a negative screen).

FIG. **9** illustrates another example of a first test results page **900**. The test results page has similar features corresponding to that of FIG. **8** shown in an alternative format. The first test results page **900** includes a first test results indication **901**. In the embodiment of first test results page **900**, a positive result was obtained. The first test results page **900** also includes the time of the test **902**, the pre-ductal results **904** and post-ductal results **910**, including the respective SpO₂ values **906**, **912** and PI values **908** and **914**, as well as the delta SpO₂ **916** illustrating the delta SpO₂ value **918**. The first test results page **900** also includes a next step indication **920**.

FIGS. **10A** and **10B** illustrate additional alternative embodiments of settings available for the CCHD monitor of the present disclosure. Referring to FIG. **10A**, screen **1000** includes settings which indicate the difference in SpO₂ readings (SpO₂ difference **1002**) between sites to indicate a CCHD problem. The screen settings also include passing SpO₂ low limit **1004**, positive result SpO₂ threshold **1006**, and enable PI **1008**. Referring to FIG. **10B**, additional settings on screen **1020** include positive result SpO₂ threshold **1022**, enable PI **1024**, positive result PI threshold **1026**, and settings defaults **1028**. The settings defaults can include different protocols for measuring CCHD including the Kemper Protocol.

FIG. **11A** illustrates additional advanced settings in screen **1100**, including a time window **1102**, a correlation window **1104**, a minimum confidence **1106**, and minimum PI **1108**. In FIG. **11B** other settings are illustrated in screen **1120**, including minimum confidence **1122**, minimum PI **1124**, minimum correlation **1126**, and mean trim **1128**.

Once all of the settings are configured, the CCHD screening device of the present disclosure provides step by step instructions for performing a CCHD test protocol. FIGS. **12A-C**, **13A-C**, and **14A-D** illustrate instructions provided for the single sensor CCHD monitor of the present disclosure. FIGS. **12A-C** illustrate an embodiment of various protocol step-by-step instructions. For the pre-ductal measurement, instruction screen **1200** instructs a caregiver to place the sensor on the right hand and press a button on the screen to move to the next instruction screen **1220**. Instruction screen **1220** indicates that a measurement is taking place. If an error occurs in the measurement, an error screen **1240** is displayed. In an embodiment, the CCHD screening system (including, for example, pulse oximeter **300** or pulse oximeter **400**) controls and tracks the implementation of the screening process, including instructions to caregivers on next steps, as described. Additional examples of instruction

may include “Attach Sensor to Right Hand,” “Attach Sensor to Alternate Site,” “Attach Sensor to Right Foot,” “Attach Sensor to Left Foot,” “Calm Patient,” “Adjust Sensor Positioning,” or the like.

Once the pre-ductal measurement is obtained, the instructions move on to the instruction screens of FIGS. 13A-C. Instruction screen 1300, labeled as step 3, instructs the caregiver to place the sensor on the either foot of the patient and press a button to continue to instruction screen 1320. In instruction screen 1320, labeled as step 4, the post-ductal measurement is obtained. If an error occurs in the post-ductal measurement, an error screen 1340 is presented to the user.

The instructions continue with FIGS. 14A-D. Once the post-ductal measurement is obtained, results screen 1400, labeled as step 5, is displayed. The results screen 1400 provides the pre- and post-ductal screening results, as well as the difference in SpO₂. The results screen 1400 illustrates an embodiment in which a passing test was obtained and instructions that the test was passed are provided on the screen at 1405. Alternatively, if a potential problem was found, as illustrated in results screen 1420 of FIG. 14B, then an instruction is provided at 1425 instructing the care giver to perform a second test in one hour. If a second screening indicates a failing test, as illustrated in results screen 1440 of FIG. 14C, then an instruction is provided to perform a third screen testing in one hour at 1445. If the results of the third test also indicate a problem, for example in screen 1460 of FIG. 14D, then an instruction is provided to refer the newborn for further medical evaluation at 1465. As described above, the CCHD screening system may include a quality indicator providing information on the confidence in the screening measurements. A quality measure may be included for each measurement, for the entire screen, or the like. For example, the display may indicate “Positive Screen, 72% Confidence.” In an embodiment, a minimum confidence threshold may be used to instruct a caregiver to repeat the measurements and/or restart the screening process. Moreover, the oximeter may produce an audio/visual alarm indicating time for a repeat screen, may accept patient information including a patient identifier, and the like.

Although the foregoing has been described in terms of certain preferred embodiments, other embodiments will be apparent to those of ordinary skill in the art from the disclosure herein. For example, other CCHD screening methodologies may take advantage of the processes for matching measurements disclosed herein. Moreover, data conditions from one screen may influence when measurements are chosen for subsequent screens. For example, if choosing according to peaks in the SpO₂ values was implemented to match measurement conditions in one screen, the same may be used or implemented in subsequent screens. Accordingly, the present disclosure is not intended to be limited by the reaction of the preferred embodiments, but is to be defined by reference to the appended claims.

In addition to the foregoing, all publications, patents, and patent applications mentioned in this specification are herein incorporated by reference to the same extent as if each individual publication, patent, or patent application was specifically and individually indicated to be incorporated by reference.

Moreover, the oximeters discussed in the foregoing may include many or all of the features of basic pulse oximeters that determine measurements for blood oxygen saturation (“SpO₂”), pulse rate (“PR”) and plethysmographic information, to read-through-motion oximeters, to co-oximeters that determine measurements of many constituents of circulating

blood. For example, Masimo Corporation of Irvine Calif. (“Masimo”) manufactures pulse oximetry systems including Masimo SET® low noise optical sensors and read through motion pulse oximetry monitors for measuring SpO₂, PR, perfusion index (“PI”) and others. Masimo sensors include any of LNOP®, LNCS®, SofTouch™ and Blue™ adhesive or reusable sensors. Masimo oximetry monitors include any of Rad-8®, Rad-5®, Rad-5v® or SatShare® monitors.

Also, many innovations improving the measurement of blood constituents are described in at least U.S. Pat. Nos. 6,770,028; 6,658,276; 6,157,850; 6,002,952; 5,769,785 and 5,758,644, and are incorporated by reference herein. Corresponding low noise optical sensors are disclosed in at least U.S. Pat. Nos. 6,985,764; 6,813,511; 6,792,300; 6,256,523; 6,088,607; 5,782,757 and 5,638,818, and are incorporated by reference herein.

Masimo also manufactures more advanced co-oximeters including Masimo Rainbow® SET, which provides measurements in addition to SpO₂, such as total hemoglobin (SpHb™), oxygen content (SpCO™), methemoglobin (Sp-Met®), carboxyhemoglobin (SpCO®) and PVI®. Advanced blood parameter sensors include Masimo Rainbow® adhesive, ReSposable™ and reusable sensors. Masimo’s advanced blood parameter monitors include Masimo Radical-7™, Rad-87™, and Rad-57™ monitors as well as Pronto and Pronto-7 spot check monitors.

Innovations relating to these more advanced blood parameter measurement systems are described in at least U.S. Pat. Nos. 7,647,083; 7,729,733; U.S. Pat. Pub. Nos. 2006/0211925; and 2006/0238358, incorporated by reference herein.

Such advanced pulse oximeters, low noise sensors and advanced blood parameter systems have gained rapid acceptance in a wide variety of medical applications, including surgical wards, intensive care and neonatal units, general wards, home care, physical training, and virtually all types of monitoring scenarios.

Thus, by employing the embodiments of the CCHD screening processes and systems disclosed herein, CCHD, particularly PDA, may be more accurately detected and diagnosed. Specifically, false positives may be reduced, variances in SpO₂ may be detected and filtered, caregivers may be more properly directed, and/or measurement confidence may be evaluated, among other advantages.

Conditional language used herein, such as, among others, “can,” “could,” “might,” “may,” “e.g.,” and the like, unless specifically stated otherwise, or otherwise understood within the context as used, is generally intended to convey that certain embodiments include, while other embodiments do not include, certain features, elements and/or states. Thus, such conditional language is not generally intended to imply that features, elements and/or states are in any way required for one or more embodiments or that one or more embodiments necessarily include logic for deciding, with or without author input or prompting, whether these features, elements and/or states are included or are to be performed in any particular embodiment.

While certain embodiments of the inventions disclosed herein have been described, these embodiments have been presented by way of example only, and are not intended to limit the scope of the inventions disclosed herein. Indeed, the novel methods and systems described herein can be embodied in a variety of other forms; furthermore, various omissions, substitutions and changes in the form of the methods and systems described herein can be made without departing from the spirit of the inventions disclosed herein. The claims and their equivalents are intended to cover such

forms or modifications as would fall within the scope and spirit of certain of the inventions disclosed herein.

What is claimed is:

1. A method of determining a likelihood of a critical congenital heart defect using a pulse oximeter, the method comprising:

displaying, on a display of a pulse oximeter, instructions to a caregiver to position a noninvasive sensor at a first measurement site on a patient predetermined to be substantially indicative of a pre-ductal blood oxygenation level and not substantially indicative of a post-ductal blood oxygenation level, wherein the noninvasive sensor is configured to output signals responsive to attenuation of light at measurement sites on the patient to the pulse oximeter, wherein the first measurement site is on a first limb of the patient;

receiving the output signals indicative of the pre-ductal blood oxygenation level at the first measurement site; displaying, on the display of the pulse oximeter, an indication to the caregiver of a success or failure of obtaining a measure of pre-ductal blood oxygenation level at the first measurement site;

displaying, on the display of the pulse oximeter, instructions to the caregiver to position the noninvasive sensor at a second measurement site on the patient predetermined to be substantially indicative of a post-ductal blood oxygenation level and not substantially indicative of a pre-ductal blood oxygenation level, wherein the second measurement site is on a second limb of the patient, and wherein the second limb of the patient is different from the first limb of the patient;

receiving the output signals indicative of the post-ductal blood oxygenation level at the second measurement site;

displaying, on the display of the pulse oximeter, an indication to the caregiver of a success or failure of obtaining a measure of post-ductal blood oxygenation level at the second measurement site;

processing, by one or more hardware processors of the pulse oximeter, the pre-ductal blood oxygenation level and the post-ductal oxygenation level to determine a likelihood that the patient has a critical congenital heart defect, wherein said processing comprises at least time-wise aligning of at least portions of the output signals indicative of the pre-ductal blood oxygenation level and the output signals indicative of the post-ductal blood oxygenation level; and

displaying, on the display of the pulse oximeter, an indication to the caregiver of the likelihood that the patient has the critical congenital heart defect.

2. The method of claim 1 further comprising:

displaying, on the display of the pulse oximeter, instructions to the caregiver to proceed with an additional screening for the critical congenital heart defect after a period of time.

3. The method of claim 1, wherein displaying the indication of the likelihood that the patient has the critical congenital heart defect comprises:

displaying, on the display of the pulse oximeter, an indication of a difference between the measure of pre-ductal blood oxygenation level and the measure of post-ductal blood oxygenation level.

4. The method of claim 3, wherein displaying the indication of the likelihood that the patient has the critical congenital heart defect further comprises:

displaying, on the display of the pulse oximeter, an indication of the measure of pre-ductal blood oxygenation level and the measure of post-ductal blood oxygenation level.

5. The method of claim 4, wherein displaying the indication of the likelihood that the patient has the critical congenital heart defect further comprises:

displaying, on the display of the pulse oximeter, respective measures of perfusion index associated with each of the measure of pre-ductal blood oxygenation level and the measure of post-ductal blood oxygenation level.

6. The method of claim 5, wherein displaying the indication of the likelihood that the patient has the critical congenital heart defect further comprises:

displaying, on the display of the pulse oximeter, an indication of a confidence in the likelihood that the patient has the critical congenital heart defect.

7. The method of claim 1 further comprising:

displaying, on the display of the pulse oximeter, an indication of a progress of obtaining the measure of pre-ductal blood oxygenation level at the first measurement site.

8. The method of claim 1 further comprising:

displaying, on the display of the pulse oximeter, an indication of a progress of obtaining the measure of post-ductal blood oxygenation level at the second measurement site.

9. The method of claim 1, wherein the first measurement site comprises a right hand of the patient.

10. The method of claim 9, wherein the second measurement site comprises at least one of a foot or left hand of the patient.

11. The method of claim 1, wherein:

the output signals indicative of the pre-ductal blood oxygenation level at the first measurement site are received during a first duration of time, and comprise a first waveform responsive to a pre-ductal oxygen saturation of the patient;

the output signals indicative of the post-ductal blood oxygenation level at the second measurement site are received during a second duration of time, and comprise a second waveform responsive to a post-ductal oxygen saturation of the patient; and

processing, by the one or more hardware processors of the pulse oximeter, the pre-ductal blood oxygenation level and the post-ductal oxygenation level to determine the likelihood that the patient has a critical congenital heart defect comprises:

electronically determining, by the one or more hardware processors, one or more waveform features in the first waveform responsive to the pre-ductal oxygen saturation;

electronically determining, by the one or more hardware processors, one or more of the one or more waveform features in the second waveform responsive to the post-ductal oxygen saturation;

electronically time-wise aligning, by the one or more hardware processors, one or more of the one or more waveform features from the first waveform with one or more of the one or more waveform features from the second waveform;

electronically determining, by the one or more hardware processors, a pre-ductal value of the first waveform at a first time; and

electronically determining, by the one or more hardware processors, a post-ductal value of the second waveform at the first time taking into account the time-wise aligning,

wherein displaying the indication of the likelihood that the patient has the critical congenital heart defect comprises: in response to determining the pre-ductal value and the post-ductal value differ by a predetermined threshold, outputting display indicia to the display indicating a likelihood that the patient has the critical congenital heart defect.

12. The method of claim 11, wherein said first time comprises a third duration of time, and wherein determining said pre-ductal value and said post-ductal value differ by the predetermined threshold comprises:

calculating an average difference between said pre-ductal value and said post-ductal value over the third duration of time.

13. The method of claim 12, wherein said first duration of time and said second duration of time each comprise at least sixty seconds.

14. The method of claim 13, wherein said third duration of time comprises at least thirty seconds.

15. The method of claim 11, wherein said one or more waveform features from said first waveform comprise at least one of: a trough, a peak, a natural high cycle of oxygen saturation, a natural low cycle of oxygen saturation, or a cyclical variation.

16. The method of claim 11, wherein processing, by the one or more hardware processors of the pulse oximeter, the pre-ductal blood oxygenation level and the post-ductal oxygenation level to determine the likelihood that the patient has a critical congenital heart defect further comprises:

discarding a waveform when a quality of said output signals is outside of a pre-determined range.

17. The method of claim 11, wherein processing, by the one or more hardware processors of the pulse oximeter, the pre-ductal blood oxygenation level and the post-ductal oxygenation level to determine the likelihood that the patient has a critical congenital heart defect further comprises:

receiving said signal outputs to obtain a third waveform responsive to a blood perfusion level of the patient, wherein a waveform is discarded when the blood perfusion level associated with the waveform is outside of a pre-determined range.

18. A pulse oximetry based system configured to determine a likelihood of a critical congenital heart defect, the system comprising:

a noninvasive sensor configured to output signals responsive to an attenuation of light at measurement sites on a patient; and

a pulse oximeter in communication with the noninvasive sensor to drive light sources of said noninvasive sensor and to receive said output signals, said pulse oximeter including one or more hardware processors configured to:

display, on a display of the pulse oximeter, instructions to a caregiver to position the noninvasive sensor at a first measurement site on a patient predetermined to be substantially indicative of a pre-ductal blood oxygenation level and not substantially indicative of a post-ductal blood oxygenation level;

receive the output signals indicative of the pre-ductal blood oxygenation level at the first measurement site;

display, on the display of the pulse oximeter, an indication to the caregiver of a success or failure of

obtaining a measure of pre-ductal blood oxygenation level at the first measurement site;

display, on the display of the pulse oximeter, instructions to the caregiver to position the noninvasive sensor at a second measurement site on the patient predetermined to be substantially indicative of a post-ductal blood oxygenation level and not substantially indicative of a pre-ductal blood oxygenation level, wherein the second measurement site is on a second limb of the patient, and wherein the second limb of the patient is different from the first limb of the patient;

receive the output signals indicative of the post-ductal blood oxygenation level at the second measurement site;

display, on the display of the pulse oximeter, an indication to the caregiver of a success or failure of obtaining a measure of post-ductal blood oxygenation level at the second measurement site;

process, by the one or more hardware processors of the pulse oximeter, the pre-ductal blood oxygenation level and the post-ductal oxygenation level to determine a likelihood that the patient has a critical congenital heart defect wherein said processing comprises at least time-wise aligning of at least portions of the output signals indicative of the pre-ductal blood oxygenation level and the output signals indicative of the post-ductal blood oxygenation level; and

display, on the display of the pulse oximeter, an indication to the caregiver of the likelihood that the patient has the critical congenital heart defect.

19. The system of claim 18, wherein:

the output signals indicative of the pre-ductal blood oxygenation level at the first measurement site are received during a first duration of time, and comprise a first waveform responsive to a pre-ductal oxygen saturation of the patient;

the output signals indicative of the post-ductal blood oxygenation level at the second measurement site are received during a second duration of time, and comprise a second waveform responsive to a post-ductal oxygen saturation of the patient; and

processing, by the one or more hardware processors of the pulse oximeter, the pre-ductal blood oxygenation level and the post-ductal oxygenation level to determine the likelihood that the patient has a critical congenital heart defect comprises:

electronically determining, by the one or more hardware processors, one or more waveform features in the first waveform responsive to the pre-ductal oxygen saturation;

electronically determining, by the one or more hardware processors, one or more of the one or more waveform features in the second waveform responsive to the post-ductal oxygen saturation;

electronically time-wise aligning, by the one or more hardware processors, one or more of the one or more waveform features from the first waveform with one or more of the one or more waveform features from the second waveform;

electronically determining, by the one or more hardware processors, a pre-ductal value of the first waveform at a first time; and

electronically determining, by the one or more hardware processors, a post-ductal value of the second waveform at the first time taking into account the time-wise aligning,
wherein displaying the indication of the likelihood that the patient has the critical congenital heart defect comprises: in response to determining the pre-ductal value and the post-ductal value differ by a predetermined threshold, outputting display indicia to the display indicating a likelihood that the patient has the critical congenital heart defect.

* * * * *

专利名称(译)	自动化CCHD筛查和检测		
公开(公告)号	US10278648	公开(公告)日	2019-05-07
申请号	US15/279134	申请日	2016-09-28
[标]申请(专利权)人(译)	梅西莫股份有限公司		
申请(专利权)人(译)	Masimo公司		
当前申请(专利权)人(译)	Masimo公司		
[标]发明人	AL ALI AMMAR MUHSIN BILAL		
发明人	AL-ALI, AMMAR MUHSIN, BILAL		
IPC分类号	A61B5/00 A61B5/0205 A61B5/02 A61B5/1455 A61B5/0295 A61B5/08 A61B5/024		
CPC分类号	A61B5/7275 A61B5/02028 A61B5/0295 A61B5/14551 A61B5/14552 A61B5/684 A61B5/6825 A61B5/6829 A61B5/7221 A61B5/7278 A61B5/7282 A61B5/743 A61B5/0205 A61B2503/04 A61B5/024 A61B5/0816		
优先权	61/703132 2012-09-19 US 61/583143 2012-01-04 US		
其他公开文献	US20170014084A1		
外部链接	Espacenet		

摘要(译)

描述了自动化的关键先天性心脏缺陷 (“CCHD”) 筛查系统和过程。可引导护理人员使用单或双传感器脉搏血氧计来获得导管前和血管后血氧测量值。测量值的增量表示CCHD可能存在或不存在。基于例如灌注指数，通过可配置的测量置信度阈值来减少测量中的误差。可以从远程数据处理中心存储和检索测量数据以重复筛选。

