



US010682061B2

(12) **United States Patent**
Reza et al.

(10) **Patent No.:** **US 10,682,061 B2**
(45) **Date of Patent:** ***Jun. 16, 2020**

- (54) **PHOTOACOUSTIC REMOTE SENSING (PARS)**
- (71) Applicant: **illumiSonics, Inc.**, Edmonton (CA)
- (72) Inventors: **Parsin Haji Reza**, Waterloo (CA); **Roger Zemp**, Edmonton (CA)
- (73) Assignee: **illumiSonics Inc.**, Edmonton, Alberta
- (*) Notice: Subject to any disclaimer, the term of this patent is extended or adjusted under 35 U.S.C. 154(b) by 0 days.

This patent is subject to a terminal disclaimer.

- (21) Appl. No.: **16/116,038**
- (22) Filed: **Aug. 29, 2018**

- (65) **Prior Publication Data**
US 2019/0104944 A1 Apr. 11, 2019
- Related U.S. Application Data**

- (63) Continuation of application No. 14/919,565, filed on Oct. 21, 2015, now Pat. No. 10,117,583.
(Continued)

- (51) **Int. Cl.**
A61B 5/00 (2006.01)
G01N 21/17 (2006.01)

- (52) **U.S. Cl.**
CPC *A61B 5/0095* (2013.01); *G01N 21/1702* (2013.01); *G01N 21/1717* (2013.01);
(Continued)

- (58) **Field of Classification Search**
CPC A61B 5/0095; G01N 21/1717; G01N 21/1702; G01N 2021/1725;
(Continued)

- (56) **References Cited**
U.S. PATENT DOCUMENTS
5,070,733 A 12/1991 Nagata et al.
5,479,259 A * 12/1995 Nakata G01K 5/52 356/432
(Continued)

- FOREIGN PATENT DOCUMENTS**
CN 101526483 A 9/2009
CN 103048271 A 4/2013
(Continued)


- OTHER PUBLICATIONS**
Beard, Paul. "Biomedical Photoacoustic Imaging." Interface Focus 1.4 (2011): 602-631. PMC. Web. Dec. 12, 2017.
(Continued)

Primary Examiner — Dominic J Bologna
(74) *Attorney, Agent, or Firm* — Bookoff McAndrews, PLLC

(57) **ABSTRACT**

A photoacoustic remote sensing system (PARS) for imaging a subsurface structure in a sample has an excitation beam configured to generate ultrasonic signals in the sample at an excitation location; an interrogation beam incident on the sample at the excitation location, a portion of the interrogation beam returning from the sample that is indicative of the generated ultrasonic signals; an optical system that focuses at least one of the excitation beam and the interrogation beam with a focal point that is below the surface of the sample; and a detector that detects the returning portion of the interrogation beam.

27 Claims, 26 Drawing Sheets



Related U.S. Application Data

- (60) Provisional application No. 62/067,197, filed on Oct. 22, 2014.
- (52) **U.S. Cl.**
CPC *G01N 2021/1706* (2013.01); *G01N 2021/1725* (2013.01); *G01N 2201/06113* (2013.01); *G01N 2201/08* (2013.01); *G01N 2201/0846* (2013.01)
- (58) **Field of Classification Search**
CPC ... *G01N 2021/1706*; *G01N 2201/0846*; *G01N 2021/06113*; *G01N 2021/08*
See application file for complete search history.

References Cited

U.S. PATENT DOCUMENTS

5,615,675	A	4/1997	O'Donnell et al.
5,991,479	A	11/1999	Kleinerman
6,016,202	A	1/2000	Fuchs et al.
6,078,397	A	6/2000	Monchalin et al.
6,256,100	B1	7/2001	Banet et al.
8,004,689	B2	8/2011	Monchalin et al.
8,180,134	B2	5/2012	Wang
8,454,512	B2	6/2013	Wang et al.
9,153,931	B2	10/2015	Ichihara et al.
2006/0184042	A1	8/2006	Wang et al.
2006/0262316	A1	11/2006	Baney
2008/0123083	A1	5/2008	Wang et al.
2008/0194929	A1	8/2008	Pesach et al.
2009/0170149	A1*	7/2009	Viator <i>G01N 21/1702</i> 435/29
2010/0268042	A1	10/2010	Wang et al.
2012/0200845	A1	8/2012	Rousseau et al.
2012/0320368	A1	12/2012	Jiao et al.
2014/0009808	A1*	1/2014	Wang <i>G02F 1/33</i> 359/10
2014/0118749	A1	5/2014	Nakajima et al.

2014/0185055	A1	7/2014	Wang
2014/0247456	A1	9/2014	Horstmann et al.
2015/0077819	A1	3/2015	Schnell et al.
2015/0148655	A1	5/2015	Haupt et al.
2015/0150465	A1	6/2015	Irisawa et al.
2015/0153269	A1	6/2015	Nakatsuka
2015/0164337	A1	6/2015	Kim et al.
2015/0185187	A1	7/2015	Wang et al.
2015/0221081	A1	8/2015	Chanaka et al.
2015/0265156	A1	9/2015	Tanaka
2016/0113507	A1	4/2016	Reza et al.

FOREIGN PATENT DOCUMENTS

DE	10 2010 012 809	A1	9/2011
WO	2009/055705	A2	4/2009
WO	2009055705	A2	4/2009
WO	2009055705	A3	6/2009
WO	2014027316	A2	2/2014
WO	2014/036405	A2	3/2014
WO	2014062529	A1	4/2014
WO	2014160116	A1	10/2014
WO	2014168930	A1	10/2014

OTHER PUBLICATIONS

International Search Report and Written Opinion dated Dec. 19, 2018 in International Application No. PCT/IB2018/057585 (25 pages).

Kevan L. Bell et al., "Coherence-gated photoacoustic remote sensing microscopy", *Optics Express*, vol. 26, No. 18, Sep. 3, 2018, 16 pp.

Zhihua Ding et al., "Real-time phase-resolved optical coherence tomography and optical Doppler tomography", *Optics Express*, vol. 10, No. 5, Mar. 11, 2002, 10 pages.

Cedric Blatter et al., "Intrasweep phase-sensitive optical coherence tomography for noncontact optical photoacoustic imaging", *Optics Letters*, vol. 37, No. 21, Nov. 1, 2012, 4 pp.

* cited by examiner




FIG. 1




FIG. 2




FIG. 3




FIG. 4




FIG. 5A




FIG. 5B




FIG. 5C




FIG. 5D




FIG. 6




FIG. 7




FIG. 8




FIG. 9




FIG. 10A




FIG. 10B




FIG. 10C




FIG. 11A

FIG. 11B

FIG. 11C




FIG. 12




FIG. 13A




FIG. 13D




FIG. 13B




FIG. 13C




FIG. 14A




FIG. 14B




FIG. 14C




FIG. 14D




FIG. 15




FIG. 16




FIG. 17




FIG. 18A




FIG. 18B




FIG. 18C




FIG. 18D

FIG. 19A




FIG. 19B





FIG. 19C




FIG. 19D




FIG. 20




FIG. 21




FIG. 22




FIG. 23A




FIG. 23B




FIG. 23C




FIG. 24

PHOTOACOUSTIC REMOTE SENSING (PARS)

CROSS-REFERENCE TO RELATED APPLICATION(S)

This patent application is a continuation of U.S. patent application Ser. No. 14/919,565, filed on Oct. 21, 2015, which claims priority to U.S. Provisional Application No. 62/067,197, filed Oct. 22, 2014, the entireties of which are incorporated herein by reference.

FIELD

This relates to the field of biomedical optics imaging and, in particular, to a laser and ultrasound-based method and system for in vivo or ex vivo, non-contact imaging of biological tissue.

BACKGROUND

Photoacoustic imaging is an emerging hybrid imaging technology providing optical contrast with high spatial resolution. Nanosecond or picosecond laser pulses fired into tissue launch thermo-elastic-induced acoustic waves which are detected and reconstructed to form high-resolution images. Photoacoustic imaging has been developed into multiple embodiments, including photoacoustic tomography (PAT), photoacoustic microscopy (PAM), optical-resolution photoacoustic microscopy (OR-PAM), and array-based PA imaging (array-PAI). In photoacoustic tomography (PAT) signals are collected from multiple transducer locations and reconstructed to form a tomographic image in a way similar to X-ray CT. In PAM, typically, a single element focused high-frequency ultrasound transducer is used to collect photoacoustic signals. A photoacoustic signal as a function of time (depth) is recorded for each position in a mechanically scanned trajectory to form a 3-D photoacoustic image. The maximum amplitude as a function of depth can be determined at each x-y scan position to form a maximum amplitude projection (MAP)C-scan image. Photoacoustic microscopy has shown significant potential for imaging vascular structures from macro-vessels all the way down to micro-vessels. It has also shown great promise for functional and molecular imaging, including imaging of nanoparticle contrast agents and imaging of gene expression. Multi-wavelength photoacoustic imaging has been used for imaging of blood oxygen saturation, by using known oxy- and deoxy-hemoglobin molar extinction spectra.

In traditional photoacoustic imaging, spatial resolution is due to ultrasonic focusing and can provide a depth-to-resolution ratio greater than 100. In OR-PAM, penetration depth is limited to ~1 mm in tissue (due to fundamental limitations of light transport) but resolution is micron-scale due to optical focusing. OR-PAM can provide micron-scale images of optical absorption in reflection-mode, in vivo, something that no other technique can provide. OR-PAM is capable of imaging blood vessels down to capillary size noninvasively. Capillaries are the smallest vessels in the body and so much crucial biology occurs at this level, including oxygen and nutrient transport. Much can go wrong at the capillary level too. In cancers, cells have an insatiable appetite for oxygen and nutrients to support their uncontrolled growth. They invoke a range of signaling pathways to spawn new vessels in a process known as angiogenesis and these vessels typically form abnormally. Tumors are often highly heterogeneous and have regions of hypoxia.

Photoacoustic imaging has demonstrated the ability to image blood oxygen saturation (SO₂) and tumor hypoxia in vivo.

In most photoacoustic and ultrasound imaging systems, piezoelectric transducers have been employed, in which an ultrasound coupling medium such as water or ultrasound gel is required. However for many clinical applications such as wound healing, burn diagnostics, surgery, and many endoscopic procedures physical contact, coupling, or immersion is undesirable or impractical.

The detection of ultrasound in photoacoustic imaging has, until recently, relied on ultrasonic transducers in contact with the biological tissue or an ultrasonic coupling agent both of which have major drawbacks as described above. Some detection strategies to solving the non-contact optical interferometric sensing problems associated with photoacoustic imaging have been reported.

Optical means of detecting ultrasound and photoacoustic signals have been investigated over a number of years; however, to date no technique has demonstrated practical non-contact in vivo microscopy in reflection mode with confocal resolution and optical absorption as the contrast mechanism.

One example of a low-coherence interferometry method for sensing photoacoustic signals was proposed in U.S. pregrant publication no. 2014/0185055 to be combined with an optical coherence tomography (OCT) system, resulting in 30 μm lateral resolution.

Another prior art system is described in U.S. pregrant publication no. 2012/0200845 entitled "Biological Tissue Inspection Method and System", which describes a non-contact photoacoustic imaging system for in vivo or ex vivo, non-contact imaging of biological tissue without the need for a coupling agent.

Other systems use a fiber based interferometer with optical amplification to detect photoacoustic signals and form photoacoustic images of phantoms with acoustic (not optical) resolution. However these systems suffer from a poor signal-to-noise ratio, other contact-based photoacoustic systems offer significantly improved detection capabilities, in vivo imaging was not demonstrated, and optical-resolution excitation was not demonstrated.

Industrial laser ultrasonics has used interferometry to detect acoustic signatures due to optical excitation of inanimate objects for non-destructive testing. This approach has been adapted to detect ultrasound ex vivo in chicken breast and calf brain specimens, however, optical-resolution focusing of the excitation light was not examined.

Laser Doppler vibrometry has been a powerful non-contact vibration sensing methodology, however, weak signal-to-noise and poor image quality have proven to be a limitation when sensing deep-tissue signals from broad-beam photoacoustic excitation.

Similarly, Mach Zehnder interferometry and two-wave mixing interferometry have been used previously for sensing photoacoustic signals. However many such techniques still require direct contact or fluid coupling; have not offered in vivo studies or optical resolution for phantom studies.

SUMMARY

According to an aspect, there is provided a photoacoustic remote sensing system (PARS) for imaging a subsurface structure in a sample, where the PARS comprises an excitation beam configured to generate ultrasonic signals in the sample at an excitation location; an interrogation beam incident on the sample at the excitation location, a portion of the interrogation beam returning from the sample that is

indicative of the generated ultrasonic signals; an optical system that focuses the excitation beam at a first focal point and the interrogation beam at a second focal point, the first and second focal points being below the surface of the sample; and an interferometer that detects the returning portion of the interrogation beam.

According to another aspect, there is provided a photoacoustic remote sensing system (PARS) for imaging a subsurface structure in a sample, where the PARS comprises an excitation beam configured to generate ultrasonic signals in the sample at an excitation location; an optical system that focuses the excitation beam with a focal point that is below the surface of the sample; an interrogation beam directed toward an optical element that is responsive to the ultrasonic signals, the optical element being interposed between the sample and the interrogation beam, wherein the generated ultrasonic signals are characterized by a returning portion of the interrogation beam; and an interferometer that detects the returning portion of the interrogation beam. The optical element may be a Fabry Perot element. The interferometer may detect ultrasonic signals to a depth of 7 cm within the sample

According to another aspect, there is provided an endoscopic device that uses a photoacoustic remote sensing system (PARS) for imaging a subsurface structure in a sample, the endoscopic device comprising a fiber optic cable having an input end and a detection end; an excitation beam coupled to the input end of the fibre optic cable, wherein in use the excitation beam generates ultrasonic signals in the sample at an excitation location that is adjacent to the detection end of the fiber optic cable, the fiber optic cable focusing the excitation beam at a first focal point that is below the surface of the sample; an interrogation beam coupled to the input end of the fibre optic cable and incident on the excitation location, the fiber optic cable focusing the excitation beam at a first focal point that is below the surface of the sample, and wherein a portion of the interrogation beam that is indicative of the generated ultrasonic signals is received by the detection end of the fiber optic cable and travels to the input end; and an interferometer that receives the returning portion of the interrogation beam at the input end of the fiber optic cable.

According to other aspects, either alone or in combination, as applicable: the first and second focal points may be within 1 mm of the surface of the sample; the first and second focal points may be greater than 1 μm below the surface of the sample; the focal point may be spaced below the surface of the sample at a depth that is greater than a focal zone of the respective at least one of the excitation beam and the interrogation beam; the excitation beam and the interrogation beam have a lateral separation of less than 1 mm or less than 0.5 mm on the sample; the excitation beam may have a focal point that is laterally within the focal zone of the interrogation beam; the interrogation beam may have a focal point that is laterally within the focal zone of the excitation beam; there may be a processor that calculates an image of the sample based on the returning portion of the interrogation beam; at least one of the first focal point and the second focal point may have a focal diameter of less than 30 μm , 10 μm , or 1 μm ; the excitation beam may be scanned through the sample while the interrogation beam is stationary; the interrogation beam may be scanned through the sample while the excitation beam is stationary; and each of the interrogation beam and the excitation beam may be scanned through the sample concurrently.

Other aspects will be apparent from the description and claims below.

BRIEF DESCRIPTION OF THE DRAWINGS

These and other features will become more apparent from the following description in which reference is made to the appended drawings, the drawings are for the purpose of illustration only and are not intended to be in any way limiting, wherein:

FIG. 1 through 4 are block diagrams of optical-resolution photoacoustic remote sensing (OR-PARS) microscopy systems

FIG. 5 is a representative drawing of the overlap between the exciter and interrogator beams on a sample.

FIG. 6 is a block diagram of a sensing system involving a Doppler vibrometry configuration.

FIG. 7 is a block diagram of a sensing system using a Fabry-Perot interferometer.

FIG. 8 is a graph of an example of the response of a Fabry-Perot interferometer.

FIG. 9 is a block diagram of a Fabry-Perot interferometer.

FIGS. 10A, 10B, and 10C are block diagrams of examples of sensing system using a Fabry-Perot interferometer.

FIGS. 11A, 11B, and 11C are block diagrams of examples of sensing systems in an endoscopy configuration.

FIG. 12 is a block diagram of a sensing system integrated with another optical imaging system.

FIG. 13A is a PARS image of a network of carbon fibres.

FIG. 13B is a graph of the FWHM obtained by fitting an individual carbon fiber signal amplitude to a Gaussian function.

FIG. 13C is a graph of the resolution using a knife edge spread function.

FIG. 13D is a comparison of images obtained by sensing systems in reflection mode and transmission mode.

FIGS. 14A, 14B, 14C, and 14D are in vivo images of CAM-membrane of 5-day chicken embryos.

FIG. 15 is a chart of the measured photoacoustic signals from various dye concentrations.

FIG. 16 is a graph of the photoacoustic signal vs. excitation energy on the sample when interrogation power is fixed at 8 mW.

FIG. 17 is a graph of the photoacoustic signal vs. interrogation power on the sample when excitation energy is fixed at 60 nJ.

FIG. 18A depicts an example of the frequency response of a PARS system.

FIG. 18B depicts an example of the photoacoustic time domain signal of an individual carbon fiber using a PARS system.

FIGS. 18C and 18D depict an example of the photoacoustic time domain signal of an individual carbon fiber using a PARS system when the excitation and interrogation beams are separated by ~ 120 and 330 μm , respectively.

FIGS. 19A, 19B, 19C, and 19D depict in vivo PARS images of a mouse ear.

FIG. 20 is a block diagram of a system using a modified version of polarization sensitive Michelson interferometry

FIG. 21 is a graph of the measure signal from an unfocused transducer at different lateral distances.

FIG. 22 is a block diagram of a scanning system with two different interferometry designs for use with the interrogation beam.

FIGS. 23A, 23B, 23C, and 24 are images of a rat's ear.

DESCRIPTION

Photoacoustic imaging is an emerging biomedical imaging modality that uses laser light to excite tissues. Energy

absorbed by chromophores or any other absorber is converted to acoustic waves due to thermo-elastic expansion. These acoustic signals are detected and reconstructed to form images with optical absorption contrast. Photoacoustic imaging (PA) has been shown to provide exquisite images of microvessels and is capable of imaging blood oxygen saturation, gene expression, and contrast agents, among other uses. In most PA and ultrasound imaging systems piezoelectric transducers have been employed, in which an ultrasound coupling medium such as water or ultrasound gel is required. However for many clinical applications such as wound healing, burn diagnostics, surgery, and many endoscopic procedures physical contact, coupling, or immersion is undesirable or impractical. The system described herein is capable of in vivo optical-resolution photoacoustic microscopy using non-contact optical interferometric sensing without use of any ultrasound medium.

The system described herein, a photoacoustic remote sensing (PARS) microscopy system, is based on the idea of focusing excitation light to a near diffraction-limited spot and detecting photoacoustic signals using a confocal interrogation beam co-focused with the excitation spot. While previous approaches used a broad excitation beam with powerful lasers delivering mJ-J of pulse energy over a broad area, the PARS microscopy technique described herein uses nJ-scale pulse energies focused to near diffraction-limited spots. When focusing into tissue, the surface fluence can be maintained below present ANSI limits for laser exposure but the ballistically-focused light beneath the tissue can create fluences transiently far above the ANSI limits (as is done in other microscopy methods). In PARS, this means that very large local fluences \sim J/cm² are created within a micron-scale spot, generating very large initial acoustic pressures. For example, at 532-nm excitation wavelength, imaging a capillary with 500 mJ/cm² local fluence would result in an initial pressure on the order of 100 MPa locally. However, because this large pressure is initially localized to a micron-scale spot, by the time the signals are detected by a fluid-coupled detector \sim 1 cm away, the signals are reduced by 1/r diffractive losses and attenuation to \sim KPa scales. Signals can be orders of magnitude less for acoustic-resolution photoacoustic imaging where unfocused excitation beams are used, ANSI limits for visible light is 20 mJ/cm², and greater imaging depths are explored. Large numerical aperture focused acoustic detection is required for optimal signal-to-noise in OR-PAM to ensure the maximal energy collection. In PARS approach, large optically-focused photoacoustic signals are detected as close to the photoacoustic source as possible, which is done optically by co-focusing an interrogation beam with the excitation spot. A long-coherence length interrogation laser is preferably used with low amplitude and phase noise to read-out the large local photoacoustic vibrations interferometrically using a novel architecture designed to optimize received signal intensities.

The high sensitivity and the fine resolution of the proposed system offer performance comparable to other in vivo optical resolution photoacoustic microscopy systems but in a non-contact reflection mode suitable for many clinical and pre-clinical applications.

Some of the possible options of the optical-resolution photoacoustic remote sensing (OR-PARS) microscopy system are depicted in FIG. 1 through 4. Variations to the depicted systems will be apparent to those skilled in the art. Referring to FIG. 1, a block diagram of PARS system 10, and in particular, an optical-resolution photoacoustic remote sensing (OR PARS) microscopy system, is shown. A multi-wavelength fiber excitation laser 12 is used in multi focus

form to generate photoacoustic signals. Excitation laser 12 preferably operates in the visible spectrum, although the particular wavelength may be selected according to the requirements of the particular application. The acoustic signatures are interrogated using a long-coherence length probe beam 16 from a detection laser 14 that is co-focused and co-aligned with the excitation spots on sample 18. The probe beam 16 passes through a beam splitter 20 that transfers a portion of the signal to a detection unit 22. Probe beam 16 passes through a polarization control/beam quality unit 24 and a second beam splitter 26 that ties in a reference beam provider 28. Probe beam 16 then passes through a beam combiner unit 30 that also directs excitation beam 32 through a scanning device 34 and a focusing device 36 before reaching sample 18. The reflected beam 38 returns along the same path and is analyzed by detection unit 22.

A modified version of polarization sensitive Michelson interferometry has been employed to remotely record the large local initial pressures from chromophores and without appreciable acoustic losses. The PARS microscopy system offers optical lateral resolution down to sub- μ m.

Referring to FIG. 2, an example of a PARS experimental setup is shown, with a multi wavelength unit 40 placed in series with the excitation laser 12 and a lens system 42. Detection laser 14 is placed in series with a lens system 42 and a beam splitter 26 that is preferably a polarized beam splitter. As depicted, detection unit 22 is made up of a photodiode 46, amplifier 48, data acquisition unit 50 and a computer 52. A portion of beam 16 is redirected by beam splitter 26 to a neutral density filter 54 after passing through a quarter wave plate 56. Excitation beam 32 and probe beam 16 are combined by beam combiner unit 30 and directed to sample 18 by an objective lens 58. Sample 18 may be placed on a scanning unit 19, allowing for sample 18 to be moved and to allow scanning. FIG. 3 depicts another example of a PARS experimental setup using common path interferometry, but without second beam splitter 26 or neutral density filter 54. FIG. 3 shows scanning unit 19 placed between beam combiner unit 30 and objective lens 58, allowing for the beam to be moved to allow scanning rather than moving sample 18. FIG. 4 depicts a further example of a PARS experimental setup using Michelson interferometry, where quarter wave plate 56 has been omitted, and beam splitter 44 is used to redirect a portion of beam 16 to both neutral density filter 54 and detection unit 22.

It will be apparent that other examples may be designed with different components to achieve similar results. For example, other examples could include aft-fiber architectures where circulators replace beamsplitters similar to optical-coherence tomography architectures. Other alternatives may include longer coherence length sources, use of balanced photodetectors, interrogation-beam modulation, incorporation of optical amplifiers in the return signal path, etc.

The OR-PARS system takes advantage of two focused laser beams on the sample which can simulate a confocal OR-PAM configuration. Since there are optical components between the objective lens 58 and the sample 18, optical aberrations can be minimized.

Unlike OCT, PARS can take advantage of a high coherence interrogation beam (HC). In the low coherence interferometry (LC), backscattering light is detected from a selected depth (via coherence gating). However in HC method signal from all depth can be detected. Combination of HC detector with multi-focus excitation improves the SNR.

The OR-PARS takes advantage of optical excitation and detection which can help dramatically reduce the footprint of the system. The absence of a bulky ultrasound transducer makes this all optical system suitable for integrating with other optical imaging systems. Unlike previous non-contact photoacoustic imaging systems, the OR-PARS system is capable of in vivo imaging. It relies on much simpler setup and takes advantage of recording the large initial ultrasound pressures without appreciable acoustic losses.

During in vivo imaging experiments, no agent or ultrasound coupling medium are required. Unlike many other interferometric sensors PARS does not require a floating table. No special holder or immobilization is required to hold the target during imaging sessions.

PARS can be used to detect ultrasound signals directly. The PARS system is capable of detecting noncontact measurement of the displacement caused by ultrasound signals from an ultrasound transducer. In one example, a small amount of water was used at the top of the transducer and the transducer was driven by a sine wave from a function generator at 10 MHz, and produced a noise equivalent pressure of 1 KPa.

Other advantages that are inherent to the structure will be apparent to those skilled in the art. The embodiments described herein are illustrative and not intended to limit the scope of the claims, which are to be interpreted in light of the specification as a whole.

A pulse laser is used to generate photoacoustic signals and the acoustic signatures are interrogated using either a long-coherence or short-coherence length probe beam co-focused with the excitation spots. The PARS system is utilized to remotely record the large local initial pressures from chromophores and without appreciable acoustic losses due to diffraction, propagation and attenuation.

The excitation beam may be any pulsed or modulated source of electromagnetic radiation including lasers or other optical sources. In one example, a nanosecond-pulsed laser was used. The excitation beam may be set to any wavelength suitable for taking advantage of optical (or other electromagnetic) absorption of the sample. The source may be monochromatic or polychromatic.

The receiver beam, or interrogation beam, may be a long-coherence or a short-coherence length probe beam. In one example discussed above, the probe beam/receiver beam had a linewidth significantly less than the frequency of signals detected. Preferably, the interrogation beam has a coherence length selected so that the line width of the laser is less than the acoustic signal bandwidth or detection bandwidth.

PARS with a long-coherence beam may be integrated with OCT to provide a complete set of information offered by both photoacoustic and OCT systems.

PARS with a short or long-coherence beam may be used for either optical resolution photoacoustic microscopy (OR-PAM) or common photoacoustic microscopy (PAM).

In one example, both excitation and receiver beam may be combined and scanned. In this way, photoacoustic excitations may be sensed in the same area as they are generated and where they are the largest. Other arrangements may also be used, including keeping the receiver beam fixed while scanning the excitation beam or vice versa. Galvanometers, MEMS mirrors and stepper/DC motors may be used as a means of scanning the excitation beam, probe/receiver beam or both.

The configurations shown in FIGS. 5A, 5B, 5C, and 5D may be used to perform PARS imaging. In the depicted embodiments, lines 502, depicted with a larger radius of

curvature, represent excitation beams and lines 504, depicted with a smaller radius of curvature, represent receiver beams. FIG. 5A offers a kind of confocal photoacoustic system where the excitation beam 502 and probing receive beam 504 are focused on the same spot, which can be on a micron- or sub-micron scale. In FIG. 5B, the optical resolution can be provided by the receiver beam 504, rather than the excitation beam 502. FIG. 5C shows excitation beam 502 and receiver beam 502 focused on different spots, and takes advantage of ultrasound time of flight in order to locate the excitation and receiver beams 502 and 504 at different positions. In FIG. 5D, optical resolution is provided by the excitation beam 502. Preferably, the focus of either or both of the excitation beam 502 or the detection beam 504 is less than 30 μm , less than 10 μm , less than 1 μm , or to the diffraction limit of light. A tighter focus results in a higher possible resolution and a better signal to noise ratio in the reflected beam that is detected. As used herein, the term "focus" is intended to refer to the focal zone of the beam, or the point at which the beam spot size is at the tightest size, and where the diameter of the focal zone is 30% greater than the diameter of the beam spot size. Also preferably, the excitation and detection beams 502 and 504 are focused on the same position, although there may be some spacing between the respective focuses as shown in FIG. 5C. In FIG. 5C, the beams may be focused at different locations, but preferably within 1 mm, 0.5 mm, 100 μm or within the range of the largest focus of the beam. In FIGS. 5A, 5B and 5D, the beams may be confocal, or may overlap within the focus of the beam with the largest focus. For example, in FIG. 5A, the excitation beam is larger than the detection beam, and the detection beam is directed at a location within the focus of the excitation beam. By moving the detection beam, the area within the excitation beam may be imaged. By having confocal beams, both beams may be moved to image the sample.

One or both of the beams are preferably focused below the surface of the sample. The beams may be focused, for example, using optics 36 shown in FIG. 1. Generally speaking, the beams may be effectively focused up to 1 mm below the surface of the sample. The beams may be focused at least 1 μm below the surface, or focused such that focal point of the beam is at least the distance of focal zone of the beam below the surface of the sample. It will be understood that, while both beams are preferably focused below the surface, in some embodiments either the excitation beam or the interrogation beam may be focused below the surface, with the other focused on, for example, the surface of the sample. In cases where only one beam is focused below the surface of the sample, the separation between the beams discussed previously will be a lateral separation, i.e. in the plane of the sample and orthogonal to the depth of the sample.

The excitation beam and sensing/receiver beam can be combined using dichroic mirrors, prisms, beamsplitters, polarizing beamsplitters etc. They can also be focused using different optical paths.

PARS can be integrated with any interferometry designs such as common path interferometer (using specially designed interferometer objective lenses), Michelson interferometer, Fizeau interferometer, Ramsey interferometer, Sagnac interferometer, Fabry-Perot interferometer and Mach-Zehnder interferometer. The basic principle is that phase (and maybe amplitude) oscillations in the probing receiver beam can be detected using interferometry and detected at AC, RF or ultrasonic frequencies using various detectors. Photoacoustic signals may also be detected using laser Doppler vibrometry configurations as shown in FIG. 6.

For clarity, the excitation beam has been omitted. In this example, the detection laser **14** passes through various optical elements, including beam splitters **602** and mirror **604** in order to provide a reference optical signal and an optical signal from sample **18** that can be compared and analyzed by photodetector **606**. A Bragg cell **608**, or other known device, is used to frequency shift the portion of probe beam **16** that is directed to sample **18**, as is known in the art.

Another interferometry example is shown in FIG. 7, which uses a Fabry-Perot interferometer (FPI) **610** in addition to the PARS system described above. FPI systems are all-optical detectors that offer high sensitivity and broad bandwidth important for photoacoustic imaging applications. When used with the PARS system, the FPI may result in an improved sensitivity of the system for deep imaging applications. As shown, the FPI **610** is reflective to the PARS interrogation beam **16** and visible to the PARS excitation beam **32** as shown in FIG. 7.

Referring to FIG. 8, an example of the response from FPI **610** is shown, in which FPI resonant peaks are formed in the range of ± 30 nm of the PARS interrogation wavelengths. For example if the PARS interrogation wavelength is 1550 nm, the range of the FPI resonant peak would be between 1520-1580 nm. The wavelength of the interrogation laser in the PARS system **10** is preferably tunable and will preferably be tuned to the sharpest slope of one the FPI resonant peaks.

Referring to FIG. 9, a detailed view of an example FPI **610** is shown. FPI **610** has mirrors **612** and **614** separated by a spacing material **616** and backed by a backing material **618**. The photoacoustic pressure generated by the excitation beam **32** shown in FIG. 7 causes changes in the optical thickness and/or physical thickness of the spacing material **616** and/or changes in the reflectivity of the FPI mirrors. These changes will cause a temporary shift of the FPI resonant peaks to a different wavelength and as a result, changes the amount of interrogation light reflected back from the FPI to the PARS system. The amount of change in the reflected interrogation light is directly proportional to the ultrasound/photoacoustic pressure. The mirrors **612** and **614** of the FPI **610** can be fabricated using various techniques, such as glancing angle deposition (GLAD) or other suitable techniques. The spacing material **616** may be various materials, such as Parylene C, PDMS, suitable polymers, or other suitable materials. The backing material **618** may be any suitable material, such as a suitable type of glass or polymer, such as PMMA. The resonant peak of the FPI **610** may be designed to work at various wavelengths, such as between 200-2500 nm, depending on the requirements of the application. An Anti-reflection (AR) coating (not shown) can be used on one side or both side of backing material **618** to improve the light coupling into the FPI **610**. It will be understood that other optical elements that are responsive to an acoustic signal may be used in place of the FPI described above, and that other optical effects other than those described above may be relied upon to enhance the detection of the ultrasonic signals in the sample.

As the PARS system **10** generally allows a user to obtain optical resolution down to sub- μm levels in a non-contact setting, adding a FPI **610** can provide acoustic resolution (i.e. greater than 30 μm) details with a penetration depth down to 7 cm. Some examples that involve the use of an FPI **610** are shown in FIGS. **10A**, **10B** and **10C**, based on the orientation of the excitation beam **32** relative to the interrogation beam **16**. As will be understood, the PARS system **10** may be used in transmission mode, such as in FIG. **10C**, using independent generation and detection units. In the

depicted configurations, both excitation and receiver beams may scan a sample together, one of the beams may be scanning with the other one is fixed, or both beams may be fixed. In addition, one or both excitation and receiver beams may be un-focused or loosely focused, or one or both beams may be focused.

As the PARS system takes advantage of an interferometry as explained herein, in which a reference beam is provided either by an external arm, or in a common mode path. The combination of the interferometry in the PARS system **10** as discussed previously and the FPI **610** can be used to improve the sensitivity of the photoacoustic imaging system.

In FPI-based photoacoustic imaging systems, the pressure of photoacoustic signals change the thickness of the FPI, optically and physically. These changes cause the shift of FPI resonant peaks and hence change the reflected light from the FPI. In the proposed configurations, the FPI **610** is a second interferometer that is added between the PARS system **10** and the sample **18**, as shown in FIG. 7. The PARS system **10** detects surface oscillations in the FPI **610**, where photoacoustic pressure physically changes the thickness of the FPI **610** and in result oscillate the FPI mirrors. The PARS system **10** as the second interferometer can pick up the oscillation of the mirrors in the FPI **610**. As such, the FPI and PARS detected signal can be combined together to improve the sensitivity of photoacoustic detection. It has been found that, under this system, a detection depth of 7 cm can be reached, although resolution is generally reduced.

In another example, low coherence probe beams can also be considered for detection of photoacoustic-induced optical phase oscillations but in this case signals from the sample beam and reference beam will interfere only if the sample and reference beam paths are equal lengths plus or minus a coherence length (as in optical coherence tomography). It may also be beneficial to scan the sample or reference beam path lengths or phase for both range gating and for measuring phase oscillations over the coherent region of interference.

The reflected light may be collected by photodiodes, avalanche photodiodes, phototubes, photomultipliers, CMOS cameras, CCD cameras (including EM-CCD, intensified-CCDs, back-thinned and cooled CCDs), etc. The detected light may be amplified by an RF amplifier, lock-in amplifier, trans-impedance amplifier, or other amplifier configuration. Also different methods may be used in order to filter the excitation beam from the receiver beam before detection. PARS may use optical amplifiers to amplify detected light prior to interferometry. It may also be beneficial to demodulate via Fabry-Perot, very narrow-line optical filters, nonlinear and photorefractive crystals and/or spectral hole burning.

An alternative to scanning vibrometry is digital holographic microscopy. Methods of digital holographic microscopy may be used to read out photoacoustic-induced optical phase oscillations. In this method, a pulsed receiver beam may be used so that the sample beam and an angled reference beam interfere on an image sensor (like a CCD or CMOS camera). Using methods of Gabor or Leith-Upatneiks holography not only the amplitude can be recovered, but also the phase of the light reflected/scattered from the sample over a wide field of view. By gating when the probe beam is pulsed onto the sample relative to the excitation beam it is possible to stroboscopically reconstruct the photoacoustic signals from each point in the sample as a function of time. Alternatively, it may be possible to time-gate the camera acquisition. The excitation spot may be optically focused or focused over a wide-field. When wide-

field excitation beams are used, optical-resolution can be achieved by receiving sensing optics and this resolution is anticipated to depths within a transport mean-free path in turbid media.

PARS can be used in many form factors, such as table top, handheld and endoscopy. Examples of endoscopy PARS are shown in FIGS. 11A, 11B and 11C with various arrangements of PARS excitation units 1102, PARS detection units 1104, fibre optics 1106 such as image-guide fibers, and lenses 1108 that focus the respective beams onto the sample 18. When excitation and detection units 1102 and 1104 are separated, there may be a separate fibre 1110 provided, such as a single mode fiber.

A table top and handheld PARS system may be constructed based on principles known in the art. The proposed PARS system takes advantage of optical excitation and detection which can help to dramatically reduce the footprint of the system. The footprint of previous systems has been much too large to use the system in all but body surfaces. For endoscopic applications, the footprint of the ultrasound detector must be minimized to make the imaging catheter small and flexible enough to navigate through small orifices and vessels. The piezoelectric receivers are not ideal candidates for endoscopic applications as there is trade-off between the sensitivity and the size of the receiver. On the other hand for many invasive applications sterilizable or disposable catheters and a non-contact approach are necessary. The system may also be used as PARS endoscopy system with a potential footprint the size of an optical fiber, as both excitation and PARS beam can be coupled into a single mode fiber or image guide fiber.

Image-guide fibers (miniaturized fiber bundles with as many as 100,000 or more individual micrometer-sized strands in a single optical fiber with diameters ranging from 200 μm to 2 mm) may be used to transmit both focused light spots. The excitation beam may be scanned either at the distal end or proximal end of the fiber using one of the scanning methods mentioned before. However, the receiver beam may be scanned or be fixed. The scanned spot is transmitted via the image-guide fiber 1106 to the output end. Therefore, it may be used to directly contact the sample, or re-focused using an attached miniature GRIN lens 1108. In one example, C-scan photoacoustic images were obtained from the fiber image-guides using an external ultrasound transducer to collect photoacoustic signals. Using an edge-spread and Gaussian function, a resolution of approximately 7 μm was obtained using the image-guide fiber 1106. It is believed that a higher resolution may also be obtained with appropriate improvements to the setup and equipment used.

PARS may be used to detect ultrasound signals generated from other sources including ultrasound transducers. The system may also be used as an optical vibrometer. Vibrometers have been used widely for assessing the operating condition of mechanical properties. Optical vibrometers (OV) offer various advantages over traditional vibration measurement techniques. The precise metrology of noncontact measurement, high sensitivity and accuracy are the major benefits of optical vibrometers. Most of the common optical vibrometer, including laser Doppler vibrometer (LDV) and Sagnac vibrometer are based on the optical interferometry, requiring two coherent light beams. OVs have been used for various applications including noncontact measurement of the displacement, the acceleration and the velocity of solid surfaces. The device size, cost and noise sensitivity of interferometry are limitations of the current OVs designs.

In one example, a vibrometry method based on PARS detection was used to detect by noncontact measurement of the displacement caused by ultrasound signals from a 10 MHz piezoelectric transducer. The reflected near-infrared beam from the sample is phase-modulated at the ultrasound frequency, and a beat-intensity can be detected. One example had a noise equivalent pressure of 1 KPa over 10 MHz bandwidth for real-time detection. This is improved orders of magnitude with lock-in detection. The measurement capability of the system may also be used to measure motion in high-frequency MEMS actuators and for optical detection of ultrasound.

The PARS system may be combined with other imaging modalities such as fluorescence microscopy, two-photon and confocal fluorescence microscopy, Coherent-Anti-Raman-Stokes microscopy, Raman microscopy, Optical coherence tomography, other photoacoustic and ultrasound systems, etc. This could permit imaging of the microcirculation, blood oxygenation parameter imaging, and imaging of other molecularly-specific targets simultaneously, a potentially important task that is difficult to implement with only fluorescence based microscopy methods. An example of a PARS system 10 integrated with another optical imaging system 1202 is shown in FIG. 12, where PARS 10 and the other optical imaging system 1202 are both connected to the sample 18 by a combiner 1204.

PARS may be used for A, B or C scan images for in vivo, ex vivo or phantom studies.

A multi-wavelength visible laser source may also been implemented to generate photoacoustic signals for functional or structural imaging.

PARS may be optimized in order to takes advantage of a multi-focus design for improving the depth-of-focus of 2D and 3D OR-PARS imaging. The chromatic aberration in the collimating and objective lens pair may be harnessed to refocus light from a fiber into the object so that each wavelength is focused at a slightly different depth location. Using these wavelengths simultaneously may be used to improve the depth of field and signal to noise ratio (SNR) of PARS images. During PARS imaging, depth scanning by wavelength tuning may be performed.

Polarization analyzers may be used to decompose detected light into respective polarization states. The light detected in each polarization state may provide information about ultrasound-tissue interaction.

Applications

It will be understood that the system described herein may be used in various ways, such as those purposes described in the prior art, and also may be used in other ways to take advantage of the aspects described above. A non-exhaustive list of applications is discussed below.

The system may be used for imaging angiogenesis for different pre-clinical tumor models.

The system may also be used for clinical imaging of micro- and macro-circulation and pigmented cells, which may find use for applications such as in (1) the eye, potentially augmenting or replacing fluorescein angiography; (2) imaging dermatological lesions including melanoma, basal cell carcinoma, hemangioma, psoriasis, eczema, dermatitis, imaging Mohs surgery, imaging to verify tumor margin resections; (3) peripheral vascular disease; (4) diabetic and pressure ulcers; (5) burn imaging; (6) plastic surgery and microsurgery; (7) imaging of circulating tumor cells, especially melanoma cells; (8) imaging lymph node angiogenesis; (9) imaging response to photodynamic therapies including those with vascular ablative mechanisms;

(10) imaging response to chemotherapeutics including anti-angiogenic drugs; (11) imaging response to radiotherapy.

The system may be useful in estimating oxygen saturation using multi-wavelength photoacoustic excitation and PARS or PARS-etalon detection and applications including: (1) 5 estimating venous oxygen saturation where pulse oximetry cannot be used including estimating cerebrovenous oxygen saturation and central venous oxygen saturation. This could potentially replace catheterization procedures which can be risky, especially in small children and infants.

Oxygen flux and oxygen consumption may also be estimated by using OR-PARS or PARS-etalon imaging to estimate oxygen saturation, and an auxiliary method to estimate blood flow in vessels flowing into and out of a region of tissue.

The system may also have some gastroenerological applications, such as imaging vascular beds and depth of invasion in Barrett's esophagus and colorectal cancers. Depth of invasion is key to prognosis and metabolic potential. Gastroenterological applications may be combined or piggy-backed off of a clinical endoscope and the miniaturized PARS/PARS-etalon system may be designed either as a standalone endoscope or fit within the accessory channel of a clinical endoscope.

The system may have some surgical applications, such as functional imaging during brain surgery, use for assessment of internal bleeding and cauterization verification, imaging perfusion sufficiency of organs and organ transplants, imaging angiogenesis around islet transplants, imaging of skin-grafts, imaging of tissue scaffolds and biomaterials to evaluate vascularization and immune rejection, imaging to aid microsurgery, guidance to avoid cutting critical blood vessels and nerves.

Other examples of applications may include PARS/PARS-etalon imaging of contrast agents in clinical or pre-clinical applications; identification of sentinel lymph nodes; non- or minimally-invasive identification of tumors in lymph nodes; imaging of genetically-encoded reporters such as tyrosinase, chromoproteins, fluorescent proteins for pre-clinical or clinical molecular imaging applications; imaging actively or passively targeted optically absorbing nanoparticles for molecular imaging; and imaging of blood clots and potentially staging the age of the clots.

PARS Mechanism

There will now be given a more detailed discussion of the PARS modulation mechanisms, and comparison with experiment.

The total intensity of light incident on the photodiode is given as the ensemble average of the squared magnitude of the sample and reference beam electric fields (taking constant factors as unity for convenience):

$$I_{PD} = \langle |E_S + E_R|^2 \rangle$$

The power of light reflected from the surface of the sample then routed to the photodiode is estimated as $0.9^2 R_1 I_0$ where I_0 is the incident intensity from the source, and R_1 is the intensity reflection coefficient at the air-sample interface and a factor of 0.9 is included for each pass through a 10:90 beamsplitter. Likewise the reference beam power is estimated as $I_R = \langle |E_E|^2 \rangle = 0.1^2 \eta_{VND F}^2 I_0$ where $\eta_{VND F}$ is the transmissivity of the variable neutral density filter (VND F).

Possible mechanisms include a pressure-induced refractive-index modulation, thermally-induced refractive index modulation, surface oscillations, and scatterer position modulation due to confined thermal expansion. Each mechanism will be discussed below.

Pressure-Induced Refractive-Index Modulation

Refractive index changes due to temperature and pressure rises may in turn affect the scattering of light. We first consider pressure changes. Local initial pressures may be calculated as very large when optical focusing and thermal confinement conditions are applied: The initial pressure is given as $P_0 = \Gamma \Phi \mu_a$ where Γ is the Gruneisen parameter. Assuming 532 nm light is focused to a micron scale spot size with a focal fluence, Φ , of 500 mJ/cm² and that this light is absorbed by oxygenated blood which has an estimated optical absorption coefficient μ_a of $0.0054 \times 43876 = 236.9$ cm⁻¹ (calculated at this wavelength assuming a hemoglobin concentration of 150 g/L and assuming no optical absorption saturation) the absorbed energy produces a transient temperature rise on a micro-scale as high as 30K. Using the above parameter estimates, we calculate an initial pressure as high as 118.5 MPa for unity Gruneisen parameter.

The optical refractive index experiences a perturbation to pressure variations estimated as

$$n(r,t) = n_0 (1 + \eta n_0^2 P(r,t) / 2\rho v_a^2)$$

where n_0 is the unperturbed optical refractive index, η is the elasto-optic coefficient (~ -0.32 for water), $P(r,t)$ is the pressure field, ρ is mass density and v_a is the speed of sound. The accumulated phase shift of light passing through a zone of enhanced pressure can be calculated by Raman Nath diffraction theory and will depend on the direction of the sound and the direction of the light as well as the pressure field inhomogeneity. For a light beam incident on a plane pressure wave where both the light and sound beams are parallel the accumulated phase shift should be zero and are rather maximum when sound fields create effective diffraction gratings orthogonal to the light propagation. Rather than calculate the phase shifts of transmitted light we are more interested in the light reflected from a refractive index mismatch. With 100 MPa, initial pressure a refractive index step of $\Delta n \sim 0.019$ is predicted in the confined excitation volume, which is a 1.4% change. This results in an optical reflection coefficient of 0.7% which is very measurable.

This mechanism will contribute to both amplitude and phase variations in the probe beam.

The electric field back-reflected from the sample and incident on the photodiode is modelled as having two components, AC and DC terms:

$$E_S = E_{DC,S} + E_{AC,S}$$

Here $E_{DC,S} = \sqrt{0.9^2 R_1 I_0}$ is the electric field magnitude of light reflected from the sample surface and $E_{AC,S} = \sqrt{0.9^2 I_0 T_1^2 e^{-2\mu_{eff}d} R_{T,P}}$ is the electric field amplitude of light reflected from the excitation volume beneath the surface due to a transient pressure induced optical index step. T_1 is the transmission intensity coefficient at the air-tissue interface. Here both ballistic and scattered photons are accounted for as reflecting from the index step hence $e^{-2\mu_{eff}d}$ is the effective light attenuation over depth d in a scattering medium with effective attenuation coefficient μ_{eff} . The factor of 2 accounts for a round trip. The fraction of light modulated F_P is calculated as the AC terms from the expansion of $I_{PD} = \langle |E_S + E_R|^2 \rangle$ divided by the DC terms, the modulated light scattered from the excitation volume contains multiply scattered photons as we have assumed, then because phases of the AC component of the sample are effectively randomized the ensemble average of products of $E_{AC,S}$ with reference beam E_R or $E_{DC,S}$ will be zero, leaving the fraction of light modulated as:

$$F_P = \frac{\langle |E_{AC,S}|^2 \rangle}{\langle |E_R|^2 \rangle + \langle |E_{DC,S}|^2 \rangle}$$

Thermally-Induced Refractive Dive Index Modulation

Thermal effects also change refractive indices. Given the volumetric thermal expansion coefficient of water is $\alpha_v=207 \times 10^{-6} \text{K}^{-1}$ at 20°C . and it's refractive index is 1.33, the refractive index would change $\sim 0.01\% / ^\circ \text{C}$., hence $\sim 0.3\%$ with a 30°C . temperature rise, which is small but potentially noticeable as a source of scattering modulation. However, this effect would only be applicable locally at the heating zone. Thermal cooling will occur on a scale of microseconds-to-milliseconds after laser-induced heating.

Surface Oscillations

The above mechanisms point to significant sources of scattering position or scattering cross-section modulation that could be readily measurable when the probe beam is focused to sense the confined excitation volume. However, these large local signals are not the only potential source of signal modulation. Acoustic signals propagating to the surface could also result in phase modulation of the reflected light. Sound pressures decay as $1/r$ due to diffractive losses so pressure signals can be significantly weaker away from the confined heating region compared to at their source. For example, assuming an isotropic spherical heating region of diameter 5 microns (e.g. an isolated red blood cell) the sound pressure level at the boundary of the sphere is estimated as $\sim 100 \text{ MPa}$ but 1 mm away this is predicted as only $2.5/1000 \times 100 \text{ MPa} \sim 0.25 \text{ MPa}$. Surface pressure modulation can cause surface oscillations and the reflection of the interrogation beam from this oscillating surface can be a source of detected signal. For a local plane wave at the tissue surface with peak pressure p the particle velocity is estimated as

$$v_p = \frac{p}{Z}$$

where Z is the characteristic acoustic impedance. For a sinusoidal pressure field the particle/surface displacements are thus $\Delta z \sim v_p / \omega_a$ which is a couple of nm for 10 MHz ultrasound with 0.25 MPa amplitude. This is a fraction of a wavelength but still a significant source of phase modulation considering the surface reflectivity may be high and will return significant amounts of incident probe-beam light back to the detector. Notice, however, that this surface modulation has a $1/\omega_a$ dependence hence there is an inherent low-pass filtering effect that could be rejecting high-frequency components. Nevertheless such signals are readily measurable by our system, as evidenced by when we position the probe beam away from the excitation beam and can still form images when we scan the excitation spot.

If we assume that only surface oscillations are considered, the electric field phasor from the sample surface is modelled as $E_S = E_{S0} \cos(kz \cos(\omega_a t)) \sim E_{S0} (J_0(kz) + J_1(kz)e^{j\omega_a t} + \dots)$ for a continuous wave modulation at acoustic angular frequency ω_a . Here $E_{S0} = \sqrt{0.9^2 R_1 P_0}$. In this case we do not need to worry about light propagating into the sample, and the fraction of light modulated by surface oscillations is estimated as:

$$F_S = \frac{2\langle E_{DC,S} E_{AC,S} \rangle + 2\langle E_R E_{AC,S} \rangle}{\langle E_{DC,S} \rangle^2 + \langle E_{AC,S} \rangle^2 + \langle E_R \rangle^2 + 2\langle E_R E_{DC,S} \rangle}$$

$$= \frac{2E_{S0}^2 J_0(kz) J_1(kz) + 2E_{S0} E_R J_1(kz)}{E_{S0}^2 J_0^2(kz) + E_{S0}^2 J_1^2(kz) + E_R^2 + 2E_{S0} E_R J_0(kz)}$$

Scatterer Position Modulation Due to Confined Thermal Expansion

The volumetric thermal expansion coefficient of water is given as $\alpha_v=207 \times 10^{-6} \text{K}^{-1}$ at 20°C . Assuming a transient temperature rise on a micro-scale as high as 30K as calculated above, a volumetric expansion

$$\frac{\Delta V}{V} = 207 \times 10^{-6} \text{K}^{-1} \times 30 \text{K} = 6.21 \times 10^{-3}$$

is predicted. For a given temperature rise, the smaller the heated volume the larger the expansion. Now if the light is absorbed from a 3-micron spot size with a volume modelled as the $1/e$ penetration depth times the cross-sectional illumination area the isotropic particle motion is modelled as

$$(6.21 \times 10^{-3} \times \pi(r)^2 \times \text{DOF}_{ex})^{1/3} = 1 \text{ } \mu\text{m}$$

which is larger than the wavelength and a very large modulation, where r is the radius of the excitation beam spot size at focus and DOF_{ex} is the depth of field of excitation beam calculated using Gaussian beam parameters, $\sim 27 \text{ } \mu\text{m}$.

Experimental Results

Experimental Setup

There will now be given an example of experimental methods and setup that was used to test the principles discussed herein. A modified version of polarization sensitive Michelson interferometry has been employed to remotely record the large local initial pressures from chromophores and without appreciable acoustic losses. The experimental setup of the optical-resolution photoacoustic remote sensing (OR-PARS) microscopy system is depicted in FIG. 20. A multi-wavelength visible laser source using stimulated Raman scattering (SRS) has been implemented to generate photoacoustic signals. Briefly, a 1 ns pulse width, frequency doubled ytterbium-doped fiber laser (IPG Photonics Inc.) with a pulse repetition rate (PRR) of 40 kHz was coupled using a fiber launch system (MBT621D/M, Thorlabs Inc.) into a 3 m polarization-maintaining single-mode fiber (PM-SW) (HB-450, Fibercore Inc., UK) to generate SRS peaks at 543, 560, 590, and 600 nm and pulse energies up to 500 nJ. The system has been optimized in order to takes advantage of a multi-focus OR-PAM for improving the depth-of-focus of 2D and 3D OR-PARS imaging. FIG. 20 shows a tunable pulsed laser source at 2202, made up of pulsed laser 12, and having polarization maintaining single mode fiber 2204, frequency doubler 2206, collimator lenses 2208, and polarization maintaining nonlinear fibre 2210. Collimator lenses 2208 and the second polarization maintaining nonlinear fibre 2210 make up the SRS peak generation region 2212. After exiting the tunable pulsed laser source 2202, the beam may encounter an optional band pass filter wheel 2214 before being reflected by mirror 2216. Dual beam combiner 2218 passes the beams to galvanometer scanning mirror 2220, through objective lens 58, and to sample 18. Galvanometer scanning mirror 2220 is controlled by controller and driver 2222, which receives instructions from function generator 2224. Detection laser 14 may be a continuous wave laser, having single mode fiber 2226 followed by collimator lens 2208. The beam from detection laser 14 is then passed through polarized beam splitter 44. A portion of the split beam passes through lens 42 to photodiode 2228, through transimpedance amplifier 2230, low-

noise amplifier **2232**, band pass filter **2234**, to analog to digital convertor **2236**, providing input for controlling the galvanometer scanning mirror **2220**, as will be understood by persons skilled in the art. The other portion of the beam is directed through quarter wave plate **56**, through **10:90** beam splitter **2238**, where the beam is either passed through variable neutral density filter **2240** to mirror **2242** to provide reference beam **2244**, or through band pass filter **2246** to combine with the beam from the pulsed laser **12** at the dual beam combiner **2218**.

The chromatic aberration in the collimating and objective lens pair was harnessed to refocus light from a fiber into the object so that each wavelength is focused at a slightly different depth location. Using these wavelengths simultaneously was previously shown to improve the depth of field and SNR for structural imaging of microvasculature with OR-PAM. The differences between the multi-focus and single wavelength in vivo images will be discussed below.

The output of the PM-SMF was collimated (F280APC-A, Thorlabs Inc.) and combined using a dichroic beam combiner (DBC) with the receiver arm of the system. For the receiver arm a continuous wavelength (CW) C-band laser with 100-kHz linewidth (UK-L1550R, Thorlabs Inc., New Jersey) was used.

The light at the laser aperture was coupled to a single mode fiber and collimated. The randomly polarized collimated interrogation beam was passed through a polarized beam splitter (VBA05-1550, Thorlabs Inc., New Jersey) to be linearly polarized and a $\frac{3}{4}$ zero order wave plate (Thorlabs Inc., New Jersey) to be circularly polarized. The circularly polarized light then passes through a beam splitter (BS) with 10:90 ratio. A variable neutral density filter (NDF) and then a mirror has been placed at the 10% output of the BS in order to provide the optimized reference power of the interferometry. The beam at the 90% output of the BS has been combined by the excitation arm and then scanned across the samples via a 2D galvanometer scanning mirror system (GVS012/M, Thorlabs Inc.). The scanning mirrors were driven by a two-channel function generator. The scanning light was then focused tightly using an objective lens (M. Plan Apo NIR 20X, Mitutoyo, Japan). The reflected light back through the wave-plate creating 90° polarization which then reflects at the polarizing beam-splitter in order to guide the maximum possible intensity of reflected light to a 150 MHz-bandwidth InGaAs photodiode (PDA10CF, Thorlabs Inc., New Jersey). A band pass filter (BPS) has been placed on the detection arm to reject the excitation beam. An objective lens (518125, LEICA, Germany) was used in front of the photodiode (not shown in the figure) in order to refocus all possible reflected interrogation light to the small photodiode aperture. The output of the photodiode was amplified using an RF amplifier (Olympus 59001'R) with a band pass filter (1 MHz-20 MHz) and 26 dB gain and then digitized using a 4-channel PCI digitizer (Gage card) at a sampling rate of 200 MSamples/s. To form images, we project the maximum amplitude of each A-scan as a pixel in a C-scan en-face image, similar to previous PAM approaches. Since there are no optical components between the objective lens and the sample (unlike other reflection mode photoacoustic systems), therefore optical aberrations can be minimized. Interferometry model for PARS microscopy has been discussed in the supplementary information section.

FIG. **22** shows two other examples of interferometry designs, having one of two continuous wave lasers **2404** and **2406**. Setup **1**, shown at **2410**, uses common path interferometry. Continuous wave laser **2404** provides a beam

through polarized beam splitter **44**, where it is split either toward photodiode **46**, or through quarter wave plate **56** and then to beam combiner unit **30**. Setup **2**, shown at **2412**, uses Michelson interferometry. Continuous wave laser **2406** provides a beam to **10:90** beam splitter **2238**, where it is sent to another photodiode **46**, neutral density filter **2408**, or to beam combiner unit **30**. In both setups shown, pulse laser **12** provides an additional beam to polarization maintaining single mode fiber **2402**, lens system **42**, and then to beam combiner unit **30**. From beam combiner unit **30**, the beam then passes through objective lens **58** and encounters sample **18**.

In a Michelson interferometry, as shown in FIG. **22**, a non-polarized beam splitter **2238** with 10:90 ratio may be used. A variable neutral density filter (NDF) **2408** on a 3-axis stage may be placed at the 10% output of the beam splitter **2238** in order to provide the optimized reference power. Like the configuration shown in FIG. **20**, the interrogation beam may be combined with the excitation beam and scanned through an objective lens **58** on the sample **18**. For both configurations, an objective lens (518125, LEICA, Germany, not shown) may be used in front of the photodiode **46** in order to refocus all possible reflected interrogation light to the small photodiode element. The output of the photodiode **46** may be amplified (Olympus 5900PR) and digitized using a 4-channel PCI digitizer (Gage card) at a sampling rate of 200 MSamples/s.

Results and Discussion

FIG. **13A** shows PARS imaging of carbon fiber networks using ~1 nJ excitation pulse energy and 6 mW interrogation power. SNR (defined as average of signal over the standard deviation of the noise) was quantified as 45 ± 3 dB. FIG. **13B** shows FWHM due to fitting individual carbon fiber (with ~6 μ m diameter) signal amplitude to a Gaussian function. FIG. **13C** shows the resolution study using a knife edge spread function. The lateral resolution of the system has been measured as $\sim 2.5 \pm 1$ μ m. FIG. **13D** compares the images of reflection mode PARS and a transmission mode OR-PAM using a 10 MHz unfocused transducer (Olympus, V312-SM). The transmission mode OR-PAM setup was not capable of imaging with 1 nJ pulse energy. FIG. **13D** was formed using 500 pulse energy and images were recorded simultaneously.

In order to validate that the ultrasound signals can be detected directly, the system shown in FIG. **13D** was used to analyze a sample **18**, which includes a PARS system **10** and an ultrasound transducer **1302** to directly detect the ultrasound signals generated by the ultrasound transducer **1302**. The PARS system **10** is also capable of detecting noncontact measurement of the displacement caused by ultrasound signals from an unfocused piezoelectric transducer (Olympus, A312-10 MHz/0.25"). A small amount of water was used at the top of the transducer **1302** and the transducer was driven by a sine wave from a function generator at 10 MHz. The proposed system has a noise equivalent pressure of ~1 KPa.

FIGS. **14A**, **14B**, **14C**, and **14D** show in vivo images of CAM-membrane of 5-day chicken embryos. FIG. **14A** shows multi focus PARS images revealing both capillary beds and bigger blood vessels. In the chicken embryo model bigger blood vessels usually are located deeper than capillaries. In order to see both deep- and shallow vessels simultaneously the multi-focus design is optimized to extend the depth-of-field to ~250 μ m. FIG. **14B** shows a zoomed-in image of both capillary beds and bigger blood vessels. FIG. **14C** shows that PARS is capable of indicating the bleeding area in the tissue. The bleeding area caused

intentionally by using very high pulse energy. FIG. 14D shows PARS images acquired with a single wavelength (532 nm) rather than multiple wavelengths. With a single wavelength, depth-of-focus is limited to $\sim 30 \mu\text{m}$, rather than $250 \mu\text{m}$ with the multi-focus approach. Hence single-wavelength excitation is better-suited for depth-sectioning. This is evident in FIG. 14D, where top capillary beds are seen but not deeper large vessels. When we scan deeper with single wavelength excitation we see larger deep vessels but not superficial capillary beds. A comparison between images in FIGS. 14A and 14B, with FIG. 14D indicate that multi-focus design helps to improve the SNR of maximum amplitude C-scan images and improve the depth of field of the system compare to a single wavelength imaging. All the images shown herein are raw data and no major image processing steps are applied.

The PARS system is capable of imaging when both beams are scanning together, or when the interrogation beam is fixed and excitation beam is scanning. The field of view will be limited in this case as the generated photoacoustic signals will experience more attenuation if they are located far from the fixed interrogation beam.

FIG. 15 is a chart of the measured photoacoustic signals from various dye concentrations. FIG. 16 shows the photoacoustic signal (V) vs. energy of the excitation laser (nJ) on the sample when the interrogation power is fixed at 8 mW. It shows a linear response as expected. FIG. 17 shows the photoacoustic signal vs. interrogation power (mW) on the sample when excitation energy is fixed at 60 nJ. It is shown that after $\sim 11 \text{ mW}$ the photodiode goes to its saturation region.

FIG. 18A depicts the PARS frequency response and FIG. 18B depicts the PARS photoacoustic time domain signal of an individual carbon fiber. Referring to FIGS. 18C and 18D, the charts represent time domain photoacoustic signal of a single carbon fiber with $\sim 7 \mu\text{m}$ diameter when the excitation and interrogation beams are separated by ~ 120 and $330 \mu\text{m}$, respectively. The -3 dB bandwidth was measured as $\sim 20 \text{ MHz}$ by imaging carbon fiber network as shown in FIG. 13A. The band pass response was expected as the RF amplifier was set to a band pass filter (1 MHz-20 MHz). The axial resolution of the system is measured $\sim 75 \mu\text{m}$. In FIG. 18b, both excitation and interrogation beam have been co-aligned in the x and y direction and co-focused together in the z direction. Therefore, the photoacoustic signal starts at the time zero. However in FIGS. 18C and 18D, a time shift has been measured as the excitation and interrogation beams are separated by ~ 120 and $330 \mu\text{m}$, respectively. The results clearly show the ultrasound time of flight variation by changing the location of detection spot.

FIGS. 19A, 19B, 19C, and 19D depict in vivo PARS images of a mouse ear, and FIGS. 23A, 23B, 23C, and 24 show in vivo PARS images of a 1.00 g rat's ear. In all in vivo images pulse energy $\sim 20\text{-}80 \text{ nJ}$ was used and the interrogation power was fixed to 6 mW. The selection of excitation and interrogation lasers will be discussed below.

Unlike OCT, PARS takes advantage of a high coherence interrogation beam in the low coherence interferometry, backscattered light is detected from a selected depth (via coherence gating). However in PARS high-coherence method signals from all depths can be detected. The depth of images shown in FIGS. 14A, 14B, 14C, and 14D was measured $\sim 1.5 \text{ mm}$. This may be reduced to $\sim 1 \text{ mm}$ in media with higher turbidity where focusing is limited to a transport mean-free path. Aside from depth considerations, high-coherence interrogation beams may offer improved signal

contrast and signal-to-noise compared to low-coherence lasers. A stochastic model is discussed in the supplementary information section.

FIG. 21 shows the measured signal from an unfocused transducer at different lateral distances. The depth of focus of the detection beam on the sample is measured as $\sim 50 \mu\text{m}$. Therefore, the focus spot diameter of the detection beam on the sample is measured as $\sim 7 \mu\text{m}$. We also measured the photoacoustic signal generated by different dye concentrations to prove the concept of optical absorption for photoacoustic imaging. It is shown in FIG. 15 that the sample with higher absorption coefficient provides bigger photoacoustic signals as expected.

Discussion

In summary the results above showed that: (1) the PARS signal strength is proportional to optical absorption; (2) the PARS signal strength increases linearly with both signal- and reference beam intensities; (3) the PARS signals are largest at the optical focal zone (4) the detected signals are indeed photoacoustic signals; (4) signal maximization occurs when excitation and detection beams are confocal; (5) long-coherence of the probe beam is important for signal-to-noise; (6) PARS signal detection is possible at superficial depths in multiply scattering tissue; (7) lateral resolution will principally be determined by the excitation spot size; (8) axial resolution will principally be determined by detection system bandwidth; and (9) depth sectioning can be achieved with high numerical aperture objectives while extended depth-of-field can be achieved by harnessing chromatic aberration with multi-spectral excitation source.

As will be understood, the high sensitivity and the fine resolution of the proposed system offer performance comparable to other in vivo optical resolution photoacoustic microscopy systems but with much higher signal to noise ratio and in a non-contact reflection mode suitable for many clinical and pre-clinical applications. In this method a multi-wavelength fiber laser in the visible range has been used in multi focus form to generate photoacoustic signals and the acoustic signatures have been interrogated using a long-coherence length probe beam co-focused and co-aligned and co-scanned with the excitation spots.

Selecting an excitation laser may involve the following considerations. First, it should be capable of producing suitable conditions of stress- and thermal-confinement that both heat- and stress-build up during the course of a laser pulse before the energy can propagate away in the form of thermal diffusion or acoustic propagation. Stress confinement is the most stringent of the two criteria. For example, for $2 \mu\text{m}$ focused excitation spot, laser pulses should be preferably shorter than $2 \mu\text{m}/1500 \text{ m/s} = 1.3 \text{ ns}$, which would require a laser with pulse widths of a nanosecond or shorter. Second, the repetition rate of the laser will determine the imaging frame rate: the faster the repetition rate the higher. However, the repetition rate is preferably not so high that signals from previous pulses overlap in time with subsequent pulses. (Even that the imaging depth is not likely to be more than about a transport mean-free path ($\sim 1 \text{ mm}$ in tissue) the maximum pulse-repetition rate is preferably on the order of 1 MHz. Pulse energy should be such that ANSI limits are met at the tissue surface, requiring sub-0 levels of pulse energy. Finally the wavelength of the excitation source should preferably be tunable for multi-spectral imaging purposes. These considerations are similar to those for OR-PAM and it is known that fiber lasers can be a good choice. In one example, a frequency doubled Ytterbium-doped fiber laser was used to achieve tenability inject μJ -scale ns-pulses at 532-nm into a length of nonlinear fiber

to generate Stimulated Raman Scattering peaks. A range of wavelengths may be generated using this technique with enough pulse energy for OR-PAM. Few other sources are capable of meeting this range of requirements.

The selection of the interrogation laser is also important. As described above, the linewidth of the probe laser should be significantly smaller than the acoustic frequencies to be detected otherwise significant noise power from the laser source could leak into the passband of the system and degrade SNR. In one example, a laser was used with a 100 KHz linewidth, which is significantly smaller than the MHz-level frequencies to be detected. The laser is preferably tunable in wavelength and power, although wavelength tuning is not critical. A wavelength of 1550 nm may be used with a 532 nm excitation light because it is spectrally different (important so that optical filters can prevent excitation light from hitting the detector) and because it is a key band in optical communications where a plethora of components are available. Water absorption at this wavelength is higher than desirable: the 1/e penetration depth is a few mm. Other wavelength bands could also be used.

In this patent document, the word "comprising" is used in its non-limiting sense to mean that items following the word are included, but items not specifically mentioned are not excluded. A reference to an element by the indefinite article "a" does not exclude the possibility that more than one of the elements is present, unless the context clearly requires that there be one and only one of the elements.

The scope of the following claims should not be limited by the preferred embodiments set forth in the examples above and in the drawings, but should be given the broadest interpretation consistent with the description as a whole.

The invention claimed is:

1. A photoacoustic remote sensing system (PARS) for imaging a subsurface structure in a sample, comprising:

a pulsed or intensity-modulated excitation beam configured to generate ultrasonic signals in the sample at an excitation location;

a pulsed, continuous, or intensity-modulated interrogation beam incident on the sample at the excitation location, a portion of the interrogation beam returning from the sample that is indicative of the generated ultrasonic signals;

an optical system configured to focus the interrogation beam at a second focal point below the surface of the sample; and

a detector that detects the returning portion of the interrogation beam.

2. The PARS of claim 1, wherein the optical system is configured to focus the excitation beam at a first focal point below the surface of the sample, and the first and second focal points are within 1 mm of the surface of the sample.

3. The PARS of claim 1, wherein the optical system is configured to focus the excitation beam at a first focal point below the surface of the sample, and the first and second focal points are greater than 1 μ m below the surface of the sample.

4. The PARS of claim 1, wherein the optical system is configured to focus the excitation beam at a first focal point below the surface of the sample, and the first focal point or the second focal point is spaced below the surface of the sample at a depth that is greater than a focal zone of the excitation beam or the interrogation beam.

5. The PARS of claim 1, wherein the excitation beam and the interrogation beam have a lateral separation of less than 1 mm on the sample, or less than 0.5 mm on the sample.

6. The PARS of claim 1, wherein the optical system is configured to focus the excitation beam at a first focal point below the surface of the sample, and the first focal point is laterally within a focal zone of the interrogation beam.

7. The PARS of claim 1, wherein the second focal point is laterally within a focal zone of the excitation beam.

8. The PARS of claim 1, further comprising a processor that calculates an image of the sample based on the returning portion of the interrogation beam.

9. The PARS of claim 1, wherein the optical system is configured to focus the excitation beam at a first focal point below the surface of the sample, and at least one of the first focal point or the second focal point has a focal diameter of less than 30 μ m, less than 10 μ m, or less than 1 μ m.

10. The PARS of claim 1, wherein the excitation beam is scanned through the sample while the interrogation beam is stationary.

11. The PARS of claim 1, wherein the interrogation beam is scanned through the sample while the excitation beam is stationary.

12. The PARS of claim 1, wherein one or more of the interrogation beam or the excitation beam is stationary.

13. The PARS of claim 1, wherein the detector is an interferometer.

14. The PARS of claim 13, wherein the interferometer is a common path interferometer, a Michelson interferometer, a Fizeau interferometer, a Ramsey interferometer, a Sagnac interferometer, a Fabry-Perot interferometer, or a Mach-Zehnder interferometer.

15. The PARS of claim 1, wherein the detector includes a photodiode, an avalanche photodiode, a phototube, a photomultiplier, a CMOS camera, a CCD camera, a EM-CCD camera, an intensified-CCD camera, a back-thinned CCD camera, or a cooled CCD camera.

16. The PARS of claim 1, wherein the excitation beam and the interrogation beam are confocal.

17. The PARS of claim 1, further including an additional imaging system selected from optical coherence tomography, optical resolution photoacoustic microscopy, common photoacoustic microscopy, fluorescence microscopy, two-photon fluorescence microscopy, confocal fluorescence microscopy, Coherent-Anti-Raman-Stokes microscopy, Raman microscopy, or other photoacoustic systems.

18. The PARS of claim 1, wherein the PARS is used in one or more of the following applications:

imaging of blood oxygen saturation;

imaging of tumor hypoxia;

imaging of wound healing, burn diagnostics, or surgery;

imaging of microcirculation;

blood oxygenation parameter imaging;

estimating blood flow in vessels flowing into and out of a region of tissue;

imaging of molecularly-specific targets;

imaging angiogenesis for pre-clinical tumor models;

clinical imaging of micro- and macro-circulation and pigmented cells;

imaging of the eye;

augmenting or replacing fluorescein angiography;

imaging dermatological lesions;

imaging melanoma;

imaging basal cell carcinoma;

imaging hemangioma;

imaging psoriasis;

imaging eczema;

imaging dermatitis;

imaging Mohs surgery;

imaging to verify tumor margin resections;

imaging peripheral vascular disease;
 imaging diabetic and/or pressure ulcers;
 burn imaging;
 plastic surgery;
 microsurgery;
 imaging of circulating tumor cells;
 imaging melanoma cells;
 imaging lymph node angiogenesis;
 imaging response to photodynamic therapies;
 imaging response to photodynamic therapies having vascular ablative mechanisms;
 imaging response to chemotherapeutics;
 imaging response to anti-angiogenic drugs;
 imaging response to radiotherapy;
 estimating oxygen saturation using multi-wavelength photoacoustic excitation;
 estimating venous oxygen saturation where pulse oximetry cannot be used;
 estimating cerebrovenous oxygen saturation and/or central venous oxygen saturation;
 estimating oxygen flux and/or oxygen consumption;
 imaging vascular beds and depth of invasion in Barrett's esophagus and/or colorectal cancers;
 functional imaging during brain surgery;
 assessment of internal bleeding and/or cauterization verification;
 imaging perfusion sufficiency of organs and/or organ transplants;
 imaging angiogenesis around islet transplants;
 imaging of skin-grafts;
 imaging of tissue scaffolds and/or biomaterials to evaluate vascularization and/or immune rejection;
 imaging to aid microsurgery;
 guidance to avoid cutting blood vessels and/or nerves;
 imaging of contrast agents in clinical or pre-clinical applications;
 identification of sentinel lymph nodes;
 non- or minimally-invasive identification of tumors in lymph nodes;
 imaging of genetically-encoded reporters, wherein the genetically-encoded reporters include tyrosinase, chromoproteins, and/or fluorescent proteins for pre-clinical or clinical molecular imaging applications;
 imaging actively or passively targeted optically absorbing nanoparticles for molecular imaging;
 imaging of blood clots; or
 staging an age of blood clots.

19. A photoacoustic remote sensing system (PARS) for imaging a subsurface structure in a sample, comprising:
 one or more lasers configured to generate a pulsed or intensity-modulated excitation beam configured to generate ultrasonic signals in the sample at an excitation location;
 wherein the one or more lasers are also configured to generate a pulsed, intensity-modulated, or continuous interrogation beam incident on the sample at the excitation location, a portion of the interrogation beam returning from the sample that is indicative of the generated ultrasonic signals;
 an optical system configured to focus the interrogation beam at a second focal point that is below the surface of the sample; and
 a detector configured to detect the returning portion of the interrogation beam.

20. The PARS of claim 19, wherein the detector detects ultrasonic signals to a depth of 7 cm within the sample.

21. The PARS of claim 19, wherein the optical system is configured to focus the excitation beam at a first focal point below the surface of the sample, and the first and second focal points are within 1 mm of the surface of the sample.

22. The PARS of claim 19, wherein the PARS is used in one or more of the following applications:
 imaging of blood oxygen saturation;
 imaging of tumor hypoxia;
 imaging of wound healing, burn diagnostics, or surgery;
 imaging of microcirculation;
 blood oxygenation parameter imaging;
 estimating blood flow in vessels flowing into and out of a region of tissue;
 imaging of molecularly-specific targets;
 imaging angiogenesis for pre-clinical tumor models;
 clinical imaging of micro- and macro-circulation and pigmented cells;
 imaging of the eye;
 augmenting or replacing fluorescein angiography;
 imaging dermatological lesions;
 imaging melanoma;
 imaging basal cell carcinoma;
 imaging hemangioma;
 imaging psoriasis;
 imaging eczema;
 imaging dermatitis;
 imaging Mohs surgery;
 imaging to verify tumor margin resections;
 imaging peripheral vascular disease;
 imaging diabetic and/or pressure ulcers;
 burn imaging;
 plastic surgery;
 microsurgery;
 imaging of circulating tumor cells;
 imaging melanoma cells;
 imaging lymph node angiogenesis;
 imaging response to photodynamic therapies;
 imaging response to photodynamic therapies having vascular ablative mechanisms;
 imaging response to chemotherapeutics;
 imaging response to anti-angiogenic drugs;
 imaging response to radiotherapy;
 estimating oxygen saturation using multi-wavelength photoacoustic excitation;
 estimating venous oxygen saturation where pulse oximetry cannot be used;
 estimating cerebrovenous oxygen saturation and/or central venous oxygen saturation;
 estimating oxygen flux and/or oxygen consumption;
 imaging vascular beds and depth of invasion in Barrett's esophagus and/or colorectal cancers;
 functional imaging during brain surgery;
 assessment of internal bleeding and/or cauterization verification;
 imaging perfusion sufficiency of organs and/or organ transplants;
 imaging angiogenesis around islet transplants;
 imaging of skin-grafts;
 imaging of tissue scaffolds and/or biomaterials to evaluate vascularization and/or immune rejection;
 imaging to aid microsurgery;
 guidance to avoid cutting blood vessels and/or nerves;
 imaging of contrast agents in clinical or pre-clinical applications;
 identification of sentinel lymph nodes;
 non- or minimally-invasive identification of tumors in lymph nodes;

25

imaging of genetically-encoded reporters, wherein the genetically-encoded reporters include tyrosinase, chromoproteins, and/or fluorescent proteins for pre-clinical or clinical molecular imaging applications;
 imaging actively or passively targeted optically absorbing nanoparticles for molecular imaging;
 imaging of blood clots; or
 staging an age of blood clots.

23. The PARS of claim 19, wherein the detector is an interferometer.

24. The PARS of claim 23, wherein the interferometer is a common path interferometer, a Michelson interferometer, a Fizeau interferometer, a Ramsey interferometer, a Sagnac interferometer, a Fabry-Perot interferometer, or a Mach-Zehnder interferometer.

25. An endoscopic device that uses a photoacoustic remote sensing system (PARS) for imaging a subsurface structure in a sample, the endoscopic device comprises:

a fiber optic cable having an input end and a detection end;

a pulsed or intensity-modulated excitation beam at the input end of the fiber optic cable, wherein, in use, the excitation beam generates ultrasonic signals in the sample at an excitation location that is adjacent to the detection end of the fiber optic cable, the fiber optic cable focusing the excitation beam at a first focal point that is below the surface of and within the sample;

a pulsed, intensity-modulated, or continuous interrogation beam at the input end of the fiber optic cable and incident on the excitation location, the fiber optic cable focusing the interrogation beam at a second focal point that is below the surface of and within the sample, wherein a portion of the interrogation beam that is indicative of the generated ultrasonic signals is received by the detection end of the fiber optic cable and travels to the input end;

a detector configured to receive the returning portion of the interrogation beam at the input end of the fiber optic cable.

26. The endoscopic device of claim 25, wherein the endoscopic device is used in one or more of the following applications:

- imaging of blood oxygen saturation;
- imaging of tumor hypoxia;
- imaging of wound healing, burn diagnostics, or surgery;
- imaging of microcirculation;
- blood oxygenation parameter imaging;
- estimating blood flow in vessels flowing into and out of a region of tissue;
- imaging of molecularly-specific targets;
- imaging angiogenesis for pre-clinical tumor models;
- clinical imaging of micro- and macro-circulation and pigmented cells;
- imaging of the eye;
- augmenting or replacing fluorescein angiography;
- imaging dermatological lesions;
- imaging melanoma;

26

- imaging basal cell carcinoma;
- imaging hemangioma;
- imaging psoriasis;
- imaging eczema;
- imaging dermatitis;
- imaging Mohs surgery;
- imaging to verify tumor margin resections;
- imaging peripheral vascular disease;
- imaging diabetic and/or pressure ulcers;
- burn imaging;
- plastic surgery;
- microsurgery;
- imaging of circulating tumor cells;
- imaging melanoma cells;
- imaging lymph node angiogenesis;
- imaging response to photodynamic therapies;
- imaging response to photodynamic therapies having vascular ablative mechanisms;
- imaging response to chemotherapeutics;
- imaging response to anti-angiogenic drugs;
- imaging response to radiotherapy;
- estimating oxygen saturation using multi-wavelength photoacoustic excitation;
- estimating venous oxygen saturation where pulse oximetry cannot be used;
- estimating cerebrovenous oxygen saturation and/or central venous oxygen saturation;
- estimating oxygen flux and/or oxygen consumption;
- imaging vascular beds and depth of invasion in Barrett's esophagus and/or colorectal cancers;
- functional imaging during brain surgery;
- assessment of internal bleeding and/or cauterization verification;
- imaging perfusion sufficiency of organs and/or organ transplants;
- imaging angiogenesis around islet transplants;
- imaging of skin-grafts;
- imaging of tissue scaffolds and/or biomaterials to evaluate vascularization and/or immune rejection;
- imaging to aid microsurgery;
- guidance to avoid cutting blood vessels and/or nerves;
- imaging of contrast agents in clinical or pre-clinical applications;
- identification of sentinel lymph nodes;
- non- or minimally-invasive identification of tumors in lymph nodes;
- imaging of genetically-encoded reporters, wherein the genetically-encoded reporters include tyrosinase, chromoproteins, and/or fluorescent proteins for pre-clinical or clinical molecular imaging applications;
- imaging actively or passively targeted optically absorbing nanoparticles for molecular imaging;
- imaging of blood clots; or
 staging an age of blood clots.

27. The endoscopic device of claim 25, wherein the detector is an interferometer.

* * * * *

专利名称(译)	光声遥感 (PARS)		
公开(公告)号	US10682061	公开(公告)日	2020-06-16
申请号	US16/116038	申请日	2018-08-29
[标]发明人	REZA PARSIN HAJI ZEMP ROGER		
发明人	REZA, PARSIN HAJI ZEMP, ROGER		
IPC分类号	A61B5/00 G01N21/17		
CPC分类号	G01N21/1702 A61B5/0095 G01N21/1717 G01N2201/06113 G01N2021/1706 G01N2201/08 G01N2201/0846 G01N2021/1725		
优先权	62/067197 2014-10-22 US		
其他公开文献	US20190104944A1		
外部链接	Espacenet		

摘要(译)

用于对样品中的地下结构成像的光声遥感系统 (PARS) 具有激发束, 该激发束被配置为在激发位置处在样品中产生超声信号; 在激发位置处入射到样品上的询问光束, 该询问光束的一部分从样品返回, 指示产生的超声信号; 一种光学系统, 其聚焦光束和激发光束和询问光束中的至少一个在样品表面以下; 探测器检测询问光束的返回部分。

