



US 20190076054A1

(19) **United States**

(12) **Patent Application Publication**
BANET et al.

(10) **Pub. No.: US 2019/0076054 A1**

(43) **Pub. Date: Mar. 14, 2019**

(54) **BODY-WORN MONITOR FOR MEASURING RESPIRATION RATE**

A61B 5/0402 (2006.01)

A61B 5/053 (2006.01)

A61B 5/024 (2006.01)

(71) Applicant: **SOTERA WIRELESS, INC.**, San Diego, CA (US)

(52) **U.S. Cl.**

CPC *A61B 5/0816* (2013.01); *A61B 2562/0219*

(2013.01); *A61B 5/721* (2013.01); *A61B 5/113*

(2013.01); *A61B 5/1118* (2013.01); *A61B*

5/1116 (2013.01); *A61B 5/0809* (2013.01);

A61B 5/0402 (2013.01); *A61B 5/6823*

(2013.01); *A61B 5/6824* (2013.01); *A61B*

5/681 (2013.01); *A61B 5/7239* (2013.01);

A61B 5/0002 (2013.01); *A61B 5/0535*

(2013.01); *A61B 5/6826* (2013.01); *A61B*

5/02416 (2013.01); *A61B 5/746* (2013.01);

A61B 5/0205 (2013.01)

(72) Inventors: **Matthew BANET**, San Diego, CA (US); **Devin McCOMBIE**, Carlsbad, CA (US); **Marshal DHILLON**, San Diego, CA (US)

(73) Assignee: **SOTERA WIRELESS, INC.**, SAN DIEGO, CA (US)

(21) Appl. No.: **16/188,125**

(22) Filed: **Nov. 12, 2018**

Related U.S. Application Data

(63) Continuation of application No. 12/559,429, filed on Sep. 14, 2009, now Pat. No. 10,123,722.

Publication Classification

(51) **Int. Cl.**

A61B 5/08 (2006.01)

A61B 5/0205 (2006.01)


A61B 5/00 (2006.01)

A61B 5/113 (2006.01)

A61B 5/11 (2006.01)

(57) **ABSTRACT**

The invention provides a multi-sensor system that uses an algorithm based on adaptive filtering to monitor a patient's respiratory rate. The system features a first sensor which is selected from the group consisting of an impedance pneumography sensor, an ECG sensor, a PPG sensor, and a motion sensor (e.g., an accelerometer) configured to attach to the patient's torso and measure therefrom a motion signal. The system further comprises (iii) a processing system, configured to operably connect to the first and motion sensors, and to determine a respiration rate value by applying filter parameters obtained from the first sensor signals to the motion sensor signals.






FIG.1B

FIG.1A




FIG. 2B

FIG. 2A




FIG. 3A




FIG. 3B




FIG. 3C




FIG. 3D




FIG. 3E




FIG. 3F




FIG. 3G



















FIG. 3H
































FIG.13









FIG. 14 From Figure 13











FIG. 15F











FIG. 16F






FIG. 17F




FIG. 18C




FIG. 19




FIG. 20A




FIG. 20B




FIG. 20C




FIG. 20D









FIG. 20E











(positioned on wrist
in wrist-worn transceiver)

FIG.25




FIG. 26A





FIG. 26B

Configuration for Indexing cNIBP Measurement
(first ~60 seconds)



Configuration for cNIBP/RR Measurements
(remaining 3 hours, 59 minutes)




FIG. 27A

FIG. 27B




FIG. 28




FIG. 29A




FIG. 29B

BODY-WORN MONITOR FOR MEASURING RESPIRATION RATE

CROSS REFERENCES TO RELATED APPLICATIONS

[0001] This application is a continuation of U.S. patent application Ser. No. 12/559,429, filed Sep. 14, 2009, now U.S. Pat. No. 10,123,722, the contents of which are incorporated herein by reference.

BACKGROUND OF THE INVENTION

Field of the Invention

[0002] The present invention relates to medical devices for monitoring vital signs, e.g., respiration rate.

Description of the Related Art

[0003] Respiration rate (RR) is a vital sign typically measured in the hospital using either an indirect electrode-based technique called 'impedance pneumography' (IP), a direct optical technique called 'end-tidal CO₂' (et-CO₂), or simply through manual counting of breaths by a medical professional. IP is typically used in lower-acuity areas of the hospital, and uses the same electrodes deployed in a conventional 'Einthoven's triangle' configuration for measuring heart rate (HR) from an electrocardiogram (ECG). One of the electrodes supplies a low-amperage (~4 mA) current that is typically modulated at a high frequency (~50-100 kHz). Current passes through a patient's chest cavity, which is characterized by a time-dependent capacitance that varies with each breath. A second electrode detects the current, which is modulated by the changing capacitance. Ultimately this yields an analog signal that is processed with a series of amplifiers and filters to detect the time-dependent capacitance change and, subsequently, the patient's RR.

[0004] In et-CO₂, a device called a capnometer features a small plastic tube that typically inserts in the patient's mouth. With each breath the tube collects expelled CO₂. A beam of infrared radiation emitted from an integrated light source passes through the CO₂ and is absorbed in a time-dependent manner that varies with the breathing rate. A photodetector and series of processing electronics analyze the transmitted signal to determine RR. et-CO₂ systems are typically used in high-acuity areas of the hospital, such as the intensive care unit (ICU), where patients often use ventilators to assist them in breathing.

[0005] In yet another technique, RR is measured from the envelope of a time-dependent optical waveform called a photoplethysmogram (PPG) that is measured from the index finger during a conventional measurement of the patient's oxygen saturation (SpO₂). Breathing changes the oxygen content in the patient's blood and, subsequently, its optical absorption properties. Such changes cause a slight, low-frequency variation in the PPG that can be detected with a pulse oximeter's optical system, which typically operates at both red and infrared wavelengths.

[0006] Not surprisingly, RR is an important predictor of a decompensating patient. For example, a study in 1993 concluded that a respiratory rate greater than 27 breaths/minute was the most important predictor of cardiac arrests in hospital wards (Fieslmann et al., 'Respiratory rate predicts cardiopulmonary arrest for internal medicine patients', *J Gen Intern Med* 1993; 8: 354-360). Subbe et al. found that,

in unstable patients, relative changes in respiratory rate were much greater than changes in heart rate or systolic blood pressure, and thus that the respiratory rate was likely to be a better means of discriminating between stable patients and patients at risk (Subbe et al., 'Effect of introducing the Modified Early Warning score on clinical outcomes, cardiopulmonary arrests and intensive care utilization in acute medical admissions', *Anaesthesia* 2003; 58: 797-802). Goldhill et al. reported that 21% of ward patients with a respiratory rate of 25-29 breaths/minute assessed by a critical care outreach service died in hospital (Goldhill et al., 'A physiologically-based early warning score for ward patients: the association between score and outcome', *Anaesthesia* 2005; 60: 547-553). Those with a higher respiratory rate had an even higher mortality rate. In another study, just over half of all patients suffering a serious adverse event on the general wards (e.g. a cardiac arrest or ICU admission) had a respiratory rate greater than 24 breaths/minute. These patients could have been identified as high risk up to 24 hours before the event with a specificity of over 95% (Cretikos et al., 'The Objective Medical Emergency Team Activation Criteria: a case-control study', *Resuscitation* 2007; 73: 62-72). Medical references such as these clearly indicate that an accurate, easy-to-use device for measuring respiratory rate is important for patient monitoring within the hospital.

[0007] Despite its importance and the large number of available monitoring techniques, RR is notoriously difficult to measure, particularly when a patient is moving. During periods of motion, non-invasive techniques based on IP and PPG signals are usually overwhelmed by artifacts and thus completely ineffective. This makes it difficult or impossible to measure RR from an ambulatory patient. Measurements based on et-CO₂ are typically less susceptible to motion, but require a plastic tube inserted in the patient's mouth, which is typically impractical for ambulatory patients.

SUMMARY OF THE INVENTION

[0008] This invention provides methods, devices, and systems for use in measuring RR using multiple input signals, including IP, PPG, and ECG waveforms, and a signal processing technique based on adaptive filtering. After being measured with a body-worn system, these waveforms are processed along with those from an accelerometer mounted on the patient's torso (most typically the chest or abdomen). The accelerometer measures small, breathing-induced movements to generate a time-dependent waveform (ACC). With adaptive filtering, an initial RR is preferably estimated from the IP waveform, and alternatively from the PPG or ECG waveform. The initial RR is then processed and used to determine parameters for a bandpass digital filter, typically implemented with a finite impulse response function. This yields a customized filtering function which then processes the ACC waveform. The filtering function generates a relatively noise-free ACC waveform with well-defined pulses corresponding to RR. Each pulse can then be further processed and counted to determine an accurate RR value, even during periods of motion.

[0009] The body-worn monitor measures IP, PPG, ECG, and ACC waveforms with a series of sensors integrated into a comfortable, low-profile system that preferably communicates wirelessly with a remote computer in the hospital. The system typically features three accelerometers, each configured to measure a unique signal along its x, y, and z

axes, to yield a total of nine ACC waveforms. In certain embodiments, the accelerometers are deployed on the patient's torso, upper arm, and lower arm, and may be embedded in the monitor's cabling or processing unit. Each ACC waveform can be additionally processed to determine the patient's posture, degree of motion, and activity level. These parameters serve as valuable information that can ultimately reduce occurrences of 'false positive' alarms/alerts in the hospital. For example, if processing of additional ACC waveforms indicates a patient is walking, then their RR rate, which may be affected by walking-induced artifacts, can be ignored by an alarm/alert engine associated with the body-worn monitor. The assumption in this case is that a walking patient is likely relatively healthy, regardless of their RR value. Perhaps more importantly, with a conventional monitoring device a walking patient may yield a noisy IP signal that is then processed to determine an artificially high RR, which then triggers a false alarm. Such a situation can be avoided with an independent measurement of motion, such as that described herein. Other heuristic rules based on analysis of ACC waveforms may also be deployed according to this invention.

[0010] Sensors attached to the wrist and bicep each measure signals that are collectively analyzed to estimate the patient's arm height; this can be used to improve accuracy of a continuous blood pressure measurement (cNIBP), as described below, that measures systolic (SYS), diastolic (DIA), and mean (MAP) arterial blood pressures. And the sensor attached to the patient's chest measures signals that are analyzed to determine posture and activity level, which can affect measurements for RR, SpO₂, cNIBP, and other vital signs. Algorithms for processing information from the accelerometers for these purposes are described in detail in the following patent applications, the contents of which are fully incorporated herein by reference: BODY-WORN MONITOR FEATURING ALARM SYSTEM THAT PROCESSES A PATIENT'S MOTION AND VITAL SIGNS (U.S. Ser. No. 12/469,182; filed May 20, 2009) and BODY-WORN VITAL SIGN MONITOR WITH SYSTEM FOR DETECTING AND ANALYZING MOTION (U.S. Ser. No. 12/469,094; filed May 20, 2009). As described therein, knowledge of a patient's motion, activity level, and posture can greatly enhance the accuracy of alarms/alerts generated by the body-worn monitor.

[0011] The body-worn monitor features systems for continuously monitoring patients in a hospital environment, and as the patient transfers from different areas in the hospital, and ultimately to the home. Both SpO₂ and cNIBP rely on accurate measurement of PPG and ACC waveforms, along with an ECG, from patients that are both moving and at rest. cNIBP is typically measured with the 'Composite Technique', which is described in detail in the co-pending patent application entitled: VITAL SIGN MONITOR FOR MEASURING BLOOD PRESSURE USING OPTICAL, ELECTRICAL, AND PRESSURE WAVEFORMS (U.S. Ser. No. 12/138,194; filed Jun. 12, 2008), the contents of which are fully incorporated herein by reference.

[0012] As described in these applications, the Composite Technique (or, alternatively, the 'Hybrid Technique' referred to therein) typically uses a single PPG waveform from the SpO₂ measurement (typically generated with infrared radiation), along with the ECG waveform, to calculate a parameter called 'pulse transit time' (PTT) which strongly correlates to blood pressure. Specifically, the ECG waveform

features a sharply peaked QRS complex that indicates depolarization of the heart's left ventricle, and, informally, provides a time-dependent marker of a heart beat. PTT is the time separating the peak of the QRS complex and the onset, or 'foot', of the PPG waveforms. The QRS complex, along with the foot of each pulse in the PPG, can be used to more accurately extract AC signals using a mathematical technique described in detail below. In other embodiments both the red and infrared PPG waveforms are collectively processed to enhance the accuracy of the cNIBP measurement.

[0013] In certain embodiments, the electrical system for measuring RR features a small-scale, low-power circuit mounted on a circuit board that fits within the wrist-worn transceiver. The transceiver additionally includes a touch-panel display, barcode reader, and wireless systems for ancillary applications described, for example, in the above-referenced applications, the contents of which have been previously incorporated herein by reference.

[0014] In one aspect, the invention provides a multi-sensor system that uses an algorithm based on adaptive filtering to monitor a patient's RR. The system features a first sensor selected from the following group: i) an IP sensor featuring at least two electrodes and an IP processing circuit configured to measure an IP signal; ii) an ECG sensor featuring at least two electrodes and an ECG processing circuit configured to measure an ECG signal; and iii) a PPG sensor featuring a light source, photodetector, and PPG processing circuit configured to measure a PPG signal. Each of these sensors measures a time-dependent signal which is sensitive to RR and is processed to determine an initial RR value. The system features a second sensor (e.g. a digital 3-axis accelerometer) that attaches to the patient's torso and measures an ACC signal indicating movement of the chest or abdomen that is also sensitive to RR.

[0015] A body-worn processing system receives a first signal representing at least one of the IP, ECG, and PPG signals, and a second signal representing the ACC signal. The processing system is configured to: i) process the first signal to determine an initial RR; ii) process the second signal with a digital filter determined from the initial RR to determine a third signal; and iii) process the third signal to determine a final value for the patient's RR.

[0016] The processing system, as described herein, can include one or more microprocessors. For example, it can include first microprocessor embedded within a single ASIC that also measures IP and ECG, or mounted on a circuit board that also contains the ASIC or an equivalent circuit made from discrete components. In these cases the first microprocessor is mounted on the patient's torso. A wrist-worn transceiver can contain the second microprocessor. In embodiments, the first microprocessor mounted on the patient's torso determines a RR from multiple time-dependent signals; this value is transmitted to the second microprocessor within the wrist-worn transceiver as a digital or analog data stream transmitted through a cable. The second microprocessor further processes the RR value alongside data describing the patient's motion and other vital signs. The secondary processing, for example, can be used to generate alarms/alerts based on RR, or suppress alarms/alerts because of the patient's motion.

[0017] In embodiments, the digital filter used for adaptive filtering is a bandpass filter or low-pass filter. Typically the digital filter is determined from a finite impulse response function. The bandpass filter typically features an upper

frequency limit determined from a multiple (e.g. 1-3×) of the initial RR. Such a digital filter is used to process time-dependent waveforms to remove noise and other artifacts to determine the initial version of RR. In this case the filter is not adaptive, and instead has a pre-determined passband. The final version of RR is determined from the adaptive filter, which as described above has a passband that depends on the initial version of RR.

[0018] In other embodiments, the processing system is further configured to determine both initial and final versions of RR by processing a filtered waveform with a mathematical derivative and then determine a zero-point crossing indicating a 'count' marking a respiratory event. Such counts are evident in the processed IP signal, which features a first series of pulses that, once analyzed by the processing system, yields the initial RR. Alternatively, the initial RR is determined from either an ECG or PPG, both of which feature a series of heartbeat-induced pulses with amplitudes characterized by a time-varying envelope, with the frequency of the envelope representing the initial RR. The waveforms used to determine the initial and final values for RR can be interchanged, e.g. the ACC waveform can be processed to determine the initial RR value, and this can then be used to design a digital filter that processes the IP, ECG, or PPG waveforms to determine the final RR value. In general, according to the invention, any combination of the above-described waveforms can be used in the adaptive filtering process to determine the initial and final RR values.

[0019] In another aspect, the invention provides a system for monitoring a patient's RR that also accounts for their posture, activity level, and degree of motion. Such patient states can result in artifacts that affect the RR measurement, and thus proper interpretation of them can reduce the occurrence of erroneous RR values and ultimately false alarms/alerts in the hospital.

[0020] In another aspect, the invention provides a cable within a body-worn monitor that includes an IP system, a motion sensor (e.g. accelerometer), and a processing system that determines RR from signals generated by these sensors. These components, for example, can be included in a terminal end of the cable, typically worn on the patient's torso, which connects to a series of disposable electrodes that attach to the patient's body. A mechanical housing, typically made of plastic, covers these and other components, such as sensors for measuring signals relating to ECG and skin temperature.

[0021] In embodiments, the cable includes at least one conductor configured to transmit both a first digital data stream representing the digital IP signal or information calculated therefrom, and a second digital data stream representing the digital motion signal or information calculated therefrom. In other embodiments these signals are processed by a microprocessor on the chest to determine an RR value, and this value is then sent in the digital data stream to another processor, such as one within the wrist-worn transceiver, where it is further processed. To transmit the serial data stream, the terminal portion of the cable can include a transceiver component, e.g. a serial transceiver configured to transmit a digital data stream according to the CAN protocol. Other properties, such as heart rate, temperature, alarms relating to ECG signals, and other information relating to the CAN communication protocol and its timing can be transmitted by the transceiver component.

[0022] In embodiments, both the IP and ECG systems are contained within a single integrated circuit. The ECG system can be modular and determine multi-lead ECG signals, such as three, five, and twelve-lead ECG signals.

[0023] In another aspect, the invention provides a method for determining RR during periods of motion. The method includes the following steps: (a) measuring a first time-dependent signal by detecting a modulated electrical current passing through the patient's torso; (b) measuring a second time-dependent signal by detecting respiration-induced movements in the patient's torso with at least one motion sensor; (c) determining a motion-related event not related to the patient's respiration rate value by processing signals from the motion sensor; and (d) collectively processing both the first and second time-dependent signals to determine a value for RR corresponding to a period when the patient's motion-related event is below a pre-determined threshold. For example, the motion-related event determined during step (c) can be the patient's posture, activity level, or degree of motion. Typically these parameters are determined from signals measured with an accelerometer mounted on the patient's torso. These signals are processed with an algorithm, described in detail below, that yields a vector indicating orientation of the patient's chest and their subsequent posture. Specifically, an angle separating the vector from a pre-determined coordinate system ultimately yields posture, as is described in detail below. Activity level (corresponding, e.g., to moving, walking, falling, convulsing) can be calculated from a mathematical transform of time-dependent variations of a motion signal that yields a frequency-domain spectrum. Portions of the spectrum (e.g. the power of specific frequency components) are compared to pre-determined frequency parameters to determine the activity level. Other operations, such as a mathematical derivative of the time-dependent motion signal, or a series of 'decision rules' based on a decision-tree algorithm, can also yield the activity level.

[0024] In another aspect, the invention provides a method for suppressing alarms related to RR by processing the patient's posture, activity level, and degree of motion as determined by the accelerometer. For example, the alarm can be suppressed if the patient is standing upright, or if their posture changes from lying down to one of sitting and standing upright. Or the alarm can be suppressed if their posture changes from either standing upright or sitting to lying down. In general, a rapid change in posture, which can be determined with the chest-worn accelerometer, may disrupt the signals used to determine RR to the point where a false alarm/alert is generated. In this embodiment, posture is determined from the vector-based analysis, described above.

[0025] In yet another aspect, the invention provides a system for monitoring a patient's RR featuring a sensor unit configured to be mounted on the patient's torso. The sensor unit features IP and motion sensors, as described above, and additionally attaches directly to an electrode that secures the unit to the patient's torso (e.g. chest or abdomen). Here, a housing comprising the IP and motion sensors additionally includes a connector featuring an opening configured to receive a metal snap on the exterior of a conventional disposable electrode. Other electrodes used for IP and ECG measurements connect to the unit through cables. The unit can additionally send a digital data stream including RR data over a CAN bus to a wrist-worn transceiver, which as

described above can further process the RR value to account for alarms/alerts, motion, etc.

[0026] In all embodiments, the wrist-worn transceiver can include a display configured to display the patient's RR and other vital signs, along with a touchpanel interface. A wireless transceiver within the wrist-worn transceiver can transmit information to a remote computer using conventional protocols such as 802.11, 802.15.4, and cellular. The remote computer, for example, can be connected to a hospital network. It can also be a portable computer, such as a tablet computer, personal digital assistant, or cellular phone.

[0027] Many advantages are associated with this invention. In general, it provides an accurate measurement of RR, along with an independent measurement of a patient's posture, activity level, and motion, to characterize an ambulatory patient in the hospital. These parameters can be collectively analyzed to improve true positive alarms while reducing the occurrence of false positive alarms. Additionally, the measurement of RR is performed with a body-worn monitor that is comfortable, lightweight, and low-profile, making it particularly well suited for patients that are moving about. Such a monitor could continuously monitor a patient as, for example, they transition from the emergency department to the ICU, and ultimately to the home after hospitalization.

[0028] Still other embodiments are found in the following detailed description of the invention and in the claims.

BRIEF DESCRIPTION OF THE DRAWINGS

[0029] FIG. 1A shows a schematic view of a patient wearing accelerometers on their abdomen (position 1) and chest (position 2) to measure ACC waveforms and RR according to the adaptive filtering process of the invention; FIG. 1B shows a schematic view of the accelerometers from FIG. 1 along with their three-dimensional measurement axes;

[0030] FIG. 2A shows a schematic view of a patient wearing ECG electrodes on their chest in a conventional Einthoven's triangle configuration to measure an IP waveform; FIG. 2B shows a schematic view of ECG and IP circuits that simultaneously process signals from each ECG electrode in FIG. 2A to determine both ECG and IP waveforms;

[0031] FIGS. 3A-D each show an ACC waveform measured with the configuration shown in FIG. 1 after processing with no filter (FIG. 3A; top), a 0.01→1 Hz bandpass filter (FIG. 3B), a 0.01→0.5 Hz bandpass filter (FIG. 3C), and a 0.01→0.1 Hz bandpass filter (FIG. 3C; bottom); FIGS. 3E-H show, respectively, time-dependent derivatives of the ACC waveforms shown in FIGS. 3A-D;

[0032] FIGS. 4A-C show an ACC waveform filtered with a 0.01→0.1 Hz bandpass filter (FIG. 4A; top), an IP waveform (FIG. 4B), and a et-CO₂ waveform (FIG. 4C; bottom) simultaneously measured from a supine patient undergoing slow, deep breaths;

[0033] FIGS. 5A-C show an ACC waveform filtered with a 0.01→0.1 Hz bandpass filter (FIG. 5A; top), an IP waveform (FIG. 5B), and a et-CO₂ waveform (FIG. 5C; bottom) simultaneously measured from a supine patient undergoing fast, deep breaths;

[0034] FIGS. 6A-C show an ACC waveform filtered with a 0.01→0.1 Hz bandpass filter (FIG. 6A; top), an IP wave-

form (FIG. 6B), and a et-CO₂ waveform (FIG. 6C; bottom) simultaneously measured from a supine patient undergoing very fast, deep breaths;

[0035] FIGS. 7A-C show an ACC waveform filtered with a 0.01→0.1 Hz bandpass filter (FIG. 7A; top), an IP waveform (FIG. 7B), and a et-CO₂ waveform (FIG. 7C; bottom) simultaneously measured from a supine patient undergoing medium, shallow breaths;

[0036] FIGS. 8A-C show an ACC waveform filtered with a 0.01→0.1 Hz bandpass filter (FIG. 8A; top), an IP waveform (FIG. 8B), and a et-CO₂ waveform (FIG. 8C; bottom) simultaneously measured from a standing patient undergoing medium, shallow breaths;

[0037] FIGS. 9A-C show an ACC waveform filtered with a 0.01→0.1 Hz bandpass filter (FIG. 9A; top), an IP waveform (FIG. 9B), and a et-CO₂ waveform (FIG. 9C; bottom) simultaneously measured from a standing patient undergoing fast, deep breaths;

[0038] FIGS. 10A-C show an ACC waveform filtered with a 0.01→0.1 Hz bandpass filter (FIG. 10A; top), an IP waveform (FIG. 10B), and a et-CO₂ waveform (FIG. 10C; bottom) simultaneously measured from a supine patient undergoing slow, deep breaths, followed by a period of apnea, followed by relatively fast, deep breaths;

[0039] FIGS. 11A-C show an ACC waveform filtered with a 0.01→0.1 Hz bandpass filter (FIG. 11A; top), an IP waveform (FIG. 11B), and a et-CO₂ waveform (FIG. 11C; bottom) simultaneously measured from a supine patient undergoing very fast, shallow breaths, followed by a period of apnea, followed by relatively slow, shallow breaths;

[0040] FIGS. 12A-C show an ACC waveform filtered with a 0.01→0.1 Hz bandpass filter (FIG. 12A; top), an IP waveform (FIG. 12B), and a et-CO₂ waveform (FIG. 12C; bottom) simultaneously measured from a walking patient undergoing fast, deep breaths;

[0041] FIG. 13 shows a flow chart along with ACC and IP waveforms used to determine RR using an adaptive filtering technique;

[0042] FIG. 14 shows a flow chart that describes details of the adaptive filtering technique shown in FIG. 13;

[0043] FIGS. 15A-E show graphs of an ACC waveform filtered initially with a 0.01→2 Hz bandpass filter (FIG. 15A; top), an IP waveform filtered initially with a 0.01→12 Hz bandpass (FIG. 15B), an ACC waveform adaptively filtered with a bandpass filter ranging from 0.01 Hz to 1.5 times the breathing rate calculated from the IP waveform in FIG. 15B (FIG. 15C), a first derivative of the filtered waveform in FIG. 15C (FIG. 15D), and the adaptively filtered waveform in FIG. 15C along with markers (FIG. 15E; bottom) indicating slow, deep breaths as determined from the algorithm shown by the flow chart in FIG. 14; FIG. 15F is a flow chart showing the algorithmic steps used to process the waveforms shown in FIGS. 15A-E;

[0044] FIGS. 16A-E show graphs of an ACC waveform filtered initially with a 0.01→2 Hz bandpass filter (FIG. 16A; top), an IP waveform filtered initially with a 0.01→12 Hz bandpass (FIG. 16B), an ACC waveform adaptively filtered with a bandpass filter ranging from 0.01 Hz to 1.5 times the breathing rate calculated from the IP waveform in FIG. 16B (FIG. 16C), a first derivative of the filtered waveform in FIG. 16C (FIG. 16D), and the adaptively filtered waveform in FIG. 16C along with markers (FIG. 16E; bottom) indicating fast, deep breaths as determined from the algorithm shown by the flow chart in FIG. 14; FIG.

16F is a flow chart showing the algorithmic steps used to process the waveforms shown in FIGS. 16A-E;

[0045] FIGS. 17A-E show graphs of an ACC waveform filtered initially with a 0.01→2 Hz bandpass filter (FIG. 17A; top), an IP waveform filtered initially with a 0.01→12 Hz bandpass (FIG. 17B), an ACC waveform adaptively filtered with a bandpass filter ranging from 0.01 Hz to 1.5 times the breathing rate calculated from the IP waveform in FIG. 17B (FIG. 17C), a first derivative of the filtered waveform in FIG. 17C (FIG. 17D), and the adaptively filtered waveform in FIG. 17C along with markers (FIG. 17E; bottom) indicating very fast, deep breaths as determined from the algorithm shown by the flow chart in FIG. 14; FIG. 17F is a flow chart showing the algorithmic steps used to process the waveforms shown in FIGS. 17A-E;

[0046] FIGS. 18A-B show graphs of an ACC waveform filtered initially with a 0.01→2 Hz bandpass filter (FIG. 18A; top), and an IP waveform filtered initially with a 0.01→12 Hz bandpass (FIG. 18B; bottom) measured from a walking patient; FIG. 18C is a flow chart showing the algorithmic steps used to process the waveforms shown in FIGS. 18A-B;

[0047] FIG. 19 is a graph showing correlation between respiratory rates measured with the adaptive filtering technique shown by the flow chart in FIG. 14 and et-CO₂;

[0048] FIG. 20A is a graph showing an unfiltered ECG waveform measured from a resting patient; FIG. 20B is a graph showing the time-dependent envelope of the ECG waveform shown in FIG. 20A; FIG. 20C is a graph showing an unfiltered PPG waveform measured simultaneously with the ECG waveform of FIG. 20A; FIG. 20D is a graph showing the time-dependent envelope of the PPG waveform shown in FIG. 20C; FIG. 20E is a graph showing an IP waveform measured simultaneously with the ECG waveform of FIG. 20A and the PPG waveform of FIG. 20C;

[0049] FIGS. 21A-C show graphs of time-dependent ECG waveforms (FIG. 21A; top), PPG waveforms (FIG. 21B), and ACC waveforms (FIG. 21C; bottom) measured along the x, y, and z-axes for a resting patient;

[0050] FIGS. 22A-C show graphs of time-dependent ECG waveforms (FIG. 22A; top), PPG waveforms (FIG. 22B), and ACC waveforms (FIG. 22C; bottom) measured along the x, y, and z-axes for a walking patient;

[0051] FIGS. 23A-C show graphs of time-dependent ECG waveforms (FIG. 23A; top), PPG waveforms (FIG. 23B), and ACC waveforms (FIG. 23C; bottom) measured along the x, y, and z-axes for a convulsing patient;

[0052] FIGS. 24A-C show graphs of time-dependent ECG waveforms (FIG. 24A; top), PPG waveforms (FIG. 24B), and ACC waveforms (FIG. 24C; bottom) measured along the x, y, and z-axes for a falling patient;

[0053] FIG. 25 shows a schematic view of the patient of FIG. 1 and a coordinate axis used with an algorithm and ACC waveforms to determine the patient's posture;

[0054] FIG. 26A shows a graph of time-dependent ACC waveforms measured from a patient's chest during different postures; FIG. 26B shows a graph of time-dependent postures determined by processing the ACC waveforms of FIG. 26A with an algorithm and coordinate axis shown in FIG. 25;

[0055] FIGS. 27A and 27B show, respectively, a three-dimensional image of the body-worn monitor of the invention attached to a patient during and after an initial indexing measurement;

[0056] FIG. 28 shows a three-dimensional image of the wrist-worn transceiver used with the body-worn monitor of FIGS. 27A and 27B;

[0057] FIG. 29A is a schematic view of a patient wearing an alternate embodiment of the invention featuring a sensor unit for measuring IP and ACC waveforms that connects directly to the patient's abdomen with an electrode; and FIG. 29B is a schematic, cross-sectional view of the sensor unit of FIG. 29A connected to the patient's abdomen with an electrode.

DETAILED DESCRIPTION OF THE INVENTION

Sensor Configuration

[0058] Referring to FIGS. 1A and 1B, a pair of accelerometers 12, 14 attach, respectively, to the chest and abdomen of a patient 10 to predict RR through the patient's torso movement and an algorithm based on adaptive filtering. Each accelerometer 12, 14 simultaneously measures acceleration (e.g. motion) along x, y, and z axes of a local coordinate system 18. As shown in FIG. 1B, the accelerometers 12, 14 are preferably aligned so the z axis points into the patient's torso. Within each accelerometer 12, 14 is an internal analog-to-digital converter that generates a digital ACC waveform 19 corresponding to each axis. Waveforms are sent as a stream of digital data to a wrist-worn transceiver (shown, for example, in FIGS. 27A, B, and 28) where they are processed using an adaptive filtering algorithm described in detail below to determine the patient's RR. Alternatively, the adaptive filtering algorithm can be performed with a microprocessor mounted proximal to the accelerometers 12, 14 on the patient's torso. Additional properties such as the patient's posture, degree of motion, and activity level are determined from these same digital ACC waveforms. As indicated by FIG. 1B, the axis within the accelerometer's coordinate system 18 that is aligned along the patient's torso (and thus orthogonal to their respiration-induced torso movement) is typically more sensitive to events not related to respiration, e.g. walking and falling.

[0059] In a preferred embodiment, digital accelerometers manufactured by Analog Devices (e.g. the ADXL345 component) are used in the configuration shown in FIG. 1A. These sensors detect acceleration over a range of +/-2 g (or, alternatively, up to +/-8 g) with a small-scale, low-power circuit.

[0060] Many patient's are classified as 'abdomen breathers', meaning during respiration their abdomen undergoes larger movements than their chest. A relative minority of patients are 'chest breathers', indicating that it is the chest that undergoes the larger movements. For this reason it is preferred that RR is determined using an ACC waveform detected along the z-axis with an accelerometer 14 positioned on the patient's abdomen. In alternate configurations the accelerometer 12 on the chest can be used in its place or two augment data collected with the abdomen-mounted sensor. Typically, ACC waveforms along multiple axes (e.g. the x and y-axes) are also modulated by breathing patterns, and can thus be used to estimate RR. In still other configurations multiple signals from one or both accelerometers 12, 14 are collectively processed to determine a single 'effective' ACC waveform representing, e.g., an average of the

two waveforms. This waveform is then processed using adaptive filtering to determine the patient's RR.

[0061] As shown in FIGS. 2A and 2B, ECG waveforms are simultaneously measured with the ACC waveforms using a trio of electrodes 20, 22, 24 typically positioned on the chest of the patient 10 in an Einthoven's triangle configuration. During a measurement, each electrode 20, 22, 24 measures a unique analog signal that passes through a shielded cable to an ECG circuit 26, which is typically mounted in a small plastic box 25 attached to the patient's chest. The ECG circuit 26 typically includes a differential amplifier and a series of analog filters with passbands that pass the high and low-frequency components that contribute to the ECG waveform 28, but filter out components associated with electrical and mechanical noise. Also within the box 25 is an accelerometer 12 and, alternatively as described above, a microprocessor for performing the adaptive filtering algorithm. A conventional analog ECG waveform 28, such as that shown in FIG. 20A, features a series of heartbeat-induced pulses, each characterized by a well-known 'QRS complex' that, informally, marks the initial depolarization of the patient's heart. To determine RR, a separate IP circuit 27 within the plastic box 25 generates a low-amperage current (typically 1-4 mA) that is modulated at a high frequency (typically 50-100 kHz). The current typically passes through electrode LL ('lower left') 24, which is located on the lower left-hand side of the patient's torso. It then propagates through the patient's chest, as indicated by the arrow 29, where a respiration-induced capacitance change modulates it according to the RR. Electrode UR ('upper right') 20 detects the resultant analog signal, which is then processed with a separate differential amplifier and series of analog filters within the IP circuit to determine an analog IP waveform 30 featuring a low-frequency series of pulses corresponding to RR. Typically the analog filters in the IP circuit 27 are chosen to filter out high-frequency components that contribute to the ECG QRS complex.

[0062] In other embodiments, the plastic box includes a temperature sensor 33, such as a conventional thermocouple, that measures the skin temperature of the patient's chest. This temperature is typically a few degrees lower than conventional core temperature, usually measured with a thermometer inserted in the patient's throat or rectum. Despite this discrepancy, skin temperature measured with the temperature sensor 33 can be monitored continuously and can therefore be used along with RR and other vital signs to predict patient decompensation.

[0063] In a preferred embodiment, both the ECG 28 and IP 30 waveforms are generated with a single application-specific integrated circuit (ASIC), or a circuit composed of a series of discrete elements which are known in the art. Preferably the ECG circuit includes an internal analog-to-digital converter that digitizes both waveforms before transmission to the wrist-worn transceiver for further processing. This circuitry, along with that associated with both the ECG and IP circuits, is contained within a single, small-scale electronic package.

[0064] Transmission of digital IP, ECG, and ACC waveforms, along with processed RR values, has several advantages over transmission of analog waveforms. First, a single transmission line in the monitor's cabling can transmit multiple digital waveforms, each generated by different sensors. This includes multiple ECG waveforms 28 (corre-

sponding, e.g., to vectors associated with three, five, and twelve-lead ECG systems) from the ECG circuit 26, the IP waveform 30 from the IP circuit 27, and ACC waveforms 19 associated with the x, y, and z axes of accelerometers 10, 12 attached to the patient's chest. Limiting the transmission line to a single cable reduces the number of wires attached to the patient, thereby decreasing the weight and cable-related clutter of the body-worn monitor. Second, cable motion induced by an ambulatory patient can change the electrical properties (e.g. electrical impedance) of its internal wires. This, in turn, can add noise to an analog signal and ultimately the vital sign calculated from it. A digital signal, in contrast, is relatively immune to such motion-induced artifacts. More sophisticated ECG circuits can plug into the wrist-worn transceiver to replace the three-lead system shown in FIG. 2A. These ECG circuits can include, e.g., five and twelve leads.

[0065] Digital data streams are typically transmitted to the wrist-worn transceiver using a serial protocol, such as a controlled area network (CAN) protocol, USB protocol, or RS-232 protocol. CAN is the preferred protocol for the body-worn monitor described in FIGS. 27A, 27B.

Determining RR from ACC Waveforms

[0066] Accelerometers positioned in the above-described locations on the patient's torso can detect respiration-induced motion associated with the chest and abdomen, and can therefore be processed to determine RR. Digital filtering is typically required to remove unwanted noise from the ACC waveform and isolate signal components corresponding to RR. Good filtering is required since respiratory-induced motions are typically small compared to those corresponding to activities (e.g. walking, falling) and posture changes (e.g. standing up, sitting down) associated with a patient's motion. Often these signals are only slightly larger than the accelerometer's noise floor.

[0067] FIGS. 3A-3D show a common, normalized ACC waveform without any filtering (FIG. 3A), and then filtered with a progressively narrow digital bandpass filter generated from a finite impulse response function featuring 1048 coefficients. FIGS. 3E-3H show the first derivative of these waveforms, and feature a zero-point crossing corresponding to a positive-to-negative slope change of a single pulse in the ACC waveform. This feature can be easily analyzed with a computer algorithm to count the various pulses that contribute to RR. As shown in FIG. 3A (the top figure), an unfiltered ACC waveform typically includes a series of respiration-induced pulses characterized by a peak amplitude which, in this case, is roughly twice that of the noise floor. This poor signal-to-noise ratio yields a derivatized signal in FIG. 3E that has no discernible zero-point crossing, thus making it nearly impossible to analyze. As shown in FIG. 3B, a relatively wide bandpass filter (0.01→1 Hz) yields an ACC waveform with a significantly improved signal-to-noise ratio. Still, as shown in FIG. 3F, the derivative of this waveform features a primary zero-point crossing occurring near 25 seconds, and a series of artificial noise-induced crossings, both before and after the primary crossing, that could be erroneously counted by an algorithm to yield an artificially high value for RR.

[0068] FIGS. 3C and 3G show, respectively, an ACC waveform and corresponding first derivative that result from a relatively narrow 0.01→0.5 Hz bandpass filter. These signals have higher signal-to-noise ratios than those shown in FIGS. 3B, 3F, but still include artificial zero-point cross-

ings on both sides of the primary zero-point crossing. While small, these features still have the potential to yield an artificially high value for RR. The signals shown in FIGS. 3D, 3H, in contrast, are ideal. Here, a narrow 0.01→0.1 Hz bandpass filter removes high-frequency components associated with artifacts in the ACC waveform, and in the process removes similar frequency components that contribute to sharp rising and falling edges of the individual breathing-induced pulses. This generates a smooth, sinusoid-shaped pulse train that once derivatized, as shown in FIG. 3H, yields a clean signal with only a single zero-point crossing. An algorithm can easily analyze this to determine RR. Importantly, as indicated by the alignment of the primary zero-point crossing in FIGS. 3F, 3G, and 3H, the finite impulse response function introduces little or no phase shift in the ACC waveforms.

[0069] As shown in FIGS. 4-9, under ideal conditions RR determined from a filtered ACC waveform agrees well with that determined from IP, which is a signal used during the adaptive filtering algorithm described herein, and et-CO₂, which represents a 'quasi' gold standard for determining RR. Data shown in each of these figures were collected simultaneously. ACC and IP waveforms were collected using an accelerometer mounted on a patient's abdomen, similar to that shown in FIG. 1A, and a trio of electrodes mounted in an Einthoven's triangle configuration, similar to that shown in FIG. 2A. The IP waveform is unfiltered, while the ACC waveform is filtered with a 0.01→0.1 Hz bandpass filter, as described with reference to FIGS. 3A, 3H. et-CO₂ was measured with a separate sensor positioned within the patient's mouth; signals from this sensor were not filtered in any way. In all cases breathing-induced pulses corresponding to RR were determined manually, and are marked accordingly in the figures. Numerical values within the markers indicate the exact number of counted pulses.

[0070] FIGS. 4-9 indicate that RR determined from both IP and ACC waveforms correlates well to absolute RR determined from et-CO₂. The correlation holds for a variety of breathing conditions, ranging from slow, deep breathing (FIGS. 4A-4C); fast, deep breathing (FIGS. 5A-5C); very fast, deep breathing (FIGS. 6A-6C); and shallow, slow breathing (FIGS. 7A-7C). Data were measured under these conditions from a patient in a prone (i.e. lying down) posture. Additionally, the agreement continues to hold for a standing patient undergoing deep, slow breathing (FIG. 8A-8C) and deep, fast breathing (FIG. 9A-9C). Even with this range of configurations, RR determined from both ACC and IP waveforms agreed to within 1 breath/minute to that determined from et-CO₂. In most cases the filtered ACC waveform appeared to have a superior signal-to-noise ratio when compared to the IP waveform, with the case for slow, deep breathing for a standing patient (FIGS. 8A-C) being the one exception.

[0071] As shown in FIGS. 10-11, agreement between RR calculated from ACC, IP, and et-CO₂ waveforms also holds before and after periods of apnea, as indicated by the shaded region 31 in FIGS. 10A-10C (lasting about 10 seconds), and region 32 in FIGS. 11A-11C (lasting about 30 seconds). As shown in FIGS. 10A-10C, for example, the patient exhibited slow, deep breaths before the period of apnea 31, and fast, deep breaths afterwards. FIGS. 11A-11C show an opposing configuration. Here, the patient exhibited fast, shallow breaths before the period of apnea, and slow, shallow breaths afterwards. In both cases agreement between RR calculated

from the three unique waveforms was maintained. These data, as described in more detail below, indicate that an adaptive filtering approach utilizing both ACC and IP waveforms can be used to predict a RR that correlates well to that measured with a gold standard, such as et-CO₂.

[0072] One confounding situation occurs when the patient is walking, as shown in FIGS. 12A-C. Here, in the ACC waveform, signals corresponding to the walking motion overwhelm those corresponding to breathing, making it impossible to selectively determine RR. However, the walking motion results in a well-defined, periodic signal characterized by a very high signal-to-noise ratio. The IP signal, in contrast, is completely corrupted by random noise, presumably caused by a combination of movements associated with the electrodes and their wires, electrical noise due to motion of the underlying muscles, and general corruption of the underlying capacitance in the patient's torso. This makes it impossible to determine RR or any other mechanical/physiological state corresponding to the patient. In this case RR determined from the et-CO₂ waveform is somewhat noisy, but still discernible.

[0073] While impossible to determine RR from the ACC and IP waveforms shown in FIG. 12A-B, the ACC waveform can be analyzed to determine walking, which it turn may be processed to avoid triggering a false alarm/alert that would normally be generated with a conventional vital sign monitor from the IP waveform, alone. For example, the ACC waveform shown in FIG. 12A, particularly when coupled with ACC waveforms corresponding to other axes of the chest-worn accelerometer as well as those from other accelerometers in the body-worn monitor, shows a clear signal indicative of walking. This determination can be corroborated with the IP waveform, which for a walking patient features an uncharacteristically low signal-to-noise ratio. Based on these signal inputs, an algorithm can determine that the patient is indeed walking, and can assume that their RR value is within normal limits, as a patient undergoing a consistent walking pattern is likely not in dire need of medical attention. For this reason an alarm/alert associated with RR is not generated. Similar alarms can be avoided when processing of the ACC waveforms determines that the patient is convulsing or falling (see, e.g., FIGS. 21-24), although in these cases a different type of alarm/alert may sound. In this way, collective processing of both the ACC and IP waveforms can help reduce false alarms/alerts associated with RR, while improving real alarms/alerts corresponding to other patient situations.

Adaptive Filtering

[0074] FIG. 13 illustrates in more detail how ACC and IP waveforms can be collectively processed to determine RR, activity levels, posture, and alarms/alerts associated with these patient states. The figure shows a flow chart describing an algorithm that would typically run using a microprocessor, such as that contained within a wrist-worn transceiver such as that shown in FIG. 28. Alternatively, the algorithm could run on a microprocessor mounted on the patient's torso with the IP and accelerometer sensors or elsewhere. The algorithm begins with steps 54, 56 that process all nine ACC waveforms, which are shown in the graph 69 on the left-hand side of the figure, to determine the patient's posture (step 54) and activity level (step 56). Both these processes are described in detail below. In general, determining posture (step 54) involves processing DC values of

the ACC waveform generated by the accelerometer mounted on the patient's chest; such signals are shown in the initial and end portions of the graph 69, which show changing DC values representing a posture change. Once sampled, the DC values are processed with an algorithm to estimate states corresponding to the patient such as standing, sitting, prone, supine, and lying on their side. This algorithm is also described with reference to FIG. 26A, 26B, below.

[0075] Once posture is determined, the algorithm then analyzes AC portions of the ACC waveforms to determine the patient's activity level (step 56). This part of the algorithm, which is also described in detail below, can be performed in several ways. For example, the AC portions of the ACC waveforms, such as the oscillating portion in the graph 69, can be processed with a Fourier Transform-based analysis to determine a frequency-dependent power-spectrum. Specific activity levels, such as walking and convulsing, involve periodic or quasi-periodic motions; these result in a well-defined power spectrum with frequency components between about 0 and 15 Hz (with this value representing the upper limit of human motion). Frequency bands in the power spectrum can be analyzed to estimate the patient's activity level. This analysis can also be combined with the posture determination from step 54 to refine the calculation for activity level. For example, a patient that is sitting down may be convulsing, but cannot be walking. Similarly, a falling event will begin with a standing posture, and end with a prone or supine posture.

[0076] Alternatively, the patient's activity level may be estimated with an algorithm based on probability and the concept of a 'logit variable', which considers a variety of time and frequency-domain parameters extracted from the AC portions of the ACC waveforms, and then processes these with a probability analysis that considers activity levels from a previously measured group of patients. An analysis based on a series of 'decision trees' can also be used to estimate the patient's activity level. Here, the decision trees feature steps that process both the AC and DC portions of the ACC waveforms to estimate the patient's activity level.

[0077] Algorithms that describe the patient's posture and activity level are described in detail in the following co-pending patent applications, the contents of which are incorporated herein by reference: VITAL SIGN MONITOR FEATURING 3 ACCELEROMETERS (U.S. Ser. No. 12/469,094; filed May 20, 2009) and METHOD FOR GENERATING ALARMS/ALERTS BASED ON A PATIENT'S POSTURE AND VITAL SIGNS (U.S. Ser. No. 12/469,236; filed May 20, 2009).

[0078] The patient's overall state is preferably grouped into one of two categories once posture and activity level are determined with steps 54 and 56. The first group involves relatively motion-free states, and includes categories such as patients that are: lying down with minimal motion (step 58), sitting up with minimal motion (step 59), and standing upright with minimal motion (step 60). Adaptive filtering that processes both ACC and IP waveforms will be effective in determining RR from this group of patients. The second group features patients that are undergoing some type of motion that will likely influence both the ACC and IP waveforms. Categories for this group include patients that are: lying down with significant motion, e.g. convulsing or talking in an animated manner (step 61), sitting up with significant motion (step 62), or standing upright with sig-

nificant motion, e.g. walking (step 63). Here, the adaptive filtering approach is abandoned, as a pair of respiratory-influenced waveforms with high signal-to-noise ratios is not available. Instead, the second group of patients is processed with a series of heuristic rules, described above, to determine whether or not to generate an alarm/alert based on their posture, activity level, and vital signs (including RR).

[0079] Patients within the first group (steps 58, 59, 60) yield ACC and IP waveforms that are collectively processed with an algorithm based on adaptive filtering to determine RR. Representative waveforms are described above and are shown, for example, by graphs 70, 71, as well as those shown in FIGS. 4-11. Details of the adaptive filtering algorithm are described below with reference to FIG. 14. This technique yields an accurate value for RR (step 66). An alarm/alert is generated if this value exceeds pre-set high and low limits for RR for a well-defined period of time (step 67).

[0080] For the second group of patients undergoing motion (steps 61, 62, 63) it is assumed that RR is normal but cannot be accurately determined (step 65). The underlying theory is that a patient that is walking or talking likely has a normal RR, and that such activity levels may result in artificially high or low values of RR that may trigger a false alarm. Still, an alarm/alert may be generated depending on the patient's posture or activity level, coupled with other vital signs and a set of heuristic rules (step 68). For example, activity levels such as convulsing or falling will automatically generate an alarm/alert. In another example, during step 68 the algorithm may ignore vital signs that are known to be strongly affected by motion (e.g. RR, blood pressure, and SpO₂), and process only those that are relatively immune to motion (e.g. heart rate and temperature). An alarm/alert may be triggered based on these parameters and the patient's motion and activity level. The set of heuristic rules used during step 68, along with a general approach for generating alarms/alerts with the body-worn monitor described herein, are described in more detail in the following co-pending patent application, the contents of which have been fully incorporated by reference above: METHOD FOR GENERATING ALARMS/ALERTS BASED ON A PATIENT'S POSTURE AND VITAL SIGNS (U.S. Ser. No. 12/469,236; filed May 20, 2009).

[0081] FIG. 14 describes in more detail an exemplary adaptive filtering algorithm used during step 64 to determine RR from the IP and ACC waveforms. The algorithm involves collecting ECG, PPG, ACC, and IP waveforms using the body-worn monitor described in FIGS. 27A, B (step 81). ECG and PPG waveforms are processed with external algorithms to determine heart rate, blood pressure, and pulse oximetry, as described in more detail below. Additionally, as described with reference to FIGS. 20A-E, these waveforms feature envelopes that are modulated by respiratory rate, and thus may be analyzed to provide an initial RR value for the adaptive filtering algorithm. Once collected, the ECG, PPG, and IP waveforms are analyzed with a series of simple metrics, such as analysis of signal-to-noise ratios and comparison of extracted RR values to pre-determined limits, to determine which one will provide the initial input to the adaptive filtering algorithm (step 82). Ideally RR is extracted from the IP waveform, as this provides a reliable initial value. If during step 82 it is determined that IP does not yield a reliable initial RR value, the envelopes of both the PPG and ECG waveforms are

extracted and analyzed as described above. If they are acceptable, RR values are then extracted from these waveforms and used for the initial value (step 89). The algorithm is terminated if each of the IP, PPG, and ECG waveforms fails to yield a reliable RR value.

[0082] If the IP waveform is deemed suitable, it is filtered with a finite impulse response filter with a bandpass of 0.01→12 Hz to remove electrical and mechanical noise that may lead to artifacts (step 83). Once filtered, the waveform is derivatized to yield a waveform similar to that shown in FIG. 3H (step 84), and then analyzed to find a zero-point crossing so that peaks corresponding to RR can be counted (step 85). During step 85 several simple signal processing algorithms may also be deployed to avoid counting features that don't actually correspond to RR, such as those shown in FIGS. 3F, 3G. For example, prior to looking for the zero-point crossing, the derivatized waveform may be squared to accentuate lobes on each side of the crossing. The resultant waveform may then be filtered again with a bandpass filter, or simply smoothed with a moving average. In other embodiments only lobes that exceed a pre-determined magnitude are considered when determining the zero-point crossing.

[0083] Once determined during step 85, the initial RR serves as the basis for the adaptive filter used in step 85. Typically this rate is multiplied by a factor (e.g. 1.5), and then used as an upper limit for a bandpass filter based on a finite impulse response function used to filter the ACC waveform (step 86). The lower limit for the bandpass filter is typically 0.01 Hz, as described above. Filtering the ACC waveform with these tailored parameters yields a resulting waveform that has a high signal-to-noise ratio, limited extraneous frequency components, and can easily be processed to determine RR. During step 87 signal processing technique similar to those described above with reference to step 84 may be used to further process the ACC waveform. These yield a smooth, derivatized waveform that is analyzed to determine a zero-point crossing and count the resulting peaks contributing to RR (step 88).

[0084] FIGS. 15, 16, and 17 illustrate how the above-described adaptive filtering algorithm can be applied to both ACC and IP waveforms. In each of the figures, the graphs show the ACC waveform filtered with an initial, non-adaptive filter (15A, 16A, 17A; 0.01→2 Hz bandpass), and the IP waveform filtered under similar conditions with a slightly larger bandpass filter (15B, 16B, 17B; 0.01→12 Hz bandpass). Typically the IP waveform is filtered with the larger bandpass so that high-frequency components composing the rising and falling edges of pulses within these waveforms are preserved.

[0085] Once filtered, the IP waveform is processed as described above to determine an initial RR. This value may include artifacts due to motion, electrical, and mechanical noise that erroneously increases or decreases the initial RR value. But typically such errors have little impact on the final RR value that results from the adaptive filter. The middle graph (FIGS. 15C, 16C, and 17C) in each figure show the ACC waveform processed with the adaptive filter. In all cases this waveform features an improved signal-to-noise ratio compared to data shown in the top graph (15A, 16A, 17A), which is processed with a non-adaptive (and relatively wide) filter. Typically the narrow bandpass on the adaptive filter removes many high-frequency components that contribute the sharp rising and falling edges of pulses in the

ACC waveforms. This slightly distorts the waveforms by rounding the pulses, giving the filtered waveform a shape that resembles a conventional sinusoid. Such distortion, however, has basically no effect on the absolute number of pulses in each waveform which are counted to determine RR.

[0086] The adaptively filtered waveform is then derivatized and graphed in FIGS. 15D, 16D, and 17D. This waveform is then processed with the above-mentioned signal processing techniques, e.g. squaring the derivative and filtering out lobes that fall beneath pre-determined threshold values, to yield an algorithm-determined 'count', indicated in FIGS. 15E, 16E, and 17E as a series of black triangles. The count is plotted along with the adaptively filtered waveforms from FIGS. 15C, 16C, and 17C. Exact overlap between each pulse in the waveform and the corresponding count indicates the algorithm is working properly. Data from each of the figures correspond to varying respiratory behavior (5, 17, and 38 breaths/minute in, respectively, FIGS. 15, 16, and 17), and indicate that this technique is effective over a wide range of breathing frequencies. The right-hand side of the figures (FIGS. 15F, 16F, and 17F) show a series of steps 90-94 that indicate the analysis required to generate the corresponding graphs in the figure.

[0087] FIG. 18 shows data collected when the patient is walking. Here, the walking motion manifests in the ACC waveform in FIG. 18A as a series of periodic pulses which look similar to RR, particularly after the initial bandpass filter of 0.01→2 Hz. However, the IP waveform shown in FIG. 18B has a poor signal-to-noise ratio, and fails to yield an accurate initial value for RR. This is indicated by step 95 in the modified flow chart shown in FIG. 18C, which highlights an alternate series of steps that are deployed when motion is present. As shown in step 96, in this case other ACC waveforms (e.g., those along the x and y-axes, indicated by ACC') are analyzed to determine that the patient is walking. In this case no value of RR is reported, and an alarm/alert is not triggered because of the above-mentioned heuristic rules (i.e. a walking patient typically has a normal RR, and is not in need of medical attention).

[0088] The efficacy of using adaptive filtering to determine RR from ACC and IP waveforms is summarized with the correlation graph in FIG. 19. The graph shows correlation with et-CO₂, which in this case represents a gold standard. Correlation is strong ($r^2=0.99$ for a RR range of 5-54 breaths/minute), and the graph includes data collected from patients in a range of postures (standing upright, lying down) and undergoing a range of breathing behaviors (deep breaths, shallow breaths). Bias calculated from these data was 0.8 breaths/minute, and the standard deviation of the differences was 1.6 breaths/minute. These statistics indicate adaptive filtering yields RR with an accuracy that is within the FDA's standards of +/-2 breaths/minute over a range of 0-70 breaths/minute.

Determining Respiratory Rate from ECG and PPG Waveforms

[0089] As described above, RR can additionally be determined from both the PPG and ECG waveforms by analyzing an envelope outlining heartbeat-induced pulses in these waveforms. Both PPG and ECG waveforms are collected with the body-worn monitor of FIGS. 27A, 27B, where they are further analyzed to continuously determine cNIBP according to the Composite Technique, as described above. FIGS. 20A-E show representative data that indicate this

technique. FIG. 20A, for example, shows an unfiltered ECG waveform featuring a train of pulses, each representing an individual QRS complex. The envelope of the QRS complexes is extracted by determining the maximum and minimum of each complex. Alternatively it can be determined with a series of digital filters that only pass very low frequencies. Comparison of the ECG envelope in FIG. 20B with the IP waveform in FIG. 20E indicates good agreement between these two approaches. Similarly, the PPG waveform shown in FIG. 20C features a train of pulses, each corresponding to a unique heartbeat, that typically follow the ECG QRS complex by a few hundred milliseconds. It is this time difference (typically called a ‘pulse transit time’, or PTT) that is sensitive to blood pressure changes, and is used during the Composite Technique to measure an absolute value for blood pressure. The PPG envelope, like the ECG envelope, is modulated by RR, and can be determined by extracting the maximum and minimum of each pulse. Alternatively this envelope can be determined with a low-pass filter similar to that used to extract the ECG envelope. As shown in FIG. 20D, the resulting envelope agrees well with the IP waveform, indicating it too is indicative of RR.

[0090] The body-worn monitor shown in FIGS. 27A, 27B measures two separate PPG waveforms (generated with red and infrared radiation) to determine the patient’s SpO₂ value. The algorithm for this calculation is described in detail in the following co-pending patent applications, the contents of which are incorporated herein by reference: BODY-WORN PULSE OXIMETER (U.S. Ser. No. 61/218, 062; filed Jun. 17, 2009). In embodiments, envelopes from both PPG waveforms can be extracted and processed to determine an initial value of RR. This value may also be calculated from the ECG waveform alone, or from this waveform and one or both PPG waveforms. As described above, this method for determining an initial RR value for the adaptive filter algorithm is less preferred than one that uses an IP waveform. Such an algorithm would be used, for example, if an IP waveform featuring a good signal-to-noise ratio was not available.

Affect of Motion on ECG, PPG, and ACC Waveforms

[0091] A patient’s activity level, as characterized by ACC waveforms, can have a significant impact on the PPG and ECG waveforms used to measure RR and cNIBP. For example, FIGS. 21-24 show time-dependent graphs of ECG, PPG, and ACC waveforms for a patient who is resting (FIG. 21), walking (FIG. 22), convulsing (FIG. 23), and falling (FIG. 24). Each graph includes a single ECG waveform, PPG waveform and three ACC waveforms. In all cases the PPG waveforms are generated with the infrared light source. The ACC waveforms correspond to signals measured along the x, y, and z axes by a single accelerometer worn on the patient’s wrist, similar to the accelerometer used within the wrist-worn transceiver shown in FIG. 28.

[0092] The figures indicate that time-dependent properties of both ECG and PPG waveforms can be strongly affected by certain patient activities, which are indicated by the ACC waveforms. Accuracy of RR and cNIBP calculated from these waveforms is therefore affected as well. FIGS. 21A-C, for example, shows data collected from a patient at rest. This state is clearly indicated by the ACC waveforms (FIG. 21C; bottom), which feature a relatively stable baseline along all three axes of the accelerometer. High-frequency noise in all the ACC waveforms shown in FIGS. 21-24 is due to

electrical noise, and is not indicative of patient motion in any way. The ECG (FIG. 21A; top) and PPG (FIG. 21B; middle) waveforms for this patient are correspondingly stable, thus allowing algorithms operating on the body-worn monitor to accurately determine SpO₂ (from the PPG waveform), along with heart rate and respiratory rate (from the ECG waveform), cNIBP (from a PTT extracted from both the ECG and PPG waveforms). Based on the data shown in FIG. 21, algorithms operating on the body-worn monitor assume that vital signs calculated from a resting patient are relatively stable; the algorithm therefore deploys normal threshold criteria for alarms/alerts, described below in Table 1, for patients in this state.

[0093] The ECG and PPG waveforms shown, respectively, in FIGS. 21A and 21B also feature envelopes indicated by the dashed lines 97a, 97b, 98 that are modulated by RR. This modulation is similar to that shown in FIGS. 20A and 20C.

[0094] FIGS. 22A-C shows ECG (FIG. 22A; top), PPG (FIG. 22B; middle), and ACC (FIG. 22C; top) waveforms measured from a walking patient wearing the body-worn monitor. In this case, the ACC waveform clearly indicates a quasi-periodic modulation, with each ‘bump’ in the modulation corresponding to a particular step. The ‘gaps’ in the modulation, shown near 10, 19, 27, and 35 seconds, correspond to periods when the patient stops walking and changes direction. Each bump in the ACC waveform includes relatively high-frequency features (other than those associated with electrical noise, described above) that correspond to walking-related movements of the patient’s wrist.

[0095] The ECG waveform measured from the walking patient is relatively unaffected by motion, other than indicating an increase in heart rate (i.e., a shorter time separation between neighboring QRS complexes) and respiratory rate (i.e. a higher frequency modulation of the waveform’s envelope) caused by the patient’s exertion. The PPG waveform, in contrast, is strongly affected by this motion, and pulses within it become basically immeasurable. Its distortion is likely due in part to a quasi-periodic change in light levels, caused by the patient’s swinging arm, and detected by the photodetector within the thumb-worn sensor. Movement of the patient’s arm additionally affects blood flow in the thumb and can cause the optical sensor to move relative to the patient’s skin. The photodetector measures all of these artifacts, along with a conventional PPG signal (like the one shown in FIG. 21B) caused by volumetric expansion in the underlying arteries and capillaries within the patient’s thumb. The artifacts produce radiation-induced photocurrent that is difficult to distinguish from normal PPG signal used to calculate SpO₂ and cNIBP. These vital signs are thus difficult or impossible to accurately measure when the patient is walking.

[0096] The body-worn monitor may deploy multiple strategies to avoid generating false alarms/alerts during a walking activity state that correspond to RR as well as all other vital signs. As described in detail below, the monitor can detect this state by processing the ACC waveforms shown in FIG. 22C along with similar waveforms measured from the patient’s bicep and chest. Walking typically elevates heart rate, respiratory rate, and blood pressure, and thus alarm thresholds for these parameters, as indicated by Table 1, are systematically and temporarily increased when this state is detected. Values above the modified thresholds are considered abnormal, and trigger an alarm. SpO₂, unlike heart rate,

respiratory rate and blood pressure, does not typically increase with exertion. Thus the alarm thresholds for this parameter, as shown in Table 1, do not change when the patient is walking. Body temperature measured with the body-worn monitor typically increases between 1-5%, depending on the physical condition of the patient and the speed at which they are walking.

TABLE 1

motion-dependent alarm/alert thresholds and heuristic rules for a walking patient			
Vital Sign	Motion State	Modified Threshold for Alarms/Alerts	Heuristic Rules for Alarms/Alerts
Blood Pressure (SYS, DIA)	Walking	Increase (+10-30%)	Ignore Threshold; Do Not Alarm/Alert
Heart Rate	Walking	Increase (+10-300%)	Use Modified Threshold; Alarm/Alert if Value Exceeds Threshold
Respiratory Rate	Walking	Increase (+10-300%)	Ignore Threshold; Do Not Alarm/Alert
SpO2	Walking	No Change	Ignore Threshold; Do Not Alarm/Alert
Temperature	Walking	Increase (+10-30%)	Use Original Threshold; Alarm/Alert if Value Exceeds Threshold

[0097] To further reduce false alarms/alerts, software associated with the body-worn monitor or remote monitor can deploy a series of heuristic rules determined beforehand using practical, empirical studies. These rules, for example, can indicate that a walking patient is likely healthy, breathing, and characterized by a normal RR. Accordingly, the rules dictate that cNIBP, RR, and SpO2 values measured during a walking state that exceed predetermined alarm/alert thresholds are likely corrupted by artifacts; the system, in turn, does not sound the alarm/alert in this case. Heart rate, as indicated by FIG. 22A, and body temperature can typically be accurately measured even when a patient is walking; the heuristic rules therefore dictate the modified thresholds listed in Table 1 be used to generate alarms/alerts for a patient in this state.

[0098] Additionally, despite the patient's walking motion, the ECG waveform shown in FIG. 22A still features an envelope shown by the dashed lines 99a, 99b that represents the patient's RR. This indicates that RR may be determined from a walking patient by processing the ECG envelope, even when other signals (e.g. IP and ACC waveforms) are corrupted. Because of the motion-induced noise in these signals, RR is typically determined directly from the ECG envelope, without using any adaptive filtering.

[0099] FIGS. 23A-C show ECG (FIG. 23A; top), PPG (FIG. 23B; middle), and ACC (FIG. 23C; bottom) waveforms measured from a patient that is simulating convulsing by rapidly moving their arm back and forth. A patient undergoing a Grand-mal seizure, for example, would exhibit this type of motion. As is clear from the waveforms, the patient is at rest for the initial 10 seconds shown in the graph, during which the ECG and PPG waveforms are uncorrupted by motion. The patient then begins a period of simulated, rapid convulsing that lasts for about 12 seconds. A brief 5-second period of rest follows, and then convulsing begins for another 12 seconds or so.

[0100] Convulsing modulates the ACC waveform due to rapid motion of the patient's arm, as measured by the

wrist-worn accelerometer. This modulation is strongly coupled into the PPG waveform, likely because of the phenomena described above, i.e.: 1) ambient light coupling into the oximetry probe's photodiode; 2) movement of the photodiode relative to the patient's skin; and 3) disrupted blood flow underneath the probe. Note that from about 23-28 seconds the ACC waveform is not modulated, indicating that the patient's arm is at rest. During this period the ambient light is constant and the optical sensor is stationary relative to the patient's skin. But the PPG waveform is still strongly modulated, albeit at a different frequency than the modulation that occurred when the patient's arm was moving, and the pulses therein are difficult to resolve. This indicates that the disrupted blood flow underneath the optical sensor continues even after the patient's arm stops moving. Using this information, both ECG and PPG waveforms similar to those shown in FIG. 23 can be analyzed in conjunction with ACC waveforms measured from groups of stationary and moving patients. These data can then be analyzed to estimate the effects of specific motions and activities on the ECG and PPG waveforms, and then deconvolute these factors using known mathematical techniques to effectively remove any motion-related artifacts. The deconvoluted ECG and PPG waveforms can then be used to calculate vital signs, as described in detail below.

[0101] The ECG waveform is modulated by the patient's arm movement, but to a lesser degree than the PPG waveform. In this case, modulation is caused primarily by electrical 'muscle noise' instigated by the convulsion and detected by the ECG electrodes, and well as by convulsion-induced motion in the ECG cables and electrodes relative to the patient's skin. Such motion is expected to have a similar affect on temperature measurements, which are determined by a sensor that also includes a cable.

[0102] Table 2, below, shows examples of the modified threshold values and heuristic rules for alarms/alerts generated by a convulsing patient. In general, when a patient experiences convulsions, such as those simulated during the two 12-second periods in FIG. 23, it is virtually impossible to accurately measure any vital signs from the ECG and PPG waveforms. For this reason the threshold values corresponding to each vital sign are not adjusted when convulsions are detected. Heart rate determined from the ECG waveform, for example, is typically erroneously high due to high-frequency convulsions, and RR is immeasurable from the distorted waveform. Strong distortion of the optical waveform also makes both SpO2 and PPT-based cNIBP difficult or impossible to measure. For this reason, algorithms operating on either the body-worn monitor or a remote monitor will not generate alarms/alerts based on vital signs when a patient is convulsing, as these vital signs will almost certainly be corrupted by motion-related artifacts.

TABLE 2

motion-dependent alarm/alert thresholds and heuristic rules for a convulsing patient			
Vital Sign	Motion State	Modified Threshold for Alarms/Alerts	Heuristic Rules for Alarms/Alerts
Blood Pressure (SYS, DIA)	Convulsing	No Change	Ignore Threshold; Generate Alarm/Alert Because of Convulsion

TABLE 2-continued

motion-dependent alarm/alert thresholds and heuristic rules for a convulsing patient			
Vital Sign	Motion State	Modified Threshold for Alarms/Alerts	Heuristic Rules for Alarms/Alerts
Heart Rate	Convulsing	No Change	Ignore Threshold; Generate Alarm/Alert Because of Convulsion
Respiratory Rate	Convulsing	No Change	Ignore Threshold; Generate Alarm/Alert Because of Convulsion
SpO2	Convulsing	No Change	Ignore Threshold; Generate Alarm/Alert Because of Convulsion
Temperature	Convulsing	No Change	Ignore Threshold; Generate Alarm/Alert Because of Convulsion

[0103] Table 2 also shows exemplary heuristic rules for convulsing patients. Here, the overriding rule is that a convulsing patient needs assistance, and thus an alarm/alert for this patient is generated regardless of their vital signs (which, as described above, are likely inaccurate due to motion-related artifacts). The system always generates an alarm/alert for a convulsing patient.

[0104] FIGS. 24A-C shows ECG (FIG. 24A; top), PPG (FIG. 24B; middle), and ACC (FIG. 24C; bottom) waveforms measured from a patient that experiences a fall roughly 13 seconds into the measuring period. The ACC waveform clearly indicates the fall with a sharp decrease in its signal, followed by a short-term oscillatory signal, due (literally) to the patient bouncing on the floor. After the fall, ACC waveforms associated with the x, y, and z axes also show a prolonged decrease in value due to the resulting change in the patient's posture. In this case, both the ECG and PPG waveforms are uncorrupted by motion prior to the fall, but basically immeasurable during the fall, which typically takes only 1-2 seconds. Specifically, this activity adds very high frequency noise to the ECG waveform, making it impossible to extract heart rate and RR during this short time period. Falling causes a sharp drop in the PPG waveform, presumably for the same reasons as described above (i.e. changes in ambient light, sensor movement, and disruption of blood flow) for walking and convulsing, making it difficult to measure SpO2 and cNIBP.

[0105] After a fall, both the ECG and PPG waveforms are free from artifacts, but both indicate an accelerated heart rate and relatively high heart rate variability for roughly 10 seconds. During this period the PPG waveform also shows distortion and a decrease in pulse amplitude. Without being bound to any theory, the increase in heart rate may be due to the patient's baroreflex, which is the body's haemostatic mechanism for regulating and maintaining blood pressure. The baroreflex, for example, is initiated when a patient begins faint. In this case, the patient's fall may cause a rapid drop in blood pressure, thereby depressing the baroreflex. The body responds by accelerating heart rate (indicated by the ECG waveform) and increasing blood pressure (indicated by a reduction in PTT, as measured from the ECG and PPG waveforms) in order to deliver more blood to the patient's extremities.

[0106] Table 3 shows exemplary heuristic rules and modified alarm thresholds for a falling patient. Falling, similar to convulsing, makes it difficult to measure waveforms and the vital signs calculated from them. Because of this and the

short time duration associated with a fall, alarms/alerts based on vital signs thresholds are not generated during an actual falls. However, this activity, optionally coupled with prolonged stationary period or convulsion (both determined from the following ACC waveform), generates an alarm/alert according to the heuristic rules.

TABLE 3

motion-dependent alarm/alert thresholds and heuristic rules for a falling patient Processing ACC Waveforms to Determine Posture			
Vital Sign	Motion State	Modified Threshold for Alarms/Alerts	Heuristic Rules for Alarms/Alerts
Blood Pressure (SYS, DIA)	Falling	No Change	Ignore Threshold; Generate Alarm/Alert Because of Fall
Heart Rate	Falling	No Change	Ignore Threshold; Generate Alarm/Alert Because of Fall
Respiratory Rate	Falling	No Change	Ignore Threshold; Generate Alarm/Alert Because of Fall
SpO2	Falling	No Change	Ignore Threshold; Generate Alarm/Alert Because of Fall
Temperature	Falling	No Change	Ignore Threshold; Generate Alarm/Alert Because of Fall

[0107] In addition to activity level, as described above and indicated in FIGS. 21-24, a patient's posture can influence how the above-described system generates alarms/alerts from RR, cNIBP, and other vital signs. For example, the alarms/alerts related to both RR and cNIBP may vary depending on whether the patient is lying down or standing up. FIG. 25 indicates how the body-worn monitor can determine motion-related parameters (e.g. degree of motion, posture, and activity level) from a patient 110 using time-dependent ACC waveforms continuously generated from the three accelerometers 112, 113, 114 worn, respectively, on the patient's chest, bicep, and wrist. The height of the patient's arm can affect the cNIBP measurement, as blood pressure can vary significantly due to hydrostatic forces induced by changes in arm height. Moreover, this phenomenon can be detected and exploited to calibrate the cNIBP measurement, as described in detail in the above-referenced patent application, the contents of which have been previously incorporated by reference: BODY-WORN VITAL SIGN MONITOR WITH SYSTEM FOR DETECTING AND ANALYZING MOTION (U.S. Ser. No. 12/469,094; filed May 20, 2009). As described in this document, arm height can be determined using DC signals from the accelerometers 113, 114 disposed, respectively, on the patient's bicep and wrist. Posture, in contrast, can be exclusively determined by the accelerometer 112 worn on the patient's chest. An algorithm operating on the wrist-worn transceiver extracts DC values from waveforms measured from this accelerometer and processes them with an algorithm described below to determine posture.

[0108] Specifically, torso posture is determined for a patient 110 using angles determined between the measured gravitational vector and the axes of a torso coordinate space 111. The axes of this space 111 are defined in a three-dimensional Euclidean space where \vec{R}_{CV} is the vertical axis, \vec{R}_{CH} is the horizontal axis, and \vec{R}_{CN} is the normal axis. These axes must be identified relative to a 'chest accelerometer coordinate space' before the patient's posture can be determined.

[0109] The first step in determining a patient's posture is to identify alignment of \vec{R}_{CV} in the chest accelerometer coordinate space. This can be determined in either of two approaches. In the first approach, \vec{R}_{CV} is assumed based on a typical alignment of the body-worn monitor relative to the patient. During a manufacturing process, these parameters are then preprogrammed into firmware operating on the wrist-worn transceiver. In this procedure it is assumed that accelerometers within the body-worn monitor are applied to each patient with essentially the same configuration. In the second approach, \vec{R}_{CV} is identified on a patient-specific basis. Here, an algorithm operating on the wrist-worn transceiver prompts the patient (using, e.g., video instruction operating on the wrist-worn transceiver, or audio instructions transmitted through a speaker) to assume a known position with respect to gravity (e.g., standing upright with arms pointed straight down). The algorithm then calculates \vec{R}_{CV} from DC values corresponding to the x, y, and z axes of the chest accelerometer while the patient is in this position. This case, however, still requires knowledge of which arm (left or right) the monitor is worn on, as the chest accelerometer coordinate space can be rotated by 180 degrees depending on this orientation. A medical professional applying the monitor can enter this information using the GUI, described above. This potential for dual-arm attachment requires a set of two pre-determined vertical and normal vectors which are interchangeable depending on the monitor's location. Instead of manually entering this information, the arm on which the monitor is worn can be easily determined following attachment using measured values from the chest accelerometer values, with the assumption that \vec{R}_{CV} is not orthogonal to the gravity vector.

[0110] The second step in the procedure is to identify the alignment of \vec{R}_{CN} in the chest accelerometer coordinate space. The monitor determines this vector in the same way it determines \vec{R}_{CV} using one of two approaches. In the first approach the monitor assumes a typical alignment of the chest-worn accelerometer on the patient. In the second approach, the alignment is identified by prompting the patient to assume a known position with respect to gravity. The monitor then calculates \vec{R}_{CN} from the DC values of the time-dependent ACC waveform.

[0111] The third step in the procedure is to identify the alignment of \vec{R}_{CH} in the chest accelerometer coordinate space. This vector is typically determined from the vector cross product of \vec{R}_{CV} and \vec{R}_{CN} , or it can be assumed based on the typical alignment of the accelerometer on the patient, as described above.

[0112] A patient's posture is determined using the coordinate system described above and in FIG. 25, along with a gravitational vector \vec{R}_G that extends normal from the patient's chest. The angle between \vec{R}_{CV} and \vec{R}_G is given by equation (1):

$$\theta_{VG}[n] = \arccos\left(\frac{\vec{R}_G[n] \cdot \vec{R}_{CV}}{\|\vec{R}_G[n]\| \|\vec{R}_{CV}\|}\right) \quad (1)$$

where the dot product of the two vectors is defined as:

$$\vec{R}_{CV} = (y_{CVx}[n] \times r_{CVx}) + (y_{CVy}[n] \times r_{CVy}) + (y_{CVz}[n] \times r_{CVz}) \quad (2)$$

The definition of the norms of \vec{R}_G and \vec{R}_{CV} are given by equations (3) and (4):

$$\|\vec{R}_G[n]\| = \sqrt{(y_{Gx}[n])^2 + (y_{Gy}[n])^2 + (y_{Gz}[n])^2} \quad (3)$$

$$\|\vec{R}_{CV}\| = (r_{CVx})^2 + (r_{CVy})^2 + (r_{CVz})^2 \quad (4)$$

[0113] As indicated in equation (5), the monitor compares the vertical angle θ_{VG} to a threshold angle to determine whether the patient is vertical (i.e. standing upright) or lying down:

$$\text{if } \theta_{VG} \leq 45^\circ \text{ then Torso State}=0, \text{ the patient is upright} \quad (5)$$

If the condition in equation (5) is met the patient is assumed to be upright, and their torso state, which is a numerical value equated to the patient's posture, is equal to 0. The patient is assumed to be lying down if the condition in equation (5) is not met, i.e. $\theta_{VG} > 45$ degrees. Their lying position is then determined from angles separating the two remaining vectors, as defined below.

[0114] The angle θ_{VG} between \vec{R}_{CN} and \vec{R}_G determines if the patient is lying in the supine position (chest up), prone position (chest down), or on their side. Based on either an assumed orientation or a patient-specific calibration procedure, as described above, the alignment of \vec{R}_{CN} is given by equation (6), where i, j, k represent the unit vectors of the x, y, and z axes of the chest accelerometer coordinate space respectively:

$$\vec{R}_{CN} = r_{CNx} \hat{i} + r_{CNy} \hat{j} + r_{CNz} \hat{k} \quad (6)$$

The angle between \vec{R}_{CN} and \vec{R}_G determined from DC values extracted from the chest accelerometer ACC waveform is given by equation (7):

$$\theta_{NG}[n] = \arccos\left(\frac{\vec{R}_G[n] \cdot \vec{R}_{CN}}{\|\vec{R}_G[n]\| \|\vec{R}_{CN}\|}\right) \quad (7)$$

The body-worn monitor determines the normal angle θ_{NG} and then compares it to a set of predetermined threshold angles to determine which position the patient is lying in, as shown in equation (8):

$$\text{if } \theta_{NG} \leq 35^\circ \text{ then Torso State}=1, \text{ the patient is supine}$$

$$\text{if } \theta_{NG} \geq 135^\circ \text{ then Torso State}=2, \text{ the patient is prone} \quad (8)$$

If the conditions in equation (8) are not met then the patient is assumed to be lying on their side. Whether they are lying on their right or left side is determined from the angle calculated between the horizontal torso vector and measured gravitational vectors, as described above.

[0115] The alignment of \vec{R}_{CH} is determined using either an assumed orientation, or from the vector cross-product of \vec{R}_{CV} and \vec{R}_{CN} as given by equation (9), where i, j, k represent the unit vectors of the x, y, and z axes of the accelerometer coordinate space respectively. Note that the orientation of the calculated vector is dependent on the order of the vectors in the operation. The order below defines the horizontal axis as positive towards the right side of the patient's body.

$$\vec{R}_{CH} = r_{CVx} \vec{i} + r_{CVy} \vec{j} + r_{CVz} \vec{k} = \vec{R}_{CV} \times \vec{R}_{CN} \quad (9)$$

The angle θ_{HG} between \vec{R}_{CH} and \vec{R}_G is determined using equation (10):

$$\theta_{HG}[n] = \arccos\left(\frac{\vec{R}_G[n] \cdot \vec{R}_{CH}}{\|\vec{R}_G[n]\| \|\vec{R}_{CH}\|}\right) \quad (10)$$

The monitor compares this angle to a set of predetermined threshold angles to determine if the patient is lying on their right or left side, as given by equation (11):

$$\begin{aligned} &\text{if } \theta_{HG} \geq 90^\circ \text{ then Torso State}=3, \text{ the patient is on} \\ &\quad \text{their right side} \\ &\text{if } \theta_{HG} < 90^\circ \text{ then Torso State}=4, \text{ the patient is on} \\ &\quad \text{their left side} \end{aligned} \quad (11)$$

Table 4 describes each of the above-described postures, along with a corresponding numerical torso state used to render, e.g., a particular icon on a remote computer:

TABLE 4

postures and their corresponding torso states	
Posture	Torso State
standing upright	0
supine: lying on back	1
prone: lying on chest	2
lying on right side	3
lying on left side	4
undetermined posture	5

[0116] FIGS. 26A and 26B show, respectively, graphs of time-dependent ACC waveforms measured along the x, y, and z-axes (FIG. 26A), and the torso states (i.e. postures; FIG. 26B) determined from these waveforms for a moving patient, as described above. As the patient moves, the DC values of the ACC waveforms measured by the chest accelerometer vary accordingly, as shown in FIG. 26A. The body-worn monitor processes these values as described above to continually determine \vec{R}_G and the various quantized torso states for the patient, as shown in FIG. 26B. The torso states yield the patient's posture as defined in Table 4. For this study the patient rapidly alternated between standing, lying on their back, chest, right side, and left side within a time period of about 160 seconds. As described above, different alarm/alert conditions (e.g. threshold values) for vital signs can be assigned to each of these postures, or the specific posture itself may result in an alarm/alert. Additionally, the time-dependent properties of the graph can be analyzed (e.g. changes in the torso states can be counted) to determine, for example, how often the patient moves in their hospital bed. This number can then be equated to various metrics, such as a 'bed sore index' indicating a patient that is so stationary in their bed that lesions may result. Such a state could then be used to trigger an alarm/alert to the supervising medical professional.

Hardware for Measuring Respiratory Rate

[0117] FIGS. 27A and 27B show how the body-worn monitor 200 described above attaches to a patient 170 to measure RR, cNIBP, and other vital signs. These figures show two configurations of the system: FIG. 27A shows the

system used during the indexing portion of the Composite Technique, and includes a pneumatic, cuff-based system 185, while FIG. 27B shows the system used for subsequent RR and cNIBP measurements. The indexing measurement typically takes about 60 seconds, and is typically performed once every 4 hours. Once the indexing measurement is complete the cuff-based system 185 is typically removed from the patient. The remainder of the time the monitor 200 performs the RR, SpO2 and cNIBP measurements.

[0118] The body-worn monitor 200 features a wrist-worn transceiver 172, described in more detail in FIG. 28, featuring a touch panel interface 173 that displays RR, blood pressure values and other vital signs. A wrist strap 190 affixes the transceiver 172 to the patient's wrist like a conventional wristwatch. A flexible cable 192 connects the transceiver 172 to a pulse oximeter probe 194 that wraps around the base of the patient's thumb. During a measurement, the probe 194 generates a time-dependent PPG waveform which is processed along with an ECG to measure cNIBP, SpO2, and possible RR. This provides an accurate representation of blood pressure in the central regions of the patient's body, as described above.

[0119] To determine ACC waveforms the body-worn monitor 200 features three separate accelerometers located at different portions on the patient's arm and chest. The first accelerometer is surface-mounted on a circuit board in the wrist-worn transceiver 172 and measures signals associated with movement of the patient's wrist. As described above, this motion can also be indicative of that originating from the patient's fingers, which will affect the SpO2 measurement. The second accelerometer is included in a small bulkhead portion 196 included along the span of the cable 182. During a measurement, a small piece of disposable tape, similar in size to a conventional bandaid, affixes the bulkhead portion 196 to the patient's arm. In this way the bulkhead portion 196 serves two purposes: 1) it measures a time-dependent ACC waveform from the mid-portion of the patient's arm, thereby allowing their posture and arm height to be determined as described in detail above; and 2) it secures the cable 182 to the patient's arm to increase comfort and performance of the body-worn monitor 200, particularly when the patient is ambulatory. The third accelerometer is mounted in a bulkhead component 174 that connects through cables 180a-c to ECG electrodes 178a-c. As described in detail above, this accelerometer, which can also be mounted closer to the patient's abdomen, measures respiration-induced motion of the patient's chest and abdomen. These signals are then digitized, transmitted through the cable 182 to the wrist-worn transceiver 172, where they are processed with an algorithm as described above to determine RR.

[0120] The cuff-based module 185 features a pneumatic system 176 that includes a pump, valve, pressure fittings, pressure sensor, analog-to-digital converter, microcontroller, and rechargeable Li:ion battery. During an indexing measurement, the pneumatic system 176 inflates a disposable cuff 184 and performs two measurements according to the Composite Technique: 1) it performs an inflation-based measurement of oscillometry to determine values for SYS, DIA, and MAP; and 2) it determines a patient-specific relationship between PTT and MAP. These measurements are described in detail in the above-referenced patent application entitled: 'VITAL SIGN MONITOR FOR MEASURING BLOOD PRESSURE USING OPTICAL, ELECTRI-

CAL, AND PRESSURE WAVEFORMS' (U.S. Ser. No. 12/138,194; filed Jun. 12, 2008), the contents of which have been previously incorporated herein by reference.

[0121] The cuff **184** within the cuff-based pneumatic system **185** is typically disposable and features an internal, airtight bladder that wraps around the patient's bicep to deliver a uniform pressure field. During the indexing measurement, pressure values are digitized by the internal analog-to-digital converter, and sent through a cable **186** according to a CAN protocol, along with SYS, DIA, and MAP blood pressures, to the wrist-worn transceiver **172** for processing as described above. Once the cuff-based measurement is complete, the cuff-based module **185** is removed from the patient's arm and the cable **186** is disconnected from the wrist-worn transceiver **172**. cNIBP is then determined using PTT, as described in detail above.

[0122] To determine an ECG, the body-worn monitor **200** features a small-scale, three-lead ECG circuit integrated directly into the bulkhead **174** that terminates an ECG cable **182**. The ECG circuit features an integrated circuit that collects electrical signals from three chest-worn ECG electrodes **178a-c** connected through cables **180a-c**. As described above, the ECG electrodes **178a-c** are typically disposed in a conventional Einthoven's Triangle configuration which is a triangle-like orientation of the electrodes **178a-c** on the patient's chest that features three unique ECG vectors. From these electrical signals the ECG circuit determines up to three ECG waveforms, which are digitized using an analog-to-digital converter mounted proximal to the ECG circuit, and sent through the cable **182** to the wrist-worn transceiver **172** according to the CAN protocol. There, the ECG and PPG waveforms are processed to determine the patient's blood pressure. Heart rate and RR are determined directly from the ECG waveform using known algorithms, such as those described above. The cable bulkhead **174** also includes an accelerometer that measures motion associated with the patient's chest as described above.

[0123] As described above, there are several advantages of digitizing ECG and ACC waveforms prior to transmitting them through the cable **182**. First, a single transmission line in the cable **182** can transmit multiple digital waveforms, each generated by different sensors. This includes multiple ECG waveforms (corresponding, e.g., to vectors associated with three, five, and twelve-lead ECG systems) from the ECG circuit mounted in the bulkhead **174**, along with waveforms associated with the x, y, and z-axes of accelerometers mounted in the bulkheads **174**, **196**. More sophisticated ECG circuits (e.g. five and twelve-lead systems) can plug into the wrist-worn transceiver to replace the three-lead system shown in FIGS. **27A** and **27B**.

[0124] FIG. **28** shows a close-up view of the wrist-worn transceiver **172**. As described above, it attaches to the patient's wrist using a flexible strap **190** which threads through two D-ring openings in a plastic housing **206**. The transceiver **172** features a touchpanel display **220** that renders a GUI **173** which is altered depending on the viewer (typically the patient or a medical professional). Specifically, the transceiver **172** includes a small-scale infrared barcode scanner **202** that, during use, can scan a barcode worn on a badge of a medical professional. The barcode indicates to the transceiver's software that, for example, a nurse or doctor is viewing the user interface. In response, the GUI **173** displays vital sign data and other medical diagnostic information appropriate for medical professionals. Using

this GUI **173**, the nurse or doctor, for example, can view the vital sign information, set alarm parameters, and enter information about the patient (e.g. their demographic information, medication, or medical condition). The nurse can press a button on the GUI **173** indicating that these operations are complete. At this point, the display **220** renders an interface that is more appropriate to the patient, such as time of day and battery power.

[0125] The transceiver **172** features three CAN connectors **204a-c** on the side of its upper portion, each which supports the CAN protocol and wiring schematics, and relays digitized data to the internal CPU. Digital signals that pass through the CAN connectors include a header that indicates the specific signal (e.g. ECG, ACC, or pressure waveform from the cuff-based module) and the sensor from which the signal originated. This allows the CPU to easily interpret signals that arrive through the CAN connectors **204a-c**, such as those described above corresponding to RR, and means that these connectors are not associated with a specific cable. Any cable connecting to the transceiver can be plugged into any connector **204a-c**. As shown in FIG. **27A**, the first connector **204a** receives the cable **182** that transports a digitized ECG waveform determined from the ECG circuit and electrodes, and digitized ACC waveforms measured by accelerometers in the cable bulkhead **174** and the bulkhead portion **196** associated with the ECG cable **182**.

[0126] The second CAN connector **204b** shown in FIG. **28** receives the cable **186** that connects to the pneumatic cuff-based system **185** used for the pressure-dependent indexing measurement (shown in FIG. **27A**). This connector **204b** receives a time-dependent pressure waveform delivered by the pneumatic system **185** to the patient's arm, along with values for SYS, DIA, and MAP values determined during the indexing measurement. The cable **186** unplugs from the connector **204b** once the indexing measurement is complete, and is plugged back in after approximately four hours for another indexing measurement.

[0127] The final CAN connector **204c** can be used for an ancillary device, e.g. a glucometer, infusion pump, body-worn insulin pump, ventilator, or et-CO₂ delivery system. As described above, digital information generated by these systems will include a header that indicates their origin so that the CPU can process them accordingly.

[0128] The transceiver includes a speaker **201** that allows a medical professional to communicate with the patient using a voice over Internet protocol (VOIP). For example, using the speaker **201** the medical professional could query the patient from a central nursing station or mobile phone connected to a wireless, Internet-based network within the hospital. Or the medical professional could wear a separate transceiver similar to the shown in FIG. **28**, and use this as a communication device. In this application, the transceiver **172** worn by the patient functions much like a conventional cellular telephone or 'walkie talkie': it can be used for voice communications with the medical professional and can additionally relay information describing the patient's vital signs and motion. The speaker can also enunciate pre-programmed messages to the patient, such as those used to calibrate the chest-worn accelerometers for a posture calculation, as described above.

Other Embodiments of the Invention

[0129] RR can also be calculated using a combination of ACC, ECG, PPG, IP, and other signals using algorithms that

differ from those described above. For example, these signals can be processed with an averaging algorithm, such as one using a weighted average, to determine a single waveform that can then be processed to determine RR. Or the ACC waveform can be used alone, without being integrated in an adaptive filtering algorithm, to determine RR without relying on IP. In this case the ACC waveform is filtered with a simple bandpass filter, e.g. a finite impulse response filter, with a set passband (e.g. 0.01→5 Hz). Similarly, multiple ACC waveforms, such as those measured along axes (e.g. the x or y-axes) orthogonal to the vector normal to the patient's chest (i.e. the z-axis), can be processed with or without adaptive filtering to determine RR. In this case the waveforms may be averaged together with a weighted average to generate a single waveform, which is then filtered, derivatized, and signal processed as described above with reference to FIG. 3 to determine RR. Similarly, envelopes associated with the ECG and PPG waveforms can be processed in a similar manner to determine RR. In still other embodiments, other sensors, such as ultra wide-band radar or acoustic sensors, can detect signals indicative of RR and used with ACC or IP waveforms and the adaptive filtering approach described above to determine RR. Here, the alternative sensors are typically used to replace measurement of the IP waveform, although they can also be used to replace measurement of the ACC waveform. An acoustic sensor suitable for this application is described, for example, in the following co-pending patent application, the contents of which are incorporated herein by reference: DEVICE FOR DETERMINING RESPIRATORY RATE AND OTHER VITAL SIGNS (U.S. Ser. No. 12/171,886; filed Jul. 12, 2008).

[0130] In addition to those methods described above, the body-worn monitor can use a number of additional methods to calculate blood pressure and other properties from the optical and electrical waveforms. These are described in the following co-pending patent applications, the contents of which are incorporated herein by reference: 1) CUFFLESS BLOOD-PRESSURE MONITOR AND ACCOMPANYING WIRELESS, INTERNET-BASED SYSTEM (U.S. Ser. No. 10/709,015; filed Apr. 7, 2004); 2) CUFFLESS SYSTEM FOR MEASURING BLOOD PRESSURE (U.S. Ser. No. 10/709,014; filed Apr. 7, 2004); 3) CUFFLESS BLOOD PRESSURE MONITOR AND ACCOMPANYING WEB SERVICES INTERFACE (U.S. Ser. No. 10/810,237; filed Mar. 26, 2004); 4) VITAL SIGN MONITOR FOR ATHLETIC APPLICATIONS (U.S.S.N.; filed Sep. 13, 2004); 5) CUFFLESS BLOOD PRESSURE MONITOR AND ACCOMPANYING WIRELESS MOBILE DEVICE (U.S. Ser. No. 10/967,511; filed Oct. 18, 2004); 6) BLOOD PRESSURE MONITORING DEVICE FEATURING A CALIBRATION-BASED ANALYSIS (U.S. Ser. No. 10/967,610; filed Oct. 18, 2004); 7) PERSONAL COMPUTER-BASED VITAL SIGN MONITOR (U.S. Ser. No. 10/906,342; filed Feb. 15, 2005); 8) PATCH SENSOR FOR MEASURING BLOOD PRESSURE WITHOUT A CUFF (U.S. Ser. No. 10/906,315; filed Feb. 14, 2005); 9) PATCH SENSOR FOR MEASURING VITAL SIGNS (U.S. Ser. No. 11/160,957; filed Jul. 18, 2005); 10) WIRELESS, INTERNET-BASED SYSTEM FOR MEASURING VITAL SIGNS FROM A PLURALITY OF PATIENTS IN A HOSPITAL OR MEDICAL CLINIC (U.S. Ser. No. 11/162,719; filed Sep. 9, 2005); 11) HAND-HELD MONITOR FOR MEASURING VITAL SIGNS (U.S. Ser. No. 11/162,742;

filed Sep. 21, 2005); 12) CHEST STRAP FOR MEASURING VITAL SIGNS (U.S. Ser. No. 11/306,243; filed Dec. 20, 2005); 13) SYSTEM FOR MEASURING VITAL SIGNS USING AN OPTICAL MODULE FEATURING A GREEN LIGHT SOURCE (U.S. Ser. No. 11/307,375; filed Feb. 3, 2006); 14) BILATERAL DEVICE, SYSTEM AND METHOD FOR MONITORING VITAL SIGNS (U.S. Ser. No. 11/420,281; filed May 25, 2006); 15) SYSTEM FOR MEASURING VITAL SIGNS USING BILATERAL PULSE TRANSIT TIME (U.S. Ser. No. 11/420,652; filed May 26, 2006); 16) BLOOD PRESSURE MONITOR (U.S. Ser. No. 11/530,076; filed Sep. 8, 2006); 17) TWO-PART PATCH SENSOR FOR MONITORING VITAL SIGNS (U.S. Ser. No. 11/558,538; filed Nov. 10, 2006); and, 18) MONITOR FOR MEASURING VITAL SIGNS AND RENDERING VIDEO IMAGES (U.S. Ser. No. 11/682,177; filed Mar. 5, 2007).

[0131] Other embodiments are also within the scope of the invention. For example, other measurement techniques, such as conventional oscillometry measured during deflation, can be used to determine SYS for the above-described algorithms. Additionally, processing units and probes for measuring pulse oximetry similar to those described above can be modified and worn on other portions of the patient's body. For example, pulse oximetry probes with finger-ring configurations can be worn on fingers other than the thumb. Or they can be modified to attach to other conventional sites for measuring SpO₂, such as the ear, forehead, and bridge of the nose. In these embodiments the processing unit can be worn in places other than the wrist, such as around the neck (and supported, e.g., by a lanyard) or on the patient's waist (supported, e.g., by a clip that attaches to the patient's belt). In still other embodiments the probe and processing unit are integrated into a single unit.

[0132] In other embodiments, a set of body-worn monitors can continuously monitor a group of patients, wherein each patient in the group wears a body-worn monitor similar to those described herein. Additionally, each body-worn monitor can be augmented with a location sensor. The location sensor includes a wireless component and a location-processing component that receives a signal from the wireless component and processes it to determine a physical location of the patient. A processing component (similar to that described above) determines from the time-dependent waveforms at least one vital sign, one motion parameter, and an alarm parameter calculated from the combination of this information. A wireless transceiver transmits the vital sign, motion parameter, location of the patient, and alarm parameter through a wireless system. A remote computer system featuring a display and an interface to the wireless system receives the information and displays it on a user interface for each patient in the group.

[0133] In embodiments, the interface rendered on the display at the central nursing station features a field that displays a map corresponding to an area with multiple sections. Each section corresponds to the location of the patient and includes, e.g., the patient's vital signs, motion parameter, and alarm parameter. For example, the field can display a map corresponding to an area of a hospital (e.g. a hospital bay or emergency room), with each section corresponding to a specific bed, chair, or general location in the area. Typically the display renders graphical icons corresponding to the motion and alarm parameters for each patient in the group. In other embodiments, the body-worn

monitor includes a graphical display that renders these parameters directly on the patient.

[0134] Typically the location sensor and the wireless transceiver operate on a common wireless system, e.g. a wireless system based on 802.11, 802.15.4, or cellular protocols. In this case a location is determined by processing the wireless signal with one or more algorithms known in the art. These include, for example, triangulating signals received from at least three different base stations, or simply estimating a location based on signal strength and proximity to a particular base station. In still other embodiments the location sensor includes a conventional global positioning system (GPS).

[0135] The body-worn monitor can include a first voice interface, and the remote computer can include a second voice interface that integrates with the first voice interface. The location sensor, wireless transceiver, and first and second voice interfaces can all operate on a common wireless system, such as one of the above-described systems based on 802.11 or cellular protocols. The remote computer, for example, can be a monitor that is essentially identical to the monitor worn by the patient, and can be carried or worn by a medical professional. In this case the monitor associated with the medical professional features a GUI wherein the user can select to display information (e.g. vital signs, location, and alarms) corresponding to a particular patient. This monitor can also include a voice interface so the medical professional can communicate directly with the patient.

[0136] FIGS. 29A, 29B show yet another alternate embodiment of the invention wherein a sensor unit 255 attaches to the abdomen of a patient 10 using an electrode 24 normally attached to the lower left-hand portion of the patient's torso. Specifically, the sensor unit 255 includes a connector 253 featuring an opening that receives the metal snap or rivet present on most disposable ECG electrodes. Connecting the connector 245 to the electrode's rivet holds the sensor unit 255 in place. This configuration reduces the number of cables in the body-worn monitor, and additionally secures an accelerometer 12 to the patient's abdomen. This is typically the part of their torso that undergoes the greatest

motion during respiration, and thus generates ACC waveforms with the highest possible signal-to-noise ratio. Also contained within the sensor unit 255 are the ECG circuit 26, the IP circuit 27, and a temperature sensor 33.

[0137] To measure IP and ECG waveforms, the sensor unit 255 connects through cables 250a, 250b to electrodes 20, 22 attached, respectively, to the upper right-hand and left-hand portions of the patient's torso. This system measures RR using the adaptive filtering approach described above, and has the additional advantage of measuring a relatively large ACC signals indicating respiration-induced motions of the patient's abdomen. As described above, these signals are typically generated by the z-axis of the accelerometer 12, which is normal to the patient's torso. ACC signals along the x and y-axes can be additionally processed to determine the patient's posture and activity level, as described above. Once RR and these motion-related properties are measured, a transceiver in the sensor unit (not shown in the figure) transmits them in the form of a digital data stream through a cable 251 to the wrist-worn transceiver for further processing.

[0138] Still other embodiments are within the scope of the following claims.

What is claimed is:

1. A method for determining a respiration rate value from a patient, comprising the following steps:

- (a) measuring a first time-dependent signal by detecting a modulated electrical current passing through the patient's torso;
- (b) measuring a second time-dependent signal by detecting respiration-induced movements in the patient's torso with at least one motion sensor;
- (c) determining a motion-related event not related to the patient's respiration rate value by processing signals from at least one motion sensor; and
- (d) collectively processing both the first and second time-dependent signals to determine a value for respiration rate corresponding to a period when at the motion-related event is below a pre-determined threshold.

* * * * *

专利名称(译)	用于测量呼吸率的体戴式监护仪		
公开(公告)号	US20190076054A1	公开(公告)日	2019-03-14
申请号	US16/188125	申请日	2018-11-12
[标]申请(专利权)人(译)	索泰拉无线公司		
申请(专利权)人(译)	SOTERA WIRELESS, INC.		
当前申请(专利权)人(译)	SOTERA WIRELESS, INC.		
[标]发明人	BANET MATTHEW MCCOMBIE DEVIN DHILLON MARSHAL		
发明人	BANET, MATTHEW MCCOMBIE, DEVIN DHILLON, MARSHAL		
IPC分类号	A61B5/08 A61B5/0205 A61B5/00 A61B5/113 A61B5/11 A61B5/0402 A61B5/024 A61B5/053		
CPC分类号	A61B5/746 A61B5/0002 A61B5/6824 A61B5/6823 A61B5/7239 A61B2562/0219 A61B5/721 A61B5/681 A61B5/113 A61B5/1118 A61B5/1116 A61B5/0816 A61B5/0809 A61B5/0535 A61B5/0402 A61B5/02416 A61B5/0205 A61B5/6826		
外部链接	Espacenet USPTO		

摘要(译)

本发明提供了一种多传感器系统，其使用基于自适应滤波的算法来监测患者的呼吸速率。该系统具有第一传感器，该第一传感器选自由阻抗呼吸描记传感器，ECG传感器，PPG传感器和运动传感器（例如，加速度计）组成的组，该运动传感器被配置为附接到患者的躯干并从中测量运动信号。该系统还包括（iii）处理系统，被配置为可操作地连接到第一和运动传感器，并通过将从第一传感器信号获得的滤波器参数应用于运动传感器信号来确定呼吸率值。

