



US008880161B2

(12) **United States Patent**  
**Warren et al.**

(10) **Patent No.:** **US 8,880,161 B2**  
(45) **Date of Patent:** **Nov. 4, 2014**

- (54) **METHODS AND DEVICES FOR ACCURATELY CLASSIFYING CARDIAC ACTIVITY**
- (71) Applicant: **Cameron Health, Inc.**, San Clemente, CA (US)
- (72) Inventors: **Jay A. Warren**, San Juan Capistrano, CA (US); **Rick Sanghera**, San Clemente, CA (US); **Venugopal Allavatam**, San Clemente, CA (US); **Surekha Palreddy**, Maplewood, MN (US)
- (73) Assignee: **Cameron Health, Inc.**, Arden Hills, MN (US)

- (52) **U.S. Cl.**  
CPC ..... *A61N 1/36592* (2013.01); *A61B 5/04525* (2013.01); *A61B 5/7264* (2013.01); *A61B 5/726* (2013.01); *A61N 1/37* (2013.01)  
USPC ..... **600/516**; 607/9
- (58) **Field of Classification Search**  
USPC ..... 600/516  
See application file for complete search history.

(\* ) Notice: Subject to any disclaimer, the term of this patent is extended or adjusted under 35 U.S.C. 154(b) by 0 days.

- (56) **References Cited**  
U.S. PATENT DOCUMENTS  
4,184,493 A 1/1980 Langer et al.  
4,300,567 A 11/1981 Kolenik et al.  
(Continued)

(21) Appl. No.: **14/058,495**

- FOREIGN PATENT DOCUMENTS  
AU 2009221694 B2 1/2014  
AU 2009221696 B2 4/2014  
(Continued)

(22) Filed: **Oct. 21, 2013**

- OTHER PUBLICATIONS  
U.S. Appl. No. 12/399,901, Final Office Action mailed Apr. 11, 2013, 11 pgs.  
(Continued)

(65) **Prior Publication Data**  
US 2014/0046396 A1 Feb. 13, 2014

- Primary Examiner* — Eric D. Bertram  
(74) *Attorney, Agent, or Firm* — Seager, Tufte & Wickhem, LLC


**Related U.S. Application Data**

- (60) Division of application No. 13/436,448, filed on Mar. 30, 2012, now Pat. No. 8,600,489, which is a continuation of application No. 12/437,547, filed on May 7, 2009, now Pat. No. 8,160,687, which is a continuation-in-part of application No. 12/399,914, filed on Mar. 6, 2009, now Pat. No. 8,160,686.
- (60) Provisional application No. 61/051,332, filed on May 7, 2008.

- (51) **Int. Cl.**  
*A61B 5/0452* (2006.01)  
*A61B 5/00* (2006.01)  
*A61N 1/365* (2006.01)  
*A61N 1/37* (2006.01)

- (57) **ABSTRACT**  
Methods, systems, and devices for signal analysis in an implanted cardiac monitoring and treatment device such as an implantable cardioverter defibrillator. In some examples, captured data including detected events is analyzed to identify likely overdetection of cardiac events. In some illustrative examples, when overdetection is identified, data may be modified to correct for overdetection, to reduce the impact of overdetection, or to ignore overdetection. Several examples emphasize the use of morphology analysis using correlation to static templates and/or inter-event correlation analysis.

**15 Claims, 15 Drawing Sheets**



	N	N-1	N-2	N-3
104	Low	Low	Low	Low
	High	Middle	High	High
	Low	Low	Low	Low
106	Middle	Low	Middle	Low



(56)

## References Cited

## U.S. PATENT DOCUMENTS

4,407,288	A	10/1983	Langer et al.	6,334,071	B1	12/2001	Lu
4,450,527	A	5/1984	Sramek	6,377,844	B1	4/2002	Graen
4,457,315	A	7/1984	Bennish	6,381,493	B1	4/2002	Stadler et al.
4,589,420	A	5/1986	Adams et al.	6,393,316	B1	5/2002	Gillberg et al.
4,595,009	A	6/1986	Leinders	6,397,100	B2	5/2002	Stadler et al.
4,679,144	A	7/1987	Cox et al.	6,411,844	B1	6/2002	Kroll et al.
4,693,253	A	9/1987	Adams	6,438,410	B2	8/2002	Hsu et al.
4,750,494	A	6/1988	King	6,449,503	B1	9/2002	Hsu
4,779,617	A	10/1988	Whigham	6,493,579	B1	12/2002	Gilkerson et al.
4,940,054	A	7/1990	Grevis et al.	6,493,584	B1	12/2002	Lu
4,979,110	A	12/1990	Albrecht et al.	6,505,068	B2	1/2003	Bonnet et al.
4,989,602	A	2/1991	Sholder et al.	6,516,225	B1	2/2003	Florio
5,000,189	A	3/1991	Throne et al.	6,561,984	B1	5/2003	Turcott
5,105,810	A	4/1992	Collins et al.	6,567,691	B1	5/2003	Stadler
5,117,824	A	6/1992	Keimel et al.	6,574,505	B1	6/2003	Warren
5,184,616	A	2/1993	Weiss	6,575,912	B1	6/2003	Turcott
5,193,535	A	3/1993	Bardy et al.	6,587,720	B2	7/2003	Hsu et al.
5,215,098	A	6/1993	Steinhaus et al.	6,625,490	B1	9/2003	McClure et al.
5,217,021	A	6/1993	Steinhaus et al.	6,643,549	B1	11/2003	Bradley et al.
5,240,009	A	8/1993	Williams	6,647,292	B1	11/2003	Bardy et al.
5,271,411	A	12/1993	Ripley et al.	6,684,100	B1	1/2004	Sweeney et al.
5,273,049	A	12/1993	Steinhaus et al.	6,699,200	B2	3/2004	Cao et al.
5,280,792	A	1/1994	Leong et al.	6,708,058	B2	3/2004	Kim et al.
5,299,119	A	3/1994	Kraf et al.	6,708,062	B2	3/2004	Ericksen et al.
5,313,953	A	5/1994	Yomtov et al.	6,721,597	B1	4/2004	Bardy et al.
5,331,966	A	7/1994	Bennett et al.	6,728,572	B2	4/2004	Hsu et al.
5,339,820	A	8/1994	Henry et al.	6,731,978	B2	5/2004	Olson et al.
5,342,402	A	8/1994	Olson et al.	6,745,068	B2	6/2004	Koyrakh et al.
5,342,407	A	8/1994	Dahl et al.	6,754,528	B2	6/2004	Bardy et al.
5,351,696	A	10/1994	Riff et al.	6,889,079	B2	5/2005	Bocek et al.
5,376,104	A	12/1994	Sakai et al.	6,892,092	B2	5/2005	Palreddy et al.
5,400,795	A	3/1995	Murphy et al.	6,909,916	B2	6/2005	Spinelli et al.
5,423,326	A	6/1995	Wang et al.	6,950,702	B2	9/2005	Sweeney
5,447,519	A	9/1995	Peterson	7,016,730	B2	3/2006	Ternes
5,464,431	A	11/1995	Adams et al.	7,020,523	B1	3/2006	Lu et al.
5,470,342	A	11/1995	Mann et al.	7,027,856	B2	4/2006	Zhou et al.
5,486,199	A	1/1996	Kim et al.	7,027,858	B2	4/2006	Cao et al.
5,513,644	A	5/1996	McClure et al.	7,027,862	B2	4/2006	Dahl et al.
5,522,852	A	6/1996	White et al.	7,031,764	B2	4/2006	Schwartz et al.
5,534,019	A	7/1996	Paspa	7,062,314	B2	6/2006	Zhu et al.
5,545,186	A	8/1996	Olson et al.	7,062,315	B2	6/2006	Koyrakh et al.
5,558,098	A	9/1996	Fain	7,062,322	B2	6/2006	Stadler et al.
5,564,430	A	10/1996	Jacobson et al.	7,076,289	B2	7/2006	Sarkar et al.
5,607,455	A	3/1997	Armstrong	7,085,599	B2	8/2006	Kim et al.
5,620,466	A	4/1997	Haefner et al.	7,103,412	B1	9/2006	Kroll
5,658,317	A	8/1997	Haefner et al.	7,117,035	B2	10/2006	Wagner et al.
5,676,690	A	10/1997	Noren	7,123,954	B2	10/2006	Narayan et al.
5,685,315	A	11/1997	McClure et al.	7,149,575	B2	12/2006	Ostroff et al.
5,702,425	A	12/1997	Wickham	7,162,301	B2	1/2007	Kim et al.
5,709,215	A	1/1998	Perttu et al.	7,167,747	B2	1/2007	Gunderson et al.
5,755,738	A	5/1998	Kim et al.	7,184,815	B2	2/2007	Kim et al.
5,776,168	A	7/1998	Gunderson	7,184,818	B2	2/2007	Kim et al.
5,779,645	A	7/1998	Olson et al.	7,191,004	B2	3/2007	Kim et al.
5,817,134	A	10/1998	Greenhut	7,194,302	B2	3/2007	Bardy et al.
5,827,197	A	10/1998	Bocek et al.	7,218,966	B2	5/2007	Haefner et al.
5,857,977	A	1/1999	Caswell et al.	7,236,819	B2	6/2007	Brockway et al.
5,876,350	A	3/1999	Lo et al.	7,248,921	B2	7/2007	Palreddy et al.
5,891,048	A	4/1999	Nigam et al.	7,266,409	B2	9/2007	Gunderson
5,991,657	A	11/1999	Kim	7,283,863	B2	10/2007	Gunderson et al.
6,029,086	A	2/2000	Kim et al.	7,302,294	B2	11/2007	Kamath et al.
6,041,251	A	3/2000	Kim et al.	7,330,757	B2	2/2008	Ostroff et al.
6,047,210	A	4/2000	Kim et al.	7,346,392	B2	3/2008	Kenknight
6,052,617	A	4/2000	Kim	7,376,458	B2	5/2008	Palreddy et al.
6,058,328	A	5/2000	Levine et al.	7,379,772	B2	5/2008	Bardy et al.
6,095,987	A	8/2000	Shmulewitz et al.	7,386,342	B1	6/2008	Falkenberg et al.
6,115,628	A	9/2000	Stadler et al.	7,392,085	B2	6/2008	Warren et al.
6,144,879	A	11/2000	Gray	7,444,182	B2	10/2008	Ostroff et al.
6,148,230	A	11/2000	Kenknight	7,447,540	B1	11/2008	Nabutovsky et al.
6,223,078	B1	4/2001	Marcovecchio	7,467,009	B2	12/2008	Palreddy et al.
6,230,055	B1	5/2001	Sun et al.	7,477,935	B2	1/2009	Palreddy et al.
6,236,882	B1	5/2001	Lee et al.	7,496,408	B2	2/2009	Ghanem et al.
6,240,313	B1	5/2001	Esler	7,496,409	B2	2/2009	Greenhut et al.
6,266,554	B1	7/2001	Hsu et al.	7,499,750	B2	3/2009	Haefner et al.
6,280,462	B1	8/2001	Hauser et al.	7,522,959	B2	4/2009	Hauser et al.
6,308,095	B1	10/2001	Hsu et al.	7,546,159	B1	6/2009	Nabutovsky et al.
				7,555,335	B2	6/2009	Kamath et al.
				7,559,900	B2	7/2009	Gillberg
				7,567,835	B2	7/2009	Gunderson et al.
				7,570,997	B2	8/2009	Lovett et al.

(56)

## References Cited

## U.S. PATENT DOCUMENTS

7,593,771 B2 9/2009 Yonce et al.  
 7,623,913 B2 11/2009 Phillips  
 7,623,916 B2 11/2009 Julian  
 7,655,014 B2 2/2010 Ko et al.  
 7,682,316 B2 3/2010 Anderson et al.  
 7,684,864 B2 3/2010 Olson et al.  
 7,715,906 B2 5/2010 Krause et al.  
 7,734,345 B2 6/2010 Cinbis  
 7,761,142 B2 7/2010 Ghanem et al.  
 7,774,049 B2 8/2010 Ghanem et al.  
 7,783,354 B2 8/2010 Gunderson  
 7,792,571 B2 9/2010 Sweeney et al.  
 7,797,036 B2 9/2010 Zhang et al.  
 7,801,594 B1 9/2010 Higham  
 7,865,233 B2 1/2011 Haefner  
 7,894,894 B2 2/2011 Stadler et al.  
 7,904,142 B2 3/2011 Kim et al.  
 7,904,153 B2 3/2011 Greenhut et al.  
 7,907,993 B2 3/2011 Ghanem et al.  
 8,160,686 B2 4/2012 Allavatam et al.  
 8,160,687 B2 4/2012 Warren et al.  
 8,265,749 B2 9/2012 Allavatam et al.  
 8,565,878 B2 10/2013 Allavatam et al.  
 8,588,896 B2 11/2013 Allavatam et al.  
 8,600,489 B2 12/2013 Warren et al.  
 8,626,280 B2 1/2014 Allavatam et al.  
 2003/0204215 A1 10/2003 Gunderson et al.  
 2004/0049120 A1 3/2004 Cao et al.  
 2004/0215239 A1 10/2004 Favet et al.  
 2004/0215240 A1 10/2004 Lovett et al.  
 2004/0220628 A1 11/2004 Wagner  
 2004/0230229 A1 11/2004 Lovett et al.  
 2004/0230233 A1 11/2004 Gunderson et al.  
 2005/0107838 A1 5/2005 Lovett et al.  
 2005/0131464 A1 6/2005 Heinrich et al.  
 2005/0154421 A1 7/2005 Ousdigian et al.  
 2006/0116595 A1 6/2006 Palreddy et al.  
 2006/0116725 A1 6/2006 Palreddy et al.  
 2006/0122676 A1 6/2006 Ko  
 2006/0167502 A1 7/2006 Haefner  
 2006/0167503 A1 7/2006 Warren et al.  
 2006/0167504 A1 7/2006 Warren et al.  
 2006/0235476 A1 10/2006 Gunderson et al.  
 2006/0241701 A1 10/2006 Markowitz et al.  
 2007/0032829 A1 2/2007 Ostroff  
 2007/0049975 A1 3/2007 Cates et al.  
 2007/0135847 A1 6/2007 Kenknight  
 2007/0142736 A1 6/2007 Cazares et al.  
 2007/0156190 A1 7/2007 Cinbis  
 2007/0179539 A1 8/2007 Degroot et al.  
 2007/0179540 A1 8/2007 Stegemann et al.  
 2007/0232944 A1 10/2007 Ghanem et al.  
 2007/0232945 A1 10/2007 Kleckner et al.  
 2007/0232948 A1 10/2007 Stadler et al.  
 2007/0233198 A1 10/2007 Ghanem et al.  
 2007/0244521 A1 10/2007 Bornzin et al.  
 2007/0255151 A1\* 11/2007 Struble et al. .... 600/513  
 2007/0276445 A1 11/2007 Sanghera et al.  
 2008/0015647 A1 1/2008 Palreddy et al.  
 2008/0077030 A1 3/2008 Ostroff  
 2008/0086174 A1 4/2008 Libbus et al.  
 2008/0091242 A1 4/2008 Kamath et al.  
 2008/0132965 A1 6/2008 Ostroff et al.  
 2008/0161870 A1 7/2008 Gunderson  
 2008/0172098 A1 7/2008 Gunderson et al.  
 2008/0183085 A1 7/2008 Van Oort et al.  
 2008/0188901 A1 8/2008 Sanghera et al.  
 2008/0215110 A1 9/2008 Gunderson  
 2008/0221632 A1 9/2008 Bardy et al.  
 2008/0228093 A1 9/2008 Dong et al.  
 2008/0243200 A1 10/2008 Scinicariello et al.  
 2008/0262559 A1 10/2008 Zhang et al.  
 2008/0275516 A1 11/2008 Ghanem et al.  
 2008/0275517 A1 11/2008 Ghanem et al.  
 2008/0275519 A1 11/2008 Ghanem et al.

2008/0275521 A1 11/2008 Warren et al.  
 2008/0288009 A1 11/2008 Kim et al.  
 2009/0036788 A1 2/2009 Nabutovsky et al.  
 2009/0043352 A1 2/2009 Brooke et al.  
 2009/0054796 A1 2/2009 Sanghera et al.  
 2009/0054938 A1 2/2009 Ostroff et al.  
 2009/0093731 A1 4/2009 Palreddy et al.  
 2009/0156957 A1 6/2009 Linder et al.  
 2009/0157128 A1 6/2009 Seim et al.  
 2009/0157132 A1 6/2009 Linder et al.  
 2009/0157137 A1 6/2009 Gilkerson et al.  
 2009/0187227 A1 7/2009 Palreddy et al.  
 2009/0228057 A1 9/2009 Allavatam et al.  
 2009/0240157 A1 9/2009 Lian et al.  
 2009/0240300 A1 9/2009 Lian et al.  
 2009/0259271 A1 10/2009 Allavatam et al.  
 2010/0004713 A1 1/2010 Warren et al.  
 2010/0094369 A1 4/2010 Allavatam et al.  
 2010/0152798 A1 6/2010 Sanghera et al.  
 2010/0152799 A1 6/2010 Sanghera et al.  
 2010/0185109 A1 7/2010 Zhang et al.  
 2010/0331904 A1 12/2010 Warren et al.  
 2011/0098585 A1 4/2011 Warren et al.  
 2011/0098775 A1 4/2011 Allavatam et al.  
 2013/0006085 A1 1/2013 Allavatam et al.  
 2014/0046204 A1 2/2014 Allavatam et al.  
 2014/0046394 A1 2/2014 Allavatam et al.  
 2014/0094868 A1 4/2014 Allavatam et al.

## FOREIGN PATENT DOCUMENTS

CN 1819855 A 8/2006  
 CN 1829554 A 9/2006  
 CN 1915166 A 2/2007  
 CN 1985750 A 6/2007  
 CN 101065059 A 10/2007  
 CN 103285513 A 9/2013  
 CN 102083496 B 10/2013  
 CN 102065948 B 12/2013  
 CN 103691061 A 4/2014  
 EP 0554208 A2 8/1993  
 EP 1774907 A1 4/2007  
 EP 2313153 B1 4/2012  
 EP 2455132 A1 5/2012  
 JP 1110344 A 1/1999  
 JP 2000023932 A 1/2000  
 JP 2006523505 A 10/2006  
 JP 2006526472 A 11/2006  
 JP 2007501099 A 1/2007  
 JP 2007510447 A 4/2007  
 JP 2008536633 A 9/2008  
 JP 2013248530 A 12/2013  
 WO WO-2004105871 A1 12/2004  
 WO WO-2006113698 A1 10/2006  
 WO WO-2009111764 A2 9/2009  
 WO WO-2009111764 A3 9/2009  
 WO WO-2009111766 A2 9/2009  
 WO WO-2009111766 A3 9/2009  
 WO WO-2009137726 A2 11/2009  
 WO WO-2009137726 A3 11/2009

## OTHER PUBLICATIONS

U.S. Appl. No. 12/399,901, Non Final Office Action mailed Sep. 17, 2012, 44 pgs.  
 U.S. Appl. No. 12/399,901, Response filed Apr. 6, 2012 to Restriction Requirement mailed Mar. 6, 2012, 25 pgs.  
 U.S. Appl. No. 12/399,901, Response filed Dec. 17, 2012 to Non Final Office Action mailed Sep. 17, 2012, 16 pgs.  
 U.S. Appl. No. 12/399,901, Restriction Requirement mailed Mar. 6, 2012, 8 pgs.  
 U.S. Appl. No. 12/399,914, Non-Final Office Action mailed Oct. 21, 2011, 34 pgs.  
 U.S. Appl. No. 12/399,914, Notice of Allowance mailed Feb. 14, 2012, 5 pgs.  
 U.S. Appl. No. 12/399,914, Response filed Oct. 7, 2011 to Restriction Requirement mailed Sep. 9, 2011, 11 pgs.

(56)

**References Cited**

## OTHER PUBLICATIONS

- U.S. Appl. No. 12/399,914, Response received Jan. 13, 2012 to Non-Final Office Action mailed Oct. 21, 2011, 8 pgs.
- U.S. Appl. No. 12/399,914, Restriction Requirement mailed Sep. 9, 2011, 6 pgs.
- U.S. Appl. No. 12/437,547, Non-Final Office Action mailed Nov. 10, 2011, 34 pgs.
- U.S. Appl. No. 12/437,547, Notice of Allowance mailed Feb. 15, 2012, 5 pgs.
- U.S. Appl. No. 12/437,547, Response filed Jan. 27, 2012 to Non-Final Office Action mailed Nov. 10, 2011, 6 pgs.
- U.S. Appl. No. 12/437,547, Response filed Oct. 10, 2011 to Restriction Requirement mailed Sep. 9, 2011, 12 pgs.
- U.S. Appl. No. 12/437,547, Restriction Requirement mailed Sep. 9, 2011, 5 pgs.
- U.S. Appl. No. 12/637,438, Non-Final Office Action mailed Dec. 7, 2011, 34 pgs.
- U.S. Appl. No. 12/637,438, Notice of Allowance mailed May 16, 2012, 5 pgs.
- U.S. Appl. No. 12/637,438, Response filed Feb. 29, 2012 to Non-Final Office Action mailed Dec. 7, 2011, 11 pgs.
- U.S. Appl. No. 12/637,438, Response filed Nov. 18, 2011 to Restriction Requirement mailed Oct. 21, 2011, 8 pgs.
- U.S. Appl. No. 12/637,438, Restriction Requirement mailed Oct. 21, 2011, 7 pgs.
- U.S. Appl. No. 13/436,398, Non Final Office Action mailed May 9, 2013, 6 pgs.
- U.S. Appl. No. 13/436,398, Response filed Apr. 11, 2013 to Restriction Requirement mailed Mar. 15, 2013, 9 pgs.
- U.S. Appl. No. 13/436,398, Restriction Requirement mailed Mar. 15, 2013, 6 pgs.
- U.S. Appl. No. 13/436,448, Non Final Office Action mailed Feb. 27, 2013, 6 pgs.
- U.S. Appl. No. 13/436,448, Notice of Allowance mailed Jul. 24, 2013, 6 pgs.
- U.S. Appl. No. 13/436,448, Response filed Feb. 13, 2013 to Restriction Requirement mailed Jan. 18, 2013, 9 pgs.
- U.S. Appl. No. 13/436,448, Response filed May 21, 2013 to Non Final Office Action mailed Feb. 27, 2013, 11 pgs.
- U.S. Appl. No. 13/436,448, Restriction Requirement mailed Jan. 18, 2013, 6 pgs.
- U.S. Appl. No. 13/607,168, Non Final Office Action mailed Mar. 26, 2013, 7 pgs.
- U.S. Appl. No. 13/607,168, Response filed Mar. 14, 2013 to Restriction Requirement mailed Feb. 27, 2013, 6 pgs.
- U.S. Appl. No. 13/607,168, Restriction Requirement mailed Feb. 27, 2013, 5 pgs.
- Chinese Application Serial No. 200980116877.4, Office Action mailed Jan. 30, 2013, (W/ English Translation), 17 pgs.
- Chinese Application Serial No. 200980116886.3, Office Action mailed Dec. 13, 2012, (W/ English Translation), 16 pgs.
- Chinese Application Serial No. 200980125869.6, Office Action mailed Jan. 10, 2013, (W/ English Translation), 12 pgs.
- European Application Serial No. 09717667.1, Office Action mailed Dec. 16, 2011, 4 pgs.
- European Application Serial No. 09717667.1, Response filed Apr. 26, 2012 to Office Action mailed Dec. 16, 2011, 44 pgs.
- European Application Serial No. 09717991.5, Office Action mailed Dec. 16, 2011, 4 pgs.
- European Application Serial No. 09717991.5, Response filed Apr. 26, 2012 to Office Action mailed Dec. 16, 2011, 41 pgs.
- European Application Serial No. 09743712.3, Office Action mailed Jun. 8, 2011, 3 pgs.
- European Application Serial No. 09743712.3, Response filed Aug. 10, 2011 to Office Action mailed Jun. 8, 2011, 9 pgs.
- European Application Serial No. 12151591.0, European Search Report mailed Mar. 21, 2012, 6 pgs.
- European Application Serial No. 12151591.0, Office Action mailed Dec. 19, 2012, 5 pgs.
- European Application Serial No. 12151591.0, Response filed Oct. 26, 2012 to European Search Report mailed Mar. 21, 2012, 10 pgs.
- European Application Serial No. 12151593.6, European Search Report mailed Mar. 21, 2012, 7 pgs.
- European Application Serial No. 12151593.6, Office Action mailed May 7, 2012, 2 pgs.
- European Application Serial No. 12151593.6, Office Action mailed Dec. 19, 2012, 5 pgs.
- European Application Serial No. 12151593.6, Response filed Apr. 19, 2013 to Office Action mailed Dec. 19, 2012, 8 pgs.
- European Application Serial No. 12151593.6, Response filed Oct. 26, 2012 to European Search Report mailed Mar. 21, 2012, 13 pgs.
- European Application Serial No. 12151595.1, European Search Report mailed Mar. 21, 2012, 6 pgs.
- European Application Serial No. 12151595.1, Response filed Oct. 26, 2012 to European Search Report mailed Mar. 21, 2012, 9 pgs.
- European Application Serial No. 12189307.7, Extended European Search Report mailed Mar. 1, 2013, 6 pgs.
- European Application Serial No. 12189311.9, Extended European Search Report mailed Mar. 1, 2013, 6 pgs.
- International Application Serial No. PCT/US2009/036432, International Report on Patentability mailed Sep. 22, 2009, 12 pgs.
- International Application Serial No. PCT/US2009/036432, International Search Report mailed Sep. 22, 2009, 5 pgs.
- International Application Serial No. PCT/US2009/036434, International Preliminary Report on Patentability mailed Sep. 7, 2010, 14 pgs.
- International Application Serial No. PCT/US2009/036434, International Search Report mailed Sep. 22, 2009, 7 pgs.
- International Application Serial No. PCT/US2009/036434, International Written Opinion mailed Sep. 7, 2010, 13 pgs.
- International Application Serial No. PCT/US2009/036434, International Written Opinion mailed Sep. 7, 2010, 11 pgs.
- International Application Serial No. PCT/US2009/043213, International Report on Patentability mailed Nov. 9, 2010, 13 pgs.
- International Application Serial No. PCT/US2009/043213, International Search Report mailed Dec. 1, 2009, 7 pgs.
- International Application Serial No. PCT/US2009/043213, International Written Opinion mailed Nov. 7, 2010, 12 pgs.
- "Methods and Devices for Identifying and Correcting Overdetection of Cardiac Events", U.S. Appl. No. 61/051,332, filed May 7, 2008, 62 pgs.
- "QT Interval", Wikipedia, [Online]. Retrieved from the Internet: <URL: [http://en.wikipedia.org/wiki/QT\\_interval](http://en.wikipedia.org/wiki/QT_interval)>, (Accessed May 11, 2011), 5 pgs.
- Gunderson, et al., "An Algorithm to Predict Implantable Cardioverter-Defibrillator Lead Failure", JACC, vol. 44, No. 9 (Nov. 2004), 1898-1902.
- Olson, Walter H, et al., "Onset and Stability for Ventricular Tachyarrhythmia Detection in an Implantable Pacer-Cardioverter-Defibrillator", IEEE, (1987), 167-170.
- Schuder, John C, "The Role of an Engineering Oriented Medical Research Group in Developing Improved Methods & Devices for Achieving Ventricular Defibrillation: The University of Missouri Experience", PACE, vol. 16, Part I, (Jan. 1993), 95-124.
- Schwake, H., et al., "Komplikationen mit Sonden bei 340 Patienten mit einem implantierbaren Kardioverter/Defibrillator", Z Kardiol, vol. 88, No. 8, (1999), 559-565.
- Swerdlow, C. D, et al., "Advanced ICD Troubleshooting: Part I", PACE, vol. 28, [Online]. Retrieved from the Internet: <[http://www.medscape.com/viewarticle/520588\\_print](http://www.medscape.com/viewarticle/520588_print)>, (Jan. 9, 2006), 1322-1346.
- Throne, Robert D, et al., "A Comparison of Four New Time-Domain Techniques for Discriminating Monomorphic Ventricular Tachycardia from Sinus Rhythm Using Ventricular Waveform Morphology", IEEE Transactions on Biomedical Engineering, vol. 38, No. 6, (Jun. 1991), 561-570.
- U.S. Appl. No. 13/436,398, Response filed Aug. 8, 2013 to Non Final Office Action mailed May 9, 2013, 13 pgs.
- U.S. Appl. No. 13/607,168, Response filed Jun. 25, 2013 to Non Final Office Action mailed Mar. 26, 2013, 12 pgs.
- U.S. Appl. No. 14/054,507, Non Final Office Action mailed May 9, 2014, 9 pgs.

(56)

**References Cited**

OTHER PUBLICATIONS

U.S. Appl. No. 14/054,507, Response filed Jan. 3, 2014 to Restriction Requirement mailed Dec. 6, 2013, 8 pgs.

U.S. Appl. No. 14/054,507, Restriction Requirement mailed Dec. 6, 2013, 6 pgs.

U.S. Appl. No. 14/096,285, Non Final Office Action mailed May 20, 2014, 6 pgs.

U.S. Appl. No. 14/096,285, Response filed May 8, 2014 to Restriction Requirement mailed May 8, 2014, 10 pgs.

U.S. Appl. No. 14/096,285, Restriction Requirement mailed Apr. 10, 2014, 5 pgs.

Australian Application Serial No. 2009221694, First Examination Report mailed Apr. 5, 2013, 4 pgs.

Australian Application Serial No. 2009221696, First Examiner Report mailed Apr. 5, 2013, 3 pgs.

Australian Application Serial No. 2009244153, First Examiner Report mailed Feb. 21, 2013, 4 pgs.

Australian Application Serial No. 2009244153, Subsequent Examiners Report mailed Aug. 30, 2013, 2 pgs.

Chinese Application Serial No. 200980116877.4, Response filed Jun. 14, 2013 to Office Action mailed Jan. 30, 2013, (W/ English Translation), 85 pgs.

Chinese Application Serial No. 200980116886.3, Office Action mailed Aug. 22, 2013, (W/ English Translation), 9 pgs.

Chinese Application Serial No. 200980116886.3, Response filed Apr. 25, 2013 to Office Action mailed Dec. 13, 2012, (W/ English Translation), 26 pgs.

Chinese Application Serial No. 200980125869.6, Response filed May 16, 2013 to Office Action mailed Jan. 10, 2013, (W/ English Translation), 8 pgs.

European Application Serial No. 09717667.1, Examination Notification Art. 94(3) mailed Sep. 12, 2013, 4 pgs.

European Application Serial No. 09743712.3, Office Action mailed Oct. 12, 2011, 4 pgs.

European Application Serial No. 09743712.3, Response filed Feb. 21, 2012 to Office Action mailed Oct. 12, 2011, 7 pgs.

Japanese Application Serial No. 2010-549930, Office Action mailed May 28, 2013, (W/ English Translation), 5 pgs.

Japanese Application Serial No. 2010-549930, Response filed Aug. 23, 2013 to Office Action mailed May 28, 2013, (W/ English Translation), 10 pgs.

Japanese Application Serial No. 2010-549931, Office Action mailed Feb. 12, 2014, (W/ English Translation), 9 pgs.

Japanese Application Serial No. 2010-549931, Office Action mailed Jun. 4, 2013, (W/ English Translation), 38 pgs.

Japanese Application Serial No. 2011-508690, Office Action mailed Apr. 8, 2014, (W/ English Translation), 7 pgs.

Japanese Application Serial No. 2011-508690, Office Action mailed Jul. 23, 2013, (W/ English Translation), 5 pgs.

Japanese Application Serial No. 2011-508690, Response filed Oct. 23, 2013 to Non Final Office Action mailed Jul. 23, 2013, (W/ English Claims), 11 pgs.

\* cited by examiner




FIG. 1




FIG. 2




FIG. 3A





FIG. 3B



102

	N	N-1	N-2	N-3
104	Low High Low	Low Middle Low	Low High Low	Low High Low
106	Middle	Low	Middle	Low





FIG. 4



124

	N	N-1	N-2	N-3
N-1	Low			
N-2	Low	Low		
N-3	<b>High</b>	Low	Low	
N-4	Low	<b>Mid / Low</b>	Low	Low
N-5		Low	<b>Mid / Low</b>	Low
N-6			Low	<b>High</b>
N-7				Low
Static	Middle	Low	Low	Middle

FIG. 5




FIG. 6




FIG. 7A

FIG. 7B




FIG. 8A




FIG. 8B




FIG. 9




FIG. 10




FIG. 11




FIG. 12




FIG. 13




FIG. 14A




FIG. 14B




FIG. 15




FIG. 16

## METHODS AND DEVICES FOR ACCURATELY CLASSIFYING CARDIAC ACTIVITY

### RELATED APPLICATIONS

The present application is a divisional of U.S. patent application Ser. No. 13/436,448, filed Mar. 30, 2012 now U.S. Pat. No. 8,600,489, which is a continuation of U.S. patent application Ser. No. 12/437,547, filed May 7, 2009 and published as US Patent Application Publication Number 2010-0004713, now U.S. Pat. No. 8,160,687, which claims the benefit of and priority to U.S. Provisional Patent Application No. 61/051,332, filed May 7, 2008 and titled METHODS AND DEVICES FOR IDENTIFYING AND CORRECTING OVERDETECTION OF CARDIAC EVENTS, and which is also a Continuation-In-Part of U.S. patent application Ser. No. 12/399,914, filed Mar. 6, 2009, published as US Patent Application Publication Number 2009-0259271 and titled METHODS AND DEVICES FOR ACCURATELY CLASSIFYING CARDIAC ACTIVITY, the disclosures of which are incorporated herein by reference.

The present application is related to U.S. patent application Ser. No. 12/399,901, filed Mar. 6, 2009 and titled ACCURATE CARDIAC EVENT DETECTION IN AN IMPLANTABLE CARDIAC STIMULUS DEVICE, published as US Patent Application Publication Number 2009-0228057, which claims the benefit of and priority to U.S. Provisional Patent Application No. 61/034,938, filed Mar. 7, 2008, the disclosures of which are incorporated herein by reference.

### FIELD

The present invention relates generally to implantable medical device systems that sense and analyze cardiac signals. More particularly, the present invention relates to implantable medical devices that capture cardiac signals within an implantee's body in order to classify cardiac activity as likely benign or malignant.

### BACKGROUND

Implantable cardiac devices typically sense cardiac electrical signals in an implantee and classify the implantee's cardiac rhythm as normal/benign or malignant. Illustrative malignant rhythms may include ventricular fibrillation and/or ventricular tachyarrhythmia. The accuracy with which an implantable medical device analyzes captured signals determines how well it makes therapy and other decisions.

New and/or alternative methods and devices for cardiac signal analysis are desired.

### SUMMARY

Some illustrative embodiments relate to the use of correlation analysis to identify overdetection of cardiac events. In one example, a High-Low-High pattern of correlation relative to a template is sought. The template may be a static template, it may be a representation of a recent captured event, or it may be an average of several recent captured events. In another example, multiple boundaries for High correlation are defined, wherein a first, higher boundary (requiring greater correlation) allows identification of overdetection based on a smaller set of detected events than a second, lower boundary. In one embodiment, a shorter sequence of High-Low-High is sufficient with the first boundary, while a longer sequence of five or more (for example, eight) alternating events is required

for the second boundary. In another embodiment, definitions of High and Low correlation are adapted to the particular signals by using average values for subsets of detected event correlations to establish boundaries.

In another embodiment, correlation analysis is performed multiple times for a given template and detected event by shifting the alignment of the template and the detected event to maximize the correlation score of the analysis. Such shifting may adjust the alignment by one or more samples away from the identified fiducial points for analysis. In another embodiment, stored templates are modified in order to accommodate changes in morphology for selected portions of the signal. In yet another embodiment, multiple features of the template and/or signal are identified and multiple correlation scores are calculated using several different features as alignment points.

When identified, overdetection can be corrected by modifying stored data in order to impact rate analysis. In one such embodiment, data correction is inhibited if the intervals surrounding a likely overdetection are longer than a predetermined threshold. In some embodiments, overdetection correction is inhibited if interval analysis relating to a likely overdetection indicates that it is unlikely to be a particular type of overdetection. In one such embodiment, the intervals surrounding a likely overdetection are analyzed to determine whether an accepted formula for estimating expected QT intervals is met and, if not, the method determines that the likely overdetection is not a T-wave, and so no data correction occurs.

### BRIEF DESCRIPTION OF THE DRAWINGS

FIG. 1 is a block diagram for an illustrative method of identifying overdetection and taking corrective action;

FIG. 2 shows an illustrative implantable cardiac stimulus system;

FIG. 3A shows an example using correlation analysis to identify overdetection;

FIG. 3B illustrates method steps for an illustrative example including rate correction;

FIG. 4 shows an example of inter-event correlation comparisons;

FIG. 5 shows another example of inter-event correlation comparisons;

FIG. 6 shows an analytical approach to short series and long series correlation analysis;

FIGS. 7A-7B illustrate examples of applying the analytical approach of FIG. 6 to series of correlation analyses;

FIGS. 8A-8B illustrate examples of tailoring correlation analysis to observed levels of correlation to a template;

FIG. 9 illustrates another method of aligning captured signal to correlation analysis templates;

FIG. 10 shows another method of storing and applying a template for correlation analysis;

FIGS. 11-12 illustrate a method of inhibiting correlation analysis identification of an overdetection;

FIG. 13 illustrates more methods for inhibiting correlation analysis identification of an overdetection;

FIGS. 14A-14B show application of a method illustrated in FIG. 13;

FIG. 15 shows a method of shock analysis for identifying shokable detected events and treatable rhythms; and

FIG. 16 illustrates a method of calculating the correlation between a captured signal and a template.

### DETAILED DESCRIPTION

The following detailed description should be read with reference to the drawings. The drawings, which are not nec-

essarily to scale, depict illustrative embodiments and are not intended to limit the scope of the invention. Some of the following examples and explanations include references to issued patents and pending patent applications. These references are for illustrative purposes and are not intended to limit the present invention to the particular methods or structures from those referenced patents and patent applications.

Unless implicitly required or explicitly stated, the methods below do not require any particular order of steps. It should be understood that when the following examples refer to a "current event," in some embodiments, this means the most recently detected cardiac event is being analyzed. However, this need not be the case, and some embodiments perform analysis that is delayed by one or more detections and or a fixed period of time. Choices shown regarding use of rectified/unrectified signals are merely illustrative, and may be changed if desired.

The nomenclature used herein indicates that a signal is sensed by an implantable cardiac device system, events are detected in the sensed signal, and cardiac activity is classified by use of the detected events (detections). Rhythm classification includes the identification of malignant rhythms, such as ventricular fibrillation or certain tachyarrhythmias, for example. Implantable therapy systems make therapy/stimulus decisions in reliance upon the classification of the cardiac rhythm.

In an illustrative example, a detected event is detected by comparing received signals to a detection threshold, which is defined by a detection profile. Any suitable detection profile may be used. Detected events are separated by intervals. Several intervals can be used to generate an average interval across a selected number of intervals, from which cardiac rate can be calculated. For example, four, eight or sixteen intervals may be used to estimate cardiac event rate as a function of average interval.

A cardiac electrogram includes several portions (often referenced as "waves") that, according to well known convention, are labeled with letters including P, Q, R, S, and T, each of which corresponds to particular physiological events. It is typical to design detection algorithms to sense the R-wave, though any portion of the cardiac cycle, if repeatedly detected, can be used to generate a beat rate. If morphology (shape) analysis is used in addition to heart rate, the system may capture and/or analyze the portion of the cycle that includes the Q, R and S waves, referred to as the QRS complex. Other portions of the patient's cardiac cycle, such as the P-wave and T-wave, are often treated as artifacts that are not sought for the purpose of estimating heart rate, though this need not be the case.

Typically, for purposes of ascertaining rate each cardiac cycle is counted only once. Overdetection (such as a double or triple detection) may occur if the device declares more than one detected event within a single cardiac cycle. Overdetection may occur if more than one portion of a single cardiac cycle is detected, or if noise causes an event to be declared when no cardiac event has taken place, for example, due to external therapy or noise, pacing artifact, skeletal muscle noise, electro-therapy, etc.

If one cardiac cycle takes place and a detection algorithm declares multiple detected events, overdetection has occurred. If the heart rate is then calculated by counting each of these detections, overcounting occurs. Calculated heart rates may be used alone or in combination with other factors to classify cardiac rhythms as malignant or benign. Overcounting in reliance on overdetected events can result in erroneously high rate calculation. Miscalculation of heart rate can lead to incorrect rhythm classification and therapy deci-

sions. Some of these concepts are further discussed and developed in U.S. patent application Ser. No. 12/399,914, titled METHODS AND DEVICES FOR ACCURATELY CLASSIFYING CARDIAC ACTIVITY, published as US Patent Application Publication Number 2009-0259271 and U.S. patent application Ser. No. 12/399,901, titled ACCURATE CARDIAC EVENT DETECTION IN AN IMPLANTABLE CARDIAC STIMULUS DEVICE, published as US Patent Application Publication Number 2009-0228057.

FIG. 1 is a process flow diagram for an illustrative method of identifying overdetection and taking corrective action. The illustrative method begins with event detection 10, where a received cardiac signal is captured and compared to a detection threshold until the received signal crosses the detection threshold, resulting in declaration of a detected event.

Next, the method performs an overdetection identification step 12. This may include one or more of several analysis methods including, as illustratively shown, morphology analysis 14, interval analysis 16 and wide QRS analysis 18. Following overdetection identification 12, if one or more overdetections are identified, the method corrects data, as shown at 20. If no data correction is needed at step 20, this step may be bypassed.

Finally, the method includes a therapy decision, as shown at 22. A therapy decision 22 may classify a cardiac rhythm of the implantee and determines whether/when therapy is to be delivered. The method then iterates to event detection 10.

The therapy decision 22 may include one or more of several forms of analysis. In one illustrative example, individual detected events are marked as shockable or non-shockable and an X-out-of-Y counter is maintained to determine whether the overall cardiac rhythm merits therapy. The marking of individual events as shockable or non-shockable may take several forms, including rate-based and/or morphology based determinations, or combinations thereof. FIG. 15, below, provides an illustrative example. Further examples are also discussed in U.S. Pat. No. 6,754,528, entitled APPARATUS AND METHOD OF ARRHYTHMIA DETECTION IN A SUBCUTANEOUS IMPLANTABLE CARDIOVERTER/DEFIBRILLATOR, and U.S. Pat. No. 7,330,757 entitled METHOD FOR DISCRIMINATING BETWEEN VENTRICULAR AND SUPRAVENTRICULAR ARRHYTHMIAS, the disclosures of which are incorporated herein by reference.

Therapy decision 22 may also take into account the persistence of a malignant condition. Some illustrative examples are shown in US Patent Application Publication Number 2006-0167503 titled METHOD FOR ADAPTING CHARGE INITIATION FOR AN IMPLANTABLE CARDIOVERTER-DEFIBRILLATOR, the disclosure of which is incorporated herein by reference. Other methods may be used as a part of the therapy decision 22.

FIG. 2 shows an illustrative implantable medical device and implant location. More particularly, an illustrative subcutaneous-only system is shown in FIG. 2. The subcutaneous system is shown relative to a heart 40, and includes a canister 42 coupled to a lead 46. The canister 42 preferably houses operational circuitry for performing analysis of cardiac activity and for providing a therapy output. The operational circuitry may include batteries, input/output circuitry, power capacitors, a high-voltage charging module, a controller, memory, telemetry components, etc., as known in the art.

Electrodes are disposed at locations throughout the system including, for example, an electrode 44 on the canister 42, and electrodes 48, 50, 52 on lead 46. The electrodes 44, 48, 50, 52 may take any suitable form and can be made of any suitable material. For example, the canister electrode 44 may be an

isolated button electrode or it may be a region or surface of the canister 42, and the electrodes 48, 50, 52 on lead 46 may be coil electrodes, ring electrodes, or other structures known in the art.

The electrodes 44, 48, 50, 52 define a plurality of sensing vectors such as V1, V2, V3 and V4. If desired, one or more vectors V1, V2, V3, and V4 may be chosen as a default sensing vector, for example, as discussed in US Patent Application Publication Number 2007-0276445 titled SYSTEMS AND METHODS FOR SENSING VECTOR SELECTION IN AN IMPLANTABLE MEDICAL DEVICE. Other uses of multiple vectors are shown, for example, in U.S. Pat. No. 7,392,085 titled MULTIPLE ELECTRODE VECTORS FOR IMPLANTABLE CARDIAC TREATMENT DEVICES. Another embodiment considers posture in vector analysis, for example, as discussed in US Patent Application Publication Number 2008-0188901 titled SENSING VECTOR SELECTION IN A CARDIAC STIMULUS DEVICE WITH POSTURAL ASSESSMENT. Multiple sensing vectors may be analyzed, sequentially or in combination, as desired.

Therapy may be applied using any chosen pair of electrodes. An illustrative example uses the can electrode 44 and the coil electrode 52 to apply therapy. Other electrode combinations may be used. Therapy may include monophasic or multiphasic defibrillation, cardioversion and/or pacing.

The present invention is not limited to any particular hardware, implant location or configuration. Instead, it is intended as an improvement upon any implantable cardiac therapy system. Some embodiments may also be used in a monitoring system to either control the monitoring functions (including annunciation and/or data recording) and/or to test the suitability of the data analysis to a particular configuration, condition or patient.

Some illustrative examples can associate with an external programmer 54 configured to communicate with the implanted device for various purposes, including, for example and without limitation, one or more of the following: device testing; upload new/revised software; modify programmable parameters such as detection or therapy settings; determine the status of device operation, battery life, or lead integrity; enable or disable functionality; and/or download data relating to the implantee's condition, prior data capture, or treatment. Any suitable communication method may be used, such as various protocols and hardware widely known in the art.

FIG. 2 omits several anatomical landmarks. The illustrative system shown may be implanted beneath the skin, outside of the ribcage of the implantee. The location illustratively shown would place the canister 42 at approximately the left axilla of the implantee, level with the cardiac apex, with the lead 46 extending medially toward the xiphoid and then toward the head of the implantee along the left side of the sternum. One illustrative example uses a method/system as shown in commonly assigned US Patent Application Publication Number 2006-0122676 entitled APPARATUS AND METHOD FOR SUBCUTANEOUS ELECTRODE INSERTION, now U.S. Pat. No. 7,655,014. Other illustrative subcutaneous systems and locations are shown in commonly assigned U.S. Pat. Nos. 6,647,292, 6,721,597 and 7,149,575.

The present invention may also be embodied in systems having various implant configurations including, for example, other subcutaneous-only, vascular-only, and/or transvenous implantation configurations/locations. The canister 42 may be placed in anterior, lateral, and/or posterior positions including, without limitation, axillary, pectoral, and sub-pectoral positions, as well as placements on either the left or right side of the implantee's torso and/or in the abdomen.

Entirely intravascular implantation of the system has also been proposed. The canister 42 and lead 46 may be placed in any of a number of suitable configurations including anterior-posterior combinations, anterior-only combinations, transvenous placement, or other vascular placements. A unitary system may omit lead 46 and instead include all electrodes on the canister 42.

FIG. 3A shows an example using correlation analysis to identify overdetection. "Correlation analysis" as used herein can take several forms. One illustrative example is shown in FIG. 16. Referring to FIG. 16, a captured signal 500 undergoes analog-to-digital conversion 502 to yield a time ordered series of samples {S1 . . . S9} that form a sampled (and usually digital) representation of the signal, as indicated at 504. The example in FIG. 16 is simplified for illustrative purposes as the number of samples for a given signal may be greater than nine. For example, in one illustrative embodiment, the captured signal 500 is about 160 milliseconds long, covering 41 samples captured at 256 Hz. Other durations and/or sampling frequencies may be selected. The signal can be windowed to approximately the QRS width, though this is not required.

The signal representation is compared to a template using correlation analysis 506. The template is shown as comprising a series of sample values {T1 . . . T9}. Prior to comparison, or as part of the comparison, the signal representation or template is scaled such that the largest peaks of the two data sets are equal in amplitude. One example of correlation analysis is correlation waveform analysis. Other examples are widely known in the art.

A simple version of correlation analysis is shown graphically in FIG. 16: the largest sample or peak of the signal representation is aligned with the peak of the template and the surrounding samples are compared to one another as shown at 508. Because the peaks are already scaled to be equal, there is no difference at the peak, but the surrounding samples may differ. Differences between the signal representation and the template are shown in cross-hatching.

Next a correlation score may be calculated as shown at 510. The sum of the absolute values of the differences between (scaled) samples of the signal representation and samples of the template is calculated and divided by the total area under the template. The quotient is subtracted from one, yielding a correlation score 512. If the correlation score is close to one, then the area of difference is small relative to the area under the template, indicating high correlation. Other methods for calculating correlation are known in the art and may be substituted; that shown in FIG. 16 is simply an example. For example, a weighted CWA may apply a weighting factor to individual sample differences in a fashion as shown in commonly assigned, copending US Patent Application Publication Number 2008-0077030, which is now U.S. Pat. No. 8,014,851.

Returning to FIG. 3A, individual events are detected by applying a detection profile 70 to a signal 72. The detection profile 70 includes a refractory period 74 followed by a constant threshold period 76 and a decay period 78. Other shapes may be used for the detection profile 70.

The signal 72 has R-waves and T-waves highlighted. In the example shown, the T-waves are large relative to the R-waves. The refractory periods shown in cross-hatching over both R-waves and T-waves indicates that each R-wave and each T-wave is being treated as a detected event. As a result, for each cardiac cycle, the detection profile 70 is detecting two events. This is one example of overdetection.

In the illustrative example, each of the individual detections is also being treated to correlation analysis relative to a template that is based on an R-wave. The results of the cor-

relation analysis are plotted at **80**. Plot **80** includes boundaries for “High” and “Low” correlation. In the example, each “X” indicates the correlation score for each detected event. A High-Low-High pattern of correlation scores occurs as shown at **82**. In the example, each High-Low-High sequence leads to a conclusion that “Low” scoring detected events are overdetected. As a result, as shown, the “Low” scoring detected event will be discarded when a High-Low-High pattern is found. In a numeric example, “High” is defined as greater than 52% correlation, while “Low” is defined as less than 25% correlation, when calculated using the form shown at **510** in FIG. **16**. Other values and analytical methods can be used.

FIG. **3B** illustrates method steps for an illustrative example including rate correction. Once a morphology overdetection pattern is found, as indicated at **90**, one or more overdetections are identified, as shown at **92**. Next, event intervals and/or rate are recalculated, as shown at **94**.

For example, as shown at **96**, a series of detections of R and T waves may result in a set of interval calculations of 225 ms (R to T) and 300 ms (T to R), which yields an average interval of 263 ms. An average interval of 263 milliseconds leads to a rate of about 229 beats-per-minute, which would be a treatable tachyarrhythmia in many patients. However, when the T-waves are identified as overdetections and the intervals on either side of the T-waves are combined, as shown at **98**, the intervals average 525 milliseconds. The rate can be recalculated to about 114 beats-per-minute, avoiding possible defibrillation, cardioversion or pacing that could result without the data correction.

FIG. **4** shows an example of inter-event correlation comparisons. An inter-event comparison is a comparison in which two individual detected events are compared to one another. The comparison may take the form of a correlation analysis, or it may make use of some other type of analysis such as wavelet transform, principal component analysis (PCA), etc., to consider the similarity between two detected events. In wavelet transform or PCA comparisons, the similarity of the results of data compression into wavelet or PCA outputs can be compared. For example, the similarity and/or order of eigenvalue outputs of PCA, or the similarity of the wavelet coefficients resulting from a wavelet transformation can be compared in a qualitative or quantitative manner.

In the example shown in FIG. **4**, a correlation analysis is performed. In the example, as shown at **108**, correlation scores are characterized as Low, Middle, or High. The “High” score zone indicates strong confidence that the compared signals are of the same character (for example, if one event is an R-wave, so is the other), while “Low” scores indicate that the compared signals are very different from one another. The “Middle” zone is intended to capture those signals that are similar but that do not create strong confidence that the two signals are of the same character. For example, in a patient who undergoes a rate-dependent morphology change (such as a rate-induced bundle block), captured R-waves may not highly correlate to a stored static template but likely fall into the Middle range relative to the template. In another example, a monomorphic VT likely has High or Middle inter-event correlation between R-waves, and Middle correlation between T-waves, while a polymorphic VT would show Middle or Low correlation between R-waves.

If desired, fuzzy logic may be applied. The use of a “Middle Zone” suggests this. For example, rather than simple “High” and “Low” characterizations, additional categories may be provided. Further, a previous measurement may be used to inform a subsequent characterization of a marginally similar or dissimilar signal.

As shown at **100**, a series of events N, N-1, N-2 and N-3 are considered as a group, with the N<sup>th</sup> detection compared to each of N-1, N-2 and N-3 via correlation analysis. The results of inter-event comparisons and comparisons to a static template are shown in a table at **102**. The inter-event comparison results are shown at **104**, and include ordered results for comparison of a given event to three prior events. Table **102** shows results for events N, N-1, N-2 and N-3. The results of the inter-event comparisons show that for any given event X, the correlation to X-2 is higher than for X-1 or X-3. This may indicate a pattern of double detection based on increased correlation between alternating events.

In the illustrative example, comparisons to a static, normal sinus rhythm template may be performed as well. Illustrative results are shown at **106**. The alternating static template results, Low-Middle-Low-Middle . . . are suggestive of possible overdetection, but because the likely R-waves do not Highly correlate, strong confidence does not result based on static template alone. However, when taken in combination with the inter-event comparison information, there is significant confidence that some events are overdetections. An applicable rule set may be as follows:

- 1) Alternating Low-High-Low for N when compared to N-1, N-2 and N-3, and
- 2) Alternating Low-High-Low for N-2 when compared to N-3, N-4 and N-5.

Conclusion: Treat N-1 and N-3 as T-waves.

A further, confirmatory rule may be:

- 3) At least “Medium” correlation for N and N-2 to static template.

Another approach is to apply only rules 1) and 3), while marking only the N-1 as an overdetection in response to the rule set being met. Once one or more events are marked as overdetections, they may be treated in the manner shown in FIG. **3B**, above.

FIG. **5** shows another example of inter-event correlation comparisons. Here the captured signal is triple-detected, as shown at **120**. In this instance, the N<sup>th</sup> detection is compared to each of N-1, N-2, N-3 and N-4. The inclusion of four individual comparisons may further assist in distinguishing a triple detection from a double detection, although some embodiments stop at three comparisons.

The results are shown in the table at **124**. For each set of comparisons, there are three Low correlations, and one Middle or one High correlation. It is likely that with triple detection, some detections will have a low correlation in each comparison. An illustrative rule set is as follows:

1. N<sup>th</sup> event has High correlation to the N-3 event;
2. N-1 and N-2 events have Low correlations to the N<sup>th</sup> event; and
3. N-1 and N-2 events have Low correlations to the Static Template.

If these three conditions are met, then N-1 and N-2 may be discarded. Further conditions may be added. For example, the static template characteristics of N and/or N-3 may be considered as well, for example:

4. N<sup>th</sup> and N-3 events have Middle or High Correlation to Static Template.

Then if all of 1-4 are met, N-1 and N-2 may be discarded and the interval from N to N-3 calculated and used in rate analysis.

In a further example, the widths of each event may also be considered, for example using this fourth condition:

5. N-1 and N-2 events are wider than a Width Threshold.

The width threshold may be set as desired; in one example the Width Threshold is in the range of 100-140 ms. This Width Threshold rule may be applied as an added layer to any

determination that an event is to be discarded as an over-detection. In another example, the polarities may be considered:

6. N-1 and N-2 each share the same polarity.

Polarity may be defined, for example, by reference to the majority of signal samples for an event, as the polarity of the sample having the greatest magnitude in the event, or by determination of which extreme, the most positive or least positive, in the event occurs first.

If desired, interval coupling may be added as another condition:

7. The combined interval N to N-3 less than Duration.

Where "Duration" is in the range of 800-1200 ms. This condition, and variants thereof, is also explained in association with FIGS. 11-13 and 14A-B below.

FIG. 6 shows an analytical approach to short series and long series correlation analysis. FIG. 6 shows a plot 140 for plotting the correlation scores for a series of detected events. The correlation scores, shown as X's, are plotted against lines 144 and 146 that define a wide band 148, and lines 150, 152 that define a narrow band 154.

The wide band 148 is applied to identify an over-detection when there are two detected events with scores above line 144 separated by a single detected event with a score below line 146, for example as shown in FIG. 7A. The narrow band is applied to identify over-detection(s) when a series of consecutive detections alternate above line 150 and below line 152, for example as shown in FIG. 7B. Numbers are shown for each threshold for illustrative purposes; these numbers may use correlation as a percentage.

The narrower band 154 applies a less stringent standard than the wider band 152 with regard to the correlation scores, and therefore more events are analyzed before making a decision to discard low scoring events. In one illustrative example, events are not discarded using the narrow band 154 until the 8 event pattern shown in FIG. 7B is met, at which point one to four of the low scoring events are discarded, with intervals around each discarded event being corrected. Subsequent to meeting the pattern in this initial step, only the newest low scoring event would be discarded. For analytical purposes, previously discarded events are used to determine whether the 8-consecutive-outside rule is met, even if those events are excluded from rate calculations. Another embodiment uses only five events, looking for a High-Low-High-Low-High sequence using the narrower band 154 and, if such a sequence is found, one or both of the Low scoring events is discarded.

The examples in FIGS. 6 and 7A-7B indicate numbers, with 50% and 20% correlations bordering the wide band 148 and 40% and 25% bordering the narrow band 154. These numbers are merely illustrative. In one example, these numbers are applied by scaling the formula shown at 510 in FIG. 16 to a percentage basis.

FIGS. 8A-8B illustrate examples of tailoring correlation analysis to observed levels of correlation to a template. Referring to FIG. 8A, a plot of correlation scores for comparing a template to a series of events is shown at 158. For purposes of identifying double detections, a mean correlation score is calculated for the odd numbered events. Clustering of the odd numbered events is then analyzed by determining whether the odd numbered events all fall within a predefined distance from the mean, for example, using the standard deviation of the set, or using a fixed distance. If the odd numbered events all fall within the predefined distance from the mean, the separation of the mean from a Low boundary is calculated. If the separation is greater than a predetermined threshold, then it is determined that the odd numbered events demonstrate monotonicity supporting a presumption that the odd numbered events are QRS complex detections. If monotonicity of

the odd numbered events is identified, one or more of the even numbered events that fall below the low threshold are marked as over-detections.

In another embodiment, before any of the even numbered events are marked as over-detections, they are all analyzed to determine whether clustering of the even numbered events has taken place, again using the mean of those events. Rather than separation of the odd-numbered event mean from a low boundary, separation between the even and odd event means is calculated to establish groupings of the events. In this embodiment, over-detection markers are applied only when sufficient clustering of the even-numbered events appears.

FIG. 8B shows another example in which the marking of over-detections is tailored to correlation scores to a static template. Here, the average correlation score for a set of 10 events is calculated. A "blank" band is then established around the average correlation score. For example, the blank band may be defined as  $\pm 15\%$ . Other "blank band" sizes may be used.

In the example of FIG. 8B, high scores are defined as those scores that fall above the blank band, and low scores are those falling below the blank band. If a pattern of High-Low-High appears around the blank band, then over-detection can be identified and one or more of the Low scoring events is marked as an over-detection.

Instead of a static template, the analysis shown by FIGS. 8A-8B may also be applied using a recently detected event as the template for comparison. The analysis noted for FIGS. 8A-8B may use calculation of the mean/average, or it may use some other predictor of a center-point for signals including the mode, median or other mathematical operation.

A further use of the inter-event comparisons shown here may be in the determination of whether a Shockable rhythm is occurring. Stimulus delivery is often used to address polymorphic conditions, such as Polymorphic Ventricular Tachycardia and Ventricular Fibrillation. Monomorphic conditions such as Monomorphic Ventricular Tachycardia (MVT) can be treated, but MVT does not always require the most energetic treatments. For example, MVT may be treated using anti-tachycardia pacing (ATP) in place of defibrillation or cardioversion, as ATP uses less energy and may be less traumatic to the patient. Patterns of correlation can be used to distinguish monomorphic arrhythmias from polymorphic arrhythmias. For example, an ongoing pattern as shown in FIG. 7A or 7B, or even FIG. 6, in which high correlations are consistently found, can be used to delay therapy, if desired.

In another example, a pattern as shown in FIG. 8A may be further analyzed by determining the size of the standard deviation for the clustered high scores. If the clustered high scores are based on a static template and show a low standard deviation, this may indicate a monomorphic condition. In some embodiments, particularly if ATP is not available, therapy may be inhibited until the monomorphic condition breaks down into a more polymorphic condition.

In one example, a system uses a tiered correlation analysis to identify treatable arrhythmias. In the example, a simple, single event correlation analysis using a static template is executed until a pattern as shown in FIG. 8A appears. Such a pattern then triggers multiple inter-event comparisons as shown in FIGS. 4-5. Then, if the inter-event comparisons show likely over-detection, interval data may be corrected. Further, if inter-event comparisons show a monomorphic condition, therapy may be inhibited.

FIG. 9 illustrates methods for aligning and realigning a captured signal to a correlation analysis template. The correlation analysis template is shown at 200, with a signal shown at 202. The correlation analysis template 200 may be a static

template or it may represent a single detected event or average of several recently detected events.

As noted in FIG. 16, correlation analysis typically uses a fiducial point as an alignment guide for an ordered series of template values and signal samples. In the example of FIG. 9, a base alignment point is identified as the sample of each of the template 200 and the signal 202 having the greatest magnitude. A series of comparisons are then made, beginning with a base aligned comparison, shown at 210, and single-sample shifts to the right, shown at 212, and the left, shown at 214. The shift one right correlation 212 is worse than the correlation score for the base comparison 210, and so the result of the shift one right correlation 212 is discarded. The shift one left correlation 214 yields a higher correlation score than the aligned correlation 210, so the result of the base correlation 210 is discarded, and another shift left correlation is calculated as shown at 216, this time offsetting the alignment points by two samples. The result at 216 shows lesser correlation than the shift-one-left correlation at 214, and so the process stops and uses the correlation score calculated for the shift-one-left correlation 214 as the correlation score for the signal 202.

When performing the shifting to the right and/or left, scaling of the signal to the template may be modified as well. For example, if scaling is initially performed by comparing the peak for the signal to the peak for the template and then equalizing the two, on shifting, the peak for the signal may instead be scaled to the point it aligns to in the template after shifting has occurred.

The method demonstrated in FIG. 9 may help to correct for noise or misalignment based on sampling artifact, slew rate, etc., that may cause the peak alignment point of the sample 202 to be less than optimal. The method includes calculating the correlation score when the fiducial points are aligned and also when the fiducial points are misaligned by one or more samples in each of two directions until a maximum correlation score is found. Limits may be placed, as desired, on the number of samples to shift to the left or right. In another embodiment, several (for example, one base, one, two, and three to the left, one, two and three to the right) scores are automatically calculated and the best is chosen.

In another embodiment highlighted in FIG. 9, plural alignment points can be defined for the template 200. Some examples include the QRS onset, the maximum amplitude, the maximum amplitude in the opposite polarity of the maximum amplitude (note the maximum amplitudes are indicated by each being a turning point where  $dV/dt=0$ ), the maximum slope point between the two major peaks (shown as  $dV/dt=MAX$ , etc.). By identifying the analogous points in the signal, the method can determine whether use of different possible alignment points would provide different correlation analysis outcomes. For example, the default may be to use the maximum amplitude point of the entire signal, but it may be that some cardiac events can be aligned instead using the maximum slope point in the monotonic segment that follows the maximum amplitude point.

FIG. 10 shows another method of storing and applying a template for correlation analysis. In this example, the signal forming a basis for a template is shown at 230. For the illustrative example, when the template is formed an interpolation region is defined between the positive peak and the negative peak of the signal 230. As a result, the stored template takes the form shown at 240: The template 240 matches the template signal 230 for regions before the positive peak and after the negative peak, but is flexible between the two peaks, as indicated by the dashed line at 242. The positive peak, in the

example shown, is the largest magnitude peak in the template, and so it is used for scaling the template to a captured signal.

Alignment to a sample 232 is then performed as shown at 244. The template is adjusted such that the positive and negative peaks are aligned with the captured signal, with a linear interpolation therebetween. Outside of the positive and negative peaks, the template continues to match the signal as shown at 230, however, the duration and slope between the positive and negative peaks are adjusted to match the captured event. The adjustment shown in FIG. 10 may avoid the difficulty of a static template being fixed in duration for a patient whose QRS width is affected by rate. The adjustment made may be limited in order to avoid excessively widening the template.

In another example, more than two template points are identified and linear interpolation may be used between them. For example, a template may be composed of five values each having a relative amplitude and relative location. When a detected event is to be compared to the template, the width and peak amplitude of the detected event are used to scale each of the values of the template, with linear interpolation between the template points.

FIGS. 11-12 illustrate a method of inhibiting data correction following identification of a likely overdetection.

As shown in FIG. 11, a QRS complex occurs at 260, followed by a premature ventricular contraction (PVC) shown at 262, following by another QRS complex at 264. The PVC is characterized, in this example, by a low correlation to the template. Thus, a High-Low-High correlation pattern appears, similar to that shown above in FIG. 3A. Some examples would therefore discard the PVC 262. Analytically, however, discarding the PVC 262 may be unnecessary since it is not actually an overdetection event. Further, the intervals around the PVC 262 are both greater than 500 milliseconds. Even without data correction, the average of the two intervals would yield an event rate of about 103 beats-per-minute, a rate that would not threaten to cause unnecessary therapy. Thus the data correction would not improve rhythm specificity in the device, while reducing beat sensitivity.

FIG. 12 illustrates a method that would avoid discarding a PVC 262 as shown in FIG. 11. Based on detected events 270, the method determines, as shown at 272, whether a correlation score sequence appears that would support a finding of double detection (DD) or overdetection. If not, the method ends, as no data correction is about to ensue. If the result from 272 is a "Yes," the method next includes determining whether the new interval that would result from data correction would be greater than a predetermined threshold, as shown at 274. In the illustrative example, the threshold is 1000 ms (60 beats-per-minute), though this number is merely illustrative. Some likely thresholds are in the range of 750-1200 milliseconds.

In another example, the order of analysis is reversed, and the overdetection analysis does not take place unless the calculated rate is high (often 150 bpm or more), or unless the intervals that could be affected are short enough to pass the applied test. In another embodiment, individual intervals are compared to a threshold (for example, in the range of 400-600 ms) and, if the individual intervals both exceed the threshold, then no interval combining occurs. In yet another example, the threshold may be a programmable parameter of an implantable system. In another example, the threshold may be scaled on the basis of a programmable VT parameter that is used to set a beat rate that the implantable system will treat as a ventricular tachycardia rate.

If the corrected interval is not longer than the threshold, the method continues to the step of combining intervals, as shown at 276, to correct for the overdetection event(s). If the cor-

rected interval would be longer than the threshold at step 274, the method simply ends without combining intervals. In this fashion, unnecessary correction of the stored data can be avoided.

FIG. 13 illustrates more methods for inhibiting correlation analysis after identification of an overdetection. The methods in FIG. 13 take advantage of known relationships between the QT interval and the RR interval of physiologic cardiac cycles. The illustrative method again begins with the identification of a pattern that suggests overdetection, as indicated at 300. As shown at 302, the possible overdetection event is then treated as a T-wave (here, the presumption is that a three-event pattern is identified, with the middle event of the three being the likely overdetection; other variants may be used) and, as shown at 304, the events on either side of the likely overdetection are treated as R-waves.

These “presumed” R and T waves from steps 302 and 304 are then used to apply a formula for calculating the QT length from the RR interval in step 306. In particular, several likely formulae are shown at 308. Examples include Bazett’s formula:

$$QT(\text{Exp})=QT*\sqrt{RR}$$

Friderica’s formula:

$$QT(\text{Exp})=QT*3\sqrt{RR}$$

And the Sagie et al. regression formula:

$$QT(\text{Exp})=QT+A*(RR-1)$$

Sagie et al. found  $A=0.154$ .

In each formula, the expected QT is shown as QT(exp), the value RR is given in seconds, and the value QT is captured during a programming session between an implant and a programmer. QT is either captured at or adjusted for a 60 beat-per-minute cardiac rate. The RR interval is found at step 304, and the measured QT interval can be captured by adding the measured width of the presumed T-wave to the interval between the first R-wave and the presumed T-wave.

The expectation is that if the likely overdetection event is an overdetection T-wave, the measured QT period will match the expected QT value given RR, using whichever formula is applied, with some band allowing for error.

If the formula applied at 306 does not yield a match, no discard occurs, as shown at 310. Alternatively, if the formula applied at 306 yields a match, then the likely overdetection is discarded as shown at 312. When the likely overdetection is discarded at 312, intervals around the overdetection are combined, as shown above in FIG. 3B. Once again, the order of analysis is reversed in other examples.

FIGS. 14A-14B show application of a method illustrated in FIG. 13. In the illustrative examples of FIGS. 14A and 14B, Friderica’s cube-root formula is applied. In each example, the previously measured QT=400 milliseconds. This value represents the estimated QT interval for the hypothetical patient that would occur at a heart rate of 60 bpm.

Referring to FIG. 14A, given three events X, Y and Z having a correlation pattern indicating overdetection, the method is applied by presuming that Y is a T-wave. The QT interval is measured for X and Y, and the RR interval is measured from X to Z, as indicated. The measured QT is referenced as well, and these values are plugged into the chosen formula. In the example, shown, using  $RR=0.8$  seconds, the expected value for QT is 371 milliseconds. Applying a  $\pm 10\%$  error band for the calculation, the acceptable range is about 334-408 milliseconds for QT. However, as shown, the measured interval is about 500 milliseconds, too long to be a QT interval for the given parameters. As a result,

the calculation suggests that the Y detection is not an overdetection T-wave, and therefore no data correction occurs. Lesser or greater error band sizes may be applied; for example,  $\pm 5\%$  error is used in another illustrative embodiment.

Referring instead to FIG. 14B, this time, the QT interval measured for X and Y is about 370 milliseconds. This value falls within the expected range, and therefore the calculation suggests that the Y detection is an overdetection T-wave. Therefore the Y detection is discarded and the interval data between X and Z is corrected.

In the examples of FIGS. 11-13 and 14A-B, if a likely overdetection is not discarded, resulting in data correction, the likely overdetection may instead be marked as a suspect detection. In an example, suspect detections are treated as unreliable, both as indicators of cardiac activity and as endpoints for intervals that can be used in rate analysis. If the likely overdetection is marked as a suspect detection, the suspect detection and each of the preceding and following intervals around the suspect detection are removed from analysis entirely.

FIG. 15 shows a method of analysis for identifying shockable detected events and treatable rhythms. FIG. 15 shows the overarching structure of an analysis method by including the steps of event detection 402, which is followed by waveform appraisal 404 and beat qualification 406. In particular, event detection 402 will typically include monitoring a captured signal to detect signal amplitude changes that indicate cardiac events. Once cardiac events are captured at block 402, waveform appraisal 404 can occur. During waveform appraisal 404, the characteristics of the signal associated with a detected event are analyzed to identify and eliminate detected events that are likely caused by noise or artifacts.

Next, detected events that pass waveform appraisal 404 undergo beat qualification 406, during which detected events are analyzed to determine whether they display morphology or interval characteristics that indicate accurate detection. This may include the correlation analyses shown above, and/or analysis of intervals or combinations of the two, for example analysis to eliminate wide complex double detection can use detected event proximity and shape characteristics to identify likely overdetections. Some further discussion appears in U.S. patent application Ser. No. 12/399,914, titled METHODS AND DEVICES FOR ACCURATELY CLASSIFYING CARDIAC ACTIVITY, and published as US Patent Application Publication Number 2009-0259271.

The architecture then turns to rhythm classification, which can begin by consideration of rate at block 408. If the rate is low, then an individual detection is marked as “Not Shockable” as indicated at 410. Alternatively, if the rate is very high, it is considered to indicate ventricular fibrillation (VF) and therefore is marked as “Shockable,” as shown at 412. Between these low and VF bands of rates is a ventricular tachycardia (VT) zone, and rates in the VT zone are analyzed using what will be referred to as Detection Enhancements, as shown at 414.

An example of a Detection Enhancement is as follows:

1. Compare to static template: If Match, not shockable; else
2. Compare to dynamic template: If no Match, shockable event; else
3. Compare to QRS width threshold: If wide, shockable, else not shockable.

Where the dynamic template can be any of the following:

- a) An average of several previous detections that correlate to one another;
- b) A set of individual events, for example  $\{N-1 \dots N-i\}$  wherein matching some or all of the individual events counts as matching the dynamic template;

c) A continually updated template.

The QRS width threshold noted above may be applied in various ways that can be tailored to the method of QRS width measurement used in a given system and/or that may be tailored to an individual patient. In one example, the following rules apply to QRS width:

- x) QRS width, during analysis, is calculated as the duration from the start of the longest monotonic segment captured during refractory before the fiducial point to the end of the longest monotonic segment captured during refractory after the fiducial point;
- y) QRS width threshold is measured for the patient during a programming session, with a maximum allowed value of 113 ms; and
- z) QRS width during analysis is considered wide if it is at least 20 ms longer than the QRS width threshold.

These rules x), y) and z) are tailored to one particular embodiment and may vary depending on the system used.

Following the marking of events as Not Shockable **410** or Shockable **412**, an X/Y counter condition is applied as indicated at **416**. The X/Y counter condition analyzes the number of Shockable events, X, that are marked during a previous set, Y, of detected events that pass both waveform appraisal **404** and beat qualification **406**. The ratio applied, and set size used, may vary. One embodiment applies an 18/24 X/Y counter condition at **416**. Other embodiments use ratios as 8 or 9 out of 12, 12 or 13 out of 16, 24/32, etc.

If the X/Y condition is not met, no shock will be delivered, as shown at **418**. If the X/Y condition is met, then the method may proceed to a charge confirmation block **420**. For example, some embodiments require that the X/Y ratio/set size be met for a selected number of consecutive events, and this condition may be tested in charge confirmation **420**. Another example condition is to determine whether a set, N, of immediately preceding detected events are all Shockable, or all have intervals that are sufficiently short to support a conclusion that the detected arrhythmia is ongoing. Other factors may also be applied in charge confirmation, for example, by observing whether overdetection has been recently noted (which may suggest that therapy should be delayed to ensure that the "arrhythmia" is not a manifestation of overcounting), or observing whether consistent long intervals have been detected (potentially suggesting spontaneous conversion to normal rhythm by the patient). For example, charge confirmation **420** may also include methods such as those shown in commonly assigned and copending U.S. patent application Ser. No. 11/042,911, titled METHOD FOR ADAPTING CHARGE INITIATION FOR AN IMPLANTABLE CARDIOVERTER-DEFIBRILLATOR, published as US Patent Application Publication Number 2006-0167503, the disclosure of which is incorporated herein by reference.

The Charge and Shock block **422** is reached if Charge Confirmation **420** is passed. Typically the process of charging takes some period of time, and so the method **400** may iterate several times before charging is completed. Some or all of the analysis used to reach an initial determination that Charging should start may be repeated during this process. Finally, if treatable conditions persist during charging, or are identified following charging, stimulus may be delivered.

With regard to the implantable system, various hardware features may be incorporated. For example, any suitable battery chemistry, such as a lithium ion battery, may be used. Therapy output can be created using a capacitive system to store energy until a stimulus level is reached using one or several capacitors. A charging circuit such as a flyback transformer circuit can be used to generate therapy voltages.

Therapy can be delivered using, for example, an H-bridge circuit or a modification thereof. Dedicated or general purpose circuitry may be used to perform analysis functions. For example, a dedicated cardiac signal analog-to-digital circuit may be used, as well as a dedicated correlation analysis block, as desired, while other functions may be performed with a microcontroller. Static and dynamic memories may be provided and used for any suitable functions. These elements may all be components of the operational circuitry for the implantable cardiac stimulus system.

Those skilled in the art will recognize that the present invention may be manifested in a variety of forms other than the specific embodiments described and contemplated herein. Accordingly, departures in form and detail may be made without departing from the scope and spirit of the present invention.

What is claimed is:

1. A method of operation in an implantable cardiac stimulus (ICS) system comprising a canister housing operational circuitry for the ICS system, the canister having a canister electrode disposed thereon, and a lead having at least a first lead electrode disposed thereon, the canister electrode and the first lead electrode electrically coupled to the operational circuitry; the method comprising:

detecting electrical events occurring within the patient; constructing event representations for the detected electrical events;

for a set of events N, N-1, N-2 and N-3, comparing the Nth event to each of the N-1, N-2 and N-3 events by correlation analysis;

determining, using the results of the correlation analysis, whether the set of events indicates one of the following: overdetection,

benign cardiac rhythm, or

likely cardiac arrhythmia; and

if likely cardiac arrhythmia is identified, determining whether stimulus is needed; and

if stimulus is needed, delivering electrical stimulus to the patient from the ICS system;

wherein overdetection is identified if the correlation analysis indicates a pattern of correlations between the events N, N-1, N-2 and N-3 that corresponds to overdetection.

2. The method of claim 1 wherein the method includes determining that overdetection is occurring if the step of comparing the Nth event to each of the N-1, N-2 and N-3 events by correlation analysis results in a pattern of correlations of the following type: low correlation to N-1 and N-3 events, and high correlation to the N-2 event.

3. The method of claim 1 further comprising:

calculating a cardiac rate using the detected electrical events; and

if the set of events, N, N-1, N-2, N-3 indicates overdetection, correcting data relating to one or more overdetected events and recalculating the calculated cardiac rate.

4. The method of claim 1 wherein the step of determining whether the set of events indicates one of the following: overdetection, benign cardiac rhythm, or likely cardiac arrhythmia includes determining whether the Nth event correlates highly to at least one of the N-1, N-2 or N-3 events and, if so, determining that a benign cardiac rhythm is occurring.

5. The method of claim 4 wherein the step of determining whether the set of events indicates one of the following: overdetection, benign cardiac rhythm, or likely cardiac arrhythmia includes determining whether the Nth event correlates closely to each of the N-1, N-2 or N-3 events and, if so, determining that a benign cardiac rhythm is occurring, further

wherein “correlates highly” indicates a relatively higher correlation than “correlates closely.”

6. An implantable cardiac stimulus (ICS) system comprising a canister housing operational circuitry for the ICS system, the ICS including a plurality of electrodes including at least a canister electrode disposed on the canister, and a lead having at least a first lead electrode disposed thereon, the canister electrode and the first lead electrode electrically coupled to the operational circuitry; wherein the operational circuitry is configured to perform a method of cardiac signal analysis comprising:

detecting electrical events occurring within the patient using the electrodes of the ICS system;

constructing event representations for the detected electrical events;

for a set of events N, N-1, N-2 and N-3, comparing the Nth event to each of the N-1, N-2 and N-3 events by correlation analysis;

determining, using the results of the correlation analysis, whether the set of events indicates one of the following: over-detection,

benign cardiac rhythm, or

likely cardiac arrhythmia; and

if likely cardiac arrhythmia is identified, determining whether stimulus is needed; and

if stimulus is needed, delivering electrical stimulus to the patient from using the electrodes of the ICS system;

wherein over-detection is identified if the correlation analysis indicates a pattern of correlations between the events N, N-1, N-2 and N-3 that corresponds to over-detection.

7. The ICS of claim 6 wherein the operational circuitry is further configured such that the method includes determining that over-detection is occurring if the step of comparing the Nth event to each of the N-1, N-2 and N-3 events by correlation analysis results in a pattern of correlations of the following type: low correlation to N-1 and N-3 events, and high correlation to the N-2 event.

8. The ICS of claim 6 wherein the operational circuitry is further configured such that the method further comprises: calculating a cardiac rate using the detected electrical events; and

if the set of events, N, N-1, N-2, N-3 indicates over-detection, correcting data relating to one or more over-detected events and recalculating the calculated cardiac rate.

9. The ICS of claim 6 wherein the operational circuitry is further configured such that the step of determining whether the set of events indicates one of the following: over-detection, benign cardiac rhythm, or likely cardiac arrhythmia includes determining whether the Nth event correlates highly to at least one of the N-1, N-2 or N-3 events and, if so, determining that a benign cardiac rhythm is occurring.

10. The ICS of claim 9 wherein the operational circuitry is further configured such that the step of determining whether the set of events indicates one of the following: over-detection, benign cardiac rhythm, or likely cardiac arrhythmia includes determining whether the Nth event correlates closely to each of the N-1, N-2 or N-3 events and, if so, determining that a benign cardiac rhythm is occurring, further wherein “correlates highly” indicates a relatively higher correlation than “correlates closely.”

11. An implantable cardiac stimulus device (ICSD) comprising a canister housing operational circuitry and having at least one canister electrode disposed thereon and being configured to couple electrically to a lead having at least one lead electrode thereon such that, when so coupled, the operational circuitry is electrically connected to the canister electrode and at least one lead electrode to form an implantable cardiac stimulus system (ICS) having a plurality of electrodes, wherein the operational circuitry is configured to perform a method of cardiac signal analysis comprising:

detecting electrical events occurring within the patient using the electrodes of the ICS system;

constructing event representations for the detected electrical events;

for a set of events N, N-1, N-2 and N-3, comparing the Nth event to each of the N-1, N-2 and N-3 events by correlation analysis;

determining, using the results of the correlation analysis, whether the set of events indicates one of the following: over-detection,

benign cardiac rhythm, or

likely cardiac arrhythmia; and

if likely cardiac arrhythmia is identified, determining whether stimulus is needed; and

if stimulus is needed, delivering electrical stimulus to the patient from using the electrodes of the ICS system;

wherein over-detection is identified if the correlation analysis indicates a pattern of correlations between the events N, N-1, N-2 and N-3 that corresponds to over-detection.

12. The ICSD of claim 11 wherein the operational circuitry is further configured such that the method includes determining that over-detection is occurring if the step of comparing the Nth event to each of the N-1, N-2 and N-3 events by correlation analysis results in a pattern of correlations of the following type: low correlation to N-1 and N-3 events, and high correlation to the N-2 event.

13. The ICSD of claim 11 wherein the operational circuitry is further configured such that the method further comprises: calculating a cardiac rate using the detected electrical events; and

if the set of events, N, N-1, N-2, N-3 indicates over-detection, correcting data relating to one or more over-detected events and recalculating the calculated cardiac rate.

14. The ICSD of claim 11 wherein the operational circuitry is further configured such that the step of determining whether the set of events indicates one of the following: over-detection, benign cardiac rhythm, or likely cardiac arrhythmia includes determining whether the Nth event correlates highly to at least one of the N-1, N-2 or N-3 events and, if so, determining that a benign cardiac rhythm is occurring.

15. The ICS of claim 14 wherein the operational circuitry is further configured such that the step of determining whether the set of events indicates one of the following: over-detection, benign cardiac rhythm, or likely cardiac arrhythmia includes determining whether the Nth event correlates closely to each of the N-1, N-2 or N-3 events and, if so, determining that a benign cardiac rhythm is occurring, further wherein “correlates highly” indicates a relatively higher correlation than “correlates closely.”

专利名称(译)	用于精确分类心脏活动的方法和装置		
公开(公告)号	<a href="#">US8880161</a>	公开(公告)日	2014-11-04
申请号	US14/058495	申请日	2013-10-21
[标]申请(专利权)人(译)	卡梅伦保健公司		
申请(专利权)人(译)	CAMERON HEALTH, INC.		
当前申请(专利权)人(译)	CAMERON HEALTH, INC.		
[标]发明人	WARREN JAY A SANGHERA RICK ALLAVATAM VENUGOPAL PALREDDY SUREKHA		
发明人	WARREN, JAY A. SANGHERA, RICK ALLAVATAM, VENUGOPAL PALREDDY, SUREKHA		
IPC分类号	A61B5/0452 A61B5/00 A61N1/365 A61N1/37		
CPC分类号	A61B5/04525 A61B5/7203 A61B5/7264 A61N1/37 A61N1/36592 A61B5/04012 A61N1/36514 A61N1/3621 A61B5/7221 A61B5/4836 A61B5/0464 A61B5/046 A61B5/0422 A61N1/3987 A61B5/0468 A61B5/0452 A61B5/726 G16H50/20		
优先权	61/051332 2008-05-07 US		
其他公开文献	US20140046396A1		
外部链接	<a href="#">Espacenet</a> <a href="#">USPTO</a>		

摘要(译)

用于植入式心脏监测和治疗设备（例如植入式心律转复除颤器）中的信号分析的方法，系统和设备。在一些示例中，分析包括检测到的事件的捕获数据以识别心脏事件的可能过度检测。在一些说明性示例中，当识别出过度检测时，可以修改数据以校正过度检测，减少过度检测的影响，或忽略过度检测的数据。几个例子强调使用与静态模板和/或事件间相关性分析相关的形态分析。

