




US 20200046234A1

(19) **United States**(12) **Patent Application Publication**
GLOVER et al.(10) **Pub. No.: US 2020/0046234 A1**(43) **Pub. Date: Feb. 13, 2020**(54) **SYSTEM AND APPARATUS COMPRISING A
MULTI-SENSOR CATHETER FOR RIGHT
HEART AND PULMONARY ARTERY
CATHETERIZATION***G01P 5/10* (2006.01)*A61B 5/0295* (2006.01)*A61M 25/10* (2006.01)*A61F 2/24* (2006.01)*A61B 5/028* (2006.01)(71) Applicant: **Three Rivers Cardiovascular Systems
Inc., Toronto (CA)**(52) **U.S. Cl.**CPC *A61B 5/02154* (2013.01); *A61B 5/024*
(2013.01); *A61B 5/01* (2013.01); *G01K 13/02*
(2013.01); *A61B 5/0215* (2013.01); *A61B*
5/743 (2013.01); *A61B 5/026* (2013.01); *G01K*
11/32 (2013.01); *A61B 5/02055* (2013.01);
G01P 5/10 (2013.01); *A61B 5/0261* (2013.01);
A61B 5/6852 (2013.01); *A61B 5/0295*
(2013.01); *A61B 5/7425* (2013.01); *A61M*
25/1025 (2013.01); *A61M 25/1018* (2013.01);
A61F 2/2427 (2013.01); *A61B 5/4836*
(2013.01); *A61B 5/028* (2013.01); *A61B*
5/6853 (2013.01); *A61B 5/6851* (2013.01);
A61B 5/02158 (2013.01); *G01L 11/02*
(2013.01)(72) Inventors: **Christopher GLOVER**, Toronto (CA);
Eric CARON, Toronto (CA); **Sylvain**
ABEL, Shawinigan (CA)(21) Appl. No.: **16/587,659**(22) Filed: **Sep. 30, 2019****Related U.S. Application Data**(63) Continuation of application No. 15/293,380, filed on
Oct. 14, 2016, now Pat. No. 10,463,259, which is a
continuation-in-part of application No. 14/874,604,
filed on Oct. 5, 2015, now Pat. No. 9,504,392, which
is a continuation of application No. 14/354,624, filed
on Apr. 28, 2014, now Pat. No. 9,149,230, filed as
application No. PCT/IB2012/055893 on Oct. 26,
2012.(60) Provisional application No. 61/552,778, filed on Oct.
28, 2011, provisional application No. 61/552,787,
filed on Oct. 28, 2011.**Publication Classification**(51) **Int. Cl.***A61B 5/0215* (2006.01)*G01L 11/02* (2006.01)*A61B 5/01* (2006.01)*G01K 13/02* (2006.01)*A61B 5/00* (2006.01)*A61B 5/026* (2006.01)*G01K 11/32* (2006.01)*A61B 5/0205* (2006.01)

(57)

ABSTRACT

A system and apparatus comprising a multi-sensor catheter for right heart and pulmonary artery catheterization is disclosed. The multi-sensor catheter comprises multi-lumen catheter tubing into which at least three optical pressure sensors, and their respective optical fibers, are inserted. The three optical pressure sensors are arranged within a distal end portion of the catheter, spaced apart lengthwise within the distal end portion for measuring pressure concurrently at each sensor location. The sensor locations are configured for placement of at least one sensor in each of the right atrium, the right ventricle and the pulmonary artery, for concurrent measurement of pressure at each sensor location. The sensor arrangement may further comprise an optical thermo-dilution sensor, and another lumen is provided for fluid injection for thermo-dilution measurements. The catheter may comprise an inflatable balloon tip and a guidewire lumen, and preferably has an outside diameter of 6 French or less.






FIG. 1 (Prior Art)

A-A CUT-THROUGH VIEW OF FIGS. 1 AND 3




FIG. 2 (Prior Art)




FIG. 3 (Prior Art)




FIG. 4 (Prior Art)




FIG. 5






FIG. 6

A-A CUT-THROUGH VIEW OF FIGS. 5, 6 AND 9



B-B CUT-THROUGH VIEW OF FIG. 6






FIG. 9




FIG. 10

A-A CUT-THROUGH VIEW OF FIG. 10




FIG. 11

B-B CUT-THROUGH VIEW OF FIG. 10




FIG. 12




FIG. 13A




FIG. 13B




FIG. 14

**SYSTEM AND APPARATUS COMPRISING A
MULTI-SENSOR CATHETER FOR RIGHT
HEART AND PULMONARY ARTERY
CATHETERIZATION**

**CROSS-REFERENCE TO RELATED
APPLICATIONS**

[0001] This application is continuation of U.S. patent application Ser. No. 15/293,380 filed Oct. 14, 2016, which is a Continuation-in-Part of U.S. patent application Ser. No. 14/874,604, filed Oct. 5, 2015, which is a Continuation of U.S. patent application Ser. No. 14/354,624, filed Apr. 28, 2014, which is a national stage entry of PCT International Application No. PCT/IB2012/055893, entitled “Apparatus, system and methods for measuring a blood pressure gradient”, filed Oct. 26, 2012, which claims priority from U.S. Provisional patent application No. 61/552,778 entitled “Apparatus, system and methods for measuring a blood pressure gradient”, filed Oct. 28, 2011 and from U.S. Provisional patent application No. 61/552,787 entitled “Fluid temperature and flow sensor apparatus and system for cardiovascular and other medical applications”, filed Oct. 28, 2011; all these applications are incorporated herein by reference, in their entirety.

[0002] This application is related to U.S. patent application Ser. No. 15/001,347, filed Jan. 20, 2016, entitled “System and Apparatus Comprising a Multisensor Guidewire for Use in Interventional Cardiology”, which is a Continuation-in-Part of PCT International Application No. PCT/IB2015/055240, of the same title, filed Jul. 10, 2015, designating the United States, and claiming priority from U.S. Provisional patent application No. 62/023,891, entitled “System And Apparatus Comprising a Multisensor Support Guidewire for Use in Trans-Catheter Heart Valve Therapies”, filed Jul. 13, 2014 and from U.S. Provisional patent application No. 62/039,952, entitled “System And Apparatus Comprising a Multisensor Support Guidewire for Use in Trans-Catheter Heart Valve Therapies”, filed Aug. 21, 2014; all said applications are incorporated herein by reference in their entirety.

TECHNICAL FIELD

[0003] The present invention relates to a system and apparatus comprising a multi-sensor catheter or a multi-sensor guidewire for use in cardiology, and more particularly to a pulmonary artery catheter for right heart catheterization and related diagnostic measurements.

BACKGROUND

[0004] The above referenced related patent applications disclose multi-sensor guidewires and multi-sensor micro-catheters for use in interventional cardiology. For example, if a heart valve is found to be malfunctioning because it is defective or diseased, minimally invasive methods are known for repair and replacement of the heart valve, by introduction of a catheter intravascularly into the heart to access the heart valve. Percutaneous procedures for minimally invasive transcatheter heart valve diagnosis, repair and replacement avoid the need for open heart surgery. These procedures may be referred to as Transcatheter Valve Therapies (TVT).

[0005] TVT for valve repair include, for example, procedures such as, balloon valvuloplasty to widen an aortic valve

which is narrowed by stenosis, or insertion of a mitral clip to reduce regurgitation when a mitral valve fails to close properly. Alternatively, if the valve cannot be repaired, a prosthetic replacement valve may be introduced. Minimally invasive Transcatheter heart Valve Replacement (TVR) procedures, including Transcatheter Aortic Valve Implantation (TAVI or TAVR) and Transcatheter Mitral Valve Implantation (TMVI), have been developed over the last decade and have become more common procedures in recent years.

[0006] While there have been many recent advances in systems and apparatus for TVT and for related diagnostic procedures, interventional cardiologists who perform these procedures have identified the need for improved apparatus for use in TVT, including apparatus for heart valve replacement. They are also seeking improved diagnostic equipment that provides real-time direct measurements, i.e. within the heart, of important hemodynamic cardiovascular parameters before, during and after TVT.

[0007] The above referenced related U.S. patent application Ser. Nos. 14/874,604 and 14/354,624 (now issued to U.S. Pat. No. 9,149,230) having common inventorship and ownership with the present application, disclose a multi-sensor microcatheter and a multi-sensor guidewire. These multi-sensor micro-catheters and guidewires comprise a distal end portion containing multiple optical sensors arranged for measuring blood pressure at several sensor locations, simultaneously, in real-time. Optionally, they include an optical or electrical sensor for measuring blood flow. The disclosed multi-sensor micro-catheters and multi-sensor guidewires can be configured for use in minimally invasive surgical procedures for measurement of intra-vascular pressure gradients, and more particularly, for direct measurement of a transvalvular pressure gradient within the heart, for any one of the four heart valves.

[0008] For example, a transvalvular measurement of pressure across the aortic valve, i.e. with pressure sensors positioned to measure pressure concurrently in the ascending aorta and left ventricle, allows for assessment of aortic regurgitation, before and after a TAVI procedure.

[0009] A need for improved diagnostic apparatus for right heart catheterization (RHC) and pulmonary artery (PA) catheterization has also been identified. For example, RHC may be performed in an Intensive Care Unit (ICU) for monitoring of critically ill patients. In a Cardiac Catheterization Lab (Cath Lab), RHC may be used for monitoring and diagnosis, and in the operating room (OR) for monitoring of important hemodynamic parameters during cardiac surgery or other high risk surgery.

[0010] During RHC and PA catheterization, a special balloon tipped catheter, which may be referred to as a pulmonary artery catheter (PA catheter) or a Swan Ganz (SG) catheter, is introduced through one of the larger veins, e.g. through an internal jugular vein, subclavian vein in the neck, or a median cubital vein in the arm, into the superior vena cava, or through a femoral vein into the inferior vena cava. The catheter tip is then introduced from the vena cava into the right atrium of the heart, advanced through the tricuspid valve into the right ventricle, and then through the pulmonic valve (alternatively called the pulmonary valve) into the PA, which is the main artery that carries de-oxygenated blood from the heart to the lungs. One lumen of the PA catheter extends from the proximal end to an opening at the distal tip. This lumen is fluid filled and is connected at its proximal end to an externally placed pressure trans-

ducer to enable the pressure at the distal tip to be monitored. Thus the pressure at the distal tip may be monitored by the catheter as it is advanced sequentially, firstly into the right atrium (RA), secondly into the right ventricle (RV) and thirdly into the PA. During these measurements the balloon may be partially inflated to allow the balloon to “float” and be drawn into the PA by the blood flow. Subsequently, after further inflating the balloon, the balloon is drawn by the blood flow and wedges in a smaller pulmonary blood vessel for measurement of a Pulmonary Capillary Wedge Pressure (PCWP). PCWP is an indirect measure of the left atrial pressure (LAP) and left ventricular end-diastolic pressure (LVEDP). At each point a characteristic pressure waveform is observed, and pressure measurements are recorded. That is, as the catheter tip is advanced, the observed waveform will change sequentially and show the transition from a RA pressure waveform, a RV pressure waveform, a PA pressure waveform and then a PCWP waveform.

[0011] Conventional four lumen/four port PA/SG catheters provide: one lumen for pressure monitoring at the catheter tip using an externally connected pressure transducer; a balloon inflation lumen; a thermistor lumen; and a fluid injection lumen for measurement of blood flow by thermo-dilution. The port for the pressure transducer may also be used for blood sampling, e.g. for measurement of mixed venous oxygen saturation (SvO₂). Some available PA catheters include two pressure sensing lumens, which can be connected to two external pressure transducers for measurement of pressures in the right atrium and in the pulmonary artery. Advanced PA catheters or SG catheters may also include several additional lumens and ports, e.g. another port for fluid infusion, and/or one or more ports for cable connections to other types of monitoring equipment, e.g. for measurement of cardiac output, oximetry, or insertion of a cardiac pacing wire.

[0012] By way of example only, the following references provide further background information and details of insertion techniques, characteristic pressure waveforms, and indications for RHC and PA catheterization:

[0013] 1) Jeremy Fernando, “Pulmonary Artery Catheters”, updated 22 Dec. 2014 (<http://lifeinthefastlane.com/ccc/pulmonary-artery-catheters/>);

[0014] 2) “Swan-Ganz—right heart catheterization”, NIH National Library of Medicine, online Medical Encyclopedia, Update Date Aug. 12, 2014 (<https://www.nlm.nih.gov/medlineplus/ency/article/003870.htm>);

[0015] 3) “Pulmonary artery catheter”, Wikipedia, version last modified 8 May 2016 (https://en.wikipedia.org/wiki/Pulmonary_artery_catheter);

[0016] 4) “Invasive Hemodynamic Monitoring: Physiological Principles and Clinical Applications” pp. 12-15; Jan M. Headley, RN, BS, CCRN, Edwards Life Sciences; Irvine, Calif., © 2002 (<http://ht.edwards.com/resourcegallery/products/swanganz/pdfs/invasivehdmphysprincbook.pdf>);

[0017] 5) B. Paunovic et al., “Pulmonary Artery Catheterization”, (c) 2011, updated 3 Jan. 2016 (<http://emedicine.medscape.com/article/1824547> (c) 2011);

[0018] 6) “Right Heart Catheterization”, Johns Hopkins Medicine Health Library, version downloaded August 2016 (http://www.hopkinsmedicine.org/healthlibrary/test_procedures/cardiovascular/right_heart_catheterization_135.40/);

[0019] 7) “Comparing FFR Tools New wires Pressure Microcatheter”, M. Kern, Cath Lab Digest, Volume 24, Issue 5, May 2016 (<http://www.cathlabdigest.com/article/Comparing-FFR-Tools-New-Wires-Pressure-Microcatheter>);

[0020] Conventional balloon tipped PA catheters that use a fluid filled catheter, which is coupled to an externally placed pressure transducer, are relatively inexpensive and durable. However, they measure pressure only at a single point, i.e. at the tip of the catheter. Thus, the cardiologist must reposition the catheter, i.e. by pushing and pulling the catheter tip back and forth to position tip to make pressure measurements at different locations within the heart. During this procedure, there is some risk that repeatedly advancing and pulling-back the catheter for repositioning tip of the catheter for pressure sensing will interfere with, or disrupt, normal operation of the heart, e.g. cause cardiac arrhythmias (such as atrial or ventricular fibrillation), interfere with opening and closing of the heart valves, or risk damage to the heart tissues.

[0021] Also, pressure sensing catheters have limited accuracy. Measurements may be affected by technical limitations such as reflection of the pressure wave at the tip and distortion if the catheter is kinked or sharply bent. Inertial artefacts and slow dynamic response (time lag, damping, hysteresis, resonances, frequency filtering) can distort the waveform, in time and amplitude, as it travels through the fluid filled lumen (de Vecchi et al., “Catheter induced Errors in Pressure Measurements in Vessels: An in-vitro and numerical study” IEEE Transaction on Biomedical Engineering, Vol. 61, No. 6, June 2014). Measurement errors as much as 20 mmHg have been reported (see ref. 2 in Robert G. Grey et al., “Feasibility of In Vivo Pressure Measurement using a pressure tip catheter via transventricular puncture”, ASAIO J. 2010 56(3) 194-199. This reference also compares limitations of a Pressure Tipped Catheter (PTC) using a piezo-electric pressure sensor and a conventional fluid filled pressure sensing catheter.

[0022] Limitations of pressure sensing catheters with externally placed transducers are also discussed in United States patent application publication no. US2011/004198, to Hoch, which discloses a central venous catheter (CVC) using a piezo-electric pressure sensor. However, electrical pressure sensors of this type have some drawbacks for in vivo applications, where long thin electrical wires are carrying small electrical signals in humid environment, e.g. requirement for electrical isolation of electrical components, significant electrical drift and temperature sensitivity and electrical interference, such as, cross-talk between wires from multiple electrical sensors and from external electromagnetic sources within the operating room.

[0023] Thus, there is a need for improved or alternative PA catheters for direct measurements of cardiovascular parameters, including blood pressure measurements, during RHC and PA catheterization procedures.

[0024] An object of the present invention is to provide for improvements or alternatives to known systems and apparatus comprising multi-sensor catheters or multi-sensor guidewires.

SUMMARY OF INVENTION

[0025] The present invention seeks to mitigate one or more disadvantages of known systems and apparatus comprising multi-sensor catheters or multi-sensor guidewires for

measuring cardiovascular parameters, and in particular to provide systems and apparatus with particular applicability for measurements of hemodynamic parameters during right heart catheterization (RHC) and PA catheterization.

[0026] In one aspect, the present invention seeks to provide a multi-sensor catheter with particular applicability for measurements of hemodynamic parameters during right heart and pulmonary artery catheterization.

[0027] A first aspect of the invention provides a multi-sensor catheter for right heart and pulmonary artery catheterization, comprising:

a length of multi-lumen catheter tubing extending between a proximal end and a distal end, having an outside diameter of ≤ 8 French, and the distal end comprising an atraumatic tip;

a plurality of optical sensors and a plurality of optical fibers; a sensor end of each optical fiber being attached and optically coupled to an individual one of the plurality of optical sensors;

each optical sensor and its optical fiber being inserted into a respective lumen of the multi-lumen catheter tubing, the sensors being spaced apart lengthwise to provide a sensor arrangement with said plurality of optical sensors positioned at respective sensor locations spaced apart lengthwise within a distal end portion of the catheter tubing;

a proximal end of each of the plurality of optical fibers being coupled to an optical input/output connector at the proximal end of the catheter for connection to an optical control system; and

the plurality of optical sensors of the sensor arrangement comprising at least three optical pressure sensors at respective sensor locations spaced apart lengthwise along said length of said distal end portion, with an aperture in the catheter tubing adjacent each optical pressure sensor for fluid contact; and

wherein the pressure sensor locations and spacings are configured to place at least one pressure sensor in each of the right atrium, the right ventricle and the pulmonary artery during right heart catheterization, for concurrent blood pressure measurements at each pressure sensor location.

[0028] The multi-sensor catheter may further comprise another lumen having a proximal port and an opening at the distal tip of the catheter tubing, said lumen having a diameter which accepts a guidewire, e.g. a standard 0.025 inch (0.625 mm) guidewire, to enable the catheter to be inserted over the guidewire. For example, to facilitate insertion into the PA, the guidewire is first steered into the RV and PA, and then the catheter is introduced over the guidewire. The latter facilitates introduction of the PA catheter by a more tortuous route, for example when it is introduced from the inferior vena cava.

[0029] The multi-sensor catheter may further comprise an inflatable balloon near the distal tip, the catheter tubing further comprising a balloon inflation lumen, the inflatable balloon being coupled by the balloon inflation lumen to a balloon inflation port at the proximal end of the catheter tubing. An inflatable balloon tip allows for introduction of the tip of the catheter to be flow directed, i.e. the balloon is floated and guided by blood flow into the PA, and for measurement of the PA pressure waveform and PCWP waveform, similar to a conventional PA catheter.

[0030] When the multi-sensor catheter comprises a balloon tip and a guidewire lumen, a guidewire may be used to facilitate insertion and/or the balloon may be inflated for

flow directed insertion. The guidewire lumen is preferably a central lumen. This lumen may also be used for blood sampling at the tip of the catheter, e.g. for measurement of SvO₂ or measurement of cardiac output by the Fick method.

[0031] In some embodiments, for direct measurement of blood flow within the PA, the plurality of optical sensors further comprises an optical flow sensor. For example, the flow sensor comprises an optical thermo-dilution or an optical thermo-convection sensor positioned proximal to the distal tip for placement in the PA and the catheter tubing further comprises a fluid injection lumen having an aperture positioned for fluid injection into the right atrium and the fluid injection lumen being coupled at the proximal end of the multi-lumen catheter to a fluid injection port.

[0032] The multi-lumen catheter tubing may, for example, have an outside diameter of between 4 and 7 French and more preferably has an outside diameter of 6 French (2.000 mm) or less. The optical pressure sensors are preferably Fabry-Perot (FP) Micro-Opto-Mechanical System (MOMS) sensors. These optical pressure sensors comprise, for example, standard optical fibers of 0.155 mm diameter and FP MOMS pressure sensors of 0.260 mm diameter at the sensor end of the fiber for sensing pressure. When the catheter has a guidewire lumen which accepts a guidewire, e.g. a standard guidewire of up to 0.025 inch (0.625 mm) diameter, and the sensor arrangement comprises three optical pressure sensors and an optical flow sensor, and a fluid injection lumen, this four optical sensor arrangement can be accommodated within a multi-lumen catheter having an outside diameter of the multi-lumen tubing of approximately 6 French. The guidewire lumen is preferably a central lumen of the multi-lumen catheter, and the other lumens are arranged symmetrically around the central lumen. With smaller diameter optical sensors and fibers, and/or for use with a smaller diameter guidewire, smaller diameter multi-lumen catheter tubing may be used. PA catheters for insertion through one of the larger veins in the neck or groin are typically in the range from 4 to 8 French, with 6 or 7 French being commonly used. Smaller gauge catheters may be preferred for convenience or patient comfort, e.g. to allow for insertion through one of the veins in the upper or lower arm, such as the median cubital vein. Smaller diameter catheters may be required for pediatric use or neonatal use.

[0033] In an embodiment, the proximal end of the multi-lumen catheter comprises a connection hub, through which each lumen of the multi-lumen catheter is connected to a "tail" comprising an individual length of flexible tubing and respective individual proximal ports. The proximal ports for each optical sensor comprising said optical input/output connector, and the optical fibers for each optical sensor extending through from the optical sensor through its respective individual lumen of the multi-lumen catheter tubing, through the hub and flexible tubing to the respective optical input/output connector for connection to the optical control system.

[0034] Optionally, the catheter comprises radiopaque markers at intervals along the length of the catheter to facilitate location of each sensor in use, e.g. by conventional radio-imaging techniques. For example, markers may be placed at regular intervals along the length, or markers are placed near each sensor. A radiopaque marker may also be placed at the tip or near the balloon.

[0035] The catheter may comprise one or more additional lumens with proximal ports, which are provided for fluid

injection or fluid infusion, with respective distal apertures located for injection or infusion of fluid the right atrium during use. One or more additional lumens and respective proximal ports may optionally be provided for other purposes, if required.

[0036] Embodiments of the present invention provide improved or alternative guidewire directed catheters and flow directed balloon catheters that enable direct measurements of cardiovascular parameters, including measurement of pressure concurrently within the RA, RV, and PA, during RHC and PA catheterization procedures.

[0037] Another aspect of the invention provides a multi-sensor catheter for right heart and pulmonary artery catheterization in pediatric or neonatal patients, comprising:

a length of multi-lumen catheter tubing extending between a proximal end and a distal end, having an outside diameter of ≤ 3 French, and the distal end comprising an atraumatic tip;

a plurality of optical sensors and a plurality of optical fibers; a sensor end of each optical fiber being attached and optically coupled to an individual one of the plurality of optical sensors;

each optical sensor and its optical fiber being inserted into a respective lumen of the multi-lumen catheter tubing, the sensors being spaced apart lengthwise to provide a sensor arrangement with said plurality of optical sensors positioned at respective sensor locations spaced apart lengthwise within a distal end portion of the catheter tubing;

a proximal end of each of the plurality of optical fibers being coupled to an optical input/output connector at the proximal end of the catheter for connection to an optical control system; and

the plurality of optical sensors of the sensor arrangement comprising at least two optical pressure sensors at respective sensor locations spaced apart lengthwise along said length of said distal end portion, with an aperture in the catheter tubing adjacent each optical pressure sensor for fluid contact; and

wherein the pressure sensor locations and spacings are configured to measure concurrent pressure waveforms for at least two locations within the heart and pulmonary artery comprising a right atrial pressure waveform, a right ventricular waveform, and a pulmonary artery waveform, a pulmonary capillary wedge position waveform, to provide concurrent blood pressure measurements at each optical pressure sensor location.

[0038] A further aspect of the invention comprises a control system for multi-sensor catheters and multi-sensor guidewires, such as, the multi-sensor PA catheters disclosed herein.

[0039] The control system comprises a light source and detector, and an optical interface for coupling, via respective input/output ports, to each of the optical fibers and optical sensors of a multi-sensor catheter or multi-sensor guidewire; data storage and processing means configured for processing optical data indicative of pressure values and optionally, optical or electrical data indicative of flow velocity values; and wherein, for right heart catheterization and pulmonary artery catheterization, the processing means is further configured for graphically displaying pressure data comprising a plurality of concurrent blood pressure waveforms. Thus, when the multi-sensor catheter comprises at least three optical pressure sensors, the system can process optical data

from each sensor and display concurrent blood pressure waveforms from the right atrium, right ventricle and pulmonary artery.

[0040] The concurrent blood pressure waveforms for each optical sensor may be displayed together for comparison, or displayed individually, for one or more time intervals, and during one or more cardiac cycles. Optionally, graphical flow velocity data may also be displayed concurrently. Advantageously, the processing means is further configured to derive and display hemodynamic parameters from the blood pressure data and flow velocity data. For example, during right heart and pulmonary artery catheterization with a multi-sensor catheter as disclosed herein comprising at least three optical pressure sensors, in addition to displaying blood pressure waveforms from the right atrium, right ventricle and pulmonary artery, a plurality of numeric values such as peak pressures, mean pressures, peak-to-peak pressure differentials for each curve, and pressure differentials or gradients between the right atrium and right ventricle, and between the right ventricle and pulmonary artery can be displayed in real time.

[0041] Correspondingly, in another exemplary embodiment, in use of a multi-sensor guidewire during TAVR, e.g. to assess functioning of the aortic valve, before and after a TAVR procedure, the system may display numeric values based on pressure waveforms from the aorta and the left ventricle. For example, in addition to displaying pressure waveforms from the aorta and the left ventricle, the system may display a plurality of numeric values such as peak pressures, mean pressures, peak-to-peak pressure differentials for each curve, and pressure differentials or gradients between the aorta and the left ventricle. The system may also compute a parameter such as an aortic regurgitation index (ARi), and display the ARi value in real time.

[0042] Accordingly, another aspect of the invention provides a computer program product embodied as a non-transitory computer readable medium storing instructions, for execution in a processor of a control system for a multi-sensor catheter or a multi-sensor guidewire, for processing optical data received concurrently from a plurality of optical sensors of the multi-sensor catheter or a multi-sensor guidewire, said optical data being indicative of blood pressure, and displaying a corresponding plurality of blood pressure waveforms, and optionally flow velocity data, and displaying numeric data relating to selected hemodynamic parameters and indexes.

[0043] Thus, apparatus and systems comprising a multi-sensor PA catheter are provided that mitigate one or more problems with known systems and apparatus for RHC and PA catheterization, which allow for diagnostic measurements and monitoring of hemodynamic parameters, including measurement of blood pressure concurrently at multiple locations within the right heart and PA.

[0044] The foregoing and other objects, features, aspects and advantages of the present invention will become more apparent from the following detailed description, taken in conjunction with the accompanying drawings, of embodiments of the invention, which description is by way of example only.

BRIEF DESCRIPTION OF THE DRAWINGS

[0045] In the drawings, identical or corresponding elements in the different Figures have the same reference numeral.

[0046] FIG. 1 (Prior Art) shows a schematic longitudinal cross-sectional view of a known type of conventional PA catheter, which may be referred to as a Swan Ganz catheter;

[0047] FIG. 2 (Prior Art) shows an enlarged transverse cross-sectional view of the PA catheter illustrated in FIG. 1 taken through plane A-A of FIG. 1 to show the lumens of the multi-lumen catheter tubing;

[0048] FIG. 3 (Prior Art) shows schematic partial cross-sectional view of a human heart to illustrate placement of a conventional PA catheter within the right heart and the PA for measurement of pressure in the PA and for measurement of blood flow by thermo-dilution;

[0049] FIG. 4 (Prior Art) shows schematic diagrams of a human heart to illustrate schematically placement of the conventional PA catheter within the right heart and PA for pressure measurements, at each the following positions: A. in the right atrium (RA); B. in the right ventricle (RV); C. in the pulmonary artery (PA); and D. in a pulmonary capillary wedge position (PCWP); the underlying plots show examples of a sequence of typical blood pressure waveform for each of positions A, B, C and D;

[0050] FIG. 5 illustrates schematically a system according to a first embodiment, comprising an apparatus for right heart and PA catheterization comprising a multi-sensor PA catheter, which is optically coupled to a control unit;

[0051] FIG. 6 shows a schematic longitudinal cross-sectional view of an apparatus for right heart and PA catheterization comprising a multi-sensor PA catheter according to the first embodiment of the present invention;

[0052] FIG. 7 shows an enlarged axial cross-sectional view of the multi-lumen catheter illustrated in FIG. 6 taken through plane A-A of FIG. 6;

[0053] FIG. 8 shows an enlarged axial cross-sectional view of the multi-lumen catheter illustrated in FIG. 6 taken through plane B-B of FIG. 6;

[0054] FIG. 9 shows a schematic partial cross-sectional diagram of a human heart to illustrate placement of the multi-sensor PA catheter of the first embodiment within the right heart and PA for diagnostic measurements of hemodynamic parameters, including concurrent measurements of blood pressure in the RA, in the RV, in the PA and the PCWP;

[0055] FIG. 10 shows a schematic longitudinal cross-sectional view of an apparatus for right heart and PA catheterization comprising a multi-sensor PA catheter according to the second embodiment of the present invention;

[0056] FIG. 11 shows an enlarged axial cross-sectional view of the multi-lumen catheter illustrated in FIG. 10 taken through plane A-A of FIG. 10; and

[0057] FIG. 12 shows an enlarged axial cross-sectional view of the multi-lumen catheter illustrated in FIG. 10 taken through plane B-B of FIG. 10.

[0058] FIG. 13A illustrates schematically a graphical display of the control system showing three concurrent blood pressure waveforms from: A. sensor position P1 in the pulmonary artery (balloon deflated); B. sensor position P2 in the right ventricle; and C. sensor position P3 in the right atrium; together with selected numeric data comprising hemodynamic parameters derived from the pressure waveforms;

[0059] FIG. 13B illustrates schematically a graphical display of the control system showing three concurrent blood pressure waveforms from: D. sensor position P1 in the

pulmonary capillary wedge position with the balloon inflated; B. sensor position P2 in the right ventricle; and C. sensor position P3 in the right atrium; together with selected numeric data comprising hemodynamic parameters derived from the pressure waveforms; and

[0060] FIG. 14 illustrates schematically a graphical display of the control system showing concurrent pressure waveforms from: A. sensor position P1 in the aorta and B. sensor position P2 in the right ventricle, with numeric data comprising hemodynamic parameters derived from the pressure waveforms, including an aortic regurgitation index (AR Index).

DETAILED DESCRIPTION OF EMBODIMENTS

[0061] As illustrated schematically in the longitudinal cross-sectional view shown in FIG. 1, a conventional PA catheter 1000, which may be referred to as a Swan Ganz catheter, comprises a length of multi-lumen catheter tubing 1002 having at its distal end 1120 an inflatable balloon 1160. The catheter typically comprises a multi-lumen catheter tubing 1002, such as illustrated in FIG. 2, which is a cross-sectional view through plane A-A of FIG. 1. By way of example, the catheter tubing shown in FIG. 2 has four lumens 1004-1, 1004-2, 1004-3 and 1004-4, and an outside diameter of 5 French. Referring back to FIG. 1, there is a hub 1006 at the proximal end of the catheter with individual proximal ports 1010-1, 1010-2, 1010-3, and 1010-4. Ports 1010-1, 1010-2, 1010-3 are coupled by individual lengths of flexible tubing 1007 through the hub 1006 to respective lumens 1004-1, 1004-2 and 1004-3 of the catheter tubing 1002. Port 1010-4 is an electrical connector for a thermistor 20 which is located within the fourth lumen 1004-4 proximal to the balloon tip 1160, towards the distal end. Electrical connections for the thermistor 20 extend through lumen 1004-4 from the thermistor 20 to an electrical cable 1009 extending from the hub 1006. For simplicity of illustration of the longitudinal cross-sectional view shown in FIG. 1, the inner walls of the lumens are not shown.

[0062] Traditionally, the multi-lumen catheter tubing 1002 of a Swan Ganz catheter is colored yellow and each of the ports 1010-1 to 1010-4 is color coded. A first lumen 1004-1 provides for inflating the balloon and has a corresponding proximal port 1010-1 for coupling to an air filled syringe for inflating and deflating the balloon. The balloon inflation port is conventionally colored red. The balloon 1160 typically has a volume of 0.5 ml to 1.5 ml and is connected to the balloon inflation lumen 1004-1. A second lumen 1004-2 is has an aperture 1012-2 opening at the distal tip 1120 and is connected at the proximal end to a proximal port 1010-2 (conventionally colored yellow) for connection to an externally placed pressure transducer, so that, when this lumen is filled with fluid, the blood pressure at the tip 1120 can be sensed. This port and lumen may also be used for sampling of blood at the tip of the catheter. For measurement of flow by thermo-dilution using the thermistor 20, there is third lumen 1004-3, which has a proximal injectate port (conventionally coloured blue) to allow for injection of a bolus of cold thermo-dilution fluid; this lumen has an injectate opening 1012-3 a distance of approximately 30 cm from the distal tip 1120. The fourth lumen 1004-4 accommodates the thermistor 20, i.e. an electrical temperature sensor, which is typically located at a distance of about 4 cm from the distal tip 1120; the electrical wires (not shown in FIG. 1) for the thermistor 20 extend through lumen 1004-4 to a proximal

port **1010-4** (conventionally colored yellow) which comprises an electrical connector for the thermistor **T**. In this example, the second lumen **1004-2**, which has an opening **1012-2** at the distal tip **1120** and respective proximal port **1010-2** has a larger diameter to allow for insertion of a guidewire, e.g. a standard 0.025 inch or 0.018 inch guidewire, to assist with introduction of the PA catheter into the right heart.

[0063] FIG. 3 illustrates schematically a partial cross-sectional view of a human heart **500** with a PA catheter **1001** positioned within the right heart and left branch of the PA. The PA catheter **1001** is similar to that shown in FIGS. 1 and 2 comprising four lumens, i.e. one for balloon inflation, one for pressure sensing, one for the thermistor and one for fluid injection for measurement of flow by thermo-dilution. Similar to the catheter shown in FIG. 1, the four lumens are connected through a hub **1006** to proximal ports **1010-1**, **1010-2**, **1010-3** and **1010-4**. The first lumen **1004-1** opens to the inflatable balloon **1160** and the first lumen is connected to the balloon inflation/deflation port **1010-1** for connection to an air filled syringe; the second lumen has an opening **1012-2** at the distal tip **1120** and a distal port **1010-2**. The distal port **1010-2** that is used for the guidewire insertion can also be used either for connection to an external pressure transducer for measurement of blood pressure at the distal tip **1120**, or, for blood sampling, such as for measuring mixed venous oxygen saturation (SvO₂). The third port **1010-3** is the proximal injection port which is connected to the third lumen **1004-3** (FIG. 2) for fluid injection into the right atrium **521** through opening **1012-3**. The fourth port **1010-4** comprises an electrical connector for the thermistor **20** which is located in the fourth lumen **1004-4** near the distal tip **1120**.

[0064] To position the PA catheter **1001** as illustrated schematically in FIG. 3, the catheter tip **1120** is introduced into the right atrium (RA) **521**, e.g. in this example upwards through the inferior vena cava **520**, and through the atrio-ventricular (tricuspid) valve **522** and into the right ventricle (RV) **523**. During introduction, partial inflation of the balloon **1160** allows for blood flow directed insertion. That is, with partial inflation of the balloon **1160** in the right ventricle **523**, the tip **1120** of the catheter tends to be drawn by the blood flow through the pulmonic valve **524** into the PA **525** and then from the PA towards the right or left branches **526L** and **526R** of the PA, and towards smaller pulmonary vessels which lead to the lungs. Pressure sensing at the tip is achieved by the fluid filled lumen **1004-2** which has opening **1012-2** at the distal tip **1120**, and a proximal port **1010-2** which is connected to the externally placed pressure transducer for monitoring pressure at the catheter tip **1120**. The port **1010-4** provides for electrical connections for the thermistor **20** (T), which located about 4 cm proximally from the balloon tip **1160**, for measurement of blood flow by thermo-dilution. As illustrated schematically in FIG. 3, the opening **1012-3** of the fluid injection lumen is positioned to allow for injection of a bolus of cold fluid into the right atrium of the heart, for detection of a temperature change by the thermistor **20** (T) when the cold fluid reaches the PA **525** and its left and right branches **526L** and **526R**, to measure blood flow by a conventional thermo-dilution technique.

[0065] In use of the PA catheter **1001**, as illustrated in FIG. 4 (Prior Art), when the tip **1120** of the catheter **1001** is first introduced into the heart **500**, e.g., through the superior vena cava as illustrated, and into the RA **521** (position A), a

pressure measurement is made in the RA to assess the RA waveform. Then the tip of the catheter is advanced through the tricuspid valve **522** into the RV **523** (position B), and a pressure measurement is made within the RV to assess the RV waveform. The balloon is partially inflated so that the blood flow draws the balloon from the right ventricle **523**, through the pulmonic valve **524** into the PA **525**. When the tip of the catheter is in the PA (position C), the balloon is deflated, and another pressure measurement is made to assess the PA waveform. The tip of the catheter is then allowed to be drawn further into the PA and the balloon at the tip of the catheter is further inflated. When the inflated balloon wedges in a smaller branch **527** of the pulmonary vessels (position D) another pressure measurement is made, called the PA occlusion pressure, or pulmonary capillary “wedge” pressure (PCWP) waveform. PCWP provides an indirect measure of the left atrial pressure (LAP) and left ventricular end-diastolic pressure (LVEDP). As illustrated by the blood pressure waveform in the low part of FIG. 4, as the catheter tip is advanced, the observed waveform changes sequentially and show the transition from a RA pressure waveform, a RV pressure waveform, a PA pressure waveform and then a PCWP waveform. The cardiac output (CO), which is the amount of blood that the heart pumps per minute, may also be determined during a right-heart and PA catheterization. Pressure measurements may be made before and after administration of intravenous (IV) heart medications. The catheter may need to be repositioned several times to allow for several pressure measurements to be made at each of the different locations A, B, C and D within the heart and PA.

[0066] Limitations of conventional PA catheters of this type include:

[0067] Pressure measurements are made by a fluid filled pressure sensing catheter

[0068] Pressure is transmitted through fluid filled lumen to a remote pressure transducer; pressure measurements are not always accurate.

[0069] Pressure is measured at the distal tip only, so pressure can be measured only at one point or location at a time.

[0070] Pressure measurements are sensitive to relative positioning and re-positioning of the catheter and pressure transducer, e.g. raising or lowering it relative to the heart in order to set the zero pressure (0 cm H₂O) on the transducer.

[0071] Kinking or bending of the catheter may dampen the characteristic waveforms seen at each position;

[0072] There is time lag (and hysteresis) between pressure being applied to the opening at the tip of the pressure sensing lumen and transmission of pressure through the fluid filled lumen to the remote transducer.

[0073] Pullback and repositioning of the catheter for multiple measurements may cause cardiac arrhythmias.

[0074] A system and apparatus comprising a multi-sensor catheter for use in cardiology, which may include diagnostic measurements of cardiovascular parameters during right heart and PA catheterization, according to an embodiment of the present invention will be illustrated and described, by way of example, with reference to a system **2000** comprising a multi-sensor PA catheter **2001**, illustrated schematically in FIGS. 5 to 9.

[0075] Firstly, referring to FIG. 5, this schematic diagram represents the system **2000** comprising an apparatus **2001**

comprising a multi-sensor catheter for right heart and PA catheterization procedures, coupled to a control system **2150**, which comprises a control unit **2151** and user interface, such as the illustrated touch screen display **2152**. The multi-sensor catheter **2001** comprises some components of a conventional PA catheter, including multi-lumen catheter tubing **2002**. In the multi-sensor catheter of this embodiment, the catheter tubing comprises seven lumens **2004-1** to **2004-7**, as illustrated in the transverse cross-sectional view in FIG. 7, which is a cross-section through A-A of FIGS. 5, 6 and 9. Typically, the catheter tubing **2002** has an outside diameter in the range of about 3-8 French (see Table 1 below for equivalent dimensions in mm and inches), and extends to an atraumatic flexible tip **2120** comprising an inflatable balloon **2160**, as shown schematically in FIGS. 5 and 6.

TABLE 1

French (Gauge)	Diameter (mm)	Diameter (inches)
2	0.667	0.027
3	1.000	0.039
4	1.333	0.053
5	1.667	0.066
6	2.000	0.079
7	2.333	0.092
8	2.667	0.105

[0076] By way of example, the catheter tubing may typically be about 110 cm in length from the distal tip **2120** to the proximal end, which comprises a connection hub **2006**. This length is suitable for introduction of the catheter into the right heart and PA through the superior vena cava (e.g. reached through the subclavian vein or interior jugular vein in the neck, or through the median cubital vein in the arm) or the inferior vena cava (e.g. reached through a femoral vein). For some applications, the catheter length may be shorter, e.g. 60 cm, or longer than 110 cm.

[0077] The PA catheter **2001** differs from a conventional PA catheter, in that, internally, as illustrated schematically in the longitudinal cross-sectional view in FIG. 6, it also contains a multi-sensor arrangement comprising a plurality of optical pressure sensors **10** and an optical temperature sensor **20**, and respective optical fibers **2011**. The optical sensors **10** and **20** are not externally visible in FIG. 5, so the sensor positions are indicated schematically by P1, P2, P3 for the optical pressure sensors and T for the optical temperature sensor. The sensor positions P1, P2, P3 and T are located along a length L distal end portion **2130**, near the distal tip **2120**. As illustrated schematically in FIG. 5, the connection hub **2006** provides connection ports **2010-n**, where n=1 . . . 7, for each of the seven lumens. Four of the connection ports, **2010-2**, **2010-3**, **2010-4**, **2010-5** are optical connectors which are coupled via flexible optical connections **2008** through the hub **2006** to the optical fibers and sensors within their respective lumens, i.e. lumens **2004-2**, **2004-3**, **2004-4**, **2004-5** shown in the transverse cross section in FIG. 7. The optical connectors each for the four optical sensors (P1, P2, P3 and T) plug into corresponding optical ports **2153** of the optical controller **2151** shown in FIG. 5.

[0078] Referring to the schematic longitudinal cross-sectional view shown in FIG. 6, and the transverse cross-sectional view shown in FIG. 7, for simplification of illustration, the internal walls of the lumens **2004-1** to **2004-7** are

omitted from FIG. 6 and only the optical fibers **2011** and sensors **10** and **20** are shown within the catheter tubing; elements shown in FIG. 6 are not drawn to scale. Thus, in description of FIG. 6, references to the lumens **2004-1** to **2004-7** are made based on FIG. 7. Also, by way of example, FIG. 7 shows typical dimensions for catheter tubing comprising seven lumens and having an outside diameter of 6 French (0.079 inch/2.00 mm), and for optical fibers **2011** having a diameter of 0.155 mm.

[0079] As is conventional, the PA catheter **2001** has an inflatable balloon **2160** connected to a balloon inflation lumen **2004-1** which is coupled through the hub **2006** and flexible tubing **2007**, to a balloon inflation/deflation port **2010-1**. Another lumen **2004-6** provides for fluid injection or infusion through a fluid injection/infusion port **2010-6** for injection of fluid through an aperture **2012-6** located close to the sensor location P3. A central lumen **2004-7**, which has an opening **2012-7** at the distal tip **2120**, has an internal diameter which is sized to receive a standard guidewire, such as a 0.025 inch guidewire, to allow for over-the-guidewire directed insertion of the PA catheter. Thus, of the seven ports **2010-n** (n=1 to 7), four of those ports **2010-2**, **2010-3**, **2010-4**, **2010-5** comprise standard optical fiber connectors and the other three ports **2010-1**, **2010-6** and **2010-7** are standard ports, such as luer fittings, i.e. for attachment of an air filled syringe for balloon inflation/deflation, for fluid injection or for guidewire insertion.

[0080] The positioning of the optical sensors **10**, **20** within the catheter tubing **2002** is illustrated in more detail in the schematic longitudinal cross-sectional view shown in FIG. 6. There are three optical sensors **10**, at sensor locations P1, P2 and P3, for measuring pressure and one optical sensor **20**, at sensor location T, for measuring temperature. Each optical sensor **10** and **20** is optically coupled to a respective individual optical fiber **2011**. That is, each optical sensor is integral with a sensor end of the fiber, or is bonded to the optical fiber, to provide an optical coupling of the sensor and fiber. Each optical fiber, carrying its sensor at the end, extends through its own individual lumen **2004-2**, **2004-3**, **2004-4**, **2004-5** of the multi-lumen catheter, as illustrated in the transverse cross-sectional view in FIGS. 7 and 8, which are taken through plane A-A and B-B, respectively, of FIG. 6. The central lumen **2004-7** is sized to accept a guidewire, such as a standard 0.025 inch guidewire to facilitate insertion of the catheter. The other six lumens are arranged symmetrically around the central lumen. A first lumen **2004-1** and its port **2010-1** provides for balloon inflation, as is conventional, with a distal opening (**2012-1** not visible) coupled to the balloon. The proximal port **2010-1** is coupled via the hub **2006** to lumen **2004-1** via a length of flexible tubing **2007**, and the port **2010-1** is typically a luer type fitting for connection to an air filled syringe (**2100** in FIG. 5) for inflation and deflation of the balloon **2160**. Second, third and fourth lumens, i.e. lumens **2004-2**, **2004-3**, **2004-4**, each accommodate one of the optical pressure sensors **10** and its optical fiber **2011**. Each of these lumens have a respective distal aperture **2012-2**, **2012-3**, **2012-4**, adjacent to the respective sensor locations P1, P2, P3 for fluid contact. A fifth lumen **2004-5** accommodates the optical temperature sensor **20** and its optical fiber **2011**. At the proximal end, the optical fibers **2011** for each sensor extend through a length of flexible tubing **2008** from the hub **2006** to the respective optical connector **2010-2**, **2010-3**, **2010-4**, **2010-5**. Thus the fibers are protected within the flexible

tubing **2008** to provide a flexible optical connection of a desired length for connection to the control system. The sixth lumen, **2004-6** is provided for fluid injection of a bolus of cold fluid, to allow for flow to be determined by thermo-dilution by detection of temperature changes by optical temperature sensor **20**. This fluid injection lumen **2004-6** has a distal opening **2012-6** positioned for delivering fluid to the right atrium, i.e. approximately 30 cm from the distal tip **2120**, close to the sensor position **P3**. When the sixth lumen is not being used for injectate for thermo-dilution measurements, it may alternatively be used for fluid infusion or injection for other purposes. As mentioned above, the central or seventh lumen **2004-7** is provided for a guidewire, if guidewire assisted insertion is required. After withdrawal of the guidewire, optionally this lumen may be used for other purposes, such as blood sampling at the tip of the PA catheter, e.g. for mixed venous oxygen saturation (SvO₂) measurements, or for measurement of cardiac output by the method of Fick.

[0081] As illustrated schematically in FIG. 6, the three optical pressure sensors **10** at sensor positions **P1**, **P2**, **P3**, and the optical temperature sensor **20** at position **T**, are provided along a length of the distal end portion **2130** spaced by distances L_1 , L_2 and L_3 from sensor position **P1**. **P1** is located at or close to the distal tip **2120**, distal to the balloon **2160**. **P2** and **P3** are arranged spaced apart lengthwise so that, in use, they can be located, respectively in the RV and the RA, when **P1** is positioned in the PA. In use, corresponding apertures **2012-2**, **2012-3**, **2012-4** near each optical pressure sensor **10** allow for fluid contact with the optical pressure sensors. By way of example, **P2** is located at a distance L_2 from **P1**, typically L_2 is about 20 cm; and **P3** is located at a distance L_3 from **P1**, typically L_3 is about 30 cm. That is, the distance between **P1** and **P2**, is about 20 cm and the distance between **P2** and **P3**, is about 10 cm. The location of the optical temperature sensor **T** is proximal to the balloon and at a distance L_1 from **P1**, where L_1 is approximately 4 cm, so that the temperature sensor **T** is located in the PA when the pressure sensors locations **P1**, **P2** and **P3** are positioned, respectively, in the PA, the RV and the RA.

[0082] The transverse cross-sectional view shown in FIG. 7 through plane A-A of FIG. 6, illustrates the arrangement of the seven lumens of the catheter tubing of the catheter **2001** of this embodiment. There is a central lumen **2004-7** for a guidewire, e.g. of a suitable diameter to receive a standard 0.025 inch guidewire. As mentioned above, the central lumen may also be used for fluid injection or infusion, or for blood sampling. The other six lumens **2004-1** to **2004-6** are arranged symmetrically around the central guidewire lumen **2004-7**. As mentioned above, three of the lumens accommodate the three pressure sensors **10** at locations **P1**, **P2** and **P3** and their optical fibers **2011**. One of the lumens accommodates the temperature sensor **20** at location **T** and its optical fiber **2011**. One lumen, for example **2004-1**, is provided for balloon inflation. The other lumen, for example **2004-6**, is provided for fluid injection or infusion. Thus, as shown schematically in the cross-sectional view in FIG. 8, through B-B of FIG. 6, an aperture or orifice **2012-4** is provided adjacent sensor **P3** for fluid contact. Correspondingly, as shown in longitudinal cross-sectional view in FIG. 6, an aperture is also provided near each of the other pressure sensors **10** for fluid contact. The lumen for fluid injection has an opening **2012-6** that is located proximal to the aperture **2012-4** for **P3** by a short distance, e.g. by 1 cm, or a distance

L_4 , of about 31 cm from **P1** at the distal tip **2120**. The spacings of the three pressure sensors **10** are selected to allow sensor locations **P2** to be placed in the RV and sensor location **P3** to be placed in the RA when sensor **P1** is in the PA, and so that the temperature sensor location **T** is positioned in the PA for measurement of blood flow in the PA by thermo-dilution.

[0083] Referring back to FIG. 5, the proximal part of the apparatus **2001** provides for optical coupling of the proximal end **2102** to the control unit **2151**. Four of the proximal ports **2010-2**, **2010-3**, **2010-4** and **2010-5**, for the four optical sensors (**P1**, **P2**, **P3** and **T**), comprise a standard type of optical fiber connector, each of which connects to a corresponding optical input/output connector **2153** of the control unit **2151**. The control unit **2151** houses a control system comprising a controller with appropriate functionality, e.g. including a processor, data storage, and optical source and optical detector, and it provides a user interface, e.g. a keypad **2154**, and touch screen display **2152**, suitable for tactile user input, and for graphical display of sensor data. The user interface connection **2155** (e.g. a standard USB cable, or alternatively, a wireless connection) is used to transfer data between the control unit **2151** to the touch screen display **2152**. The control unit **2151** and touch screen display **2152** may optionally be integrated within a single housing or module. As illustrated schematically in FIG. 5, each of the optical sensors is connected by an individual optical connector to the control unit. Alternatively, in other embodiments, the proximal ends of the optical fibers for each of the optical sensors may be bundled together and coupled into a single multi-fiber proximal optical connector for connection to a multi-fiber optical port of the control unit. At the proximal end **2102** of apparatus **2001**, a hub **2006** couples each lumen of the catheter via a respective length of flexible tubing to an individual proximal port. The other three ports, **2010-1**, **2010-6** and **2010-7** comprise standard connections, such as Luer fittings, for attachment of a syringe or tubing. When not in use, the latter ports would typically be supplied with sterile plugs. In use, port **2010-1** is connected to a small air filled syringe **2100** for inflation/deflation of the balloon tip **2160**. Port **2010-6** is used for fluid injection or infusion. Port **2010-7** is sized to accept a guidewire **2200**, e.g. a standard 0.025 inch guidewire, or this port can be used for blood sampling at the tip of the catheter.

[0084] The optical pressure sensors **10** (**P1**, **P2**, **P3**) are preferably Fabry-Perot (FP) Micro-Opto-Mechanical System (MOMS) sensors, such as described by FISO Technologies (E. Pinet, "Pressure measurement with fiber-optic sensors: Commercial technologies and applications" 21st International Conference on Optical Fiber Sensors, edited by Wojtek J. Bock, Jacques Albert, Xiaoyi Bao, Proc. of SPIE Vol. 7753, (2011)). These optical pressure sensors comprise an optical fiber having a FP MOMS sensor at the sensor end of the fiber for sensing pressure. By way of example, for standard diameter optical fibers, each fiber (e.g., fibers **2011** in FIGS. 7 and 8) has a diameter of 0.155 mm (0.006 inch) and each optical pressure sensor (e.g., sensors **10** in FIG. 8) has a diameter of 0.260 mm (0.010 inch).

[0085] For measurement of flow by thermo-dilution, the optical temperature sensor **20** (**T**) may, for example, be a GaAs (Gallium Arsenide) fiber optic temperature sensor, as described by FISO technologies (E. Pinet et al., "Temperature fiber-optic point sensors: Commercial technologies and

industrial applications”, MIDEM Conference Proceedings, Sep. 29-Oct. 1, 2010, Radenci, Slovenia).

[0086] A typical material for fabrication of the multi-lumen catheter is a flexible polymer, such as, 4033 Pebax® (a Polyether block amide or PEBA, or other suitable thermoplastic elastomer (TPE)), which has regulatory approval for fabrication of conventional PA catheters. The wall thickness of the tubing may be ~0.005 inch. The guidewire lumen has a diameter, for example, of 0.029 inch to accommodate a standard 0.025 inch guidewire. Conventional coloring of the standard ports may be provided. A different color coding may be provided for the optical ports to facilitate quick recognition and connection to correspondingly color coded ports of the optical controller. As illustrated schematically the transverse cross-sectional views in FIGS. 7 and 8, sensors and fibers of these dimensions can be accommodated within catheter tubing **2002** having an outside diameter of 6 French (0.079 inch or 2.0 mm—see Table 1).

[0087] For some applications a larger diameter catheter, e.g. 7 French, may be acceptable.

[0088] For smaller fibers, e.g. 0.100 mm fibers, and smaller diameter sensors, if a guidewire lumen is not required, or if the guidewire to be used is smaller than 0.025 inch, e.g. 0.018 inch, the dimensions of the lumens and the outside diameter of the catheter tubing may be reduced in size accordingly, e.g. to 5 French or less.

[0089] It is preferable that the arrangement of the lumens has rotational symmetry about the longitudinal axis, and the wall thickness of each lumen is selected to provide the required mechanical characteristics, such as an appropriate degree of flexibility and stiffness, with symmetric torque characteristics along its length. For over the guidewire insertion, a more flexible catheter may be selected. For insertion without a guidewire, a stiffer catheter may be desirable. For example, while the catheter requires sufficient flexibility to traverse from the RA into the RV and then be guided into the PA, it is also desirable that the catheter has sufficient stiffness or rigidity (i.e. is not too floppy) to withstand turbulent blood flow within the ventricles, to withstand distortion or kinking, and to maintain a minimum bend radius of the optical fibers.

[0090] When the optical pressure sensors are FP MOMS sensors, they measure pressure at point locations of the sensor at the end of the fiber, i.e. pressure exerted on the FP membrane, and optical measurements are based on interference measurements, i.e. frequency shifts, rather than amplitude measurements. Blood pressure measurements are made with greater accuracy and reliability compared to conventional pressure sensing with a fluid filled catheter and an external pressure transducer. FP MOMS sensors can provide significantly more accurate pressure measurements, with minimal drift, compared to electrical pressure sensors, such as piezo-electric sensors. Optical pressure sensors avoid the need for multiple long thin electrical connections, which not only have significant electrical drift, but are subject to cross-talk and electro-magnetic interference. For similar reasons, it is also preferable that for measurement of flow by thermo-dilution or thermo-convection, the temperature sensor is preferably also an optical sensor rather than an electrical sensor. For example, for thermo-dilution measurements, the temperature sensor may be a fiber optic sensor which measures temperature based on the temperature

dependence of a GaAs sensor at the tip of the fiber, i.e. a temperature dependent shift in the peak wavelength of light reflected from the sensor.

[0091] FIG. 9 provides a schematic diagram of the heart **500** to illustrate placement of the multi-sensor catheter **2001** of the first embodiment during a right heart and PA catheterization procedure to measure pressure concurrently in the right atrium **521**, right ventricle **523** and PA **525**. In this example, the multi-sensor catheter **2001** is introduced into the heart through the inferior vena cava **520**, through the right atrium **521**, through the tricuspid valve **522** into the right ventricle **523**, through the pulmonic valve **524** and into the PA **525**. The tip **2120** of the catheter extends into the left branch **526L** of the PA. The aperture **2012-2** for pressure sensor P1, which is distal to the balloon **2160** is located in the left branch of the PA **526L**. The temperature sensor T is also in the left branch of the PA **526L**, proximal to the balloon. Aperture **2012-3** for pressure sensor P2 is located in the right ventricle **523**. Aperture **2012-4** for pressure sensor P3 is located in the right atrium **521**. The proximal fluid injection port **2012-6** is also located in the right atrium **521**. Thus, when the optical connectors for the three optical pressure sensors P1, P2 and P3, and the optical temperature sensor T are connected to the optical control system, the pressure measurements in the RA, RV and PA can be made concurrently. Also, the optical temperature sensor T can be used for measurement of flow by a conventional thermo-dilution technique.

[0092] For some applications, a temperature sensor for measurement of flow by thermo-dilution may not be required, and it may be omitted.

[0093] A multi-sensor PA catheter **3001** of a second embodiment is illustrated in FIGS. 10, 11 and 12. As shown in a longitudinal cross-sectional view of FIG. 10, in this embodiment, the multi-sensor catheter comprises three optical pressure sensors **10**, at sensor locations similar to those of the multi-sensor PA catheter **2001** of the first embodiment, but there is no temperature sensor. As illustrated in the transverse cross-sectional views in FIGS. 11 and 12, in this simplified version of the multi-sensor PA catheter, the catheter tubing **3002** comprises five lumens, i.e. a first lumen **3004-1** for balloon inflation; three lumens **3004-2**, **3004-3**, **3004-4**, one for each of the three pressure sensors **10** and its respective optical fiber **3011**; and the central lumen **3004-7** is provided for a guidewire, and/or for blood sampling, as described for the multi-sensor PA catheter of the first embodiment. As shown in FIG. 10, the optical fibers **3011** for the optical pressure sensors **10** are connected from the respective lumens, through connection hub **3006**, and via flexible tubing **3008** to respective ports **3010-2**, **3010-3** and **3010-4**. Ports **3010-2**, **3010-3**, **3010-4** comprise optical connectors for the three optical pressure sensors **10** in the second, third and fourth lumens. Each of the three lumens **3004-2**, **3004-3**, **3004-4** has a respective aperture **3012-2**, **3012-3** and **3012-4** for fluid contact with the respective optical pressure sensors **10**. Port **3010-1** connects via tubing **3007** to the hub **3006** and the first lumen for inflation/deflation of the balloon **3160**. Port **3010-7** connects via tubing **3007** to the central lumen which has an opening **3012-7** at the distal tip **3120** for a guidewire, or for blood sampling at the tip of the catheter. Radiopaque markers **3014** are provided along the length of the catheter tubing to assist in positioning the sensors **10** within the heart. As illustrated, by way of example, in FIGS. 11 and 12, the catheter tubing

has an external diameter of 6 French, and provides a central lumen **3004-7** having an inside diameter which can accept a guidewire, e.g. a standard 0.025 inch guidewire, and four lumens arranged symmetrically around the central lumen. That is, one lumen for balloon inflation, and three lumens for three optical fibers of 0.155 mm diameter and FP-MOMS optical pressure sensors of 0.260 mm diameter. The outside diameter of the PA catheter tubing may be reduced if smaller diameter optical fibers and sensors are used, or for use with a smaller diameter guidewire.

[0094] In comparing the multi-sensor catheters of the first and second embodiments, comprising 7 lumens and 5 lumens respectively, it will be appreciated that the dimensions of each catheters, such as the external diameter, the number of lumens and the thicknesses of the internal walls of the catheter defining each lumen are described by way of example only. As mentioned above, smaller optical sensors and smaller optical fibers may be accommodated within smaller lumens, to provide a catheter having a smaller outside diameter. This may be desirable for some applications. The material from which the catheter is made, and the wall thicknesses defining the lumens, may be selected to provide the catheter with a required stiffness or flexibility, and size.

[0095] In other alternative embodiments, when the catheter is to be flow directed by the balloon tip, and introduction over a guidewire is not required, the guidewire lumen may be omitted.

[0096] For some applications, for example for pediatric or neonatal use, a significantly smaller diameter catheter may be required, e.g. 3 French. Correspondingly, the spacings of the optical pressure sensors would be closer together, i.e. matched to the smaller dimensions of the chambers of patient's heart, for placement of one sensor in the RA, one in the RV and one in the PA. In such a case the guidewire lumen may be omitted so that three pressure sensors can be accommodated within a multi-lumen catheter of the required diameter. While three pressure sensors are desirable for concurrent measurements of RA, RV and PA pressure waveforms, when a guidewire lumen is required for a smaller diameter catheter, it may only be possible to accommodate two optical pressure sensors. In this arrangement, the two sensors would be spaced apart so that initially, one sensor can be positioned in the RA and one in the RV for concurrent measurement of RA and RV pressure waveforms, and then subsequently the catheter would be advanced to position one sensor in the RV and one in the PA for concurrent measurement of RV and PA pressure waveforms, and for RV and PCWP pressure waveforms.

[0097] The lengthwise spacings (L_2 and L_3) of the optical pressure sensors at locations **P1**, **P2** and **P3** described with respect to the multi-sensor catheters of the first and second embodiments, i.e. for measurement of pressure waveforms concurrently in the RA, RV and PA refer to typical spacings required for an adult human heart, where the distance from the RA to the RV is about 10 cm and the distance from the RV to the PA, in a region downstream of the pulmonic valve, is about 10 cm. The PCWP position is typically a further 10 cm into one of the right or left branches of the pulmonary artery, i.e. about 20 cm from the RV. Thus, to position **P1** in the PA near the wedge position, **P2** in the RV, and **P3** in the RA, L_2 is about 20 cm and L_3 is about 30 cm. The location of the temperature sensor **T** is typically positioned between **P1** and **P2**, spaced a distance L_1 from **P1**, for measurement

of blood flow within the PA, where L_1 is e.g. about 4 cm to 10 cm. For pediatric and neonatal use, i.e. for smaller sized hearts, the spacings of the sensors would be reduced accordingly.

[0098] While it is envisaged that multi-sensor catheters for right heart and PA catheterization may comprise more than three optical pressure sensors, there is a practical limit to how many sensors can be accommodated within a multi-lumen catheter of a particular outside diameter.

[0099] Since multi-sensor PA catheters are intended as disposable, single use catheters, in practice, the number of optical sensors may also be limited by component costs and fabrication costs. Currently, standard diameter optical fibers and optical pressure sensors are lower cost than smaller diameter optical fibers and optical pressure sensors. Since each pressure sensor is in an individual lumen of the multi-lumen catheter, the available space for each lumen is also limited by the wall thickness and tolerances for each lumen of a multi-lumen catheter. As described herein, it is currently feasible to manufacture a multi-sensor PA catheter with three optical pressure sensors and one optical temperature sensor, within a 6 French multi-lumen catheter. Use of smaller fibers and sensors or smaller guidewire may allow the diameter to be reduced to 5 French or less.

[0100] In a multi-sensor catheter of yet another embodiment, instead of three optical pressures sensors and one optical temperature sensor, it may be desirable to have four optical pressure sensors to enable concurrent pressure measurements in the right atrium, in the right ventricle, in the PA near the pulmonic valve and also in a branch of PA for measurement of the PCWP.

[0101] For example, if a temperature sensor is not required for blood flow measurements, for example where blood flow is measured by an alternative technique, e.g. by the Fick method, a fourth optical pressure sensor may be provided instead of the optical temperature sensor, so that the multi-sensor catheter can be introduced so as to position one sensor in the RA, one sensor in the RV, one sensor in the PA, and one sensor for measuring PCWP when the balloon is inflated. By way of example, in such an arrangement, four pressure sensors **P1**, **P2**, **P3** and **P4** are spaced at intervals of ~10 cm, i.e. the distance **P1** to **P2** (L_1) is 10 cm, **P1** to **P3** (L_2) is 20 cm, and **P1** to **P4** (L_3) is 30 cm.

[0102] This arrangement may be desirable for longer term monitoring of pressure waveforms in the RA, RV, PA, as well PCWP pressure waveforms. That is the catheter may be positioned in a fixed and stable location, to enable observation of pressure waveforms at each sensor location over an extended time period, e.g. for ICU patients requiring monitoring over several days or more.

[0103] Alternatively, where it is not feasible to accommodate an optical temperature sensor as well as the desired number of optical pressure sensors, or for cost reasons, a conventional small sized, low cost, electrical flow sensor, i.e. a thermistor, may be used, with conventional electrical connections to the control system.

[0104] In the embodiments described above, radiopaque markers may be provided near the balloon, and optionally near each sensor, to assist in locating the tip and positioning the sensors in use, i.e. using conventional radio-imaging techniques, when introducing the guidewire and positioning the pressure sensors in the right atrium, right ventricle and PA. The radiopaque markers typically comprise a suitable heavy metal e.g. barium, tantalum, gold or platinum. Alter-

natively, markers are provided at regular intervals, e.g. at 10 cm intervals along the length of the catheter tubing as is conventional for PA catheters.

[0105] Preferably that the optical fibers have some freedom to move or slide within the lumen when the catheter is flexed. The fibers are of the appropriate length so that the sensors at the sensor end (distal end) of the fibers are appropriately positioned at sensor locations in the distal end portion of the catheter. Each of fibers may be secured near the proximal end, e.g. by adhesive bonding where they pass through the hub. Each fiber may also be secured in its lumen, near the sensor location, e.g. by injection of a medical grade adhesive through the wall of the lumen.

[0106] If required, in use, the lumens containing the fibers and sensors may be flushed with fluid, e.g. saline solution, to remove air from the catheter lumens. Alternatively, an adhesive, or a medical grade gel, may serve to plug the lumen each side of the aperture surrounding the optical sensor, while leaving the sensor exposed for fluid contact. For example, a bolus of medical grade adhesive may be injected through the tubing to secure the fiber near each sensor and to plug the lumen around the fiber. Similarly, the adhesive may also be injected into the lumen distal to the aperture. Also, if required, components of the multi-sensor catheter may be coated to reduce blood clotting, for example, if the multi-sensor catheter is to be left in place for an extended period.

[0107] The optical pressure sensors **10** at locations **P1**, **P2**, **P3** are preferably Fabry-Perot Micro-Opto-Mechanical-Systems (FP MOMS) pressure sensors. As an example, a suitable commercially available FP MOMS pressure sensor is the Fiso FOP-M260. These FP MOMS sensors meet specifications for an appropriate pressure range and sensitivity for blood pressure measurements. They have an outside diameter of 0.260 mm (260 μ m). Typically, they would be attached and optically coupled (i.e. integral with or bonded to) to a sensor end of an optical fiber with an outside diameter of 0.100 mm (100 μ m) to 0.155 mm (155 μ m). Optical fibers and FP MOMS sensors of smaller diameter tend to be more expensive, and may be used, when appropriate.

[0108] The optional optical flow sensor **20** may comprise an optical thermo-dilution or an optical thermo-convection flow sensor, e.g. as described in U.S. patent application Ser. No. 14/354,588.

[0109] For operation of the optical sensors, the optical output ports **2010-2**, **2010-3**, **2010-4**, **2012-5** couple to the respective optical ports of the control unit **2151** (e.g. see FIG. 5) for controlling operation of the optical sensors **10** and **20**. The flexible tubing **2008** surrounding each of the optical fibers extends between the hub **2006** and the optical connector of the control unit to provide a flexible optical coupling from the hub **2006** to the control unit **2151**. This tubing is provided to protect the optical fibers and can have any appropriate diameter and flexibility. Since the catheter is intended for single use only, preferably the optical connectors are standard low cost optical connectors. Similarly, the flexible tubing, and other connectors for the other ports are preferably standard materials and components, such as luer fittings or other medical standard fluid ports and electrical connectors, as appropriate, which can be sterilized, and so that the multi-sensor catheters can be provided in single-use sterile packaging, using conventional standard processes for medical devices.

[0110] For protection of the sensors during assembly, it may be preferred to insert optical fibers and optical sensors through the respective lumen from the distal end of the catheter and subsequently form the optical connector at the proximal end, and then close the lumen at the distal tip. On the other hand, when the optical connector is pre-formed at the proximal end of the sub-assembly before insertion, the sensor end of the sub-assembly is inserted into the catheter from proximal end of the respective lumen. In either case, it is preferable that the catheter lumens have smooth rounded surfaces, with non-stick internal surfaces, i.e. to avoid sharp edges, so that the sensors and optical fibers can slide smoothly into their catheter lumen without catching on sharp edges or corners, to avoid mechanical damage to the sensors or optical fibers. However, in use of the multi-sensor catheter, it is preferable that the fibers are fixed at the proximal end only of the catheter tubing so that the fibers have some freedom to move or slide within the lumens when the catheter is flexed. If required, the optical fibers may also be adhesively bonded near the aperture to secure the sensor at the appropriate sensor location.

[0111] As mentioned above, it is desirable that the multi-sensor PA catheter has mechanical characteristics, such as stiffness and flexibility, similar to a standard PA catheter. The optical fibers and optical sensors do not add significant stiffness to the catheter, and thus these characteristics are primarily determined by the type of material and wall thickness used for the multi-lumen catheter tubing.

[0112] Other factors for consideration are: regulatory requirements for medical devices, ease of use and safety. For these reasons, it is desirable that the materials for fabrication of a multi-sensor PA catheter are based on a conventional tried and tested PA catheter or other medical device, i.e. based on a predicate device structure which has regulatory approval and which is fabricated with materials and components which already have FDA and/or CE mark regulatory approval.

[0113] It will be appreciated that in alternative embodiments or variants of the multi-sensor catheters of the embodiments described in detail above, different combinations of one or more features disclosed herein, and features disclosed in the related patent applications referenced herein, may provide multi-sensor catheters of further alternative embodiments

[0114] As disclosed herein, the cardiologist is offered multi-sensor catheters which have particular application for right heart and PA catheterization. These multi-sensor catheters are configured for monitoring and diagnostic measurements of hemodynamic parameters, including concurrent measurement of blood pressure within the RA, RV and PA.

[0115] Control System and Graphical Display of Pressure Waveforms and Associated Hemodynamic Parameters.

[0116] Referring to the control system, which was described above with reference to FIG. 5, and as described in the above referenced related patent applications, it will be apparent that the control system may be used with multi-sensor catheters and multi-sensor guidewires for concurrent blood pressure measurements at each pressure sensor location, using, two, three or more optical pressure sensors. Optionally, for flow measurements, the multi-sensor catheter or guidewire is further equipped with an optical or electrical flow sensor. The control system comprises a light source and detector and an optical interface for coupling, via respective input/output ports, to each of the optical fibers and optical

sensors of the multi-sensor catheter or guidewire. The control system also comprises data storage and processing means configured for processing optical data indicative of pressure values and optionally, optical or electrical data indicative of flow velocity values. For use with multi-sensor catheters and multi-sensor guidewires, for example, the multi-sensor catheter disclosed herein having three optical pressure sensors, and an optional optical flow sensor, the control system would have a corresponding number of signal processing channels with optical inputs for each of the optical sensors. The signal processing elements for each channel may be referred to as a signal conditioning unit. For measuring blood pressure and flow within the heart and blood vessels, in particular for measuring intravascular or transvalvular blood pressure gradients, the processing means is further configured for graphically displaying pressure data, and optionally flow velocity data, comprising a plurality of blood pressure waveforms, i.e. a pressure waveform from each optical pressure sensor. The concurrent blood pressure waveforms for each of the optical sensors may be displayed for one or more time intervals, and during one or more cardiac cycles. Preferably, the processing means is further configured to derive hemodynamic parameters from the blood pressure data, and optionally from flow velocity data, and display numeric values of the parameters as well as display the pressure waveforms from each sensor. By way of example only, some schematic representations of pressure waveforms and associated numeric data, for a patient with a healthy or normally functioning heart, are shown in FIGS. 13A, 13B and 14. In practice, pressure waveforms and pressure values vary from patient to patient and may be dependent on a number of factors, such as, whether or not the patient has a healthy or diseased heart, or other conditions that may affect functioning of the heart. Skilled medical practitioners will recognize characteristic variations in each pressure waveform and associated pressure values, indicative of e.g. valvular stenosis or other patient physiology. Advantageously, in use of a multi-sensor catheter, concurrent pressure measurements from multiple optical pressure sensors enable the cardiologist to directly compare multiple pressure waveforms, in real-time, from the RA, RV and PA/PCWP.

[0117] As an example, FIG. 13A represents a graphical display 2152-1 showing three blood pressure waveforms from a three sensor PA catheter: A. pressure sensor location P1 in the pulmonary artery; B. pressure sensor location P2 in the right ventricle; and C. pressure sensor location P3 in the right atrium. Pressure units are displayed in mmHg. Optionally, it may be desirable to select and display one or multiple pressure waveforms and related parameters in different formats. For example, during right heart and pulmonary artery catheterization with a multi-sensor catheter as disclosed herein, in addition to displaying concurrent blood pressure waveforms from each the right atrium, right ventricle and pulmonary artery, a plurality of numeric values, such as, peak pressures, mean pressures, peak-to-peak (PK-PK) pressure differentials for each curve, and pressure differentials or gradients between the right atrium and right ventricle, and between the right ventricle and pulmonary artery may be displayed in real time. For example, as illustrated schematically by the numeric data to the right of the graphical display shown in FIG. 13A, numeric data selected for display may include the mean pressure of the RA, systolic/diastolic pressure for RV/PA, a peak-to-peak

gradient (PPG) for RV to PA, a PPG for RA to RV, and heart rate HR. As is conventional, the user interface may include a number of buttons or keys, such as shown at the bottom of the display in FIG. 13A, e.g. to select parameters for display, change display modes, and input identification. FIG. 13B represents a graphical display 2152-2 for three pressure waveforms, similar to that shown in FIG. 13A, except that waveform A for the PA pressure is replaced with waveform D for the pressure sensor location P1 is in the pulmonary capillary wedge position with the balloon inflated.

[0118] Correspondingly, in use of a multi-sensor guidewire during TAVR, e.g. a multi-sensor guidewire with two optical pressure sensors to assess functioning of the aortic valve, before and after a TAVR procedure. As illustrated schematically in FIG. 14, the system comprises a graphical display 2052-3 which shows two pressure waveforms: A. from pressure sensor location P1 in the left ventricle and B. from pressure sensor location P2 in the aorta, together with numeric values derived from the pressure waveforms from the left ventricle and aorta, e.g. a transvalvular pressure gradient. For example, the system may automatically compute an aortic regurgitation index (ARI or AR Index) and display the ARI value in real time. The ARI may be computed from measured values of the left ventricular end-diastolic pressure (LVEDP), diastolic blood pressure (DBP), and systolic blood pressure (SBP), as illustrated in FIG. 14, and is defined as:

$$ARI = ((DBP - LVEDP) / SBP) \times 100$$

[0119] In one embodiment, the control system comprises a signal processing unit for receiving optical data and optionally electrical data, from a multi-sensor catheter or guidewire. The signal processing unit is coupled by a data connection to a general purpose computer system, which may be personal computer (PC), such as a laptop or tablet PC, comprising processing means, i.e. one or more processors and a computer program product, embodied in a non-transitory computer readable medium storing instructions, in the form of code, for execution by the processing means. The computer program product is, for example, a software application comprising instructions for execution in a processor of the tablet PC for receiving or retrieving data, and displaying a plurality of concurrent pressure waveforms from the optical pressure sensors, and for computing, and displaying in real-time, associated hemodynamic parameters or an index, such as ARI.

[0120] In another example of a TAVT procedure, a multi-sensor guidewire or catheter may be used for measurement of concurrent pressure waveforms upstream and downstream of the mitral valve, e.g. to assess functioning of the mitral valve before and after TMVI.

[0121] For example, the tablet PC is configured for graphically displaying pressure data, and optionally flow velocity data, e.g. comprising a plurality of blood pressure waveforms. The concurrent blood pressure waveforms for each optical sensor may be displayed for one or more time intervals, and during one or more cardiac cycles. The processing means is further configured to derive and display hemodynamic parameters from the blood pressure data and flow velocity data. For example, during right heart and pulmonary artery catheterization with a multi-sensor catheter as disclosed herein, in addition to displaying blood pressure waveforms from the right atrium, right ventricle and pulmonary artery, a selected plurality of numeric values

such as peak pressures, mean pressures, peak to peak pressure differentials for each curve, and pressure differentials or gradients between the right atrium and right ventricle, and between the right ventricle and pulmonary artery can be displayed in real time.

[0122] As is conventional, the system may comprise a user interface, such as a keyboard or touchscreen, to allow the operator to select from available information which waveforms or parameters are to be displayed. The interface may allow the operator to input user data such as patient identification, and data interfaces may be provided to output data to other devices or systems, or receive data from other sources, such as from other sensors or monitoring systems, which are typically used in an ICU or OR. For example, in a cardiac catheterization laboratory, the control system for a multi-sensor catheter or guidewire may be coupled to, or part of, a computing system controlling other equipment, and which is equipped with one or more large screen displays close to the operating table, and other remote displays in a monitoring area. The latter are used to display various forms of data, sequentially, concurrently, or on demand. Such data may include, e.g. fluoroscopic imaging, with or without contrast media, and transesophageal echo-cardiography (TEE) images, as well as sensor data comprising pressure waveforms from the multi-sensor catheter or guidewire and associated hemodynamic parameters calculated or derived from the received optical pressure sensor data.

[0123] While a specialized signal processing unit or interface, which may be referred to as a “signal conditioner”, is used to receive optical data from the multi-sensor catheter or multi-sensor guidewire, and generate output data indicative of pressure for display of pressure waveforms, the output data may be fed by a standard data connection, wired or wireless, to a processor, such as a general purpose computer, which is configured to provide the required functionality. For example, the system includes a processor and a computer program product (typically referred to as a software application or computer code), embodied in a non-transitory computer readable medium storing instructions, for execution in a processor of a control system for a multi-sensor catheter or a multi-sensor guidewire, for processing optical data received concurrently from a plurality of optical pressure sensors indicative of blood pressure, displaying a corresponding plurality of blood pressure waveforms, and optionally flow velocity data, and displaying numeric data relating to selected hemodynamic parameters and indexes.

TABLE 2

Abbreviations or acronyms	
ARi or AR Index	Aortic Regurgitation Index
Cath Lab	Cardiac Catheterization Laboratory
CO	Cardiac Output
CVC	Central Venous Pressure catheter
DBP	Diastolic Blood Pressure
FP MOMS Sensor	Fabry-Pérot Micro-Opto-Mechanical-System Sensor
ICU	Intensive Care Unit
LAP	Left Atrial Pressure
LVEDP	Left Ventricular End-Diastolic Pressure
OR	Operating Room
PA	Pulmonary Artery
PCWP	Pulmonary Capillary Wedge Position/Pressure
PEBA	PolyEther Block Amide
RA	Right Atrium
RHC	Right Heart Catheterization

TABLE 2-continued

Abbreviations or acronyms	
RV	Right Ventricle
SvO ₂	Mixed venous oxygen saturation
SBP	Systolic Blood Pressure
TAVI or TAVR	Transcatheter Aortic Valve Implantation or Replacement
TMVI or TMVR	Transcatheter Mitral Valve Implantation or Replacement
TPE	Thermoplastic elastomer
TVR	Transcatheter heart Valve Replacement
TVT	Transcatheter Valve Therapies (TVT).

INDUSTRIAL APPLICABILITY

[0124] Multi-sensor PA catheters according to embodiments of the invention disclosed herein provide real-time, concurrent, multi-chamber (RA and RV) pressure measurements within the right heart and also the PA. Each pressure measurement is taken concurrently under identical and stable conditions, while the multi-sensor catheter is positioned to locate one sensor in each of the RA, RV and PA.

[0125] In contrast, during RHC using a conventional PA catheter, it is necessary to move the catheter to get each pressure measurement, so each measurement is taken at a different time, under different conditions. For example, withdrawing a PA catheter from the PA to the RV may cause cardiac arrhythmia or premature ventricular contraction (PVC). That is, the instantaneous condition for each pressure measurement is impacted by moving the catheter.

[0126] In some instances, such as monitoring in an ICU, it may be too risky to move the PA catheter once the catheter tip is placed in the PA. For a conventional PA catheter, it would then only provide the PA pressure, not a RV and RA pressure. Thus, where appropriate, the multi-sensor catheter offers continuous real-time and concurrent monitoring of all of RA, RV and PA pressures for an extended time, e.g. over a period of days, for an ICU patient. In this scenario, RA pressure monitoring provides alternative to monitoring of central venous pressure (CVP).

[0127] For post-operative monitoring, there may be risks in use of a Swan Ganz catheter, and there is a need to appropriately select patients who need Swan Ganz catheterization. For patients where Swan Ganz catheterization is appropriate, the multi-sensor PA catheter offers continuous real-time and concurrent monitoring of all of RA, RV and PA pressures.

[0128] Since these pressure measurements also provide an indirect measure of left heart hemodynamic parameters, these measurements can help to identify pathology and physiological problems and select appropriate therapies, drugs, and procedures. For example, pressure measurements may help to differentiate patient physiologies, and identify a filling problem vs. a valve problem, such as, an obstruction of the tricuspid valve.

[0129] For example, multiple concurrent blood pressure measurements during RHC may show, e.g., a high RA pressure and a low RV pressure, which may indicate tricuspid valve stenosis/obstruction. In conditions such as pulmonary edema, concurrent pressure measurements of the RA, RV, PA and PCWP may provide information which helps to determine or differentiate, e.g., whether symptoms are caused by a RH or LH problem, a valve problem (stenosis or regurgitation), a muscle problem, cardiac restriction or

constriction, PA hypertension, or a primary lung problem (such as Acute Respiratory Distress Syndrome).

[0130] The number of TAVR procedures per year is projected to increase dramatically in the next few years. During left heart catheterization for TAVR, the Aortic Regurgitation index (ARi) is a widely used parameter for assessing function of the aortic valve before and after valve repair or replacement. It is also expected that the mitral valve replacement will increase in future years, but aortic valve replacement is much more common.

[0131] By comparison, pulmonic valve interventions are rare and there is not a commonly used PA index. However, multi-sensor catheters providing concurrent pressure measurements in the RA, RV and PA may allow for an appropriate index to be adopted for RHC.

[0132] Although embodiments of the invention have been described and illustrated in detail, it is to be clearly understood that the same is by way of illustration and example only and not to be taken by way of limitation, the scope of the present invention being limited only by the appended claims.

1. A multi-sensor catheter for right heart and pulmonary artery catheterization comprising:

a length of catheter tubing extending between a proximal end and a distal end, and the distal end comprising a distal tip;

a plurality of optical sensors and a plurality of optical fibers; a sensor end of each optical fiber being attached and optically coupled to an individual one of the plurality of optical sensors;

the plurality of optical sensors and optical fibers extending within the catheter tubing, the sensors being spaced apart lengthwise to provide a sensor arrangement with said plurality of optical sensors positioned at respective sensor locations spaced apart lengthwise within a distal end portion of the catheter tubing;

a proximal end of each of the plurality of optical fibers being coupled to an optical input/output connector at the proximal end of the multi-sensor catheter for connection to an optical control system; and

the plurality of optical sensors of the sensor arrangement comprising at least two optical pressure sensors, with an aperture in the catheter tubing adjacent each optical pressure sensor for fluid contact; and

wherein the sensor locations and spacings of the at least two optical pressure sensors are configured to measure concurrent pressure waveforms for at least two locations within the right heart and pulmonary artery.

2. The multi-sensor of claim 1, wherein the sensor locations and spacings of the at least two optical pressure sensors are configured to measure concurrent pressure waveforms for at least two locations within the right heart and pulmonary artery comprising a right atrial pressure waveform, a right ventricular waveform, a pulmonary artery waveform and pulmonary capillary wedge position waveform, to provide concurrent blood pressure measurements at each optical pressure sensor location.

3. The multi-sensor catheter of claim 1, having an outside diameter sized to allow insertion through a vein in an upper or lower arm.

4. The multi-sensor catheter of claim 1, comprising an inflatable balloon near the distal tip, the inflatable balloon

being coupled by a balloon inflation lumen of the catheter tubing to a balloon inflation port at the proximal end of the catheter tubing.

5. The multi-sensor catheter of claim 1, further comprising a guidewire lumen.

6. The multi-sensor catheter of claim 1, further comprising:

a guidewire lumen;

an inflatable balloon near the distal tip, the inflatable balloon being coupled by a balloon inflation lumen of the catheter tubing to a balloon inflation port at the proximal end of the catheter tubing;

wherein the catheter tubing has an outside diameter sized to allow insertion into the heart through a vein in an upper or lower arm; and

wherein the sensor locations and spacings of the at least two optical pressure sensors are configured to measure concurrent pressure waveforms for at least two locations within the right heart and pulmonary artery comprising a right atrial pressure waveform, a right ventricular waveform, a pulmonary artery waveform and pulmonary capillary wedge position waveform, to provide concurrent blood pressure measurements at each optical pressure sensor location.

7. A control system for a multi-sensor catheter for right heart catheterization and pulmonary artery catheterization, wherein said plurality of optical pressure sensors for measuring blood pressure concurrently at each optical pressure sensor location are configured to measure concurrent pressure waveforms for at least two locations within the right heart and pulmonary artery comprising a right atrial pressure waveform, a right ventricular waveform, a pulmonary artery waveform and a pulmonary capillary wedge pressure, to provide concurrent blood pressure measurements at each optical pressure sensor location, the control system comprising:

a light source and detector, and an optical interface for coupling, via respective input/output ports, to each of the optical fibers and optical sensors of the multi-sensor catheter;

data storage and a processor configured for processing optical data indicative of pressure values obtained from each of said optical pressure sensors to generate data for pressure blood pressure waveforms; and

wherein the processor is further configured for outputting said data to a patient monitoring system to display concurrent blood pressure waveforms for at least two of the right atrium, right ventricle and pulmonary artery, and pulmonary capillary wedge pressure.

8. A control system for multi-sensor catheters or multi-sensor guidewires comprising a plurality of optical pressure sensors for measuring blood pressure concurrently at each optical pressure sensor location, comprising:

a light source and detector, and an optical interface for coupling, via respective input/output ports, to each of the optical fibers and optical sensors of the multi-sensor catheter or multi-sensor guidewire;

data storage and processing means configured for processing optical data indicative of pressure values; and

wherein the processing means is further configured for graphically displaying pressure data comprising a plurality of concurrent blood pressure waveforms derived from each of the plurality of pressure sensors.

9. The control system of claim 8, wherein for right heart catheterization and pulmonary artery catheterization using a multi-sensor catheter comprising at least three optical pressure sensors,

the processing means is further configured for graphically displaying pressure data comprising a plurality of concurrent blood pressure waveforms from each of the optical pressure sensors, comprising displaying concurrent blood pressure waveforms from the right atrium, right ventricle and pulmonary artery.

10. The control system of claim 9, wherein the pressure waveforms for each optical pressure sensor may be displayed together for comparison, or displayed individually, for one or more time intervals, and during one or more cardiac cycles.

11. The control system of claim 10, further configured to derive and display hemodynamic parameters from the blood pressure data, and optionally from flow velocity data obtained from a flow sensor.

12. The control system of claim 1, wherein said hemodynamic parameters comprise one or more of a plurality of numeric values comprising peak pressures, mean pressures, peak-to-peak pressure differentials for each curve, and pressure differentials or gradients.

13. The control system of claim 8, wherein for use of a multi-sensor guidewire during transcatheter aortic valve replacement (TAVR) wherein a sensor arrangement of the multi-sensor guidewire is configured for placement of at least one optical pressure sensor in the aorta and at least one optical pressure sensor in the right ventricle, the system is configured to display numeric values based on pressure waveforms from the aorta and the left ventricle, and to display a plurality of numeric values comprising peak pressures, mean pressures, peak-to-peak pressure differentials for each pressure waveform, and transvalvular pressure differentials or gradients between the aorta and the left ventricle; to compute parameters comprising an aortic regurgitation index (ARi); and to display an ARi value and other parameter values in real-time.

* * * * *

专利名称(译)	系统和装置，包括用于右心和肺动脉导管插入术的多传感器导管		
公开(公告)号	US20200046234A1	公开(公告)日	2020-02-13
申请号	US16/587659	申请日	2019-09-30
申请(专利权)人(译)	三江心血管SYSTEMS INC.		
当前申请(专利权)人(译)	三江心血管SYSTEMS INC.		
[标]发明人	GLOVER CHRISTOPHER CARON ERIC ABEL SYLVAIN		
发明人	GLOVER, CHRISTOPHER CARON, ERIC ABEL, SYLVAIN		
IPC分类号	A61B5/0215 G01L11/02 A61B5/01 G01K13/02 A61B5/00 A61B5/026 G01K11/32 A61B5/0205 G01P5/10 A61B5/0295 A61M25/10 A61F2/24 A61B5/028		
CPC分类号	A61B2090/306 G01K11/32 A61B5/028 A61M25/1018 A61M2025/1079 A61F2/2427 A61B2562/04 A61B5/6851 A61M25/1025 A61B2562/0247 A61B5/7425 A61B5/6852 A61B5/02055 A61B5/743 A61B2090/3966 A61B5/026 A61B5/02154 A61B5/0215 A61B5/01 A61B5/0295 A61B5/0002 G01L11/02 G16H40/63 A61B5/4836 A61B5/024 A61B5/6853 G01P5/10 A61B5/0261 A61B2562/028 A61B2562/0266 A61B2562/0233 G01K13/02 A61B5/02158 G16H20/40		
优先权	14/354624 2014-04-28 US PCT/IB2012/055893 2012-10-26 WO 61/552778 2011-10-28 US 61/552787 2011-10-28 US		
外部链接	Espacenet USPTO		

摘要(译)

公开了一种包括用于右心和肺动脉导管插入术的多传感器导管的系统和设备。多传感器导管包括多腔导管，至少三个光学压力传感器及其各自的光纤插入该多腔导管中。三个光学压力传感器布置在导管的远端部分内，在远端部分内纵向间隔开，以同时测量每个传感器位置处的压力。传感器位置被配置用于在右心房，右心室和肺动脉的每一个中放置至少一个传感器，以同时测量每个传感器位置的壓力。传感器装置可以进一步包括光学热稀释传感器，并且提供另一个内腔用于流体注入以进行热稀释测量。导管可以包括可膨胀的球囊尖端和导丝管腔，并且优选地具有6弗伦或更小的外径。

