



US 20180263522A1

(19) **United States**

(12) **Patent Application Publication**
Ghosh et al.

(10) **Pub. No.: US 2018/0263522 A1**

(43) **Pub. Date: Sep. 20, 2018**

(54) **QRS OFFSET AND ONSET
DETERMINATION**

(71) Applicant: **Medtronic, Inc.**, Minneapolis, MN
(US)

(72) Inventors: **Subham Ghosh**, Blaine, MN (US);
D'Anne Kudlik, Lino Lakes, MN (US);
Ya-Jian Cheng, Lino Lakes, MN (US)

(21) Appl. No.: **15/915,995**

(22) Filed: **Mar. 8, 2018**

Related U.S. Application Data

(60) Provisional application No. 62/471,938, filed on Mar.
15, 2017.

Publication Classification

(51) **Int. Cl.**

A61B 5/0456 (2006.01)

A61B 5/0408 (2006.01)

A61B 5/00 (2006.01)

(52) **U.S. Cl.**

CPC **A61B 5/0456** (2013.01); **A61B 5/6801**
(2013.01); **A61B 5/0408** (2013.01)

(57)

ABSTRACT

The exemplary systems and methods may be configured to generate a dispersion signal from a plurality of cardiac signals and determine a QRS onset time value and a QRS offset time value from the plurality of cardiac signals. The QRS onset and offset time values may be used to measure, or capture, activation times.




FIG. 1




FIG. 2




FIG. 3




FIG. 4





FIG. 5

FIG. 6




FIG. 7





FIG. 8






FIG. 10




FIG. 11




FIG. 12




FIG. 13




FIG. 14




FIG. 15




FIG. 16




FIG. 17




FIG. 18




FIG. 19




FIG. 20A




FIG. 20B




FIG. 21A





FIG. 21B



QRS OFFSET AND ONSET DETERMINATION

[0001] This application claims the benefit of U.S. Provisional Application Ser. No. 62/471,938, filed Mar. 15, 2017, the entire contents of which are incorporated by reference herein in their entirety.

[0002] The disclosure herein relates to systems and methods for use in the determination of QRS onset and offset time values from a plurality of cardiac signals.

SUMMARY

[0003] The exemplary systems, methods, and interfaces described herein may be configured to assist a user (e.g., a physician) in evaluating a patient and/or evaluating cardiac therapy (e.g., cardiac therapy being performed on a patient during and/or after implantation of cardiac therapy apparatus). In one or more embodiments, the systems, methods, and interfaces may be described as being noninvasive. For example, in some embodiments, the systems, methods, and interfaces may not need, or include, implantable devices such as leads, probes, sensors, catheters, implantable electrodes, etc. to monitor, or acquire, a plurality of cardiac signals from tissue of the patient for use in evaluating the patient and/or cardiac therapy. Instead, the systems, methods, and interfaces may use electrical measurements taken noninvasively using, e.g., a plurality of external electrodes attached to the skin of a patient about the patient's torso.

[0004] In one or more embodiments, the systems, methods, and interfaces may use, or include, implantable devices such as leads, probes, implantable electrodes, etc. to monitor, or acquire, a plurality of cardiac signals from tissue of the patient for use in evaluating the patient and/or cardiac therapy. For example, the systems, methods, and interfaces may use electrical measurements taken using, e.g., a plurality of implantable electrodes either implanted long term or temporarily within a patient.

[0005] One exemplary system may include electrode apparatus and computing apparatus comprising one or more processors and coupled to the electrode apparatus. The electrode apparatus may include a plurality of electrodes to monitor electrical activity from tissue of a patient. The computing apparatus may be configured to monitor electrical activity using the plurality of electrodes to generate a plurality of cardiac signals over time and generate a dispersion signal from the plurality of cardiac signals. The dispersion signal may be representative of the dispersion of the plurality of cardiac signals over time. The computing apparatus may be further configured to determine a QRS onset time value and a QRS offset time value for each of at least one QRS complex within the monitored electrical activity based on the dispersion signal.

[0006] One exemplary method may include monitoring electrical activity from tissue of a patient using a plurality of electrodes to generate a plurality of cardiac signals over time and generating a dispersion signal from the plurality of cardiac signals. The dispersion signal may be representative of the dispersion of the plurality of cardiac signals over time. The exemplary method may further include determining a QRS onset time value and a QRS offset time value for each of at least one QRS complex within the monitored electrical activity based on the dispersion signal.

[0007] In one or more embodiments, the computing apparatus may be further configured to execute or the method

may further include determining an activation time for each of the at least one QRS complex within each of the plurality of cardiac signals within the QRS duration between the QRS onset and offset time values corresponding to the at least one QRS complex. Further, the computing apparatus may be further configured to execute or the method may further include determining at least one metric of electrical heterogeneity based on the activation time for each of the at least one QRS complex within each of the plurality of cardiac signals.

[0008] In one or more embodiments, the plurality of electrodes may include a plurality of external electrodes to be located proximate the patient's skin and/or a plurality of internal electrodes to be located within the patient. In one or more embodiments, generating a dispersion signal from the plurality of cardiac signals may include determining a standard deviation of the plurality of cardiac signals over time.

[0009] In one or more embodiments, the computing apparatus may be further configured to execute or the method may further include selecting a portion of the dispersion signal over a selected time period. Determining a QRS onset time value and a QRS offset time value for each of at least one QRS complex within the monitored electrical activity based on the dispersion signal may include comprise determining the QRS onset time value and the QRS offset time value for each of the at least one QRS complex within the monitored electrical activity of the selected time period based on the selected portion of the dispersion signal. In one or more embodiments, the selected time period may be less than or equal to 5 seconds.

[0010] In one or more embodiments, determining a QRS onset time value and a QRS offset time value for each of at least one QRS complex within the monitored electrical activity based on the dispersion signal may further include determining a maximum dispersion value of the dispersion signal and determining the QRS onset time value and the QRS offset time value using the maximum dispersion value. In one or more embodiments, determining a QRS onset time value and a QRS offset time value for each of at least one QRS complex within the monitored electrical activity based on the dispersion signal may further include filtering the dispersion signal to remove one or more values of the dispersion signal that are less than or equal to a selected percentage of the maximum dispersion value of the dispersion signal. In one or more embodiments, determining a QRS onset time value and a QRS offset time value for each of at least one QRS complex within the monitored electrical activity based on the dispersion signal may further include filtering the dispersion signal to remove one or more dispersion values of the dispersion signal corresponding to at least one of a T-wave and a P-wave in the plurality of cardiac signals.

[0011] In one or more embodiments, determining a QRS onset time value and a QRS offset time value for each of at least one QRS complex within the monitored electrical activity based on the dispersion signal may further include determining a minimum dispersion value of the dispersion signal within a first time period before a maximum dispersion value of the dispersion signal to determine the QRS onset time value and determining a minimum dispersion value of the dispersion signal within a second time period after the maximum dispersion value of the dispersion signal to determine the QRS offset time value. In one or more

embodiments, each of the first and second time periods may be less than or equal to 120 milliseconds. In one or more embodiments, each of the first and second time periods may be less than or equal to 200 milliseconds.

[0012] In one or more embodiments, determining a QRS onset time value and a QRS offset time value for each of at least one QRS complex within the monitored electrical activity based on the dispersion signal may further include determining a latest dispersion value of the dispersion signal within a first time period before a maximum dispersion value of the dispersion signal that is less than or equal to a selected percentage of a minimum dispersion value of the dispersion signal within the first time period to determine the QRS onset time value and determining an earliest dispersion value of the dispersion signal within a second time period after the maximum dispersion value of the dispersion signal that is less than or equal to a selected percentage of a minimum dispersion value of the dispersion signal within the second time period to determine the QRS onset time value

[0013] In one or more embodiments, the computing apparatus may be further configured to execute or the method may further include determining an absolute amplitude between the QRS onset time value and the QRS offset time value for each of the at least QRS complex for each of the plurality of cardiac signals, selecting one or more cardiac signals of the plurality of cardiac signals based on the absolute amplitude, and determining an activation time for each of the at least one QRS complex within each of the selected one or more cardiac signals within the QRS duration between the QRS onset and offset time values corresponding to the at least one QRS complex. In one or more embodiments, selecting one or more cardiac signals of the plurality of cardiac signals based on the absolute amplitude may include selecting one or more cardiac signals of the plurality of cardiac signals defining an absolute amplitude greater than or equal to a threshold value, and the threshold value may be less than the median absolute amplitude of the plurality of cardiac signals.

[0014] In one or more embodiments, the computing apparatus may be further configured to execute or the method may further include removing one or more noisy cardiac signals from the plurality of cardiac signals prior to determining the dispersion signal. Further, in one or more embodiments, the one or more noisy cardiac signals may be identified by identifying cardiac signals comprising absolute areas-under-the-curve greater than or equal to a threshold value. In one or more embodiments, the at least one QRS complex may include a plurality of QRS complexes

[0015] In one or more embodiments, the computing apparatus may be further configured to execute or the method may further include adjusting cardiac therapy in response to one or more metrics derived using the determined QRS onset and offset time values and/or displaying on a graphical user interface one or more metrics derived using the determined QRS onset and offset time values. Further, the one or more metrics may include activation time of each of the plurality of cardiac signals.

[0016] Electrode apparatus (e.g., an "ECG belt") may be used for various cardiac resynchronization therapy (CRT) applications including optimizing lead location/pacing parameters, etc. based on one or more metrics of electrical heterogeneity derived from activation times computed from the multiple electrodes on the body-surface.

Determination of activation times may include determination of the earliest onset and the latest offset of QRS complex (depolarization activity) based on cardiac signals, or electrocardiogram (ECG) signals, from all the electrodes on the body-surface.

[0017] Various systems and methods may identify a QRS onset/offset from a single cardiac signal or channel of an ECG. To repeat the single-channel QRS onset/offset identification, or determination, for a multitude of channels may involve considerable computational expense. The exemplary systems and methods disclosed herein may be described as enabling automatic determination of QRS onset and offset time values based on a single, composite time-varying signal which at any time-instant represents the dispersion (e.g., standard deviation) of amplitudes across all, or at least more than one, cardiac signals/ECG channels on the body-surface of the patient.

[0018] Further, the exemplary systems and methods may isolate the QRS complex using a dispersion measurement across all, or more than one, electrodes determined to be in, or make, contact with the body-surface (e.g., based on impedance) at each sampled time point. The dispersion measurement may refer to any measurement that takes advantage of the differences in signal magnitude and polarity across multiple channels collected during body surface mapping—including but not limited to—standard deviation, coefficient of variance, range, and absolute deviations, often taken about a measure of central tendency like mean, mode or median. Still further, the exemplary systems and methods may establish a threshold criterion to eliminate noisy channels from electrocardiographic measurements from the body surface mapping to, e.g., bolster the accuracy and confidence of the determinations.

[0019] The exemplary systems and methods may be used for applications where cardiac depolarization onset and offset time values are identified from data involving a multitude of ECG/far-field EGMs for determination of activation times (e.g., body-surface mapping) and mapping from a constellation catheter-basket or a balloon with multiple unipolar electrodes inside a heart cavity (e.g. RV or LV) for beat-by-beat, simultaneous mapping of electrical activity.

[0020] In one or more embodiments, the exemplary systems and methods may be described as using QRS peak detection using a dispersion measurement. At each time point, the dispersion measurement may be calculated across all cardiac signals, or ECG channels. The maximum dispersion in a local window of time may be calculated, and a threshold dependent on the maximum dispersion value (e.g., a local maximum dispersion value within a time window) may be calculated. The first point in the window where the signal reaches the locally-determined dispersion threshold may be used to determine the peak, or maximum value, of the QRS complex. It may be described that the exemplary systems and methods may obtain, or get, a window, or local area of time, find a maximum dispersion within the window, and set a threshold based on the maximum (e.g., 75% of max). At the first point where this threshold is reached in the window, the next local maximum may be determined, which will be identified as the QRS peak or maximum dispersion value for the window.

[0021] In another embodiment, which does not use a threshold, the signals from the electrodes are high-pass filtered to reduce T-wave and P-wave activity, the single

dispersion signal may be generated from the filtered signals, and the signal acquisition may be performed in short time periods (e.g., 1000 milliseconds) to include only one or two cardiac cycles whose peak QRS can be identified by finding the maximum amplitude of the dispersion signal.

[0022] In one or more embodiments, the exemplary systems and methods may be described as detecting QRS complex onset and offset using a dispersion measurement. Using the previously-determined QRS peaks, each QRS complex onset may be determined by identifying the time of minimum dispersion or of a selected percentage of the minimum dispersion (e.g., with 10% of the minimum dispersion) in a local time frame (e.g., window of time, selected time period, etc.) prior to each QRS peak. Similarly, each QRS offset may be determined by identifying the time of minimum dispersion or of a selected percentage of the minimum dispersion (e.g., with 10% of the minimum dispersion) in a local time frame (e.g., window of time, selected time period, etc.) after each QRS peak.

[0023] In one or more embodiments, the exemplary systems and methods may be described as being configured to eliminate noisy channels. To reduce the number of noisy channels, a software-level threshold of QRS complex amplitude may be established. After the QRS onset and offset are determined, the absolute (e.g., peak-to-peak) amplitude between the onset and offset values may be calculated for each cardiac signal or ECG channel. A threshold may then be established at some fraction of the median absolute amplitude value. Any cardiac signal or ECG channel with an amplitude less than the determined, or calculated, threshold may not be used for computation of activation times.

[0024] Automatic signal feature isolation and noise reduction for body surface mapping calculations may be described as streamlining calculations for various cardiac patient evaluation and/or therapy evaluation tools. As a result, patient-level data analysis may be more accurate and consistent and may further simplify the role of the physician, which may increase the efficiency of patient assessment. Further, the exemplary systems and methods may be described as being useful for automated processing of electrical activation maps that can be integrated with other automated routines in other devices.

[0025] In one or more embodiments, the exemplary systems and methods may be used to adjust cardiac therapy configurations and determine onset, offset and duration of QRS complexes and the resulting heterogeneity of activation resulting from the new cardiac device configuration. Further, any changes to the electrical heterogeneity from a baseline condition may be displayed on a graphical user interface. Further, users may select one or more cardiac therapy configurations based on changes in heterogeneity, and the exemplary systems and methods may assist the users in doing so.

[0026] The above summary is not intended to describe each embodiment or every implementation of the present disclosure. A more complete understanding will become apparent and appreciated by referring to the following detailed description and claims taken in conjunction with the accompanying drawings.

BRIEF DESCRIPTION OF THE DRAWINGS

[0027] FIG. 1 is a diagram of an exemplary system including electrode apparatus, display apparatus, and computing apparatus.

[0028] FIGS. 2-3 are diagrams of exemplary external electrode apparatus for measuring torso-surface potentials.

[0029] FIG. 4 is a block diagram of an exemplary method for determining QRS duration from a plurality of cardiac signals.

[0030] FIG. 5 is a more detailed block diagram of determining QRS duration from a dispersion signal of FIG. 4.

[0031] FIG. 6 is an exemplary graph of a plurality of cardiac signals monitored from surface electrodes.

[0032] FIG. 7 is an exemplary graph of a subset of the plurality of cardiac signals of FIG. 6.

[0033] FIG. 8 is an exemplary graph of a cardiac signal including low frequency noise.

[0034] FIG. 9 is an exemplary graph of a dispersion signal overlaid over a plurality of cardiac signals that were unfiltered and an exemplary graph of a dispersion signal overlaid over the plurality of cardiac signals after filtering.

[0035] FIG. 10 is an exemplary graph of a dispersion signal overlaid over the subset of the plurality of cardiac signals of FIG. 7.

[0036] FIG. 11 is an exemplary graph of the dispersion signal of FIG. 10.

[0037] FIG. 12 is an exemplary graph of a portion of the dispersion signal of FIG. 11 including identification of a maximum, or peak, dispersion value.

[0038] FIG. 13 is the exemplary graph of the portion of the dispersion signal of

[0039] FIG. 12 including identification of an onset dispersion value prior to the maximum value corresponding to the onset, or onset time value, of the QRS duration.

[0040] FIG. 14 is the exemplary graph of the portion of the dispersion signal of

[0041] FIG. 12 including identification of an offset dispersion value after the maximum value corresponding to the offset, or offset time value, of the QRS duration.

[0042] FIG. 15 is an exemplary graph of a plurality of cardiac signals including a plurality of QRS complexes and identified QRS onsets and offsets using the exemplary methods and systems described herein.

[0043] FIG. 16 is an exemplary graph of a plurality of cardiac signals including a single QRS complex, an identified QRS onset, and an identified offset using the exemplary methods and systems described herein.

[0044] FIG. 17 is an exemplary graph of the differences of standard deviation of activation-times (SDAT) generated using manually determined QRS durations and QRS durations determined by the exemplary methods and systems described herein.

[0045] FIG. 18 is an exemplary graph of the differences of average left surrogate electrical activation times (LVAT) generated using manually determined QRS durations and QRS durations determined by the exemplary methods and systems described herein.

[0046] FIG. 19 is a diagram of an exemplary system including an exemplary implantable medical device (IMD).

[0047] FIG. 20A is a diagram of the exemplary IMD of FIG. 20.

[0048] FIG. 20B is a diagram of an enlarged view of a distal end of the electrical lead disposed in the left ventricle of FIG. 20A.

[0049] FIG. 21A is a block diagram of an exemplary IMD, e.g., of the systems of FIGS. 19-20.

[0050] FIG. 21B is another block diagram of an exemplary IMD (e.g., an implantable pulse generator) circuitry and associated leads employed in the systems of FIGS. 19-20.

DETAILED DESCRIPTION OF EXEMPLARY EMBODIMENTS

[0051] In the following detailed description of illustrative embodiments, reference is made to the accompanying figures of the drawing which form a part hereof, and in which are shown, by way of illustration, specific embodiments which may be practiced. It is to be understood that other embodiments may be utilized and structural changes may be made without departing from (e.g., still falling within) the scope of the disclosure presented hereby.

[0052] Exemplary systems and methods shall be described with reference to FIGS. 1-21. It will be apparent to one skilled in the art that elements or processes from one embodiment may be used in combination with elements or processes of the other embodiments, and that the possible embodiments of such methods and systems using combinations of features set forth herein is not limited to the specific embodiments shown in the Figures and/or described herein. Further, it will be recognized that the embodiments described herein may include many elements that are not necessarily shown to scale. Still further, it will be recognized that timing of the processes and the size and shape of various elements herein may be modified but still fall within the scope of the present disclosure, although certain timings, one or more shapes and/or sizes, or types of elements, may be advantageous over others.

[0053] Cardiac electrical activation times can be detected or estimated in proximity of a reference location (e.g., which can be a chosen location for the left ventricle lead during implant) using unipolar electrocardiogram (ECG) recordings. Such electrical activation times may be measured and displayed, or conveyed, to an implanter by a system which acquires the ECG signals and generates the metric of electrical activation times (e.g., depolarization) measured from various ECG locations.

[0054] Various exemplary systems, methods, and interfaces may be configured to use electrode apparatus including external electrodes, display apparatus, and computing apparatus to noninvasively assist a user (e.g., a physician) in the evaluation of a patient's condition and/or cardiac therapy being performed on a patient. An exemplary system **100** including electrode apparatus **110**, display apparatus **130**, and computing apparatus **140** is depicted in FIG. 1.

[0055] The electrode apparatus **110** as shown includes a plurality of electrodes incorporated, or included, within a band wrapped around the chest, or torso, of a patient **14**. The electrode apparatus **110** is operatively coupled to the computing apparatus **140** (e.g., through one or wired electrical connections, wirelessly, etc.) to provide electrical signals from each of the electrodes to the computing apparatus **140** for analysis, evaluation, etc. Exemplary electrode apparatus may be described in U.S. Pat. No. 9,320,446 entitled "Bio-electric Sensor Device and Methods" and issued on Apr. 26, 2016, which is incorporated herein by reference in its entirety. Further, exemplary electrode apparatus **110** will be described in more detail in reference to FIGS. 2-3.

[0056] Although not described herein, the exemplary system **100** may further include imaging apparatus. The imaging apparatus may be any type of imaging apparatus configured to image, or provide images of, at least a portion of

the patient in a noninvasive manner. For example, the imaging apparatus may not use any components or parts that may be located within the patient to provide images of the patient except noninvasive tools such as contrast solution. It is to be understood that the exemplary systems, methods, and interfaces described herein may further use imaging apparatus to provide noninvasive assistance to a user (e.g., a physician) to locate or select a pacing electrode or pacing vector proximate the patient's heart in conjunction with the evaluation of cardiac therapy.

[0057] For example, the exemplary systems, methods, and interfaces may provide image guided navigation that may be used to navigate leads including electrodes, leadless electrodes, wireless electrodes, catheters, etc., within the patient's body while also providing noninvasive cardiac therapy evaluation including determining location information (e.g., location information for the electrodes). Exemplary systems and methods that use imaging apparatus and/or electrode apparatus may be described in U.S. patent application Ser. No. 13/916,353 filed on Jun. 12, 2013 and entitled "Implantable Electrode Location Selection," U.S. patent application Ser. No. 13/916,377 filed on Jun. 12, 2013 and entitled "Implantable Electrode Location Selection," U.S. patent application Ser. No. 14/227,955 filed on Mar. 27, 2014 and entitled "Systems, Methods, and Interfaces for Identifying Effective Electrodes," U.S. patent application Ser. No. 14/227,919 filed on Mar. 27, 2014 and entitled "Systems, Methods, and Interfaces for Identifying Optical Electrical Vectors," each of which is incorporated herein by reference in its entirety.

[0058] Exemplary imaging apparatus may be configured to capture x-ray images and/or any other alternative imaging modality. For example, the imaging apparatus may be configured to capture images, or image data, using isocentric fluoroscopy, bi-plane fluoroscopy, ultrasound, computed tomography (CT), multi-slice computed tomography (MSCT), magnetic resonance imaging (MM), high frequency ultrasound (HIFU), optical coherence tomography (OCT), intra-vascular ultrasound (IVUS), two dimensional (2D) ultrasound, three dimensional (3D) ultrasound, four dimensional (4D) ultrasound, intraoperative CT, intraoperative MRI, etc. Further, it is to be understood that the imaging apparatus may be configured to capture a plurality of consecutive images (e.g., continuously) to provide video frame data. In other words, a plurality of images taken over time using the imaging apparatus may provide video frame, or motion picture, data. Additionally, the images may also be obtained and displayed in two, three, or four dimensions. In more advanced forms, four-dimensional surface rendering of the heart or other regions of the body may also be achieved by incorporating heart data or other soft tissue data from a map or from pre-operative image data captured by MM, CT, or echocardiography modalities. Image datasets from hybrid modalities, such as positron emission tomography (PET) combined with CT, or single photon emission computer tomography (SPECT) combined with CT, could also provide functional image data superimposed onto anatomical data, e.g., to be used to navigate treatment apparatus proximate target locations within the heart or other areas of interest.

[0059] Systems and/or imaging apparatus that may be used in conjunction with the exemplary systems and method described herein are described in U.S. Pat. App. Pub. No. 2005/0008210 to Evron et al. published on Jan. 13, 2005,

U.S. Pat. App. Pub. No. 2006/0074285 to Zarkh et al. published on Apr. 6, 2006, U.S. Pat. App. Pub. No. 2011/0112398 to Zarkh et al. published on May 12, 2011, U.S. Pat. App. Pub. No. 2013/0116739 to Brada et al. published on May 9, 2013, U.S. Pat. No. 6,980,675 to Evron et al. issued on Dec. 27, 2005, U.S. Pat. No. 7,286,866 to Okerlund et al. issued on Oct. 23, 2007, U.S. Pat. No. 7,308,297 to Reddy et al. issued on Dec. 11, 2011, U.S. Pat. No. 7,308,299 to Burrell et al. issued on Dec. 11, 2011, U.S. Pat. No. 7,321,677 to Evron et al. issued on Jan. 22, 2008, U.S. Pat. No. 7,346,381 to Okerlund et al. issued on Mar. 18, 2008, U.S. Pat. No. 7,454,248 to Burrell et al. issued on Nov. 18, 2008, U.S. Pat. No. 7,499,743 to Vass et al. issued on Mar. 3, 2009, U.S. Pat. No. 7,565,190 to Okerlund et al. issued on Jul. 21, 2009, U.S. Pat. No. 7,587,074 to Zarkh et al. issued on Sep. 8, 2009, U.S. Pat. No. 7,599,730 to Hunter et al. issued on Oct. 6, 2009, U.S. Pat. No. 7,613,500 to Vass et al. issued on Nov. 3, 2009, U.S. Pat. No. 7,742,629 to Zarkh et al. issued on Jun. 22, 2010, U.S. Pat. No. 7,747,047 to Okerlund et al. issued on Jun. 29, 2010, U.S. Pat. No. 7,778,685 to Evron et al. issued on Aug. 17, 2010, U.S. Pat. No. 7,778,686 to Vass et al. issued on Aug. 17, 2010, U.S. Pat. No. 7,813,785 to Okerlund et al. issued on Oct. 12, 2010, U.S. Pat. No. 7,996,063 to Vass et al. issued on Aug. 9, 2011, U.S. Pat. No. 8,060,185 to Hunter et al. issued on Nov. 15, 2011, and U.S. Pat. No. 8,401,616 to Verard et al. issued on Mar. 19, 2013, each of which is incorporated herein by reference in its entirety.

[0060] The display apparatus **130** and the computing apparatus **140** may be configured to display and analyze data such as, e.g., electrical signals (e.g., electrocardiogram data), cardiac information representative of at least one of mechanical cardiac functionality and electrical cardiac functionality, etc. Cardiac information may include, e.g., electrical heterogeneity information or electrical dyssynchrony information, surrogate electrical activation information or data, etc. that is generated using electrical signals gathered, monitored, or collected, using the electrode apparatus **110**. In at least one embodiment, the computing apparatus **140** may be a server, a personal computer, or a tablet computer. The computing apparatus **140** may be configured to receive input from input apparatus **142** and transmit output to the display apparatus **130**. Further, the computing apparatus **140** may include data storage that may allow for access to processing programs or routines and/or one or more other types of data, e.g., for driving a graphical user interface configured to noninvasively assist a user in evaluating a pacing location (e.g., the location of an implantable electrode used for pacing, the location of pacing therapy delivered by a particular pacing vector, etc.).

[0061] The computing apparatus **140** may be operatively coupled to the input apparatus **142** and the display apparatus **130** to, e.g., transmit data to and from each of the input apparatus **142** and the display apparatus **130**. For example, the computing apparatus **140** may be electrically coupled to each of the input apparatus **142** and the display apparatus **130** using, e.g., analog electrical connections, digital electrical connections, wireless connections, bus-based connections, network-based connections, internet-based connections, etc. As described further herein, a user may provide input to the input apparatus **142** to manipulate, or modify, one or more graphical depictions displayed on the display apparatus **130** and to view and/or select one or more pieces of information related to the cardiac therapy.

[0062] Although as depicted the input apparatus **142** is a keyboard, it is to be understood that the input apparatus **142** may include any apparatus capable of providing input to the computing apparatus **140** to perform the functionality, methods, and/or logic described herein. For example, the input apparatus **142** may include a mouse, a trackball, a touchscreen (e.g., capacitive touchscreen, a resistive touchscreen, a multi-touch touchscreen, etc.), etc. Likewise, the display apparatus **130** may include any apparatus capable of displaying information to a user, such as a graphical user interface **132** including cardiac information, textual instructions, graphical depictions of electrical activation information, graphical depictions of anatomy of a human heart, images or graphical depictions of the patient's heart, graphical depictions of locations of one or more electrodes, graphical depictions of a human torso, images or graphical depictions of the patient's torso, graphical depictions or actual images of implanted electrodes and/or leads, etc. Further, the display apparatus **130** may include a liquid crystal display, an organic light-emitting diode screen, a touchscreen, a cathode ray tube display, etc.

[0063] The processing programs or routines stored and/or executed by the computing apparatus **140** may include programs or routines for computational mathematics, matrix mathematics, dispersion determinations (e.g. standard deviations, variances, ranges, interquartile ranges, mean absolute differences, average absolute deviations, etc.), filtering algorithms, maximum value determinations, minimum value determinations, threshold determinations, moving windowing algorithms, decomposition algorithms, compression algorithms (e.g., data compression algorithms), calibration algorithms, image construction algorithms, signal processing algorithms (e.g., various filtering algorithms, Fourier transforms, fast Fourier transforms, etc.), standardization algorithms, comparison algorithms, vector mathematics, or any other processing required to implement one or more exemplary methods and/or processes described herein. Data stored and/or used by the computing apparatus **140** may include, for example, electrical signal/waveform data from the electrode apparatus **110**, dispersions signals, windowed dispersions signals, parts or portions of various signals, electrical activation times from the electrode apparatus **110**, graphics (e.g., graphical elements, icons, buttons, windows, dialogs, pull-down menus, graphic areas, graphic regions, 3D graphics, etc.), graphical user interfaces, results from one or more processing programs or routines employed according to the disclosure herein (e.g., electrical signals, cardiac information, etc.), or any other data that may be necessary for carrying out the one and/or more processes or methods described herein.

[0064] In one or more embodiments, the exemplary systems, methods, and interfaces may be implemented using one or more computer programs executed on programmable computers, such as computers that include, for example, processing capabilities, data storage (e.g., volatile or non-volatile memory and/or storage elements), input devices, and output devices. Program code and/or logic described herein may be applied to input data to perform functionality described herein and generate desired output information. The output information may be applied as input to one or more other devices and/or methods as described herein or as would be applied in a known fashion.

[0065] The one or more programs used to implement the systems, methods, and/or interfaces described herein may be

provided using any programmable language, e.g., a high level procedural and/or object orientated programming language that is suitable for communicating with a computer system. Any such programs may, for example, be stored on any suitable device, e.g., a storage media, that is readable by a general or special purpose program running on a computer system (e.g., including processing apparatus) for configuring and operating the computer system when the suitable device is read for performing the procedures described herein. In other words, at least in one embodiment, the exemplary systems, methods, and/or interfaces may be implemented using a computer readable storage medium, configured with a computer program, where the storage medium so configured causes the computer to operate in a specific and predefined manner to perform functions described herein. Further, in at least one embodiment, the exemplary systems, methods, and/or interfaces may be described as being implemented by logic (e.g., object code) encoded in one or more non-transitory media that includes code for execution and, when executed by a processor, is operable to perform operations such as the methods, processes, and/or functionality described herein.

[0066] The computing apparatus 140 may be, for example, any fixed or mobile computer system (e.g., a controller, a microcontroller, a personal computer, mini computer, tablet computer, etc.). The exact configuration of the computing apparatus 140 is not limiting, and essentially any device capable of providing suitable computing capabilities and control capabilities (e.g., graphics processing, etc.) may be used. As described herein, a digital file may be any medium (e.g., volatile or non-volatile memory, a CD-ROM, a punch card, magnetic recordable medium such as a disk or tape, etc.) containing digital bits (e.g., encoded in binary, trinary, etc.) that may be readable and/or writable by computing apparatus 140 described herein. Also, as described herein, a file in user-readable format may be any representation of data (e.g., ASCII text, binary numbers, hexadecimal numbers, decimal numbers, graphically, etc.) presentable on any medium (e.g., paper, a display, etc.) readable and/or understandable by a user. Further, the computing apparatus 140 may be described as including one or more processors or processing circuitry configured to execute the illustrative processes, methods, and algorithms described herein.

[0067] In view of the above, it will be readily apparent that the functionality as described in one or more embodiments according to the present disclosure may be implemented in any manner as would be known to one skilled in the art. As such, the computer language, the computer system, or any other software/hardware which is to be used to implement the processes described herein shall not be limiting on the scope of the systems, processes or programs (e.g., the functionality provided by such systems, processes or programs) described herein.

[0068] Electrical activation times of the patient's heart may be useful to evaluate a patient's cardiac condition and/or cardiac therapy being delivered to a patient. Surrogate electrical activation information or data of one or more regions of a patient's heart may be monitored, or determined, using electrode apparatus 110 as shown in FIG. 1 and in FIG. 2-3. The exemplary electrode apparatus 110 may be configured to measure body-surface potentials of a patient 14 and, more particularly, torso-surface potentials of a patient 14. As shown in FIG. 2, the exemplary electrode apparatus 110 may include a set, or array, of electrodes 112,

a strap 113, and interface/amplifier circuitry 116. The electrodes 112 may be attached, or coupled, to the strap 113 and the strap 113 may be configured to be wrapped around the torso of a patient 14 such that the electrodes 112 surround the patient's heart. As further illustrated, the electrodes 112 may be positioned around the circumference of a patient 14, including the posterior, lateral, posterolateral, anterolateral, and anterior locations of the torso of a patient 14.

[0069] Further, the electrodes 112 may be electrically connected to interface/amplifier circuitry 116 via wired connection 118. The interface/amplifier circuitry 116 may be configured to amplify the signals from the electrodes 112 and provide the signals to the computing apparatus 140. Other exemplary systems may use a wireless connection to transmit the signals sensed by electrodes 112 to the interface/amplifier circuitry 116 and, in turn, the computing apparatus 140, e.g., as channels of data. For example, the interface/amplifier circuitry 116 may be electrically coupled to each of the computing apparatus 140 and the display apparatus 130 using, e.g., analog electrical connections, digital electrical connections, wireless connections, bus-based connections, network-based connections, internet-based connections, etc.

[0070] Although in the example of FIG. 2 the electrode apparatus 110 includes a strap 113, in other examples any of a variety of mechanisms, e.g., tape or adhesives, may be employed to aid in the spacing and placement of electrodes 112. In some examples, the strap 113 may include an elastic band, strip of tape, or cloth. In other examples, the electrodes 112 may be placed individually on the torso of a patient 14. Further, in other examples, electrodes 112 (e.g., arranged in an array) may be part of, or located within, patches, vests, and/or other manners of securing the electrodes 112 to the torso of the patient 14.

[0071] The electrodes 112 may be configured to surround the heart of the patient 14 and record, or monitor, the electrical signals associated with the depolarization and repolarization of the heart after the signals have propagated through the torso of a patient 14. Each of the electrodes 112 may be used in a unipolar configuration to sense the torso-surface potentials that reflect the cardiac signals. The interface/amplifier circuitry 116 may also be coupled to a return or indifferent electrode (not shown) that may be used in combination with each electrode 112 for unipolar sensing. In some examples, there may be about 12 to about 50 electrodes 112 spatially distributed around the torso of patient. Other configurations may have more or fewer electrodes 112.

[0072] The computing apparatus 140 may record and analyze the torso-surface potential signals sensed by electrodes 112 and amplified/conditioned by the interface/amplifier circuitry 116. The computing apparatus 140 may be configured to analyze the signals from the electrodes 112 to provide surrogate electrical activation information or data such as surrogate cardiac electrical activation times, e.g., representative of actual, or local, electrical activation times of one or more regions of the patient's heart as will be further described herein. For example, electrical signals measured at the left anterior surface location of a patient's torso may be representative, or surrogates, of electrical signals of the left anterior left ventricle region of the patient's heart, electrical signals measured at the left lateral surface location of a patient's torso may be representative, or surrogates, of electrical signals of the left lateral left

ventricle region of the patient's heart, electrical signals measured at the left posterolateral surface location of a patient's torso may be representative, or surrogates, of electrical signals of the posterolateral left ventricle region of the patient's heart, and electrical signals measured at the posterior surface location of a patient's torso may be representative, or surrogates, of electrical signals of the posterior left ventricle region of the patient's heart. In one or more embodiments, measurement of activation times can be performed by measuring the period of time between an onset of cardiac depolarization (e.g., onset of QRS complex) and the next onset of cardiac depolarization. In one or more embodiments, measurement of activation times can be performed by measuring the period of time between an onset of cardiac depolarization (e.g., onset of QRS complex) and conclusion of cardiac repolarization (e.g., offset of QRS complex). In one or more embodiments, measurement of activation times can be performed by picking an appropriate fiducial point (e.g., peak values, minimum values, minimum slopes, maximum slopes, zero crossings, threshold crossings, etc. of a near or far-field EGM) and measuring time between fiducial points (e.g., within the electrical activity).

[0073] Additionally, the computing apparatus 140 may be configured to provide graphical user interfaces depicting the surrogate electrical activation times obtained using the electrode apparatus 110. Exemplary systems, methods, and/or interfaces may noninvasively use the electrical information collected using the electrode apparatus 110 to evaluate a patient's cardiac condition and/or cardiac therapy being delivered to the patient.

[0074] FIG. 3 illustrates another exemplary electrode apparatus 110 that includes a plurality of electrodes 112 configured to surround the heart of the patient 14 and record, or monitor, the electrical signals associated with the depolarization and repolarization of the heart after the signals have propagated through the torso of the patient 14. The electrode apparatus 110 may include a vest 114 upon which the plurality of electrodes 112 may be attached, or to which the electrodes 112 may be coupled. In at least one embodiment, the plurality, or array, of electrodes 112 may be used to collect electrical information such as, e.g., surrogate electrical activation times. Similar to the electrode apparatus 110 of FIG. 2, the electrode apparatus 110 of FIG. 3 may include interface/amplifier circuitry 116 electrically coupled to each of the electrodes 112 through a wired connection 118 and be configured to transmit signals from the electrodes 112 to computing apparatus 140. As illustrated, the electrodes 112 may be distributed over the torso of a patient 14, including, for example, the anterior, lateral, posterolateral, anterolateral, and posterior surfaces of the torso of the patient 14.

[0075] The vest 114 may be formed of fabric with the electrodes 112 attached to the fabric. The vest 114 may be configured to maintain the position and spacing of electrodes 112 on the torso of the patient 14. Further, the vest 114 may be marked to assist in determining the location of the electrodes 112 on the surface of the torso of the patient 14. In some examples, there may be about 25 to about 256 electrodes 112 distributed around the torso of the patient 14, though other configurations may have more or fewer electrodes 112.

[0076] As described herein, the electrode apparatus 110 may be configured to measure electrical information (e.g., electrical signals) representing different regions of a

patient's heart. For example, activation times of different regions of a patient's heart can be approximated from surface electrocardiogram (ECG) activation times measured using surface electrodes in proximity to surface areas corresponding to the different regions of the patient's heart.

[0077] The exemplary systems, methods, and interfaces may be used to provide noninvasive assistance to a user in the evaluation of a patient's cardiac health or status, and/or the valuation of cardiac therapy (e.g., cardiac therapy being presently-delivered to a patient during implantation or after implantation). Further, the exemplary systems, methods, and interfaces may be used to assist a user in the configuration of the cardiac therapy being delivered to a patient.

[0078] Exemplary method 200 depicted in FIG. 4 may be used to determine onsets and offsets of QRS complexes from a plurality of cardiac signals. The exemplary method 200 may be generally described to be used in the evaluation of a patient's cardiac health and/or cardiac therapy being delivered to the patient.

[0079] The exemplary method 200 may include monitoring, or measuring, electrical activity using a plurality of electrodes 202. As described herein, the plurality of electrodes may include external, surface electrodes coupled to, or attached, to patient's skin, e.g., using the apparatus and systems described herein with respect to FIGS. 1-3. Additionally, the plurality of electrodes may include implantable electrodes implanted in one or more locations with a patient, e.g., using the apparatus and systems described herein with respect to FIGS. 19-21. Still further, the plurality of electrodes may further include temporarily-implanted (e.g., non-permanent, not long term, etc.) electrodes such as, e.g., electrodes located in the cavity of a patient's heart using basket catheter or the like, electrodes located outside the heart but around the heart (e.g., epicardial space) using a "sock mapping" tool or the like. Further the electrodes, or electrode apparatus, may include a mix, or combination, of external and implanted electrodes. Regardless of the electrode apparatus utilized, the exemplary method 200 includes monitoring electrical activity, and more specifically, cardiac signals or ECG signals, from at least two electrodes or channels. In other words, electrical signals may be collected from multiple (n) electrodes such as, e.g., body surface electrodes.

[0080] An exemplary graph of a plurality of cardiac signals 300 (e.g., voltage signals measured in millivolts) monitored from surface electrodes is depicted in FIG. 6. The plurality of cardiac signals 300, as shown, were monitored over a time period of at least 5000 milliseconds. Each line represents a different cardiac signal of the plurality of signals. As shown, there are 37 signals (n=37) 302.

[0081] In one or more embodiments, not all of the cardiac signals monitored 202 may be used by the remainder of the exemplary method 200. For example, two or more cardiac signals may be selected from the plurality of cardiac signals to be further used by the method 200. An exemplary graph of a subset of signals 301 of the plurality of cardiac signals 301 of FIG. 6 is depicted in FIG. 7. As shown, there are 32 signals (n=32) 301.

[0082] In one or more embodiments, the exemplary method 200 may filter, or remove, one or more cardiac signals from the plurality of cardiac signals monitored 204. In other words, one or more noisy cardiac signals may be removed from the plurality of cardiac signals. The one or more noisy cardiac signals may be identified by identifying

cardiac signals including, or defining, absolute areas-under-the-curve greater than or equal to a pre-filter threshold value.

[0083] An exemplary graph of a cardiac signal **304** including low frequency noise is depicted in FIG. 8. The graph depicts a common ECG artifact **305**, which may be referred to as low frequency noise. This type of noise may be observed in subcutaneous and cutaneous ECG recording devices such as, e.g., an implanted monitoring device and/or external electrode apparatus. Temporary and/or slight contact problems (e.g., electrodes in contact with tissue) may cause, or generate, a low frequency baseline drift, which, in turn, may cause R-wave sensing processes, or algorithms, to detect the noise as an R-wave and/or fail to sense the following two or three R-waves because, e.g., the noise may be larger than the subsequent R-waves. The exemplary methods and systems described herein may detect the low frequency noise **305** so as to be able to reject, or filter, this type of noise from the plurality of cardiac signals being monitored.

[0084] The exemplary method **200** may filter **204** the plurality of signals using area-under-curve (AUC) to detect the low frequency noise, such that, e.g., when the noise is detected, the monitoring apparatus may stop monitoring, or recording, cardiac signals at least temporally. More specifically, the filter **204** may evaluate the AUC for a selected portion of the cardiac signal and may determine whether the AUC of the selected portion is greater than a threshold value, which may be determined by all of the signal data within the selected portion.

[0085] For example, the AUC **307** of the noise artifact **305** is greater than a threshold AUC value (e.g., a baseline voltage-time value). A first moving window, or selected portion, may be used, which may be about 0.01 to about 3 seconds long prior to the current time point. A baseline signal value (e.g., voltage) may be determined based on the signal within the first moving window, or selected portion. The threshold value may be a median, mean, or other statistic metric of the AUC measurements within the moving window or selected portion.

[0086] In one embodiment, the AUC **307** of the signal within a second moving window, or selected portion, (e.g., voltage) after the current time point may be calculated. Then calculate the AUC by subtracting the baseline values from the first moving window from the values in the second moving window (e.g., $AUC = EGM \text{ voltages in the second window} - \text{baseline EGM voltages in the first window}$). As shown in FIG. 8, the AUC **307** of the low frequency noise **305** may be several times higher than the AUC of typical R-waves or T-waves.

[0087] In another embodiment, a single window up to 5 seconds long of the signal may be evaluated. The AUC **307** may be determined, or calculated, for each signal within this window with opportunity to subtract baseline ECG signal value if desired. A threshold may be selected relative to some statistic of the AUC (e.g., 3 times the median AUC in this window). The signals above the threshold may be removed or “filtered out.”

[0088] Additionally, the same filtering **204** of the plurality of signals **310** using area-under-curve (AUC) to detect the low frequency noise is depicted in FIG. 9 to, e.g., remove signals monitored using external electrodes having to detached or loosely coupled electrodes to improve R-wave timing accuracy. More specifically, an exemplary graph of a plurality of cardiac signals **310** prior to such filtering **204**

and an exemplary graph of a dispersion signal **314** overlaid over the plurality of cardiac signals **312** after such filtering **204** are depicted in FIG. 9. As shown, one or more cardiac signals were removed, or filtered, from the first, or top, graph (“Before Filtering”) to the second, or bottom, graph (“After Filtering”), which may result a better dispersion signal describe further herein and better identification of peak, or maximum, values associated with, or corresponding, to R-waves of QRS complexes within the plurality of cardiac signals.

[0089] The exemplary method **200** may further include generating a dispersion signal **206** from the plurality of cardiac signals monitored **202** and potentially filtered **204**. The dispersion signal may be representative of the dispersion of the plurality of cardiac signals over time. The dispersion signal may be the standard deviation of the plurality of cardiac signals over time. In other embodiments, the dispersion signal may be variance, a coefficient of variance, range, mean absolute deviation, a measure of central tendency (e.g., like the mean), interquartile deviation of amplitudes, median absolute deviation taken about a measure of central tendency like the mean, median, mode, and/or another statistical measure of the plurality of cardiac signals over time.

[0090] An exemplary graph of a dispersion signal **320** overlaid over the subset of the plurality of cardiac signals **301** of FIG. 7 is depicted in FIG. 10. The standard deviation across all signals **301** in the subset at each time point was calculated to generate the dispersion signal **320**. As shown, the thin lines are the signals from the subset **301**, and the thick black line is the dispersion signal **320** (e.g., a standard deviation) that was calculated across all signals at each time point.

[0091] An exemplary expanded graph of the dispersion signal **320** of FIG. 10 without the plurality of cardiac signals **301** is depicted in FIG. 11, which will be further described herein with reference to the user of the dispersion signal **320**.

[0092] The dispersion signal may be used by the method **200** to determine a QRS onset time value and a QRS offset time value **208** for each of at least one QRS complex within the monitored electrical activity **202** based on the generated dispersion signal **206**. In other words, the generated dispersion signal may be used to determine onset time value and a QRS offset time value **208**, which may be later used to conduct various measurements such as, e.g., activation time measurements, using the plurality of monitored cardiac signals. The QRS onset time value may be described as a time value, or point in time, that is indicative of the start, or beginning, of the QRS complex of the cardiac signals when taking into consideration the plurality of monitored cardiac signals. The QRS offset time value may be described as a time value, or point in time, that is indicative of the end of the QRS complex of the cardiac signals when taking into consideration the plurality of monitored cardiac signals. The QRS onset and offset time values may be used to calculate one or more values or metrics from or using the plurality of cardiac signals as will be described further herein.

[0093] Determination of the QRS onset time value and the QRS offset time value **208** is further depicted in the flow diagram of FIG. 5. As shown, determination of the QRS onset time value and the QRS offset time value **208** may include windowing the dispersion signal **230**, filtering the windowed signal **232**, determining a maximum dispersion value **234**, determining a minimum dispersion values **236**

about the maximum dispersion value, and determining the QRS onset time value and the QRS offset time value using the onset and offset dispersion values about the maximum dispersion value **238** (e.g., the onset and offset dispersion values may be related to the minimum dispersion values about the maximum dispersion value **238** such as a percentage thereof).

[0094] Determination of the QRS onset time value and the QRS offset time value **208** may be begin may selecting a portion **321** of the dispersion signal **320** over a selected time period **322**. The selected time period **322** may be described as a moving time window, which may move along (representative by arrow **325**) along the dispersion signal **320** such that the analysis may be performed along the dispersion signal **320** in discrete slices or portions **321** of the dispersion signal **320**. The QRS onset and offset time values for each QRS complex may be determined, or identified, within the monitored electrical activity of the selected time period **322** based on the selected portion **321** of the dispersion signal **320**.

[0095] The selected time period, or moving time window, may be less than or equal to about 5 seconds. In other embodiments, the selected time period, or moving time window, may be less than or equal to about 4.5 seconds, less than or equal to about 4 seconds, less than or equal to about 3.5 seconds, less than or equal to about 3 seconds, less than or equal to about 2.5 seconds, less than or equal to about 2 seconds, less than or equal to about 1 second, etc. and/or greater than or equal to about 0.25 seconds, greater than or equal to about 0.5 seconds, greater than or equal to about 1.25 seconds, greater than or equal to about 2.75 seconds, greater than or equal to about 3.25 seconds, greater than or equal to about 3.75 seconds, greater than or equal to about 4.25 second, etc. Further, in one or more embodiments, the selected time period, or moving time window, may be selected, or adjusted, based on the patient's heart rate so as to capture, e.g., one QRS complex per selected time period, two or more QRS complexes per selected time period, etc. In at least one embodiment, the selected time period, or moving time window, may be defined to capture a single QRS complex in most instances.

[0096] In the example depicted in FIG. 11, a 1000 millisecond selected time period, or moving time window, **322** is used that shifts (represented by arrow **325**) as individual dispersion values are evaluated.

[0097] Prior to determining a maximum dispersion value within the selected time period **322**, the exemplary method **208** may selectively filter the windowed signal **232** using one or more filtering techniques or processes to, e.g., avoid signal artifacts that do not correspond to the QRS complex such as the P-wave, the T-wave, and noise. To do so, in one embodiment, a threshold value **327** may be determined, and only the dispersion values greater than threshold value **327** may be considered when determining a maximum dispersion value within the selected time period **322**. As shown in FIG. 11, the threshold value is represented by the horizontal line **327** within the selected time period **322**. The threshold value may be based one or more statistics or metrics designed to allow peak, or maximum, dispersion values to remain while eliminating dispersion values that may not be relevant when identifying peak, or maximum, values. The threshold value may be based on, or defined by, a selected percentage of the maximum dispersion value of the dispersion signal **320**. In other words, the dispersion signal may be filtered **232** to

remove one or more values of the dispersion signal **320** that are less than or equal to a selected percentage of the maximum dispersion value of the dispersion signal. The selected percentage may be greater than or equal to about 50%, greater than or equal to about 65%, greater than or equal to about 75%, greater than or equal to about 85%, etc. and/or less than or equal to about 90%, less than or equal to about 80%, less than or equal to about 70%, less than or equal to about 60%, etc.

[0098] In other embodiments, high pass filtering may be used to filter the dispersion signal **320** to remove one or more dispersion values of the dispersion signal corresponding to at least one of a T-wave and a P-wave in the plurality of cardiac signals. Further, the filtering the plurality of cardiac signals through a filter (e.g., high-pass) may to remove the lower frequency components like P- and T-waves, while preserving the higher frequency QRS. Additional illustrative filtering processes that may be used with the illustrative systems and methods described herein may be described in U.S. patent application Ser. No. 15/802,615 entitled "Detection of Noise Signals in Cardiac Signals" and filed on Nov. 3, 2017, which is incorporated by reference herein in its entirety.

[0099] For use in determining QRS onset and offset time values **208**, a maximum dispersion value may be determined **234** within the selected time period **322**. An exemplary graph of the selected portion **321** of the dispersion signal **320** of FIG. 11 within the selected time period **322** is depicted in FIG. 12. In this embodiment, the maximum dispersion value **330** may be determined by analyzing each of the dispersion values after the first dispersion value exceeds the threshold **327** for a selected time period. In other embodiments, all dispersion values over the threshold **327** may be evaluated. As show, the selected time period is 120 milliseconds. In another embodiment, the selected time period may be 200 milliseconds. In other embodiments, the selected time period may be greater than or equal to about 80 milliseconds (ms), greater than or equal to about 100 ms, greater than or equal to about 140 ms, greater than or equal to about 160 ms, greater than or equal to about 180 ms, greater than or equal to about 220 ms, greater than or equal to about 240 ms, etc. and/or less than or equal to about 250 ms, less than or equal to about 230 ms, less than or equal to about 210 ms, less than or equal to about 190 ms, less than or equal to about 150 ms, less than or equal to about 130 ms, etc.

[0100] As shown, the maximum dispersion value **330** is determined. Additionally, as shown, a blanking period **331** extends for a selected blanking time period beyond the determined maximum dispersion value **330**, within which no further peak, or maximum, dispersion values will be identified (e.g., to avoid noise, T-waves, etc.). In this example, the blanking time period **321** is 200 milliseconds. In other embodiments, the blanking time period **321** may be greater than or equal to about 50 milliseconds, greater than or equal to about 150 milliseconds, greater than or equal to about 300 milliseconds, greater than or equal to about 400 milliseconds, etc. and/or less than or equal to about 500 milliseconds, less than or equal to about 350 milliseconds, less than or equal to about 250 milliseconds, less than or equal to about 100 milliseconds, etc.

[0101] After the maximum, or peak, dispersion value **330** is determined **234**, onset and offset dispersion values **236** may be determined proximate the maximum dispersion value **330**, which may be used to determine QRS onset and

offset time values. For example, an onset dispersion value **332** may be determined for the QRS onset time value, and an offset dispersion value **334** may be determined for the QRS offset time value. The onset dispersion value **332** corresponding to the QRS onset time value may occur prior to the maximum dispersion value **330**, and the offset dispersion value **334** corresponding to the QRS offset time value may occur after the maximum dispersion value **330**.

[0102] In one or more embodiments, the onset dispersion value **332** of the dispersion signal **320** corresponding to the QRS onset time value may be determined within a first time period **340** before the maximum dispersion value **330** of the dispersion signal **320** as shown in FIG. **13**. Likewise, the offset dispersion value **334** of the dispersion signal **320** corresponding to the QRS offset time value may be determined within a second time period **342** after, or following, the maximum dispersion value **330** of the dispersion signal **320** as shown in FIG. **14**.

[0103] The onset and offset dispersion values prior to and occurring after the maximum dispersion value **330** that is within 10%, 20%, 30%, 40%, 50%, 60%, or another selected percentage or statistical metric, of the minimum dispersion values within the first and second time periods **340**, **342**, respectively, may be used to determine QRS onset and offset time values. For example, the onset dispersion value **332** may be the latest dispersion value of the dispersion signal **320** within the first time period **340** before a maximum dispersion value **330** of the dispersion signal **320** that is less than or equal to 110% of a minimum dispersion value of the dispersion signal **320** within the first time period **322** may be used to determine the QRS onset time value. Further, the offset dispersion value **334** may be the earliest dispersion value of the dispersion signal **320** within the second time period **342** after the maximum dispersion value **330** of the dispersion signal **320** that is less than or equal to 110% of a minimum dispersion value of the dispersion signal **320** within the second time period **342**. The horizontal line **344** represents 10% greater than the minimum dispersion value within the first time period **340**, and the horizontal line **346** represents 10% greater than the minimum dispersion value within the second time period **342**.

[0104] In another embodiment, the minimum dispersion values within the first and second time periods **340**, **342**, may be used as the onset and offset dispersion values respectively.

[0105] The onset and offset dispersion values **332**, **334** may be used to determine the QRS onset and offset time values, respectively, **238** because the time values at which the onset and offset dispersion values **332**, **334** occur within the first and second time periods **340**, **342** may be the QRS onset and QRS offset time values, respectively. In other words, the time at which the onset dispersion value **332** occurs is the QRS onset time value, and the time at which the offset dispersion values **334** occurs is the QRS offset time value.

[0106] An exemplary graph of a plurality of cardiac signals including a plurality of QRS complexes and identified QRS onset time values **350** and offset time values **352** using the exemplary methods and systems described herein is depicted in FIG. **15**, and an exemplary graph of a plurality of cardiac signals including a single QRS complex, an identified QRS onset time value **350**, and an identified QRS offset time value **352** using the exemplary methods and systems described herein is depicted in FIG. **16**.

[0107] The QRS onset and offset time values may be used to determine activations times with each of the QRS complexes of each of the plurality of cardiac signals **212**. More specifically, activation times can be determined using one or more fiducial points within the QRS duration defined by the QRS onset time value and the QRS offset time value of each of the plurality of cardiac signals using the systems, methods, and processes described herein with respect to FIGS. **1-3**.

[0108] Prior to determining activation times **212** based on the cardiac signal and the QRS onset and offset time values, noisy signals may be filtered **210** based on one or more metrics or statistics corresponding to the cardiac signal values between the QRS onset time value and the QRS offset time value within each of the plurality of cardiac signals. For example, an absolute amplitude may be determined between the QRS onset time value and the QRS offset time value for each of the QRS complexes for each of the plurality of cardiac signals, and one or more cardiac signals of the plurality of cardiac signals may be selected, or filtered, based on the absolute amplitude. More specifically, in at least one embodiment, selecting one or more cardiac signals of the plurality of cardiac signals may define an absolute amplitude greater than or equal to a threshold value, and the threshold value is less than the median absolute amplitude of the plurality of cardiac signals. Then, the activation times may be generated based on the selected, or filtered, one or more cardiac signals.

[0109] One or more metrics such as metrics of electrical and mechanical cardiac heterogeneity may be generated **214** based on the activation times. Electrical heterogeneity information and other cardiac therapy information may be described in U.S. Provisional Patent Application No. 61/834,133 entitled "METRICS OF ELECTRICAL DYS-SYNCHRONY AND ELECTRICAL ACTIVATION PATTERNS FROM SURFACE ECG ELECTRODES" and filed on Jun. 12, 2013, which is hereby incorporated by reference in its entirety.

[0110] Electrical heterogeneity information may be defined as information indicative of at least one of mechanical synchrony or dyssynchrony of the heart and/or electrical synchrony or dyssynchrony of the heart. In other words, electrical heterogeneity information may represent a surrogate of actual mechanical and/or electrical functionality. In at least one embodiment, electrical heterogeneity information may be used to determine a surrogate value representative of the maximum left ventricular pressure rate. The left ventricular pressure may be typically monitored invasively with a pressure sensor located in the left ventricular of a patient's heart. As such, the use of electrical heterogeneity information to determine a surrogate value representative of the maximum left ventricular pressure rate may avoid invasive monitoring using a left ventricular pressure sensor.

[0111] In at least one embodiment, the electrical heterogeneity information may include a standard deviation of ventricular activation times corresponding to some or all of the external electrodes, e.g., of the electrode apparatus **110**. Further, regional electrical heterogeneity information may include standard deviations and/or averages of activation times corresponding to electrodes located in certain anatomic areas of the torso. For example, external electrodes on the left side of the torso of a patient may be used to compute regional left electrical heterogeneity information.

[0112] The electrical heterogeneity information may be generated using one or more various systems and/or methods. Electrical heterogeneity information may be generated using an array, or a plurality, of surface electrodes and/or imaging systems as described in U.S. Pat. App. Pub. No. 2012/0283587 A1 published Nov. 8, 2012 and entitled “ASSESSING INTRA-CARDIAC ACTIVATION PATTERNS AND ELECTRICAL DYSSYNCHRONY,” U.S. Pat. App. Pub. No. 2012/0284003 A1 published Nov. 8, 2012 and entitled “ASSESSING INTRA-CARDIAC ACTIVATION PATTERNS”, and U.S. Pat. No. 8,180,428 B2 issued May 15, 2012 and entitled “METHODS AND SYSTEMS FOR USE IN SELECTING CARDIAC PACING SITES,” each of which is incorporated herein by reference in its entirety.

[0113] Electrical heterogeneity information may include one or more metrics or indices. For example, one of the metrics, or indices, of electrical heterogeneity may be a standard deviation of activation-times (SDAT) measured by some or all of the electrodes on the surface of the torso of patient. In some examples, the SDAT may be calculated using the estimated cardiac activation times over the surface of a model heart.

[0114] Another metric, or index, of electrical heterogeneity may be a left standard deviation of surrogate electrical activation times (LVED) monitored by external electrodes located proximate the left side of patient. Further, another metric, or index, of electrical heterogeneity may include an average of surrogate electrical activation times (LVAT) monitored by external electrodes located proximate the left side of patient. The LVED and LVAT may be determined (e.g., calculated, computed, etc.) from electrical activity measured only by electrodes proximate the left side of the patient, which may be referred to as “left” electrodes. The left electrodes may be defined as any surface electrodes located proximate the left ventricle, which includes region to left of the patient’s sternum and spine. In one embodiment, the left electrodes may include all anterior electrodes on the left of the sternum and all posterior electrodes to the left of the spine. In another embodiment, the left electrodes may include all anterior electrodes on the left of the sternum and all posterior electrodes. In yet another embodiment, the left electrodes may be designated based on the contour of the left and right sides of the heart as determined using imaging apparatus (e.g., x-ray, fluoroscopy, etc.).

[0115] Another exemplary metric, or index, of dyssynchrony may be a range of activation times (RAT) that may be computed as the difference between the maximum and the minimum torso-surface or cardiac activation times, e.g., overall, or for a region. The RAT reflects the span of activation times while the SDAT gives an estimate of the dispersion of the activation times from a mean. The SDAT also provides an estimate of the heterogeneity of the activation times, because if activation times are spatially heterogeneous, the individual activation times will be further away from the mean activation time, indicating that one or more regions of heart have been delayed in activation. In some examples, the RAT may be calculated using the estimated cardiac activation times over the surface of a model heart.

[0116] Another exemplary metric, or index, of electrical heterogeneity information may include estimates of a percentage of surface electrodes located within a particular region of interest for the torso or heart whose associated

activation times are greater than a certain percentile, such as, for example the 70th percentile, of measured QRS complex duration or the determined activation times for surface electrodes. The region of interest may, e.g., be a posterior, left anterior, and/or left-ventricular region. The exemplary metric, or index, may be referred to as a percentage of late activation (PLAT). The PLAT may be described as providing an estimate of percentage of the region of interest, e.g., posterior and left-anterior area associated with the left ventricular area of heart, which activates late. A large value for PLAT may imply delayed activation of a substantial portion of the region, e.g., the left ventricle, and the potential benefit of electrical resynchronization through CRT by pre-exciting the late region, e.g., of left ventricle. In other examples, the PLAT may be determined for other subsets of electrodes in other regions, such as a right anterior region to evaluate delayed activation in the right ventricle. Furthermore, in some examples, the PLAT may be calculated using the estimated cardiac activation times over the surface of a model heart for either the whole heart or for a particular region, e.g., left or right ventricle, of the heart.

[0117] In one or more embodiments, additional cardiac information may include indicators of favorable changes in global cardiac electrical activation such as, e.g., described in Sweeney et al., “Analysis of Ventricular Activation Using Surface Electrocardiography to Predict Left Ventricular Reverse Volumetric Remodeling During Cardiac Resynchronization Therapy,” *Circulation*, 2010 Feb. 9, 121(5): 626-34 and/or Van Deursen, et al., “Vectorcardiography as a Tool for Easy Optimization of Cardiac Resynchronization Therapy in Canine LBBB Hearts,” *Circulation Arrhythmia and Electrophysiology*, 2012 Jun. 1, 5(3): 544-52, each of which is incorporated herein by reference in its entirety. Further, such additional cardiac information may also include measurements of improved cardiac mechanical function measured by imaging or other systems to track motion of implanted leads within the heart as, e.g., described in Ryu et al., “Simultaneous Electrical and Mechanical Mapping Using 3D Cardiac Mapping System: Novel Approach for Optimal Cardiac Resynchronization Therapy,” *Journal of Cardiovascular Electrophysiology*, 2010 February, 21(2): 219-22, Sperzel et al., “Intraoperative Characterization of Interventricular Mechanical Dyssynchrony Using Electroanatomic Mapping System—A Feasibility Study,” *Journal of Interventional Cardiac Electrophysiology*, 2012 November, 35(2): 189-96, and/or U.S. Pat. App. Pub. No. 2009/0099619 A1 entitled “METHOD FOR OPTIMIZING CRT THERAPY” and published on Apr. 16, 2009, each of which is incorporated herein by reference in its entirety.

[0118] An exemplary graph of the differences of standard deviation of activation-times (SDAT) generated using manually determined QRS durations and QRS durations determined by the exemplary methods and systems described herein is depicted in FIG. 17. And, an exemplary graph of the differences of average left surrogate electrical activation times (LVAT) generated using manually determined QRS durations and QRS durations determined by the exemplary methods and systems described herein is depicted in FIG. 18. As shown, the exemplary systems and methods were used to provide activation times using external electrode apparatus (e.g., ECG belt) from 138 data files. SDAT and LVAT were calculated using the QRS onset and offset times values determined using the exemplary systems, methods, and processes. “Manual” measurements, where a user

attempted to identify the QRS onset and offset times values “by hand” were used for comparison, and the SDAT and LVAT measurements were calculated at the time of data collection. The SDAT difference as shown equals SDAT determined using the exemplary systems and methods minus SDAT from the “manual” measurement, and the LVAT difference equals LVAT determined using the exemplary systems and methods minus LVAT from the “manual” measurements. As shown, the SDAT difference and LVAT difference over the 138 files indicates that the differences between the exemplary systems and methods and the “manual” calculations are small.

[0119] The exemplary method 200 may further include adjusting cardiac therapy 216 based on, or in response to, one or more metrics derived using the determined QRS onset and offset time values. Exemplary cardiac therapy may be further described herein with reference to FIGS. 19-21, which may be adjusted using the exemplary method 200. Additionally, the exemplary systems and methods described herein may further use or incorporate the systems and methods described in U.S. Pat. App. Pub. No. 2016/0045737 to Ghosh et al. published on Feb. 18, 2016, U.S. Pat. App. Pub. No. 2016/0045738 to Ghosh et al. published on Feb. 18, 2016, U.S. Pat. App. Pub. No. 2016/0045744 to Gillberg et al. published on Feb. 18, 2016, all of which are incorporated herein by reference in their entireties.

[0120] Additionally, the exemplary method 200 may be used with systems that include graphical user interfaces for use by users to evaluate a patient’s cardiac health and/or adjust cardiac therapy. As described herein with reference to FIG. 1, the exemplary systems and methods described herein may use display apparatus 130 including a graphical user interface 132. The graphical user interface 132 may be configured to, among other things, present information for use in assisting a user in evaluating a cardiac pacing, or therapy, location, in assisting a user in assessing a patient’s cardiac health, in assisting a user in navigating at least one implantable electrode to a region of the patient’s heart, etc. For example, the graphical user interface 132 may be configured to display activation time information and heterogeneity metrics determined by method 200. Further, for example, the graphical user interface 132 may be configured to display a spatial map of electrical activation times based on geometric location of the electrodes relative to each other.

[0121] Further, the graphical user interface 132 may be configured to depict, or display, a graphical representation of cardiac electrical activation times from the monitored electrical activity about a portion of human anatomy monitored during intrinsic rhythm and/or during the delivery of cardiac therapy. In at least one embodiment, the cardiac electrical activation times may be represented about a portion of human anatomy by color scaling a portion of human anatomy on the graphical user interface 132 according to the surrogate cardiac electrical activation times. Additional exemplary graphical representations of surrogate electrical activation times that may utilize or be utilized by the exemplary systems and methods described herein may be described in U.S. Pat. App. Pub. No. 2015/0157231 to Gillberg et al. published on Jun. 11, 2015, U.S. Pat. App. Pub. No. 2015/0157232 to Gillberg et al. published on Jun. 11, 2015, U.S. Pat. App. Pub. No. 2015/0157865 to Gillberg et al. published on Jun. 11, 2015, and U.S. Pat. App. Pub. No. 2016/0030751 to Ghosh et al. published on Feb. 4, 2016, U.S. Patent Application Publication No. 2012/0284003 A1

published on Nov. 8, 2012 and entitled “Assessing Intra-Cardiac Activation Patterns” and U.S. Patent Application Publication No. 2012/0283587 A1 published on Nov. 8, 2012 and entitled “Assessing Intra-Cardiac Activation Patterns and Electrical Dyssynchrony,” each of which is hereby incorporated by reference in its entirety.

[0122] The exemplary systems, methods, and graphical user interfaces described herein may be used with respect to the implantation and configuration of an implantable medical device (IMD) and/or one or more leads configured to be located proximate one or more portions of a patient’s heart. For example, the exemplary systems, methods, and interfaces may be used in conjunction with an exemplary therapy system 10 described herein with reference to FIGS. 9-11.

[0123] FIG. 19 is a conceptual diagram illustrating an exemplary therapy system 10 that may be used to deliver pacing therapy to a patient 14. Patient 14 may, but not necessarily, be a human. The therapy system 10 may include an implantable medical device 16 (IMD), which may be coupled to leads 18, 20, 22. The IMD 16 may be, e.g., an implantable pacemaker, cardioverter, and/or defibrillator, that delivers, or provides, electrical signals (e.g., paces, etc.) to and/or senses electrical signals from the heart 12 of the patient 14 via electrodes coupled to one or more of the leads 18, 20, 22.

[0124] The leads 18, 20, 22 extend into the heart 12 of the patient 14 to sense electrical activity of the heart 12 and/or to deliver electrical stimulation to the heart 12. In the example shown in FIG. 19, the right ventricular (RV) lead 18 extends through one or more veins (not shown), the superior vena cava (not shown), and the right atrium 26, and into the right ventricle 28. The left ventricular (LV) coronary sinus lead 20 extends through one or more veins, the vena cava, the right atrium 26, and into the coronary sinus 30 to a region adjacent to the free wall of the left ventricle 32 of the heart 12. The right atrial (RA) lead 22 extends through one or more veins and the vena cava, and into the right atrium 26 of the heart 12.

[0125] The IMD 16 may sense, among other things, electrical signals attendant to the depolarization and repolarization of the heart 12 via electrodes coupled to at least one of the leads 18, 20, 22. In some examples, the IMD 16 provides pacing therapy (e.g., pacing pulses) to the heart 12 based on the electrical signals sensed within the heart 12. The IMD 16 may be operable to adjust one or more parameters associated with the pacing therapy such as, e.g., AV delay and other various timings, pulse width, amplitude, voltage, burst length, etc. Further, the IMD 16 may be operable to use various electrode configurations to deliver pacing therapy, which may be unipolar, bipolar, quadripolar, or further multipolar. For example, a multipolar lead may include several electrodes that can be used for delivering pacing therapy. Hence, a multipolar lead system may provide, or offer, multiple electrical vectors to pace from. A pacing vector may include at least one cathode, which may be at least one electrode located on at least one lead, and at least one anode, which may be at least one electrode located on at least one lead (e.g., the same lead, or a different lead) and/or on the casing, or can, of the IMD. While improvement in cardiac function as a result of the pacing therapy may primarily depend on the cathode, the electrical parameters like impedance, pacing threshold voltage, current drain, longevity, etc. may be more dependent on the pacing vector, which includes both the cathode and the anode. The

IMD 16 may also provide defibrillation therapy and/or cardioversion therapy via electrodes located on at least one of the leads 18, 20, 22. Further, the IMD 16 may detect arrhythmia of the heart 12, such as fibrillation of the ventricles 28, 32, and deliver defibrillation therapy to the heart 12 in the form of electrical pulses. In some examples, IMD 16 may be programmed to deliver a progression of therapies, e.g., pulses with increasing energy levels, until a fibrillation of heart 12 is stopped.

[0126] FIGS. 20A-20B are conceptual diagrams illustrating the IMD 16 and the leads 18, 20, 22 of therapy system 10 of FIG. 19 in more detail. The leads 18, 20, 22 may be electrically coupled to a therapy delivery module (e.g., for delivery of pacing therapy), a sensing module (e.g., for sensing one or more signals from one or more electrodes), and/or any other modules of the IMD 16 via a connector block 34. In some examples, the proximal ends of the leads 18, 20, 22 may include electrical contacts that electrically couple to respective electrical contacts within the connector block 34 of the IMD 16. In addition, in some examples, the leads 18, 20, 22 may be mechanically coupled to the connector block 34 with the aid of set screws, connection pins, or another suitable mechanical coupling mechanism.

[0127] Each of the leads 18, 20, 22 includes an elongated insulative lead body, which may carry a number of conductors (e.g., concentric coiled conductors, straight conductors, etc.) separated from one another by insulation (e.g., tubular insulative sheaths). In the illustrated example, bipolar electrodes 40, 42 are located proximate to a distal end of the lead 18. In addition, bipolar electrodes 44, 45, 46, 47 are located proximate to a distal end of the lead 20 and bipolar electrodes 48, 50 are located proximate to a distal end of the lead 22.

[0128] The electrodes 40, 44, 45, 46, 47, 48 may take the form of ring electrodes, and the electrodes 42, 50 may take the form of extendable helix tip electrodes mounted retractably within the insulative electrode heads 52, 54, 56, respectively. Each of the electrodes 40, 42, 44, 45, 46, 47, 48, 50 may be electrically coupled to a respective one of the conductors (e.g., coiled and/or straight) within the lead body of its associated lead 18, 20, 22, and thereby coupled to a respective one of the electrical contacts on the proximal end of the leads 18, 20, 22.

[0129] Additionally, electrodes 44, 45, 46 and 47 may have an electrode surface area of about 5.3 mm² to about 5.8 mm². Electrodes 44, 45, 46, and 47 may also be referred to as LV1, LV2, LV3, and LV4, respectively. The LV electrodes (i.e., left ventricle electrode 1 (LV1) 44, left ventricle electrode 2 (LV2) 45, left ventricle electrode 3 (LV3) 46, and left ventricle 4 (LV4) 47 etc.) on the lead 20 can be spaced apart at variable distances. For example, electrode 44 may be a distance of, e.g., about 21 millimeters (mm), away from electrode 45, electrodes 45 and 46 may be spaced a distance of, e.g. about 1.3 mm to about 1.5 mm, away from each other, and electrodes 46 and 47 may be spaced a distance of, e.g. 20 mm to about 21 mm, away from each other.

[0130] The electrodes 40, 42, 44, 45, 46, 47, 48, 50 may further be used to sense electrical signals (e.g., morphological waveforms within electrograms (EGM)) attendant to the depolarization and repolarization of the heart 12. The electrical signals are conducted to the IMD 16 via the respective leads 18, 20, 22. In some examples, the IMD 16 may also deliver pacing pulses via the electrodes 40, 42, 44, 45, 46, 47, 48, 50 to cause depolarization of cardiac tissue of the

patient's heart 12. In some examples, as illustrated in FIG. 20A, the IMD 16 includes one or more housing electrodes, such as housing electrode 58, which may be formed integrally with an outer surface of a housing 60 (e.g., hermetically-sealed housing) of the IMD 16 or otherwise coupled to the housing 60. Any of the electrodes 40, 42, 44, 45, 46, 47, 48, 50 may be used for unipolar sensing or pacing in combination with the housing electrode 58. It is generally understood by those skilled in the art that other electrodes can also be selected to define, or be used for, pacing and sensing vectors. Further, any of electrodes 40, 42, 44, 45, 46, 47, 48, 50, 58, when not being used to deliver pacing therapy, may be used to sense electrical activity during pacing therapy.

[0131] As described in further detail with reference to FIG. 20A, the housing 60 may enclose a therapy delivery module that may include a stimulation generator for generating cardiac pacing pulses and defibrillation or cardioversion shocks, as well as a sensing module for monitoring the electrical signals of the patient's heart (e.g., the patient's heart rhythm). The leads 18, 20, 22 may also include elongated electrodes 62, 64, 66, respectively, which may take the form of a coil. The IMD 16 may deliver defibrillation shocks to the heart 12 via any combination of the elongated electrodes 62, 64, 66 and the housing electrode 58. The electrodes 58, 62, 64, 66 may also be used to deliver cardioversion pulses to the heart 12. Further, the electrodes 62, 64, 66 may be fabricated from any suitable electrically conductive material, such as, but not limited to, platinum, platinum alloy, and/or other materials known to be usable in implantable defibrillation electrodes. Since electrodes 62, 64, 66 are not generally configured to deliver pacing therapy, any of electrodes 62, 64, 66 may be used to sense electrical activity and may be used in combination with any of electrodes 40, 42, 44, 45, 46, 47, 48, 50, 58. In at least one embodiment, the RV elongated electrode 62 may be used to sense electrical activity of a patient's heart during the delivery of pacing therapy (e.g., in combination with the housing electrode 58, or defibrillation electrode-to-housing electrode vector).

[0132] The configuration of the exemplary therapy system 10 illustrated in FIGS. 19-21 is merely one example. In other examples, the therapy system may include epicardial leads and/or patch electrodes instead of or in addition to the transvenous leads 18, 20, 22 illustrated in FIG. 19. Additionally, in other examples, the therapy system 10 may be implanted in/around the cardiac space without transvenous leads (e.g., leadless/wireless pacing systems) or with leads implanted (e.g., implanted transvenously or using approaches) into the left chambers of the heart (in addition to or replacing the transvenous leads placed into the right chambers of the heart as illustrated in FIG. 19). Further, in one or more embodiments, the IMD 16 need not be implanted within the patient 14. For example, the IMD 16 may deliver various cardiac therapies to the heart 12 via percutaneous leads that extend through the skin of the patient 14 to a variety of positions within or outside of the heart 12. In one or more embodiments, the system 10 may utilize wireless pacing (e.g., using energy transmission to the intracardiac pacing component(s) via ultrasound, inductive coupling, RF, etc.) and sensing cardiac activation using electrodes on the can/housing and/or on subcutaneous leads.

[0133] In other examples of therapy systems that provide electrical stimulation therapy to the heart 12, such therapy

systems may include any suitable number of leads coupled to the IMD 16, and each of the leads may extend to any location within or proximate to the heart 12. For example, other examples of therapy systems may include three transvenous leads located as illustrated in FIGS. 9-11. Still further, other therapy systems may include a single lead that extends from the IMD 16 into the right atrium 26 or the right ventricle 28, or two leads that extend into a respective one of the right atrium 26 and the right ventricle 28.

[0134] FIG. 21A is a functional block diagram of one exemplary configuration of the IMD 16. As shown, the IMD 16 may include a control module 81, a therapy delivery module 84 (e.g., which may include a stimulation generator), a sensing module 86, and a power source 90.

[0135] The control module 81 may include a processor 80, memory 82, and a telemetry module 88. The memory 82 may include computer-readable instructions that, when executed, e.g., by the processor 80, cause the IMD 16 and/or the control module 81 to perform various functions attributed to the IMD 16 and/or the control module 81 described herein. Further, the memory 82 may include any volatile, non-volatile, magnetic, optical, and/or electrical media, such as a random-access memory (RAM), read-only memory (ROM), non-volatile RAM (NVRAM), electrically-erasable programmable ROM (EEPROM), flash memory, and/or any other digital media. An exemplary capture management module may be the left ventricular capture management (LVCM) module described in U.S. Pat. No. 7,684,863 entitled "LV THRESHOLD MEASUREMENT AND CAPTURE MANAGEMENT" and issued Mar. 23, 2010, which is incorporated herein by reference in its entirety.

[0136] The processor 80 of the control module 81 may include any one or more of a microprocessor, a controller, a digital signal processor (DSP), an application specific integrated circuit (ASIC), a field-programmable gate array (FPGA), and/or equivalent discrete or integrated logic circuitry. In some examples, the processor 80 may include multiple components, such as any combination of one or more microprocessors, one or more controllers, one or more DSPs, one or more ASICs, and/or one or more FPGAs, as well as other discrete or integrated logic circuitry. The functions attributed to the processor 80 herein may be embodied as software, firmware, hardware, or any combination thereof.

[0137] The control module 81 may control the therapy delivery module 84 to deliver therapy (e.g., electrical stimulation therapy such as pacing) to the heart 12 according to a selected one or more therapy programs, which may be stored in the memory 82. More, specifically, the control module 81 (e.g., the processor 80) may control various parameters of the electrical stimulus delivered by the therapy delivery module 84 such as, e.g., AV delays, VV delays, pacing pulses with the amplitudes, pulse widths, frequency, or electrode polarities, etc., which may be specified by one or more selected therapy programs (e.g., AV and/or VV delay adjustment programs, pacing therapy programs, pacing recovery programs, capture management programs, etc.). As shown, the therapy delivery module 84 is electrically coupled to electrodes 40, 42, 44, 45, 46, 47, 48, 50, 58, 62, 64, 66, e.g., via conductors of the respective lead 18, 20, 22, or, in the case of housing electrode 58, via an electrical conductor disposed within housing 60 of IMD 16. Therapy delivery module 84 may be configured to generate and deliver electrical stimulation therapy such as pacing therapy

to the heart 12 using one or more of the electrodes 40, 42, 44, 45, 46, 47, 48, 50, 58, 62, 64, 66.

[0138] For example, therapy delivery module 84 may deliver pacing stimulus (e.g., pacing pulses) via ring electrodes 40, 44, 45, 46, 47, 48 coupled to leads 18, 20, 22 and/or helical tip electrodes 42, 50 of leads 18, 22. Further, for example, therapy delivery module 84 may deliver defibrillation shocks to heart 12 via at least two of electrodes 58, 62, 64, 66. In some examples, therapy delivery module 84 may be configured to deliver pacing, cardioversion, or defibrillation stimulation in the form of electrical pulses. In other examples, therapy delivery module 84 may be configured deliver one or more of these types of stimulation in the form of other signals, such as sine waves, square waves, and/or other substantially continuous time signals.

[0139] The IMD 16 may further include a switch module 85 and the control module 81 (e.g., the processor 80) may use the switch module 85 to select, e.g., via a data/address bus, which of the available electrodes are used to deliver therapy such as pacing pulses for pacing therapy, or which of the available electrodes are used for sensing. The switch module 85 may include a switch array, switch matrix, multiplexer, or any other type of switching device suitable to selectively couple the sensing module 86 and/or the therapy delivery module 84 to one or more selected electrodes. More specifically, the therapy delivery module 84 may include a plurality of pacing output circuits. Each pacing output circuit of the plurality of pacing output circuits may be selectively coupled, e.g., using the switch module 85, to one or more of the electrodes 40, 42, 44, 45, 46, 47, 48, 50, 58, 62, 64, 66 (e.g., a pair of electrodes for delivery of therapy to a bipolar or multipolar pacing vector). In other words, each electrode can be selectively coupled to one of the pacing output circuits of the therapy delivery module using the switching module 85.

[0140] The sensing module 86 is coupled (e.g., electrically coupled) to sensing apparatus, which may include, among additional sensing apparatus, the electrodes 40, 42, 44, 45, 46, 47, 48, 50, 58, 62, 64, 66 to monitor electrical activity of the heart 12, e.g., electrocardiogram (ECG)/electrogram (EGM) signals, etc. The ECG/EGM signals may be used to measure or monitor activation times (e.g., ventricular activations times, etc.), heart rate (HR), heart rate variability (HRV), heart rate turbulence (HRT), deceleration/acceleration capacity, deceleration sequence incidence, T-wave alternans (TWA), P-wave to P-wave intervals (also referred to as the P-P intervals or A-A intervals), R-wave to R-wave intervals (also referred to as the R-R intervals or V-V intervals), P-wave to QRS complex intervals (also referred to as the P-R intervals, A-V intervals, or P-Q intervals), QRS-complex morphology, ST segment (i.e., the segment that connects the QRS complex and the T-wave), T-wave changes, QT intervals, electrical vectors, etc.

[0141] The switch module 85 may also be used with the sensing module 86 to select which of the available electrodes are used, or enabled, to, e.g., sense electrical activity of the patient's heart (e.g., one or more electrical vectors of the patient's heart using any combination of the electrodes 40, 42, 44, 45, 46, 47, 48, 50, 58, 62, 64, 66). Likewise, the switch module 85 may also be used with the sensing module 86 to select which of the available electrodes are not to be used (e.g., disabled) to, e.g., sense electrical activity of the patient's heart (e.g., one or more electrical vectors of the patient's heart using any combination of the electrodes 40,

42, 44, 45, 46, 47, 48, 50, 58, 62, 64, 66), etc. In some examples, the control module 81 may select the electrodes that function as sensing electrodes via the switch module within the sensing module 86, e.g., by providing signals via a data/address bus.

[0142] In some examples, sensing module 86 includes a channel that includes an amplifier with a relatively wider pass band than the R-wave or P-wave amplifiers. Signals from the selected sensing electrodes may be provided to a multiplexer, and thereafter converted to multi-bit digital signals by an analog-to-digital converter for storage in memory 82, e.g., as an electrogram (EGM). In some examples, the storage of such EGMs in memory 82 may be under the control of a direct memory access circuit.

[0143] In some examples, the control module 81 may operate as an interrupt driven device, and may be responsive to interrupts from pacer timing and control module, where the interrupts may correspond to the occurrences of sensed P-waves and R-waves and the generation of cardiac pacing pulses. Any necessary mathematical calculations may be performed by the processor 80 and any updating of the values or intervals controlled by the pacer timing and control module may take place following such interrupts. A portion of memory 82 may be configured as a plurality of recirculating buffers, capable of holding one or more series of measured intervals, which may be analyzed by, e.g., the processor 80 in response to the occurrence of a pace or sense interrupt to determine whether the patient's heart 12 is presently exhibiting atrial or ventricular tachyarrhythmia.

[0144] The telemetry module 88 of the control module 81 may include any suitable hardware, firmware, software, or any combination thereof for communicating with another device, such as a programmer. For example, under the control of the processor 80, the telemetry module 88 may receive downlink telemetry from and send uplink telemetry to a programmer with the aid of an antenna, which may be internal and/or external. The processor 80 may provide the data to be uplinked to a programmer and the control signals for the telemetry circuit within the telemetry module 88, e.g., via an address/data bus. In some examples, the telemetry module 88 may provide received data to the processor 80 via a multiplexer.

[0145] The various components of the IMD 16 are further coupled to a power source 90, which may include a rechargeable or non-rechargeable battery. A non-rechargeable battery may be selected to last for several years, while a rechargeable battery may be inductively charged from an external device, e.g., on a daily or weekly basis.

[0146] FIG. 21B is another embodiment of a functional block diagram for IMD 16. FIG. 21B depicts bipolar RA lead 22, bipolar RV lead 18, and bipolar LV CS lead 20 without the LA CS pace/sense electrodes and coupled with an implantable pulse generator (IPG) circuit 31 having programmable modes and parameters of a bi-ventricular DDD/R type known in the pacing art. In turn, the sensor signal processing circuit 91 indirectly couples to the timing circuit 43 and via data and control bus to microcomputer circuitry 33. The IPG circuit 31 is illustrated in a functional block diagram divided generally into a microcomputer circuit 33 and a pacing circuit 21. The pacing circuit 21 includes the digital controller/timer circuit 43, the output amplifiers circuit 51, the sense amplifiers circuit 55, the RF

telemetry transceiver 41, the activity sensor circuit 35 as well as a number of other circuits and components described below.

[0147] Crystal oscillator circuit 89 provides the basic timing clock for the pacing circuit 21 while battery 29 provides power. Power-on-reset circuit 87 responds to initial connection of the circuit to the battery for defining an initial operating condition and similarly, resets the operative state of the device in response to detection of a low battery condition. Reference mode circuit 37 generates stable voltage reference and currents for the analog circuits within the pacing circuit 21. Analog-to-digital converter (ADC) and multiplexer circuit 39 digitize analog signals and voltage to provide, e.g., real time telemetry of cardiac signals from sense amplifiers 55 for uplink transmission via RF transmitter and receiver circuit 41. Voltage reference and bias circuit 37, ADC and multiplexer 39, power-on-reset circuit 87, and crystal oscillator circuit 89 may correspond to any of those used in exemplary implantable cardiac pacemakers.

[0148] If the IPG is programmed to a rate responsive mode, the signals output by one or more physiologic sensors are employed as a rate control parameter (RCP) to derive a physiologic escape interval. For example, the escape interval is adjusted proportionally to the patient's activity level developed in the patient activity sensor (PAS) circuit 35 in the depicted, exemplary IPG circuit 31. The patient activity sensor 27 is coupled to the IPG housing and may take the form of a piezoelectric crystal transducer. The output signal of the patient activity sensor 27 may be processed and used as a RCP. Sensor 27 generates electrical signals in response to sensed physical activity that are processed by activity circuit 35 and provided to digital controller/timer circuit 43. Activity circuit 35 and associated sensor 27 may correspond to the circuitry disclosed in U.S. Pat. No. 5,052,388 entitled "METHOD AND APPARATUS FOR IMPLEMENTING ACTIVITY SENSING IN A PULSE GENERATOR" and issued on Oct. 1, 1991 and U.S. Pat. No. 4,428,378 entitled "RATE ADAPTIVE PACER" and issued on Jan. 31, 1984, each of which is incorporated herein by reference in its entirety. Similarly, the exemplary systems, apparatus, and methods described herein may be practiced in conjunction with alternate types of sensors such as oxygenation sensors, pressure sensors, pH sensors, and respiration sensors, for use in providing rate responsive pacing capabilities. Alternately, QT time may be used as a rate indicating parameter, in which case no extra sensor is required. Similarly, the exemplary embodiments described herein may also be practiced in non-rate responsive pacemakers.

[0149] Data transmission to and from the external programmer is accomplished by way of the telemetry antenna 57 and an associated RF transceiver 41, which serves both to demodulate received downlink telemetry and to transmit uplink telemetry. Uplink telemetry capabilities may include the ability to transmit stored digital information, e.g., operating modes and parameters, EGM histograms, and other events, as well as real time EGMs of atrial and/or ventricular electrical activity and marker channel pulses indicating the occurrence of sensed and paced depolarizations in the atrium and ventricle.

[0150] Microcomputer 33 contains a microprocessor 80 and associated system clock and on-processor RAM and ROM chips 82A and 82B, respectively. In addition, microcomputer circuit 33 includes a separate RAM/ROM chip 82C to provide additional memory capacity. Microprocessor

80 normally operates in a reduced power consumption mode and is interrupt driven. Microprocessor **80** is awakened in response to defined interrupt events, which may include A-TRIG, RV-TRIG, LV-TRIG signals generated by timers in digital timer/controller circuit **43** and A-EVENT, RV-EVENT, and LV-EVENT signals generated by sense amplifiers circuit **55**, among others. The specific values of the intervals and delays timed out by digital controller/timer circuit **43** are controlled by the microcomputer circuit **33** by way of data and control bus from programmed-in parameter values and operating modes. In addition, if programmed to operate as a rate responsive pacemaker, a timed interrupt, e.g., every cycle or every two seconds, may be provided in order to allow the microprocessor to analyze the activity sensor data and update the basic A-A, V-A, or V-V escape interval, as applicable. In addition, the microprocessor **80** may also serve to define variable, operative AV delay intervals, V-V delay intervals, and the energy delivered to each ventricle and/or atrium.

[0151] In one embodiment, microprocessor **80** is a custom microprocessor adapted to fetch and execute instructions stored in RAM/ROM unit **82** in a conventional manner. It is contemplated, however, that other implementations may be suitable to practice the present invention. For example, an off-the-shelf, commercially available microprocessor or microcontroller, or custom application-specific, hardwired logic, or state-machine type circuit may perform the functions of microprocessor **80**.

[0152] Digital controller/timer circuit **43** operates under the general control of the microcomputer **33** to control timing and other functions within the pacing circuit **21** and includes a set of timing and associated logic circuits of which certain ones pertinent to the present invention are depicted. The depicted timing circuits include URI/LRI timers **83A**, V-V delay timer **83B**, intrinsic interval timers **83C** for timing elapsed V-EVENT to V-EVENT intervals or V-EVENT to A-EVENT intervals or the V-V conduction interval, escape interval timers **83D** for timing A-A, V-A, and/or V-V pacing escape intervals, an AV delay interval timer **83E** for timing the A-LVp delay (or A-RVp delay) from a preceding A-EVENT or A-TRIG, a post-ventricular timer **83F** for timing post-ventricular time periods, and a date/time clock **83G**.

[0153] The AV delay interval timer **83E** is loaded with an appropriate delay interval for one ventricular chamber (e.g., either an A-RVp delay or an A-LVp) to time-out starting from a preceding A-PACE or A-EVENT. The interval timer **83E** triggers pacing stimulus delivery, and can be based on one or more prior cardiac cycles (or from a data set empirically derived for a given patient).

[0154] The post-event timer **83F** times out the post-ventricular time period following an RV-EVENT or LV-EVENT or a RV-TRIG or LV-TRIG and post-atrial time periods following an A-EVENT or A-TRIG. The durations of the post-event time periods may also be selected as programmable parameters stored in the microcomputer **33**. The post-ventricular time periods include the PVARP, a post-atrial ventricular blanking period (PAVBP), a ventricular blanking period (VBP), a post-ventricular atrial blanking period (PVARP) and a ventricular refractory period (VRP) although other periods can be suitably defined depending, at least in part, on the operative circuitry employed in the pacing engine. The post-atrial time periods include an atrial refractory period (ARP) during which an A-EVENT is

ignored for the purpose of resetting any AV delay, and an atrial blanking period (ABP) during which atrial sensing is disabled. It should be noted that the starting of the post-atrial time periods and the AV delays can be commenced substantially simultaneously with the start or end of each A-EVENT or A-TRIG or, in the latter case, upon the end of the A-PACE which may follow the A-TRIG. Similarly, the starting of the post-ventricular time periods and the V-A escape interval can be commenced substantially simultaneously with the start or end of the V-EVENT or V-TRIG or, in the latter case, upon the end of the V-PACE which may follow the V-TRIG. The microprocessor **80** also optionally calculates AV delays, VV delays, post-ventricular time periods, and post-atrial time periods that vary with the sensor based escape interval established in response to the RCP(s) and/or with the intrinsic atrial and/or ventricular rate.

[0155] The output amplifiers circuit **51** contains a RA pace pulse generator (and a LA pace pulse generator if LA pacing is provided), a RV pace pulse generator, a LV pace pulse generator, and/or any other pulse generator configured to provide atrial and ventricular pacing. In order to trigger generation of an RV-PACE or LV-PACE pulse, digital controller/timer circuit **43** generates the RV-TRIG signal at the time-out of the A-RVp delay (in the case of RV pre-excitation) or the LV-TRIG at the time-out of the A-LVp delay (in the case of LV pre-excitation) provided by AV delay interval timer **83E** (or the V-V delay timer **83B**). Similarly, digital controller/timer circuit **43** generates an RA-TRIG signal that triggers output of an RA-PACE pulse (or an LA-TRIG signal that triggers output of an LA-PACE pulse, if provided) at the end of the V-A escape interval timed by escape interval timers **83D**.

[0156] The output amplifiers circuit **51** includes switching circuits for coupling selected pace electrode pairs from among the lead conductors and the IND-CAN electrode **20** to the RA pace pulse generator (and LA pace pulse generator if provided), RV pace pulse generator and LV pace pulse generator. Pace/sense electrode pair selection and control circuit **53** selects lead conductors and associated pace electrode pairs to be coupled with the atrial and ventricular output amplifiers within output amplifiers circuit **51** for accomplishing RA, LA, RV and LV pacing.

[0157] The sense amplifiers circuit **55** contains sense amplifiers for atrial and ventricular pacing and sensing. High impedance P-wave and R-wave sense amplifiers may be used to amplify a voltage difference signal that is generated across the sense electrode pairs by the passage of cardiac depolarization wavefronts. The high impedance sense amplifiers use high gain to amplify the low amplitude signals and rely on pass band filters, time domain filtering and amplitude threshold comparison to discriminate a P-wave or R-wave from background electrical noise. Digital controller/timer circuit **43** controls sensitivity settings of the atrial and ventricular sense amplifiers **55**.

[0158] The sense amplifiers may be uncoupled from the sense electrodes during the blanking periods before, during, and after delivery of a pace pulse to any of the pace electrodes of the pacing system to avoid saturation of the sense amplifiers. The sense amplifiers circuit **55** includes blanking circuits for uncoupling the selected pairs of the lead conductors and the IND-CAN electrode **20** from the inputs of the RA sense amplifier (and LA sense amplifier if provided), RV sense amplifier and LV sense amplifier during the ABP, PAVBP and VBP. The sense amplifiers circuit **55**

also includes switching circuits for coupling selected sense electrode lead conductors and the IND-CAN electrode **20** to the RA sense amplifier (and LA sense amplifier if provided), RV sense amplifier and LV sense amplifier. Again, sense electrode selection and control circuit **53** selects conductors and associated sense electrode pairs to be coupled with the atrial and ventricular sense amplifiers within the output amplifiers circuit **51** and sense amplifiers circuit **55** for accomplishing RA, LA, RV, and LV sensing along desired unipolar and bipolar sensing vectors.

[0159] Right atrial depolarizations or P-waves in the RA-SENSE signal that are sensed by the RA sense amplifier result in a RA-EVENT signal that is communicated to the digital controller/timer circuit **43**. Similarly, left atrial depolarizations or P-waves in the LA-SENSE signal that are sensed by the LA sense amplifier, if provided, result in a LA-EVENT signal that is communicated to the digital controller/timer circuit **43**. Ventricular depolarizations or R-waves in the RV-SENSE signal are sensed by a ventricular sense amplifier result in an RV-EVENT signal that is communicated to the digital controller/timer circuit **43**. Similarly, ventricular depolarizations or R-waves in the LV-SENSE signal are sensed by a ventricular sense amplifier result in an LV-EVENT signal that is communicated to the digital controller/timer circuit **43**. The RV-EVENT, LV-EVENT, and RA-EVENT, LA-SENSE signals may be refractory or non-refractory, and can inadvertently be triggered by electrical noise signals or aberrantly conducted depolarization waves rather than true R-waves or P-waves.

[0160] The techniques described in this disclosure, including those attributed to the IMD **16**, the computing apparatus **140**, and/or various constituent components, may be implemented, at least in part, in hardware, software, firmware, or any combination thereof. For example, various aspects of the techniques may be implemented within one or more processors, including one or more microprocessors, DSPs, ASICs, FPGAs, or any other equivalent integrated or discrete logic circuitry, as well as any combinations of such components, embodied in programmers, such as physician or patient programmers, stimulators, image processing devices, or other devices. The term “module,” “processor,” or “processing circuitry” may generally refer to any of the foregoing logic circuitry, alone or in combination with other logic circuitry, or any other equivalent circuitry.

[0161] Such hardware, software, and/or firmware may be implemented within the same device or within separate devices to support the various operations and functions described in this disclosure. In addition, any of the described units, modules, or components may be implemented together or separately as discrete but interoperable logic devices. Depiction of different features as modules or units is intended to highlight different functional aspects and does not necessarily imply that such modules or units must be realized by separate hardware or software components. Rather, functionality associated with one or more modules or units may be performed by separate hardware or software components, or integrated within common or separate hardware or software components.

[0162] When implemented in software, the functionality ascribed to the systems, devices and techniques described in this disclosure may be embodied as instructions on a computer-readable medium such as RAM, ROM, NVRAM, EEPROM, FLASH memory, magnetic data storage media, optical data storage media, or the like. The instructions may

be executed by one or more processors to support one or more aspects of the functionality described in this disclosure

[0163] Embodiment 1: A system comprising:

[0164] electrode apparatus comprising a plurality of electrodes to monitor electrical activity from tissue of a patient; and

[0165] computing apparatus comprising processing circuitry and coupled to the electrode apparatus, the computing apparatus configured to:

[0166] monitor electrical activity using the plurality of electrodes to generate a plurality of cardiac signals over time,

[0167] generate a dispersion signal from the plurality of cardiac signals, wherein the dispersion signal is representative of the dispersion of the plurality of cardiac signals over time, and

[0168] determine a QRS onset time value and a QRS offset time value for each of at least one QRS complex within the monitored electrical activity based on the dispersion signal.

[0169] Embodiment 2: A method comprising:

[0170] monitoring electrical activity from tissue of a patient using a plurality of electrodes to generate a plurality of cardiac signals over time;

[0171] generating a dispersion signal from the plurality of cardiac signals, wherein the dispersion signal is representative of the dispersion of the plurality of cardiac signals over time; and

[0172] determining a QRS onset time value and a QRS offset time value for each of at least one QRS complex within the monitored electrical activity based on the dispersion signal.

[0173] Embodiment 3: The system of embodiment 1 or the method of embodiment 2, wherein the computing apparatus is further configured to execute or the method further comprises determining an activation time for each of the at least one QRS complex within each of the plurality of cardiac signals within the QRS duration between the QRS onset and offset time values corresponding to the at least one QRS complex.

[0174] Embodiment 4: The system or method of embodiment 3, wherein the computing apparatus is further configured to execute or the method further comprises determining at least one metric of electrical heterogeneity based on the activation time for each of the at least one QRS complex within each of the plurality of cardiac signals.

[0175] Embodiment 5: The systems or methods as set forth in any one of embodiments 1-4, wherein the plurality of electrodes comprises a plurality of external electrodes to be located proximate the patient's skin.

[0176] Embodiment 6: The systems or methods as set forth in embodiments 1-5, wherein the plurality of electrodes comprises a plurality of internal electrodes to be located within the patient.

[0177] Embodiment 7: The systems or methods as set forth in any one of embodiments 1-6, wherein generating a dispersion signal from the plurality of cardiac signals comprises determining a standard deviation of the plurality of cardiac signals over time.

[0178] Embodiment 8: The systems or methods as set forth in any one of embodiments 1-7, wherein the computing apparatus is further configured to execute or the method further comprises selecting a portion of the dispersion signal over a selected time period, and

[0179] wherein determining a QRS onset time value and a QRS offset time value for each of at least one QRS complex within the monitored electrical activity based on the dispersion signal comprises determining the QRS onset time value and the QRS offset time value for each of the at least one QRS complex within the monitored electrical activity of the selected time period based on the selected portion of the dispersion signal.

[0180] Embodiment 9: The systems or methods of embodiment 8, wherein the selected time period is less than or equal to 5 seconds.

[0181] Embodiment 10: The systems or methods as set forth in any one of embodiments 1-9, wherein determining a QRS onset time value and a QRS offset time value for each of at least one QRS complex within the monitored electrical activity based on the dispersion signal further comprises:

[0182] determining a maximum dispersion value of the dispersion signal; and

[0183] determining the QRS onset time value and the QRS offset time value using the maximum dispersion value.

[0184] Embodiment 11: The systems or methods of embodiment 10, wherein determining a QRS onset time value and a QRS offset time value for each of at least one QRS complex within the monitored electrical activity based on the dispersion signal further comprises:

[0185] filtering the dispersion signal to remove one or more values of the dispersion signal that are less than or equal to a selected percentage of the maximum dispersion value of the dispersion signal.

[0186] Embodiment 12: The systems or methods of embodiment 10, wherein determining a QRS onset time value and a QRS offset time value for each of at least one QRS complex within the monitored electrical activity based on the dispersion signal further comprises:

[0187] filtering the dispersion signal to remove one or more dispersion values of the dispersion signal corresponding to at least one of a T-wave and a P-wave in the plurality of cardiac signals.

[0188] Embodiment 13: The systems or methods as set forth in any one of embodiments 1-12, wherein determining a QRS onset time value and a QRS offset time value for each of at least one QRS complex within the monitored electrical activity based on the dispersion signal further comprises:

[0189] determining a minimum dispersion value of the dispersion signal within a first time period before a maximum dispersion value of the dispersion signal to determine the QRS onset time value, and

[0190] determining a minimum dispersion value of the dispersion signal within a second time period after the maximum dispersion value of the dispersion signal to determine the QRS offset time value.

[0191] Embodiment 14: The systems or methods of embodiment 13, wherein each of the first and second time periods are less than or equal to 200 milliseconds.

[0192] Embodiment 15: The systems or methods as set forth in any one of embodiments 1-14, wherein determining a QRS onset time value and a QRS offset time value for each of at least one QRS complex within the monitored electrical activity based on the dispersion signal further comprises:

[0193] determining a latest dispersion value of the dispersion signal within a first time period before a maximum dispersion value of the dispersion signal that is less than or equal to a selected percentage of a minimum dispersion

value of the dispersion signal within the first time period to determine the QRS onset time value, and

[0194] determining an earliest dispersion value of the dispersion signal within a second time period after the maximum dispersion value of the dispersion signal that is less than or equal to a selected percentage of a minimum dispersion value of the dispersion signal within the second time period to determine the QRS offset time value

[0195] Embodiment 16: The systems or methods as set forth in any one of embodiments 1-15, wherein the computing apparatus is further configured to execute or the method further comprises:

[0196] determining an absolute amplitude between the QRS onset time value and the QRS offset time value for each of the at least QRS complex for each of the plurality of cardiac signals;

[0197] selecting one or more cardiac signals of the plurality of cardiac signals based on the absolute amplitude; and

[0198] determining an activation time for each of the at least one QRS complex within each of the selected one or more cardiac signals within the QRS duration between the QRS onset and offset time values corresponding to the at least one QRS complex.

[0199] Embodiment 17: The systems or methods of embodiment 16, wherein selecting one or more cardiac signals of the plurality of cardiac signals based on the absolute amplitude comprises selecting one or more cardiac signals of the plurality of cardiac signals defining an absolute amplitude greater than or equal to a threshold value, wherein the threshold value is less than the median absolute amplitude of the plurality of cardiac signals.

[0200] Embodiment 18: The systems or methods as set forth in any one of embodiments 1-17, wherein the computing apparatus is further configured to execute or the method further comprises removing one or more noisy cardiac signals from the plurality of cardiac signals prior to determining the dispersion signal.

[0201] Embodiment 19: The systems or methods of embodiment 18, wherein the one or more noisy cardiac signals are identified by identifying cardiac signals comprising absolute areas-under-the-curve greater than or equal to a threshold value.

[0202] Embodiment 20: The systems or methods as set forth in any one of embodiments 1-19, wherein the at least one QRS complex comprises a plurality of QRS complexes.

[0203] Embodiment 21: The systems or methods as set forth in any one of embodiments 1-20, wherein the computing apparatus is further configured to execute or the method further comprises adjusting cardiac therapy in response to one or more metrics derived using the determined QRS onset and offset time values, wherein the one or more metrics comprise activation time of each of the plurality of cardiac signals.

[0204] Embodiment 22: The systems or methods as set forth in any one of embodiments 1-21, wherein the computing apparatus is further configured to execute or the method further comprises displaying on a graphical user interface one or more metrics derived using the determined QRS onset and offset time values, wherein the one or more metrics comprise activation time of each of the plurality of cardiac signals.

[0205] This disclosure has been provided with reference to illustrative embodiments and is not meant to be construed in a limiting sense. As described previously, one skilled in the

art will recognize that other various illustrative applications may use the techniques as described herein to take advantage of the beneficial characteristics of the apparatus and methods described herein. Various modifications of the illustrative embodiments, as well as additional embodiments of the disclosure, will be apparent upon reference to this description.

What is claimed:

1. A system comprising:
 - electrode apparatus comprising a plurality of electrodes to monitor electrical activity from tissue of a patient; and
 - computing apparatus comprising processing circuitry and coupled to the electrode apparatus, the computing apparatus configured to:
 - monitor electrical activity using the plurality of electrodes to generate a plurality of cardiac signals over time,
 - generate a dispersion signal from the plurality of cardiac signals, wherein the dispersion signal is representative of the dispersion of the plurality of cardiac signals over time, and
 - determine a QRS onset time value and a QRS offset time value for each of at least one QRS complex within the monitored electrical activity based on the dispersion signal.
2. The system of claim 1, wherein the computing apparatus is further configured to determine an activation time for each of the at least one QRS complex within each of the plurality of cardiac signals within the QRS duration between the QRS onset and offset time values corresponding to the at least one QRS complex.
3. The system of claim 2, wherein the computing apparatus is further configured to determine at least one metric of electrical heterogeneity based on the activation time for each of the at least one QRS complex within each of the plurality of cardiac signals.
4. The system of claim 1, wherein the plurality of electrodes comprises a plurality of external electrodes to be located proximate the patient's skin.
5. The system of claim 1, wherein the plurality of electrodes comprises a plurality of internal electrodes to be located within the patient.
6. The system of claim 1, wherein generating a dispersion signal from the plurality of cardiac signals comprises determining a standard deviation of the plurality of cardiac signals over time.
7. The system of claim 1, wherein the computing apparatus is further configured to select a portion of the dispersion signal over a selected time period, and
 - wherein determining a QRS onset time value and a QRS offset time value for each of at least one QRS complex within the monitored electrical activity based on the dispersion signal comprises determining the QRS onset time value and the QRS offset time value for each of the at least one QRS complex within the monitored electrical activity of the selected time period based on the selected portion of the dispersion signal.
8. The system of claim 1, wherein determining a QRS onset time value and a QRS offset time value for each of at least one QRS complex within the monitored electrical activity based on the dispersion signal further comprises:
 - determining a maximum dispersion value of the dispersion signal; and
 - determining the QRS onset time value and the QRS offset time value using the maximum dispersion value.
9. The system of claim 1, wherein determining a QRS onset time value and a QRS offset time value for each of at least one QRS complex within the monitored electrical activity based on the dispersion signal further comprises:
 - determining a minimum dispersion value of the dispersion signal within a first time period before a maximum dispersion value of the dispersion signal to determine the QRS onset time value, and
 - determining a minimum dispersion value of the dispersion signal within a second time period after the maximum dispersion value of the dispersion signal to determine the QRS offset time value.
10. The system of claim 1, wherein determining a QRS onset time value and a QRS offset time value for each of at least one QRS complex within the monitored electrical activity based on the dispersion signal further comprises:
 - determining a latest dispersion value of the dispersion signal within a first time period before a maximum dispersion value of the dispersion signal that is less than or equal to a selected percentage of a minimum dispersion value of the dispersion signal within the first time period to determine the QRS onset time value, and
 - determining an earliest dispersion value of the dispersion signal within a second time period after the maximum dispersion value of the dispersion signal that is less than or equal to a selected percentage of a minimum dispersion value of the dispersion signal within the second time period to determine the QRS offset time value.
11. The system of claim 1, wherein the computing apparatus is further configured to:
 - determine an absolute amplitude between the QRS onset time value and the QRS offset time value for each of the at least one QRS complex for each of the plurality of cardiac signals;
 - select one or more cardiac signals of the plurality of cardiac signals based on the absolute amplitude; and
 - determine an activation time for each of the at least one QRS complex within each of the selected one or more cardiac signals within the QRS duration between the QRS onset and offset time values corresponding to the at least one QRS complex.
12. The system of claim 1, wherein the computing apparatus is further configured to remove one or more noisy cardiac signals from the plurality of cardiac signals prior to determining the dispersion signal, wherein the one or more noisy cardiac signals are identified by identifying cardiac signals comprising absolute areas-under-the-curve greater than or equal to a threshold value.
13. The system of claim 1, wherein the computing apparatus is further configured to display on a graphical user interface one or more metrics derived using the determined QRS onset and offset time values, wherein the one or more metrics comprise activation time of each of the plurality of cardiac signals.
14. A method comprising:
 - monitoring electrical activity from tissue of a patient using a plurality of electrodes to generate a plurality of cardiac signals over time;

generating a dispersion signal from the plurality of cardiac signals, wherein the dispersion signal is representative of the dispersion of the plurality of cardiac signals over time; and

determining a QRS onset time value and a QRS offset time value for each of at least one QRS complex within the monitored electrical activity based on the dispersion signal.

15. The method of claim **14** further comprising determining an activation time for each of the at least one QRS complex within each of the plurality of cardiac signals within the QRS duration between the QRS onset and offset time values corresponding to the at least one QRS complex.

16. The method of claim **15** further comprising determining at least one metric of electrical heterogeneity based on the activation time for each of the at least one QRS complex within each of the plurality of cardiac signals.

17. The method of claim **14**, wherein generating a dispersion signal from the plurality of cardiac signals comprises determining a standard deviation of the plurality of cardiac signals over time.

18. The method of claim **14** further comprising selecting a portion of the dispersion signal over a selected time period, and

wherein determining a QRS onset time value and a QRS offset time value for each of at least one QRS complex within the monitored electrical activity based on the dispersion signal comprises determining the QRS onset time value and the QRS offset time value for each of the at least one QRS complex within the monitored electrical activity of the selected time period based on the selected portion of the dispersion signal.

19. The method of claim **14**, wherein determining a QRS onset time value and a QRS offset time value for each of at least one QRS complex within the monitored electrical activity based on the dispersion signal further comprises:

determining a maximum dispersion value of the dispersion signal; and

determining the QRS onset time value and the QRS offset time value using the maximum dispersion value.

20. The method of claim **19**, wherein determining a QRS onset time value and a QRS offset time value for each of at least one QRS complex within the monitored electrical activity based on the dispersion signal further comprises:

filtering the dispersion signal to remove one or more values of the dispersion signal that are less than or equal to a selected percentage of the maximum dispersion value of the dispersion signal.

21. The method of claim **19**, wherein determining a QRS onset time value and a QRS offset time value for each of at least one QRS complex within the monitored electrical activity based on the dispersion signal further comprises:

filtering the dispersion signal to remove one or more dispersion values of the dispersion signal correspond-

ing to at least one of a T-wave and a P-wave in the plurality of cardiac signals.

22. The method of claim **14**, wherein determining a QRS onset time value and a QRS offset time value for each of at least one QRS complex within the monitored electrical activity based on the dispersion signal further comprises:

determining a latest dispersion value of the dispersion signal within a first time period before a maximum dispersion value of the dispersion signal that is less than or equal to a selected percentage of a minimum dispersion value of the dispersion signal within the first time period to determine the QRS onset time value, and determining an earliest dispersion value of the dispersion signal within a second time period after the maximum dispersion value of the dispersion signal that is less than or equal to a selected percentage of a minimum dispersion value of the dispersion signal within the second time period to determine the QRS onset time value

23. The method of claim **22**, wherein each of the first and second time periods are less than or equal to 200 milliseconds.

24. The method of claim **14** further comprising:

determining an absolute amplitude between the QRS onset time value and the QRS offset time value for each of the at least QRS complex for each of the plurality of cardiac signals;

selecting one or more cardiac signals of the plurality of cardiac signals based on the absolute amplitude; and determining an activation time for each of the at least one QRS complex within each of the selected one or more cardiac signals within the QRS duration between the QRS onset and offset time values corresponding to the at least one QRS complex.

25. The method of claim **14** further comprising adjusting cardiac therapy in response to one or more metrics derived using the determined QRS onset and offset time values, wherein the one or more metrics comprise activation time of each of the plurality of cardiac signals.

26. A system comprising:

electrode apparatus comprising a plurality of electrodes to monitor electrical activity from tissue of a patient; and computing apparatus comprising processing circuitry and coupled to the electrode apparatus, the computing apparatus configured to:

monitor electrical activity using the plurality of electrodes to generate a plurality of cardiac signals over time,

determine a standard deviation of the plurality of cardiac signals over time, and

determine a QRS onset time value and a QRS offset time value for each of at least one QRS complex within the monitored electrical activity based on the determined standard deviation.

* * * * *

专利名称(译)	QRS OFFSET和ONSET确定		
公开(公告)号	US20180263522A1	公开(公告)日	2018-09-20
申请号	US15/915995	申请日	2018-03-08
[标]申请(专利权)人(译)	美敦力公司		
申请(专利权)人(译)	美敦力公司, INC.		
当前申请(专利权)人(译)	美敦力公司, INC.		
[标]发明人	GHOSH SUBHAM KUDLIK DANNE CHENG YA JIAN		
发明人	GHOSH, SUBHAM KUDLIK, D'ANNE CHENG, YA-JIAN		
IPC分类号	A61B5/0456 A61B5/0408 A61B5/00		
CPC分类号	A61B5/0456 A61B5/0408 A61B5/6801 A61B5/04085 A61B5/0422 A61B5/0428 A61B5/0452 A61B5/04525 A61B5/0472 A61B5/6823 A61B5/6852 A61B5/686 A61B5/7203 A61B5/7217 A61B5/7264 A61B5/7278 A61N1/3622 A61N1/365 A61N1/3704 A61N1/371 G16H50/20		
优先权	62/471938 2017-03-15 US		
外部链接	Espacenet USPTO		

摘要(译)

示例性系统和方法可以被配置为从多个心脏信号生成色散信号, 并从多个心脏信号确定QRS开始时间值和QRS偏移时间值。QRS起始和偏移时间值可用于测量或捕获激活时间。

