



US010368787B2

(12) **United States Patent**  
**Reichgott et al.**

(10) **Patent No.:** **US 10,368,787 B2**  
(45) **Date of Patent:** **\*Aug. 6, 2019**

(54) **FLOWOMETRY IN OPTICAL COHERENCE TOMOGRAPHY FOR ANALYTE LEVEL ESTIMATION**

(58) **Field of Classification Search**  
CPC ..... A61B 5/0066; A61B 5/026-0265; A61B 5/027; A61B 5/0285; A61B 5/029;  
(Continued)

(71) Applicant: **Masimo Corporation**, Irvine, CA (US)

(56) **References Cited**

(72) Inventors: **Samuel Reichgott**, Coopersburg, PA (US); **Walter J. Shakespeare**, Macungie, PA (US); **George Kechter**, Peoria, IL (US); **Phillip William Wallace**, Bernardsville, NJ (US); **Matthew J. Schurman**, Somerset, NJ (US)

U.S. PATENT DOCUMENTS

3,826,905 A 7/1974 Valkama et al.  
3,958,560 A 5/1976 March  
(Continued)

(73) Assignee: **MASIMO CORPORATION**, Irvine, CA (US)

FOREIGN PATENT DOCUMENTS

EP 0282234 9/1988  
EP 0160768 5/1989  
(Continued)

(\*) Notice: Subject to any disclaimer, the term of this patent is extended or adjusted under 35 U.S.C. 154(b) by 918 days.

This patent is subject to a terminal disclaimer.

OTHER PUBLICATIONS

US 8,845,543 B2, 09/2014, Diab et al. (withdrawn)  
(Continued)

(21) Appl. No.: **14/711,610**

*Primary Examiner* — Adam J Eiseman

(22) Filed: **May 13, 2015**

(74) *Attorney, Agent, or Firm* — Knobbe Martens Olson & Bear LLP

(65) **Prior Publication Data**

(57) **ABSTRACT**

US 2016/0058347 A1 Mar. 3, 2016

Optical coherence tomography (herein "OCT") based analyte monitoring systems are disclosed. In one aspect, techniques are disclosed that can identify fluid flow in vivo (e.g., blood flow), which can act as a metric for gauging the extent of blood perfusion in tissue. For instance, if OCT is to be used to estimate the level of an analyte (e.g., glucose) in tissue, a measure of the extent of blood flow can potentially indicate the presence of an analyte correlating region, which would be suitable for analyte level estimation with OCT. Another aspect is related to systems and methods for scanning multiple regions. An optical beam is moved across the surface of the tissue in two distinct manners. The first can be a coarse scan, moving the beam to provide distinct scanning positions on the skin. The second can be a fine scan where the beam is applied for more detailed analysis.


**Related U.S. Application Data**

(63) Continuation of application No. 14/063,977, filed on Oct. 25, 2013, now Pat. No. 9,060,721, which is a  
(Continued)

(51) **Int. Cl.**  
**A61B 5/00** (2006.01)  
**A61B 5/026** (2006.01)  
(Continued)

(52) **U.S. Cl.**  
CPC ..... **A61B 5/14551** (2013.01); **A61B 5/0066** (2013.01); **A61B 5/0073** (2013.01);  
(Continued)

**18 Claims, 16 Drawing Sheets**



<b>Related U.S. Application Data</b>						
	continuation of application No. 12/397,593, filed on Mar. 4, 2009, now Pat. No. 8,571,617.					
(60)	Provisional application No. 61/068,058, filed on Mar. 4, 2008, provisional application No. 61/033,584, filed on Mar. 4, 2008.					
(51)	<b>Int. Cl.</b>					
	<i>A61B 5/145</i> (2006.01)					
	<i>A61B 5/1455</i> (2006.01)					
(52)	<b>U.S. Cl.</b>					
	CPC ..... <i>A61B 5/026</i> (2013.01); <i>A61B 5/0261</i> (2013.01); <i>A61B 5/1455</i> (2013.01); <i>A61B 5/14532</i> (2013.01); <i>A61B 5/14546</i> (2013.01); <i>A61B 5/4875</i> (2013.01); <i>A61B 5/725</i> (2013.01); <i>A61B 5/7221</i> (2013.01); <i>A61B 5/7278</i> (2013.01); <i>A61B 2576/00</i> (2013.01)					
(58)	<b>Field of Classification Search</b>					
	CPC ..... A61B 5/14532; A61B 5/14546; A61B 5/1455-14557					
	See application file for complete search history.					
(56)	<b>References Cited</b>					
	U.S. PATENT DOCUMENTS					
	4,014,321 A	3/1977	March	5,349,953 A	9/1994	McCarthy et al.
	4,476,875 A	10/1984	Nilsson et al.	5,361,758 A	11/1994	Hall et al.
	4,590,948 A	5/1986	Nilsson	D353,195 S	12/1994	Savage et al.
	4,606,351 A	8/1986	Lubbers	D353,196 S	12/1994	Savage et al.
	4,655,225 A	4/1987	Dahne et al.	5,370,114 A	12/1994	Wong et al.
	4,704,029 A	11/1987	Van Heuvelen	5,372,136 A	12/1994	Steuer et al.
	4,731,363 A	3/1988	Hamilton et al.	5,376,336 A	12/1994	Lubbers et al.
	4,743,604 A	5/1988	Alig et al.	5,377,676 A	1/1995	Vari et al.
	4,746,211 A	5/1988	Ruth et al.	5,379,238 A	1/1995	Stark
	4,750,830 A	6/1988	Lee	5,383,452 A	1/1995	Buchert
	4,834,111 A	5/1989	Khanna et al.	5,398,681 A	3/1995	Kupershmidt
	4,871,755 A	10/1989	Alig et al.	D359,546 S	6/1995	Savage et al.
	4,873,989 A	10/1989	Einzig	5,431,170 A	7/1995	Mathews
	4,882,492 A	11/1989	Schlager	5,433,197 A	7/1995	Stark
	4,883,953 A	11/1989	Koashi et al.	5,435,309 A	7/1995	Thomas et al.
	4,890,621 A	1/1990	Hakky et al.	D361,840 S	8/1995	Savage et al.
	4,901,728 A	2/1990	Hutchinson	D362,063 S	9/1995	Savage et al.
	4,948,248 A	8/1990	Lehman	5,448,992 A	9/1995	Kupershmidt
	4,960,128 A	10/1990	Gordon et al.	5,452,716 A	9/1995	Clift
	4,964,408 A	10/1990	Hink et al.	5,452,717 A	9/1995	Branigan et al.
	4,979,509 A	12/1990	Hakky	D363,120 S	10/1995	Savage et al.
	4,989,978 A	2/1991	Groner	5,456,252 A	10/1995	Vari et al.
	5,025,785 A	6/1991	Weiss	5,457,535 A	10/1995	Schmidtke et al.
	5,028,787 A	7/1991	Rosenthal et al.	5,459,570 A	10/1995	Swanson et al.
	5,041,187 A	8/1991	Hink et al.	5,479,934 A	1/1996	Imran
	5,054,487 A	10/1991	Clarke	5,482,036 A	1/1996	Diab et al.
	5,069,213 A	12/1991	Polczynski	5,490,505 A	2/1996	Diab et al.
	5,070,874 A	12/1991	Barnes et al.	5,492,118 A	2/1996	Gratton et al.
	5,101,814 A	4/1992	Palti	5,494,043 A	2/1996	O'Sullivan et al.
	5,112,124 A	5/1992	Harjunmaa et al.	5,501,226 A	3/1996	Petersen et al.
	5,115,133 A	5/1992	Knudson	5,533,511 A	7/1996	Kaspari et al.
	5,163,438 A	11/1992	Gordon et al.	5,534,851 A	7/1996	Russek
	5,168,325 A	12/1992	Yoder-Short	5,535,743 A	7/1996	Backhaus et al.
	5,178,153 A	1/1993	Einzig	5,549,114 A	8/1996	Petersen et al.
	5,209,231 A	5/1993	Cote et al.	5,551,422 A	9/1996	Simonsen et al.
	5,222,495 A	6/1993	Clarke et al.	5,553,616 A	9/1996	Ham et al.
	5,222,496 A	6/1993	Clarke et al.	5,561,275 A	10/1996	Savage et al.
	5,243,983 A	9/1993	Tarr et al.	5,562,002 A	10/1996	Lalin
	5,267,152 A	11/1993	Yang et al.	5,582,168 A	12/1996	Samuels et al.
	5,277,181 A	1/1994	Mendelson et al.	5,582,171 A	12/1996	Chornenky et al.
	5,278,627 A	1/1994	Aoyagi et al.	5,590,649 A	1/1997	Caro et al.
	5,313,941 A	5/1994	Braig et al.	5,602,924 A	2/1997	Durand et al.
	5,319,355 A	6/1994	Russek	5,632,272 A	5/1997	Diab et al.
	5,321,501 A	6/1994	Swanson et al.	5,638,816 A	6/1997	Kiani-Azarbayjany et al.
	5,337,744 A	8/1994	Branigan	5,638,818 A	6/1997	Diab et al.
	5,341,805 A	8/1994	Stavridi et al.	5,645,440 A	7/1997	Tobler et al.
	5,348,003 A	9/1994	Caro	5,685,299 A	11/1997	Diab et al.
				5,710,630 A	1/1998	Essenpreis et al.
				D393,830 S	4/1998	Tobler et al.
				5,743,262 A	4/1998	Lepper, Jr. et al.
				5,758,644 A	6/1998	Diab et al.
				5,760,910 A	6/1998	Lepper, Jr. et al.
				5,769,785 A	6/1998	Diab et al.
				5,782,757 A	7/1998	Diab et al.
				5,785,659 A	7/1998	Caro et al.
				5,791,347 A	8/1998	Flaherty et al.
				5,795,295 A	8/1998	Hellmuth et al.
				5,810,734 A	9/1998	Caro et al.
				5,823,950 A	10/1998	Diab et al.
				5,830,131 A	11/1998	Caro et al.
				5,833,618 A	11/1998	Caro et al.
				5,860,919 A	1/1999	Kiani-Azarbayjany et al.
				5,890,929 A	4/1999	Mills et al.
				5,904,654 A	5/1999	Wohlmann et al.
				5,919,134 A	7/1999	Diab
				5,934,925 A	8/1999	Tobler et al.
				5,940,182 A	8/1999	Lepper, Jr. et al.
				5,995,855 A	11/1999	Kiani et al.
				5,997,343 A	12/1999	Mills et al.
				6,002,952 A	12/1999	Diab et al.
				6,011,986 A	1/2000	Diab et al.
				6,027,452 A	2/2000	Flaherty et al.
				6,036,642 A	3/2000	Diab et al.
				6,036,919 A	3/2000	Thym et al.
				6,045,509 A	4/2000	Caro et al.
				6,067,462 A	5/2000	Diab et al.
				6,081,735 A	6/2000	Diab et al.
				6,088,607 A	7/2000	Diab et al.
				6,110,522 A	8/2000	Lepper, Jr. et al.

(56)

## References Cited

## U.S. PATENT DOCUMENTS

6,124,597 A	9/2000	Shehada	6,697,658 B2	2/2004	Al-Ali
6,128,521 A	10/2000	Marro et al.	RE38,476 E	3/2004	Diab et al.
6,129,675 A	10/2000	Jay	6,699,194 B1	3/2004	Diab et al.
6,144,868 A	11/2000	Parker	6,707,554 B1	3/2004	Miltner et al.
6,147,108 A	11/2000	Hauptman	6,714,804 B2	3/2004	Al-Ali et al.
6,151,516 A	11/2000	Kiani-Azarbayjany et al.	RE38,492 E	4/2004	Diab et al.
6,152,754 A	11/2000	Gerhardt et al.	6,721,582 B2	4/2004	Trepagnier et al.
6,157,850 A	12/2000	Diab et al.	6,721,585 B1	4/2004	Parker
6,165,005 A	12/2000	Mills et al.	6,725,073 B1	4/2004	Motamedi et al.
6,173,197 B1	1/2001	Boggett et al.	6,725,075 B2	4/2004	Al-Ali
6,184,521 B1	2/2001	Coffin, IV et al.	6,728,560 B2	4/2004	Kollias et al.
6,206,830 B1	3/2001	Diab et al.	6,735,459 B2	5/2004	Parker
6,229,856 B1	5/2001	Diab et al.	6,745,060 B2	6/2004	Diab et al.
6,232,609 B1	5/2001	Snyder et al.	6,760,607 B2	7/2004	Al-Ali
6,236,872 B1	5/2001	Diab et al.	6,770,028 B1	8/2004	Ali et al.
6,241,683 B1	6/2001	Macklem et al.	6,771,994 B2	8/2004	Kiani et al.
6,253,097 B1	6/2001	Aronow et al.	6,780,651 B2	8/2004	Douglas et al.
6,256,523 B1	7/2001	Diab et al.	6,788,965 B2	9/2004	Ruchti et al.
6,263,222 B1	7/2001	Diab et al.	6,792,300 B1	9/2004	Diab et al.
6,278,522 B1	8/2001	Lepper, Jr. et al.	6,813,511 B2	11/2004	Diab et al.
6,280,213 B1	8/2001	Tobler et al.	6,816,741 B2	11/2004	Diab
6,285,896 B1	9/2001	Tobler et al.	6,822,564 B2	11/2004	Al-Ali
6,294,062 B1	9/2001	Bruck, Jr. et al.	6,826,419 B2	11/2004	Diab et al.
6,301,493 B1	10/2001	Marro et al.	6,830,711 B2	12/2004	Mills et al.
6,308,089 B1	10/2001	von der Ruhr et al.	6,836,337 B2	12/2004	Cornsweet
6,317,627 B1	11/2001	Ennen et al.	6,837,337 B2	1/2005	Thomas et al.
6,321,100 B1	11/2001	Parker	6,847,449 B2	1/2005	Bashkansky et al.
6,325,761 B1	12/2001	Jay	6,850,787 B2	2/2005	Weber et al.
6,334,065 B1	12/2001	Al-Ali et al.	6,850,788 B2	2/2005	Al-Ali
6,343,224 B1	1/2002	Parker	6,852,083 B2	2/2005	Caro et al.
6,349,228 B1	2/2002	Kiani et al.	6,861,639 B2	3/2005	Al-Ali
6,360,114 B1	3/2002	Diab et al.	6,898,452 B2	5/2005	Al-Ali et al.
6,368,283 B1	4/2002	Xu et al.	6,920,345 B2	7/2005	Al-Ali et al.
6,371,921 B1	4/2002	Caro et al.	6,931,268 B1	8/2005	Kiani-Azarbayjany et al.
6,377,829 B1	4/2002	Al-Ali	6,934,570 B2	8/2005	Kiani et al.
6,388,240 B2	5/2002	Schulz et al.	6,939,305 B2	9/2005	Flaherty et al.
6,397,091 B2	5/2002	Diab et al.	6,943,348 B1	9/2005	Coffin, IV
6,425,863 B1	7/2002	Werner et al.	6,950,687 B2	9/2005	Al-Ali
6,430,437 B1	8/2002	Marro	6,961,598 B2	11/2005	Diab
6,430,525 B1	8/2002	Weber et al.	6,970,792 B1	11/2005	Diab
6,443,881 B1	9/2002	Finger	6,979,812 B2	12/2005	Al-Ali
6,463,311 B1	10/2002	Diab	6,985,764 B2	1/2006	Mason et al.
6,470,199 B1	10/2002	Kopotic et al.	6,990,364 B2	1/2006	Ruchti et al.
6,501,975 B2	12/2002	Diab et al.	6,993,371 B2	1/2006	Kiani et al.
6,505,059 B1	1/2003	Kollias et al.	6,996,427 B2	2/2006	Ali et al.
6,515,273 B2	2/2003	Al-Ali	6,999,904 B2	2/2006	Weber et al.
6,519,487 B1	2/2003	Parker	7,003,338 B2	2/2006	Weber et al.
6,525,386 B1	2/2003	Mills et al.	7,003,339 B2	2/2006	Diab et al.
6,526,300 B1	2/2003	Kiani et al.	7,015,451 B2	3/2006	Dalke et al.
6,541,216 B1	4/2003	Wilsey et al.	7,020,506 B2	3/2006	Fine et al.
6,541,756 B2	4/2003	Schulz et al.	7,024,233 B2	4/2006	Ali et al.
6,542,764 B1	4/2003	Al-Ali et al.	7,027,849 B2	4/2006	Al-Ali
6,544,212 B2	4/2003	Galley et al.	7,030,749 B2	4/2006	Al-Ali
6,556,853 B1	4/2003	Cabib et al.	7,039,449 B2	5/2006	Al-Ali
6,572,566 B2	6/2003	Effenhauser	7,041,060 B2	5/2006	Flaherty et al.
6,580,086 B1	6/2003	Schulz et al.	7,044,918 B2	5/2006	Diab
6,584,336 B1	6/2003	Ali et al.	7,048,687 B1	5/2006	Reuss et al.
6,595,316 B2	7/2003	Cybulski et al.	7,067,893 B2	6/2006	Mills et al.
6,597,932 B2	7/2003	Tian et al.	7,096,052 B2	8/2006	Mason et al.
6,597,933 B2	7/2003	Kiani et al.	7,096,054 B2	8/2006	Abdul-Hafiz et al.
6,606,511 B1	8/2003	Ali et al.	7,132,641 B2	11/2006	Schulz et al.
6,632,181 B2	10/2003	Flaherty et al.	7,142,901 B2	11/2006	Kiani et al.
6,639,668 B1	10/2003	Trepagnier	7,149,561 B2	12/2006	Diab
6,640,116 B2	10/2003	Diab	7,186,966 B2	3/2007	Al-Ali
6,643,530 B2	11/2003	Diab et al.	7,190,261 B2	3/2007	Al-Ali
6,650,917 B2	11/2003	Diab et al.	7,190,464 B2	3/2007	Alphonse
6,654,624 B2	11/2003	Diab et al.	7,215,984 B2	5/2007	Diab
6,658,276 B2	12/2003	Kiani et al.	7,215,986 B2	5/2007	Diab
6,661,161 B1	12/2003	Lanzo et al.	7,221,971 B2	5/2007	Diab
6,671,531 B2	12/2003	Al-Ali et al.	7,225,006 B2	5/2007	Al-Ali et al.
6,678,543 B2	1/2004	Diab et al.	7,225,007 B2	5/2007	Al-Ali
6,684,090 B2	1/2004	Ali et al.	RE39,672 E	6/2007	Shehada et al.
6,684,091 B2	1/2004	Parker	7,239,905 B2	7/2007	Kiani-Azarbayjany et al.
6,697,656 B1	2/2004	Al-Ali	7,245,953 B1	7/2007	Parker
6,697,657 B1	2/2004	Shehada et al.	7,254,429 B2	8/2007	Schurman et al.
			7,254,431 B2	8/2007	Al-Ali
			7,254,433 B2	8/2007	Diab et al.
			7,254,434 B2	8/2007	Schulz et al.
			7,272,425 B2	9/2007	Al-Ali

(56)

## References Cited

## U.S. PATENT DOCUMENTS

7,274,955	B2	9/2007	Kiani et al.	7,884,945	B2	2/2011	Srinivasan et al.
D554,263	S	10/2007	Al-Ali	7,891,355	B2	2/2011	Al-Ali et al.
7,280,858	B2	10/2007	Al-Ali et al.	7,894,868	B2	2/2011	Al-Ali et al.
7,280,860	B2	10/2007	Ikeda et al.	7,899,507	B2	3/2011	Al-Ali et al.
7,289,835	B2	10/2007	Mansfield et al.	7,899,518	B2	3/2011	Trepagnier et al.
7,292,883	B2	11/2007	De Felice et al.	7,904,132	B2	3/2011	Weber et al.
7,295,866	B2	11/2007	Al-Ali	7,909,772	B2	3/2011	Popov et al.
7,307,734	B2	12/2007	Dogariu	7,910,875	B2	3/2011	Al-Ali
7,328,053	B1	2/2008	Diab et al.	7,919,713	B2	4/2011	Al-Ali et al.
7,332,784	B2	2/2008	Mills et al.	7,937,128	B2	5/2011	Al-Ali
7,340,287	B2	3/2008	Mason et al.	7,937,129	B2	5/2011	Mason et al.
7,341,559	B2	3/2008	Schulz et al.	7,937,130	B2	5/2011	Diab et al.
7,343,186	B2	3/2008	Lamego et al.	7,941,199	B2	5/2011	Kiani
D566,282	S	4/2008	Al-Ali et al.	7,951,086	B2	5/2011	Flaherty et al.
7,353,055	B2	4/2008	Hogan	7,957,780	B2	6/2011	Lamego et al.
7,355,512	B1	4/2008	Al-Ali	7,962,188	B2	6/2011	Kiani et al.
7,356,365	B2	4/2008	Schurman	7,962,190	B1	6/2011	Diab et al.
7,371,981	B2	5/2008	Abdul-Hafiz	7,976,472	B2	7/2011	Kiani
7,373,193	B2	5/2008	Al-Ali et al.	7,988,637	B2	8/2011	Diab
7,373,194	B2	5/2008	Weber et al.	7,990,382	B2	8/2011	Kiani
7,376,453	B1	5/2008	Diab et al.	7,991,446	B2	8/2011	Ali et al.
7,377,794	B2	5/2008	Al Ali et al.	8,000,761	B2	8/2011	Al-Ali
7,377,899	B2	5/2008	Weber et al.	8,008,088	B2	8/2011	Bellott et al.
7,383,070	B2	6/2008	Diab et al.	RE42,753	E	9/2011	Kiani-Azarbayjany et al.
7,415,297	B2	8/2008	Al-Ali et al.	8,019,400	B2	9/2011	Diab et al.
7,428,432	B2	9/2008	Ali et al.	8,028,701	B2	10/2011	Al-Ali et al.
7,438,683	B2	10/2008	Al-Ali et al.	8,029,765	B2	10/2011	Bellott et al.
7,440,787	B2	10/2008	Diab	8,036,727	B2	10/2011	Schurman et al.
7,454,240	B2	11/2008	Diab et al.	8,036,728	B2	10/2011	Diab et al.
7,467,002	B2	12/2008	Weber et al.	8,046,040	B2	10/2011	Ali et al.
7,469,157	B2	12/2008	Diab et al.	8,046,041	B2	10/2011	Diab et al.
7,471,969	B2	12/2008	Diab et al.	8,046,042	B2	10/2011	Diab et al.
7,471,971	B2	12/2008	Diab et al.	8,048,040	B2	11/2011	Kiani
7,483,729	B2	1/2009	Al-Ali et al.	8,050,728	B2	11/2011	Al-Ali et al.
7,483,730	B2	1/2009	Diab et al.	RE43,169	E	2/2012	Parker
7,489,958	B2	2/2009	Diab et al.	8,118,620	B2	2/2012	Al-Ali et al.
7,496,391	B2	2/2009	Diab et al.	8,126,528	B2	2/2012	Diab et al.
7,496,393	B2	2/2009	Diab et al.	8,128,572	B2	3/2012	Diab et al.
D587,657	S	3/2009	Al-Ali et al.	8,130,105	B2	3/2012	Al-Ali et al.
7,499,741	B2	3/2009	Diab et al.	8,145,287	B2	3/2012	Diab et al.
7,499,835	B2	3/2009	Weber et al.	8,150,487	B2	4/2012	Diab et al.
7,500,950	B2	3/2009	Al-Ali et al.	8,175,672	B2	5/2012	Parker
7,509,154	B2	3/2009	Diab et al.	8,180,420	B2	5/2012	Diab et al.
7,509,494	B2	3/2009	Al-Ali	8,182,443	B1	5/2012	Kiani
7,510,849	B2	3/2009	Schurman et al.	8,185,180	B2	5/2012	Diab et al.
7,526,328	B2	4/2009	Diab et al.	8,190,223	B2	5/2012	Al-Ali et al.
7,530,942	B1	5/2009	Diab	8,190,227	B2	5/2012	Diab et al.
7,530,949	B2	5/2009	Al Ali et al.	8,203,438	B2	6/2012	Kiani et al.
7,530,955	B2	5/2009	Diab et al.	8,203,704	B2	6/2012	Merritt et al.
7,557,929	B2	7/2009	Fang-Yen et al.	8,204,566	B2	6/2012	Schurman et al.
7,563,110	B2	7/2009	Al-Ali et al.	8,219,168	B2	7/2012	Lowery
7,596,398	B2	9/2009	Al-Ali et al.	8,219,172	B2	7/2012	Schurman et al.
7,618,375	B2	11/2009	Flaherty	8,224,411	B2	7/2012	Al-Ali et al.
D606,659	S	12/2009	Kiani et al.	8,228,181	B2	7/2012	Al-Ali
7,647,083	B2	1/2010	Al-Ali et al.	8,229,533	B2	7/2012	Diab et al.
D609,193	S	2/2010	Al-Ali et al.	8,233,955	B2	7/2012	Al-Ali et al.
D614,305	S	4/2010	Al-Ali et al.	8,244,325	B2	8/2012	Al-Ali et al.
RE41,317	E	5/2010	Parker	8,255,026	B1	8/2012	Al-Ali
7,729,733	B2	6/2010	Al-Ali et al.	8,255,027	B2	8/2012	Al-Ali et al.
7,734,320	B2	6/2010	Al-Ali	8,255,028	B2	8/2012	Al-Ali et al.
7,761,127	B2	7/2010	Al-Ali et al.	8,260,577	B2	9/2012	Weber et al.
7,761,128	B2	7/2010	Al-Ali et al.	8,265,723	B1	9/2012	McHale et al.
7,764,982	B2	7/2010	Dalke et al.	8,274,360	B2	9/2012	Sampath et al.
D621,516	S	8/2010	Kiani et al.	8,280,473	B2	10/2012	Al-Ali
7,791,155	B2	9/2010	Diab	8,301,217	B2	10/2012	Al-Ali et al.
7,801,581	B2	9/2010	Diab	8,306,596	B2	11/2012	Schurman et al.
7,822,452	B2	10/2010	Schurman et al.	8,310,336	B2	11/2012	Muhsin et al.
RE41,912	E	11/2010	Parker	8,315,683	B2	11/2012	Al-Ali et al.
7,844,313	B2	11/2010	Kiani et al.	RE43,860	E	12/2012	Parker
7,844,314	B2	11/2010	Al-Ali	8,337,403	B2	12/2012	Al-Ali et al.
7,844,315	B2	11/2010	Al-Ali	8,346,330	B2	1/2013	Lamego
7,865,222	B2	1/2011	Weber et al.	8,353,842	B2	1/2013	Al-Ali et al.
7,873,497	B2	1/2011	Weber et al.	8,355,766	B2	1/2013	MacNeish, III et al.
7,880,606	B2	2/2011	Al-Ali	8,359,080	B2	1/2013	Diab et al.
7,880,626	B2	2/2011	Al-Ali et al.	8,364,223	B2	1/2013	Al-Ali et al.
				8,364,226	B2	1/2013	Diab et al.
				8,374,665	B2	2/2013	Lamego
				8,385,995	B2	2/2013	Al-Ali et al.
				8,385,996	B2	2/2013	Smith et al.

(56)

## References Cited

## U.S. PATENT DOCUMENTS

8,388,353	B2	3/2013	Kiani et al.	8,755,856	B2	6/2014	Diab et al.
8,399,822	B2	3/2013	Al-Ali	8,755,872	B1	6/2014	Marinow
8,401,602	B2	3/2013	Kiani	8,761,850	B2	6/2014	Lamego
8,405,608	B2	3/2013	Al-Ali et al.	8,764,671	B2	7/2014	Kiani
8,414,499	B2	4/2013	Al-Ali et al.	8,768,423	B2	7/2014	Shakespeare et al.
8,418,524	B2	4/2013	Al-Ali	8,771,204	B2	7/2014	Telfort et al.
8,423,106	B2	4/2013	Lamego et al.	8,777,634	B2	7/2014	Kiani et al.
8,428,967	B2	4/2013	Olsen et al.	8,781,543	B2	7/2014	Diab et al.
8,430,817	B1	4/2013	Al-Ali et al.	8,781,544	B2	7/2014	Al-Ali et al.
8,437,825	B2	5/2013	Dalvi et al.	8,781,549	B2	7/2014	Al-Ali et al.
8,455,290	B2	6/2013	Siskavich	8,788,003	B2	7/2014	Schurman et al.
8,457,703	B2	6/2013	Al-Ali	8,790,268	B2	7/2014	Al-Ali
8,457,707	B2	6/2013	Kiani	8,801,613	B2	8/2014	Al-Ali et al.
8,463,349	B2	6/2013	Diab et al.	8,821,397	B2	9/2014	Al-Ali et al.
8,466,286	B2	6/2013	Bellot et al.	8,821,415	B2	9/2014	Al-Ali et al.
8,471,713	B2	6/2013	Poeze et al.	8,830,449	B1	9/2014	Lamego et al.
8,473,020	B2	6/2013	Kiani et al.	8,831,700	B2	9/2014	Schurman et al.
8,483,787	B2	7/2013	Al-Ali et al.	8,840,549	B2	9/2014	Al-Ali et al.
8,489,364	B2	7/2013	Weber et al.	8,847,740	B2	9/2014	Kiani et al.
8,498,684	B2	7/2013	Weber et al.	8,849,365	B2	9/2014	Smith et al.
8,504,128	B2	8/2013	Blank et al.	8,852,094	B2	10/2014	Al-Ali et al.
8,509,867	B2	8/2013	Workman et al.	8,852,994	B2	10/2014	Wojtczuk et al.
8,515,509	B2	8/2013	Bruinsma et al.	8,868,147	B2	10/2014	Stippick et al.
8,523,781	B2	9/2013	Al-Ali	8,868,150	B2	10/2014	Al-Ali et al.
8,529,301	B2	9/2013	Al-Ali et al.	8,870,792	B2	10/2014	Al-Ali et al.
8,532,727	B2	9/2013	Ali et al.	8,886,271	B2	11/2014	Kiani et al.
8,532,728	B2	9/2013	Diab et al.	8,888,539	B2	11/2014	Al-Ali et al.
D692,145	S	10/2013	Al-Ali et al.	8,888,708	B2	11/2014	Diab et al.
8,547,209	B2	10/2013	Kiani et al.	8,892,180	B2	11/2014	Weber et al.
8,548,548	B2	10/2013	Al-Ali et al.	8,897,847	B2	11/2014	Al-Ali
8,548,549	B2	10/2013	Schurman et al.	8,909,310	B2	12/2014	Lamego et al.
8,548,550	B2	10/2013	Al-Ali	8,911,377	B2	12/2014	Al-Ali
8,560,032	B2	10/2013	Al-Ali et al.	8,912,909	B2	12/2014	Al-Ali et al.
8,560,034	B1	10/2013	Diab et al.	8,920,317	B2	12/2014	Al-Ali et al.
8,570,167	B2	10/2013	Al-Ali	8,921,699	B2	12/2014	Al-Ali et al.
8,570,503	B2	10/2013	Vo et al.	8,922,382	B2	12/2014	Al-Ali et al.
8,571,617	B2	10/2013	Reichgott et al.	8,929,964	B2	1/2015	Al-Ali et al.
8,571,618	B1	10/2013	Lamego et al.	8,942,777	B2	1/2015	Diab et al.
8,571,619	B2	10/2013	Al-Ali et al.	8,948,834	B2	2/2015	Diab et al.
8,584,345	B2	10/2013	Al-Ali et al.	8,948,835	B2	2/2015	Diab
8,577,431	B2	11/2013	Lamego et al.	8,965,471	B2	2/2015	Lamego
8,581,732	B2	11/2013	Al-Ali et al.	8,983,564	B2	3/2015	Al-Ali
8,588,880	B2	11/2013	Abdul-Hafiz et al.	8,989,831	B2	3/2015	Al-Ali et al.
8,600,467	B2	12/2013	Al-Ali et al.	8,996,085	B2	3/2015	Kiani et al.
8,606,342	B2	12/2013	Diab	8,998,809	B2	4/2015	Kiani
8,626,255	B2	1/2014	Al-Ali et al.	9,028,429	B2	5/2015	Telfort et al.
8,630,691	B2	1/2014	Lamego et al.	9,037,207	B2	5/2015	Al-Ali et al.
8,634,889	B2	1/2014	Al-Ali et al.	9,060,721	B2	6/2015	Reichgott et al.
8,641,631	B2	2/2014	Sierra et al.	9,066,666	B2	6/2015	Kiani
8,652,060	B2	2/2014	Al-Ali	9,066,680	B1	6/2015	Al-Ali et al.
8,663,107	B2	3/2014	Kiani	9,072,474	B2	7/2015	Al-Ali et al.
8,666,468	B1	3/2014	Al-Ali	9,078,560	B2	7/2015	Schurman et al.
8,667,967	B2	3/2014	Al-Ali et al.	9,084,569	B2	7/2015	Weber et al.
8,670,811	B2	3/2014	O'Reilly	9,095,316	B2	8/2015	Welch et al.
8,670,814	B2	3/2014	Diab et al.	9,106,038	B2	8/2015	Telfort et al.
8,676,286	B2	3/2014	Weber et al.	9,107,625	B2	8/2015	Telfort et al.
8,682,407	B2	3/2014	Al-Ali	9,107,626	B2	8/2015	Al-Ali et al.
RE44,823	E	4/2014	Parker	9,113,831	B2	8/2015	Al-Ali
RE44,875	E	4/2014	Kiani et al.	9,113,832	B2	8/2015	Al-Ali
8,690,799	B2	4/2014	Telfort et al.	9,119,595	B2	9/2015	Lamego
8,700,112	B2	4/2014	Kiani	9,131,881	B2	9/2015	Diab et al.
8,702,627	B2	4/2014	Telfort et al.	9,131,882	B2	9/2015	Al-Ali et al.
8,706,179	B2	4/2014	Parker	9,131,883	B2	9/2015	Al-Ali
8,712,494	B1	4/2014	MacNeish, III et al.	9,131,917	B2	9/2015	Telfort et al.
8,715,206	B2	5/2014	Telfort et al.	9,138,180	B1	9/2015	Coverston et al.
8,718,735	B2	5/2014	Lamego et al.	9,138,182	B2	9/2015	Al-Ali et al.
8,718,737	B2	5/2014	Diab et al.	9,138,192	B2	9/2015	Weber et al.
8,718,738	B2	5/2014	Blank et al.	9,142,117	B2	9/2015	Muhsin et al.
8,720,249	B2	5/2014	Al-Ali	9,153,112	B1	10/2015	Kiani et al.
8,721,541	B2	5/2014	Al-Ali et al.	9,153,121	B2	10/2015	Kiani et al.
8,721,542	B2	5/2014	Al-Ali et al.	9,161,696	B2	10/2015	Al-Ali et al.
8,723,677	B1	5/2014	Kiani	9,161,713	B2	10/2015	Al-Ali et al.
8,740,792	B1	6/2014	Kiani et al.	9,167,995	B2	10/2015	Lamego et al.
8,754,776	B2	6/2014	Poeze et al.	9,176,141	B2	11/2015	Al-Ali et al.
8,755,535	B2	6/2014	Telfort et al.	9,186,102	B2	11/2015	Bruinsma et al.
				9,192,312	B2	11/2015	Al-Ali
				9,192,329	B2	11/2015	Al-Ali
				9,192,351	B1	11/2015	Telfort et al.
				9,195,385	B2	11/2015	Al-Ali et al.

(56)

## References Cited

## U.S. PATENT DOCUMENTS

9,211,072	B2	12/2015	Kiani	9,750,461	B1	9/2017	Telfort
9,211,095	B1	12/2015	Al-Ali	9,775,545	B2	10/2017	Al-Ali et al.
9,218,454	B2	12/2015	Kiani et al.	9,775,546	B2	10/2017	Diab et al.
9,226,696	B2	1/2016	Kiani	9,775,570	B2	10/2017	Al-Ali
9,241,662	B2	1/2016	Al-Ali et al.	9,778,079	B1	10/2017	Al-Ali et al.
9,245,668	B1	1/2016	Vo et al.	9,782,077	B2	10/2017	Lamego et al.
9,259,185	B2	2/2016	Abdul-Hafiz et al.	9,782,110	B2	10/2017	Kiani
9,267,572	B2	2/2016	Barker et al.	9,787,568	B2	10/2017	Lamego et al.
9,277,880	B2	3/2016	Poeze et al.	9,788,735	B2	10/2017	Al-Ali
9,289,167	B2	3/2016	Diab et al.	9,788,768	B2	10/2017	Al-Ali et al.
9,295,421	B2	3/2016	Kiani et al.	9,795,300	B2	10/2017	Al-Ali
9,307,928	B1	4/2016	Al-Ali et al.	9,795,310	B2	10/2017	Al-Ali
9,323,894	B2	4/2016	Kiani	9,795,358	B2	10/2017	Telfort et al.
D755,392	S	5/2016	Hwang et al.	9,795,739	B2	10/2017	Al-Ali et al.
9,326,712	B1	5/2016	Kiani	9,801,556	B2	10/2017	Kiani
9,333,316	B2	5/2016	Kiani	9,801,588	B2	10/2017	Weber et al.
9,339,220	B2	5/2016	Lamego et al.	9,808,188	B1	11/2017	Perea et al.
9,341,565	B2	5/2016	Lamego et al.	9,814,418	B2	11/2017	Weber et al.
9,351,673	B2	5/2016	Diab et al.	9,820,691	B2	11/2017	Kiani
9,351,675	B2	5/2016	Al-Ali et al.	9,833,152	B2	12/2017	Kiani et al.
9,364,181	B2	6/2016	Kiani et al.	9,833,180	B2	12/2017	Shakespeare et al.
9,368,671	B2	6/2016	Wojtczuk et al.	9,839,379	B2	12/2017	Al-Ali et al.
9,370,325	B2	6/2016	Al-Ali et al.	9,839,381	B1	12/2017	Weber et al.
9,370,326	B2	6/2016	McHale et al.	9,847,002	B2	12/2017	Kiani et al.
9,370,335	B2	6/2016	Al-Ali et al.	9,847,749	B2	12/2017	Kiani et al.
9,375,185	B2	6/2016	Ali et al.	9,848,800	B1	12/2017	Lee et al.
9,386,953	B2	7/2016	Al-Ali	9,848,806	B2	12/2017	Al-Ali et al.
9,386,961	B2	7/2016	Al-Ali et al.	9,848,807	B2	12/2017	Lamego
9,392,945	B2	7/2016	Al-Ali et al.	9,861,298	B2	1/2018	Eckerbom et al.
9,397,448	B2	7/2016	Al-Ali et al.	9,861,304	B2	1/2018	Al-Ali et al.
9,408,542	B1	8/2016	Kinast et al.	9,861,305	B1	1/2018	Weber et al.
9,436,645	B2	9/2016	Al-Ali et al.	9,867,578	B2	1/2018	Al-Ali et al.
9,445,759	B1	9/2016	Lamego et al.	9,872,623	B2	1/2018	Al-Ali
9,466,919	B2	10/2016	Kiani et al.	9,876,320	B2	1/2018	Coverston et al.
9,474,474	B2	10/2016	Lamego et al.	9,877,650	B2	1/2018	Muhsin et al.
9,480,422	B2	11/2016	Al-Ali	9,877,686	B2	1/2018	Al-Ali et al.
9,480,435	B2	11/2016	Olsen	9,891,079	B2	2/2018	Dalvi
9,492,110	B2	11/2016	Al-Ali et al.	9,895,107	B2	2/2018	Al-Ali et al.
9,510,779	B2	12/2016	Poeze et al.	9,913,617	B2	3/2018	Al-Ali et al.
9,517,024	B2	12/2016	Kiani et al.	9,924,893	B2	3/2018	Schurman et al.
9,532,722	B2	1/2017	Lamego et al.	9,924,897	B1	3/2018	Abdul-Hafiz
9,538,949	B2	1/2017	Al-Ali et al.	9,936,917	B2	4/2018	Poeze et al.
9,538,980	B2	1/2017	Telfort et al.	9,943,269	B2	4/2018	Muhsin et al.
9,549,696	B2	1/2017	Lamego et al.	9,949,676	B2	4/2018	Al-Ali
9,554,737	B2	1/2017	Schurman et al.	9,955,937	B2	5/2018	Telfort
9,560,996	B2	2/2017	Kiani	9,965,946	B2	5/2018	Al-Ali
9,560,998	B2	2/2017	Al-Ali et al.	9,980,667	B2	5/2018	Kiani et al.
9,566,019	B2	2/2017	Al-Ali et al.	D820,865	S	6/2018	Muhsin et al.
9,579,039	B2	2/2017	Jansen et al.	9,986,919	B2	6/2018	Lamego et al.
9,591,975	B2	3/2017	Dalvi et al.	9,986,952	B2	6/2018	Dalvi et al.
9,622,692	B2	4/2017	Lamego et al.	9,989,560	B2	6/2018	Poeze et al.
9,622,693	B2	4/2017	Diab	9,993,207	B2	6/2018	Al-Ali et al.
D788,312	S	5/2017	Al-Ali et al.	10,007,758	B2	6/2018	Al-Ali et al.
9,636,055	B2	5/2017	Al-Ali et al.	D822,215	S	7/2018	Al-Ali et al.
9,636,056	B2	5/2017	Al-Ali	D822,216	S	7/2018	Barker et al.
9,649,054	B2	5/2017	Lamego et al.	10,010,276	B2	7/2018	Al-Ali et al.
9,662,052	B2	5/2017	Al-Ali et al.	10,032,002	B2	7/2018	Kiani et al.
9,668,679	B2	6/2017	Schurman et al.	10,039,482	B2	8/2018	Al-Ali et al.
9,668,680	B2	6/2017	Bruinsma et al.	10,052,037	B2	8/2018	Kinast et al.
9,668,703	B2	6/2017	Al-Ali	10,058,275	B2	8/2018	Al-Ali et al.
9,675,286	B2	6/2017	Diab	10,064,562	B2	9/2018	Al-Ali
9,687,160	B2	6/2017	Kiani	10,130,291	B2	11/2018	Schurman et al.
9,693,719	B2	7/2017	Al-Ali et al.	2002/0016533	A1	2/2002	Marchitto et al.
9,693,737	B2	7/2017	Al-Ali	2004/0162471	A1	8/2004	Ikeda et al.
9,697,928	B2	7/2017	Al-Ali et al.	2004/0167382	A1	8/2004	Gardner et al.
9,717,425	B2	8/2017	Kiani et al.	2004/0260158	A1	12/2004	Hogan
9,717,458	B2	8/2017	Lamego et al.	2005/0043597	A1	2/2005	Xie
9,724,016	B1	8/2017	Al-Ali et al.	2005/0070771	A1	3/2005	Rule et al.
9,724,024	B2	8/2017	Al-Ali	2005/0101846	A1	5/2005	Fine et al.
9,724,025	B1	8/2017	Kiani et al.	2005/0105097	A1	5/2005	Fang-Yen et al.
9,730,640	B2	8/2017	Diab et al.	2005/0226912	A1	10/2005	Lowery
9,743,887	B2	8/2017	Al-Ali et al.	2005/0254061	A1	11/2005	Alphonse
9,749,232	B2	8/2017	Sampath et al.	2006/0063988	A1	3/2006	Schurman et al.
9,750,442	B2	9/2017	Olsen	2006/0161054	A1	7/2006	Reuss et al.
9,750,443	B2	9/2017	Smith et al.	2006/0187462	A1	8/2006	Srinivasan
				2006/0264719	A1	11/2006	Schurman et al.
				2006/0276696	A1	12/2006	Schurman et al.
				2007/0249917	A1	10/2007	Ballerstadt et al.
				2007/0282478	A1	12/2007	Al-Ali et al.

(56)

## References Cited

## U.S. PATENT DOCUMENTS

2009/0247924 A1	10/2009	Lamego et al.	2014/0135588 A1	5/2014	Al-Ali et al.
2009/0247984 A1	10/2009	Lamego et al.	2014/0142401 A1	5/2014	Al-Ali et al.
2009/0275813 A1	11/2009	Davis	2014/0142402 A1	5/2014	Al-Ali et al.
2009/0275844 A1	11/2009	Al-Ali	2014/0163344 A1	6/2014	Al-Ali
2009/0299157 A1	12/2009	Telfort et al.	2014/0163402 A1	6/2014	Lamego et al.
2010/0004518 A1	1/2010	Vo et al.	2014/0166076 A1	6/2014	Kiani et al.
2010/0030040 A1	2/2010	Poeze et al.	2014/0171763 A1	6/2014	Diab
2010/0261979 A1	10/2010	Kiani	2014/0180038 A1	6/2014	Kiani
2011/0001605 A1	1/2011	Kiani et al.	2014/0180154 A1	6/2014	Sierra et al.
2011/0082711 A1	4/2011	Poeze et al.	2014/0180160 A1	6/2014	Brown et al.
2011/0105854 A1	5/2011	Kiani et al.	2014/0187973 A1	7/2014	Brown et al.
2011/0125060 A1	5/2011	Telfort et al.	2014/0194709 A1	7/2014	Al-Ali et al.
2011/0208015 A1	8/2011	Welch et al.	2014/0194711 A1	7/2014	Al-Ali
2011/0209915 A1	9/2011	Telfort et al.	2014/0194766 A1	7/2014	Al-Ali et al.
2011/0213212 A1	9/2011	Al-Ali	2014/0200420 A1	7/2014	Al-Ali
2011/0230733 A1	9/2011	Al-Ali	2014/0200422 A1	7/2014	Weber et al.
2011/0237911 A1	9/2011	Lamego et al.	2014/0206963 A1	7/2014	Al-Ali
2012/0059267 A1	3/2012	Lamego et al.	2014/0213864 A1	7/2014	Abdul-Hafiz et al.
2012/0116175 A1	5/2012	Al-Ali et al.	2014/0243627 A1	8/2014	Diab et al.
2012/0165629 A1	6/2012	Merritt et al.	2014/0266790 A1	9/2014	Al-Ali et al.
2012/0179006 A1	7/2012	Jansen et al.	2014/0275808 A1	9/2014	Poeze et al.
2012/0209082 A1	8/2012	Al-Ali	2014/0275835 A1	9/2014	Lamego et al.
2012/0209084 A1	8/2012	Olsen et al.	2014/0275871 A1	9/2014	Lamego et al.
2012/0227739 A1	9/2012	Kiani	2014/0275872 A1	9/2014	Merritt et al.
2012/0265039 A1	10/2012	Kiani	2014/0275881 A1	9/2014	Lamego et al.
2012/0283524 A1	11/2012	Kiani et al.	2014/0288400 A1	9/2014	Diab et al.
2012/0286955 A1	11/2012	Welch et al.	2014/0296664 A1	10/2014	Bruinsma et al.
2012/0296178 A1	11/2012	Lamego et al.	2014/0303520 A1	10/2014	Telfort et al.
2012/0319816 A1	12/2012	Al-Ali	2014/0309506 A1	10/2014	Lamego et al.
2012/0330112 A1	12/2012	Lamego et al.	2014/0309559 A1	10/2014	Telfort et al.
2013/0023775 A1	1/2013	Lamego et al.	2014/0316217 A1	10/2014	Purdon et al.
2013/0045685 A1	2/2013	Kiani	2014/0316218 A1	10/2014	Purdon et al.
2013/0046204 A1	2/2013	Lamego et al.	2014/0316228 A1	10/2014	Blank et al.
2013/0041591 A1	3/2013	Lamego	2014/0323825 A1	10/2014	Al-Ali et al.
2013/0060108 A1	3/2013	Schurman et al.	2014/0323897 A1	10/2014	Brown et al.
2013/0060147 A1	3/2013	Welch et al.	2014/0323898 A1	10/2014	Purdon et al.
2013/0096405 A1	4/2013	Garfo	2014/0330092 A1	11/2014	Al-Ali et al.
2013/0096936 A1	4/2013	Sampath et al.	2014/0330098 A1	11/2014	Merritt et al.
2013/0109935 A1	5/2013	Al-Ali et al.	2014/0330099 A1	11/2014	Al-Ali et al.
2013/0162433 A1	6/2013	Muhsin et al.	2014/0333440 A1	11/2014	Kiani
2013/0190581 A1	7/2013	Al-Ali et al.	2014/0336481 A1	11/2014	Shakespeare et al.
2013/0197328 A1	8/2013	Diab et al.	2014/0343436 A1	11/2014	Kiani
2013/0211214 A1	8/2013	Olsen	2014/0357966 A1	12/2014	Al-Ali et al.
2013/0243021 A1	9/2013	Siskavich	2015/0005600 A1	1/2015	Blank et al.
2013/0253334 A1	9/2013	Al-Ali et al.	2015/0011907 A1	1/2015	Purdon et al.
2013/0274571 A1	10/2013	Diab et al.	2015/0018650 A1	1/2015	Al-Ali et al.
2013/0296672 A1	11/2013	O'Neil et al.	2015/0032029 A1	1/2015	Al-Ali et al.
2013/0296713 A1	11/2013	Al-Ali et al.	2015/0038859 A1	2/2015	Dalvi et al.
2013/0317370 A1	11/2013	Dalvi et al.	2015/0080754 A1	3/2015	Purdon et al.
2013/0324808 A1	12/2013	Al-Ali et al.	2015/0087936 A1	3/2015	Al-Ali et al.
2013/0331660 A1	12/2013	Al-Ali et al.	2015/0094546 A1	4/2015	Al-Ali
2013/0331670 A1	12/2013	Kiani	2015/0099950 A1	4/2015	Al-Ali et al.
2013/0338461 A1	12/2013	Lamego et al.	2015/0101844 A1	4/2015	Al-Ali et al.
2014/0012100 A1	1/2014	Al-Ali et al.	2015/0106121 A1	4/2015	Muhsin et al.
2014/0025306 A1	1/2014	Weber et al.	2015/0112151 A1	4/2015	Muhsin et al.
2014/0034353 A1	2/2014	Al-Ali et al.	2015/0165312 A1	6/2015	Kiani
2014/0051952 A1	2/2014	Reichgott et al.	2015/0196249 A1	7/2015	Brown et al.
2014/0051953 A1	2/2014	Lamego et al.	2015/0216459 A1	8/2015	Al-Ali et al.
2014/0051954 A1	2/2014	Al-Ali et al.	2015/0238722 A1	8/2015	Al-Ali
2014/0058230 A1	2/2014	Abdul-Hafiz et al.	2015/0245773 A1	9/2015	Lamego et al.
2014/0066783 A1	3/2014	Kiani et al.	2015/0245794 A1	9/2015	Al-Ali
2014/0077956 A1	3/2014	Sampath et al.	2015/0257689 A1	9/2015	Al-Ali et al.
2014/0081100 A1	3/2014	Muhsin et al.	2015/0272514 A1	10/2015	Kiani et al.
2014/0081175 A1	3/2014	Telfort	2015/0351697 A1	12/2015	Weber et al.
2014/0094667 A1	4/2014	Schurman et al.	2015/0359429 A1	12/2015	Al-Ali et al.
2014/0100434 A1	4/2014	Diab et al.	2015/0366507 A1	12/2015	Blank
2014/0114199 A1	4/2014	Lamego et al.	2016/0029932 A1	2/2016	Al-Ali
2014/0120564 A1	5/2014	Workman et al.	2016/0066824 A1	3/2016	Al-Ali et al.
2014/0121482 A1	5/2014	Merritt et al.	2016/0081552 A1	3/2016	Wojtczuk et al.
2014/0121483 A1	5/2014	Kiani	2016/0095543 A1	4/2016	Telfort et al.
2014/0125495 A1	5/2014	Al-Ali	2016/0095548 A1	4/2016	Al-Ali et al.
2014/0127137 A1	5/2014	Bellott et al.	2016/0103598 A1	4/2016	Al-Ali et al.
2014/0128696 A1	5/2014	Al-Ali	2016/0166182 A1	6/2016	Al-Ali et al.
2014/0128699 A1	5/2014	Al-Ali et al.	2016/0166183 A1	6/2016	Poeze et al.
2014/0129702 A1	5/2014	Lamego et al.	2016/0196388 A1	7/2016	Lamego
			2016/0197436 A1	7/2016	Barker et al.
			2016/0213281 A1	7/2016	Eckerbom et al.
			2016/0228043 A1	8/2016	O'Neil et al.
			2016/0233632 A1	8/2016	Scruggs et al.

(56)

## References Cited

## U.S. PATENT DOCUMENTS

2016/0234944 A1 8/2016 Schmidt et al.  
 2016/0270735 A1 9/2016 Diab et al.  
 2016/0283665 A1 9/2016 Sampath et al.  
 2016/0287090 A1 10/2016 Al-Ali et al.  
 2016/0287786 A1 10/2016 Kiani  
 2016/0296169 A1 10/2016 McHale et al.  
 2016/0310052 A1 10/2016 Al-Ali et al.  
 2016/0314260 A1 10/2016 Kiani  
 2016/0324488 A1 11/2016 Olsen  
 2016/0327984 A1 11/2016 Al-Ali et al.  
 2016/0331332 A1 11/2016 Al-Ali  
 2016/0367173 A1 12/2016 Dalvi et al.  
 2017/0000394 A1 1/2017 Al-Ali et al.  
 2017/0007134 A1 1/2017 Al-Ali et al.  
 2017/0007198 A1 1/2017 Al-Ali et al.  
 2017/0014083 A1 1/2017 Diab et al.  
 2017/0014084 A1 1/2017 Al-Ali et al.  
 2017/0042488 A1 2/2017 Muhsin  
 2017/0055851 A1 3/2017 Al-Ali  
 2017/0055882 A1 3/2017 Al-Ali et al.  
 2017/0055887 A1 3/2017 Al-Ali  
 2017/0055896 A1 3/2017 Al-Ali et al.  
 2017/0079594 A1 3/2017 Telfort et al.  
 2017/0086723 A1 3/2017 Al-Ali et al.  
 2017/0143281 A1 5/2017 Olsen  
 2017/0147774 A1 5/2017 Kiani  
 2017/0156620 A1 6/2017 Al-Ali et al.  
 2017/0173632 A1 6/2017 Al-Ali  
 2017/0187146 A1 6/2017 Kiani et al.  
 2017/0188919 A1 7/2017 Al-Ali et al.  
 2017/0196464 A1 7/2017 Jansen et al.  
 2017/0196470 A1 7/2017 Lamego et al.  
 2017/0224262 A1 8/2017 Al-Ali  
 2017/0228516 A1 8/2017 Sampath et al.  
 2017/0245790 A1 8/2017 Al-Ali et al.  
 2017/0251974 A1 9/2017 Shreim et al.  
 2017/0251975 A1 9/2017 Shreim et al.  
 2017/0258403 A1 9/2017 Abdul-Hafiz et al.  
 2017/0311851 A1 11/2017 Schurman et al.  
 2017/0311891 A1 11/2017 Kiani et al.  
 2017/0325728 A1 11/2017 Al-Ali et al.  
 2017/0332976 A1 11/2017 Al-Ali et al.  
 2017/0340293 A1 11/2017 Al-Ali et al.  
 2017/0360310 A1 12/2017 Kiani et al.  
 2017/0367632 A1 12/2017 Al-Ali et al.  
 2018/0008146 A1 1/2018 Al-Ali et al.  
 2018/0014752 A1 1/2018 Al-Ali et al.  
 2018/0028124 A1 2/2018 Al-Ali et al.  
 2018/0055385 A1 3/2018 Al-Ali  
 2018/0055390 A1 3/2018 Kiani et al.  
 2018/0055430 A1 3/2018 Diab et al.  
 2018/0064381 A1 3/2018 Shakespeare et al.  
 2018/0069776 A1 3/2018 Lamego et al.  
 2018/0070867 A1 3/2018 Smith et al.  
 2018/0082767 A1 3/2018 Al-Ali et al.  
 2018/0085068 A1 3/2018 Telfort  
 2018/0087937 A1 3/2018 Al-Ali et al.  
 2018/0103874 A1 4/2018 Lee et al.  
 2018/0103905 A1 4/2018 Kiani  
 2018/0110478 A1 4/2018 Al-Ali  
 2018/0116575 A1 5/2018 Perea et al.  
 2018/0125368 A1 5/2018 Lamego et al.  
 2018/0125430 A1 5/2018 Al-Ali et al.  
 2018/0125445 A1 5/2018 Telfort et al.  
 2018/0130325 A1 5/2018 Kiani et al.  
 2018/0132769 A1 5/2018 Weber et al.  
 2018/0132770 A1 5/2018 Lamego  
 2018/0146901 A1 5/2018 Al-Ali et al.  
 2018/0146902 A1 5/2018 Kiani et al.  
 2018/0153442 A1 6/2018 Eckerbom et al.  
 2018/0153446 A1 6/2018 Kiani  
 2018/0153447 A1 6/2018 Al-Ali et al.  
 2018/0153448 A1 6/2018 Weber et al.  
 2018/0161499 A1 6/2018 Al-Ali et al.  
 2018/0168491 A1 6/2018 Al-Ali et al.

2018/0174679 A1 6/2018 Sampath et al.  
 2018/0174680 A1 6/2018 Sampath et al.  
 2018/0182484 A1 6/2018 Sampath et al.  
 2018/0184917 A1 7/2018 Kiani  
 2018/0192953 A1 7/2018 Shreim et al.  
 2018/0192955 A1 7/2018 Al-Ali et al.  
 2018/0199871 A1 7/2018 Pauley et al.  
 2018/0206795 A1 7/2018 Al-Ali  
 2018/0206815 A1 7/2018 Telfort  
 2018/0213583 A1 7/2018 Al-Ali  
 2018/0214031 A1 8/2018 Kiani et al.  
 2018/0214090 A1 8/2018 Al-Ali et al.  
 2018/0218792 A1 8/2018 Muhsin et al.  
 2018/0225960 A1 8/2018 Al-Ali et al.  
 2018/0238718 A1 8/2018 Dalvi  
 2018/0242853 A1 8/2018 Al-Ali  
 2018/0242921 A1 8/2018 Muhsin et al.  
 2018/0242926 A1 8/2018 Muhsin et al.  
 2018/0247353 A1 8/2018 Al-Ali et al.  
 2018/0247712 A1 8/2018 Muhsin et al.  
 2019/0029574 A1 1/2019 Schurman et al.

## FOREIGN PATENT DOCUMENTS

EP	0127947	8/1990
EP	0280986	7/1992
EP	0317121	2/1994
EP	0536187	9/1994
EP	0589191	3/1997
EP	0603658	2/1999
EP	0631137	3/2002
EP	0670143	5/2003
WO	WO 88/06726	9/1988
WO	WO 89/10087	11/1989
WO	WO 91/18548	12/1991
WO	WO 92/10131	6/1992
WO	WO 92/17765	10/1992
WO	WO 93/00855	1/1993
WO	WO 93/07801	4/1993
WO	WO 93/09421	5/1993
WO	WO 93/16629	9/1993
WO	WO 94/04070	3/1994
WO	WO 94/13193	6/1994
WO	WO 95/32416	11/1995
WO	WO 02/65090	8/2002

## OTHER PUBLICATIONS

Arnold, V.W. et al., "Fourier Transformation Infrared Spectrometry—A New (Old) Method of Detection in Forensic Chemistry and Criminal Investigation," *Beitr Gerichl Med.* 47:123-47 (1989).

Arnold, M.A. et al., "Determination of Physiological Levels of Glucose in an Aqueous Matrix with Digitally Filtered Fourier Transform Near-Infrared Spectra," *Anal. Chem.* 64(14):1457-64 (1990).

Bruulsema, J.T. et al., "Correlation Between Blood Glucose Concentration in Diabetics and Noninvasively Measured Tissue Optical Scattering Coefficient," *Opt. Lett.* 22(3):190-93 (1997).

Burritt, M.F., "Current Analytical Approaches to Measuring Blood Analytes," *Clin. Chem.* 36(8 pt.2):1562-66 (1990).

Chira, I.S. et al., "Light Scattering by Blood Components After Supplying Glucose," *Biomed. Tech.* 35(5): 102-06 (1990).

Christison, G.B. et al., "Laser Photoacoustic Determination of Physiological Glucose Concentrations in Human Whole Blood," *Med. Biol. Eng. Comput.* 31(3):284-90 (1993).

Cote, G.L. et al., "Noninvasive Optical Polarimetric Glucose Sensing Using a True Phase Measurement Technique," *IEEE Trans. Biomed. Eng.* 39(7):752-56 (1992).

CRC Press, "Handbook of Chemistry and Physics" 64th ed. pp. D-223, D-224, and D. 235.

Drezek, R. et al., "Light Scattering From Cell: Finite Difference Time-Domain Simulations and Goniometric Measurements," *Appl. Opt.* 38(16):3651-61 (1999).

Duck, F. A., *Physical Properties of Tissue*, (Academic London 1990).

(56)

## References Cited

## OTHER PUBLICATIONS

- Dyer, D.G. et al., "Accumulation of Maillard Reaction Products in Skin Collagen in Diabetes and Aging," *J. Clin. Invest.* 91:2463-69 (1993).
- Esenaliev, R.O. et al., "Noninvasive Monitoring of Glucose Concentration with Optical Coherence Tomography," *Optics Lett.* 26(13):992-94 (2001).
- Faber, D.J. et al., "Light Absorption of (oxy-)Hemoglobin Assessed by Spectroscopic Optical Coherence Tomography," *Optics Lett.* 28(16):1436-38 (2003).
- Fercher, A. et al., "In Vivo Optical Coherence Tomography," *Amer. J. Ophthalmol.* 116(1):113-14 (1993).
- Flock, S.T. et al., "Total Attenuation Coefficients and Scattering Phase Functions of Tissues and Phantom Materials at 633 nm," *Med. Phys.* 14(5):835-41 (1987).
- Fogt, E.J., "Continuous Ex Vivo and In Vivo Monitoring with Chemical Sensors," *Clin. Chem.* 36(8 pt.2):1573-80 (1990).
- Frank, K.H. et al., "Measurements of Angular Distributions of Rayleigh and Mie Scattering Events in Biological Models," *Phys. Med. Biol.* 34(8):1901-16 (1989).
- Gabriely, I. et al., "Transcutaneous Glucose Measurement Using Near-Infrared Spectroscopy During Hypoglycemia," *Diabetes Care* 22(12):2026-32 (1999).
- Galanza, E.I. et al., "Skin Backreflectance and Microvascular System Functioning at the Action of Osmotic Agents," *J. Phys. D. Appl. Phys.* 36:1739-46 (2003).
- Gilbert, J.W. et al., "A Cerebrospinal Fluid Glucose Biosensor for Diabetes Mellitus," *ASAIO J.* 38(2):82-87 (1992).
- Goetz, M.J. et al., "Application of a Multivariate Technique to Raman Spectra for Quantification of Body Chemicals," *IEEE Trans. Biomed. Eng.* 42:728-31 (1995).
- Goodman, J.W., "Some Fundamental Properties of Speckle," *J. Optical Soc. of America* 66(11):1145-50 (1976).
- Gough, D.A., "The Composition and Optical Rotary Dispersion of Bovine Aqueous Humor," *Diabetes Care* 5(3):266-70 (1982).
- Gunby, P., "Laser-Implant Contact Lens Could be Glucose Monitor," *JAMA* 243(4):317 (1980).
- Guyton, A.C., *Textbook of medical physiology*, (W.B. Saunders Company 1992).
- Huang, D. et al., "Optical Coherence Tomograph," *Science* 254:1178-81 (1991).
- Huang, Y.L. et al., "On-Line Determination of Glucose Concentration Throughout Animal Cell Cultures Based on Chemiluminescent Detection of Hydrogen Peroxide Coupled with Flow-Injection Analysis," *J. Biotechnol.* 18(1-2):161-72 (1991).
- International Search Report from PCT/US05/26744, dated Oct. 25, 2006.
- International Search Report, from corresponding PCT/US06/13775, dated Sep. 13, 2006.
- International Search Report, from corresponding PCT/US06/21535, dated Feb. 21, 2008.
- Kaiser, N., "Laser Absorption Spectroscopy with an ATR Prism—Noninvasive in Vivo Determination of Glucose," *Horm. Metab. Res. Suppl.* 8:30-33 (1979).
- Kajiwara, K. et al., "Spectroscopic Quantitative Analysis of Blood Glucose by Fourier Transform Infrared Spectroscopy with an Attenuated Total Reflection Prism," *Med. Prog. Technol.* 18(3):181-89 (1992).
- Khalil, O.S., "Spectroscopic and Clinical Aspects of Noninvasive Glucose Measurements," *Clin. Chem.* 45(2):165-77 (1999).
- Kholodnykh, A.I. et al., "Precision of Measurement of Tissue Optical Properties with Optical Coherence Tomography," *Appl. Optics* 42(16):3027-37 (2003).
- King, T.W. et al., "Multispectral Polarimetric Glucose Detection Using a Single Pockels Cell," *Optical Engineering* 33(8):2746-53 (1994).
- Kohl, M. et al., "Influence of Glucose Concentration on Light Scattering in Tissue-Simulating Phantoms," *Optics Letters* 19(24):2170-72 (1994).
- Kohl, M. et al., "The Influence of Glucose Concentration Upon the Transport of Light in Tissue-Simulating Phantoms," *Phys. Med. Biol.* 40:1267-87(1995).
- Kruse-Jarres, J.D., "Physicochemical Determinations of Glucose in Vivo," *J. Clin. Chem. Clin. Biochem.* 26(4):201-08 (1988).
- Larin, K.V. et al., "Noninvasive Blood Glucose Monitoring With Optical Coherence Tomography," *Diabetes Care* 25(12):2263-67 (2002).
- Larin, K.V. et al., "Phase-Sensitive Optical Low-Coherence Reflectometry for the Detection of Analyte Concentrations," *Appl. Optics* 43(17):3408-14 (2004).
- Larin, K.V. et al., "Specificity of Noninvasive Blood Glucose Sensing Using Optical Coherence Tomography Technique: A Pilot Study," *Physics in Med. & Biol.* 48:1371-90 (2003).
- Larin, K.V. et al., "Optoacoustic Signal Profiles for Monitoring Glucose Concentration in Turbid Media," *SPIE Proc.* 3726:576-83 (1988).
- Lide, D.R., *CRC Handbook of Chemistry and Physics*, 79<sup>th</sup> ed. (CRC Press, Boca Raton, Florida, 1998).
- Mackenzie, H.A. et al., "Advances in Photoacoustic Noninvasive Glucose Testing," *Clin. Chem.* 45(9):1587-95 (1999).
- Maier, J.S. et al., "Possible Correlation Between Blood Glucose Concentration and the Reduced Scattering Coefficient of Tissues in the Near Infrared," *Optics Lett.* 19(24):2062-64 (1994).
- March, W. et al., "Optical Monitor of Glucose," *Trans. Am. Soc. Artif. Intern. Organs* 25:28-31 (1979).
- March, W.F. et al., "Noninvasive Glucose Monitoring of the Aqueous Humor of the Eye: Part II. Animal Studies and the Scleral Lens," *Diabetes Care* 5(3):259-65 (1982).
- Mendelson, Y. et al., "Blood Glucose Measurement by Multiple Attenuated Total Reflection and Infrared Absorption Spectroscopy," *IEEE Trans. Biomed. Eng.* 37(5):458-65 (1990).
- Moreno-Bondi, M.C. et al., "Oxygen Optrode for Use in a Fiber-Optic Glucose Biosensor," *Anal. Chem.* 62(21):2377-80 (1990).
- Muller, A., "In Vivo Measurement of Glucose Concentration with Lasers," *Harm. Metab. Res. Suppl.* 8:33-35 (1979).
- Narayanaswamy, R., "Current Developments in Optical Biochemical Sensors," *Biosens. Bioelectron.* 6(6):467-75 (1991).
- Pan, S. et al., "Near-Infrared Spectroscopic Measurement of Physiological Glucose Levels in Variable Matrices of Protein and Triglycerides," *Anal. Chem.* 68:1124-35 (1996).
- Peterson, J.I. et al., "A Miniature Fiberoptic pH Sensor Potentially Suitable for Glucose Measurements," *Diabetes Care* 5(3):272-74 (1982).
- Quan, K.M. et al., "Glucose Determination by a Pulsed Photoacoustic Technique—An Experimental Study Using a Gelatin-Based Tissue Phantom," *Phys. Med. Biol.* 38(12): 1911-22 (1993).
- Rabinovitch, B. et al., "Noninvasive Glucose Monitoring of the Aqueous Humor of the Eye: Part I. Measurement of Very Small Optical Rotations," *Diabetes Care* 5(3):254-58 (1982).
- Robinson, M.R. et al., "Noninvasive Glucose Monitoring in Diabetic Patients: A Preliminary Evaluation," *Clin. Chem.* 38(9):1618-22 (1992).
- Robinson, R.J. et al., "Glucose-Sensitive Membrane and Infrared Absorption Spectroscopy for Potential Use as an Implantable Glucose Sensor," *ASAIO J.* 38(3):M458-62 (1992).
- Rusch, T.L. et al., "Signal Processing Methods for Pulse Oximetry," *Comput. Biol. Med.* 26(2):143-59 (1996).
- Ruth, Bernhard, et al., "Noncontact determination of skin blood flow using the laser speckle method: Application to patients with peripheral arterial occlusive disease (PAOD) and to type-1 diabetics." 1993. Wiley-Liss. *Lasers in Surgery and Medicine*. 13: 179-188.
- Schmitt, J.M. et al., "Measurement of Optical Properties of Biological Tissues by Low-Coherence Reflectometry," *Appl. Optics* 32(30):6032-42 (1993).
- Schmitt, J.M. et al., "Speckle in Optical Coherence Tomography," *J. Biomed. Optics* 4(1 ):95-105 (1999).
- Schmitt, J.M. et al., "Optical Coherence Tomography (OCT): A Review," *IEEE J. Selected Topics in Quantum Electronics* 5(4):1205-15 (1999).
- Sevick, E.M. et al., "Near-Infrared Optical Imaging of Tissue Phantoms with Measurement in the Change of Optical Path Lengths," *Adv. Exp. Med. Biol.* 345:815-23 (1994).

(56)

**References Cited**

## OTHER PUBLICATIONS

- Sodickson, L.A. et al., "Kromoscopic Analysis: A Possible Alternative to Spectroscopic Analysis for Noninvasive Measurement of Analytes in Vivo," *Clin. Chern.* 40(9):1838-44 (1994).
- Star, W.M. et al., "Light Dosimetry: Status and Prospects," *J. Photochem. Photobiol.* 1(2):149-67 (1987).
- Stoddart, S. et al., "Pulse Oximetry: What it is and How to Use it," *J. Neonatal Nursing* 10:12-14 (1997).
- Takai, N. et al., "Studies of the Development of Optical Fiber Sensors for Biochemical Analysis," *Artif. Organs* 15(2):86-89 (1991).
- Tunchin, V.V. et al., "Light Propagation in Tissues with Controlled Optical Properties," *J. Biomed. Opt.* 2(4):401-17 (1997).
- Wang, L. et al., "Speckle Reduction in Laser Projection Systems by Diffractive Optical Elements," *Appl. Optics* 37(10):1770-75 (1998).
- Weast, R.C., et al. *CRC Handbook of Chemistry and Physics*, 70<sup>th</sup> ed., (CRC Cleveland, Ohio, 1989).
- Welch, A.J. et al., "Practical Models for Light Distribution in Laser-Irradiated Tissue," *Lasers Surg. Med.* 6(6):488-93 (1987).
- Wicksted, J.P. et al., "Monitoring of Aqueous Humor Metabolites Using Raman Spectroscopy," *SPIE Proc.* 2135:264-74 (1994).
- Zeller, H. et al., "Blood Glucose Measurement by Infrared Spectroscopy," *J. Artif. Organs* 12(2):129-35 (1989).
- Zhao, Yonghua, et al. "Doppler standard deviation imaging for clinical monitoring of in vitro human skin blood flow." *Optics Letters*. Sep. 15, 2000. vol. 25, No. 18 pp. 1358-1360.
- U.S. Appl. No. 14/737,242, filed Jun. 11, 2015 including its ongoing prosecution history, including without limitation Office Actions, Amendments, Remarks, and any other potentially relevant documents, GLT Acquisition Corp.
- 2014/0094667, U.S. Appl. No. 14/036,786, filed Sep. 25, 2013 including its ongoing prosecution history, including without limitation Office Actions, Amendments, Remarks, and any other potentially relevant documents, Apr. 3, 2014, GLT Acquisition Corp.
- 2014/0336481, U.S. Appl. No. 14/280,294, filed May 16, 2014 including its ongoing prosecution history, including without limitation Office Actions, Amendments, Remarks, and any other potentially relevant documents, Nov. 13, 2014, GLT Acquisition Corp.
- U.S. Appl. No. 15/605,460, Method for Data Reduction and Calibration of an OCT-Based Physiological Monitor, filed May 25, 2017.
- U.S. Appl. No. 15/799,689, Multispot Monitoring for Use in Optical Coherence Tomography, filed Oct. 31, 2017.

FIG. 1




FIG. 2




FIG. 3A




FIG. 3B




FIG. 3C




FIG. 4




FIG. 5




FIG. 6A




FIG. 6B




FIG. 7A




FIG. 7B




FIG. 8




FIG. 9





FIG. 10







FIG. 11


FIG. 12




**FIG. 13**

2289 SITE 5


TISSUE DISTORTION AND BLOOD GLUCOSE OVER TIME



**FIG. 14**  
2289 SITE 5  
PERFUSION SIGNAL vs GLUCOSE SIGNAL - PERFUSION DETECTOR 1.0



**FIG. 15**  
PULSATILE BLOOD FLOW FROM PERIPHERAL PERFUSION



## FLOWOMETRY IN OPTICAL COHERENCE TOMOGRAPHY FOR ANALYTE LEVEL ESTIMATION

### CROSS-REFERENCE TO RELATED APPLICATIONS

The present application is a continuation of U.S. patent application Ser. No. 14/063,977, filed Oct. 25, 2013, titled "Flowometry in Optical Coherence Tomography for Analyte Level Estimation," which is a continuation of U.S. patent application Ser. No. 12/397,593, filed Mar. 4, 2009, titled "Flowometry in Optical Coherence Tomography for Analyte Level Estimation," which claims the benefit of U.S. Provisional Application No. 61/068,058, filed Mar. 4, 2008, titled "Flowometry in Optical Coherence Tomography for Analyte Level Estimation," and U.S. Provisional Application No. 61/033,584, filed Mar. 4, 2008, titled "Multispot Monitoring for Use in Optical Coherence Tomography." The present application is also related to U.S. patent application Ser. No. 12/397,577, filed Mar. 4, 2009, titled "Multispot Monitoring for Use in Optical Coherence Tomography." The entirety of each of the above-mentioned applications is hereby incorporated by reference herein for all purposes.

### BACKGROUND

The present application relates to methods and devices for performing measurements with optical coherence tomographic techniques, and in some instances particularly for detecting analytes using optical coherence tomography.

Optical coherence tomography (herein "OCT") is an optical imaging technique that has shown great promise for use in many applications such as glucose monitoring. In this sensor modality, the high depth resolution of this technique is able to measure, with a high degree of accuracy, the scattering properties of a subject's tissue (e.g., the skin). These scattering properties can be sensitive to the changes in analyte levels in a subject's tissue. e.g., blood glucose levels in a subject. Thus, changes in the OCT signal can be correlated with changes in the analyte levels and thus serve as a prospective predictor of changes in that subject's analyte levels.

While OCT-based devices and techniques have shown promise in detecting analytes such as glucose, improvements are still needed. For instance, improvements in the speed and ease with which such techniques can be applied could help commercial acceptability of such devices. Accordingly, a need persists to provide improved OCT devices and techniques, which can provide accelerated estimates of analyte levels in a subject's tissue in an easy to use and reliable manner.

### SUMMARY

Optical coherence tomography (herein "OCT") based analyte monitoring systems are disclosed. In one aspect, techniques are disclosed that can identify fluid flow in vivo (e.g., blood flow), which can act as a metric for gauging the extent of blood perfusion in tissue. For instance, if OCT is to be used to estimate the level of an analyte (e.g., glucose) in tissue, a measure of the extent of blood flow can potentially indicate the presence of an analyte correlating region, which would be suitable for analyte level estimation with OCT. Another aspect is related to systems and methods for scanning multiple regions. An optical beam is moved across the surface of the tissue in two distinct manners. The first can

be a coarse scan, moving the beam to provide distinct scanning positions on the skin. The second can be a fine scan where the beam is applied for more detailed analysis.

Some aspects of the present invention are directed to techniques which can detect fluid flow in vivo (e.g., blood flow) using optical coherence tomography (herein "OCT"). Such techniques can utilize the speckle from an OCT scan and/or the fringe-frequency modulation to identify flow. Detection of flow by OCT can be used in a variety of applications such as heart monitoring or imaging of fluid flow in a non-invasive manner. In some embodiments, OCT is applied to detect blood flow, which can act as a metric for gauging the extent of blood perfusion in tissue. For instance, if OCT is to be used to estimate the level of an analyte (e.g., glucose) in tissue, a measure of the extent of blood flow can potentially indicate the presence of an analyte correlating region (herein "ACR") which would be suitable for analyte level estimation with OCT. While not necessarily being bound by any particular theory, if a particular analyte is typically associated with blood (e.g., related to the presence of arterioles and venules), the identification of blood perfused tissue can act as a marker of where detailed OCT measurements and/or analysis of measurements should be performed to obtain estimates of analyte levels. Accordingly, some embodiments are directed to methods and device for analyte level estimation using OCT, where OCT scanning can identify tissue regions having blood flow that are suitable for analyte estimation.

For example, exemplary embodiments are drawn toward systems and methods of determining blood analyte levels in a tissue sample using optical coherence tomography (OCT). In an exemplary method, a plurality of regions in the tissue sample are scanned using OCT to provide corresponding intensity data measurements. Each tissue region can be raster scanned to provide the intensity data. The tissue regions can optionally be scanned over multiple depths. The intensity measurements can include signals from backscattered light and/or can exhibit speckle. The intensity data measurements, and/or speckle, can be analyzed to determine whether blood flow is present in the corresponding region. Blood analyte levels can be obtained from one or more of the regions having blood flow present as analyzed from the OCT measurements.

Scanning of a tissue sample can be performed in a variety of manners. In some instances, multiple regions are scanned over multiple moments in time to provide the intensity measurements. Comparison of intensity data, which can exhibit speckle, taken over the multiple time moments can be used to determine whether blood flow is present. In some instances, intensity data measurements are compared over multiple tissue regions to determine whether blood flow is present.

Other aspects of the invention are directed to systems for determining at least one blood characteristic in a tissue sample. The system can comprise an OCT apparatus configured to obtain intensity data measurements from a plurality of regions in the tissue sample. Such systems can be consistent with any of the OCT systems disclosed in the present application and/or those within the knowledge of one skilled in the art. The system can further comprise a processor configured to analyze the intensity data measurements from the OCT apparatus to determine whether blood flow is present in at least one region. In some instances, the system can be configured to provide different intensity data measurements over a plurality of regions and a plurality of depths. The system can also be configured to provide raster scans of a plurality of regions. Systems that provide scan-

ning over different regions can utilize a beam scanner, which can be appropriately configured to achieve the desired scanning modality. In some embodiments, the OCT apparatus can be configured to provide different intensity data measurements over a plurality of moments in time.

In related embodiments, the processor can be configured to convert the intensity data measurements into analyte levels (e.g., blood glucose levels). The processor can be configured to analyze speckle in the intensity data measurements and optionally compare those measurements to determine whether blood flow is present. In other embodiments, the processor can analyze fringe-modulated data, such as high frequency components and/or amplitude modulation. The processor can also comprise a signal filter for converting at least a portion of the intensity data measurement into a reduced intensity signal, which can indicate the presence of blood flow through the correlation with fringe frequency modulation. The signal filter can comprise a narrow and/or wide band filter to convert at least a portion of the intensity data measurements into an attenuated output, with the processor optionally configured to compare the reduced intensity signal and the attenuated output to determine the presence of blood flow.

In some exemplary embodiments, a method is utilized to determine the presence of fluid flow in tissue using OCT. These embodiments can optionally be combined with other steps to determine blood analyte levels, as described herein. A tissue sample can be scanned over each of one or more tissue regions, for example at various depths, to obtain corresponding fringe-modulated data, which can be analyzed to determine the presence of blood flow. The fringe-modulated data to be analyzed can include determining a peak fringe frequency shift, or analyzing the amplitude modulation in the data, to determine whether blood flow is present. In the latter case, selected high frequency components of the intensity data, which can be related to speckle, can be analyzed to determine whether blood flow is present. In some instances, the intensity data measurements having fringe-modulated data can be passed through a filter, wherein the filter is configured to produce a filtered signal indicating fringe frequency modulation indicative of blood flow's presence. For example, the intensity data measurements can be passed through a narrow band filter, which can be configured to produce a reduced intensity signal output from an intensity data measurement exhibiting blood flow. Optionally, the intensity data measurements can also be passed through a wide band filter, which can be configured to produce an attenuated output exhibiting at least one fringe-modulated feature from an intensity data measurement exhibiting blood flow. Subsequently, the reduced intensity output and the attenuated output can be compared to determine whether blood flow is present. Digital signal processing, analog signal processing, or a combination of the two techniques can be utilized to analyze the intensity data measurements in any fashion.

Other aspects of the present invention are directed to methods, devices, and/or systems that can scan multiple tissue sites. Tissue sites can each be scanned using an optical coupler, which can be attached to a single location on a subject's skin; each tissue site can also, or alternatively, be associated with a separate potential analyte correlating region (herein "ACR") or be separated by a distance at least as large as some distance associated with an ACR. An assessment of the quality of each tissue site to provide acceptable analyte levels (i.e., that the site is a ACR) can be made, followed by analyzing OCT data at one or more of the identified ACRs to provide an estimate of an analyte level

in the scanned tissue. Other aspects of devices and methods that embody some or all of these features are included in the present application.

For instance, some embodiments are drawn toward OCT-based analyte monitoring systems in which the optical interrogation beam is moved across the surface of the tissue in two distinct ways. The first is a coarse tune of the beam position, moving the beam over a large distance in order to provide positions on the skin that are spatially distinct from each other. The second is a fine scan where the beam is applied over one or more smaller areas that are selected by a controller for more detailed analysis.

Coarse position tuning over a large area of tissue can be useful when the tissue is inhomogeneous so that regions of little or no analytic value can be ignored or discounted, and/or regions of high analytic content can be identified for future analysis. For example, hairs with their associated structures (i.e., sebaceous glands, erector muscles, etc.) can be avoided. These structures are large and, due to the depth at which their blood vessels lie, have little to no associated perfusion in the dermis. Thus, coarse scanning over a large spatial area allows the sensor to avoid such structures. Another example is an arteriole and/or venule plexus, a highly perfused region of tissue, and a desirable scanning location for the sensor. These structures are more sparsely distributed and thus coarse scanning over a large area is desirable to ensure that such a plexus can be located. In both cases, the location of these structures cannot easily be discerned by the naked eye and thus the sensor can also accommodate for sub-optimal sensor placement by the user.

The fine or raster portion of the scan can provide a smaller region of tissue from which data is obtained to correlate to glucose changes. The fine scan takes advantage of the high spatial resolution of the system and allows regions of the tissue with a high density of analyte correlating regions (ACRS) to be monitored, improving the system response to glucose.

Some exemplary embodiments are drawn toward optical coherence tomography (OCT) systems for determining an analyte level in tissue. The systems, which can optionally be a non-imaging system unlike many other OCT systems, can include a light source for generating skin-penetrating radiation suitable for OCT measurements and a detector for receiving reflected light from the skin-penetrating radiation. An interferometer can be coupled to the system to form a combined signal from the reflected light from the skin and a reference beam, the combined signal being received by the detector. A phase shifter can be included and configured to shift the phase of the reference beam, which can provide tissue depth scanning.

A beam scanner can be included for directing a beam of radiation from the source to a plurality of sites to provide a measure of analyte levels (e.g., glucose levels) in the tissue at one or more of the sites. Each site can be associated with a different analyte correlating region. For instance, each site can include a non-overlapping area (e.g., each site being a unique spatial area) relative to all other sites. The sites can form a selected pattern and/or substantially cover a designated larger tissue area. The beam scanner can scan within a tissue site in a variety of manners. For example, the scanner can scan the entire area of at least one of the plurality of sites and/or scan in a manner to provide an aggregated OCT measurement that has reduced speckle content. Scanning within a tissue area can be conducted to provide OCT data sufficient to provide an analyte level estimation and/or sufficient to provide a measure of validity of a tissue site as an ACR.

In some embodiments, a beam scanner, which can be part of an OCT apparatus for detecting analyte levels in scanned tissue, can be configured to adjust movement of a beam on at least two-different length scales. One of the length scales can be a site-scale where the beam moves over a distance large enough such that the beam scanner probes a plurality of tissue sites. Another length scale can be a measurement-scale wherein the beam moves over a distance smaller than the site-scale. OCT measurements can be taken over the measurement-scale such that the data obtained can provide analyte measurement correlation.

Beam scanners can include one or more optical elements (e.g., mirror or rotatable prism) configured to move a beam of radiation. For instance a rotatable prism can be used to move the beam to a variety of tissue sites, while a moveable mirror can provide scanning within a tissue site. Other types of devices, such as a flexure, can also be used to control optical element movement.

Also, a controller can be included for selecting sites from the plurality of sites for monitoring. The controller can be integrated into a single unit with a beam scanner, or be a separate device coupled to the beam scanner. The controller can be configured to associate OCT measurements with each of the plurality of sites, and can subsequently provide an aggregated measure of the analyte level using OCT measurements from a plurality of sites. OCT measurements can correspond with a measure of an analyte level in the scanned tissue. In some instances, the controller can be configured to select one or more sites based on whether a site is validated for analyte level measurement (e.g., whether the site is validated based upon a tissue hydration level). OCT measurements from one or more validated site scan be used to calibrate an OCT system.

Some embodiments are directed toward a structure for providing a coupling between an optic and a subject during OCT measurements. One or more optical elements can be optically couplable to an OCT system to guide reflected light. A patch for aligning the optical element with skin of the subject can be included. The patch can include a rigid body coupled to the optical element. The rigid body can be configured to stabilize optical path lengths between the optical element and the skin. A plurality of pliable extensions can be coupled to the rigid body. Each pliable extension can extend away from the rigid body, and can be configured to hinder movement of the rigid body relative to the skin. In some instances, a patch for providing a coupling between an optic and a subject during OCT measurements can have an optical element and a moisture-removing structure configured to contact skin of the subject. The moisture-removing structure can be configured to transport moisture away from the skin of the subject to hinder moisture build up at an interface between the optical element and the skin. Moisture removing structures can include a moisture-absorbent material, and/or a perforated hydrophilic material.

Other embodiments are directed to methods for performing OCT measurements to determine an analyte level in tissue, such as a blood glucose level. Such methods can be implemented without the need to form and/or process an image. An OCT system can be aligned to a subject through an optical coupler. Then a beam can be scanned over a first tissue site using the attached optical coupler to collect a first set of OCT measurements. Such scanning can combine light reflected from scanning a designated tissue site with a reference beam to produce an interference signal, which can be processed to aid determination of the analyte level in the tissue. The beam can then be scanned over one or more other sites to provide corresponding OCT measurements. Each

tissue site can be spatially distinct from the one or more of the other tissue sites. The tissue sites can also form a selected pattern. Scanning within a tissue site can include scanning the beam over the entire spatial area of the tissue site and/or scanning to provide data to obtain an analyte level which can reduce the effect of speckle. A plurality of depths can also be scanned.

Specific sites can be selected from the scanned sites for monitoring analyte levels, with the corresponding OCT measurements being processed to measure the analyte level in the tissue. The selection of sites can include designating whether one or more sites is associated with a validated site. This can be determine, for example, by determining whether corresponding OCT measurements indicate a selected level of tissue hydration. If a two or more sites are validated and corresponded with a measure of an analyte level, the plurality of levels can be aggregated into an aggregated measure of analyte levels in the tissue, which can be used to calibrate an OCT system. Alternatively, data from validated sites can be used to calibrate an OCT system without aggregating the data.

#### BRIEF DESCRIPTION OF THE DRAWINGS

The present application will be more fully understood from the following detailed description taken in conjunction with the accompanying drawings (not necessarily drawn to scale), in which:

FIG. 1 is a schematic diagram of an OCT system, consistent with an embodiment of the present invention;

FIG. 2 is a schematic diagram of a beam scanner implemented to scan tissue, according to one embodiment of the present invention;

FIG. 3A is a schematic diagram of a number of tissue sites to be scanned oriented around a circular path, according to one embodiment of the present invention;

FIG. 3B is a schematic diagram of a number of tissue sites to be scanned oriented in a gridded pattern to substantially cover a tissue area, according to another embodiment of the present invention;

FIG. 3C is a schematic diagram of the tissue site distribution around a circular pattern overlaid with varying analyte correlating regions, according to some embodiments of the present invention;

FIG. 4 is a cross-sectional perspective view of a beam scanner with a scanning optic, according to one embodiment of the present invention;

FIG. 5 is a cross-sectional perspective view of a beam scanner with a two scanning optics, according to another embodiment of the present invention;

FIG. 6A is a schematic diagram corresponding to locations scanned within a tissue site to substantially scan the tissue site, according to one embodiment of the present invention;

FIG. 6B is a schematic diagram corresponding to locations scanned within a tissue site to scan a plurality of locations in the tissue site, according to another embodiment of the present invention;

FIG. 7A is a side view of a skin contacting device, consistent with an embodiment of the present invention;

FIG. 7B is a top view of the skin contacting device of FIG. 7A;

FIG. 8 is a schematic of a flow diagram for detecting blood analyte levels in tissue, according to an embodiment of the present invention;

FIG. 9 is a flow diagram of a method for processing OCT measurements, according to an embodiment of the present invention;

FIG. 10 is a flow diagram of a method for determining whether a site is valid, consistent with one embodiment of the present invention;

FIG. 11 is a flow diagram showing a variety of techniques for aggregating OCT data measurements, according to one embodiment of the present invention;

FIG. 12 is a schematic flow diagram of a multi-filtering method for detecting fluid flow, according to an embodiment of the present invention;

FIG. 13 provides graphs of a difference signal from OCT scanning at varying tissue depths as a function of time, and a corresponding graph of measurements of blood glucose values as a function of time, according to some embodiments of the present invention;

FIG. 14 provides a graphs showing overlapping traces of a perfusion signal and a glucose signal derived from the data shown in FIG. 13; and

FIG. 15 provides a graph of pulsatile blood flow from OCT blood flow measurements, according to an embodiment of the present invention.

#### DETAILED DESCRIPTION

Aspects of the present application that are consistent with some embodiments described herein can improve the use of OCT to determine analyte levels. The type of analytes that can be identified can include any analyte capable of being detected with OCT. In many instances, these analytes are water-based analytes and/or analytes that are present in blood perfused tissue. Examples include oxygen, hemoglobin, glucose, water, components of blood, and other analytes present in tissues or other samples. In many embodiments herein, glucose is an analyte of interest. An analyte correlating region (ACR) is a region of a subject's tissue in which OCT can be applied to provide an accurate estimate of an analyte's level in the tissue. Accordingly, using OCT to provide measurements substantially within a ACR can yield data allowing an accurate estimate of the analyte level.

For a selected analyte, regions of the tissue in which an OCT signal is sensitive to changes in the analyte level are not typically uniformly distributed, either as a function of depth or laterally across the tissue. For example, analytes that are present in blood perfused tissue (e.g., glucose) can be non-uniformly distributed due to the non-uniform distribution of blood vessels in the tissue. In particular, the potentially high resolution of OCT, both depth and lateral, enables the location of ACRS, for example glucose correlating regions (herein "GCRs"), to be identified. However, since the resolution of an individual OCT measurement is typically much higher than the scale of many ACRS, it can be difficult to insure that these regions will exist at a given tissue location which is scanned using conventional techniques. Thus, for example, an OCT glucose monitor's performance in the past has been highly dependent on the location of the skin upon which the sensor is placed.

Though OCT can be applied using a brute force technique, analyzing a large tissue region in depth and extent followed by extensive data analysis to isolate portions of the data that do correspond with good ACRS, such a methodology is costly in terms of time and effort expended to produce accurate analyte level readings. As well, the equipment and data processing commensurate with such effort can be a financial burden. Accordingly, if one can utilize an OCT technique to quickly identify tissue regions with substantial

blood flow, more extensive OCT measurements can potentially be performed, and data analyzed, therewith to yield faster and more accurate results. In addition, the identification of ACRS can be beneficial by reducing the effort needed to utilize OCT to accurately identify an analyte level. Furthermore, if ACRS can be identified more readily, OCT measurements can be less dependent upon the placement of the sensor vis-à-vis the patient's skin location. Some embodiments of the present invention can potentially address some of these problems. By utilizing embodiments disclosed herein, such as techniques that provide multiple tissue site OCT sensing and/or assessing whether the spots are ACRS, (e.g., by blood flow detection) an OCT technique can focus data gathering or analysis on the ACR regions to collect an analyte level without performing unneeded data collection and/or analysis on regions which do not correlate analyte levels acceptably.

Other potential advantages can also be accrued. For instance, sensor placement by the user can be less critical. Since the ACRS within the tissue are not visible to the unaided eye, it is not possible for the user to place the sensor on a region of tissue that has these structures. Further, there are structures within tissue that do not correlate with analyte changes, such as sebaceous glands associated with hairs in skin when attempting to sense glucose. In this case again, sebaceous glands are not visible to the user, thus a sensor with the ability to monitor multiple tissue sites over a wide spatial extent provides less sensitivity to sensor placement.

In another instance, monitoring multiple tissue sites potentially provides multiple signals that can be used to improve the sensitivity, stability, and signal-to-noise levels of an OCT system relative to the analyte changes in the subject.

Also, diversity of tissue sites can provide stability to the OCT system's analyte correlation by minimizing the effects of physiologic variances in a subject during monitoring. An example of this would be changes in microcirculation on the subject from pharmacologically or temperature induced changes. In this case, tissue sites where the blood perfusion changes dramatically (up or down) can be identified and weighted by the sites where the perfusion remained more constant generating a more stable output of the sensor. Systems and Devices for OCT Scanning

Some embodiments are directed to systems for performing OCT to determine the presence of blood flow in a scanned tissue region. Some additional embodiments are directed to systems for performing OCT to determine an analyte level in tissue. Other embodiments combine such detection with other features to perform analyte level estimation with an OCT system. Such embodiments can utilize an OCT system to provide data intensity measurements that are analyzed by a processor. Exemplary OCT systems can include a number of components such as a light source for generating skin-penetrating radiation suitable for OCT measurements, interferometers, and/or a detector for receiving reflecting light from the skin-penetrating radiation. The system can also include a beam scanner, which can be used to direct a beam of radiation from the source to one or more tissue sites. Beam scanners can include devices that operate in accord with the disclosure of the present application and/or operate in a conventional manner. Data collected from one or more of the sites can be processed by a processor to provide a measure of the presence of blood flow or the measure of analyte levels at the respective sites using any combination of the methodologies disclosed herein and/or known to those skilled in the art. The term "processor" is used generally to include any actual number of

devices configured to perform the designated processes. Accordingly, a processor can be embodied as any combination of a programmable unit, a microprocessor, an embedded preprogrammed processor, a stand-alone computing unit, or other devices and configurations as understood by those skilled in the art. The system can also include a controller, which can be used to designate a tissue site associated with blood flow or can be used to select one or more of the multiple sites for further processing (e.g., extracting analyte data therefrom).

As utilized herein, a tissue site can be a selected tissue region (e.g., an area or volume of the tissue) which can be probed using OCT. In many instances, a tissue site has a geometrical size larger than the resolution of the OCT system. Accordingly, in some instances, the tissue site is at least large enough to allow an OCT system to perform scanning within the site, which can potentially lead to data of an analyte level within the site.

Some exemplary embodiments of OCT systems can be described with reference to the schematic diagram of an OCT system shown in FIG. 1. A light source 110 is used to generate skin-penetrating radiation 115 suitable for OCT measurements. A interferometer is employed to perform the measurement. For example, the interferometer can include a beam splitter 140 which can direct a measurement light beam 141 to an optical element 160 and a reference light beam 142 to a phase shifter 130 (e.g., a mirror). The sample beam 141 is directed by the optical element 160 to a tissue sample 120 (e.g., a skin portion of a subject). Light 121 is reflected from the tissue 120 and directed to the interferometer. The reference beam 142 is reflected by the phase shifter 130 back to the interferometer, where the phase shifter can be positioned to change the phase of the reflected light. The interferometer can combine the light from the phase shifter and tissue sample, and direct the light to a detector 150. The constructively interfered light will provide a signal to the detector 150. Such constructively interfered light thus interrogates a depth of the tissue 120 selected by the distance traveled by the sample beam 142 to the phase shifter 130. As used herein, tissue depth is typically the shortest distance from a location in the tissue to a tangent plane on the tissue surface.

In some embodiments where fringe frequency modulation is utilized, the phase shifter 130 is moved at a constant velocity  $V$ , which results in a fringe frequency proportional to the ratio of  $V$  to the wavelength of the radiation. When the sample beam strikes a moving target, this fringe frequency is modulated. Accordingly, some embodiments herein utilize the modulation of the fringe frequency to detect blood flow.

The general use of OCT systems to provide estimates of analyte levels in tissues is described in a number of references. For example, U.S. Pat. No. 7,254,429 provides descriptions of methods, devices, and systems which can be used to perform OCT measurements in tissue; U.S. Patent Application Publication No. U.S. 2006/0264719 A1 provides techniques for calibrating an OCT system for performing analyte measurements; and U.S. Patent Application No. U.S. 2006/0276696 A1 provides techniques for performing analyte measurements and calibration that include the use of multiple wavelengths of radiation. The teachings of each of these references is included herein in their entirety. Accordingly, OCT systems consistent with embodiments herein can utilize any one or combination of features of OCT systems described in the references herein. For example, in some embodiments, the light source is a low coherence source, and/or the system can be specifically configured for detecting glucose (e.g., blood glucose). OCT systems can be

configured as non-imaging systems, which can optionally allow scanning of a tissue site such as to reduce the effect of speckle. In another example, the light source of an OCT system can utilize one or more wavelengths to enable glucose reading detection. Other aspects of OCT systems that can be utilized with embodiments herein can include features understood by those skilled in the art.

#### Beam Scanners and Controllers

Some embodiments herein include a beam scanner which can be configured to direct beams of radiation to one or more tissue sites for an OCT system. Such beam directing can be achieved in a variety of manners. For instance, as shown in FIG. 1, a beam scanner 170 can be coupled to one or more optical elements 160 to direct the beam to a particular site. A beam scanner, however, can also be coupled to any combination of optics and/or other light directing devices to provide the scanning. For example, the beam splitter, optionally in conjunction with one or more other optical elements, can also be coupled to the beam scanner such that the scanner can direct scanning of light impinging on a sample region. A controller 175 can also be included. The controller, which can be coupled to the beam scanner, can be used to facilitate operation of the beam scanner in a variety of manners (e.g., identify and/or direct which tissue sites should be utilized to provide an analyte measurement). A controller can be an integral portion of a beam scanner (e.g., within the inner workings of a single device) or can be embodied separately (e.g., as a processor of a microcomputer which is in electrical communication with a beam scanner). It should be understood that a beam scanner coupled with a controller can be part of an OCT system, or alternatively constitute an embodiment by itself.

The imaging optics of a beam scanner can be configured to have a variety of characteristics. First the optics can be configured to reduce variations in focal depth during the scanning at widely differing spatial positions. Also, the optical train of a beam scanner can be configured to limit unwanted spurious reflections that can corrupt the interferometer signal such as from optical surfaces at critical positions in the optical path. The numerical aperture of the optics can be selected to reduce the returned scattering signal while minimizing the confocal effect through the depth of interest in the skin. Aberrations due to far off axis scanning through the optic can also be desirably reduced.

Beam scanners can be configured to scan one or more tissue sites where the distinction between sites and the spatial extent of a particular tissue site can also be varied. For instance, each site can be chosen to have a unique region relative to every other site (i.e., none of the sites overlap). The sites can substantially cover a tissue section to be investigated, or can be distributed within a tissue section with space therebetween. The sites can form a selected pattern or can be randomly distributed therein. In some embodiments, the sites can be spaced farther apart than some characteristic length scale of an ACR. Such embodiments can allow an OCT system to quickly assess a tissue sample to locate ACRs where further measurements can be taken and/or analysis of particular measurements is to be performed to reduce the effort of data gathering and analysis. The characteristic length scale of an ACR can depend upon the analyte to be estimated and/or the type of tissue to be scanned. For example, with glucose correlating regions (GCRs), the characteristic length scale can be on a length associated with the distribution of blood vessels in a dermal region.

The spatial extent of a tissue site can also selected in a variety of geometries. For instance, different tissue sites can

be distributed over a certain area with all the sites being substantially at the same depth. In other situations, the tissue sites can span a variety of depths and/or spatial extents, encompassing various volumes.

With regard to the size of a tissue site, as mentioned earlier, each tissue site can be a tissue region large enough that an OCT system can probe a variety of locations in the tissue site to determine an analyte level in the region (e.g., the region can be large enough so that analyte correlation in the region is possible). As well, the tissue site can be a region large enough to encompass an ACR, or to perform some other functionality associated with analyte detection (e.g., large enough to allow scanning therein to reduce speckle in an OCT system in which an analyte level is being estimated as described in U.S. Pat. No. 7,254,429). The tissue site can also be limited in size such that scanning in the site can be relatively rapid vis-à-vis scanning the entire area covered by a plurality of sites. The tissue size can also be limited such that such that the effects of structures that do not change with an analyte (such as a hair on glucose sensing) on one tissue site does not influence the signal received in another tissue site. The relative sizes each of the tissue sites can also vary, or be substantially uniform.

As well, the beam scanning can be performed using various data acquisition techniques to provide OCT data that can be used for a variety of purposes. For example, the scanning within a tissue sample can be sufficient to provide data allowing an analyte level measurement within the tissue sample. In another example, the scanning can provide data that allow validation of a site, as described herein, to determine whether a tissue site is an ACR which can provide good analyte level estimation. For instance, validation of a site can depend upon detecting the presence of blood flow. The latter example, can potentially lead to a more rapid data collection methodology where sites can initially be scanned for validation—and the validated sites are rescanned to provide more detailed OCT data for analyte level measurements within each validated site. Other varieties of scanning within a tissue site are also contemplated.

In some embodiments, beam scanners for use with OCT systems can adjust movement of a beam over multiple length scales, e.g., at least two different length scales. For instance, the beam scanner can move a beam over a site-scale, where the beam is moved over a distance large enough that the beam scanner probes a plurality of tissue sites within the distance of a coarse scale (e.g., site scale). Accordingly, the beam can interrogate a multiplicity of tissue sites (e.g., non-overlapping sites, and/or sites separated by a distance larger than a characteristic length of an ACR). The beam scanner can also be configured to move the beam over a measurement scale. Typically, the measurement scale is smaller than the coarse scale. OCT measurements can be taken within the measurement scale to provide an analyte level estimate in the tissue (e.g., the measurement scale has a size sufficient to allow analyte measurement correlation). It is understood that scanning over a coarse scale can incorporate any of the features regarding different tissue sites (e.g., the site-scale scans over tissue sites that do overlap one another), while scanning over the measurement scale can incorporate any of the features described herein regarding scanning within a tissue site (e.g., scanning in a tissue site to reduce speckle from a composite signal). It is understood that scanning can also be performed over additional length scales as well.

Some illustrations of features previously described with respect to beam scanners are described with reference to FIGS. 2 and 3A-3C. FIG. 2 provides an illustration of some

features of particular embodiments of the present invention which utilize a beam scanner. The scanner 210 can comprise a body of a device with one or more optics 261, 262. The optics 261, 262 can be configured to direct light from an interferometer 240 to an imaging optic 263, which further directs the light to tissue 250. By configuring the optics 261, 262 appropriately, a beam that impinges upon tissue can be manipulated to provide multispot scanning. For example, one optic 261 can position the beam over multiple tissue sites, while another optic 262 can provide scanning within a tissue site, which can provide data for analyte measurements. Alternatively, a single optic can provide both the coarse and fine scanning functions (e.g., a galvanic reflecting mirror driven for large scale beam displacements in the coarse tuning mode and fine dithering in the analytic mode).

FIGS. 3A-3C provide some exemplary orientations for tissue sites which can be scanned with an OCT device. As shown in FIG. 3A, a beam scanner can be configured to distribute tissue sites to be scanned around a circular path. The beam scanner can thus move the beam around a diameter 13 to distribute tissue sites therearound. As well, the beam scanner can move an OCT beam within a length scale 23 to provide OCT measurements which can be correlated to an ACR potentially. FIG. 3B presents another pattern of tissue sites that can be scanned with an exemplary beam scanner. The tissue section 310 can be gridded into a set of adjacent sites 311 that substantially cover the tissue section 310. Accordingly, a beam scanner can be configured to move the beam such that tissue sites are distributed over a length scale 14, while each tissue site is interrogated on a smaller length scale 24. It is understood that for any tissue site pattern, the number of tissue sites to be distributed, the shape of the pattern on which tissue sites are distributed, and the scanning within the tissue site can all be varied. For example, the tissue site need not be square shaped, and the entire tissue site need not be interrogated to obtain data for OCT scanning.

FIG. 3C schematically depicts a tissue site 325, 326, 327, 328 distribution pattern to be interrogated by a beam scanner overlaid with the analyte correlating regions 330, 331, 332 of a tissue sample 320. As depicted, some tissue sites 325, 326, 328 intersection portions of ACRS 330, 332, 331, while one tissue site 327 does not interrogate an ACR. Accordingly, by identifying the tissue sites that intersect ACRS, a more targeted analysis and gathering of OCT data can be performed to obtain analyte levels in the tissue 320.

FIGS. 4 and 5 depict embodiments of beam scanners that have been reduced to practice. FIG. 4 depicts a beam scanner 400 which utilizes two discrete optical elements. One optical element can locate tissue sites by utilizing a rotating wedge prism 410, which can displace the tissue sites around a circle, as depicted in FIGS. 3A and 3C. The other optical element can provide the scanning within a tissue site (e.g., raster scanning). This element can be embodied as a moveable mirror, for example a mirror coupled to a flexure. The flexure can be attached to electromagnetic coils to affect its positioning of the mirror, which allows scanning within the tissue site.

FIG. 5 depicts another embodiment of a beam scanner, which utilizes a single optical element for beam steering. The optical element is coupled to a flexure, where the flexure is configured to move the optic (e.g., a mirror) with a greater deflection angle, relative to the flexure used in the device shown in FIG. 4. In this particular flexure, all positions of within a selected spatial extent are scannable, akin to what is shown in FIG. 3B. Accordingly, a greater variation in designation of tissue sites, and scanning within tissue sites,

can be achieved with this design. It is understood that other beam scanners can utilize any number of optical elements and/or devices to manipulate the optical elements, in a manner consistent with the functionality of the described devices, including using such elements within the knowl- 5 edge of one skilled in the art.

With regard to scanning within a tissue site, beam scanners can be configured to perform such scanning using various methodologies. The scanning can be sufficient to substantially cover an entirety of a tissue site as shown in FIG. 6A, or can scan a variety of locations in the tissue site that are spaced apart. For instance, the scanning can be such that enough locations are scanned in a manner to allow reduction of speckle when the OCT measurements are combined to estimate an analyte level. In some embod- 15 iments, a raster scan is performed within the tissue site. The raster scanning can be performed in any designated pattern, with as many or few locations actually probed as desired; one example is shown in FIG. 6B. For example, the raster scanning can be of an area, a line, or a volume of locations. 20 In many instances, the raster scanning is done at a plurality of depths of tissue.

As discussed earlier, a controller can be incorporated with an OCT system and/or a beam scanner to aid in the scanner's operation. For example, a controller can be used to associate each of a plurality of tissue sites with particular OCT measurements taken therefrom. The OCT measurements can be used to estimate selected analyte levels therefrom, or can be used to "validate" the site for estimating selected analyte levels (e.g., using the data to determine whether the site 30 covers a sufficient amount of an ACR to extract analyte level data therefrom). In some embodiments, the controller can select the tissue sites (e.g., the validated tissue sites) that will be interrogated in further detail, for instance taking further OCT measurements to determine analyte levels from data 35 taken therefrom. Alternatively, if data for determining analyte levels has been extracted from a set of tissue sites previously, the controller can act to determine which tissue site(s) should be examined to determine analyte levels.

Validation of a tissue site for OCT analyte measurement can be performed using any number of techniques including those known to one skilled in the art for analyzing OCT data. Some embodiments utilize OCT measures of tissue hydration as a technique for validating a tissue site. When an analyte is associated with the level of hydration in a tissue 45 sample (e.g., the amount of blood perfused tissue), one measure of whether the site encompasses an ACR is a measure of the hydration level in the tissue. As discussed in U.S. Pat. No. 7,254,429 and U.S. Patent Application Publication No. U.S. 2006/0276696 A1, a section of tissue can be 50 scanned at two wavelengths, one wavelength that has a low absorption coefficient for water/blood components and a high scattering coefficient for such components (e.g., 1310 nm), and one wavelength that has a high absorption coefficient for the water/blood components (e.g., 1450 nm). By comparing the intensity of backscattered light over a tissue sample at each of these components, one can establish a measure of hydration in the tissue sample. For example, the difference in the average intensity level of OCT backscattering of a tissue site between two wavelengths can be calculated and serve as a validation metric. When the average difference exceeds some threshold value, the site can be designated "valid," while a measure below the threshold value can be designated "not valid." Since the average intensity value depends upon the spatially distributed backscattering intensity, validation of a tissue site can also depend upon the extent to which an ACR overlaps the

interrogated tissue site. Referring back to FIG. 3C, a tissue site 327 which does not intersect an ACR will have a low difference value. Tissue sites 325, 326, 328 that intersect an ACR will provide a difference value that depends not only on the analyte level in the ACR but the relative overlap of the ACR with the tissue site. Accordingly, one tissue site 326 can exhibit a higher difference value than other tissue sites 325, 328 that intersect ACRs even if all the ACRs have the same analyte levels therein. Variations on such techniques are also included within the scope of the present application. For example, validation of a tissue site can be performed by examining OCT data taken at two different wavelengths where the wavelengths are selected to provide a contrasting signal indicative of the presence of a particular analyte, for example oxygenated hemoglobin vis-à-vis deoxygenated hemoglobin as discussed in U.S. Patent Application Publication No. U.S. 2005/0059868 A1. It is also understood that combinations of validation techniques can be implemented to provide tissue site validation.

Some embodiments are directed to beam scanners and/or controllers that are configured to validate and/or select a site based upon an OCT measurement indicating the presence of blood flow. The measure of blood flow can be identified in a variety of ways, as described within the present application. Accordingly, embodiments of the invention are directed to devices (e.g., OCT systems, beam scanner, and/or controllers) that are configured to detect fluid flow (e.g., blood flow) using any combination of the techniques discussed herein. Such blood flow can be associated with an analyte, and therefore the validated tissue site with blood flow can be subjected to further OCT measurements and/or data analysis to extract analyte levels.

As previously mentioned, validated tissue site(s) can be interrogated and/or data analyzed therefrom to associate a corresponding analyte level therewith. In some instances, a controller can aggregate the analyte levels from two or more validated sites to provide an aggregated measure of the analyte level in the tissue. Such aggregation can be an average value, a median value, or other aggregated value as understood by one skilled in the art; some further aspects of how aggregation can occur, and thus be implemented in a controller, are discussed herein. Alternatively, or additionally, the analyte levels corresponding with one or more validated sites can be used to calibrate an OCT system such that future OCT measurements utilize the calibration. Such calibration can utilize any of the techniques described in U.S. Patent Application Publication Nos. U.S. 2006/0264719 A1 and U.S. 2006/0276696 A1. 65 Optic/Skin Contacting Devices

Some embodiments of the present invention are directed to a structure for coupling an optic to a subject, which is configured for use with an OCT measurement. Some exemplary embodiments include the structure itself, an OCT system with such a structure, or a patch. Embodiments can include one or more features as described herein. Such coupling structures can be configured to reduce mechanical distortion of the skin at a measuring site. The distortion includes two types. First, the local distortion of the skin due to the protrusion of the optic surface into the skin. Reducing this effect can hinder longer term stretching and/or the need to accommodate a continuous change in the baseline signal due to the stretching. Such changes, referred to as settling, are not correlated with analyte changes and represent the viscoelastic response of the skin to the pressure of the optic. At the same time, however, protrusion of the optic can be desirable to insure stable continuous contact of the measuring surface with the skin under all conditions of force on the

sensor or assembly. The second category of distortion is due to forces on the sensor such as sensor weight, inertia during movement, or brushing the sensor against an object. Also, a sensor cable can transmit some force to the assembly in the form of a torque or a pull, causing motion that deforms the skin macroscopically over the disposable attachment region. Examples are a twisting of the skin that may rearrange the analyte correlating structures in depth, or a pull that due to the varying modulus of the features in the skin causes a change to the baseline interferometer signal.

Some exemplary embodiments are directed to a structure for coupling an optic to a subject that include one or more optical elements that can be coupled to an OCT system to guide reflect light from skin. One particular embodiment is shown in FIGS. 7A and 7B. An optic 740 can include a protruding portion 745 for contacting the skin of a subject. A patch can be used to align the optical element on the subject's skin. The patch can include a rigid body 720 coupled to the optical element 740 to stabilize the optical path lengths between the element and the skin. The optical element 740 can optionally be removably couplable with the rigid element. Accordingly, in some embodiments, the patch section can be configured as a disposable element, with the optic being reusable. The patch can further include a plurality of pliable extensions 710 coupled to the rigid body 720. Each extension can be configured to extend away from the rigid body, and configured to hinder movement of the rigid body relative to the skin. Extensions can be constructed with any suitable material having a flexible quality; non-limiting examples include polymer-based materials that can optionally have elastic qualities. Adhesive 730 can be applied to couple the structure to the skin. It is understood that the precise configuration of such a structure as shown in FIGS. 7A and 7B does not limit the scope of such structures.

Other embodiments are directed to a optic-skin interface that can help remove moisture from the interface. If moisture from the skin is allowed to build up at the interface, the optical interface to the skin can be distorted and the characteristic shape of the baseline scan can change. Accordingly, one exemplary embodiment is directed to a patch for OCT measurements that includes an optical element and a moisture removing structure 750 as shown in FIG. 7A; it is understood that the moisture removing structure need not be incorporated with all the exact features depicted in FIG. 7A. The moisture removing structure can be configured to transport moisture away from the skin to hinder moisture build up at the interface between the optical element and the skin. The moisture removing structure can utilize any number of moisture-absorbent materials and/or hydrophilic materials. For example, a hydrophilic foam can be used which allows moisture to be removed by capillary action from the skin and transported to an exterior surface of the foam for evaporation to the external environment. A moisture removing structure can also be embodied as a perforated structure around the optic, which can allow moisture to evaporate directly through the perforations.

It is understood that other features for coupling OCT optics to a subject can also be incorporated within the scope of the present invention. Some of these features are described in U.S. Patent Application Publication No. U.S. 2007/0219437 A1.

#### Methods of Estimating Analyte Levels Using OCT

Some embodiments are directed to methods of determining or estimating analyte levels in a tissue using OCT measurements. These methods can be practiced using any number of devices including any appropriate combination of the devices revealed in the present application. However, it

is understood that such methods can also be implemented using other devices, including those known to one skilled in the art, albeit specifically configured to practice the embodied method. Accordingly, these methods are not limited by the specific devices, and their delineated operation, revealed elsewhere in the present application. Likewise, the methods described herein, and portions thereof, can be implemented on devices described in the present application. Accordingly, some devices, such as a beam scanner and/or controller, can be configured to carry out portions or the entirety of a method consistent with some embodiments of the present application.

Some embodiments of methods for estimating analyte levels in tissue are consistent with the flow diagram depicted in FIG. 8. One or more tissue sites can be scanned using OCT 810. Scanning can occur over a two-dimensional region (e.g., at a specified depth), or can be scanning over a volumetric region (e.g., scanning a designated area at a multiplicity of depths). A tissue site can be defined as discussed with respect to beam scanners herein. For example, the tissue site can be spatially distinct from any other (e.g., not overlapping), and/or can be spaced apart from one another. Next, the data collected from the scanning can be analyzed to determine whether blood flow is present in the corresponding tissue site 820. The data can correspond with backscattered light from the OCT measurements, which can potentially carry information about the presence of blood flow. Finally, a measure of a blood analyte level can be obtained from one or more of the tissue sites that exhibits blood flow from the analyzed data 830. Measurement can include the steps of obtaining OCT data from the sites where blood flow is present to obtain data sufficient to estimate an analyte level, and/or analyzing pre-obtained OCT data in a manner to extract analyte levels.

Some embodiments of methods for estimating analyte levels in tissue are consistent with the flow diagram depicted in FIG. 9. An OCT system can be aligned using an optical coupler to a subject 910. For instance, a coupling can be attached to a subject through which light is transmitted and reflected to obtain OCT measurements.

A plurality of tissue sites (e.g., at least two) can be scanned using the optical coupler 920. Scanning can occur over a two-dimensional region (e.g., at a specified depth,) or can be scanning over a volumetric region (e.g., scanning a designated area at a multiplicity of depths). The tissue sites can be defined as discussed with respect to beam scanners herein. For example, the tissue sites can be spatially distinct from one another (e.g., not overlapping), or can be spaced apart using some characteristic length scale, as discussed earlier. As discussed earlier, any combination of the OCT measurement techniques discussed by the references disclosed herein, or as utilized by embodiments including a beam scanner as discussed herein, can be utilized to perform the scanning step. In one example, the entirety or portions of a tissue can be scanned as depicted in the selected patterns of FIGS. 3A and 3B, or any other selected pattern. As well, within a tissue site the entirety or a portion of the tissue site can be scanned. In another example, scanning can be performed within a tissue site to provide an OCT measure of an analyte level that is reduced in speckle content and/or is oriented toward a non-imaging technique. Scanning of a tissue site can be performed to render a variety of data types as well. For example, the scanning can provide data that can be converted to an estimate of analyte levels, and/or the scanning can provide data that allows validation of a tissue site as described in the present application.

One or more of the scanned tissue sites can be selected **930** for further processing. Such selection can be performed either after scanning of all tissue sites is completed, or at any time after a particular tissue site has been scanned. In some embodiments, a survey of tissue sites can be initially scanned to determine which will be selected for OCT analyte estimation processing.

Alternatively, or in addition, sites can be investigated in an on-going manner (e.g., after a selected number of interrogations for analyte level estimation on a cyclic basis) to continue to validate that the tissue site is appropriate for analyte level estimation. Such on-going selection of tissue sites can act to monitor the potential perturbations and/or physiologic drift that can occur in tissue to alter the location of an ACR. Use of an on-going site selection monitoring technique can result in an adaptive technique that helps maintain the accuracy of an analyte level estimator. In some embodiments, the selection is performed by validating the tissue site using any of the variations discussed within the present application (e.g., using a measure of tissue hydration level, or using a multiple wavelength measuring technique that highlights the presence or absence of a selected analyte).

OCT measurements at the selected tissue sites can be processed to estimate analyte levels **940**. For example, processing can include using data at validated sites to calibrate an OCT sensor. Such calibration can be as previously mentioned in the present application. In another example, when more than one site has been selected (e.g., validated) analyte levels corresponding with each of the validated sites can be aggregated to provide an aggregated measure of the analyte level in a tissue. Aggregation can also include aggregating all the data associated with selected tissue sites at once without determining an individual analyte level estimate for each tissue site. Aggregation can be performed in a variety of manners. Some non-limiting examples include: simple or trimmed mean/median of all signals, weighted averaging based on signal properties, weighted averaging based on input from the site selection algorithm, or some kind of cross-correlation/convolution of the signals.

FIG. **10** provides one example of a method for performing site selection. Scans (e.g., depth scans) are performed for a number of tissue sites, with depths being probed by a raster scan **1010**. Next, site metrics are applied **1020** to determine if a site is acceptable. Such site metrics can include site validation determination using OCT scans at different wavelengths of light to characterize tissue. If a site is acceptable **1030**, **1040**, its OCT measurements can be used to identify an analyte level, such as by using its measurements in an aggregation algorithm with other acceptable site data. If the site is not validated **1030**, **1050** it can be considered unacceptable for measuring analyte levels. It is understood that site that are designated unacceptable may be acceptable at later times, for example when implementing an on-going validation site methodology.

Aggregation can take place at several points in the data process as exemplified in the flow diagram of FIG. **11**: First, depth scans can be collected **1110** from raster scanning a number of tissue sites. In one instance, the individual depth scans can be validated using a site selection algorithm **1120**, **1121**, **1122** with the selected sites being aggregate to form a base scan **1130**. An analyte converting technique (e.g., glucose conversion technique **1140**) can be applied to the aggregated base scan to form an analyte level measure, which can subsequently be used to calibrate an OCT system **1170**. In another instance, the individual depth scans can be averaged to form corresponding base scans **1150**. The base

scans can each be subjected to a site selection algorithm **1121** to determine which are valid. The valid scans can then be further analyzed to report analyte levels for calibration of an OCT system. The analyte levels for corresponding to each valid site can be aggregated **1180**, or the OCT data of the valid sites can first be combined into a single scan **1160** before the analyte level is determined for the base scan.

Fluid flow detection using OCT can be achieved using a variety of methodologies. One methodology involves detecting changes in a speckle pattern indicative of the presence of fluid flow. In many instances, each raw OCT scan can exhibit a high degree of speckle in the signal due to the micro-roughness of the skin and the coherence of the optical source. Some aspects of this effect are described in U.S. Pat. No. 7,254,429. This speckle "noise" on the underlying depth intensity signal is a function of the optical coherence length, wavelength, distance from the focal point, size and shape of the scattering centers, and optical properties of the scatterers and the surrounding media. In speckle flowometry, the particles in the flowing fluid (e.g., red blood cells in blood) can cause the speckle pattern in the region of the flow to change in time more rapidly relative to the change in the speckle of the surrounding tissue. Tissue sites with a high degree of speckle temporal variations are sites where blood flow exists. Thus, when a blood analyte (e.g., blood glucose) is sought to be identified, the presence of blood flow can correspond with the presence of an ACR.

Accordingly, some embodiments are directed to techniques that utilize OCT intensity data measurements that exhibit speckle. The speckle in these measurements can be analyzed to determine whether blood flow is present, for example by comparing two or more different intensity data measurements. Comparison of intensity data measurements for speckle differences can occur in a number of ways. In one instance, two or more scans at one location at a particular tissue depth are repeated at different moments in time (e.g., over a time period of less than about a few seconds). Comparison of the corresponding intensity data sets can indicate the presence of fluid flow if the changes in speckle are substantial. If this is also performed at varying depths, providing intensity data sets of a set of locations at different times and depths, comparison of the intensity sets can also indicate where blood flow is more of less prevalent depending upon the relative temporal changes in the speckle pattern as different times (e.g., more change corresponding with more flow). It is understood that more than one location could potentially be scanned, with the trade-off of requiring more time to complete which can complicate a temporal analysis.

In another instance, a set of scans (e.g., raster scans) is performed at a variety of locations (e.g., linear raster scans) at varying depths. The depths that are more similar to one another in speckle pattern can be indicative of less blood perfused tissue, while depths that differ in speckle pattern can be more indicative of the presence of blood perfused tissue. This technique can potentially require less data than the formerly described technique, but with a trade off of signal to noise ratio.

Another methodology for detecting fluid flow using OCT relies on techniques associated with Doppler OCT, or more particularly in examining the frequency modulation in an OCT fringe signal due to the interaction with fluid flow. As mentioned earlier, when the phase shifter in a reference arm of an OCT system moves at a constant velocity, a corresponding fringe frequency is established. In fringe frequency modulation, moving particles in a flowing fluid can cause a shift in the observed frequency of the interferogram

in the region of the blood flow. Accordingly, this technique utilizes some principles like what is exploited in ultrasonic imaging of blood flow. Previous work in Doppler flowometry has adopted this technique for OCT, and relies on analyzing the shift in the peak frequency. Examples of such work include Zvyagin, et al., "Real-time detection technique for Doppler optical coherence tomography," *Optics Letters*, Vol. 25, No. 22, Nov. 15, 2000, pp. 1645-47; Davé et al., "Doppler-angle measurement in highly scattering media," *Optics Letters*, Vol. 25, No. 20, Oct. 15, 2000, pp. 1523-25; Zhao et al., "Doppler standard deviation imaging for clinical monitoring of in vivo human skin blood flow," *Optics Letters*, Vol. 25, No. 18, Sep. 15, 2000, pp. 1358-60; Zhao et al., "Real-time phase-resolved functional optical coherence tomography by use of optical Hilbert transformation," *Optics Letters*, Vol. 27, No. 2, Jan. 15, 2002, pp. 98-100; Barton et al., "Flow measurement without phase information in optical coherence tomography images," *Optics Express*, Vol. 13, No. 14, Jul. 11, 2005, pp. 5234-39. Doppler flowometry can be very sensitive to orientation of the flow relative to the incoming beam from the OCT sensor which provides a challenge in extracting quantitative blood flow measurements.

Accordingly, some embodiments are directed to techniques for detecting fluid flow (e.g., blood flow) by scanning one or more tissue sites to collect fringe modulated data and analyzing the fringe modulated data for the presence of blood flow in the tissue site. Such measurements can be made at varying depths, which can allow the fringe frequency to be determined.

Determination of the presence of fluid flow can be achieved in a variety of manners. A shift in fringe frequency components (e.g., shift in the peak fringe frequency) relative to what is expected from a moving phase shifter can indicate the presence of fluid flow. Using other techniques known to those skilled in the art, further data can be extracted including directionality of the scatterers and a velocity profile. Such detailed information, however, can require substantial data processing. Accordingly, some embodiments utilize a digital signal processing to extract the shift in a peak frequency. For example, when a phase shifter is configured to provide a 600 kHz baseline signal, oversampling by a factor of 5 or more will typically utilize a 3 Msps system with at least 12 bits of data conversion resolution.

Some embodiments are directed to techniques that examine the amplitude modulation of an OCT signal (e.g., vis-à-vis a tissue sample with no fluid flow) to determine if fluid flow is present. Amplitude modulation is distinct from looking at the shift in the peak fringe frequency, and can include such observables as the examining the intensity changes at selected frequencies, the frequency broadening, and/or a shape change in a fringe-frequency envelope. For instance, the backscattered OCT signal can be examined for a change in total intensity of the signal (e.g., over a given bandwidth), and/or a measure of the broadening of the fringe-frequency envelope relative to the initial OCT signal. As an OCT beam interacts with moving particles, not only does the main frequency shift due to the movement but the breadth of the frequency components can also grow. Without necessarily being bound by any particular theory, it is believed that speckle plays a role in frequency broadening. Since speckle has a tendency to create high frequency components in the data, such components are more likely to be shifted by interaction with scatterers. Accordingly, detection of the amplitude modulation of the envelope of frequencies can be an indicator of the presence of fluid flow. For example, digital signal processing can be applied to a

backscattered signal to create a measure of the amplitude modulation (e.g., looking at a measure of the shape change in the envelope relative to what is expected for tissue without fluid flow or relative to an initial signal).

In some embodiments, examination of the amplitude modulation can also allow the use of less data intensive techniques for signal processing relative to known techniques that follow the Doppler shift in a peak frequency. If properties such as directionality and relative velocity are not to be determined, simpler signal processing techniques (e.g., analog signal processing) can be utilized, though more advanced data processing is not precluded. For instance, the integrated power of an interferogram can be examined to determine the presence of fluid flow. Thus, detection of the envelop shape of a return signal, relative to an expected shape, can be enough to determine the presence of fluid flow. Thus even though directionality information is lost, an indication of fluid flow is more easily identified with less data processing. It is understood, however, that techniques susceptible to use with analog signal processing can also substitute digital signal processing, consistent with some embodiments of the present invention.

For instance, in some embodiments, a backscattered wide band signal from an OCT system is filtered to determine the presence of fluid flow. The filtered signal of a backscattered wide band signal can be used, in some embodiments, to produce a filtered signal indicating fringe frequency modulation indicative of blood flow's presence. Such a system can utilize either analog or digital signal filtering. One filter applied to the backscattered signal can be a narrow band filter, which can be centered at the base fringe frequency set by the phase shifter. The width of the filter frequencies can be chosen such that the filtered backscattered signal is indicative of the presences of fluid flow. For example, the filtered backscattered signal from tissue with little to no fluid flow has substantially higher intensity than the filtered backscattered signal from tissue with fluid flow present. Accordingly, in some situations, the bandwidth of the narrow band filter can act to reduce the intensity of a backscattered OCT signal by eliminating higher and lower frequency components that shift out of the bandwidth due to the presence of fluid flow.

Another filter can also be applied to the raw backscattered OCT signal, where this filter is a wide band filter, relative to the narrow band filter, which can also be centered at the base fringe frequency set by the phase shifter. The wide band filter can be configured with a bandwidth such that a portion of backscattered OCT signal, when scattering is from tissue having fluid flow present, is still in the filtered signal (e.g., at least one fringe-modulated feature). Accordingly, comparing the intensity data from the wideband filter output and the narrow band filter output can be indicative of the presence of fluid flow. In particular, the use of the wide band filter can act to allow adjustments due to tissue movement or other changes not due to detecting fluid flow, thus decreasing the possibility of false positives.

It is understood that digital or analog filters for amplitude modulation examination can have a number of other configurations that differ from what has just been described. For example, a filter need not be symmetric about a centered frequency, and/or the filter need not be centered at the frequency of the initial fringe-frequency. Indeed, the filter(s) can be chosen in any suitable manner to accentuate identification of amplitude modulation to aid fluid flow determination.

One exemplary embodiment of the use of a multifiltered backscattered OCT signal system for fluid flow detection is

shown in FIG. 12. A backscattered signal can be received from reflected light from the tissue of a subject interrogated using an OCT system. The OCT system can utilize a base phase fringe signal having a frequency of around 600 kHz centered. The wide band output of an optical receiver 1210 that accepts the reflected light can be routed through at least two filter channels. The narrowband channel contains a narrow band pass filter 1222, centered on the original fringe rate of the interferometer. The filtered signal is then integrated 1232 and passed through a data acquisition board (DAQ) 1242 to be further processed by a computer 1270 (e.g., for obtaining glucose level estimations). The narrow filter signal can be used to compute structural changes induced by glucose variations. As a side effect, this filter 1222 will reduce the contribution of scattering by blood and blood related components as Doppler induced shift from the blood flow will move the signal outside of the filter band pass.

The wideband channel of FIG. 12 passes the backscattered OCT signal through a wide band pass filter 1221 which is also centered at the original fringe rate of the interferometer. The wide band filtered signal is also integrated 1231 and then digitized by the DAQ 1241. The wideband channel contains higher or lower additional frequency components of the signal. Due to the Doppler shift and spectral broadening generated by the passing blood cells, the wide band signal will increase relative to the narrow band one whenever flowing blood is encountered. Differences between these narrow band and wide band signals, as determined by a comparator 1250 for example, can then be used as indicators of the depth and magnitude of blood flow. An indication of the blood perfusion in the scanned tissue 1260 can then be provided. While the directional information is lost with this technique, the simplicity of implementation, and the natural conversion speed of analog processing make such an embodiment potentially useful.

As mentioned earlier, the use of amplitude modulation in an OCT signal to detect fluid flow can also utilize analysis of the speckle in the backscattered OCT signal. For instance, the multi filter approach described above can also be applied to the speckle analysis. Since a narrow band filter can be configured to reduce the intensity of the observed speckle due to fringe frequency shifting, the variance of the speckle can also be lower. Thus, the speckle variance of the resulting OCT signals from the two filters can be compared to identify a difference in the speckle variance in space, in time, or both.

OCT identification of blood flow in a tissue can be used in a number of ways. As alluded to earlier, blood flow identification can provide a measure of blood perfusion in tissue, which can identify whether a tissue section exhibits characteristics for analyte level estimation (e.g., identifying that the tissue has an ACR at a particular set of locations such as at particular depths or lateral extents). The blood flow identification can also serve as an indicator of heart rate by choosing a blood flow location and monitoring the signal's sinusoidal behavior, correlating the oscillatory behavior with heart rate. In some embodiments, during an analyte monitoring session, these blood flow identification techniques can also provide a timely indication of significant sensor disturbance (i.e., the loss of a signal indicating the presence of blood flow), which otherwise may render the analyte measurement inaccurate.

#### Example Embodiments

The following example is provided to illustrate some embodiments of the invention. The example is not intended to limit the scope of any particular embodiment(s) utilized.

An OCT system was implemented to carry out the method diagrammed in FIG. 12 using commercially available equipment.

A test is carried out at a specific tissue site. An initial OCT scan of the tissue at varying depths is performed. The data over each depth is averaged to give an averaged intensity data at that point. Subsequently, OCT scans of the same tissue site are performed at varying moments in time, each of which is designated by a scan number in FIG. 13. The averaged value of the OCT scan versus depth for each time was subtracted from the initial OCT averaged scan value versus depth. The differences are plotted in the pixel plot shown in the top portion of FIG. 13, where a measure of the difference is provided by the shading of the pixel. The lower portion of FIG. 13 provides a corresponding time graph of measured glucose values in the blood stream of the subject whose tissue was scanned. Accordingly, the comparison of the two plots shows that the OCT method can identify regions of tissue that are susceptible to change due to glucose and the presence of blood flow.

FIG. 14 aggregates the pixel data of FIG. 13 by summing all the difference data at a given depth location over all times. That summation is plotted as a function of depth to provide an indicator of the glucose signal, which is the thick line shown in FIG. 14. The data is normalized on the maximum summation in the data set. The thin line corresponds to a perfusion signal measure obtained using the signal processing exemplified in FIG. 12. The backscattered data from a scanned tissue sample at a given depth was sent through a wide band filter and a narrow band filter. An integrator aggregated the data. The differences between the filter outputs at varying depths was plotted as shown by the thin line of FIG. 14. Again, the data is normalized on the maximum difference value. The correspondence between the lines as a function of time shows that a measure of tissue perfusion can act to identify glucose distortion regions, and can provide a glucose independent site metric.

FIG. 15 is a trace showing the pulsatile blood flow indicative of heart rate that is derived from the OCT data. In particular, the perfusion signal of a well-perfused tissue site is aggregated and followed as a function of time. Since the blood flow rate changes through a cycle of a heart beat, the perfusion signal provides a corresponding change, which is shown in FIG. 14. Accordingly, the heart rate of a subject can be followed.

It is understood that a number of variations on the methods described herein are possible, including variations within the knowledge of one skilled in the art. Indeed, numerous other steps can be added, or particular step omitted (e.g., omitting the aligning step; or simply performing the site selection step on existing OCT data). Such variations are within the scope of the present invention.

#### Additional Embodiments

While the present invention has been described in terms of specific methods, structures, and devices it is understood that variations and modifications will occur to those skilled in the art upon consideration of the present invention. For example, the methods and compositions discussed herein can be utilized beyond the preparation of metallic surfaces for implants in some embodiments. As well, the features illustrated or described in connection with one embodiment can be combined with the features of other embodiments. Such modifications and variations are intended to be included within the scope of the present invention. Those skilled in the art will appreciate, or be able to ascertain using no more than routine experimentation, further features and advantages of the invention based on the above-described

embodiments. Accordingly, the invention is not to be limited by what has been explicitly shown and described.

All publications and references are herein expressly incorporated by reference in their entirety. The terms “a” and “an” can be used interchangeably, and are equivalent to the phrase “one or more” as utilized in the present application. The terms “comprising,” “having,” “including,” and “containing” are to be construed as open-ended terms (i.e., meaning “including, but not limited to,”) unless otherwise noted. Recitation of ranges of values herein are merely intended to serve as a shorthand method of referring individually to each separate value falling within the range, unless otherwise indicated herein, and each separate value is incorporated into the specification as if it were individually recited herein. All methods described herein can be performed in any suitable order unless otherwise indicated herein or otherwise clearly contradicted by context. The use of any and all examples, or exemplary language (e.g., “such as”) provided herein, is intended merely to better illuminate the invention and does not pose a limitation on the scope of the invention unless otherwise claimed. No language in the specification should be construed as indicating any non-claimed element as essential to the practice of the invention.

What is claimed is:

1. A method comprising:
  - receiving, from an optical coherence tomography apparatus, an intensity data measurement for a region of tissue;
  - passing the intensity data measurement through a first frequency band filter having a first bandwidth to produce a first output;
  - passing the intensity data measurement through a second frequency band filter having a second bandwidth that is wider than the first bandwidth to produce a second output;
  - determining, by a processor, an indication of fluid flow in the region of tissue by comparing the first output and the second output; and
  - in response to determining a positive indication of fluid flow in the region of tissue, determining, by the processor, a blood analyte measurement based on at least some of the intensity data measurement.
2. The method of claim 1, wherein both the first bandwidth and the second bandwidth are centered at a same base frequency.
3. The method of Claim 1, wherein the blood analyte measurement comprises a measurement of at least one of: blood glucose, blood hemoglobin, or blood oxygen.
4. The method of claim 1, wherein determining the indication of fluid flow in the region of tissue comprises:
  - analyzing fringe frequency data in the first output and the second output.
5. The method of claim 4, wherein analyzing fringe frequency data comprises:
  - analyzing frequency modulation in the fringe frequency data in the first output and the second output.
6. The method of claim 4, wherein analyzing fringe frequency data comprises:

analyzing amplitude modulation in the fringe frequency data in the first output and the second output.

7. The method of claim 6, wherein analyzing fringe frequency data further comprises:
  - comparing speckle intensity in the first output and the second output.
8. The method of claim 1, wherein determining the indication of fluid flow in the region of tissue comprises:
  - comparing speckle intensity in the first output and the second output.
9. The method of claim 2, wherein the intensity data measurement comprises a backscattered wide band signal.
10. A method comprising:
  - receiving, from an optical coherence tomography apparatus, an intensity data measurement for a region of tissue;
  - passing the intensity data measurement through a first filter to produce a first output;
  - passing the intensity data measurement through a second filter to produce a second output;
  - analyzing, by a processor, fringe frequency data of the first output and the second output to determine an indication of fluid flow in the region of tissue; and
  - in response to determining a positive indication of fluid flow in the region of tissue, analyzing, by the processor, the intensity data measurement to determine a blood analyte measurement.
11. The method of claim 10 further comprising:
  - analyzing a plurality of intensity data measurements for a same region of tissue to determine a blood analyte measurement.
12. The method of claim 11, wherein the plurality of intensity data measurements are obtained at a plurality of depths in the same region of tissue.
13. The method of claim 11, wherein the plurality of intensity data measurements are obtained at a plurality of times for the same region of tissue.
14. The method of claim 10, wherein the blood analyte measurement comprises a measurement of at least one of: blood glucose, blood hemoglobin, or blood oxygen.
15. The method of claim 10, wherein analyzing fringe frequency data comprises:
  - analyzing frequency modulation in the fringe frequency data in the first output and the second output.
16. The method of claim 10, wherein analyzing fringe frequency data comprises:
  - analyzing amplitude modulation in the fringe frequency data in the first output and the second output.
17. The method of claim 16, wherein analyzing fringe frequency data further comprises:
  - comparing speckle intensity in the first output and the second output.
18. The method of claim 17, wherein determining the indication of fluid flow in the region of tissue comprises:
  - comparing speckle intensity in the first output and the second output.

\* \* \* \* \*

专利名称(译)	用于分析物水平估计的光学相干断层扫描中的流量计		
公开(公告)号	<a href="#">US10368787</a>	公开(公告)日	2019-08-06
申请号	US14/711610	申请日	2015-05-13
[标]申请(专利权)人(译)	GLT收购		
申请(专利权)人(译)	GLT收购CORP.		
当前申请(专利权)人(译)	Masimo公司		
[标]发明人	REICHGOTT SAMUEL SHAKESPEARE WALTER J KECHTER GEORGE WALLACE PHILLIP WILLIAM SCHURMAN MATTHEW J		
发明人	REICHGOTT, SAMUEL SHAKESPEARE, WALTER J. KECHTER, GEORGE WALLACE, PHILLIP WILLIAM SCHURMAN, MATTHEW J.		
IPC分类号	A61B5/00 A61B5/145 A61B5/1455 A61B5/026		
CPC分类号	A61B5/14551 A61B5/0066 A61B5/0073 A61B5/026 A61B5/0261 A61B5/1455 A61B5/14546 A61B5/4875 A61B5/725 A61B5/7221 A61B5/7278 A61B5/14532 A61B2576/00		
优先权	61/068058 2008-03-04 US 61/033584 2008-03-04 US		
其他公开文献	US20160058347A1		
外部链接	<a href="#">Espacenet</a>		

摘要(译)

公开了基于光学相干断层扫描（此处为“OCT”）的分析物监测系统。在一个方面，公开了可以识别体内流体流动（例如，血流）的技术，其可以充当用于测量组织中血液灌注程度的度量。例如，如果要使用OCT来估计组织中分析物（例如葡萄糖）的水平，则血流量的测量可能潜在地指示存在分析物相关区域，这将适合于分析物水平。与OCT估计。另一方面涉及用于扫描多个区域的系统和方法。光束以两种不同的方式在组织表面上移动。第一种可以是粗略扫描，移动光束以在皮肤上提供不同的扫描位置。第二种可以是精细扫描，其中应用光束以进行更详细的分析。

