



US009168015B2

(12) **United States Patent**
Stearns et al.

(10) **Patent No.:** **US 9,168,015 B2**
(45) **Date of Patent:** **Oct. 27, 2015**

(54) **METHOD AND APPARATUS FOR GATE SPECIFIC MR-BASED ATTENUATION CORRECTION OF TIME-GATED PET STUDIES**

(71) Applicants: **Charles William Stearns**, Milwaukee, WI (US); **Scott David Wollenweber**, Waukesha, WI (US); **David M. Goldhaber**, Shorewood, WI (US)

(72) Inventors: **Charles William Stearns**, Milwaukee, WI (US); **Scott David Wollenweber**, Waukesha, WI (US); **David M. Goldhaber**, Shorewood, WI (US)

(73) Assignee: **GENERAL ELECTRIC CORPORATION**, Schenectady, NY (US)

(*) Notice: Subject to any disclaimer, the term of this patent is extended or adjusted under 35 U.S.C. 154(b) by 119 days.

(21) Appl. No.: **14/138,311**

(22) Filed: **Dec. 23, 2013**

(65) **Prior Publication Data**

US 2015/0178906 A1 Jun. 25, 2015

(51) **Int. Cl.**
A61B 6/00 (2006.01)
G06T 5/50 (2006.01)

(Continued)

(52) **U.S. Cl.**
CPC **A61B 6/5264** (2013.01); **A61B 5/0033** (2013.01); **A61B 5/0522** (2013.01); **A61B 5/085** (2013.01); **A61B 6/03** (2013.01); **G06T 5/50** (2013.01); **G06T 7/0026** (2013.01); **G01T 1/1611** (2013.01); **G01T 1/2985** (2013.01); **G06T 2207/10088** (2013.01); **G06T 2207/10104** (2013.01); **G06T 2207/20048** (2013.01); **G06T 2207/20172** (2013.01); **G06T 2207/30004** (2013.1)

(58) **Field of Classification Search**

CPC G06T 5/50; G06T 7/0026; G06T 2207/20048; G06T 2207/10088; G06T 2207/10104; G06T 2207/20172; G06T 2207/30004; G06T 2207/30061; G06T 2211/40; A61B 5/0033; A61B 5/0035; A61B 5/0073; A61B 5/0522; A61B 5/055; A61B 5/085; A61B 6/025; A61B 6/03; A61B 6/06; A61B 2019/5236; A61B 6/5258; A61B 6/5264; A61B 8/13; A61B 2019/524; G01T 1/1611; G01T 1/2985

See application file for complete search history.

(56) **References Cited**

U.S. PATENT DOCUMENTS

5,251,128 A * 10/1993 Crawford 600/425
6,377,833 B1 * 4/2002 Albert 600/410

(Continued)

OTHER PUBLICATIONS

Boucher, et al., Respiratory Gating for 3-Dimensional PET of the Thorax: Feasibility and Initial Results, The Journal of Nuclear Medicine, vol. 45, No. 2, Feb. 2004.

(Continued)


Primary Examiner — Manav Seth

(74) Attorney, Agent, or Firm — Grogan, Tuccillo & Vanderleeden, LLP

(57) **ABSTRACT**

A method for attenuation correction of a plurality of time-gated PET images of a moving structure includes establishing an ACF matrix, based on a baseline MR image of the structure; developing a plurality of image transforms, each of the image transforms registering the baseline MR image to a respective cine MR image of the moving structure; applying each of the image transforms to the ACF matrix to generate a corresponding cine ACF matrix; time-correlating each of the cine MR images, and its corresponding cine ACF matrix, to one of the plurality of time-gated PET data sets; and producing from each of the time-gated PET data sets an attenuation-corrected time-gated PET image.

20 Claims, 3 Drawing Sheets



(51) **Int. Cl.**

G06T 7/00 (2006.01)
A61B 5/00 (2006.01)
A61B 5/05 (2006.01)
A61B 5/085 (2006.01)
A61B 6/03 (2006.01)
G01T 1/161 (2006.01)
G01T 1/29 (2006.01)

(56)

References Cited

U.S. PATENT DOCUMENTS

6,990,368 B2 * 1/2006 Simon et al. 600/425
 7,117,026 B2 * 10/2006 Shao et al. 600/411
 7,180,074 B1 * 2/2007 Crosetto 250/370.09
 7,787,675 B2 * 8/2010 Pan et al. 382/128
 8,098,916 B2 * 1/2012 Thielemans et al. 382/131
 8,108,024 B2 * 1/2012 Carlsen et al. 600/407
 8,208,709 B2 * 6/2012 Ding et al. 382/131
 8,218,848 B2 * 7/2012 Lenglet et al. 382/131
 8,535,061 B2 * 9/2013 Boutchko et al. 434/267
 8,600,132 B2 * 12/2013 Razifar et al. 382/128
 8,768,432 B2 * 7/2014 Ladebeck 600/411
 8,866,087 B2 * 10/2014 Pistorius et al. 250/363.04
 8,942,445 B2 * 1/2015 Foo et al. 382/128
 2003/0233039 A1 * 12/2003 Shao et al. 600/407

2006/0237652 A1 * 10/2006 Kimchy et al. 250/363.02
 2009/0110256 A1 * 4/2009 Thielemans et al. 382/131
 2010/0204563 A1 * 8/2010 Stodilka et al. 600/411
 2011/0007958 A1 * 1/2011 Salomon et al. 382/131
 2011/0123083 A1 * 5/2011 Ojha et al. 382/131
 2013/0320973 A1 12/2013 Fenchel et al.

OTHER PUBLICATIONS

Woo, et al., Development of a Motion Detecting System for Respiratory-Gated PET Using Laser Optical Displacement Sensor, 2004 IEEE.

International Search Report and Written Opinion for PCT/US2014/049885, mail date Feb. 6, 2015, 13 pages.

Buerger et al., Investigation of Mr-Based Attenuation Correction and Motion Compensation for Hybrid PET/MR, IEEE Transactions on Nuclear Science, vol. 59, No. 5, Oct. 2012, 10 pages.

Maes et al., Medical Image Registration Using Mutual Information, Proceedings of the IEEE, vol. 91, No. 10, Oct. 2003, 24 pages.

Mattes et al., PET-CT Image Registration in the Chest Using Free-form Deformations, IEEE Transactions on Medical Imaging, vol. 22, No. 1, Jan. 2003, 9 pages.

Wu et al., Low-dose interpolated average CT for attenuation correction in cardiac PET/CT, Nuclear Instruments and Methods in Physics Research, vol. 619, No. 1-3, 4 pages.

* cited by examiner




FIG. 1




FIG. 2




FIG. 3

$$S(\mu) = - \sum_l \sum_k p(l, k | \mu) \log \frac{p(l, k | \mu)}{p_T(l | \mu) p_R(k)}$$

FIG. 4




FIG. 5




FIG. 6

1

METHOD AND APPARATUS FOR GATE SPECIFIC MR-BASED ATTENUATION CORRECTION OF TIME-GATED PET STUDIES

BACKGROUND

1. Technical Field

Embodiments of the invention relate generally to enhancing quality of images obtained from PET-MR scanners. Particular embodiments relate to attenuation correction of PET images.

2. Discussion of Art

Positron emission tomography ("PET") machines use one or more rings of scintillators or other detectors to generate electrical signals from gamma rays (photon pairs) that produced from the recombination of electrons, within a target material, and positrons, emitted from decay of a radionuclide packaged in a tracer compound. Typically, recombination events occur within about 1 mm from the radionuclide decay event, and the recombination photons are emitted in generally opposite directions to arrive at different detectors. Paired photon arrivals that occur within a detection window (usually less than a few nanoseconds apart) are counted as indicating a recombination event, and, on this basis, computed tomography algorithms are applied to the scintillator position and detection data in order to locate the various recombination events, thereby producing three-dimensional images of the tracer disposition within the target material.

Typically, the target material is body tissue, the tracer compound is a liquid analogue to a biologic fluid, and the radionuclide is disposed primarily in body tissues that make use of the biologic fluid. For example, a common form of PET makes use of fluorodeoxyglucose (^{18}F), which is analogous to glucose with the ^{18}F radionuclide substituted for one of the hydroxyl groups ordinarily composing glucose. Brain matter, kidneys, and growing cells (e.g., metastasizing cancer cells) preferentially absorb both glucose and fluorodeoxyglucose. Therefore, PET is quite useful in oncologic studies, for localizing particular organs, and for studying metabolic processes.

One challenge in obtaining desired PET image quality is that gamma rays, in the energy spectrum produced by positron-electron interactions, are easily attenuated by typical body tissues and are differently attenuated by different body tissues. Varying attenuation will change the likelihood of detecting the recombination events from the patient, thereby confounding the process of making an image from the PET data. Accordingly, it is highly desirable to provide means for attenuation correction ("AC").

For example, PET often is combined with computed tomography ("CT"), which uses a moving X-ray source and detectors to obtain images of internal structures. X-rays are photons and are attenuated much like the higher-energy photons produced from positron-electron recombination events. However, unlike PET photons, the source strength of CT photons is known. Thus, CT image data is a direct measurement of photon attenuation between source and detectors. Therefore, CT image data provides a useful basis for AC of concurrent PET imaging.

Combined PET-CT scans have proven particularly useful in respiratory-gated studies, for example studies undertaken during normal breathing in order to diagnose or evaluate lung cancer. Respiratory-gated images can reduce motion blurring that would otherwise result from acquisition of the PET image throughout a respiratory cycle. However, accurate reconstruction of respiratory-gated PET studies requires

2

gate-specific attenuation correction information. The simultaneous acquisition of CT and PET data supports gate-specific AC.

For example, consider two elements of a PET detector where the line of response between the detectors passes through the lower thorax of the patient. At peak inspiration that line may pass through the lower lobes of the lung, which will attenuate the 511 keV photons by a relatively small amount. As the patient exhales the liver moves up into the thorax, so that at the end of expiration the same line of response may now pass through soft tissue, which will attenuate the 511 keV photons much more than the lung. If the PET data is gated so that there are different images being acquired at peak-inspiration and end-exhalation, each will benefit from an attenuation correction that corresponds to the actual distribution of attenuating tissue during that phase of the respiratory cycle.

Increasingly, and for a variety of reasons including lifetime radiation dose reduction goals, PET/CT scans are being replaced by combining PET scans with magnetic resonance imaging ("MR"). This new combination ("PET-MR") presents novel technical issues. For example, whereas CT forms an image based on detection of X-rays emitted from a source through a target, MR forms an image based on detection of rotating "relaxation" magnetic fields that are produced within a target by nuclei that have odd atomic numbers, i.e., total number of neutrons and protons not divisible by two, in response to fluctuation of an imposed magnetic field. Thus, MR measures a phenomenon fundamentally different from the photons detected by CT and PET.

One advantage of MR is that magnetic fields do not attenuate in body tissues, so that nucleus location can be determined (using Fourier analysis) based solely on frequency shifting between the imposed magnetic field and the response field. Another advantage is that by careful selection of pulse sequence, distinct tissues or materials can be highlighted. Accordingly, MR frequently is used for differentiating tissue types within a patient, and also is used for identifying fine detail structures. Typically, different pulse sequences are used for tissue differentiation. For example, a T1 pulse sequence can be used to obtain images with water appearing darker and fat brighter. On the other hand, a T2 pulse sequence can be used to obtain an image with fat darker, and water lighter.

Thus, a single apparatus that combines PET and MR (a "PET-MR scanner") can provide fine detail, tissue differentiation, and metabolic data. However, because MR signals do not attenuate in the same way as PET or CT signals attenuate, and because MR signal return is highly dependent on the type of pulse sequence used (with each pulse sequence emphasizing a different material), whereas the PET signal is attenuated by every layer of material intervening between a recombination event and a pair of detectors, single-scan MR image data does not necessarily provide a reliable basis for AC of concurrent PET imaging.

MR imaging has the capability to acquire a volume of images over the respiratory cycle, in a manner known as retrospectively gated cine-MRI. However, the types of MR images that can be acquired in a cine fashion have not been considered suitable for producing PET attenuation correction values. The MR protocols conventionally deemed effective for producing PET attenuation correction values, such as 2-point Dixon water-fat scans, typically require more than 10 seconds to acquire, thus, are not usable for capturing multiple images during a respiratory cycle. Instead, in order to minimize motion blur and other breathing artifacts, these MR image data typically have been acquired during a long breath-

hold. However, a breath-held MR image does not show the intermediate lung positions that are needed for accurate AC of a PET scan.

BRIEF DESCRIPTION

Embodiments of the invention provide a method for attenuation correction of a plurality of time-gated PET images of a moving structure. The method includes establishing an ACF matrix, based on a baseline MR image of the structure; developing a plurality of image transforms, each of the image transforms registering the baseline MR image to a respective cine MR image of the moving structure; applying each of the image transforms to the ACF matrix to generate a corresponding cine ACF matrix; time-correlating each of the cine MR images, and its corresponding cine ACF matrix, to one of the plurality of time-gated PET data sets; and producing from each of the time-gated PET data sets an attenuation-corrected time-gated PET image.

Aspects of the invention provide an apparatus for generating a plurality of attenuation corrected time-gated PET images. The apparatus includes a PET-MR scanner; a controller connected in communication with the PET-MR scanner, and an image processor connected in communication with the controller. The controller is configured to acquire a baseline MR image of a structure, acquire a sequence of cine MR images of the structure, and acquire a plurality of time-gated PET data sets of the structure. The image processor is configured to generate a baseline ACF matrix based on the baseline MR image, register the baseline MR image with at least some of the cine MR images, generate cine ACF matrices based on registration of the baseline MR image with the cine MR images, and attenuation correct the time-gated PET data sets based on the cine ACF matrices.

Other aspects of the invention provide a method for attenuation correction of PET imaging. The method includes obtaining, in a first MR acquisition, a baseline MR image suitable for deriving PET attenuation correction factors; and obtaining, in a second MR acquisition, a cine sequence of MR images during a respiratory cycle, concurrent with the respiratory-gated PET study. The method further comprises non-rigidly registering the baseline MR image to each of the cine MR images in order to produce a group of registered MR images that correspond to PET data sets from the respiratory-gated PET study and that are usable for deriving attenuation correction factors. The registered MR images then are used to form attenuation correction factors for their respective time-gated PET data sets.

DRAWINGS

The present invention will be better understood from reading the following description of non-limiting embodiments, with reference to the attached drawings, wherein below:

FIG. 1 shows a conventional method for attenuation correction of PET images based on concurrent CT images.

FIG. 2 shows an inventive method for attenuation correction of PET images based on concurrent and baseline MR images, according to a first embodiment of the invention.

FIG. 3 shows an exemplary process for registering a baseline MR image to a cine MR image, according to an embodiment of the invention.

FIG. 4 shows an equation for mutual information of a baseline MR image and a cine MR image.

FIG. 5 shows another method for attenuation correction of PET images based on concurrent and baseline MR images, according to a second embodiment of the invention.

FIG. 6 shows an apparatus for implementing inventive methods for attenuation correction of PET images based on concurrent and baseline MR images, according to embodiments of the invention.

DETAILED DESCRIPTION

Reference will be made below in detail to exemplary embodiments of the invention, examples of which are illustrated in the accompanying drawings. Wherever possible, the same reference characters used throughout the drawings refer to the same or like parts, without duplicative description. Exemplary embodiments of the present invention are described with respect to combined PET-MR scanners, although embodiments may be adapted for use with other imaging systems.

Aspects of the invention relate to improving the simultaneous collection of PET and MR images by providing attenuation correction (AC) of the PET images based on the MR image data.

In PET/CT studies, a conventional method **100** of respiratory-gated attenuation (as shown in FIG. 1) includes obtaining **110** a sequence of cine CT images **112** over the same anatomic region and during the same time period as acquiring **120** a sequence of time-gated PET data sets **122**. The sequence of cine CT images **112** can be acquired at rates on the order of one image per second, allowing acquisition of six to ten images during a typical respiratory cycle, so that the set of CT images then can be correlated to the gated set of PET raw data files **120**. The sequence of CT images **110** are used to derive **130** a sequence of attenuation correction factor matrices ("ACF matrices") **132**, and then each ACF matrix is time-correlated **140** to one of the time-gated PET data sets **122** in order to reconstruct **150** a set of attenuation corrected ("AC") PET images **152**.

FIG. 2 shows a method **200**, according to a first embodiment of the invention, for MR-based attenuation correction **210** of a sequence of respiratory-gated PET data sets **222**, in order to produce attenuation-corrected PET images **212**. In addition to acquiring **220** the PET data sets **222**, the method **200** also includes acquiring **230** a tissue-distinguishing or "baseline" MR image **232** that is suitable for later generating **240** a three-dimensional matrix **242** of photon attenuation correction factors ("baseline ACF matrix"). In an embodiment, the baseline MR image **232** is obtained **230** as a substantially static image using a patient breath-hold protocol and a scan protocol suitable for tissue-differentiation study—e.g., a 2-point Dixon scan, which clearly highlights water and fat, and requires in excess of ten seconds to complete. Although, in certain embodiments, the baseline MR image **232** may be obtained as an averaged image using a tissue-differentiation study during normal respiration, or can be compiled during respiration or during a breath-hold as a composite image from several distinct single-tissue scans, it can be expected that an averaged image or a composite image will not be optimal for subsequently generating the static ACF matrix **242**.

Based on the tissues identified in the baseline MR image **232**, standard photon attenuation calculations are undertaken for the typical positron-recombination energies in order to generate **240** the baseline ACF matrix **242**.

In addition to generating **240** the baseline ACF matrix **242**, the method **200** also includes acquiring **250** a sequence of cine MR images **252**, and registering **260** each of the cine MR images **252** back to the static MR image **232**. In the process of registering **260**, a plurality of image transforms **262** are produced, each of the image transforms registering a respective

one of the cine MR images **252** with the static MR image **232**. In an embodiment, the image transforms are non-rigid. For example, Vemuri, et al., Med. Image Anal. 7:1-20 (2003), describe a “level-sets” algorithm for non-rigid image registration. Other mutual-information-based algorithms (e.g., as taught by Woods (1992) or by Viola and Wells (1994)) can also be used. In most embodiments, the cine MR images **252** are obtained using different (faster) scan protocols than was used for obtaining the baseline MR image **232**—e.g., zero-TE or ultra-short-TE scan protocols. Also, in certain embodiments, different scan protocols may be used for different cine MR images **252**. Accordingly, registering **260** may also include pixel intensity scaling or shifting to match tissue types between the different scan protocols.

For example, a locally perturbed (non-rigid) rigid body registration **260**, based on mutual information, can proceed as shown in FIG. 3, in which one of the cine images **252** is represented as a differentiable cubic spline function $f_T(x)$, mapping pixel intensities to a linear pixel array, which can be interpolated at any between-pixel position within the set of sampled values composing the image. Similarly, the baseline image **232** is represented as another differentiable cubic spline function $f_R(x)$. A smoothed joint histogram of

$$\mu = \arg \min S(f_R f_T^{-1} g(\bullet | \mu)),$$

which relates the cine image **252** back to the baseline image **232**, is defined as a cross product of the two spline functions. The smoothed joint histogram is transformed by applying to the baseline image **232** a set of transformation parameters $\mu = \{\gamma, \theta, \phi, t_x, t_y, t_z, \delta_j\}$, which are chosen to optimize mutual information $S(\mu)$ according to a hierarchical multi-resolution optimization scheme. In the set of transformation parameters μ , $\{\gamma, \theta, \phi\}$ are bulk roll-pitch-yaw Euler angles, $[t_x, t_y, t_z]$ is a bulk translation vector, δ_j is a set of deformation coefficients each corresponding to one of a plurality of control points that are defined by a sparse grid that is overlaid on the baseline image **232**. Typically, the sparse grid is a regular grid. However, in other embodiments, the sparse grid can be weighted according to pixel intensity or other image metrics. For any given set of non-rigid transformation parameters, mutual information $S(\mu)$ is calculated (as shown in FIG. 4), and mutual information gradients $\nabla S(\mu)$ are calculated, based on probability distributions p , p_T , and p_R that respectively correspond to: the smoothed joint histogram; a marginal smoothed histogram for the cine image **252**; and a marginal smoothed histogram for the baseline image **232**. The probability distributions are presumed independent of the transformation parameters. Parameters τ and κ , shown in FIG. 4, are pixel intensity histogram bin indexes for the cine image **252** and for the baseline image **232**, respectively. Varying the histogram bin sizes can adjust the quality of image registration. The registration is described as locally-perturbed rigid body because it combines local deformation coefficients δ_j with rigid body transforms $\{\gamma, \theta, \phi, [t_x, t_y, t_z]\}$. Evaluating quality of fit, the registration process **260** either iterates on revised μ , or returns μ (the image transform **262** that registers the baseline image **232** to the particular cine image **252**) for a satisfactory value of $S(\mu)$.

Then, the method **200** includes applying each of the image transforms **262** to the static ACF matrix **242** in order to generate **270** a plurality of cine ACF matrices **272**. Each of the cine ACF matrices **272** is time-correlated **280** to one of the plurality of PET data sets **222** within the respiratory-gated PET sequence **220**, and is applied **210** to the corresponding PET data set **222** in order to produce a reconstructed and attenuation-corrected PET image **212**. In certain embodiments, each of the PET data sets **222** is time-correlated to

more than one of the cine ACF matrices, and one of a simple average, a time-weighted average, a velocity-weighted average, or a time-velocity-weighted average is used in applying **210** the plural cine ACF matrices **272** to the PET data set **222**.

Although the static MR image acquisition **230** is shown as preceding the cine MR sequence acquisition **250**, these steps equally can be accomplished in the other order. In certain embodiments, the cine MR sequence may be duplicated (e.g., before and after the breath-held static image) or the breath-held static image may be duplicated (e.g., before and after the cine sequence). Similarly, registering **260** may be accomplished concurrent with acquiring **250**, or after all images have been acquired.

Although the plurality of registered ACF matrices **272** are described as being obtained by applying image transforms **262** to the baseline ACF matrix **242**, registered ACF matrices **272** also could be calculated from the baseline MR image **232**, as transformed according to the plurality of image transforms **262**. Thus, an alternate method **500** as shown in FIG. 5, according to other embodiments of the invention, includes a first MR acquisition **230** to obtain a static (breath-held) MR image **232** suitable for deriving **240** baseline PET attenuation correction factors **242**; and a second MR acquisition **250** to obtain a sequence of cine MR images **252** acquired during a respiratory cycle, concurrent with a respiratory-gated PET study **220**. The static MR image **232** is non-rigidly registered to each of the cine MR images **252** in order to produce **560** a group of transformed MR images **562** that correspond to PET data sets **222** from the respiratory-gated PET study and that are usable for deriving **570** attenuation correction factors **572**. The registered MR images then are used to form **570** the attenuation correction factors **572**, which are time-correlated **280** for producing **210** attenuation-corrected PET images **212** from their respective time-gated PET data sets **222**.

FIG. 6 shows schematically an apparatus **600**, according to an embodiment of the invention, for accomplishing the method **200** as described above. The apparatus **600** comprises a combined PET-MR scanner **610**, which is connected in communication with a controller **620**, the controller in turn connected in communication with an image processor **630**. The controller and image processor can both be implemented as separate software in a single computer, or separately in computers co-located with the PET-MR scanner **610**; or either or both of the controller **620** and the image processor **630** can be located off-site (remotely) from the PET-MR scanner and/or each other. The controller **620** is configured to accomplish at least acquiring **220** the PET data sets **222**, acquiring **230** the baseline MR image **232**, and acquiring **250** the sequence of cine MR images **252**. The image processor **630** is configured to accomplish at least generating **240** the baseline ACF matrix **242**, registering **260** the baseline MR image **232** with each of the sequence of cine MR images **252**, generating **270** the plurality of cine ACF matrices **272**, and attenuation-correcting **210** the time-gated PET data sets **222** based on the plurality of cine ACF matrices **272**. Either the controller **620** or the image processor **630** may be configured to accomplish time-correlating **280** the cine MR images **252** with the time-gated PET data sets **222**. Alternatively, the controller **620** and the image processor **630** can be integrated into single software/single processor, or can be implemented in a cloud (SaaS) paradigm.

Thus, embodiments of the invention provide gate-specific attenuation correction factors for time-gated PET data, based on attenuation correction data that is not acquired within the duration of a single gate.

For example, some embodiments of the invention provide a method for attenuation correction of a plurality of time-

gated PET images of a moving structure. The method comprises establishing an ACF matrix, based on a baseline MR image of the structure; developing a plurality of image transforms, each of the image transforms registering the baseline MR image to a respective cine MR image of the moving structure; applying each of the image transforms to the ACF matrix to generate a corresponding cine ACF matrix; time-correlating each of the cine MR images, and its corresponding cine ACF matrix, to one of the plurality of time-gated PET data sets; and producing from each of the time-gated PET data sets an attenuation-corrected time-gated PET image. In certain embodiments, the baseline MR image is a composite image. In some other embodiments, the baseline MR image is an averaged image. In some embodiments, the baseline MR image is obtained while the structure is substantially static. In some embodiments, the baseline MR image is obtained using a tissue differentiating scan protocol. In some embodiments, each of the cine MR images uses a same scan protocol as each other of the cine MR images. In certain embodiments, developing the plurality of image transforms comprises pixel intensity scaling or pixel intensity shifting. In certain embodiments, developing the plurality of image transforms comprises iteratively calculating mutual information of a joint pixel intensity histogram constructed from the baseline MR image and from an intermediate transform of the cine MR image. In certain embodiments, developing the plurality of image transforms comprises establishing for each cine MR image an image transform μ that comprises rigid transformation parameters $\{\gamma, \theta, \phi, [tx, ty, tz]\}$ and deformation coefficients $\{\delta_j\}$ for registering the baseline MR image to that cine MR image. In certain embodiments, producing an attenuation-corrected time-gated PET image comprises applying to one of the plurality of time-gated PET data sets at least one cine ACF matrix that is time-correlated to that PET data set. In certain embodiments, producing an attenuation-corrected time-gated PET image comprises applying to one of the plurality of time-gated PET data sets one of a simple average, a time-weighted average, a velocity-weighted average, or a time-velocity-weighted average of plural cine ACF matrices that are time-correlated to that PET data set.

Other aspects of the invention provide an apparatus for generating a plurality of attenuation corrected time-gated PET images. The apparatus comprises a PET-MR scanner; a controller connected in communication with the PET-MR scanner, and an image processor connected in communication with the controller. The controller is configured to acquire a baseline MR image of a structure, acquire a sequence of cine MR images of the structure, and acquire a plurality of time-gated PET data sets of the structure. The image processor is configured to generate a baseline ACF matrix based on the baseline MR image, register the baseline MR image with at least some of the cine MR images, generate cine ACF matrices based on registration of the baseline MR image with the cine MR images, and attenuation correct the time-gated PET data sets based on the cine ACF matrices. In certain aspects, the controller is configured to time-correlate each of the cine ACF matrices to one of the time-gated PET data sets. In certain aspects, the image processor is configured to time-correlate each of the cine ACF matrices to one of the time-gated PET data sets. In certain aspects, the image processor is configured to register the baseline MR image to each of the cine MR images using a locally perturbed rigid registration. In certain aspects, the image processor is configured to register the baseline MR image to each of the cine MR images using a mutual information algorithm. In certain aspects, the controller is configured to obtain the baseline MR image while the structure is substantially static. In certain

aspects, the controller is configured to concurrently obtain the cine MR images and the PET data sets while the structure is moving. In certain aspects, the image processor is configured to attenuation correct at least one of the plurality of time-gated PET data sets using one of a simple average, a time-weighted average, a velocity-weighted average, or a time-velocity-weighted average of plural cine ACF matrices that are time-correlated to that PET data set.

Other aspects of the invention provide a method for attenuation correction of PET imaging. The method comprises obtaining, in a first MR acquisition, a baseline MR image suitable for deriving PET attenuation correction factors; and obtaining, in a second MR acquisition, a cine sequence of MR images during a respiratory cycle, concurrent with the respiratory-gated PET study. The method further comprises non-rigidly registering the baseline MR image to each of the cine MR images in order to produce a group of registered MR images that correspond to PET data sets from the respiratory-gated PET study and that are usable for deriving attenuation correction factors. The registered MR images then are used to form attenuation correction factors for their respective time-gated PET data sets.

It is to be understood that the above description is intended to be illustrative, and not restrictive. For example, the above-described embodiments (and/or aspects thereof) may be used in combination with each other. In addition, many modifications may be made to adapt a particular situation or material to the teachings of the invention without departing from its scope. While the dimensions and types of materials described herein are intended to define the parameters of the invention, they are by no means limiting and are exemplary embodiments. Many other embodiments will be apparent to those of skill in the art upon reviewing the above description. The scope of the invention should, therefore, be determined with reference to the appended claims, along with the full scope of equivalents to which such claims are entitled. In the appended claims, the terms “including” and “in which” are used as the plain-English equivalents of the respective terms “comprising” and “wherein.” Moreover, in the following claims, terms such as “first,” “second,” “third,” “upper,” “lower,” “bottom,” “top,” etc. are used merely as labels, and are not intended to impose numerical or positional requirements on their objects. Further, the limitations of the following claims are not written in means-plus-function format and are not intended to be interpreted based on 35 U.S.C. §112, sixth paragraph, unless and until such claim limitations expressly use the phrase “means for” followed by a statement of function void of further structure.

This written description uses examples to disclose several embodiments of the invention, including the best mode, and also to enable one of ordinary skill in the art to practice embodiments of the invention, including making and using any devices or systems and performing any incorporated methods. The patentable scope of the invention is defined by the claims, and may include other examples that occur to one of ordinary skill in the art. Such other examples are intended to be within the scope of the claims if they have structural elements that do not differ from the literal language of the claims, or if they include equivalent structural elements with insubstantial differences from the literal language of the claims.

As used herein, an element or step recited in the singular and proceeded with the word “a” or “an” should be understood as not excluding plural of the elements or steps, unless such exclusion is explicitly stated. Furthermore, references to “one embodiment” of the present invention are not intended to be interpreted as excluding the existence of additional

embodiments that also incorporate the recited features. Moreover, unless explicitly stated to the contrary, embodiments “comprising,” “including,” or “having” an element or a plurality of elements having a particular property may include additional such elements not having that property.

Since certain changes may be made in the above-described embodiments, without departing from the spirit and scope of the invention herein involved, it is intended that all of the subject matter of the above description or shown in the accompanying drawings shall be interpreted merely as examples illustrating the inventive concept herein and shall not be construed as limiting the invention.

What is claimed is:

1. A method for attenuation correction of a plurality of time-gated PET images of a moving structure, comprising:

establishing an ACF matrix, based on a baseline MR image of the structure;

developing a plurality of image transforms, each of the image transforms registering the baseline MR image to a respective cine MR image of the moving structure;

applying each of the image transforms to the ACF matrix to generate a corresponding cine ACF matrix;

time-correlating each of the cine MR images, and its corresponding cine ACF matrix, to one of the plurality of time-gated PET data sets; and

producing from each of the time-gated PET data sets an attenuation-corrected time-gated PET image.

2. The method as claimed in claim 1, wherein the baseline MR image is a composite image.

3. The method as claimed in claim 1, wherein the baseline MR image is an averaged image.

4. The method as claimed in claim 1, further comprising a step of obtaining the baseline MR image while the structure is substantially static.

5. The method as claimed in claim 1, further comprising a step of obtaining the baseline MR image using a tissue differentiating scan protocol.

6. The method as claimed in claim 1, wherein each of the cine MR images uses a same scan protocol as each other of the cine MR images.

7. The method as claimed in claim 1, wherein developing the plurality of image transforms comprises pixel intensity scaling or pixel intensity shifting.

8. The method as claimed in claim 1, wherein developing the plurality of image transforms comprises, for each of the cine MR images, iteratively calculating mutual information of a joint pixel intensity histogram constructed from the baseline MR image and from an intermediate transform of the cine MR image.

9. The method as claimed in claim 1, wherein developing the plurality of image transforms comprises establishing for each cine MR image an image transform μ that comprises rigid transformation parameters $\{\gamma, \theta, \phi, [t_x, t_y, t_z]\}$ and deformation coefficients $\{\delta_j\}$ for registering the baseline MR image to that cine MR image.

10. The method as claimed in claim 1, wherein producing an attenuation-corrected time-gated PET image comprises applying to one of the plurality of time-gated PET data sets at least one cine ACF matrix that is time-correlated to that PET data set.

11. The method as claimed in claim 10, wherein producing an attenuation-corrected time-gated PET image comprises applying to one of the plurality of time-gated PET data sets one of a simple average, a time-weighted average, a velocity-

weighted average, or a time-velocity-weighted average of plural cine ACF matrices that are time-correlated to that PET data set.

12. An apparatus for generating a plurality of attenuation corrected time-gated PET images, said apparatus comprising:

a PET-MR scanner;

a controller connected in communication with the PET-MR scanner, wherein the controller is configured to acquire a baseline MR image of a structure, acquire a sequence of cine MR images of the structure, and acquire a plurality of time-gated PET data sets of the structure; and

an image processor connected in communication with the controller, wherein the image processor is configured to generate a baseline ACF matrix based on the baseline MR image, register the baseline MR image with at least some of the cine MR images, generate cine ACF matrices based on registration of the baseline MR image with the cine MR images, and attenuation correct the time-gated PET data sets based on the cine ACF matrices.

13. The apparatus as claimed in claim 12, wherein the controller is configured to time-correlate each of the cine ACF matrices to one of the time-gated PET data sets.

14. The apparatus as claimed in claim 12, wherein the image processor is configured to time-correlate each of the cine ACF matrices to one of the time-gated PET data sets.

15. The apparatus as claimed in claim 12, wherein the image processor is configured to register the baseline MR image to each of the cine MR images using a locally perturbed rigid registration.

16. The apparatus as claimed in claim 12, wherein the image processor is configured to register the baseline MR image to each of the cine MR images using a mutual information algorithm.

17. The apparatus as claimed in claim 12, wherein the controller is configured to obtain the baseline MR image while the structure is substantially static.

18. The apparatus as claimed in claim 12, wherein the controller is configured to concurrently obtain the cine MR images and the PET data sets while the structure is moving.

19. The apparatus as claimed in claim 12, wherein the image processor is configured to attenuation correct at least one of the plurality of time-gated PET data sets using one of a simple average, a time-weighted average, a velocity-weighted average, or a time-velocity-weighted average of plural cine ACF matrices that are time-correlated to that PET data set.

20. A method for attenuation correction of PET imaging in a respiratory-gated PET study, comprising:

obtaining, in a first MR acquisition, a baseline MR image suitable for deriving PET attenuation correction factors;

obtaining, in a second MR acquisition, a cine sequence of MR images during a respiratory cycle, concurrent with the respiratory-gated PET study; and

non-rigidly registering the baseline MR image to each of the cine MR images in order to produce a group of registered MR images that correspond to PET data sets from the respiratory-gated PET study and that are usable for deriving attenuation correction factors,

wherein the registered MR images then are used to form attenuation correction factors for their respective time-gated PET data sets.

* * * * *

专利名称(译)	用于门控特定的基于MR的时间门控PET研究的衰减校正的方法和装置		
公开(公告)号	US9168015	公开(公告)日	2015-10-27
申请号	US14/138311	申请日	2013-12-23
[标]申请(专利权)人(译)	STEARNS CHARLES WILLIAM WOLLENWEBER SCOTT DAVID GOLDHABER DAVID中号		
申请(专利权)人(译)	STEARNS , CHARLES WILLIAM WOLLENWEBER , DAVID SCOTT GOLDHABER , DAVID M.		
当前申请(专利权)人(译)	GENERAL ELECTRIC CORPORATION		
[标]发明人	STEARNS CHARLES WILLIAM WOLLENWEBER SCOTT DAVID GOLDHABER DAVID M		
发明人	STEARNS, CHARLES WILLIAM WOLLENWEBER, SCOTT DAVID GOLDHABER, DAVID M.		
IPC分类号	A61B6/00 A61B5/085 A61B5/05 A61B6/03 A61B5/00 G06T5/50 G06T7/00 G01T1/161 G01T1/29		
CPC分类号	A61B6/5264 A61B5/0033 A61B5/0522 A61B5/085 A61B6/03 G06T5/50 G06T7/0026 G01T1/1611 G01T1/2985 G06T2207/10088 G06T2207/10104 G06T2207/20048 G06T2207/20172 G06T2207/30004 A61B6/5282 A61B6/037 A61B6/5247 A61B6/541 G01R33/481 G06T7/32 G06T2207/10084		
其他公开文献	US20150178906A1		
外部链接	Espacenet USPTO		

摘要(译)

一种用于对移动结构的多个时间选通PET图像进行衰减校正的方法包括基于该结构的基线MR图像建立ACF矩阵;开发多个图像变换,每个图像变换将基线MR图像记录到移动结构的相应电影MR图像;将每个图像变换应用于ACF矩阵以生成相应的电影ACF矩阵;将每个电影MR图像及其相应的电影ACF矩阵时间相关到多个时间选通的PET数据集之一;并且从每个时间选通的PET数据集产生衰减校正的时间选通PET图像。

