



US009848795B2

(12) **United States Patent**
Marecki et al.

(10) **Patent No.:** **US 9,848,795 B2**
(45) **Date of Patent:** **Dec. 26, 2017**

(54) **ELECTRODE ASSEMBLY**

(71) Applicant: **BOSTON SCIENTIFIC SCIMED, INC.**, Maple Grove, MN (US)

(72) Inventors: **Andrew T. Marecki**, Shrewsbury, MA (US); **Paul Hultz**, Brookline, MA (US); **Michael C. Kozlowski**, Wakefield, MA (US); **Brian Stewart**, North Reading, MA (US)

(73) Assignee: **Boston Scientific Scimed Inc.**, Maple Grove, MN (US)

(*) Notice: Subject to any disclaimer, the term of this patent is extended or adjusted under 35 U.S.C. 154(b) by 244 days.

(21) Appl. No.: **14/723,408**

(22) Filed: **May 27, 2015**

(65) **Prior Publication Data**

US 2015/0351652 A1 Dec. 10, 2015

Related U.S. Application Data

(60) Provisional application No. 62/007,716, filed on Jun. 4, 2014.

(51) **Int. Cl.**
A61B 5/042 (2006.01)
A61B 18/14 (2006.01)
(Continued)

(52) **U.S. Cl.**
CPC **A61B 5/0422** (2013.01); **A61B 5/6856** (2013.01); **A61B 5/6858** (2013.01); **A61B 5/6859** (2013.01); **A61B 18/1492** (2013.01); **A61B 5/6855** (2013.01); **A61B 5/6857** (2013.01); **A61B 2017/00044** (2013.01); **A61B 2018/0016** (2013.01); **A61B 2018/00267** (2013.01); **A61B 2562/0209** (2013.01); **A61B 2562/125** (2013.01); **A61B 2562/164** (2013.01);
(Continued)

(58) **Field of Classification Search**

CPC ... A61B 5/0422; A61B 5/6859; A61B 5/6869; A61B 2018/00267
See application file for complete search history.

(56) **References Cited**

U.S. PATENT DOCUMENTS

4,649,924 A 3/1987 Taccardi
4,674,518 A 6/1987 Salo
(Continued)

FOREIGN PATENT DOCUMENTS

CA 2719329 A1 10/2009
CN 203017083 U 6/2013
(Continued)

OTHER PUBLICATIONS

Lorensen et al., "Marching Cubes: A High Resolution 3D Surface Construction Algorithm", *Computer Graphics* 21 (4):163-169, Jul. 1987.


(Continued)

Primary Examiner — Lee S Cohen
(74) *Attorney, Agent, or Firm* — Faegre Baker Daniels LLP

(57) **ABSTRACT**

An expandable electrode assembly for use in a cardiac mapping procedure includes multiple bipolar electrode pairs including a first electrode located on an outer surface and a second electrode located on an inner surface of the individual splines forming the expandable electrode assembly. Such an electrode arrangement may produce improved electrical activation signals which may be used to produce a more accurate map of the electrical activity of a patient's heart.

18 Claims, 15 Drawing Sheets



(51)	Int. Cl.								
	<i>A61B 5/00</i>	(2006.01)		6,631,290	B1	10/2003	Guck et al.		
	<i>A61B 17/00</i>	(2006.01)		6,640,119	B1	10/2003	Budd et al.		
	<i>A61B 18/00</i>	(2006.01)		6,650,927	B1	11/2003	Keidar		
(52)	U.S. Cl.			6,690,963	B2	2/2004	Ben-Haim et al.		
	CPC <i>A61B 2562/166</i>	(2013.01); <i>Y10T 29/49119</i>		6,701,176	B1	3/2004	Halperin et al.		
		(2015.01); <i>Y10T 29/49126</i>		6,728,562	B1	4/2004	Budd et al.		
				6,773,402	B2	8/2004	Govari et al.		
				6,807,439	B2	10/2004	Edwards et al.		
				6,839,588	B1	1/2005	Rudy		
(56)	References Cited			6,847,839	B2	1/2005	Ciaccio et al.		
	U.S. PATENT DOCUMENTS			6,872,428	B2	3/2005	Yang et al.		
				6,892,090	B2	5/2005	Verard et al.		
				6,892,091	B1	5/2005	Ben-Haim et al.		
				6,893,588	B2	5/2005	Lawson et al.		
				6,939,309	B1	9/2005	Beatty et al.		
				6,957,101	B2	10/2005	Porath et al.		
				6,978,168	B2	12/2005	Beatty et al.		
				6,990,370	B1	1/2006	Beatty et al.		
				7,016,719	B2	3/2006	Rudy et al.		
				7,043,292	B2	5/2006	Tarjan et al.		
				7,198,635	B2	4/2007	Danek et al.		
				7,263,397	B2	8/2007	Hauck et al.		
				7,505,810	B2	3/2009	Harlev et al.		
				7,515,954	B2	4/2009	Harlev et al.		
				7,729,752	B2	6/2010	Harlev et al.		
				8,103,327	B2	1/2012	Harlev et al.		
				8,137,343	B2	3/2012	Harlev et al.		
				8,364,235	B2	1/2013	Kordis et al.		
				8,447,377	B2	5/2013	Harlev et al.		
				8,463,368	B2	6/2013	Harlev et al.		
				8,538,509	B2	9/2013	Harlev et al.		
				8,725,240	B2	5/2014	Harlev et al.		
				8,728,075	B2	5/2014	Wu et al.		
				8,755,861	B2	6/2014	Harlev et al.		
				8,825,130	B2	9/2014	Just et al.		
				9,014,793	B2	4/2015	Harlev et al.		
				9,585,588	B2	3/2017	Marecki et al.		
				9,687,166	B2	6/2017	Subramaniam et al.		
				2002/0151807	A1	10/2002	Goldin		
				2002/0177766	A1	11/2002	Mogul		
				2003/0018251	A1	1/2003	Solomon		
				2003/0065271	A1	4/2003	Khoury		
				2003/0076277	A1	4/2003	Muramatsu et al.		
				2003/0078509	A1	4/2003	Panescu		
				2003/0216630	A1	11/2003	Jersey-Willuhn et al.		
				2004/0077942	A1	4/2004	Hall et al.		
				2004/0097806	A1	5/2004	Hunter et al.		
				2004/0243015	A1	12/2004	Smith et al.		
				2004/0254437	A1	12/2004	Hauck et al.		
				2005/0033136	A1	2/2005	Govari et al.		
				2005/0038337	A1	2/2005	Edwards		
				2005/0054918	A1	3/2005	Sra		
				2005/0107834	A1	5/2005	Freeman et al.		
				2005/0154282	A1	7/2005	Li et al.		
				2005/0288599	A1	12/2005	MacAdam et al.		
				2006/0085049	A1	4/2006	Cory et al.		
				2006/0116575	A1	6/2006	Willis		
				2006/0122526	A1	6/2006	Berenfeld et al.		
				2006/0173251	A1	8/2006	Govari et al.		
				2006/0178587	A1	8/2006	Khoury		
				2006/0241401	A1	10/2006	Govari et al.		
				2007/0016007	A1	1/2007	Govari et al.		
				2007/0038078	A1	2/2007	Osadchy		
				2007/0049821	A1	3/2007	Willis		
				2007/0083194	A1	4/2007	Kunis et al.		
				2007/0197929	A1	8/2007	Porath et al.		
				2007/0265539	A1	11/2007	Hastings et al.		
				2007/0270703	A1	11/2007	He et al.		
				2007/0287902	A1	12/2007	Fuimaono et al.		
				2007/0299351	A1	12/2007	Harlev et al.		
				2007/0299352	A1	12/2007	Harlev et al.		
				2007/0299353	A1	12/2007	Harlev et al.		
				2008/0190438	A1	8/2008	Harlev et al.		
				2008/0221566	A1	9/2008	Krishnan		
				2008/0234588	A1	9/2008	Feldman et al.		
				2008/0249424	A1	10/2008	Harlev et al.		
				2009/0171274	A1	7/2009	Harlev et al.		
				2009/0177072	A1	7/2009	Harlev et al.		
				2009/0253976	A1	10/2009	Harlev et al.		

(56) **References Cited**

U.S. PATENT DOCUMENTS

2009/0299355 A1 12/2009 Bencini et al.
 2010/0286551 A1 11/2010 Harlev et al.
 2012/0271135 A1 10/2012 Burke et al.
 2012/0277567 A1 11/2012 Harlev et al.
 2013/0172715 A1 7/2013 Just et al.
 2013/0274582 A1 10/2013 Afonso et al.
 2013/0345538 A1 12/2013 Harlev et al.
 2014/0018880 A1 1/2014 Zarins et al.
 2014/0200442 A1 7/2014 Harlev et al.
 2014/0238175 A1 8/2014 Huszar et al.
 2014/0275921 A1 9/2014 Harlev et al.
 2015/0223726 A1 8/2015 Harlev
 2015/0342491 A1 12/2015 Marecki et al.
 2015/0374252 A1 12/2015 de la Rama et al.
 2017/0143227 A1 5/2017 Marecki et al.

FOREIGN PATENT DOCUMENTS

EP 0779059 B1 6/1997
 EP 1484026 A1 12/2004
 EP 2265172 A2 12/2010
 EP 2269505 B1 5/2012
 JP H08511438 A 12/1996
 JP H11504541 A 4/1999
 JP 2002078694 A 3/2002
 JP 2011507656 A 3/2011
 JP 2014518654 A 8/2014
 JP 2016521180 A 7/2016
 WO WO9725917 A1 7/1997
 WO 2008097767 A2 8/2008
 WO 2009085108 A1 7/2009
 WO 2009123819 A2 10/2009
 WO 2013028998 A2 2/2013
 WO 2014110579 A1 7/2014
 WO 2015187386 A1 12/2015
 WO 2015187430 A2 12/2015

OTHER PUBLICATIONS

Makela et al., "A Review of Cardiac Image Registration Methods", *IEEE Transaction on Medical Imaging*, 21 (9):1011-1021, 2002.
 Malladi, R. et al., "A Geometric Approach to Segmentation and Analysis of 3D Medical Images", *Mathematical Methods in Biomedical Image Analysis*, Proceedings of the Workshop on, Jun. 21-22, 1996, pp. 244-252.
 Mangan, Alan et al., "Partitioning 3D Surface Meshes Using Watershed Segmentation", *IEEE Transactions on Visualization and Computer Graphics*, 5(4):308-321, 1999.
 Meininger et al., "Initial Experience with a Novel Focused Ultrasound Ablation System for Ring Ablation Outside the Pulmonary Vein", *Journal of Interventional Cardiac Electrophysiology*, 8:141-148, 2003.
 Merrill, Daniel R. et al., "Electrical stimulation of excitable tissue: design of efficacious and safe protocols", *Journal of Neuroscience Methods*, 141:171-198, 2005.
 Miller, "Editor's Forum—Application of Registration for Ablation: A Marriage of Technologies", *Journal of Interventional Cardiac Electrophysiology*, 11:87-89, 2004.
 Nademanee et al., "A New Approach for Catheter Ablation of Atrial Fibrillation: Mapping of the Electrophysiologic Substrate", *Journal of the American College of Cardiology*, 43(11):2044-2053, 2004.
 Non-Final Office Action in U.S. Appl. No. 11/451,908, dated Sep. 4, 2008, 12 pages.
 Non-Final Office Action issued in U.S. Appl. No. 11/451,898 dated Sep. 25, 2008, 13 pages.
 Noseworthy et al., "The Impact of Respiration on Left Atrial and Pulmonary Venous Anatomy: Implications for Image-Guided Intervention", *Heart Rhythm*, 2:1173-1178, 2005.
 Pappone et al., "Robotic Magnetic Navigation for Atrial Fibrillation Ablation", *Journal of the American College of Cardiology*, 47(7):1390-1400, 2006.

Paragios, "A Level Set Approach for Shape-Driven Segmentation and Tracking of the Left Ventricle", *IEEE Transactions on Medical Imaging*, 22(6):773-776, 2003.

Persson et al., "A Simple Mesh Generator in MATLAB", *SIAM Review*, 46(2):329-345, 2004.

Persson, "Mesh Generation for Implicit Geometries", Massachusetts Institute of Technology—Thesis, Feb. 5, 2006.

Pham, Dzung, et al., "Current Methods in Medical Image Segmentation", *Annu. Rev. Biomed. Eng.*, 02:315-337, 2000.

Rao et al., "Novel Noncontact Catheter System for Endocardial Electrical and Anatomical Imaging", *Annals of Biomedical Engineering*, 32(4):573-584, 2004.

Reddy et al., "Use of a Diode Laser Balloon Ablation Catheter to Generate Circumferential Pulmonary Venous Lesions in an Open-Thoracotomy Caprine Model", *PACE*, 27:52-57, 2004.

Reddy et al., "Integration of Cardiac Magnetic Resonance Imaging with Three-Dimensional Electroanatomic Mapping to Guide Left Ventricular Catheter Manipulation—Feasibility in a Porcine Model of Healed Myocardial Infarction", *Journal of the American College of Cardiology*, 44(11):2202-2213, 2004.

Sanders et al., "Spectral Analysis identifies Sites of High-Frequency Activity Maintaining Atrial Fibrillation in Humans", *Circulation*, 112:789-797, 2005.

Sethian, "Level Set Methods and Fast Marching Methods: Evolving Interfaces in Computational Geometry, Fluid Mechanics, Computer Vision, and Materials Science", Department of Mathematics—University of California, Berkeley, Cambridge University Press, 1999.

Simon et al. "Electroanatomic Mapping of the Right Arm With a Right Atrial Basket Catheter and Three-Dimensional Intracardiac Echocardiography", *PACE*, 27: 318-326, 2004.

Smits et al., "Catheter-Based Intramyocardial Injection of Autologous Skeletal Myoblasts as a Primary Treatment of Ischemic Heart Failure", *Journal of the American College of Cardiology*, 42(12):2063-2069, 2003.

Solomon et al., "Real-Time Cardiac Catheter Navigation on Three-Dimensional CT Images", *Journal of Interventional Cardiac Electrophysiology*, 8:27-36, 2003.

Sra et al., "Registration of Three-Dimensional Left Atrial Computed Tomographic Images With Projection Images Obtained Using Fluoroscopy", *Circulation*, 112:3763-3768, 2005.

Sra, Jasbir et al., "Registration of 3D Computed Tomographic Images With Interventional Systems: Implications for Catheter Ablation of Atrial Fibrillation", *J Intery Card Electrophysiol*, 16:141-148, 2006.

Stevenson, "Radiofrequency Catheter Ablation of Ventricular Tachycardia After Myocardial Infarction", *Circulation*, 98:308-314, 1998.

Taccardi et al., "A New Intracavitary Probe for Detecting the Site of the Origin of Ectopic Ventricular Beats During One Cardiac Cycle", *Circulation*, 75(1):272-281, 1987.

Thal et al., "Novel Applications in Catheter Ablation", *Journal of Interventional Cardiac Electrophysiology*, 13:17-21, 2005.

Thiagalingam et al., "Noncontact Mapping of the Left Ventricle: Insights from Validation With Transmural Contact Mapping", *PACE*, 27:570-578, 2004.

Voth, "The Inverse Problem of Electrocardiography: Industrial Solutions and Simulations", BEM and NFSI Conference Proceedings, Minneapolis, MN, May 12-15, 2005, pp. 191-194.

Wittkampf et al., "LocaLisa: New Technique for Real-Time 3-Dimensional Localization of Regular Intracardiac Electrodes", *Circulation*, 99:1312-1317, 1999.

Written Opinion of the International Searching Authority issued in PCT/US208/13553, dated Feb. 5, 2009, 6 pages.

Yezzi, Anthony et al., "A Geometric Snake Model for Segmentation", *IEEE Transactions on Medical Imaging*, 16(2) Apr. 1997.

Written Opinion of International Searching Authority issued in PCT/US2009/061277, dated Apr. 8, 2010, 10 pages.

Adams et al., "Seeded Region Growing", *IEEE Transactions on Pattern Analysis and Machine Intelligence*, 16 (6):641-647, 1994.

Arthur, "Clinical Use of Intracardiac Impedance: Current Applications and Future Perspectives", *PACE*, vol. 24:500-506, Apr. 2001.

(56)

References Cited

OTHER PUBLICATIONS

- Authorized officer Carl H. Layno, International Search Report and the Written Opinion in PCT/US07/70854 dated Sep. 12, 2008, 15 pages.
- Authorized officer Lee W. Young, International Search Report and the Written Opinion in PCT/US08/52385 dated Aug. 8, 2008, 11 pages.
- Authorized officer, Blaine R. Copenheaver, International Search Report and the Written Opinion in PCT/US2009/061277 dated Apr. 8, 2010, 13 pages.
- Baan, Jan et al., "Continuous Measurement of Left Ventricular Volume in Animals and Humans by Conductance Catheter", *Circulation*, 07(5):812-823, 1984.
- Badics, "Real-Time Reconstruction of Endocardial Potential Maps in Non-Contact Cardiac Mapping", *International Journal for computation and Mathematics in Electrical Engineering (COMPEL)*, vol. 28, No. 4, 2009.
- Ben-Haim et al., "Nonfluoroscopic, in Vivo Navigation and Mapping Technology", *Nature Medicine*, 2(12):1393-1395, 1996.
- Besl et al., "A Method for Registration of 3-D Shapes", *IEEE Transaction on Pattern Analysis and Machine Intelligence*, 14(2):239-256, 1992.
- Blomstrom-Lundqvist et al., "ACC/AHA/ESC Guidelines for the Management of Patients with Supraventricular Arrhythmias—Executive Summary", *Journal of the American College of Cardiology*, 42(8):1493-1531, 2003.
- Breithardt et al., "AHA Medical/Scientific Statement—Special Report: Standards for Analysis of Ventricular Late Potentials Using High-Resolution or Signal/Averaged Electrocardiography", *Circulation*, 83(4):1481-1488, 1991.
- Brooks et al., "Electrical Imaging of the Heart", *IEEE Signal Processing Magazine*, pp. 24-42, 1997.
- Caspi et al., "Stem Cell Research: Regenerating the Heart Using Human Embryonic Stem Cells—from Cell to Bedside", *IMAJ* 8:208-214, 2006.
- Cheney et al., "Electrical Impedance Tomography," *SIAM Review* 41, pp. 85-101, 1999.
- Communication pursuant to Article 94(3) EPC in European Application No. 07798369, dated Nov. 17, 2011, 5 pages.
- De Groot et al., "Three-Dimensional Catheter Positioning During Radiofrequency Ablation in Patients: First Application of a Real-Time Position Management System", *Journal of Cardiovascular Electrophysiology*, 11:1183-1192, 2000.
- Donahue et al., "Focal Modification of Electrical Conduction in the Heart by Viral Gene Transfer", *Nature Medicine*, 6(12):1395-1398, 2000.
- Dong et al., "Integrated Electroanatomic Mapping With Three-Dimensional Computed Tomographic Images for Real-Time Guided Ablations", *Circulation* 113:186-194, 2006.
- Durrer et al., "Total Excitation of the Isolated Human Heart", *Circulation*, XL1:899-912, 1970.
- Ector, Joris et al., "Cardiac Three-Dimensional Magnetic Resonance Imaging and Fluoroscopy Merging", *Circulation*, 112:3769-3776, 2005.
- European Search Report issued in EP Application No. 12815179.2, dated Apr. 28, 2015, 6 pages.
- Extended European Search Report issued in EP Application No. 09727423.7, dated May 15, 2012, 5 pages.
- Friedman, "Catheter Cryoablation of Cardiac Arrhythmias", *Current Opinion in Cardiology*, 20:48-54, 2005.
- Friedman, "Novel Mapping Techniques for Cardiac Electrophysiology", *Heart*, 87:575-582, 2002.
- Geddes et al., "Criteria for the Selection of Materials for Implanted Electrodes," *Annals of Biomedical Engineering* 31:879-890, 2003.
- Gepstein et al., "A Novel Method for Nonfluoroscopic Catheter-Based Electroanatomical Mapping of the Heart", *Circulation* 95:1611-1622, 1997.
- Haug, E. J. et al.: *Design Sensitivity Analysis of Structural Systems*, Mathematics in Science and Engineering, vol. 177 (1986).
- Huang, Yi-Chih et al., "Development of a Third Generation Intraventricular Impedance Imaging (Iii) System Evaluation of Hardware Design", *Engineering in Medicine and Biology Society, Proceedings of the 19th Annual Internal Conference of the IEEE/EMBS*, 6:2442-2444 Oct. 30-Nov. 2, 1997.
- International Preliminary Report on Patentability issued in PCT/US2008/013553, dated Feb. 5, 2009, 6 pages.
- International Preliminary Report on Patentability issued in PCT/US2008/052385, dated Aug. 8, 2008, 6 pages.
- International Preliminary Report on Patentability issued in PCT/US2009/036099, dated Oct. 14, 2010, 20 pages.
- International Search Report and Written Opinion issued in PCT/US2009/036099, dated Apr. 28, 2009, 21 pages.
- International Search Report and Written Opinion issued in PCT/US2014/060137, dated Dec. 10, 2014, 11 pages.
- International Search Report and Written Opinion issued in PCT/US2015/032753, dated Mar. 9, 2016, 17 pages.
- Jain et al., "Cell Therapy Attenuates Deleterious Ventricular Remodeling and Improves Cardiac Performance after Myocardial Infarction", *Circulation*, 103:1920-1927, 2001.
- Jalife, "Rotors and Spiral Waves in Atrial Fibrillation", *Journal of Cardiovascular Electrophysiology*, 14:776-780, 2003.
- Jane et al., Alignment Methods for Averaging of High-Resolution Cardiac Signals: A Comparative Study of Performance, *IEEE Transaction on Biomedical Engineering*, 38(6):571-579, 1991.
- Jia et al., "Electrophysiologic Endocardial Mapping from a Noncontact Nonexpandable Catheter: A Validation Study of a Geometry-Based Concept". *Journal of Cardiovascular Electrophysiology*, 11:1238-1251, 2000.
- Kikuchi et al., "Targeted Modification of Atrial Electrophysiology by Homogeneous Transmural Atrial Gene Transfer", *Circulation*, 111:264-270, 2005.
- Kistler et al., "Validation of Three-Dimensional Cardiac Image integration: Use of Integrated CT Image into Electroanatomic Mapping System to Perform Catheter Ablation of Atrial Fibrillation", *Journal of Cardiovascular Electrophysiology*, 17:341-348, 2006.
- Kuklik et al., The reconstruction, from a set of points, and analysis of the interior surface of the heart chamber, *Physiological Measurement* 25, pp. 617-627, 2004.
- Kun, Stevan et al., "Conductance Volumetric Model of an Eccentrically Positioned Catheter within a Three-Compartment Ellipsoidal Ventricle", *U, IEEE Transactions on Jun. 1993*, 40(6); 589-592.
- L. Piegi, W. Tiller: *The NURBS Book*, 2nd Edition, Springer (1997).
- Laciar et al., "Improved Alignment Method for Noisy High-Resolution ECG and Holter Records Using Multiscale Cross-Correlation", *IEEE Transactions on Biomedical Engineering*, 50(3):344-353, 2003.
- Liu et al., "Endocardial Potential Mapping from a Noncontact Nonexpandable Catheter: A Feasibility Study", *Annals of Biomedical Engineering*, 26:994-1009, 1998.
- International Preliminary Report on Patentability issued in PCT/US2015/032004, dated Dec. 6, 2016, 7 pages.
- International Preliminary Report on Patentability issued in PCT/US2015/032753, dated Dec. 6, 2016, 10 pages.
- International Preliminary Report on Patentability issued in PCT/US2014/060137, dated Apr. 28, 2016, 9 pages.
- International Search Report and Written Opinion issued in PCT/US2015/032004, dated Sep. 4, 2015, 8 pages.




FIG. 1




FIG. 2A




FIG. 2B




FIG. 3A




FIG. 3B




FIG. 3D




FIG. 3C




FIG. 3F




FIG. 3E




FIG. 3G




FIG. 4A




FIG.4B




FIG. 4C




FIG. 4D




FIG. 5B




FIG. 5A




FIG. 6




FIG. 7B




FIG. 7A




FIG. 8B




FIG. 8A




FIG. 9A

FIG. 9B

FIG. 9C

FIG. 9D

1

ELECTRODE ASSEMBLY**CROSS-REFERENCE TO RELATED APPLICATIONS**

This application claims priority under 35 U.S.C. §119 to U.S. Provisional Application Ser. No. 62/007,716, filed Jun. 4, 2014, the entirety of which is incorporated herein by reference.

TECHNICAL FIELD

The present disclosure generally relates to catheters and electrode assemblies for use in cardiac procedures and more particularly, to an electrode assembly that may be utilized in a cardiac mapping procedure.

BACKGROUND

Mapping the electrical activity of the heart is a critical component for the diagnosis and treatment of heart disease. Many advanced therapies (such as ablation for the treatment of arrhythmias) require detailed electroanatomic mapping. Currently, mapping is performed in an electrophysiology (EP) lab, during which mapping catheters are inserted into the heart and carefully moved to various locations around the heart to map and identify the origins of the arrhythmia. Once the origin of the arrhythmia is identified, the specific tissue may be destroyed by ablation.

SUMMARY

The present disclosure generally relates to catheter and electrode assemblies for use in cardiac procedures and more particularly, to an electrode assembly that may be utilized in a cardiac mapping procedure.

In one example, a catheter includes: an elongate catheter body extending from a proximal end to a distal end; and an expandable electrode assembly disposed at the distal end of the catheter body, the electrode assembly configured to transition from a collapsed configuration to an expanded configuration and comprising at least one flexible member having an outer surface and an inner surface, wherein the at least one flexible member comprises a first electrode disposed on the outer surface of the flexible member and a second electrode disposed on the inner surface of the flexible member.

In addition or alternatively to any one or more of the above, and in another example, the first and second electrodes are configured to form a bipolar electrode pair.

In addition or alternatively to any one or more of the above, and in another example, the first electrode is located directly opposite the second electrode.

In addition or alternatively to any one or more of the above, and in another example, the first electrode is offset from the second electrode.

In addition or alternatively to any one or more of the above, and in another example, the flexible member comprises at least one flexible printed circuit.

In addition or alternatively to any one or more of the above, and in another example, the flexible member comprises a single, dual sided flexible printed circuit wherein the first electrode is formed on an outer surface of the flexible printed circuit and the second electrodes is formed on an inner surface of the flexible printed circuit.

In addition or alternatively to any one or more of the above, and in another example, the flexible member com-

2

prises a first flexible printed circuit defining the first electrode formed on an upper surface of a substrate and a second flexible printed circuit defining the second electrode formed on a lower surface of the substrate.

5 In addition or alternatively to any one or more of the above, and in another example, a distance between the first electrode and the second electrode is less than about 0.5 mm.

In addition or alternatively to any one or more of the above, and in another example, the flexible member comprises multiple bipolar electrode pairs defined by a first electrode disposed on the outer surface of the flexible member and a second electrode disposed on the inner surface of the flexible member.

10 In addition or alternatively to any one or more of the above, and in another example, further comprising two or more flexible members, each of the two or more flexible members comprising at least a first electrode disposed on the outer surface of the flexible member and at least a second electrode disposed on the inner surface of the flexible member.

15 In addition or alternatively to any one or more of the above, and in another example, wherein the first and second electrodes form a bipolar electrode pair across the outer and inner surface of the flexible member.

20 In another example, a method of forming a flexible electrode assembly includes: forming a flexible electrode assembly comprising at least one flexible member having an outer surface and an inner surface, wherein the at least one flexible member comprises a first electrode disposed on the outer surface of the flexible member and a second electrode disposed on the inner surface of the flexible member and wherein the flexible electrode assembly is configured to transition from a collapsed configuration to an expanded configuration; and coupling the flexible electrode assembly to a distal end of an elongate catheter body.

25 In addition or alternatively to any one or more of the above, and in another example, the method further includes: forming a flexible layered sheet comprising at least one flexible printed circuit defining a first electrode on an outer surface of the flexible layered sheet and a second electrode on an inner surface of the flexible layered sheet; separating the flexible layered sheet into two or more flexible members, each flexible member having a first electrode located on an outer surface and a second electrode located on an inner surface; and forming an expandable electrode assembly from at least one of the flexible members.

30 In addition or alternatively to any one or more of the above, and in another example, the method further includes forming an expandable electrode assembly from two or more flexible members by joining the two or more flexible members together at a first end of each of the two or more flexible members.

35 In addition or alternatively to any one or more of the above, and in another example, the method further includes joining the two or more flexible members together at a second end of each of the two or more flexible members.

40 In another example, a catheter includes an elongate catheter body extending from a proximal end to a distal end; and an expandable electrode assembly disposed at the distal end of the catheter body, the electrode assembly configured to transition from a collapsed configuration to an expanded configuration and comprising at least one flexible member having an outer surface and an inner surface, wherein the at least one flexible member comprises a first electrode disposed on the outer surface of the flexible member and a second electrode disposed on the inner surface of the flexible member.

In addition or alternatively to any one or more of the above, and in another example, the first and second electrodes are configured to form a bipolar electrode pair.

In addition or alternatively to any one or more of the above, and in another example, the first electrode is located directly opposite the second electrode.

In addition or alternatively to any one or more of the above, and in another example, the first electrode is offset from the second electrode.

In addition or alternatively to any one or more of the above, and in another example, the catheter further includes two or more flexible members, each of the two or more flexible members comprising at least a first electrode disposed on the outer surface of the flexible member and at least a second electrode disposed on the inner surface of the flexible member.

In addition or alternatively to any one or more of the above, and in another example, the flexible member comprises at least one flexible printed circuit.

In addition or alternatively to any one or more of the above, and in another example, the flexible member comprises a single, dual sided flexible printed circuit wherein the first electrode is formed on an outer surface of the flexible printed circuit and the second electrodes is formed on an inner surface of the flexible printed circuit.

In addition or alternatively to any one or more of the above, and in another example, the flexible member comprises a first flexible printed circuit defining the first electrode formed on an upper surface of a substrate and a second flexible printed circuit defining the second electrode formed on a lower surface of the substrate.

In addition or alternatively to any one or more of the above, and in another example, a distance between the first electrode and the second electrode is less than about 0.5 mm.

In addition or alternatively to any one or more of the above, and in another example, the flexible member comprises multiple bipolar electrode pairs defined by a first electrode disposed on the outer surface of a flexible member and a second electrode disposed on the inner surface each of the flexible member.

In another example, a catheter includes: an elongate catheter body extending from a proximal end to a distal end; and an expandable electrode assembly disposed at the distal end of the catheter body, the electrode assembly configured to transition from a collapsed configuration to an expanded configuration and comprising two or more flexible splines having an outer surface and an inner surface, wherein at least one of the two or more flexible splines comprises at least a first electrode disposed on the outer surface of the flexible spline and at least a second electrode disposed on the inner surface of the flexible spline.

In addition or alternatively to any one or more of the above, and in another example, the first and second electrodes are configured to form a bipolar electrode pair.

In addition or alternatively to any one or more of the above, and in another example, the first electrode is located directly opposite the second electrode.

In addition or alternatively to any one or more of the above, and in another example, the first electrode is offset from the second electrode.

In addition or alternatively to any one or more of the above, and in another example, each of the two or more splines comprises multiple bipolar electrode pairs defined by a first electrode disposed on the outer surface of a flexible spline and a second electrode disposed on the inner surface of the flexible spline.

In addition or alternatively to any one or more of the above, and in another example, each of the two or more flexible splines comprises at least one flexible printed circuit.

In addition or alternatively to any one or more of the above, and in another example, the at least one flexible circuit is a single, dual sided flexible printed circuit having a first electrode is formed on an upper surface of the flexible printed circuit and a second electrode formed on a lower surface of the flexible printed circuit.

In addition or alternatively to any one or more of the above, and in another example, each of the two or more flexible splines comprises a first flexible printed circuit defining a first electrode formed on an upper surface of a substrate and a second flexible printed circuit defining a second electrode formed on a lower surface of the substrate.

In yet another example, a method of forming a flexible electrode assembly includes: forming a flexible electrode assembly comprising at least one flexible member having an outer surface and an inner surface, wherein the at least one flexible member comprises a first electrode disposed on the outer surface of the flexible member and a second electrode disposed on the inner surface of the flexible member and wherein the flexible electrode assembly is configured to transition from a collapsed configuration to an expanded configuration; and coupling the flexible electrode assembly to a distal end of an elongate catheter body.

In addition or alternatively to any one or more of the above, and in another example, the method further includes forming a flexible layered sheet comprising at least one flexible printed circuit defining a first electrode on an outer surface of the flexible layered sheet and a second electrode on an inner surface of the flexible layered sheet; separating the flexible layered sheet into two or more flexible members, each flexible member having a first electrode located on an outer surface and a second electrode located on an inner surface; and forming the expandable electrode assembly from at least one of the flexible members.

In addition or alternatively to any one or more of the above, and in another example, the method further includes forming an expandable electrode assembly from two or more flexible members by joining the two or more flexible members together at least at a first end of each of the two or more flexible member.

In still another example, a method of forming a flexible electrode assembly is disclosed. The method includes: forming a first flexible printed circuit comprising one or more electrodes on an upper surface of a substrate and forming a second flexible printed circuit comprising one or more electrodes on a lower surface of the substrate to produce a flexible layered sheet; separating the flexible layered sheet into two or more splines extending longitudinally from a proximal end of the flexible layered sheet to a distal end of the flexible layered sheet, wherein the two or more splines are fully separated from one another such that they are not connected and wherein each of the two or more splines comprises at least one bipolar electrode pair defined by a first electrode from the first flexible printed circuit disposed on the upper surface of the substrate and a second electrode from the second flexible printed circuit disposed on the lower surface of the substrate, each electrode located on opposite sides of each of the two or more splines; mechanically joining the fully separated two or more flexible splines together to form an expandable electrode assembly.

In addition or alternatively to any one or more of the above, the substrate comprises Nitinol.

In addition or alternatively to any one or more of the above, the step of separating the flexible layered sheet into two or more splines comprises laser cutting the flexible layered sheet into two or more splines.

In addition or alternatively to any one or more of the above, the step of separating the flexible layered sheet into two or more splines comprises die cutting the flexible layered sheet into two or more splines.

In addition or alternatively to any one or more of the above, the method further includes securing a second end of the first spline and a second end of the second spline to a distal end of a catheter body.

In addition or alternatively to any one or more of the above, wherein the fully separated splines are mechanically joined together by inserting their respective distal ends into corresponding slots provided in a distal cap.

The above summary of some embodiments is not intended to describe each disclosed embodiment or every implementation of the present disclosure. The Figures, and Detailed Description, which follow, more particularly exemplify these embodiments.

BRIEF DESCRIPTION OF THE DRAWINGS

The disclosure may be more completely understood in consideration of the following detailed description of various embodiments in connection with the accompanying drawings, in which:

FIG. 1 is a schematic diagram showing a catheter in the context of a system;

FIGS. 2A-2B are schematic views of an exemplary catheter that may be utilized in the system shown in FIG. 1;

FIGS. 3A-3G are schematic views of exemplary expandable electrode assemblies;

FIG. 4A is an isometric view of an expandable electrode assembly shown in a collapsed configuration;

FIG. 4B is an isometric view of the expandable electrode assembly of FIG. 4A shown in an expanded configuration;

FIG. 4C is a close-up, isometric view of the expandable electrode assembly of FIG. 4B in the expanded configuration;

FIG. 4D is a top plan view of the expandable electrode assembly shown in FIGS. 4B and 4C in the expanded configuration;

FIGS. 5A and 5B are close-up schematic views of exemplary individual splines of an expandable electrode assembly including multiple bipolar electrode pairs;

FIG. 6 is a side, cross-sectional view of an exemplary multi-layered flexible sheet including at least one flexible printed circuit used to construct an expandable electrode assembly;

FIGS. 7A and 7B are schematic views of the inner and outer surfaces, respectively, a portion of a flexible spline incorporating a single flexible printed circuit;

FIG. 8A shows a top plan view of a multi-layered flexible sheet used to construct an electrode assembly;

FIG. 8B is a bottom plan view of the multi-layered flexible sheet shown in FIG. 8A; and

FIGS. 9A-9D provide a stepwise illustration of a method of constructing the expandable electrode assembly from a multi-layered flexible sheet including at least one flexible printed circuit.

While the disclosure is amenable to various modifications and alternative forms, specifics thereof have been shown by way of example in the drawings and will be described in detail. It should be understood, however, that the intention is not to limit aspects of the disclosure to the particular

embodiments described. On the contrary, the intention is to cover all modifications, equivalents, and alternatives falling within the spirit and scope of the disclosure.

DETAILED DESCRIPTION

The following detailed description should be read with reference to the drawings in which similar elements in different drawings are numbered the same. The detailed description and the drawings, which are not necessarily to scale, depict illustrative embodiments and are not intended to limit the scope of the disclosure. The illustrative embodiments depicted are intended only as exemplary. Selected features of any illustrative embodiment may be incorporated into an additional embodiment unless clearly stated to the contrary.

For the following defined terms, these definitions shall be applied, unless a different definition is given in the claims or elsewhere in this specification.

All numeric values are herein assumed to be modified by the term “about”, whether or not explicitly indicated. The term “about” generally refers to a range of numbers that one of skill in the art would consider equivalent to the recited value (i.e., having the same function or result). In many instances, the term “about” may be indicative as including numbers that are rounded to the nearest significant figure.

The recitation of numerical ranges by endpoints includes all numbers within that range (e.g., 1 to 5 includes 1, 1.5, 2, 2.75, 3, 3.80, 4, and 5).

Although some suitable dimensions, ranges and/or values pertaining to various components, features and/or specifications are disclosed, one of skill in the art, incited by the present disclosure, would understand desired dimensions, ranges and/or values may deviate from those expressly disclosed.

As used in this specification and the appended claims, the singular forms “a”, “an”, and “the” include plural referents unless the content clearly dictates otherwise. As used in this specification and the appended claims, the term “or” is generally employed in its sense including “and/or” unless the content clearly dictates otherwise.

Mapping the electrophysiology of heart rhythm disorders often involves the introduction of a constellation catheter or other mapping/sensing device having a plurality of electrodes and/or sensors (e.g., CONSTELLATION®, commercially available from Boston Scientific) into a cardiac chamber. The sensors detect the electric activity of the heart at sensor locations. It may be desirable to have the electric activity processed into electrogram signals that accurately represent cellular excitation through cardiac tissue relative to the sensor locations. A processing system may then analyze and output the signal to a display device. Further, the processing system may output the signal as an activation or vector field map. The physician may use the activation or vector field map to perform a diagnostic procedure.

FIG. 1 is a high level, schematic view of an overall system 2 that includes a physician, a patient, catheters, including a mapping catheter 10, and related electrophysiology equipment located within an operating room. A physician 16 introduces the catheter 10 into the vasculature of the patient 11 at the patient's leg and advances it along a blood vessel ultimately, entering the patient's heart 12. As will be described in greater detail herein, the catheter 10 may include an electrode assembly having multiple sensing electrodes for sensing the electrical activity of the heart. Other catheters that may be used in the procedure are represented by companion catheter 18. Each catheter 10, 18 is coupled

to a processing system **20** using appropriate catheter cabling typified by connection cable **17**. If the companion catheter **18** is an ablation catheter, then processing system **20** also forms an interface to an RF ablation unit (not illustrated).

Processing system **20** may include dedicated circuitry (e.g., discrete logic elements and one or more microcontrollers; a memory or one or more memory units, application-specific integrated circuits (ASICs); and/or specially configured programmable devices, such as, for example, programmable logic devices (PLDs) or field programmable gate arrays (FPGAs)) for receiving and/or processing the acquired activation signals. In at least some embodiments, processing system **20** includes a general purpose microprocessor and/or a specialized microprocessor (e.g., a digital signal processor, or DSP, which may be optimized for processing activation signals) that executes instructions to receive, analyze and display information associated with the received activation signals. In such implementations, processing system **20** can include program instructions, which when executed, perform part of the signal processing. Program instructions can include, for example, firmware, microcode or application code that is executed by microprocessors or microcontrollers. In addition, the processing system **20** may include suitable signal conditioning circuitry including signal amplifiers, rectifiers, filters, etc. for improving the quality of the incoming activation signal. The above-mentioned implementations are merely exemplary. A variety of processing systems **20** are contemplated.

In some embodiments, processing system **20** may be configured to measure the electrical activity in the myocardial tissue adjacent to one or more electrodes located on the electrode assembly. For example, in some embodiments, processing system **20** may be configured to detect electrical activity associated with a dominant rotor or divergent activation pattern in the anatomical feature being mapped. Dominant rotors and/or divergent activation patterns may have a role in the initiation and maintenance of atrial fibrillation, and ablation of the rotor path, rotor core, and/or divergent foci may be effective in terminating the atrial fibrillation. In either situation, processing system **20** processes the sensed activation signals to generate a display of relevant characteristics, such as an isochronal map, activation time map, action potential duration (APD) map, a vector field map, a contour map, a reliability map, an electrogram, a cardiac action potential, and/or the like. The relevant characteristics may be used by the physician to identify a site suitable for ablation therapy.

In use, the physician looks at a computer display **26**. Present on the display **26** is a substantial amount of information. A large window presents an image of the heart chamber **13** along with an image of the catheter **10**. The physician will manipulate and control the catheter **10** based in part on the images and other data presented on the display **26**. The representation of the heart chamber **13** may use color, wire frame, or other techniques to depict the structure of the heart chamber **13** and to simultaneously portray electrical activity of the patient's heart. In some cases, it may be useful to display chamber geometry, catheter location, and electrical activity in an integrated fashion on the display **26**. In use, the physician will observe the display **26** and interact with the workstation **24** and the catheters **10** and **18**, to direct a medical procedure such as, for example, a cardiac mapping procedure.

FIGS. **2A** and **2B** are schematic views of an exemplary intravascular catheter **10** that may be utilized in the context of the system **2** shown in FIG. **1**. The catheter **10** may be deployed at a target location within a patient's heart. The

catheter **10** may be used to map electro-anatomical characteristics of the heart and/or to locate and position other catheters within the heart. Electrode stability and the known spatial geometry of the electrodes may improve the accuracy of the mapping device. In some cases, the catheter **10** may include an expandable electrode assembly **30** including one or more electrodes that may be used for cardiac mapping or diagnosis, ablation and/or other therapies involving the application of energy to a patient's heart. The expandable one or more electrodes may be located on the inner and/or outer surfaces of at least one flexible member forming the expandable electrode assembly **30**.

As shown in FIGS. **2A** and **2B**, the catheter **10** includes an elongate catheter body **34** extending from a proximal end **38** to a distal end **42**. In addition, the catheter body **34** may include a lumen (not shown) extending there through, but this is not required in all embodiments. The catheter body **34** may have sufficient flexibility so as to navigate the tortuous pathways of a patient's vasculature system. The catheter **10** can include a handle assembly **46** coupled to the proximal end **38** of the catheter body **34**. A physician may manipulate the handle assembly **46** to deliver, steer, rotate, deploy and/or deflect the catheter **10** when performing a medical procedure.

In some cases, the handle assembly **46** may include a first actuation mechanism **48** that may be manipulated to transition the expandable electrode assembly **30** from a collapsed configuration (shown in FIG. **2A**) suitable for delivery of the catheter **10** to a target location within a patient's body (e.g. the heart) and an expanded configuration (shown in FIG. **2B**) suitable for use in a diagnostic procedure and/or delivery of a therapy. In some cases, the actuation mechanism **48** may include a pull wire that may be coupled to the expandable electrode assembly **30** that, when actuated in a proximal direction as indicated by the arrow shown in FIG. **2B**, causes the expandable electrode assembly **30** to transition from the collapsed configuration to the expanded configuration. In other cases, the actuation mechanism **48** may include a retractable sheath that, when retracted in a proximal direction as indicated by the arrow shown in FIG. **2B**, may permit the expandable electrode assembly **30** to self-expand from the collapsed configuration to the expanded configuration. These are just some examples of exemplary actuation mechanisms that may be utilized to facilitate expansion of the expandable electrode assembly **30** when the catheter **10** is in use. In some cases, the catheter body **34** may include a deflectable distal portion **52** that a physician may manipulate using a second actuation mechanism **54** provided in the handle assembly **46** to position the electrode assembly **30** nearer or adjacent to tissue of interest.

As discussed herein, the expandable electrode assembly **30** may include one or more electrodes that may be used for cardiac mapping or diagnosis, ablation and/or other therapies. In use, the expandable electrode assembly **30** may be expanded and used to position the one or more electrodes adjacent and/or in contact with the target tissue of interest to measure an electrical signal. The expandable electrode assembly **30** may include at least one flexible member or spline on which the one or more electrodes may be located. In some cases, the expandable electrode assembly may include two or more flexible members as the size and geometry of the expandable electrode assembly **30** may permit. For example, the expandable electrode assembly may include four, five, six, seven, and in some cases as many as eight flexible members, but not limited to this. In some cases, the size, geometry and number of flexible members or splines may be dependent the location of the body in which

the device is to be deployed to investigate the target tissue of interest. One or more electrodes 64 may be disposed on at least one of the flexible members forming the expandable electrode assembly 30. The electrodes 64 may be located on an inner surface, an outer surface or both the inner and outer surfaces of the at least one flexible member. In some cases, at least a first electrode 64 is located on an outer surface and at least a second electrode 64 is located on an inner surface of a flexible member. In many cases, the at least one flexible member 60 is substantially planar or flat. FIGS. 3A-3G are schematic views of exemplary expandable electrode assemblies 30a-30g including at least one flexible member 60 that may be used to support a plurality of electrodes 64 as described herein according to the various embodiments.

FIGS. 4A-4D show different views of an exemplary expandable electrode assembly 30 that may be used to support a plurality of electrodes. As shown in FIGS. 4A and 4B, the expandable electrode assembly 30 is capable of being transitioned from a generally collapsed configuration (FIG. 4A) suitable for delivery of the catheter 10 and the electrode assembly 30 to a target location within the patient's heart and an expanded configuration (FIG. 4B) suitable for use in a desired cardiac procedure such as, for example, a cardiac mapping or ablation procedure.

As shown in FIGS. 4A and 4B, the expandable electrode assembly 30 may include two or more flexible members or splines 60 which may be capable of being flexed outwardly and away from a longitudinal axis of the electrode assembly 30. In some cases, as discussed herein, an actuation mechanism may be utilized to transition the electrode assembly 30 including the two or more flexible splines 60 from the collapsed configuration (FIG. 4A) to the expanded configuration (FIG. 4B). In other cases, the flexible splines 60 may incorporate a shape memory material that may facilitate self-expansion of the flexible splines 60 and consequently, the electrode assembly 30, from the collapsed configuration to the expanded configuration. The flexible splines 60 may be relatively stiff such that the electrode assembly 30 may be expanded into a set of known, reproducible shapes capable of retaining a known spatial geometry when in use which, in some cases, may be aided by the incorporation of a shape-memory material or other stiff polymeric material such as, for example, a polyimide or PEEK into the flexible splines 60. Alternatively, depending upon the desired application, the flexible splines 60 may be fabricated such that they are somewhat compliant so as to conform to a surface of a patient's heart when placed into intimate contact with the patient's heart. In addition, in some cases, the flexible members or splines 60 may be fabricated such that they are substantially planar or flat.

The expandable electrode assembly 30 may include a number of electrodes 64 located on each of the flexible splines 60 forming an electrode array. In many cases, the electrodes 64 may be sensing electrodes. In addition, the electrode assembly 30 may include at least some current injection locator electrodes. The electrode assembly 30 may also include a tip electrode which may be used for cardiac stimulation, ablation or as a locator electrode.

Each electrode 64 may be electrically connected to the cabling in the handle assembly 46. In some cases, the signal from each individual electrode 64 may be independently available at the processing system 20 (FIG. 1). This may be achieved by passing a conductor for each electrode through the connection cable 17 (FIG. 1).

The number of electrodes 64 distributed throughout the electrode assembly 30 and the stability of the shape of electrode assembly 30, when expanded, may affect the

overall performance of the mapping system. In some cases, the electrodes 64 may have a uniform and symmetrical distribution throughout the expandable electrode assembly 30. In other cases, the electrodes 64 may have an asymmetrical distribution throughout the expandable electrode assembly 30 which may be advantageous for non-contact cardiac mapping procedures. An electrode assembly 30 having an asymmetrical distribution of electrodes 64 throughout the expandable electrode assembly 30 may also be useful for contact mapping.

The electrodes 64 may be located on the outer surfaces 66 of each of the splines 60, the inner surfaces 68 of each of the splines 60, or both the outer and inner surfaces 66, 68 of each of the flexible splines 60 as shown in FIGS. 4B and 4C. In some cases, up to sixty-four sensing electrodes 64 may be distributed over and along the various splines 60 including both the outer and inner surfaces 66, 68 of the splines 60. Depending upon the application, the electrode assembly 30 may include fewer or greater than sixty-four electrodes.

In many cases, the electrodes 64 may form at least one bipolar electrode pair. In some cases, the electrodes 64 may form multiple bipolar electrode pairs. The bipolar electrode pairs may be distributed throughout the expandable electrode assembly 30. In some cases, the bipolar electrode pairs may be formed between first and second electrodes 64 located on the same surface of a flexible member or spline 60, between first and second electrodes 64 located on opposite surfaces of a flexible member or spline 60, or between a first electrode 64 located on a first spline 60 and a second electrode 64 located on a second spline 60. In the example in which the bipolar electrode pair is formed between electrodes 64 located on different splines 60, the individual electrodes 64 forming the bipolar electrode pair may be both located on the inner surface of their respective splines 60, the outer surface of their respective splines 60 or, alternatively, one electrode 64 may be located on an outer surface of its respective spline 60 and the other electrode 64 forming the bipolar electrode pair may be located on the inner surface of its respective spline 60. These are just some examples.

In some cases, each of the flexible splines 60 may include at least one bipolar electrode pair. In some cases, all of the electrodes 64 located on the flexible splines 60 may be paired together to form a plurality of electrode pairs distributed along the length of the individual flexible splines 60. In some cases, the electrode pairs may be located equidistant from one another along the length of each of the flexible splines 60. Alternatively, the electrode pairs may have a varied spacing forming an electrode array having an asymmetrical distribution. Up to thirty-two bipolar electrode pairs may be distributed throughout the electrode assembly 30 for a total of up to sixty-four electrodes 64 depending upon the overall size and geometry of the electrode assembly 30. However, it is contemplated that the electrode assembly 30 may be configured such that it is capable of carrying fewer or greater than thirty-two bipolar electrode pairs, depending upon the overall size and geometry of the electrode assembly 30 and the desired application.

FIGS. 5A and 5B are close-up schematic views of two illustrative individual flexible members or splines 60 including a plurality of bipolar electrode pairs 72 distributed along their length. The orientation of the dipole of the bipolar electrode pair 72 does not affect the ability of the bipolar electrode pair 72 in sensing local electrical activity of the heart. However, it may be useful for the clinical or other individual performing the procedure to know whether the positive or negative electrode of the bipolar pair is located

on the outer surface 66 of the flexible splines. It will be generally understood by those skilled in the art that each bipolar electrode pair 72 formed across the flexible splines will have the same electrical orientation. In addition or alternatively, at least some bipolar electrode pairs may be formed between adjacent electrodes on the outer or inner surface 66, 68 of the splines 60 (not shown). It will be generally understood that each of the flexible splines 60 forming the electrode assembly 30 will have a similar if not, the same construction, and in some cases, may be substantially planar or flat.

As shown in FIGS. 5A and 5B, the bipolar electrode pairs 72 may be spaced an equal distance from one another along the length of the spline 60. In some cases, all of the bipolar electrode pairs 72 located along the length of each of the flexible splines 60 of the electrode assembly 30 include a first electrode 74a located on the outer surface 66 and a second electrode 74b located opposite the first electrode 74a on the inner surface 68 of the flexible splines 60. In some cases, as shown in FIG. 5A, the electrodes 74a, 74b forming the bipolar electrode pair 72 may be located directly opposite one another across the spline 60. In addition or alternatively, as shown in FIG. 5B, the electrodes 74a, 74b forming the bipolar electrode pair 72 may be offset from one another such that the electrodes 74a, 74b forming the bipolar electrode pairs 72 along the length of the flexible member or spline 60 have a staggered configuration. In some cases, a first electrode 74a located on an outer surface 66 of a spline 60 may be offset from a second electrode 74b located on the inner surface 68 of the spline 60 by one half the measurable distance between adjacent electrodes 74b located on the inner surface of the spline 60. The reverse may also be true. A first electrode 74b located on the inner surface 68 of the spline 60 may be offset from a second electrode 74a located on the outer surface 66 of the spline 60 by one half the measurable distance between adjacent electrodes 74a located on the outer surface 66 of the spline 60. In other cases, a first electrode 74a, 74b of a bipolar electrode pair 72 may be offset from a second electrode 74a, 74b located on the opposite surface 66 or 68 of the spline 60 by one quarter, one third, two thirds or three quarters of the distance between two adjacent electrodes 74 or 74b located on the opposite surface 66 or 68 of the spline 60. These are just some examples.

Referring back to the example shown in FIG. 5A, a number of advantages may be associated with utilizing multiple bipolar electrode pairs 72 located on each of the splines 60 of the electrode assembly 30 having a first electrode 74a located on the outer surface 66 and a second electrode 74b located opposite the first electrode 74a on the inner surface 68 of each of the flexible splines 60. Because the electrodes 74a, 74b forming a bipolar electrode pair 72 are located on opposite surfaces 66, 68 of the flexible spline 60, they may be spaced closely together, separated only by the thickness of the spline 60. In some cases, the distance between the two electrodes located on the outer and inner surfaces 66, 68 of the splines 60 may be less than about 1.5 mm, less than about 1.0 mm, less than about 0.8 mm, less than about 0.5 mm and more particularly, less than about 0.4 mm. Such minimal spacing between the electrodes 74a, 74b of a bipolar electrode pair 72 may be not be possible if the electrodes 74a, 74b are located on the same surface (outer or inner) of a spline 60.

Placing the electrodes 74a, 74b on opposite surfaces 66, 68 may avoid problems such as increased impedance and susceptibility to noise associated with reducing the size of the electrodes to minimize the distance between adjacent

electrodes. For example, if the individual electrodes forming the bipolar pair are located on the same surface (outer or inner surface) of the spline 60, the spacing between the electrodes may be decreased by decreasing the size of the electrodes such that they may be spaced more closely together. However, the reduction in electrode surface area becomes problematic because the reduced electrode surface area results in an increase in impedance. Placing the electrodes on opposite surfaces of the splines may mitigate impedance concerns by allowing a suitable electrode surface area to be maintained while decreasing the spacing between electrodes.

Additionally, the reduced spacing between the electrodes of the bipolar electrode pair 72 resulting from their location on opposite surfaces of a spline 60 may improve the ability of the bipolar electrode pair 72 to reject far field noise, and may facilitate an improved reduction in noise even from nearby, adjacent tissue. The ability to reject far field signals, even those generated by adjacent tissue, may improve the output signal generated by the bipolar pair of electrodes by reducing and localizing the sensing area to the tissue directly adjacent the bipolar electrode pair and more particularly, to the tissue adjacent the electrode located on the outer surface 66 of the spline 60. For example, when the electrodes 74a, 74b of a bipolar electrode pair 72 are located on opposite surfaces of a spline 60, they may sense nearly the identical far field signal such that when an activation signal sensed by the first electrode is subtracted from an activation signal generated by the second electrode of the bipolar electrode pair, any noise or other signal pollution resulting from the far field signal is removed from the resulting bipolar electrogram. In addition, any differences between the two activation signals sensed by the first and second electrodes 74a, 74b of the bipolar electrode pair may be emphasized such as when the electrode located on the outer surface 66 of the spline 60 is in contact with heart tissue. The remaining signal after subtraction may be indicative of the local electrical activation directly adjacent the electrode in contact with the heart tissue. Because the signal may be more localized, this may increase the spatial responsivity of the electrode system which includes multiple bipolar electrode pairs including a first electrode located on an outer surface opposite a second electrode located on an inner surface of spline which are used to sense multiple activation signals in a similar manner to produce a map of the electrical activity of the patient's heart. The improved output signal indicative of the sensed electrical activity generated by the bipolar electrode pair may, in turn, produce an improved bipolar electrogram, and may provide a better representation of the electrodes' location in a three-dimensional space for mapping the electrical activity of the patient's heart.

In addition, because each of the bipolar electrode pairs may be orientated substantially perpendicular to the direction of wavefront propagation which may reduce the sensitivity of the bipolar signal to the direction of wavefront propagation. Additionally, the perpendicular orientation to the direction of wavefront propagation may cause the electrode located on the outer surface of the spline to behave in a unipolar fashion with the electrode located on the inner surface serving substantially as a reference electrode. This phenomenon may be true for each bipolar electrode pair 72 located along the length of the spline 60. Finally, because each of the electrodes of the bipolar electrode pairs 72 are substantially co-located in space, this may result in an improved spatial response pattern to the intrinsic electrical

activity of the patient's heart resulting in a more accurate representation (map) of the electrical activity of the patient's heart.

In use, according to some embodiments, bipolar electrograms measured using electrodes **74a**, **74b** on opposing surfaces **66**, **68** of each of the splines **60** can also be used for detecting contact with viable tissue. When the tissue nearest the electrode **74a** located on the outer surface **66** is not electrically active or is located far away from the electrode **74a**, the signals from the two opposing electrodes **74a**, **74b** may be substantially identical, and the bipolar electrogram generated by the bipolar electrode pair **72** will have a small amplitude. When tissue is nearest the electrode **74a** located on the outer surface **66** and is electrically active, the rapid spatial decay of the local activation will result in a bipolar signal having a large amplitude.

In some embodiments, impedance measurements can also be used to detect and/or confirm tissue contact. The impedance measured separately through opposing electrodes of a bipolar electrode pair **72** may be different because of the medium through which the signal must travel to reach each sensing electrode of the bipolar electrode pair **72**. For example, electrode **74a**, located on the outer surface **66**, may exhibit a greater impedance than electrode **74b**, located on the inner surface **68**, which may be in contact with mostly blood. A bipolar impedance measurement between the two electrodes may therefore rise sharply with tissue contact, providing a clear and localized indication of tissue contact.

In still other embodiments, to detect tissue proximity rather than just contact, opposing electrodes can be used for a four wire (or more) impedance measurement, with a first pair of electrodes driving current and a second pair of electrodes measuring voltage. For example, current can be driven between at least a first adjacent pair of interior electrodes, and a voltage drop can be measured across a second adjacent pair of exterior electrodes. Tissue proximity will increase the measured voltage due to the lower conductivity of tissue relative to blood. In another example, a current can be driven between a first pair of opposing electrodes and a voltage can be measured using one or more nearby opposing pair of electrodes. The impedance measurements may be repeated for different combinations of electrode pairs. These are just some examples.

The impedance measurements obtained according to the different methods as described herein may be used to determine a distance between the electrode assembly **30** and the heart tissue. Additionally or alternatively, the impedance measurements may be used to weight the different activation times obtained from the bipolar electrograms, and may be used to indicate which activation times correspond to good tissue contact. The impedance measurements also may be used to characterize the tissue being contacted by the different electrodes such as, for example, to confirm whether or not a tissue area of interest has been successfully ablated using the ablation catheter **18**.

FIGS. **6-9D** relate to a method of forming an electrode assembly **30** as described herein. In many cases, the expandable electrode assembly **30**, including the flexible splines **60** having multiple electrodes located on the outer and inner surfaces **66**, **68** may be constructed from a multi-layered flexible sheet **80** including a first flexible printed circuit **82** adhesively bonded to an upper surface of a substrate and a second flexible printed circuit **84** bonded to a lower surface of the same substrate. The first and second flexible printed circuits **82**, **84** define the outer and inner surfaces **66**, **68** of each of the splines **60**. FIG. **6** is a cross-sectional view of an exemplary multi-layered flexible sheet **80** including the first

and second flexible printed circuits **82**, **84** from which an expandable electrode assembly **30**, as described herein, may be constructed. The total thickness **L** of the flexible sheet **80**, including the two flexible printed circuits **82**, **84**, is less than about 0.4 mm and defines the thickness of each of the individual splines **60**. It will be generally understood that the flexible sheet **80**, including the two flexible printed circuits **82**, **84**, may be fabricated using a variety of suitable known techniques for producing circuits, including flexible printed circuits, having multiple layers. These techniques may include but are not limited to: laminating, masking, photolithography, etching, plating, sputtering, vapor deposition, and/or the like. Similar techniques or combination of techniques may be used to fabricate a single, dual sided flexible printed circuit as is also described herein.

In some cases, the multi-layered flexible sheet **80** may include a relatively stiff substrate **86**. The substrate **86** may be constructed from Nitinol or some other stiff material such as a polyimide or polyether ether ketone (PEEK) that may facilitate shape retention of the electrode assembly **30**. Alternative materials such as, for example, a compliant material may be used to form the substrate **86** to obtain the desired mechanical characteristics. A first flexible printed circuit **82** defining at least a first electrode may be formed on the upper surface of the substrate **86** and a second flexible printed circuit **84** defining at least a second electrode may be formed on the lower surface of the substrate. In forming each of the first and second flexible printed circuits **82**, **84**, first metallization layer **88a**, **88b** may be bonded to the upper and lower surfaces of the substrate **86** using an adhesive layer **92a**, **92b**. An insulating layer **96a**, **96b** may be deposited over the first metallization layer **88a**, **88b**. A second metallization layer **102a**, **102b** may be formed over the insulating layer **96a**, **96b**. A connection can be formed by constructing a via between the two metallization layers **88a** and **102a** and **88b**, **102b**. A via can be formed by creating a hole through both metallization layers **88**, **102** and the insulating layer **96** and then plating the walls of the hole between the two metallization layers **88**, **102**, to form a metal connection **108a**, **108b**. A topcoat layer **110a**, **110b** may then be provided over the outer metallization layers **102a**, **102b**. The topcoat layer **110a**, **110b** serves to insulate portions of the outer metallization layer **102a**, **102b** from external contact. Portions of the topcoat layer **110a**, **110b** may be removed at selected locations along the flexible printed circuits **82**, **84** and an additional metal layer **112a**, **112b** may be sputter-coated or plated onto the exposed portion of the outer metallization layers **102a**, **102b** to form the electrodes **64**.

The material used to form the electrodes **64** may be selected to reduce impedance of the electrochemical interface between the electrode **64** and blood. Reducing impedance may reduce overall system noise. Exemplary electrode materials include, but are not limited to gold, stainless steel, platinum, platinum-iridium, titanium nitride, platinum black or iridium oxide. In some cases, the electrodes **64** may be fabricated from a gold metal layer coated with iridium oxide.

In some cases, each flexible spline **60** used to form an expandable electrode assembly **30**, such as described herein, may be formed from an individual flexible sheet having a first flexible printed circuit formed on an upper surface of substrate and a second flexible printed circuit formed on a lower surface of a substrate. The individual flexible splines **60** may be mechanically joined together to form an expandable electrode assembly **30**, as described herein.

Alternatively, the flexible splines **60** may be formed from a single, dual-sided flexible printed circuit having an upper

surface including a first electrode of at least one bipolar electrode pair formed therein and a lower surface including a second electrode of the bipolar electrode pair formed therein. It will be generally understood that the single, dual-sided flexible printed circuit may include multiple electrodes formed in the upper and lower surfaces thereof. The electrodes located on opposite surfaces of the splines may define multiple bipolar electrode pairs having a first electrode located on the upper surface of the spline and a second electrode located opposite the first electrode on the lower surface. In some cases, the single, dual-sided flexible printed circuit may include a stiffened core layer. The stiffened core layer may incorporate a shape memory material (e.g. Nitinol) or some other stiff material such as a polyimide or polyether ether ketone (PEEK) that may facilitate shape retention of the electrode assembly **30**. Alternative materials such as, for example, a compliant material may be used to form the core layer to obtain the desired mechanical characteristics.

In other cases, the flexible members or splines can also be manufactured from a single, multilayered flexible printed circuit laminated to a mechanical stiffener. FIGS. **7A** and **7B** show schematic views of the inner and outer surfaces **166**, **168** of a portion of a flexible spline **160** incorporating a single flexible printed circuit and a mechanical stiffener **172**. The flexible printed circuit may have electrodes **164** formed on both sides of a substrate (e.g. polyimide or PEEK), and may contain multiple conductive layers (e.g. copper or gold or another suitable conductive metal or metal alloy) to route the signal lines from each electrode **160** along the spline **160**. A mechanical stiffener **172**, fabricated from a suitable metal or plastic material, may be laminated to one side of the flexible printed circuit. The stiffener **172** may feature openings **176** that correspond with the inside electrodes of the flexible circuit, allowing the electrodes **164** to be exposed to the blood pool in vivo and acquire electrical signals.

FIG. **8A** shows a top plan view and FIG. **8B** is a bottom plan view of a multi-layered flexible sheet used to construct an electrode assembly. FIGS. **9A-9D** provide a stepwise illustration of an exemplary method of constructing the expandable electrode assembly from a flexible sheet **80** including two flexible printed circuits **82**, **84**, as described herein. Starting with a planar, flexible sheet **80**, as shown in FIG. **9A**, a series of apertures **124** may be formed in a distal region of the flexible sheet **80** using laser cutting, die cutting, chemical etching, or another precision cutting means. Together the plurality of apertures **124** forms a bonding region **116**. A termination section **120** may be formed at the opposite end of the flexible sheet **80**. The termination section **120** may be used to bond the electrode assembly **30** to the catheter body. Next, as shown in FIG. **9B**, the planar flexible sheet **80** may be wound around a major axis **130**, bringing first edge **132** toward a second edge **134** of the flexible sheet **80**. FIG. **9C** shows the two edges **132**, **134** juxtaposed with both edges **132**, **134** fixed to define a generally cylindrical structure. In some cases, the edges of the distal bonding region **116** may be secured by encapsulating the distal bonding region **116** with an adhesive, and the edges of termination section **120** may be secured by anchoring or bonding it to a distal portion of the catheter body. In some cases, as shown in FIGS. **9A-9D**, the termination band may include first and second tabs **136**, **138** which may be inserted into a lumen or slot provided in the catheter body and then secured within and adhesive or other suitable bonding technique, and which may include electrical connections to route the electrode signals to the processing system **20** (FIG. **1**). FIG. **9D** shows the completed electrode assembly **30**.

In some cases, the flexible splines **60** may be fully separated from one another such that they are not connected. The distal ends of each of the flexible splines may be mechanically joined together using a band, ring or cap provided for that purpose. In one example, each of the distal ends of the separated splines may be inserted into a corresponding slot provided in a distal cap. The distal ends of the separated splines may be mechanically interlocked with the cap by a locking feature formed at the distal end of the spline. The cap may form an atraumatic tip of the electrode assembly. The proximal ends of the separated splines may be anchored or bonded to a distal end of a catheter body. In some cases, the proximal ends of the separated splines may be first joined together using a band or ring before anchoring or bonding the splines to the distal end of catheter body using an adhesive or potting material.

Those skilled in the art will recognize that the present disclosure may be manifested in a variety of forms other than the specific embodiments described and contemplated herein. Accordingly, departure in form and detail may be made without departing from the scope and spirit of the present disclosure as described in the appended claims.

What is claimed is:

1. A catheter comprising:

an elongate catheter body extending from a proximal end to a distal end; and

an expandable electrode assembly disposed at the distal end of the catheter body, the electrode assembly configured to transition from a collapsed configuration to an expanded configuration and comprising at least one flexible member having an outer surface and an inner surface, wherein the at least one flexible member comprises a first electrode disposed on the outer surface of the flexible member and a second electrode disposed on the inner surface of the flexible member, wherein the first electrode and the second electrode are configured to form a bipolar electrode pair and wherein a distance between the first electrode and the second electrode is less than 1.5 mm.

2. The catheter of claim 1, wherein the first electrode is located directly opposite the second electrode.

3. The catheter of claim 1, wherein the first electrode is offset from the second electrode.

4. The catheter of claim 1, further comprising two or more flexible members, each of the two or more flexible members comprising at least a first electrode disposed on the outer surface of the flexible member and at least a second electrode disposed on the inner surface of the flexible member.

5. The catheter of claim 1, wherein the flexible member comprises at least one flexible printed circuit.

6. The catheter of claim 5, wherein the flexible member comprises a single, dual sided flexible printed circuit wherein the first electrode is formed on an outer surface of the flexible printed circuit and the second electrode is formed on an inner surface of the flexible printed circuit.

7. The catheter of claim 1, wherein the distance between the first electrode and the second electrode is less than about 0.5 mm.

8. The catheter of claim 1, wherein the flexible member comprises multiple bipolar electrode pairs defined by a first electrode disposed on the outer surface of a flexible member and a second electrode disposed on the inner surface of the flexible member.

9. A catheter comprising:

an elongate catheter body extending from a proximal end to a distal end;

17

an expandable electrode assembly disposed at the distal end of the catheter body, the electrode assembly configured to transition from a collapsed configuration to an expanded configuration and comprising two or more flexible splines having an outer surface and an inner surface, wherein at least one of the two or more flexible splines comprises at least a first electrode disposed on the outer surface of the flexible spline and at least a second electrode disposed on the inner surface of the flexible spline, wherein the first electrode and the second electrode are configured to form a bipolar electrode pair and wherein a distance between the first electrode and the second electrode is less than 1.5 mm.

10. The catheter of claim 9, wherein the first electrode is located directly opposite the second electrode.

11. The catheter of claim 9, wherein the first electrode is offset from the second electrode.

12. The catheter of claim 9, wherein each of the two or more splines comprises multiple bipolar electrode pairs defined by a first electrode disposed on the outer surface of a flexible spline and a second electrode disposed on the inner surface of the flexible spline.

13. The catheter of claim 9, wherein each of the two or more flexible splines comprises at least one flexible printed circuit.

14. The catheter of claim 13, wherein the at least one flexible circuit is a single, dual sided flexible printed circuit having a first electrode is formed on an upper surface of the flexible printed circuit and a second electrode formed on a lower surface of the flexible printed circuit.

15. The catheter of claim 13, wherein each of the two or more flexible splines comprises a first flexible printed circuit defining a first electrode formed on an upper surface of a substrate and a second flexible printed circuit defining a second electrode formed on a lower surface of the substrate.

18

16. A method of forming a flexible electrode assembly, the method comprising:

forming a flexible electrode assembly comprising at least one flexible member having an outer surface and an inner surface, wherein the at least one flexible member comprises a first electrode disposed on the outer surface of the flexible member and a second electrode disposed on the inner surface of the flexible member and wherein the flexible electrode assembly is configured to transition from a collapsed configuration to an expanded configuration, wherein the first electrode and the second electrode are configured to form a bipolar electrode pair and wherein a distance between the first electrode and the second electrode is less than 1.5 mm; and coupling the flexible electrode assembly to a distal end of an elongate catheter body.

17. The method of claim 16, further comprising:

forming a flexible layered sheet comprising at least one flexible printed circuit defining a first electrode on an outer surface of the flexible layered sheet and a second electrode on an inner surface of the flexible layered sheet;

separating the flexible layered sheet into two or more flexible members, each flexible member having a first electrode located on an outer surface and a second electrode located on an inner surface; and

forming the expandable electrode assembly from at least one of the flexible members.

18. The method of claim 16, further comprising forming an expandable electrode assembly from two or more flexible members by joining the two or more flexible members together at least at a first end of each of the two or more flexible members.

* * * * *

专利名称(译)	电极组装		
公开(公告)号	US9848795	公开(公告)日	2017-12-26
申请号	US14/723408	申请日	2015-05-27
[标]申请(专利权)人(译)	波士顿科学西美德公司		
申请(专利权)人(译)	BOSTON SCIENTIFIC SCIMED , INC.		
当前申请(专利权)人(译)	BOSTON SCIENTIFIC SCIMED INC.		
[标]发明人	MARECKI ANDREW T HULTZ PAUL KOZLOWSKI MICHAEL C STEWART BRIAN		
发明人	MARECKI, ANDREW T. HULTZ, PAUL KOZLOWSKI, MICHAEL C. STEWART, BRIAN		
IPC分类号	A61B5/042 A61B18/14 A61B5/00 A61B18/00 A61B17/00		
CPC分类号	A61B5/0422 A61B5/6858 A61B5/6859 A61B18/1492 A61B5/6856 Y10T29/49126 A61B5/6855 A61B5/6857 A61B2017/00044 A61B2018/0016 A61B2018/00267 A61B2562/0209 A61B2562/125 A61B2562/164 A61B2562/166 Y10T29/49119		
代理机构(译)	FAEGRE BAKER丹尼尔斯律师事务所		
优先权	62/007716 2014-06-04 US		
其他公开文献	US20150351652A1		
外部链接	Espacenet USPTO		

摘要(译)

一种用于心脏标测过程的可扩展电极组件，包括多个双极电极对，包括位于外表面上的第一电极和位于形成可扩展电极组件的各个花键的内表面上的第二电极。这种电极布置可以产生改进的电激活信号，其可以用于产生患者心脏的电活动的更准确的图。

