




US 20180103839A1

(19) **United States**(12) **Patent Application Publication**
Kurtz et al.(10) **Pub. No.: US 2018/0103839 A1**(43) **Pub. Date: Apr. 19, 2018**(54) **NETWORKED SYSTEM OF MOBILE
COMMUNICATION PLATFORMS FOR
NONPHARMACOLOGIC CONSTRICTION
OF A PUPIL**(71) Applicants: **Ronald Michael Kurtz**, Irvine, CA
(US); **Gergely T. Zimanyi**, Berkeley,
CA (US)(72) Inventors: **Ronald Michael Kurtz**, Irvine, CA
(US); **Gergely T. Zimanyi**, Berkeley,
CA (US)(21) Appl. No.: **15/375,149**(22) Filed: **Dec. 11, 2016****Related U.S. Application Data**(63) Continuation-in-part of application No. 15/293,269,
filed on Oct. 13, 2016.**Publication Classification**(51) **Int. Cl.**
A61B 3/00 (2006.01)
A61B 3/14 (2006.01)
A61N 5/06 (2006.01)
A61B 5/00 (2006.01)(52) **U.S. Cl.**
CPC **A61B 3/0008** (2013.01); **A61B 3/14**
(2013.01); **A61B 3/0025** (2013.01); **A61N**
5/0625 (2013.01); **A61B 2576/00** (2013.01);
A61B 5/0022 (2013.01); **A61B 5/6898**
(2013.01); **A61B 2560/0431** (2013.01); **A61B**
5/0013 (2013.01)(57) **ABSTRACT**

A networked system of ophthalmic stimulators for temporarily constricting eye-pupils comprises a set of ophthalmic stimulators, each ophthalmic stimulator including a mobile irradiation control system, to generate an irradiation control signal; an irradiation source, coupled to the irradiation control system, to generate an irradiation; and an irradiation delivery system, coupled to the mobile irradiation control system, to receive the irradiation from the irradiation source, and to deliver a patterned irradiation to an iris of the eye; wherein the mobile irradiation control system controls at least one of the irradiation source and the irradiation delivery system with the irradiation control signal so that the patterned irradiation causes a temporary constriction of the pupil of the eye, without causing a permanent constriction of the pupil; and a central station, including a central image processor, wherein the mobile irradiation control systems and the central station are configured to communicate through a communication network.






FIG. 1




FIG. 2A




FIG. 2B




FIG. 3A




FIG. 3B




FIG. 4A






FIG. 4B






FIG. 6A




FIG. 6B




FIG. 6C




FIG. 6D




FIG. 7A




FIG. 7B




FIG. 7C




FIG. 7D




FIG. 7E




FIG. 7F




FIG. 8A




FIG. 8B




FIG. 9A




FIG. 9B




FIG. 9C




FIG. 9D




FIG. 9E




FIG. 10




FIG. 11A




FIG. 11B




FIG. 11C




FIG. 11D




FIG. 12A




FIG. 12B




FIG. 12C




FIG. 12D




FIG. 12E




FIG. 13A




FIG. 13C




FIG. 13B




FIG. 14




FIG. 15





FIG. 16A






FIG. 16C




FIG. 16D




FIG. 16E




FIG. 17B

FIG. 17A




FIG. 17C




FIG. 17D

**NETWORKED SYSTEM OF MOBILE
COMMUNICATION PLATFORMS FOR
NONPHARMACOLOGIC CONSTRICTION
OF A PUPIL**

CROSS-REFERENCE TO RELATED PATENT
DOCUMENTS

[0001] This application is a continuation-in-part of U.S. Patent Application: "System for temporary nonpharmacologic constriction of the pupil", by Ronald M. Kurtz and Gergely T. Zimanyi, with application Ser. No. 15/293,269, filed on Oct. 13, 2016; hereby incorporated by reference in its entirety.

FIELD OF INVENTION

[0002] This invention relates to a system for pupil constriction, more precisely, to a system of temporary, non-pharmacological construction of a pupil of an eye.

BACKGROUND

[0003] A number of devices that make use of the increased depth of field of a small aperture have been proposed for use in ophthalmology, and developed to improve vision. These devices are particularly promising to improve near vision for those who have presbyopia. Examples of such devices include small aperture corneal inlays, reduced-aperture intraocular lenses, as well as other aperture implants that are meant to impact light propagation along the visual axis. While effective, these surgically implanted permanent inlays carry the risk inherent with any implantable device, such as inflammation, infection, or displacement that may require secondary surgical procedures to remove the implant and may necessitate performing other procedures.

[0004] Pharmacological methods have also been proposed using medications such as pilocarpine and other agents to temporarily constrict the pupil. While these drugs can temporarily improve vision, they generally require frequent instillation of drops, and can be associated with undesirable side effects, such as headaches.

[0005] An alternative approach has been proposed by Hickenbotham in US patent application 2013/0226161, which utilizes a laser to cauterize certain portions of the iris to cause a permanent constriction of the pupil. While this approach offers some advantages over implants and medications, the permanent constriction of the pupil, achieved by a controlled damaging of the iris dilator muscle, does not allow for a trial of the effect, and once performed, leaves the patient with a permanent deficit in iris function. In addition, the exact shape of the constricted pupil may be difficult to control, and may result in odd, irregular, oval, or other undesired pupil shapes. Therefore, the medical need persists to develop a non-pharmacological, non-permanent vision improvement that does not involve inserting a small-aperture object surgically into the eye.

[0006] In some embodiments, an ophthalmic stimulator for temporarily constricting a pupil of an eye comprises an irradiation control system, to generate an irradiation control signal; an irradiation source, coupled to the irradiation control system, to generate an irradiation; and an irradiation delivery system, coupled to the irradiation control system, to receive the irradiation from the irradiation source, and to deliver a patterned irradiation to an iris of the eye; wherein the irradiation control system controls at least one of the

irradiation source and the irradiation delivery system with the irradiation control signal so that the patterned irradiation causes a temporary constriction of the pupil of the eye, without causing a permanent constriction of the pupil.

[0007] In some embodiments, a method for temporarily constricting a pupil of an eye by an ophthalmic stimulator comprises generating an irradiation control signal by an irradiation control system; generating an irradiation by an irradiation source, coupled to the irradiation control system; receiving the irradiation and delivering a patterned irradiation to an iris of the eye with an irradiation delivery system; and controlling at least one of the irradiation source and the irradiation delivery system by the irradiation control signal of the irradiation control system so that the patterned irradiation is causing a temporary constriction of the pupil of the eye, without causing a permanent constriction of the pupil.

[0008] In some embodiments, an ophthalmic stimulator for constricting a pupil of an eye comprises an irradiation control system, to generate an irradiation control signal; an irradiation source, coupled to the irradiation control system, to generate an irradiation; and an irradiation delivery system, coupled to the irradiation control system, to receive the irradiation from the irradiation source, and to deliver a patterned irradiation to an iris of the eye; wherein the irradiation control system controls at least one of the irradiation source and the irradiation delivery system with the irradiation control signal so that the patterned irradiation causes a long-term constriction of the pupil of the eye.

[0009] In some embodiments, an ophthalmic stimulator for temporarily constricting a pupil of an eye comprises an irradiation control system, to generate an irradiation control signal; a light source, coupled to the irradiation control system, to generate a light beam; and a beam-shaping optics, coupled to the irradiation control system, to receive the light beam from the light source, and to deliver a light ring to an iris of the eye; wherein the irradiation control system controls at least one of the light source and the beam-shaping optics with the irradiation control signal so that the light ring causes a temporary constriction of the pupil of the eye, without causing a permanent constriction of the pupil.

[0010] In some embodiments, an ophthalmic stimulator for temporarily constricting a pupil of an eye comprises a digital beam controller, to generate a digital beam-control signal; a light source, coupled to the beam controller, to generate a light beam; and a digitally controlled beam modulator, to receive the digital beam-control signal from the beam controller, to receive the light beam from the light source, and to modulate the received light beam into a patterned light, delivered to an iris of the eye; wherein the beam controller controls at least one of the light source and the digitally controlled beam modulator with the digital beam-control signal so that the patterned light causes a temporary constriction of the pupil of the eye, without causing a permanent constriction of the pupil.

[0011] In some embodiments, an ophthalmic stimulator for temporarily constricting a pupil of an eye comprises an irradiation control system, having a feedback system, to generate an irradiation control signal using a feedback of the feedback system; an irradiation source, coupled to the irradiation control system, to generate an irradiation; and an irradiation delivery system, having a targeting system and coupled to the irradiation control system, to receive the irradiation from the irradiation source, and to direct a patterned irradiation in a pattern to a treatment region of an

iris of the eye using the targeting system; wherein the irradiation control system controls at least one of the irradiation source and the irradiation delivery system with the irradiation control signal so that the patterned irradiation causes a temporary constriction of the pupil, without causing a permanent constriction of the pupil.

[0012] In some embodiments, an ophthalmic stimulator for temporarily constricting a pupil of an eye comprises a mobile irradiation control system, to generate an irradiation control signal; an irradiation source, coupled to the irradiation control system, to generate an irradiation; and an irradiation delivery system, coupled to the mobile irradiation control system, to receive the irradiation from the irradiation source, and to deliver a patterned irradiation to an iris of the eye; wherein the mobile irradiation control system controls at least one of the irradiation source and the irradiation delivery system with the irradiation control signal so that the patterned irradiation causes a temporary constriction of the pupil of the eye, without causing a permanent constriction of the pupil.

[0013] In some embodiments, a networked system of ophthalmic stimulators for temporarily constricting eye-pupils comprises a set of ophthalmic stimulators, each ophthalmic stimulator including a mobile irradiation control system, to generate an irradiation control signal; an irradiation source, coupled to the irradiation control system, to generate an irradiation; and an irradiation delivery system, coupled to the mobile irradiation control system, to receive the irradiation from the irradiation source, and to deliver a patterned irradiation to an iris of the eye; wherein the mobile irradiation control system controls at least one of the irradiation source and the irradiation delivery system with the irradiation control signal so that the patterned irradiation causes a temporary constriction of the pupil of the eye, without causing a permanent constriction of the pupil; and a central station, including a central image processor, wherein the mobile irradiation control systems of the ophthalmic stimulators and the central station are configured to communicate through a communication network.

BRIEF DESCRIPTION OF THE FIGURES

[0014] FIG. 1 illustrates an eye 1.

[0015] FIGS. 2A-B illustrate the pupil under different illuminations.

[0016] FIGS. 3A-B illustrate an effect of applying an irradiation to the iris.

[0017] FIGS. 4A-B illustrate the effect of irradiation on the muscle response.

[0018] FIGS. 5A-B illustrate embodiments of the ophthalmic stimulator 100, and the permanent ophthalmic stimulator 100'.

[0019] FIGS. 6A-D illustrate embodiments of the ophthalmic stimulator 100.

[0020] FIGS. 7A-F illustrate embodiments of the ophthalmic stimulator 100 with a beam shaping optics 134.

[0021] FIGS. 8A-B illustrate embodiments of the ophthalmic stimulator 100 with a digital beam controller 110.

[0022] FIGS. 9A-E illustrate embodiments of the beam modulator 134.

[0023] FIG. 10 illustrates an irradiation controller 112.

[0024] FIGS. 11A-D illustrate steps of the methods 300, 300', 302, and 304.

[0025] FIGS. 12A-E illustrate embodiments of the alignment system 135.

[0026] FIGS. 13A-C illustrate mobile embodiment of the ophthalmic stimulator 100.

[0027] FIG. 14 illustrates a mobile network of self-treatment stimulators 400.

[0028] FIG. 15 illustrates an embodiment of the feedback system, 116.

[0029] FIGS. 16A-E illustrate methods 510-550.

[0030] FIGS. 17A-D illustrate various irradiation patterns 210.

DETAILED DESCRIPTION

[0031] Embodiments of the invention address the above described needs in the following manner. Some embodiments provide systems and methods for a temporary constriction of the pupil without the need of medication therapy. The duration of the constriction can be controlled by a selection of treatment parameters. In a suitable range of treatment parameters, the procedure can be fully reversible: after a characteristic time, the pupils return to essentially their original diameter without further treatment. The pupils can be re-constricted by applying the treatment repeatedly. Therefore, the here-described methods and devices provide the advantages of a temporary, but long lasting vision improvement, while avoid the hazards associated with (1) aperture implants and inlays, inserted by a surgical procedure, (2) permanent destruction of tissue, and (3) pharmaceutical approaches and their undesirable side-effects.

[0032] Some embodiments achieve these advantages by heating the iris by an irradiation to a suitable temperature range, (1) to cause a temporary inactivation of the iris dilator muscle, and, in some cases, (2) to enhance an action of the iris constrictor sphincter muscle. This irradiative heat treatment can be applied for a time sufficiently long to cause a reduction in contractile activity, but short enough to avoid causing permanent tissue damage. While the detailed mode of action is yet to be clarified, this effect may be mediated by inactivation of the actin-myosin complex in the exposed muscle.

[0033] FIG. 1 shows a cross section of an eye 1. The eye 1 includes the well known constituents: a cornea 5, where light enters the eye 1, and a sclera 7, an opaque, fibrous protective outer layer of the eye 1 that contains collagen and elastic fibers. Distal to the cornea 5 is an anterior chamber 9 that contains an aqueous humor. The anterior chamber 9 is separated from a posterior chamber 15 by an iris 11. An opening at a center of the iris 11 is a pupil 13 that allows the light to proceed toward the posterior segment of the eye 1. Behind the pupil 13, ciliary muscles 17 hold a lens 19 in a central position. These ciliary muscles 17 can also delimit the lens 19 as part of accommodating the vision to the distance of the target the eye is looking at. With advancing age, the ciliary muscles 17 slowly lose their ability to deform and adapt the lens 19 to varying vision distances: a condition typically referred to as presbyopia. Behind the lens 19 is a vitreous 21. As the light crosses the vitreous 21, it eventually hits the retina 23. The electric stimuli, generated by the incoming light in the retina 23, are transmitted by the optic nerve 25 towards the brain.

[0034] FIG. 2A-B illustrate that the iris 11 includes a circular sphincter muscle 40 around the pupil 13, capable of shrinking the perimeter of the pupil 13, thus constricting it. At the same time, the iris 11 also includes radial dilator muscles 30 that specialize in expanding, or enlarging, the pupil 13. The competition of the sphincter muscles 40 and

dilator muscles **30** determines the eventual radius of the pupil **13**. FIG. 2A illustrates in its left panel that in strong light the contracting sphincter muscles **40** constrict the pupil **13**. FIG. 2A illustrates in its middle panel the pupil **13** in an average light. FIG. 2A illustrates in its right panel that in low light conditions, the radial dilator muscles **30** dominate the sphincter muscles **40** and dilate the pupil to enhance the amount of light directed to the retina **23**.

[0035] FIG. 2B illustrates a cross section of the iris **11** from the side. It is well visible that the sphincter pupillae **40** is positioned along the edge of the pupil **13**, the pupillary ruff while the radial dilator pupillae **30** are located radially outward, farther from the edge of the pupil **13**.

[0036] The anatomy of the muscles of the iris **11** is also important. The dilator muscle **30** fibers are typically located near the distal portion of the iris **11**, adjacent to the iris pigmented epithelium. In contrast, the constrictor sphincter muscles **40** are more superficial and central, located towards the pupil's edge or margin. Details of the anatomy of these muscles can be found in much greater detail in Junqueira L. C., Carneiro J. 2005. Basic Histology, Eleventh Edition. The McGraw-Hill Companies, Inc. United States of America.

[0037] FIGS. 3A-B illustrate a principle of embodiments of the invention. FIG. 3A illustrates that a patterned irradiation can be applied to the iris **11** for a limited time period, such as 1-100 seconds, with less time required when higher temperatures are applied. The pattern is typically a ring of light, or light-ring. The irradiation raises the temperature of the iris **11** in a treatment region. The tissue of the iris **11** can be heated to temperatures that are not sufficient to cauterize or destroy the tissue, but are capable of reducing an activity, or responsiveness of the targeted tissues.

[0038] FIG. 3B illustrates the outcome of the irradiation. The heat treatment reduces the activity of the iris dilator muscle and this allows the pupillary constrictor, or sphincter, muscle to reduce the pupil's diameter. Reducing the pupil's diameter reduces the aberrations of the imaging of the eye, sometimes referred to as the pinhole effect in optics. Reducing the aberrations extends the depth of focus, and thereby compensates the emergence of presbyopia in an aging eye. Since this method utilizes the natural constrictor muscle to effect the pupil size change, the risk of pupil de-centration is less than in the case of surgical implants, discussed previously.

[0039] FIGS. 4A-B illustrate that heat treatments have been already studied and demonstrated to reduce muscle activity in human tissues, such as in the lung and the prostate, which have smooth muscle tissues similar to that of the iris. The heat treatment can reduce, or inhibit, muscle activity in these tissues. The duration of inactivity can last for hours to days in these systems (see Am. J. Respir. Cell Mol. Biol. Vol 44. pp 213-221, 2011). FIG. 4A illustrates the effect of heat treatments on lung smooth muscle. The muscle tissue was heated for a treatment time between 5 s and 60 s. After the heat treatment, a test stimulus was administered to the heat-treated and the untreated muscles. The graph reports the ratio of responses to this test as a function of the treatment temperature of the tissue. Visibly, as the treatment temperature exceeded 50 Celsius, or Centigrade, the response of the treated muscle to the test stimulus gradually decreased. For heat treatments above 55-60 Celsius, the response became negligible: the muscle was deactivated by the treatment.

[0040] FIG. 4B illustrates the same ratio of responses of treated muscles to non-treated muscles, with the difference that it indicates how long the effect lasted. As the curves show, the de-activation of the smooth muscle with heat treatments raising the muscle temperature above 50-55 Celsius lasted at least for 28 hours, and possibly longer. This remarkably long-lasting deactivation of smooth muscle in response to such a mild and short temperature increase is utilized by embodiments described in this document.

[0041] FIG. 5A illustrates an ophthalmic stimulator **100** for temporarily constricting a pupil **13** of an eye **1**, building on the just-described observations, comprising an irradiation control system **110**, to generate an irradiation control signal; an irradiation source **120**, coupled to the irradiation control system **110**, to generate an irradiation **200**; and an irradiation delivery system **130**, coupled to the irradiation control system **110**, to receive the irradiation **200** from the irradiation source **120**, and to deliver a patterned irradiation **200p** to an iris **11** of the eye **1** in a pattern **210**. In embodiments, the irradiation control system **110** controls at least one of the irradiation source **120** and the irradiation delivery system **130** with the irradiation control signal so that the patterned irradiation **200p** causes a temporary constriction of the pupil **13** of the eye **1**, without causing a permanent constriction of the pupil **13**.

[0042] The irradiation control system **110** can include a memory, to store executable programs and applications; a processor, to execute at least one of a stored program and an installed application; and a user interface, to receive input from a user in relation to an operation of the memory and the processor.

[0043] In some embodiments of the ophthalmic stimulator **100**, the irradiation source **120** can include an incoherent light source, such as a light source, a LED, a lamp, an infrared source, a broad-band source, a narrow-band source, a radio-frequency source, an electromagnetic radiation source, or a sound source, to generate a light beam, an electromagnetic irradiation, an infrared beam, a LED light, or a sound. A separate class of irradiation sources can include a coherent light source, such as a laser, a pulsed laser, or a continuous wave (CW) laser.

[0044] The just discussed classes of incoherent and coherent irradiation sources have different advantages and drawbacks. Lasers offer good control and unparalleled precision. At the same time, laser beams have a very small diameter, often less than 100 microns. Therefore, to affect larger treatment regions, they require a complex and expensive, digitally controlled optical system, such as a scanning system. These laser-plus-scanning systems offer great control and precision. At the same time, they can be expensive, and can introduce multiple sources of unreliability and performance degradation, a potential problem in medical applications, where high reliability is essential. Using lasers and scanners may therefore necessitate regular maintenance. Also, laser beams can be very intense, thus if a laser gets pointed to an unintended part of an ophthalmic tissue, it can cause substantial damage. Therefore, much stronger safety systems and precautions are needed in laser systems.

[0045] In contrast, non-coherent light sources, such as LEDs, infrared sources, lamps, infrared sources, and others may offer less precision and control. However, this control may be sufficient for the purposes of the here-described treatment. Also, incoherent light sources can make the ophthalmic stimulator **100** much simpler, lighter, and

smaller at the same time. Since they typically do not require a digitally controlled scanning system, these incoherent light sources can also be cheaper to maintain and can be more robust and reliable. Finally, since these light sources are less intense, systems with incoherent light sources may require less stringent safety systems and measures. All in all, a comparative analysis of the competing advantages and disadvantages is performed when a system designer decides whether to use a coherent, or an incoherent light source as the irradiation source **120** of the ophthalmic stimulator **100**.

[0046] Embodiments of the ophthalmic stimulator **100** can be characterized by numerous treatment parameters. These treatment parameters can include the followings. A power density of the patterned irradiation **200p** of the irradiation delivery system **130** can be in the range of 0.1-1000 mW/cm², in some designs in the range of 1-100 mW/cm². A total power delivered by the patterned irradiation **200p** to the iris can be in the range of: 0.1-1,000 mW, in some designs in the range of 1-100 mW. A total energy, deposited by the patterned irradiation **200p** during the treatment can be in the range of 10 microJ-10 J, in some designs in the range of 100 microJ-100 mJ.

[0047] A wavelength of the irradiation source **120** can be in the range of 400-4,000 nm, in some designs, in the range of 600-1,500 nm. The wavelength of some stimulators **100** can be selected by noting in FIG. 2B, that the muscle fibers of the radial dilators **30** are located in the proximity of the pigmented epithelium of the iris **11**. This fact can be used to selectively target and heat the dilator muscles **30** indirectly. The pigmented epithelium layers may not have essential functions that would be negatively affected by heating, such as undergoing an irrecoverable reactivity change. To build on this, irradiation sources **120** can emit the irradiation **200** with a wavelength close to the wavelength where the absorption of the pigmented epithelium shows a maximum, or is at least greatly enhanced. Such irradiation sources **120** can heat the pigments particularly efficiently, possibly to temperatures 55 C, 60 C, possibly even to 60-65 Celsius. The heated pigmented epithelia can then provide a secondary, or indirect heating to the dilator muscles **30**, located in their immediate proximity, to the medically preferred 50-55 Celsius temperatures.

[0048] FIG. 2B also illustrates that the dilator muscles **30** are in the distal region of the iris **11**. Therefore, irradiation with wavelengths that penetrate the iris tissue more efficiently and to greater depths can be favored to make sure that the dilator muscles **30** are well heated. In several ophthalmologic studies, irradiation with longer wavelengths showed greater penetration into ophthalmic tissues. Therefore, some irradiation sources **120** may emit irradiation **200** with longer wavelengths to penetrate more deeply into the iris, with eventual absorption by the pigmented epithelium, to achieve secondary heating of the dilator muscle fibers **30**. Accordingly, a depth of a treated tissue within the iris can be in some designs in the range of 10 microns-3,000 microns, in some designs, in the range of 500-2,000 microns.

[0049] Some irradiation sources may emit a continuous, or continuous wave (CW) irradiation **200**. Others, such as lasers, or LEDs, may emit pulsed irradiation. A frequency of the pulsed irradiation **200** can be in a range of 1 Hz to 1 MHz, in some designs, in the range of 100 Hz to 100 kHz. The length of the emitted pulses can vary from 10 femtoseconds to 1 second, in some designs from 1 microsecond to 1 millisecond. The total treatment time can be in the range

of 1 sec to 300 sec, in some embodiments in the range of 10 sec to 100 sec. A beam profile of the patterned irradiation **200p** can be a rectangular, a flat top, a smoothed, a Gaussian, or a Lorentzian profile.

[0050] An inner radius Rp(inner) of the pattern **210** can be in the range of 2-10 mm, in some designs in the range of 3-6 mm. An outer radius Rp(outer) of the pattern **210** can be in the range of 3-15 mm, in some designs, in the range of 5-10 mm. The pattern **210** can be such that a treated fraction of the iris has an area that is 10-80% of the total area of the iris **11**, in some design, this fraction can be in the range of 20-50%.

[0051] In some embodiments, the irradiation delivery system **130** can include a pattern generator, an optical beam shaper, a patterning optics, a beam profiler, or a digitally controlled irradiation optics. Some of these elements can be built mostly from passive optical elements, such as lenses and mirrors, with some system characteristics controlled electronically, such as a telescopic distance between two lenses. In other embodiments, the irradiation delivery system **130** can include optical elements that are actively operated and controlled by electronic or digital circuitry, as described below.

[0052] Some embodiments of the ophthalmic stimulator **100** can be configured to increase a temperature of a treatment region of the iris to a range of 45-60 degrees Celsius. Other embodiments can increase the temperature of the treatment region of the iris to a range of 50-55 degrees Celsius. As discussed in relation to FIGS. 4A-B, treatments with temperatures in these ranges have been demonstrated to impact the responsiveness of smooth muscle tissue temporarily, in a reversible and repeatable manner.

[0053] The actual effect of the heat treatment depends on several factors, since different temperatures and treatment durations can have a multitude of effects on smooth muscle cells and function. On the cellular level, first, a heat treatment can induce biochemical changes and secretions that can affect the functioning of the treated tissue, such as heat shock proteins. Second, it can cause loss of cells through various mechanisms, such as apoptosis, or programmed cell death. Finally, on a much shorter time scale, heat treatment can lead to specific loss of contractility due to denaturation of myosin molecules or inhibition of ion channels.

[0054] On a higher, physiological level, the effect of the heat treatment on the pupil may depend on factors such as dilator muscle fiber orientation, and on opposing, constrictor, muscle action. Finally, the heat treatment can change the physical properties of the muscles in different aspects as well, including shrinking or expanding the length of the muscle strands, making the strands more or less aligned, and changing of the elastic moduli of the muscles, among others.

[0055] For all these reasons, the iris of the individual patients can be analyzed by the ophthalmologist before the treatment with the ophthalmic stimulator **100**. Based on the analysis, the desired medical outcomes can be cross-referenced with the patient data of the individual patients. Subsequently, the treatment region, treatment parameters and specifically the treatment temperatures can be set. As discussed further below, for some medical outcomes heating the radial dilator muscles **30** can be preferable, for others, heating the circular sphincter muscles **40** can be preferable. The treatment regions can be set according to these medical considerations.

[0056] FIGS. 6A-D illustrate that in some embodiments of the ophthalmic stimulator **100**, the irradiation control system

110 can include an irradiation controller 112, an imaging system 114 and a user interface 118. The imaging system 114 can be electronically coupled to the irradiation controller 112, to relay images, image-related data, and control information. The imaging system 114 can include an image processor 114ip, whose functions will be described later.

[0057] FIGS. 6A-B illustrate two implementations of the imaging system 114. In FIG. 6A, an imaging light 220 is reflected out from the optical pathway of the patterned irradiation 200p by a beam splitter 131 towards the imaging system 114 that is positioned outside the irradiation optical pathway. In FIG. 6B, a small imaging system, such as a small CCD camera 114 can be placed on the distal end of the irradiation delivery system 130, directly receiving the imaging light 220. The imaging light 220 can be a reflection of an imaging light, projected on the iris 11 by an imaging light source. In other designs, the imaging light 220 can be simply the ambient light reflected from the iris 11.

[0058] The imaging system 114 can be any one of the well known ophthalmic imaging systems, including a CCD camera, feeding into a video monitor, any other electronic or digital imaging system, a video microscope, or a surgical microscope.

[0059] The irradiation control system 110 can generate the irradiation control signal by generating an image of the iris 11 of the eye with the imaging system 114 for a user, followed by receiving an image-based input from the user through the user interface 118, and generating the irradiation control signal to control the irradiation delivery system 130 to deliver the patterned irradiation 200p in accordance with the received image-based input.

[0060] In a typical example, the patterned irradiation 200p can impact the iris 11 in a ring pattern 210 with an inner radius $R_{p(inner)}$ and an outer radius $R_{p(outer)}$. In this embodiment, the user of the system, such as ophthalmologist, or an ophthalmic surgeon, can be prompted via the user interface 118 to enter the image-based input, which in this case can be a selection of the inner radius $R_{p(inner)}$ and the outer radius $R_{p(outer)}$ of the ring pattern 210, based on the surgeon analyzing the image, relayed by the imaging system 114.

[0061] FIGS. 12D-E illustrate that setting these radii $R_{p(inner)}$ and $R_{p(outer)}$ determines whether the ring pattern 210, and thus the treatment region, is the region of the radial dilator muscles 30, or the circular sphincter muscles 40. Denoting the outer radius of the sphincter muscles with $R(sphincter)$, if the surgeon selects the inner radius $R_{p(inner)}$ of the ring pattern 210 to be greater than $R(sphincter)$: $R_{p(inner)} > R(sphincter)$, then the ring pattern 210 will fall on the radial dilator muscles 30, and those muscles will receive the heat treatment. Whereas, if the surgeon selects the outer radius $R_{p(outer)}$ of the ring pattern 210 to be smaller than $R(sphincter)$: $R_{p(outer)} < R(sphincter)$, then the ring pattern 210 will fill on the circular sphincter muscles 40, and the circular sphincter muscles 40 will be treated by the patterned irradiation 200p. As discussed, an ophthalmologist can select either treatment region based on a prior analysis of the patient's specific data, and the desired medical outcomes.

[0062] In some embodiments, the irradiation control system 110 can include an image processor 114ip in the imaging system 114. The image processor 114ip can be integrated with the imaging system 114, can be partially integrated, or can be a separate electronic or computational system. In

these embodiments, the irradiation control system 110 can generate the irradiation control signal by generating an image of the iris 11 with the imaging system 114 for the image processor 114ip, receiving an image-based input from the image processor 114ip, and generating the irradiation control signal to control the irradiation delivery system 130 to deliver the patterned irradiation 200p in accordance with the received image-based input.

[0063] In a representative embodiment, the patterned irradiation 200p can impact the iris 11 in a ring pattern 210 with inner and outer radii $R_{p(inner)}$ and $R_{p(outer)}$. The imaging system 114 can image the iris 11, and relay this image to the image processor 114ip, in response, the image processor 114ip can run an image recognition program, possibly including an edge-recognition software, and identify the inner and outer radii of the iris 11, and the radius $R(sphincter)$ that demarcates the radial dilator muscles 30 from the circular sphincter muscles 40. Then, the image processor 114ip can generate the image-based input that sets, or suggests to set, the $R_{p(inner)}$ and $R_{p(outer)}$ radii of the ring pattern 210. The effect of these choices on the treatment region and the corresponding medical effects have been explained earlier.

[0064] FIG. 12A-C illustrate that in some embodiments of the ophthalmic stimulator 100, the irradiation control system 110 can include an alignment system 135.

[0065] FIG. 12A illustrates that in some embodiments the ophthalmic stimulator 100 can include an objective 133, the last optical element that guides the patterned irradiation 200p toward the eye 1. In these embodiments, the alignment system 135 can include a frame, or chin-rest 136, on which the patient can rest her/his chin to minimize the motion of the eye 1 relative to the stimulator 100. The alignment system 135 can also include a patient interface 137 that contacts the eye 1 of the patient. Many types of patient interfaces 137 are known in the art and can be used here. FIG. 12A illustrates a patient interface 137, whose proximal end is attached to the objective 133 of the ophthalmic stimulator 100, and whose distal end the patient presses her eyes against. The patient interface 137 can ensure a firm coupling, or docking, to the eye by involving a vacuum suction system, or a forceps. The patient interface 137 can be a one-piece or a two-piece patient interface. The distal end of the patient interface 137 can include a contact lens, to ensure a smoother, softer connection to the eye. Such a contact lens also minimizes the optical distortions of the patterned irradiation 200p as it exits the patient interface 137 and enters the cornea 5 of the eye 1.

[0066] FIG. 12B illustrates another embodiment of the alignment system 135, where the patient interface 137 is coupled to the frame 136 instead of the stimulator 100. Since the frame 136 is rigidly coupled to the ophthalmic stimulator 100, the optical pathway of the patterned light 200p is similarly secure from the objective to the eye 1 in this embodiment as well. One of the differences is that there is a distance between the stimulator 100 and the patient interface, 137, so the patient does not have to lean forward to receive the treatment, and the doctor sees where the patterned light 200p hits the patient interface 137. As before, this patient interface 137 can also be a one-piece and a two-piece patient interface 137.

[0067] The patient interfaces 137 of either FIG. 12A or 12B is preferably aligned and centered with the eye 1 before coupling, or docking them to the eye 1. FIG. 12C illustrates

a corresponding aligning, or centering, pattern **138** of the alignment system **135**. This centering pattern, or aligning pattern, can include an aligning ring **138a**, or an aligning cross-hair **138b**, or both. This aligning pattern **138** can be formed in, projected into, or digitally overlaid, the image formed by the imaging system **114**, in a position that is concentric with the optical axis of the objective **133**. The ophthalmic surgeon, or any other user or operator, can dock the patient interface **137** of the stimulator **100** to the eye with increased precision, with aligning, or centering, the aligning element **138** with the pupil **13** during the docking procedure.

[0068] In a video-monitor-based embodiment, the surgeon can make the centering of the aligning ring **138a** on the video image with the edge of the pupil **13** part of the docking. During the docking, the surgeon can instruct the patient to move her/his head and eye around, until the circular edge of the pupil **13** is concentric with the aligning ring **138a**. Then the surgeon can complete the docking of the patient interface **137** to the eye **1**. Further embodiments of the alignment system **135** will be described later.

[0069] In some designs, the stimulator **100** can include a fixation light **202**, and the surgeon can instruct the patient to stare at the fixation light **202** during docking. The patient staring, or fixating at the fixation light **202** can further help centering the patient interface **137** with the pupil **13** during the docking.

[0070] In these embodiments, the irradiation control system **110** can generate the irradiation control signal by processing alignment data with the alignment system **135**, and generating the irradiation control signal to control the irradiation delivery system **130** to deliver the patterned irradiation **200p** to the iris in a pattern **210** aligned with the pupil **13** of the eye.

[0071] In some embodiments of the ophthalmic stimulator **100**, the processing alignment data can include generating an image of the iris **11** with the imaging system **114**, and overlaying an alignment pattern **138** on the generated image. The generating the irradiation control signal can include generating a misalignment-warning signal, or generating an alignment-guidance signal, if a misalignment is detected during the processing of the alignment data that is part of the docking. The misalignment-warning signal can alert the operating surgeon to instruct the patient to move his/her head, eye, or both to improve the alignment to help making the docking precise. Also, for stimulator designs where the stimulator **100** or the patient interface **137** itself can be moved or adjusted, the misalignment-warning signal can alert the surgeon for the need to adjust the stimulator **100** or the patient interface **137**.

[0072] An example for an adjustable patient interface **137** is a two-piece patient interface **137**, where one piece of the patient interface **137** can be attached to the stimulator **100** at its objective **133**, the other piece of the patient interface **137** can be coupled to the eye with vacuum-suction, or pressing, and the docking includes the surgeon maneuvering the two pieces of the patient interface **137** to dock to each other.

[0073] FIGS. 6C-D also show a feedback system **116**. This system will be described in detail below.

[0074] FIG. 10 illustrates that the irradiation controller **112** can include a number of blocks. These blocks can be implemented as a dedicated processor or circuitry, or can be implemented as a software, code, program, or application, implemented on a computer of the irradiation controller **112**,

or a combination of hardware and software blocks. In various embodiments, the irradiation controller **112** can include:

[0075] a feedback block **112a**, to receive feedback data and to send a feedback signal to a processor **113**;

[0076] an imaging block **112b**, to receive imaging data and to send an imaging signal to the processor **113**;

[0077] an alignment block **112c**, to receive alignment data and to send an alignment signal to the processor **113**;

[0078] a memory block **112d**, to receive patterns for storage and patient data, to store algorithms and codes, and to send stored patterns, patient data, or executable algorithms to the processor **113**;

[0079] a pattern generator block **112e**, to receive pattern parameters and to send generated patterns to the processor **113**;

[0080] a user interface block **112f**, to receive a user input, for example through a user interface **118**, that can be patterns, commands, and irradiation parameters, and to send the received patterns, commands and irradiation data as a user input signal to the processor **113**.

[0081] Each of these blocks can receive, their input from corresponding hardware blocks, such as sensors, controllers, hardware blocks and user interfaces. For example, the feedback block **112a** can be a dedicated circuitry that receives the feedback data from the feedback system **116**, as described below. The imaging block **112b** can be a software algorithm, implemented on a processor that receives the imaging data from the imaging system **114** that can include a CCD camera, a video monitor, or a surgical microscope.

[0082] In response to signals, received from any of the blocks **112a-f**, the processor **113** can send an irradiation control signal to the irradiation source **120**, or to the irradiation delivery system **130**, or to both.

[0083] In some detail, in embodiments of the ophthalmic stimulator **100** the irradiation control system **110** can include the memory **112d**, and the generating the irradiation control signal can include recalling stored data from the memory **112d**, representing at least one of an irradiation pattern and patient data, and generating the irradiation control signal to control the irradiation delivery system **130** to deliver the patterned irradiation **200p** to the iris **11** in accordance with the recalled stored data.

[0084] In embodiments, the irradiation control system **110** can include a pattern generator; and the generating the irradiation control signal can include generating an electronic representation of the irradiation pattern **210**; and generating the irradiation control signal to control the irradiation delivery system **130** to deliver the patterned irradiation **200p** with the generated irradiation pattern **210**.

[0085] Returning to the medial effects and treatments, embodiments of the ophthalmic stimulator **100** can cause a temporary constriction of the pupil **13** of the eye that includes an at least 5% reduction of a radius of the pupil **13** that lasts less than one hour. In some cases, the reduction of the radius of the pupil can last for a time interval more than one hour and less than one day. In other embodiments, the temporary constriction of the pupil of the eye includes an at least 5% reduction of the radius of the pupil that lasts for a time interval between one day and one week; or between one week and one month; or between one month and three months; or between three months and one year.

[0086] Each of these time intervals has their own medical and patient advantages. The longer the pupil constriction lasts, the less often the treatment may need to be applied, which can be preferred by patients. Also, the overall paradigm of use of the ophthalmic stimulator 100 depends on the duration of the constriction. Stimulators that constrict a pupil for a month or longer can be deployed in the offices of ophthalmologists, and patients can schedule regular visits for re-constriction treatments on a monthly basis. Stimulators that constrict the pupil for a day or longer could be tabletop systems that the individual patients buy, or lease, and they self-administer the treatment, for example, as part of a daily routine. Finally, stimulators that constrict the pupil for an hour, or for a few hours, can be mobile systems which the patient can carry with themselves and apply the treatment on demand. Obviously, stimulators operated by untrained patients have to have much more robust safety, monitoring and control systems to prevent undesirable medical outcomes. In sum, embodiments that constrict the pupil for different time intervals can offer very different medical outcomes, may be operated by very different personnel, and may need very different safety, monitoring and control systems.

[0087] FIG. 11A illustrates embodiments of a method 300, related for the preceding description, for temporarily constricting a pupil 13 of an eye by an ophthalmic stimulator 100. The method 300 includes the following steps:

[0088] generating 310 an irradiation control signal by an irradiation control system 110;

[0089] generating 320 an irradiation 200 by an irradiation source 120, coupled to the irradiation control system 110;

[0090] receiving 330 the irradiation 200, and delivering 332 a patterned irradiation 200p to an iris 11 of the eye with an irradiation delivery system 130; and

[0091] controlling 340 at least one of the irradiation source 120 and the irradiation delivery system 130 by the irradiation control signal of the irradiation control system 110 so that the patterned irradiation is causing a temporary constriction of the pupil of the eye, without causing a permanent constriction of the pupil.

[0092] In embodiments, the generating 320 the irradiation 200 can include generating a light beam, an electromagnetic irradiation, a LED light, a narrow-band light, a broad-band light, an infrared beam, an incoherent light, a radio-frequency beam, or a sound by the irradiation source 120. Another class of irradiation sources 120 can include a coherent light source, a laser beam, a continuous wave laser beam, or a pulsed laser beam. Marked differences between the preceding incoherent irradiation sources and the just-listed coherent and laser sources will be discussed below.

[0093] The delivering 332 of the patterned irradiation 200p can include patterning the irradiation 200 by at least one of a pattern generator 112e, an optical beam shaper 132, a patterning optics, a beam profiler, a beam scanner 134, and a digitally controlled irradiation optics.

[0094] In embodiments, the causing the temporary constriction of the pupil can include increasing a temperature of a treatment region of the iris to a range of 45-60 degrees Celsius. In some embodiments, the temperature of the treatment region of the iris can be raised into a range of 50-55 degrees Celsius.

[0095] FIG. 10 illustrates, that in some embodiments of the method 300, the irradiation control system 110 can

include an imaging system 114, in some cases with a corresponding imaging block 112b in the irradiation controller 112, and a user interface 118, in some cases with a corresponding user interface block 112f in the irradiation controller 112. In these embodiments, the generating 310 of the irradiation control signal can include generating an image of the iris 11 of the eye with the imaging system 114 for a user, receiving an image-based input from the user through the user interface 118, and generating the irradiation control signal to control the irradiation delivery system 130 to deliver the patterned irradiation 200p in accordance with the received input. In embodiments, the patterned irradiation 200p can impact the iris in a ring pattern 210; and the image-based input can be an inner radius Rp(inner) and an outer radius Rp(outer) of the ring pattern 210, selected by the user.

[0096] In some embodiments of the method 300, the irradiation control system 110 can include an imaging system 114, and an image processor 114ip, in some cases implemented in the imaging block 112b of the irradiation controller 112. The generating 310 of the irradiation control signal can include generating an image of the iris of the eye with the imaging system 114 for the image processor 114ip; processing the image of the iris and generating an image-based input by the image processor 114ip; receiving the image-based input from the image processor 114ip; and generating 310 the irradiation control signal to control the irradiation delivery system 130 to deliver the patterned irradiation 200p in accordance with the received image-based input. In some designs, the patterned irradiation 200p can impact the iris 11 in a ring pattern 210; and the image-based input can be an inner radius Rp(inner) and an outer radius Rp(outer) of the ring pattern.

[0097] In some embodiments of the method 300, the irradiation control system 110 can include an alignment system 135, in some cases with its alignment block 112c in the irradiation controller 112; and the generating 310 of the irradiation control signal can include processing alignment data with the alignment system 135, and generating the irradiation control signal to control the irradiation delivery system 130 to deliver the patterned irradiation 200p to the iris in a pattern 210 aligned with the pupil 13 of the iris 11.

[0098] In some embodiments of the method 300, the processing alignment data can include generating an image of the iris with an imaging system 114, and overlaying an alignment pattern 138 on the image, in some cases with the alignment block 112c, or with the image processor 114ip; and the generating 310 the irradiation control signal can include generating a misalignment warning signal, or generating an alignment-guidance signal.

[0099] In some embodiments, the irradiation control system 110 can include a memory block 112d; and the generating the irradiation control signal 310 can include recalling stored data from the memory block 112d, representing at least one of an irradiation pattern 210 and patient data; and generating 310 the irradiation control signal to control the irradiation delivery system 130 to deliver the patterned irradiation 200p to the iris 11 in accordance with the recalled stored data. In some designs, the irradiation control system can include the pattern generator 112e; and the generating 310 of the irradiation control signal can include generating the irradiation pattern 210; and generating the irradiation

control signal to control the irradiation delivery system **130** to deliver the patterned irradiation **200p** with the generated irradiation pattern **210**.

[0100] Some embodiments of the method **300** can include acquiring and analyzing patient data; selecting a treatment region based on the analyzing of the patient data; and delivering the patterned irradiation **200p** to the selected treatment region. A notable embodiment of this step is the ophthalmologist analyzing patient data and deciding whether the treatment radiation shall be applied to the radial dilator muscles **30**, or to the circular sphincter muscles **40**. This analysis and decision can involve selecting the appropriate treatment parameters among the large number of treatment parameters described previously.

[0101] In some cases, the selecting the treatment region can include selecting a ring pattern **210r** with an inner radius $R_p(\text{inner})$ larger than $R(\text{sphincter})$, a radius of a region of the circular sphincter muscles **40**.

[0102] In some cases, the selecting the treatment region can include selecting a ring pattern **210r** with an outer radius $R_p(\text{outer})$ smaller than $R(\text{sphincter})$, the radius of a region of the circular sphincter muscles **40**.

[0103] Some embodiments of the method **300** can include controlling the irradiation source **120**, or the irradiation delivery system **130**, or both, so that the patterned irradiation **200p** is causing a temporary constriction of the pupil of the eye that includes an at least 5% reduction of a radius of the pupil that lasts less than one hour.

[0104] In some cases, the temporary constriction of the pupil can last between one hour and one day. In some cases, the temporary constriction of the pupil can last between one day and one week; in some cases between one week and one month; in some cases between one month and three months; and in some cases between three months and one year. The medical, patient, implementation, and safety differences between embodiments involving temporary constrictions of different duration have been discussed earlier.

[0105] The ophthalmic stimulators **100** described up to now shared a common trait: they caused a temporary constriction of the pupil.

[0106] FIG. 5B illustrates a distinct class of permanent ophthalmic stimulators **100'** that can cause a long-term, or even a permanent constriction of the pupil. These ophthalmic stimulators **100'** share some of the major engineering elements with the temporary constriction stimulators **100**, but have different medical modes of action, different irradiation sources, and stronger safety systems, among others.

[0107] In some embodiments, an ophthalmic stimulator **100'** for constricting a pupil of an eye can include an irradiation control system **110'**, to generate an irradiation control signal; an irradiation source **120'**, coupled to the irradiation control system **110'**, to generate an irradiation **200'**; and an irradiation delivery system **130'**, coupled to the irradiation control system **110'**, to receive the irradiation **200'** from the irradiation source **120'**, and to deliver a patterned irradiation **200p'** to the iris **11** of the eye **1**; wherein the irradiation control system **110'** controls the irradiation source **120'**, or the irradiation delivery system **130'**, or both, with the irradiation control signal so that the patterned irradiation **200p'** causes a long-term constriction of the pupil of the eye.

[0108] In a class of the ophthalmic stimulator **100'**, the irradiation source **120'** can include an incoherent light source, such as a lamp, a LED, an infrared light source, a

radiofrequency source, an electromagnetic source and a sound source. In another class, the irradiation source **120'** can include a coherent light source, such as laser, a pulsed laser and a continuous wave laser. There are substantial differences between irradiation sources that employ incoherent light sources and those that employ coherent light sources, as discussed above.

[0109] In some embodiments, the irradiation delivery system **130'** can include an optical beam shaper and a patterning optics.

[0110] In some embodiments, the ophthalmic stimulator **100'** can be configured to increase a temperature of a treatment region of the iris to a range of 50-80 degrees Celsius. In some embodiments, the ophthalmic stimulator **100'** can be configured to increase a temperature of the treatment region of the iris to a range of 55-70 degrees Celsius.

[0111] Some embodiments of the ophthalmic stimulator **100'** can cause a long-term constriction of the pupil that lasts longer than a year. In some cases, the ophthalmic stimulator **100'** can be designed to cause an irreversible change in the iris of the eye. This long-term, or permanent, change can be a change of the length, or spatial extent, of the treated muscle tissue. In other cases, it can be a reduced, or enhanced, elasticity, or flexibility. In some cases, it can be an altered stiffness. In some cases, it can be an altered reactivity to stimuli.

[0112] The ophthalmic stimulator **100'** achieves the long-term reduction of constriction of the pupil by applying the irradiation **200'** with treatment parameters critically different from the ones used by the temporary stimulator **100**. The critical difference can be one of many factors that cause permanent, or long-term constriction of the pupil, including the followings. Beams with wavelength short enough to cause permanent change. Beams with intensity per area high enough to cause long-term change. Beams with total deposited energy high enough to cause permanent change. Beams with treatment times long enough to cause permanent change. Beams with beam pulses long enough, and frequencies high enough to cause permanent change. Which specific parameters are sufficient to make the change permanent is patient specific and is selected by the surgeon.

[0113] In some embodiments, the irradiation control system **110'** can include an imaging system **114'** and a user interface **118'**. In these embodiments, the irradiation control system **110'** can, generate the irradiation control signal by generating an image of the iris of the eye with the imaging system **114'** for a user, receiving an image-based input from the user through the user interface **118'**, and generating the irradiation control signal to control the irradiation delivery system **130'** to deliver the patterned irradiation **200p'** in accordance with the received input.

[0114] Some of the engineering details of the permanent ophthalmic stimulator **100'** are analogous to that of the temporary ophthalmic stimulator **100'**. To contain the length of this document, some of these details of the stimulator **100'** will not be provided with their own figures, but the corresponding figures in the description of the stimulator **100** will be referenced, with the understanding that those need to be modified to cause a long term, not temporary constriction of the pupil.

[0115] In some embodiments of the ophthalmic stimulator **100'**, the irradiation control system **110'** can include an alignment system **135'**; and the irradiation control system

110' can generate the irradiation control signal by processing alignment data with the alignment system 135', and generating the irradiation control signal to control the irradiation delivery system 130' to deliver the patterned irradiation 200p' to the iris in a pattern 210, aligned with a pupil 13 of the iris 11.

[0116] FIG. 11B illustrates a related method 300' for causing a long-term constriction of a pupil of an eye by the ophthalmic stimulator 100'. The method 300' can include the following steps:

[0117] generating 310' an irradiation control signal by an irradiation control system 110';

[0118] generating 320' an irradiation by an irradiation source 120', coupled to the irradiation control system 110';

[0119] receiving 330' the irradiation and delivering 332' a patterned irradiation to an iris of the eye with an irradiation delivery system 130'; and

[0120] controlling 340' at least one of the irradiation source 120' and the irradiation delivery system 130' by the irradiation control signal of the irradiation control system 110' so that the patterned irradiation causes a long-term constriction of the pupil of the eye.

[0121] In the method 300', the causing the long-term constriction of the pupil can include increasing a temperature of a treatment region of the iris to a range of 50-80 degrees Celsius. In some cases, the method 300' can include increasing a temperature of the treatment region of the iris to a range of 55-70 degrees Celsius. While these ranges have some overlap with temperature ranges described in relation to the temporary stimulator 100, for a particular patient the temperature range where the constriction is temporary can be quite well separated from the temperature range, where the constriction is permanent. For example, for a particular patient, temperatures in the range of 50-55 C may constrict the pupil for a day or less; temperatures in the 55-60 C range may cause the pupil to constrict for a time between a week and a month, temperatures in the 60-65 C range can cause the pupil to constrict for a time between a month and a year, and temperatures in the 65-70 C range may cause the pupil to constrict for a time longer than a year. These long-term changes can very well be associated with an irreversible change in the iris of the eye.

[0122] As before, in some embodiments of the method 300' the irradiation control system 110' can include an imaging system 114 and a user interface 118; and the generating the irradiation control signal can include generating an image of the iris of the eye with the imaging system 114 for a user, receiving an image-based input from the user through the user interface 118, and generating the irradiation control signal to control the irradiation delivery system 130' to deliver the patterned irradiation 200p' in accordance with the received input.

[0123] In some embodiments of the method 300', the irradiation control system 110' can include an alignment system 135; and the generating the irradiation control signal can include processing alignment data with the alignment system 135, and generating the irradiation control signal to control the irradiation delivery system 130' to deliver the patterned irradiation 200p' to the iris in a pattern 210 aligned with a pupil of the iris.

[0124] As discussed, the ophthalmologist operating the stimulator 100' can analyze several factors when practicing the method 300'. The analysis can include the determination

what treatment parameters to use to achieve a long-term or permanent constriction change, to go beyond the previously described temporal changes. The analysis can also be focused at which treatment regions to irradiate. As discussed before, some vision-improvement goals can be better achieved by heat-treating the radial dilator muscles 30, others by heat-treating the circular sphincter muscles 40.

[0125] Both of these analyses can involve acquiring and analyzing patient data. In a typical example, a patient may have used the temporary ophthalmic stimulator 100 by practicing the method 300 repeatedly and for an extended period, and may have grown comfortable with its effect to the degree that she/he decided to make the constriction of the pupil permanent. During these preceding temporary treatments, the irradiation controller 110 of the stimulator 100, or its operator may have acquired and collected a substantial amount of data about the particular patient. An ophthalmologist, who is planning administering a higher energy irradiation by practicing the method 300' with a permanent ophthalmic stimulator 100' to permanently change the constriction of the pupil, may evaluate and analyze the data that was collected during the previous, repeated temporary constrictions of the pupil of this particular patient. This analysis can be followed by selecting a treatment region based on the analyzing of the patient data; and delivering the patterned irradiation 200p' to the selected treatment region to cause the long-term constriction of the pupil.

[0126] FIG. 7A illustrates that some embodiments of the ophthalmic stimulator 100 may include an irradiation control system 110, to generate an irradiation control signal; a light source 120, coupled to the irradiation control system 110, to generate a light beam 200; and a beam-shaping optics 132, coupled to the irradiation control system 110, to receive the light beam 200 from the light source 120, and to deliver a light ring 200r to an iris 11 of the eye in a ring pattern 210r. In embodiments, the irradiation control system 110 can control the light source 120, or the beam-shaping optics 132, or both, with the irradiation control signal so that the light ring 200r causes a temporary constriction of the pupil of the eye, without causing a permanent constriction of the pupil. The beam-shaping optics can also include an objective 133, to direct the light ring 200p towards the iris of the eye, to provide additional control.

[0127] Embodiments of the here-described ophthalmic stimulator 100 can be analogous, or equivalent to the embodiments described in relation to the stimulator 100 in relation to FIGS. 5A-B and 6A-D. In particular, the embodiments of the irradiation source 120 can also serve as the light source 120 here. For example, the light source can be an infrared light source. Also, the beam-shaping optics 132 can be an embodiment of the irradiation delivery system 130.

[0128] FIG. 7B illustrates that the beam-shaping optics 132 can include a proximal axicon lens 140, positioned with its base-plane oriented toward the light source 120, to transform the received light beam 200 into the light ring 200r.

[0129] Here it is recalled that an axicon lens is a glass cone with a circle as its base. An axicon lens can be also visualized as an isosceles triangle, rotated around its axis of symmetry. Direct ray tracing establishes that axicon lenses transform a regular, full light beam into a light ring 200r. Generating the light ring 200r "passively", without any scanners, or other digitally controlled active optics with moving parts, makes an axicon lens a very useful, simple,

and reliable implementation of the beam-shaping optics 132 for the purposes of the stimulator 100.

[0130] However, it is also noted that the radius $r(\text{ring})$ of the light ring 200r increases with the distance $d(\text{target})$ from the axicon lens 140. Therefore, if the patient moves her/his head along the optical axis, doing so changes the radius $r(\text{ring})$ of the light ring 200r and can have undesirable medical effect.

[0131] FIG. 7C illustrates an embodiment of the beam-shaping optics 132 that resolves this problem. This embodiment includes the proximal axicon lens 140-1, with its base-plane oriented towards the light source 120. It further includes a second, distal, “complementary” collimating axicon lens 140-2, that is co-axial with the proximal axicon lens 140-1, positioned with its cone-tip oriented toward a cone-tip of the proximal axicon lens 140-1, to collimate the light ring with the increasing radius into a light ring with a constant radius, independent of the distance $d(\text{target})$.

[0132] Embodiments with such a complementary axicon lens pair 140-1 and 140-2 can further include a lens position actuator 141, to adjust an axicon distance $d(\text{axicon})$ between the proximal axicon lens 140-1 and the distal axicon lens 140-2. Changing the axicon distance $d(\text{axicon})$ can be used to adjust the radius $r(\text{ring})$ of the light ring 210 as part of the setting of the overall ring pattern 210 by the ophthalmic surgeon in FIGS. 12D-E.

[0133] Additional optical solutions may be needed to tune $R_p(\text{inner})$ independently from $R_p(\text{outer})$, to tune the radius of the ring independently from its width. Examples of such solutions include (a) a beam blocker to block out part of the light ring; (b) a deformable axicon lens 140, capable of changing the angle of the cone of the axicon lens; and (c) a deformable mirror, in some cases a deformable conical mirror.

[0134] An important aspect of ophthalmic irradiation systems is to ensure that the patient's eye is aligned with the optical axis of the irradiation system. Previously, various alignment systems 135 have been already described. A particularly useful element of such alignment systems 135 can be a fixation light 202, as mentioned. The surgeon may instruct the patient to stare, or fixate, on a centrally positioned fixation light. Such fixation lights 202 can be provided by a small bright LED, positioned centrally, projected into, or superimposed into the optical pathway.

[0135] FIG. 7D shows that the beam-shaping optics 132 that uses an axicon lens 140 offers a particularly simple implementation of the fixation light 202. In some embodiments, the tip of the cone of the axicon lens 140 may be flattened. Such flattened tip axicon lenses 140 do not redirect or refract the small central portion of the incoming light 200, so that they propagate centrally and thus can act as the fixation light 202. Such embodiments are attractive because the fixation light 202 is naturally centered with the beam-shaping optics 132, without the need to introduce any additional structures to hold the fixation light in place that can at the same time block part of the light 200, and without the need of centering the fixation light 202 by a finely adjustable system.

[0136] In the case when the treatment light 200 is an infrared light, the flattened tip can be covered by a luminescent material, a phosphor, an upconverting material, a higher harmonic generating material, a multi-photon induced fluorescence material, or any optical material or

structure that converts the infrared light 200 into a visible light, needed as a fixation light 202.

[0137] FIG. 7E illustrates an embodiment of the beam-shaping optics 132. The incoming light 200 can be guided through a pair of beam-expanding lenses: a diverging, lens 142, followed by a collimating lens 143. This 142-143 lens combination can expand the radius of the incoming beam 200 to the $R_p(\text{outer})$ outer radius, set or desired by the ophthalmic surgeon. The beam radius can be adjusted by adjusting the distance of the diverging lens 142 from the collimating lens 143 by an actuator. Next, the expanded beam can be directed at an adjustable beam stop 144 that can block out a central portion of the expanded beam so that the transmitted beam has an inner radius equaling $R_p(\text{inner})$ as set by the surgeon. The radius of the adjustable beam stop 144 can be adjusted by a number of known mechanical designs. Further, since the stopped beam carries an energy with it that can undesirably heat the beam-shaping optics 132, a heat sink 145 can be employed, configured to absorb, or guide away the energy of the stopped beam. Many heat sinks are known, such as metallic ribs, and air-cooled systems. It is also possible to reflect the stopped beam out of the beam-shaping optics 132 and absorb it or release it peripherally. These solutions reduce the need for heat management greatly.

[0138] FIG. 7F illustrates another embodiment of the beam-shaping optic 132. This embodiment 132 can be configured to generate a light beam directly with a ring shape, without the need of an optics that would transform the generated light. In a typical embodiment, the light source 200 can include a ring of LEDs 146-1, 146-2, . . . 146-N, collectively referenced as 146-i, to generate light beamlets; and a ring-shaped diffuser 147, to transform the light beamlets into a light beam with a well-distributed intensity profile to form the light ring 200r. In some embodiments, there can be more than one ring of LEDs 146-i. Activating a different number of LED rings can be one way to adjust the radius of the light ring 200r. (Throughout this document, elements “x-1, x-2, . . . , x-N” will be sometimes collectively referenced as “x-i”, for brevity.)

[0139] Finally, the embodiments of FIGS. 7B-F can be combined. One such combination was already mentioned. The embodiment based on the axicon-lens 140, or 140-1/140-2, may need additional optical elements to adjust the inner and outer radii $R_p(\text{inner})$ and $R_p(\text{outer})$ independently. In some cases, the beam stop 144 can be used to adjust the inner radius $R_p(\text{inner})$ of the light ring 200r that was generated by the axicon lens 140.

[0140] FIG. 11C illustrates a method 302 for temporarily constricting a pupil of an eye by an ophthalmic stimulator 100. The method 302 can include the following steps:

[0141] generating 302a an irradiation control signal by an irradiation control system 110;

[0142] generating 302b a light beam 200 by a light source 120, coupled to the irradiation control system 110;

[0143] receiving 302c the light beam 200, and delivering 302d a light ring 200r to an iris of the eye with a beam-Shaping optics 132; and

[0144] controlling 302e at least one of the light source 120 and the beam-shaping optics 132 by the irradiation control signal of the irradiation control system 110 so that the light ring 200r is causing a temporary constrict-

tion of the pupil of the eye, without causing a permanent constriction of the pupil.

[0145] In some embodiments of the method 302, the delivering 302d the light ring 200r can include transforming the received light beam 200 into the light ring 200r by a proximal axicon lens 140, positioned with its base-plane oriented toward the light source 120, wherein the light ring 200r has an increasing radius $r(\text{ring})$ with increasing distance $d(\text{target})$ from the axicon lens 140.

[0146] In some cases, the delivering 302d the light ring 200r can include collimating the light ring with the increasing radius into a light ring 200r with a constant radius by a distal collimating axicon lens 140-2, co-axial with the proximal axicon lens 140-1, positioned with its cone-tip oriented toward a cone-tip of the proximal axicon lens 140-1. In these embodiments, the delivering the light ring can include adjusting an axicon distance $d(\text{axicon})$ between the proximal axicon lens 140-1 and the distal axicon lens 140-2 by a lens position actuator 141, thereby adjusting the radius of the light ring, $r(\text{ring})$.

[0147] The method 302 can also include generating a fixation light 202 by selectively transmitting a small fraction of the received light beam 200 by a flattened cone-tip of the proximal axicon lens 140-1. In embodiments where the light is an infrared light, the small, flattened tip of the axicon lens 140-1 can be covered by an optical material that can transform the infrared light into a visible light.

[0148] In some embodiments of the method 302, the delivering 302d the light ring can include utilizing a beam stop 144 to generate the light ring 200r by blocking a central portion of the received light beam 200.

[0149] Finally, in some embodiments of the method 302, the generating 302b a light beam can include generating the light beam with a ring shape by the light source including a ring of LEDs 146-i.

[0150] FIG. 8A illustrates that embodiments of the ophthalmic stimulator 100 for temporarily constricting a pupil of an eye can include a digital beam controller 110, to generate a digital beam-control signal; a light source 120, coupled to the digital beam controller 110, to generate a light beam 200; and a digitally controlled beam modulator 134, for example a beam scanner 134, to receive the digital beam-control signal from the digital beam controller 110, to receive the light beam from the light source 120, and to modulate the received light beam into a modulated light, or modulated light 200m, delivered to an iris of the eye. In embodiments, the digital beam controller 110 can control the light source 120, the digitally controlled beam modulator 134, or both, with the digital beam-control signal so that the modulated light 200m causes a temporary constriction of the pupil of the eye, without causing a permanent constriction of the pupil.

[0151] As before, embodiments of the here-described ophthalmic stimulator 100 can be analogous, or equivalent to the embodiments described in relation to the ophthalmic stimulator 100 in relation to FIGS. 5A-B, 6A-D, and 7A-F. In particular, the embodiments of the irradiation control system 110 can be analogous, or equivalent, to the embodiments of the digital beam controller 110, the irradiation source 120 can also serve as the light source 120 here, and the digitally controlled beam controller 134 can be an embodiment of the irradiation delivery system 130.

[0152] In what follows, numerous examples of the digitally controlled beam modulator 134 will be described. To

emphasize that all these are embodiments of the same block, they are all labeled with 134 or as a variant of label 134.

[0153] For example, FIG. 8A illustrates a beam scanner 134 as an embodiment of the digitally controlled beam modulator 134, to scan the received light beam 200 according to a pattern 210 on the iris. Embodiments described in relation to FIGS. 8A-B, and FIGS. 9A-E can be different from the embodiments described in relation to FIGS. 7A-F in that the latter embodiments utilize dominantly “passive” optical elements, such as lenses and mirrors, and do not need elaborate digital control signals and moving parts, with the possible exception of the lens position actuator 141. Also, the systems of FIGS. 7A-F typically irradiate the pattern 210 simultaneously.

[0154] In contrast, the digitally controlled embodiments of FIGS. 8A-B, and FIGS. 9A-E can involve active elements, where extensive digital control signals move or adjust a number of active optical elements. These embodiments typically irradiate the iris on a point-by-point basis, with the help of various types of scanners and optical arrays. As such, these digitally controlled embodiments can offer higher precision and control, at the same time, they can be more complex, raising issues of reliability, maintenance and costs, and the irradiation treatment can take longer. Also, the points of the pattern 210 are often irradiated sequentially, instead of simultaneously.

[0155] FIG. 8B illustrates one embodiment of the digitally controlled beam modulator 134 in a laser-based ophthalmic stimulator 100. The light source 120 can be a laser source 120L, emitting a laser beam 200L. The digitally controlled beam modulator 134 can be an X-Y scanner 134L, to scan the received laser beam 200L as a scanned laser beam 200sL according to a pattern on the iris. A large number of laser scanners are known that can scan the scanned laser beam 200sL with a wide variety of complex patterns 210.

[0156] FIG. 9A illustrates the first of a set of reflection mode beam modulators 134r. While the scanner embodiments in FIGS. 8A-B irradiate the iris in a pattern 210 sequentially, the embodiments of FIGS. 9A-E can irradiate the pattern 210 either sequentially, or simultaneously. The embodiment of FIG. 9A includes a reflective LCD array 150 with an addressable array of LCD pixels 152. Switching the LCD pixels 152 on-off can control how much of an incoming light the LCD array reflects from through the LCD pixels.

[0157] FIG. 9B illustrates another embodiment of the reflection-mode beam modulator 134r that includes a deformable reflector 160, with a substrate 162; a mechanical actuator array 164, positioned on the substrate 162; and a deformable mirror 166, positioned to be deformable by the mechanical actuator array 164 according to the digital beam-control signal.

[0158] FIG. 9C illustrates yet another embodiment of a reflection-mode beam modulator 134r. This is an acousto-optical modulator 170 that includes a set of acoustic piezo transducers 172, to deform a deformable reflector 174, according to the beam-control signal. This embodiment has similarities to the previous one in FIG. 9B. One of the differences is that the deformation is performed not by an array that can be controlled point-by-point, but in a global manner, where the transducers are operated to form patterns across the entire deformable reflector 174 simultaneously.

[0159] FIG. 9D illustrates yet another embodiment of a reflection-mode beam modulator 134r. This is a digital

mirror device **180** that includes a substrate **182**; an array of mechanical actuators **184-i**, positioned on the substrate **182**; and an array of rotatable mirrors **186-i**, where the rotatable mirrors **186-i** are rotatable individually by the actuators **184-i** according to the beam-control signal. Such digital mirror arrays are well known in digital projectors, for example.

[0160] FIG. 9E illustrates a different, transmission-mode beam modulator **134t**. This embodiment can include an addressable pixel array **190** of variable transparency pixels **192-i**. This embodiment **190** has design aspects analogous to the embodiment **150** in FIG. 9A, as it also builds on the principle of individual pixels changing their optical (reflective or transmissive) properties under electric, control, thereby modulating the beam on a pixel-by-pixel basis. As indicated earlier, the irradiation can be either sequential or in parallel, the latter type embodiments requiring much less moving parts and allowing shorter irradiation times.

[0161] FIGS. 17A-D illustrate various patterns **210** the digitally controlled ophthalmic stimulators **100** can irradiate on the iris with the modulated beam **200m**. In embodiments, illustrated in FIG. 17A, the digitally controlled beam modulator **134** can be controlled by the beam controller **110** to modulate the received light beam **200** into a modulated light **200m**, so that it irradiates a pattern **210** that is a ring, or multiple rings. FIG. 17B illustrates a pattern **210** that is a segmented ring. FIG. 17C illustrates a pattern **210** that includes radial spokes. Finally, FIG. 17D illustrates a pattern that is a combination of ring segments and spokes.

[0162] FIG. 11D illustrates a method **304** that is related to operating the digitally controlled ophthalmic stimulators **100**. The method **304** can include the following steps:

[0163] generating **304a** a digital beam-control signal by a digital beam controller **110**;

[0164] generating **304b** a light beam **200** by a light source **120**, coupled to the digital beam controller **110**;

[0165] receiving **304c** the light beam **200**, modulating **304d** the light beam **200** into a modulated light **200m**, and delivering **304e** the modulated light **200m** to an iris of the eye, with a digitally controlled beam modulator **134**; and

[0166] controlling **304f** the light source **110**, the digitally controlled beam modulator **134**, or both, by the digital beam-control signal of the digital beam controller **110** so that the modulated light **200m** is causing a temporary constriction of the pupil of the eye, without causing a permanent constriction of the pupil.

[0167] In embodiments, the modulating **304d** can include scanning the received light beam on the iris according to a pattern by a beam scanner **134**. In other embodiments, the modulating **304d** can include modulating the light by a reflection-mode beam modulator **134r**. The reflection-mode beam modulator **134r** can be a reflective LCD array **150**, with an addressable array of LCD pixels, a deformable reflector **160**, an acousto-optical modulator **170**, and a digital mirror device **180**. In some embodiments of the method, the modulating **304d** can include modulating the light by a transmission-mode beam modulator **134t**.

[0168] Finally, the modulating **304d** can include modulating the received light beam into the modulated light **200m** with the pattern being one of a ring, multiple rings, a segmented ring, a pattern of radial spokes, and a combination of ring segments and spokes.

[0169] FIG. 15 illustrates other embodiments of an ophthalmic stimulator **100** for temporarily constricting a pupil of an eye that includes an irradiation control system **110**, having a feedback system, to generate an irradiation control signal using a feedback of the feedback system; an irradiation source **120**, coupled to the irradiation control system **110**, to generate an irradiation; and an irradiation delivery system **130**, coupled to the irradiation control system **110**, to receive the irradiation **200** from the irradiation source **120**, and to direct a patterned irradiation **200p** in a pattern to a treatment region of an iris of the eye, guided by the feedback-based irradiation control signal; wherein the irradiation control system **110** controls at least one of the irradiation source **120** and the irradiation delivery system **130** with the feedback-based irradiation control signal so that the patterned irradiation **200p** causes a temporary constriction of the pupil, without causing a permanent constriction of the pupil. As before, the here-described embodiments can be analogous to the ones described in relation to FIGS. 5A-B, FIGS. 6A-D, FIGS. 7A-F, and FIGS. 8A-B, and analogously labeled elements can serve analogous functions.

[0170] In what follows, various embodiments and blocks of the feedback system **116** will be described. These embodiments and blocks typically include a hardware block, such as an imaging system, or a temperature sensor. They are coupled to the irradiation controller **112**, which processes their feedback and generates irradiation control signals, to be transmitted to the irradiation source **120** and to the irradiation delivery system **130**. As discussed in relation to FIG. 10, the irradiation controller **112** can include corresponding blocks that are dedicated to receive the feedback. For example, the irradiation controller **112** can include the dedicated feedback block **112a** to receive the feedback from an embodiment of the feedback system **116**. These receiving blocks can be implemented in hardware, such as an application specific integrated circuit ASIC; or they can be implemented in a software form, such as a piece of code or application, implemented in the processor **113** of the irradiation controller **112**; or in a shared processor, or input/output controller. In yet other embodiments, the feedback can be coupled straight into the central processor **113**, whose code can process the feedback directly. A particularly simple implementation can be a simple “stop” feedback signal, triggered by a security concern, which can be directly executed by the processor by shutting down the irradiation source **120** with a control signal, without the need of any intermediate processing.

[0171] In some embodiments, the feedback system **116** can include at least one of a pupillometer **116a** and an imaging system **114**, to sense a diameter of the pupil, and to generate a feedback according to the sensed pupil diameter. As discussed just now, this feedback can be received and processed either by a dedicated feedback block **112a** that is implemented inside the irradiation controller **112**, or can be received by the processor **113** of the irradiation controller **112** itself. In some embodiments, the pupillometer **116a** can be coupled to the irradiation controller **112** directly, in others, through a user interface **118-1a**. Similarly, the imaging system **114** can be coupled to the irradiation controller **112** directly, or through a user interface **118-2**.

[0172] FIGS. 16A-E illustrate methods, or processes, **510-550** that operate in relation to the embodiments **116a-f** of the feedback system **116**. In general, the methods, or processes, **510-550** can include the following steps:

[0173] generating a feedback-based irradiation control signal by an irradiation control system 110, using a feedback of a feedback system of the irradiation control system;

[0174] generating an irradiation 200 by an irradiation source 120, coupled to the irradiation control system;

[0175] receiving the irradiation 200 and directing a patterned irradiation 200p to a treatment region of an iris of the eye with an irradiation delivery system 130, guided by the feedback-based irradiation-control signal; and

[0176] controlling at least one of the irradiation source 120 and the irradiation delivery system 130 with the feedback-based irradiation control signal of the irradiation control system 110 so that the patterned irradiation 200p causes a temporary constriction of the pupil, without causing a permanent constriction of the pupil.

[0177] The description continues with details of the processes, or methods, 510-550. FIG. 16A illustrates that in a representative case, the feedback can be generated through the following sequence, method, or process 510. A short time after starting to apply the patterned irradiation 200p to the iris 11 according to the pattern 210, in step 511, the pupillometer 116a, or the imaging system 114, can sense that “Pupil radius is large relative to a reference or target”, or “Target radius not reached”. This can be followed by a step 512, generating the feedback, or feedback signal: “Carry on irradiation”, as indicated by the eye and steps on the left side of FIG. 16A. Here and in what follows, each “step x of generating feedback signal” may also be referred to with the shorter form of “feedback signal x”, for brevity. Also, the feedback signal can include not only the command to continue or to stop the irradiation, but it can also include the sensed information as well. In the present example, the step 512 of generating a feedback signal can include sending the feedback signal “Target radius not reached. Carry on.”

[0178] With the passing of time, the irradiation increases the temperature of a portion of the iris 11, as indicated by the denser dot-filling of the pattern 210 on the right. The increased temperature induces the constriction of the pupil 13, as indicated by the eye 1 having a smaller pupil 13 on the right of FIG. 16A. In step 513, the pupillometer 116a, or the imaging system 114, can sense that “Pupil radius is sufficiently close to the reference”, or “Target pupil radius sensed”. This can be followed by the generation of the feedback in step 514: “Power down irradiation”, or “Target radius reached. Power down”. This feedback, or feedback signal 514 can be transmitted by the feedback system 116 to the irradiation controller 112. In response, the irradiation control system 110 can send a corresponding feedback-based irradiation control signal to the irradiation source 120 to power down. The top graph of FIG. 16A illustrates that the feedback-induced irradiation control signal 514 causes the powering down of the irradiation after the receiving of the feedback signal 514.

[0179] It is mentioned here that pupillometers reached a high level of sophistication and can provide a variety of useful, actionable information. For a review of the field, see Olson D, Stutzman S, Saju C, Wilson M, Zhao, W., Aiyagari V. Interrater of Papillary Assessments, Neurocrit Care, Published online: 17 Sep. 2015. These pupillometers can assess pupil size, and shape with very high accuracy and reproducibility. In addition, such devices can measure parameters such as onset and peak constriction, constriction and dilation

velocity, and latency using various light stimuli, both before and after treatment to assess effects that may not be apparent simply based on pupil diameter.

[0180] FIG. 15 illustrates that in some embodiments, the feedback system 116 can include a pupillometer 116a, and at least one of an infrared sensor or camera 116b to sense a temperature of the treatment region, and to generate a feedback according to the sensed temperature. As before, the infrared camera 116b can be coupled to the irradiation controller 112 directly, or via a user interface 118-1b.

[0181] FIG. 16B illustrates the corresponding process 520, or method 520, or sequence of operation of this infrared sensor/camera 116b. At an early time during the irradiation, in a step 521, the infrared sensor/camera 116b can sense “Temperature low relative to a reference”, or simply “Low temperature”. In a typical case, a temperature T sensed to be less than 45 C can be classified as “low temperature”. This can prompt generating the feedback in step 522: “Carry on irradiation”. Visibly this feedback leads to the maintaining the power of the irradiation, as indicated by the graph on top of FIG. 16B.

[0182] With the progression of the irradiation time, the target region, irradiated according to the pattern 210, starts warming up. This is indicated by the dotting of the pattern 210 getting denser. After some time, in step 523, the infrared (thermal) sensor/camera 116b can sense “a medium temperature relative to the reference”, or simply “medium temperature”. In a typical example, this can be a temperature in the 45 C-55 C range. In response, a feedback signal can be generated in step 524, sent from the feedback system 116 to the irradiation controller 112: “Start power down the irradiation”, or “Medium temperature. Power down”. As indicated, the irradiation control system 110 can generate a feedback-based irradiation control signal to the irradiation source 120, which in response can start powering down the power of the irradiation gradually, as indicated by the dashed line in the top graph.

[0183] In some embodiments, the settings and thresholds can be chosen differently. In such cases, the IR camera 116b can wait until it senses a “high temperature relative to the reference” in step 525, such as the IR sensor/camera 116b senses the temperature T that exceeds 55 C. Such a sensing by the IR sensor/camera 116b can prompt the generation of the feedback “Stop the irradiation” in step 526, to be sent to the irradiation controller 112. Analogously to earlier steps of the process, the irradiation control system 110 can generate a feedback-based irradiation control signal for the irradiation source 120 to discontinue the irradiation with a hard stop, as indicated by the solid line in the top graph of FIG. 16B.

[0184] One such scenario can be associated with an irregular, or unexpected progress of the irradiation, when, for whatever reason, the iris heats faster than expected. This can be a consequence of an unexpected patient response, or an incorrect calibration of the irradiation’s treatment parameters. Once the IR camera 116b senses that the temperature rose to a value high relative to a reference, such as to above 55 C, for safety reasons the feedback-based irradiation control signal can bring the irradiation power to zero via a hard stop.

[0185] FIG. 15 illustrates that the feedback system 116 can further include at least one of an alignment system 116c, an eye tracker 116d, a wavefront sensor 116e, an iris scanner 116f, and an imaging system 114. The alignment system 116c can be related to, combined with, or analogous to any

embodiment of the alignment system 135, described earlier, for example in relation to FIGS. 12A-C. Any of these feedback implementations can sense an alignment of one of the iris and the pupil relative to the irradiation delivery system 130, as discussed earlier.

[0186] FIG. 16C illustrates a mode of operation, or method 530 for such alignment-related feedback implementations. While FIG. 16C specifically refers to the eye tracker feedback 116d, an analogous process can be practiced with the analogous feedback alignment system 116c, iris scanner 116f, or imaging system 114. In a step 531, the eye tracker 116d can “sense alignment” between the iris 11, the pupil 13 and the irradiation delivery system 130. Sensing alignment in step 531 can lead to generating, in step 532, the feedback “Eye aligned. Carry on irradiation”, which results in the irradiation source 120 maintaining the power of the irradiation, as shown by the top graph of FIG. 16C.

[0187] A central concern for the efficacy and safety of the irradiation treatment is that the eye 1, iris 11, and pupil 13 remain aligned with the irradiation delivery system throughout the irradiation. However, there is a possibility that the eye, iris, and pupil become misaligned. This can be caused by an involuntary eye movement by the patient, a reaction to a sensation of discomfort or pain by the patient, or a problem developing with the patient interface 137, such as the breaking of a vacuum suction, among others. Also, misalignment can be the natural consequence of the ophthalmologist not using a firm eye-fixation method, such as physically restraining the eyeball only by hand, or by pressure with a forceps. In these cases, the gaze of the eye can naturally drift away to a degree that it becomes misaligned with the pattern 210 and the irradiation delivery system 130.

[0188] FIG. 16C illustrates that the eye can get misaligned to a degree that the patterned irradiation 200p may reach the edge of the pupil 13. In such cases, the irradiation may start hitting the retina, a much more light-sensitive tissue. This raises a higher level of safety concerns. Embodiments of the feedback system 116 can handle such developments by the eye tracker 116d “sensing a misalignment”, or “misalignment sensed” in step 533. This can lead to a generation of a feedback signal “Eye misaligned! Safety Stop!” in step 534. The irradiation control system 110 can generate a corresponding feedback-based irradiation control signal for the irradiation source 120, which in response can execute a hard stop of the irradiation, as shown by the top graph. The step 534 can be accompanied with a signal to an operator, or user: “Realign at least one of the irradiation delivery system, the iris, and the pupil.”

[0189] FIG. 16D illustrates a process, or method 540 that flexibly manages naturally occurring misalignments. Steps 541-544 are analogous to steps 531-534, in relation to the eye losing alignment with the irradiation delivery system 120. However, the process 540 can dynamically manage if the misalignment developed not as a safety-threatening problem that required an irreversible hard stop, but as a consequence of a naturally shifting eye, which can be followed by the eye realigning with the irradiation delivery system 130. A typical situation can be when the eye is not docked to the ophthalmic stimulator 100 in a fixed manner with a patient interface 137, but is left free. In such embodiments, the patient may be fixating on a fixation light, but her gaze can be distracted for a short period by natural processes such as a mild discomfort or blinking, after which the patient

re-fixates on the fixation light, thus realigning the eye with the irradiation delivery system 130. Such scenarios can be managed by the process 540 via step 545, where the eye tracker 116d can “sense a realignment”, followed by step 546, where a feedback signal is generated confirming “Eye realigned. Resume irradiation”. The irradiation control system 110 can then generate a feedback-based irradiation control signal that makes the irradiation source 120 to resume the irradiation.

[0190] In some cases, the “stop irradiation 544—resume irradiation 546” sequence can be repeated several times. A notable embodiment can be a hand-held, mobile ophthalmic stimulator 100m, described below in relation to FIGS. 13A-C, where the eye can fall out from alignment relative to the irradiation delivery system 130m repeatedly, followed by the eye getting realigned with the irradiation delivery system 130m of the mobile ophthalmic stimulator 100m repeatedly, since the eyes of the patient are not held firmly in place by an immobilizing system.

[0191] Finally, FIG. 16E illustrates yet another feedback method or process 550. In this method, or process, 550, the feedback system 116 can include at least one of the pupilometer 116a and the imaging system 114, to sense at least one of a pupil characteristic or an iris characteristic, and to generate a feedback according to the sensed characteristic.

[0192] In step 551, the imaging system 114 may sense that the pupil 13 has a regular shape. In response, it may generate the feedback signal: “Progress regular. Carry on,” in step 552. However, in some cases, in step 553 the imaging system 114 may sense, or image, that an “irregular pupil shape” is emerging as a consequence of the irradiation. In other embodiments, the imaging system 114 may sense, or image, that at least one of a pupil characteristic and an iris characteristic is becoming unacceptable relative to a reference as a consequence of the irradiation. These situations can arise, when the pupil does not react according to medical expectations to the irradiation. A simple example can be that the pupil starts to lose its circular shape, and evolve toward an elongated, or irregular shape. A non-circular pupil can be perceived as an undesirable treatment outcome and therefore necessitates safety protocols within the feedback system 116 to manage or to counter-act it.

[0193] A corresponding step 554 can include the generation of a “modify irradiation pattern” feedback signal, possibly preceded by a “safety stop” feedback signal 554. The process 550 can be continued by the pattern generator 112e actually modifying the irradiation pattern 210 in step 555, followed by generating a “Pattern modified. Resume irradiation,” feedback 556.

[0194] FIG. 16E illustrates a characteristic example, where the pupil 13 starts to evolve from circular towards an elongated oval shape because of the irradiation. This undesirable process can be detected by the imaging system 114 in step 553. In response to the corresponding, feedback-based irradiation control signal, the irradiation delivery system 130 may change the irradiation pattern 210 from a circle into an oval that is oriented 90 degree opposite to the pupil’s oval. Such a modified irradiation pattern 210p may be successful to counter-act the development of the undesirable oval pupil.

[0195] Irradiation delivery systems 130 and 134 that are digitally controlled and active systems, like the beam modulators and beam scanners 134 of FIGS. 8A-B, and the digitally controlled beam modulators 134 of FIGS. 9A-E,

can modify the irradiation patterns **210** relatively easily. The beam-shaping optics **132** of the optical systems in FIGS. 7A-F with little or no digital, point-by-point control can also have some such functionalities. A simple embodiment can be the beam-shaping optics **132** including deformable mirrors. Actuators along the periphery, or along the perimeter of such deformable mirrors can elongate a circular pattern **210** into an oval pattern **210** by a simple cylindrical deformation of the mirror. Other low order wavefront deformations can be also introduced by deforming such a deformable mirror. Such deformable mirrors were also described earlier as systems that can enable the independent tuning of the $R_p(\text{inner})$ and the $R_p(\text{outer})$ radii of the pattern **210**, and also in relation to FIGS. 9B-C.

[0196] Further embodiments can include further methods or processes, where the feedback system **116** includes the wavefront sensor **116e**, or the iris scanner **116f**, and the method includes generating a feedback based on a condition of at least one of the iris and the pupil, sensed by the wavefront sensor **116e**, or the iris scanner **116f**.

[0197] In yet other embodiments, the feedback system **116** can be configured to carry out a test and then generate a feedback signal based on the test. In a simple embodiment, during the treatment, a short light pulse can be sent to the eye, and the reaction time, or the reaction radius-change of the pupil can be measured and assessed by the feedback system **116**. A feedback-based irradiation control signal can then be generated based on this assessment.

[0198] In some cases, the feedback by the feedback system **116** can serve only a diagnostic purpose, not necessarily leading to the generation of a feedback signal to impact the irradiation. This feedback can be a visual feedback for the operator, or user of the ophthalmic stimulator **100** via a user interface **118-1a** to **118-1f** or **118-2**. The user may, in response to this visual feedback, then modify the treatment. The feedback can be a wide variety of information, from pupil size to sensed temperature, to a pupil shape or alignment.

[0199] In some embodiments, an ophthalmic stimulator **100** for constricting the pupil **13** of an eye **1** can include an irradiation control system **110**, having a safety feedback system **116**, to generate a feedback-based irradiation control signal using a feedback of the safety feedback system **116**; an irradiation source **120**, coupled to the irradiation control system **110**, to generate an irradiation **200**; and an irradiation delivery system **130**, coupled to the irradiation control system **110**, to receive the irradiation **200** from the irradiation source **120**, and to direct a patterned irradiation **200p** in a pattern **210** to a treatment region of an iris **11** of the eye **1**, guided by the feedback-based irradiation control signal. In these embodiments, the irradiation control system **110** can control at least one of the irradiation source **120** and the irradiation delivery system **130** with the feedback-based irradiation control signal so that the patterned irradiation **200p** causes a long-term constriction of the pupil **13**.

[0200] Including the safety feedback system **116** can be critically important to guarantee the safety of the operation of the ophthalmic stimulators **100**, especially those that are powerful enough to cause a long-term constriction of the pupil **13**. In the here-described embodiments, the safety feedback systems **116** are safety-oriented feedback systems, whose role can be critically different from simple guiding or aligning feedback systems which may only help the targeting of the irradiation, but are not part of ensuring the safety

of the irradiation. Various embodiments of the safety feedback system **116** can provide feedback when the desired result has been achieved, as determined from a pupil size measurement, for example. This feedback ensures the safety of the treatment by signaling that further irradiation is not necessary and in fact may be harmful. In other embodiments, a feedback can be generated by the safety feedback system **116** if a misalignment has been detected, and the irradiation by the laser source **120L** may be inadvertently directed at the retina of the eye **1**. Yet other examples include generating a feedback to indicate that a temperature of the target tissue rose to a reference level, or exceeded a critical level. In these embodiments, the feedback of the safety feedback system **116** can indicate that the treatment goal has been achieved, or the possibility of an undesirable medical outcome, such as retinal exposure, or overheating the iris tissue. Receiving such a feedback from the safety feedback system **116** can prompt the irradiation control system **110** to generate a feedback-based irradiation control signal to stop, or to power down the irradiation. Employing such a safety feedback system **116** can ensure that the irradiation is stopped when the treatment goal is achieved, or when there is a possibility of an undesirable medical outcome. As such, including the safety feedback system **116** into the ophthalmic stimulator **100** can be critical to ensure its safe operation.

[0201] FIG. 15 illustrates that in some embodiments of the ophthalmic stimulator **100** the safety feedback system **116** can include at least one of a pupilometer **116a** and an imaging system **114**, to sense a diameter of the pupil, and to generate a feedback according to the sensed pupil diameter. The generated feedback can include at least one of “pupil radius is large relative to a reference, carry on irradiation”; and “pupil radius is sufficiently close to the reference, power down irradiation”.

[0202] In some embodiments of the ophthalmic stimulator **100**, the safety feedback system **116** can include at least one of an infrared camera and a thermal sensor **116b**, to sense a temperature of the treatment region, and to generate a feedback according to the sensed temperature. The feedback can include at least one of “sensed temperature is low relative to a reference, carry on irradiation”; “sensed temperature is medium relative to the reference, start power down of irradiation”; and “sensed temperature is high relative to the reference, stop irradiation”.

[0203] In some embodiments of the ophthalmic stimulator **100**, the safety feedback system **116** can include at least one of an alignment system **116c**, an eye tracker **116d**, an iris scanner **116f**, and an imaging system **114**, to sense an alignment of one of the iris **11** and the pupil **13** with the irradiation delivery system **130**. The feedback can include at least one of: “the irradiation delivery system is aligned with one of the iris and the pupil, carry on irradiation”; and “the irradiation delivery system is misaligned with one of the iris and the pupil, stop irradiation, and realign at least one of the irradiation delivery system, the iris, and the pupil”.

[0204] With reference to FIG. 8B, in some embodiments of the ophthalmic stimulator **100**, the irradiation source **120** can include a laser source **120L** that can be a continuous wave laser, a pulsed laser, or a scanned laser. A wide variety of laser sources are known in the art that are useful for ophthalmic applications and can be employed as the irradiation laser source **120L**. In particular, femtosecond lasers, with pulse length of 1-1,000 femtosecond and frequencies of

1 kHz. 1 MHz have been widely used in the art, and can be used for the laser source **120**. Other lasers with longer pulse lengths, in the picosecond or even nanosecond range can be utilized as well.

[0205] In some embodiments of the ophthalmic stimulator **100**, the irradiation source **120**, the irradiation delivery system **130**, and the irradiation control system **110** can be configured so that the patterned irradiation causing the long-term constriction of the pupil includes cauterizing iris tissue. Such embodiments can cause permanent alteration of the iris tissue, and therefore can cause permanent shrinking of the pupil.

[0206] FIGS. 13A-C illustrate a class of mobile implementations of the ophthalmic stimulator **100m**, indicated by the label “m”. Some of these embodiments will be referred to as a mobile ophthalmic stimulator **100m**. FIG. 13A illustrates that this class of embodiments can include a mobile irradiation control system **110m**, to generate an irradiation control signal; an irradiation source **120m** coupled to the mobile irradiation control system **110m**, to generate an irradiation **200**; and an irradiation delivery system **130m**, coupled to the mobile irradiation control system **110m**, to receive the irradiation from the mobile irradiation source **120m**, and to deliver a patterned irradiation **200p** to an iris of the eye. In embodiments, the mobile irradiation control system **110m** can control at least one of the irradiation source **120m** and the irradiation delivery system **130m** with the irradiation control signal so that the patterned irradiation **200p** causes a temporary constriction of the pupil of the eye, without causing a permanent constriction of the pupil.

[0207] In embodiments, the mobile irradiation control system **110m** can include a mobile communication platform **111m**, or simply mobile platform **111m** that can be a mobile telephone **111m**, a mobile communication device, and a mobile tablet; and a mobile irradiation controller **110cm**, installed on the mobile communication platform **111m**, to generate the irradiation control signal. In a characteristic embodiment, the mobile irradiation controller **110cm** can be a software application, downloaded from a provider over the internet and installed or implemented on a mobile phone **111m**. In other embodiments, the mobile irradiation controller **110cm** can be a dedicated processor, for example, in a separate box that can be installed on the mobile communication platform **111m** by plugging it into the mobile communication platform **111m** through a USB port, headphone jack, or charging port. For brevity, the mobile irradiation controller **110cm** is sometimes simply referred to as irradiation controller **110cm**, where the “m” label indicates the mobile nature of this irradiation controller. The mobile phone **111m** itself then can be attached to the remainder of the mobile ophthalmic stimulator **100m**, which can be a table-top system that includes the mobile irradiation source **120m**, and the mobile irradiation delivery system **130m**, installed either in an office of an ophthalmologist, or in a user's residence. In some embodiments, the mobile phone **111m** can be coupled to the rest of the ophthalmic stimulator **100m** by an electric connector or docking station. In other embodiments, the coupling and communication between the mobile phone **111m** and the rest of the ophthalmic stimulator **100m** can be a wireless communication, for example through a Bluetooth, or a wi-fi system or channel.

[0208] The mobile communication platform **111m** can include a memory, to store the above mentioned software

implementation of the mobile irradiation controller **110cm**; a processor, to execute the stored software implementation of the mobile irradiation controller **110cm**; and a user interface, to receive input from a user in relation to an operation of the memory and the processor.

[0209] Once the mobile platform **111m** or mobile phone **111m** is coupled to the rest of the mobile ophthalmic stimulator **100m**, a calibration process can be carried out, so that the mobile irradiation control system **110m** acquires information about the type and characteristics of the rest of the mobile ophthalmic stimulator **100m**. For example, information regarding the power and type of the light beam **200** generated by the irradiation source **120m**, and information regarding the type of signaling, communication and control protocols needed for the communication between the mobile platform **111m** and the rest of the mobile ophthalmic stimulator **100m**.

[0210] The irradiation delivery system **130m** can include at least one of a pattern generator, an optical beam shaper, a patterning optics, a beam profiler, and a digitally controlled irradiation optics. As in the other related embodiments, the mobile ophthalmic stimulator **100m** can be configured to increase a temperature of a treatment region of the iris to a range of 45-60 degrees Celsius.

[0211] The mobile irradiation control system **110m** can include a mobile imaging system **114m**, such as a mobile camera **114m**, to generate the irradiation control signal by generating an image of the iris of the eye by the mobile imaging system **114m**, receiving an image-based input, and generating the irradiation control signal to control at least one of the irradiation source **120m** and the irradiation delivery system **130m** to deliver the patterned irradiation according to the received image-based input.

[0212] In a characteristic example, the mobile irradiation control system **110m** can include a mobile phone **111m** that can be attached to the rest of the ophthalmic stimulator **100m** that is installed in a medical office as a desktop office device. As such, the irradiation source **120m** and the irradiation delivery system **130m** can themselves be a movable, light bench-top device that is mobile, but less mobile than the fully mobile platform **111m**, or mobile phone **111m**. Accordingly, in some embodiments they can be referred to as the mobile irradiation source **120m**, and the mobile irradiation delivery system **130m**.

[0213] The mobile camera **114m** of the mobile phone **111m** can image the iris **11** and pupil **13** of a patient who is looking into the camera **114m**. The mobile irradiation controller **110cm**, implemented on the mobile phone **111m**, can display the image of the pupil on the screen of the mobile phone **111m**, and also electronically overlay a proposed irradiation pattern **210**. The irradiation control application can then invite the doctor, or user, to modify the pattern within some limits of safety as the image-based input, such as to move the inner and the outer radii $R_p(\text{inner})$ and $R_p(\text{outer})$, while making sure that the pattern **210** remains on the iris **11**. Once the modification input is received, possibly together with some treatment parameters, the irradiation control application on the mobile phone **111m** can send an irradiation control signal to the irradiation source **120m** and the irradiation delivery system **130m** wirelessly with a Bluetooth channel. In response, the irradiation source **120m** and the irradiation delivery system **130m** can generate and deliver the patterned irradiation **200p** onto the imaged iris **11**.

[0214] FIG. 13A illustrates that in some embodiments, the mobile irradiation control system **110m** can include an image processor **114ipm**, to receive the image of the iris from the imaging system **114m**, and to generate the image-based input based on a processing of the image of the iris. In some designs, this image processor **114ipm** can determine the inner and outer radii $R_{p(inner)}$ and $R_{p(outer)}$ of the ring pattern **210**, as well as the treatment parameters. There can be hybrid systems, where the image processor **114ipm** performs the above determinations, however, a user interface **118** of the mobile telephone **111m** still prompts a surgeon or operator to approve the displayed choices of the imager processor **114ipm**, as a safety measure.

[0215] In some implementations, the image processor **114ipm** can generate the image-based input by correlating an alignment pattern **138** with the generated image of the iris, in analogy to the alignment system **135** in FIG. 12C. Subsequently, the mobile irradiation control system **110m** can be configured to generate the irradiation control signal according to the received image-based input that includes a misalignment-warning signal, an alignment-guidance signal, or an irradiation-stop signal, if a misalignment is detected. In a characteristic case, the mobile phone **111m** can alert the ophthalmologist that a misalignment was detected, possibly also generating an alignment-guidance signal, such as which way to move the eye **1**, or the irradiation delivery system **130** to realign the eye and the irradiation delivery system **130**.

[0216] The ability of the mobile platform **111m** to communicate can play a very useful role in some implementations. In these designs, the mobile irradiation control system **110m** can include an on-board communication application, to receive the image of the iris from the imaging system **114m**, to communicate the received image to a central station **410** having an image processor, and to receive the image-based input from image processor of the central station **410**.

[0217] FIGS. 13B-C illustrate advanced embodiments, where not the mobile phone **111m** is attached to the rest of the ophthalmic stimulator **100m**, but the rest of the ophthalmic stimulator **100m** is attached to the mobile phone **111m**, to create a fully mobile ophthalmic stimulator **100m**.

[0218] FIG. 13B illustrates a design of the mobile ophthalmic stimulator **100m**, wherein the irradiation source **120m** and the irradiation delivery system **130m** are part of a small, compact irradiation device **120m/130m**; and the mobile irradiation control system **110m** is coupled to the irradiation device **120m/130m** to send the irradiation control signal by at least one of an electronic coupling, an electric coupling, a wireless coupling, and an optical coupling.

[0219] In the shown example, the irradiation source **120m** and the irradiation delivery system **130m** are configured to be electrically coupled to, and mechanically attached to the mobile irradiation control system **110m**. For example, the irradiation device **120m/130m** can be plugged into one of the ports of the mobile phone **111m**, such as into the USB port, or into the headphone jack, or the power charging port. In another example, the irradiation device **120m/130m** can be attached to the mobile phone **111m** by a clip, mini-pliers, or pincer.

[0220] FIG. 13C illustrates an example, in which the irradiation source **120m** is a light source of the mobile irradiation control system **110m**, and the irradiation delivery system **130m** is mechanically attached to the mobile irra-

diation control system **110m**, to receive a light, generated by the irradiation source **120m**. In a characteristic example, the flashlight of the mobile phone **111m** itself can be used as the irradiation source **120m**. The flashlight of the mobile phone **111m**, of course, needs to be calibrated to gain control over the power irradiated by its irradiation **200**, and possibly filtered or dampened. Nevertheless, using the imaging capabilities of the mobile phones **111m** and their flashlight can make the mobile ophthalmic stimulator **100m** much cheaper and compact, and therefore suitable for being carried by a patient as a personal accessory. This aspect can be very useful if irradiation treatments are utilized that cause a temporary constriction of the pupil that lasts less than a day, and thus a once-a-day application in the morning does not secure the pupil constriction for the entire day. Such treatments may need to be refreshed as the day goes on. A portable, personalized, mobile phone-based ophthalmic stimulator **100m** can be the answer for the need for refreshing treatments during the day.

[0221] Obviously, safety is a high priority consideration for the mobile embodiments of the stimulator **100m** that are not operated by trained ophthalmologists. Moreover, achieving and preserving alignment for the duration of the treatment also becomes an elevated challenge for mobile stimulator **100m**. Mobile stimulators **100m** can address these concerns, by practicing the method, or process **540**, illustrated in FIG. 16D. It is recalled here, that in step **543**, if the imaging system **114**, the alignment system **116c**, or the eye tracker **116d** sense a misalignment, then they can induce the generation of a feedback-based irradiation control signal that makes the irradiation source **120m** stop the irradiation **200**. Implementing this imaging-triggered “safety stop” process makes mobile stimulators **100m** safe, and minimizes undesirable retinal exposure.

[0222] Moreover, if the eye gets realigned, for example, because the user moves either the hand-held mobile phone **111m**, or moves her/his gaze, then the imaging system **114**, or its equivalents, can sense the realignment in step **545**, and the irradiation controller **110m** can cause the restart the irradiation. These stop **543**—restart **545** steps can be performed repeatedly, as, for example, the handheld mobile phone **111m** is moving in the patient’s hand.

[0223] An interrupted, or multiply interrupted irradiation treatment may take longer to achieve the temperature rise required for the desired pupil constriction, and to administer the treatment for the time necessary for efficacy. Therefore, mobile stimulators **100m** can include at least one of a thermal camera, an infrared camera and a thermal sensor **116b**, to track an amount of time a treatment region of the iris had a temperature in a predetermined range. In an example, the irradiation controller **110m** may add up the multiply interrupted time-segments, when the treatment region of the iris was at the prescribed temperature, and ensure that the treatment region has been held at the prescribed temperature range for the time interval necessary to achieve the targeted pupil constriction. For example, the IR sensor **116b** can track that the treated ring **210r** of the iris **11** remains at 55 Celsius for a prescribed time, such as for 20 seconds, or for 40 seconds, in order to achieve a pupil constriction that will last all day.

[0224] In some embodiments, the safety stop **543**—restart **545** steps can be also performed under the control of the central station **410**. In such embodiments, it can be the image processor of the central station **410** that senses the misalign-

ment of the patterned irradiation relative to the iris or the pupil, as well as that senses the realignment, prompting the generation of the restart command.

[0225] Finally, the central station 410 can perform monitoring functions over a series of treatments performed by the mobile ophthalmic stimulator 100*m*. In some embodiments, the mobile stimulator 100*m* can be configured to take and send the image of the iris to the central station 410 for monitoring, to receive a monitoring-based control signal from the central station, and to generate the irradiation control signal in accordance with the received monitoring-based control signal. For example, the images, sent by the stimulator 100*m*, can be analyzed by the central station 410. This analysis can recognize that the treatment is inducing an undesirable effect in the retina over the term of several treatments. In such case, the central station may send out a monitoring-based control signal to the mobile stimulator 100*m* to either prevent the user from administering further treatments, or to change a treatment parameter, such as to reduce a power or intensity of the patterned irradiation 200*p*. Such central station-related systems are described next.

[0226] FIG. 14 illustrates a networked system 400 of ophthalmic stimulators for temporarily constricting eye-pupils. The networked system 400, or mobile network 400, can include a set of mobile ophthalmic stimulators 100*m*-1, 100*m*-2, . . . 100*m*-N, collectively referred to as mobile ophthalmic stimulators 100*m*-i, each mobile ophthalmic stimulator 100*m*-i including a mobile irradiation control system 110*m*-i, to generate an irradiation control signal; an irradiation source 120*m*-i, coupled to the irradiation control system 110*m*-i, to generate an irradiation; and an irradiation delivery system 130*m*-i, coupled to the mobile irradiation control system 110*m*-i, to receive the irradiation from the irradiation source 120*m*-i, and to deliver a patterned irradiation 200*p* to an iris of the eye; wherein the mobile irradiation control system 110*m*-i controls at least one of the irradiation source 120*m*-i and the irradiation delivery system 130*m*-i with the irradiation control signal so that the patterned irradiation 200*p* causes a temporary constriction of the pupil of the eye, without causing a permanent constriction of the pupil.

[0227] The networked system 400 further includes a central station 410, including a central image processor, wherein the mobile irradiation control systems 110*m*-i of the mobile ophthalmic stimulators 100*m*-i and the central station 410 are configured to communicate through a communication network. In this section, the term mobile ophthalmic stimulator 100*m*-i encompasses all embodiments described in relation to FIGS. 13A-C.

[0228] In embodiments of the networked system 400, each mobile irradiation control system 110*m*-i can include a mobile communication platform 111*m*-i, including at least one of a mobile telephone, a mobile communication device, and a mobile tablet; and a mobile irradiation controller 110*cm*-i, implemented on the mobile communication platform 111*m*-i, to generate the irradiation control signal. In embodiments, the mobile communication platforms 111*m*-i can include a memory, to store a software implementation of the mobile irradiation controller 110*cm*-i; a processor, to execute the stored software implementation of the mobile irradiation controller 110*cm*-i; and a user interface, to receive input from a user in relation to an operation of the memory and the processor. In embodiments, the mobile

ophthalmic stimulators can be configured to increase a temperature of a treatment region of the iris to a range of 45-60 degrees Celsius.

[0229] Each mobile irradiation control system 110*m*-i can include an imaging system 114*m*-i, to generate the irradiation control signal by generating an image of the iris of the eye by the imaging system 114*m*-i, receiving an image-based input, and generating the irradiation control signal to control at least one of the irradiation 120*m*-i source and the irradiation delivery system 130*m*-i to deliver the patterned irradiation 200*p* according to the received image-based input.

[0230] In some characteristic embodiments, the mobile irradiation control systems 110*m*-i can include an image processor 114*ipm*-i, to receive the image of the iris from the imaging system, and to generate the image-based input based on a processing of the image of the iris. In FIG. 14, the image processors 114*ipm*-i are not shown for scarcity of space. Embodiments of the image processors 114*ipm*-i have been shown and described earlier, such as in FIG. 6A. As described earlier, the on-board image processors 114*ipm*-i, can generate an image-based input for the irradiation control systems 110*m*-i, which then can control the rest of the mobile ophthalmic stimulators 100*m*-i accordingly. In such embodiments, the communications of the mobile ophthalmic stimulators 100*m*-i with the central station 410 can be a recording of the results of the image processing, and the record of the treatments performed by the mobile ophthalmic stimulators 100*m*-i.

[0231] In other embodiments, each ophthalmic stimulator 100*m*-i can be configured to send the image of the iris to the central station 410; and the central station 410 can be configured to analyze the received image by a central image processor 410*ip*, and to respond to the sending mobile ophthalmic stimulator 100*m*-i with the image-based input based on the analysis. This communication and analysis can be real-time, actionable. In other cases, it can be a post-treatment, recording the actions type communication.

[0232] In real-time embodiments, each ophthalmic stimulator 100*m*-i can be configured to generate and to send the image of the iris to the central station 410 before the irradiation delivery system delivers 130*m*-i the patterned irradiation to the iris; and the central station 410 can be configured to respond to the sending ophthalmic stimulator 100*m*-i with the image-based input that indicates whether the central station 410 authorizes the irradiation delivery system 130-i of the ophthalmic stimulator 100*m*-i to deliver the patterned irradiation to the iris.

[0233] Clearly, such preauthorization-based networked systems 400 have safety benefits, as when the patient intends to use the mobile ophthalmic stimulator 100*m*-i, the stimulator 100*m*-i first needs to send an image of the iris to be treated to the central station 410. This gives a chance for the central image processor 410*ip* to analyze the image of the iris, and if it finds anything that raises a medical concern, such as a shape change, or an unexpected discoloration, the central station 410 can communicate a "Treatment not authorized" imaging-based input to the mobile stimulator 100*m*-i, which then prevents the mobile stimulator 100*m*-i from irradiating the iris when medical concerns have been raised by the image analysis.

[0234] In a related embodiment of the networked system 400, each ophthalmic stimulator 100*m*-i can be configured to generate, and to send, the image of the iris to the central

station **410** before the irradiation delivery system **130m-i** delivers the patterned irradiation **200p** to the iris; and the central station **410** can be configured to respond to the sending mobile ophthalmic stimulator **100m-i** with the image-based input that indicates irradiation parameters to be used by the irradiation delivery system **130m-i** of the mobile ophthalmic stimulator **100m-i** when delivering the patterned irradiation to the iris.

[0235] The safety aspects of this embodiment are quite similar to the previous one. One of the differences is that the imaging-based input from the central station is not a binary “authorized-not authorized” input, but a quantitative input, nuanced input. In a characteristic example, the central image processor **410ip** can notice a small discoloration of the iris in the image, sent in by the mobile stimulator **100m-i**. However, the discoloration may be small enough so that a hard-stop “Treatment not authorized” input may be excessive. In such cases, the central image processor **410ip** can respond instead by a message of “Reduce power of irradiation in next treatment” input. In some embodiments, the central image processor **410ip** can even schedule a follow-up imaging, to check how the iris reacted to the reduced power irradiation: was the reduction sufficient to eliminate the discoloration, or further analysis is needed.

[0236] In some embodiments, the central imaging processor **410ip** of the central station **410** can be configured to perform a medical analysis of the image of the iris, and to respond to the sending ophthalmic stimulator **100m-i** with the image-based input that indicates if a negative medical condition was found by the analysis. The medical analysis can take place in a number of ways. The central station **410** can engage in an automated medical analysis, where for example past images of the iris, recalled from a memory, are compared to the present image. Or, the image of the iris can be compared to a database, compiled from tracking a large number of irises. Some embodiments can use artificial intelligence systems to recognize, and to evaluate the negative medical condition, such as an inflammation of the iris. Or, the image processor can flag the image, and request an opinion or analysis by a human specialist.

[0237] The negative medical condition can also be a wide range of conditions, including a change of color of the iris, a change of an optical characteristic, and a change of shape of the iris.

[0238] In some advanced embodiments, the mobile ophthalmic stimulators **100m-i** can be configured to test the iris **11** and to send a test result to the central station **410**; and the central station **410** can be configured to perform a medical analysis of the test result, and to respond to the sending mobile ophthalmic stimulators **100m-i** with the image-based input that indicates if a negative medical condition was found by the analysis. The mentioned test of the iris can include irradiating the iris with a test irradiation, and measuring a constriction of the pupil in response to the test irradiation. The performing a medical analysis can include recalling a previous test result, as mentioned. Finally, the detection of a negative medical condition can include comparing the test result with the previous test result, and finding the test result less acceptable than the previous test result. In other embodiments, the comparison can be made not with past measurements or tests on the same iris, but to a database of a large number of irises. This database can be organized

into groups according to many shared traits, so that patients with comparable medical situations and characteristics are compared by the database.

[0239] As mentioned in relation to the mobile stimulators **100m** of FIGS. 13A-C earlier, in another class of embodiments, the mobile ophthalmic stimulators **100m-i** can be configured to send alignment data to the central station **410** regarding an alignment of the patterned irradiation **200p** with at least one of the iris and the pupil; and the central station **410** can be configured to evaluate the alignment data; and to send a control signal to stop the patterned irradiation **200p** when the patterned irradiation **200p** is evaluated to be misaligned with at least one of the iris and the pupil. In some embodiments, the alignment data can be generated by the imaging system **114m**. In others, by various embodiments of the alignment system **135**, possibly using the patient interface **137** and the alignment pattern **138**. In imaging-based embodiments, the control signal can be analogous to the image-based input, described earlier.

[0240] Generally speaking, in some embodiments of the networked system **400** the mobile irradiation control systems **110m-i** of the mobile ophthalmic stimulators **100m-i** and the central station **410** can be configured to communicate regarding safety monitoring of the irradiations and treatments by an interface, or dedicated block or code **413**. This is a generic concept that encompasses communication regarding all major safety monitoring channels, including expected and unexpected medical outcomes, treatment parameters, proper alignment, and test results, from the viewpoint of safety. As described, the safety monitoring can result prompting a dedicated block, processor, or code **416** to signal or order preventive shutdowns of the mobile stimulators.

[0241] Analogous communications can be performed by a treatment outcome monitoring block, dedicated processor, or code **412**. Communications about treatment outcomes can then be used by a block, dedicated processor, or piece of code **415**, to develop and assemble a statistics of the treatment outcomes with the purpose of improving the understanding and the operations of the networked system **400** for the benefit of the patients. This communication channel can, of course, also be useful for pushing out new versions of treatment software from the central station **410** to the individual mobile stimulators **100m-i**.

[0242] These communications may not be real time, or actionable. In some embodiments, for example, the mobile irradiation control systems **110m-i** of the mobile ophthalmic stimulators **100m-i** and the central station **410** can be configured to communicate treatment outcomes after an irradiation has been performed. In other embodiments, they can be configured to communicate regarding patient data, which then can be stored in a dedicated processor and memory **411**.

[0243] While this document contains many specifics, these should not be construed as limitations on the scope of an invention or of what may be claimed, but rather as descriptions of features specific to particular embodiments of the invention. Certain features that are described in this document in the context of separate embodiments can also be implemented in combination in a single embodiment. Conversely, various features that are described in the context of a single embodiment can also be implemented in multiple embodiments separately or in any suitable subcombination. Moreover, although features may be described above as acting in certain combinations and even initially claimed as

such, one or more features from a claimed combination can in some cases be excised from the combination, and the claimed combination may be directed to a subcombination or a variation of a subcombination.

1. A networked system of mobile ophthalmic stimulators for temporarily constricting eye-pupils, comprising:

- a set of mobile ophthalmic stimulators, each mobile ophthalmic stimulator including
 - a mobile irradiation control system, to generate an irradiation control signal;
 - an irradiation source, coupled to the irradiation control system, to generate an irradiation; and
 - an irradiation delivery system, coupled to the mobile irradiation control system,
 - to receive the irradiation from the irradiation source, and
 - to deliver a patterned irradiation to an iris of the eye; wherein

the mobile irradiation control system controls at least one of the irradiation source and the irradiation delivery system with the irradiation control signal so that the patterned irradiation causes a temporary constriction of the pupil of the eye, without causing a permanent constriction of the pupil; and

a central station, wherein

the mobile irradiation control systems of the of the ophthalmic stimulators and the central station are configured to communicate through a communication network.

2. The networked system of claim 1, each mobile irradiation control system comprising:

- a mobile communication platform, including at least one of a mobile telephone, a mobile communication device, and a mobile tablet; and
- a mobile irradiation controller, installed on the mobile communication platform, to generate the irradiation control signal.

3. The networked system of claim 2, the mobile communication platforms comprising:

- a memory, to store a software implementation of the mobile irradiation controller;
- a processor, to execute the software implementation of the mobile irradiation controller; and
- a user interface, to receive input from a user in relation to an operation of the memory and the processor.

4. The networked system of claim 1, wherein:

the mobile ophthalmic stimulator is configured to increase a temperature of a treatment region of the iris to a range of 45-60 degrees Celsius.

5. The networked system of claim 1, the mobile irradiation control systems comprising:

- an imaging system, to generate the irradiation control signal by
 - generating an image of the iris of the eye by the imaging system,
 - receiving an image-based input, and
 - generating the irradiation control signal to control at least one of the irradiation source and the irradiation delivery system to deliver the patterned irradiation according to the received image-based input.

6. The networked system of claim 5, the mobile irradiation control systems comprising:

- an image processor,
 - to receive the image of the iris from the imaging system, and
 - to generate the image-based input based on a processing of the image of the iris.

7. The networked system of claim 5, wherein;

each mobile ophthalmic stimulator is configured to send the image of the iris to the central station; and the central station is configured

- to analyze the received image by a central image processor, and
- to respond to the sending ophthalmic stimulator with the image-based input based on the analysis.

8. The networked system of claim 7, wherein;

each mobile ophthalmic stimulator is configured to generate and to send the image of the iris to the central station before the irradiation delivery system delivers the patterned irradiation to the iris; and

the central station is configured to respond to the sending mobile ophthalmic stimulator with the image-based input that indicates whether the central station authorizes the irradiation delivery system of the ophthalmic stimulator to deliver the patterned irradiation to the iris.

9. The networked system of claim 7, wherein:

each mobile ophthalmic stimulator is configured to generate and to send the image of the iris to the central station before the irradiation delivery system delivers the patterned irradiation to the iris; and

the central station is configured to respond to the sending mobile ophthalmic stimulator with the image-based input that indicates irradiation parameters to be used by the irradiation delivery system of the ophthalmic stimulator when delivering the patterned irradiation to the iris.

10. The networked system of claim 7, wherein:

the central imaging processor of the central station is configured

- to perform a medical analysis of the image of the iris, and
- to, respond to the sending mobile ophthalmic stimulator with the image-based input that indicates if a negative medical condition was found by the analysis.

11. The networked system of claim 10, wherein:

the medical analysis includes recalling a previous image of the imaged iris from a memory; and

the negative medical condition is one of a change of color of the iris, a change of an optical characteristic, and a change of shape of the iris.

12. The networked system of claim 5, wherein:

the mobile ophthalmic stimulators are configured to test the iris and to send a test result to the central station; and

the central station is configured

- to perform a medical, analysis of the test result, and
- to respond to the sending mobile ophthalmic stimulators with the image-based input that indicates if a negative medical condition was found by the analysis.

13. The networked system of claim 12, wherein:
the test, of the iris includes
irradiating the iris with a test irradiation, and
measuring a constriction of the pupil in response to the
test irradiation;
the performing a medical analysis includes
recalling a previous test result; and
the detection of a negative medical condition includes
comparing the test result with the previous test result,
and
finding the test result less acceptable than the previous
test result.

14. The networked system of claim 1, wherein:
the mobile ophthalmic stimulators are configured to send
alignment data to the central station regarding an
alignment of the patterned irradiation with at least one
of the iris and the pupil and

the central station is configured
to evaluate the alignment data; and
to send a control signal to stop the patterned irradiation
when the patterned irradiation is evaluated to be
misaligned with at least one of the iris and the pupil.

15. The networked system of claim 1, wherein:
the mobile irradiation control systems of the mobile
ophthalmic stimulators and the central station are con-
figured to communicate treatment outcomes.

16. The networked system of claim 1, wherein:
the mobile irradiation control systems of the mobile
ophthalmic stimulators and the central station are con-
figured to communicate regarding safety monitoring of
treatments.

17. The networked system of claim 1, wherein:
the mobile irradiation control systems of the mobile
ophthalmic stimulators and the central station are con-
figured to communicate regarding patient data.

* * * * *

专利名称(译)	用于非瞳孔收缩的移动通信平台的网络系统		
公开(公告)号	US20180103839A1	公开(公告)日	2018-04-19
申请号	US15/375149	申请日	2016-12-11
[标]申请(专利权)人(译)	ZIMANYI GERGELY†		
申请(专利权)人(译)	KURTZ, RONALD MICHAEL ZIMANYI, GERGELY T.		
当前申请(专利权)人(译)	KURTZ, RONALD MICHAEL ZIMANYI, GERGELY T.		
[标]发明人	KURTZ RONALD MICHAEL ZIMANYI GERGELY T		
发明人	KURTZ, RONALD MICHAEL ZIMANYI, GERGELY T.		
IPC分类号	A61B3/00 A61B3/14 A61N5/06 A61B5/00		
CPC分类号	A61B3/0008 A61B3/14 A61B3/0025 A61N5/0625 A61B2576/00 A61B5/0022 A61B5/6898 A61B2560/0431 A61B5/0013 A61F2009/00876		
外部链接	Espacenet USPTO		

摘要(译)

用于暂时收缩眼睛瞳孔的眼科刺激器的联网系统包括一组眼科刺激器，每个眼科刺激器包括移动照射控制系统，以产生照射控制信号；照射源，与照射控制系统连接，产生照射；和照射传送系统，连接到移动照射控制系统，以接收来自照射源的照射，并将图案化的照射传递到眼睛的虹膜；其中，所述移动照射控制系统利用所述照射控制信号控制所述照射源和所述照射传送系统中的至少一个，使得所述图案化照射引起所述眼睛的瞳孔的暂时收缩，而不引起所述瞳孔的永久性收缩；和中心站，包括中央图像处理器，其中移动辐射控制系统和中心站被配置为通过通信网络进行通信。

