(19)	Europäisches Patentamt European Patent Office Office européen des brevets	(11) EP 1 582 150 A1				
(12)	EUROPEAN PATE					
(43)	Date of publication: 05.10.2005 Bulletin 2005/40	(51) Int CI. ⁷ : A61B 8/00 , G01S 15/89, G06T 15/00				
(21)	Application number: 05000206.2					
(22)	Date of filing: 07.01.2005					
(84)	Designated Contracting States: AT BE BG CH CY CZ DE DK EE ES FI FR GB GR HU IE IS IT LI LT LU MC NL PL PT RO SE SI SK TR Designated Extension States: AL BA HR LV MK YU	 Kwak, Byoung Joo Gwangjin-gu Seoul 143-140 (KR) Kim, Nam Chul Suseong-gu Daegu 706-811 (KR) Kim, Sang Hyun Yangsan Campus Yangsan-si Gyeongsangnam-do 626-847 (KR) 				
(30)	Priority: 01.04.2004 KR 2004022567	(74) Representative: Schmid Wolfgang				
(71)	Applicant: MEDISON CO., LTD. Kangwon-do 250-870 (KR)	Lorenz & Kollegen Patent- und Rechtsanwaltskanzlei Alte Ulmer Strasse 2				
(72) •	Inventors: Shim, Jaeyoon Seongdong-gu Seoul 133-769 (KR)	D-89522 Heidenheim (DE)				

(54) Apparatus and method for forming a 3D ultrasound image

(57) The present invention relates to a 3D ultrasound diagnostic forming a 3D ultrasound image only with volume data exiting within contour by automatically detecting the contour of a target object, comprising: a first unit for generating a region of interest (ROI) box on a 2D ultrasound image; a second unit for detecting a contour of a target object in the ROI box; and a third unit for forming a 3D ultrasound image by rendering volume data existing in the detected contour.





Description

5

30

[0001] The present invention generally relates to a 3-dimensional (3D) ultrasound diagnostic system, and more particularly to an apparatus and a method for automatically detecting a contour from a 2-dimensional (2D) ultrasound image of a target object and forming a 3D ultrasound image with volume data within the contour.

- **[0002]** A 3-dimensional (3D) ultrasound diagnostic system is a medical equipment for providing clinical information such as spatial information, anatomical information and the like, which cannot be provided from a conventional 2-dimensional image. The 3D ultrasound diagnostic system acquires volume data from signals received from a target object through a probe, and performs a scan conversion process for the acquired volume data. A 3D ultrasound image
- 10 of the target object is displayed on a display device such as a monitor, a screen or the like by performing a rendering process upon images obtained from the scan-converted data. This is so that a user can obtain clinical information of the target object.

[0003] As is well-known in the art, the probe typically has a plurality of transducers, wherein the respective timing of inputting pulse signals to each transducer is appropriately delayed. This is so that a focal ultrasound beam is transmitted

- ¹⁵ into the target object along a transmit scan line. Each transducer receives echo signals reflected from a focal point on the transmit scan line in a different reception time and converts the echo signal to reception signals of an electrical signal. The reception signals are transmitted to a beam former. The reception signals are appropriately delayed, wherein the delayed reception signals are summed in the beam former. This is so that the reception focal beam representing an energy level reflected from the focal point on the transmit scan line is outputted. Until a 2D slice image of the target
- 20 object formed by the reception focal beams for a plurality of scan lines is generated, the above process is repeatedly carried out.

[0004] A volume data acquisition unit outputs the volume data by synthesizing 2D ultrasound images, which represent sectional planes of the target object, inputted from the beam former. The volume data are generated from signals reflected from the target object existing in a 3D space and defined in torus coordinates. Therefore, in order to perform

²⁵ a rendering process for the volume data in a display device having Cartesian coordinates (e.g., monitor, screen and the like), the scan conversion for performing coordinate conversion of the volume data is required. The scan conversion is implemented in a scan converter.

[0005] Scan-converted volume data in the scan converter are rendered through a typical volume rendering process so that the 3D ultrasound image is displayed. The user obtains clinical information of the target object through the 3D ultrasound image displayed on the display device.

- **[0006]** The 3D ultrasound diagnostic system is primarily utilized for displaying a shape of a fetus with the 3D ultrasound image in the fields of obstetrics and gynecology. After acquiring volume data by scanning an abdominal region of a pregnant woman, the volume rendering process is performed upon the acquired volume data. This is so that the shape of the fetus can be displayed with the 3D ultrasound image. However, since the volume data includes mixed
- ³⁵ data of uterus tissue, adipose tissue, amniotic fluid, floating matters and the fetus, if the rendering process is directly applied to the volume data, it is difficult to clearly display the shape of the fetus with the 3D ultrasound image. Therefore, in order to display the shape of the fetus with the 3D ultrasound image, it is required to segment the fetus region from neighboring regions such as the amniotic fluid and the like.
- [0007] Accordingly, through the use of external interface devices (e.g., a mouse, a keyboard and the like) connected to the 3D ultrasound diagnostic system, a region of interest (ROI) box enclosing the shape of a fetus in a 2D ultrasound image, which is displayed on the display device, can be generated as illustrated in Fig. 1A. Thereafter, a final ROI box is generated by finely operating the external interfaces as illustrated in Fig. 1B. The volume data existing in a contour detected from an image in the ROI box are rendered such that the 3D ultrasound image of the fetus can be displayed.
 [0008] However, since the generation of ROI box generation and the detection of contour for the target object image
- ⁴⁵ are manually operated by the user in the 3D ultrasound diagnostic system, the quality of the finally displayed 3D ultrasound image depends on the expertise of the user. That is, the size of the ROI box is not consistent according to the user generating the ROI box. As such, there is often a problem since a desired 3D ultrasound image of the target object cannot be accurately displayed.
- [0009] Also, even if the user is an expert, there is a problem in that it takes a long working time to generate the ROI box and detect the contour of the target object image from the 2D ultrasound image. This is because the user generates the ROI box directly on the 2D ultrasound image. Moreover, if the size of the ROI box is not accurate for the desired 3D ultrasound image of the target object, then there is a problem in that an error may be generated in the volume data rendering process or the contour detection process of the target object in the ROI box.
- [0010] Therefore, it is an objective of the present invention to provide an apparatus and a method for forming an accurate 3D ultrasound image of a target object while reducing errors, which may be generated in a rendering process of a volume data and a contour detection process of the target object. It is a further objective of the present invention to reduce the time consumed in the process of generating a region of interest (ROI) box and the process of detecting contour of the target object by automatically generating the ROI box and detecting the contour of the target object

existing in the ROI box.

5

15

20

[0011] In accordance with an aspect of the present invention, there is provided an apparatus for forming a 3-dimensioanl (3D) ultrasound image, comprising: a first unit for generating a region of interest (ROI) box on a 2D ultrasound image; a second unit for detecting contour of a target object in the ROI box; and a third unit for forming a 3D ultrasound image by rendering volume data existing in the detected contour.

[0012] In accordance with another aspect of the present invention, there is provided a method for forming a 3-dimensional (3D) ultrasound image, comprising the steps of: a) generating a region of interest (ROI) box on a 2D ultrasound image; b) detecting contour of a target object in the ROI box; and c) forming a 3D ultrasound image by rendering volume data existing in the detected contour.

¹⁰ **[0013]** The above and other objects and features of the present invention will become apparent from the following description of preferred embodiments given in conjunction with the accompanying drawings, in which:

Figs. 1A and 1B are diagrams depicting region of interest (ROI) boxes generated on 2D ultrasound images;

Fig. 2 is a schematic block diagram illustrating a device for forming a 3D ultrasound image, which is constructed in accordance with the preferred embodiment of the invention;

Fig. 3 is a detailed block diagram illustrating a ROI box generation unit of Fig. 2;

Figs. 4A to 4C represent examples of showing the results of performing image segmentation for a 2D ultrasound image;

Figs. 5A to 5F show a process for automatically adjusting a size of the ROI box for a fetal face image;

Fig. 6A illustrates a mask used for determining the capability of the contour detection for the target object image; Fig. 6B is a diagram showing a 2D ultrasound image applying the mask;

Figs. 7A and 7B show examples of performing smoothing and brightness contrast for the target object image in the ROI box;

Figs. 8A and 8B show examples of performing binarization for the target object image in the ROI box; and

²⁵ Figs. 9A and 9B show examples of detecting contour at a top surface of a fetal image.

[0014] Fig. 2 is a schematic block diagram showing a 3-dimensional (3D) ultrasound image forming device 3 in a 3D ultrasound diagnostic device constructed in accordance with the preferred embodiment of the invention. The 3D ultrasound image forming device 3 includes a region of interest (ROI) box generation unit 31, a first determination unit 33,

³⁰ a second determination unit 35, a contour detection unit 37 and a 3D image processing unit 39. As the 3D key of a control panel (not shown) mounted in the 3D ultrasound image forming device 3 is activated, the 3D ultrasound image forming device 3 starts to operate.

[0015] If a user activates the 3D key, then one of the 2D ultrasound images, which are acquired through a probe and a beam former in the 3D ultrasound diagnostic system, is displayed with a brightness mode (B-mode) on a display

- ³⁵ device (not shown). In accordance with the present invention, the 2D ultrasound image displayed on the display device is a 2D ultrasound image representing a central slice of the target object. The B-mode represents that energies of signals reflected from the target object are displayed with a brightness level. The ROI box generation unit 31 automatically generates the ROI box on the 2D ultrasound image displayed on the display device.
- [0016] The first determination unit 33 determines whether the size of the ROI box, which is automatically generated, is suitable for that of the 2D ultrasound image of the target object. The second determination unit 35 determines whether the contour of the target object, which exists in the ROI box, can be detected. The contour detection unit 37 detects the contour of the target object existing in the ROI box. The 3D image processing unit 39 selects the volume data, which exist in the contour of the target object, among volume data stored in a volume data acquisition unit and forms a 3D ultrasound image by rendering them.
- ⁴⁵ **[0017]** Hereinafter, the ROI box generation unit 31 for automatically generating the ROI box on the 2D ultrasound image, which is displayed on the display device of the 3D ultrasound diagnostic system, will be described in detail by referring to Figs. 3 to 5F.

[0018] Fig. 3 is a detailed block diagram showing the ROI box generation unit 31 of Fig. 2. As shown in Fig. 3, the ROI box generation unit 31 includes an image segmentation unit 311, a ROI box setting unit 313 and a ROI box

- ⁵⁰ adjustment unit 315. The image segmentation unit 311 segments the 2D ultrasound image into a target object image and a background image neighbored with the target object. The ROI box setting unit 313 sets a ROI box having a predetermined size on the 2D ultrasound image, while the ROI box adjustment unit 315 adjusts the size of the ROI box. [0019] Since some factors, which make it difficult to clearly display the 3D ultrasound image (e.g. a speckle noise and the like), exist in the 2D ultrasound image for forming the volume data of the target object, a process for removing
- ⁵⁵ such factors should be first carried out. The image processing unit 311 removes the speckle noise existing in the 2D ultrasound image through a filtering process. The filtering process is carried out by using a Lee filter in order to remove only the speckle noise, while preserving edge information of the target object within the 2D ultrasound image, in accordance with the preferred embodiment of the present invention.

[0020] The image segmentation unit 311 sets a threshold value for binarization of the 2D ultrasound image in which the speckle noise has been removed. The 2D ultrasound image is binarized by referring to the threshold value so that the 2D ultrasound image can be segmented into two classes. Since the 2D ultrasound image is displayed with the B-mode, the threshold value is set according to the brightness of the 2D ultrasound image. For example, if the threshold value of the 2D ultrasound image having pixel values of 0 to 255 is 't', the 2D ultrasound image is segmented into a

- value of the 2D ultrasound image having pixel values of 0 to 255 is 't', the 2D ultrasound image is segmented into a first class having pixel values of {0, 1, 2, ..., t} and a second class having pixel values of {t, t+1, t+2, ..., 255}.
 [0021] The image segmentation unit 311 performs a first segmentation process for the 2D ultrasound image based on a first binarization threshold value t1. This is so that the 2D ultrasound image can be segmented into a target object image and a background image. A second segmentation process for the background image, which is segmented
- through the first segmentation process, is performed by referring to a second threshold value t2. This is so that the background image can be segmented into a target object image and a background image. Thereafter, a third segmentation process for the background image, which is segmented through the second segmentation process, is performed by referring to a third threshold value t3 so that the background image is finally segmented into a target object image and a background image. As these segmentation processes are repeatedly performed, the image segmentation unit
- ¹⁵ 311 more clearly segments the 2D ultrasound image into the target object image and the background image neighboring the target object image. The resulting values segmenting the image in the image segmentation unit 311 are outputted to the ROI box adjustment unit 315.

[0022] Fig. 4A shows an original 2D ultrasound image. Fig. 4B shows a 2D ultrasound image in which the image segmentation process is performed once. Fig. 4C shows a 2D ultrasound image in which the image segmentation process is performed twenty times.

20

[0023] The ROI box setting unit 313 sets the size of the ROI box configured with four bounds generated on the 2D ultrasound image originally displayed on the display device and outputs the set ROI box to the ROI box adjustment unit 315. Generally, the size of the ROI box is previously set according to the types of target objects to be displayed with the 3D ultrasound image.

- **[0024]** The ROI box adjustment unit 315 adjusts the size of the ROI box set by the ROI box setting unit 313 so as to be suitable for the size of the target object image exiting in the ROI box. During this time, the characteristic of the target object image becomes an important factor for adjusting the size of the ROI box. A fetus is an example of the target object. A fetus ultrasound image has a valley at the boundary between a head and a body, wherein a front surface of the head, which is a surface of a face, has more curvedness than that of the body. Also, a characteristic approximately a surface is longer than the body.
- ³⁰ exists in which the face is longer than the body.
 [0025] Referring to Fig. 5, there will be described an example which automatically adjusts the size of the ROI box for the overall fetus to the size of the ROI box for the face in the ROI box adjustment unit 315.
 [0026] First, as shown in Fig. 4C, the ROI box adjustment unit 315 selects an image existing in the ROI box set by the ROI box setting unit 313 from the image-segmented 2D ultrasound image. As shown in Fig. 5A, a binarization
- ³⁵ threshold value is set for the selected fetal image, wherein a binarization process is applied to the selected fetal image by referring to the threshold value. Next, the ROI box adjustment unit 315 removes the noise regions from the binarization regions of the fetal image. This is so that the binarization regions, which are determined with a portion of the fetal image, are detected as shown in Fig. 5B. More specifically, the ROI box adjustment unit 315 examines the brightness for each binarization region and calculates the mean of the brightness. When the brightness of the binarization
- ⁴⁰ regions is lower than the mean of the brightness, the binarization region is considered as a noise region and thereby removed. Also, even if the brightness is greater than the mean of the brightness, the noise regions may exist. Therefore, the ROI box adjustment unit 315 sets a threshold value at the binarization region. As such, when a pixel value of the binarization region is lower than the threshold value, the binarization region is considered as a noise region and thereby removed.
- 45 [0027] After the noise regions are removed from the binarization regions, the ROI box adjustment unit 315 assigns a pixel value of "255" to the overall pixels existing from pixels positioned at a top surface of the detected binarization region to pixels positioned at a bottom bound of the ROI box set by the ROI box setting unit 313. This is so that a binarization image of the fetus can be generated as shown in Fig. 5C. Next, the ROI box adjustment unit 315 generates a surface curved line on the binarization image as illustrated in Fig. 5D. Then, the ROI box adjustment unit 315 searches
- ⁵⁰ relative maximum points and relative minimum points among pixels neighboring each other at the surface curved line and appoints the relative minimum point corresponding to the valley between the head and the body of the fetus. [0028] As shown in Fig. 5D, after the relative minimum point corresponding to the valley between the head and the body of the fetus is appointed as N2, the ROI box adjustment unit 315 appoints a starting point N1 and an ending point N3 of the surface curved line. This is so that the surface curved line is segmented into two areas of [N1, N2] and [N2,
- N3]. The ROI box adjustment unit 315 calculates the number of the relative maximum points existing at a first surface curved line area of [N1, N2] and a second surface curved line area of [N2, N3]. Further, an area having a greater number of relative maximum points is determined as the face surface of the fetus. Since the curvedness of the surface of the fetal face is greater than that of the fetal body, the number of relative maximum points at the fetal face is greater

than that of the fetal body.

5

[0029] After the face area of the fetus is determined, the ROI box adjustment unit 315 moves the left bound of the ROI box to the relative minimum point N2 of the surface curved line. It further moves the right bound of the ROI box to the right bound of the surface curved line. Therefore, the ROI box adjustment unit 315 automatically adjusts the positions of left/right bounds of the ROI box set by the ROI box setting unit 313 so as to be suitable for a face size of the fetal image. Fig. 5E shows an example wherein the left/right bounds of the ROI box set on the 2D ultrasound image displayed on the displayed o

displayed on the display device are automatically adjusted by the ROI box adjustment unit 315.
 [0030] After adjusting the left/right bounds of the ROI box for the fetal face image, the ROI box adjustment unit 315 performs the binarization, noise region removal, binarization image generation and surface curved line generation processes for the fetal face image existing in the adjusted ROI box. In order to generate a binarization image from the

- ¹⁰ processes for the fetal face image existing in the adjusted ROI box. In order to generate a binarization image from the fetal image, the ROI box adjustment unit 315 assigns a pixel value of "255" to the overall pixels existing from the left most surface of the binarization regions detected from the fetal face image to the right bound of the adjusted ROI box. [0031] The ROI box adjustment unit 315 moves the top bound of the ROI box toward the bottom bound of the ROI box until the top bound meets with the surface curved line. It then determines the position, which the top bound is met
- ¹⁵ with the surface curved line, as the top bound position of the ROI box. Also, the ROI box adjustment unit 315 moves the bottom bound of the ROI box toward the top bound of the ROI box until the bottom bound reaches the surface curved line. It then determines the position, which the bottom bound of the ROI box is met with the surface curved line, as the bottom bound position of the ROI box. Fig. 5F shows an example wherein the ROI box set on the 2D ultrasound image is finally adjusted for the fetal face image by the ROI box adjustment unit 315.
- 20 **[0032]** For the sake of convenience, while the ROI box generation process in the ROI box generation unit 31 is described for the fetal image (which is an example), it will be apparent that the size of a ROI box set on an arbitrary target object of a 2D ultrasound image can be adjusted to be suitable for the size of the arbitrary target object by automatically adjusting the bounds of the ROI box according to the above ROI box generation process.
- [0033] The first determination unit 33 illustrated in Fig. 2 determines the suitability of the size of ROI box for the target object image generated in the ROI box generation unit 31. Generally, the ROI box generated from the ROI box generation unit 31 has a similar size for the same type of target object image. Accordingly, the first determination unit 33 selects one of the standard histograms, which are previously set for each type of target object images, from a memory (not shown) built in the 3D ultrasound diagnostic system. Then, a square mean error value of the selected histogram and histograms for the target object image, which exists in the ROI box inputted from the ROI box generation unit 31, is calculated.

[0034] As a result, if the calculated square mean error value is greater than the previously predetermined value, the first determination unit 33 determines that the size of the ROI box generated from the ROI box generation unit 31 is not suitable and then stops the operation of the 3D ultrasound diagnostic system. Thereafter, it notifies that the size of the generated ROI box is not suitable to the user. On the other hand, if the calculated square mean error value is

- ³⁵ less than the predetermined value, the first determination unit 33 determines that the size of the ROI box generated from the ROI box generation unit 31 is suitable. It then outputs the 2D ultrasound image inputted from the ROI box generation unit 31 to the second determination unit 35. The histogram for the target object image in the ROI box, which is inputted from the ROI box generation unit 31, is transmitted to the memory so as to renew the standard histogram of the target object image.
- 40 [0035] The second determination unit 35 shown in Fig. 2 determines whether the contour of the target object image in the ROI box can be detected. That is, as the second determination unit 35 determines the capability of the contour detection of the target object image in the ROI box, it improves the efficiency for displaying the 3D ultrasound image in the 3D ultrasound diagnostic system. Since the absorption and reflection of the ultrasound are different according to each part of the target object (e.g., the fetus and the amniotic fluid enclosing the fetus), boundaries are formed in
- 45 the target object of the 2D ultrasound image due to an edge effect clearly representing the differences in brightness. The second determination unit 35 determines the capability of the contour detection for the target object on the basis of the edge effect.

[0036] Fig. 6A shows a mask used for determining whether the contour of the 2D ultrasound image can be detected. Fig. 6B shows a 2D ultrasound image applying the mask of Fig. 6A. A mask configured with pixels of 3×3 is commonly

- ⁵⁰ used so as to detect the boundaries of the target object in the 2D ultrasound image. In designing the mask for the 2D ultrasound image, (1) a pixel value corresponding to a dark portion is set with "-1", (2) a pixel value corresponding to a bright portion is set with "1", and (3) a pixel value forming the boundary between the dark portion and the bright portion is set with "0". In accordance with the preferred embodiment of the present invention, a mask configured with pixels of 5 × 5 expanding a pixel region forming the boundary is used as shown in Fig. 6A. The size of the pixel configuring the mask is identical to that of the pixel configuring the 2D ultrasound image.
- **[0037]** The second determination unit 35 detects pixels forming the boundary by matching a pixel positioned at a center of the mask (illustrated in Fig. 6A) with all pixels existing in a region "A" corresponding to a portion from the top bound of the ROI box to half of the fetal face image one-to-one. The reason why the detection process of pixels forming

the boundary is carried out only for such portion (from the top bound of the ROI box to half of the fetal face image) is to improve the detection speed and to prevent the edge region from being generated at undesired regions.

- [0038] More specifically, the central pixel of the mask of 5 × 5 is first matched with an arbitrary pixel existing in the region "A". Next, the second determination 35 selects the pixels in the region "A", which are distributed adjacent to the arbitrary pixel and matched with the pixels configuring the mask of 5 × 5 one to one. Thereafter, the second determination unit 35 multiplies the pixel values of the selected pixels by one pixel value of pixels configuring the mask of 5 × 5, i.e., "-1", "1" or "0", respectively. It then sums up the resulting values so that the second determination unit 35 finally determines the summed value as a new pixel value for the arbitrary pixel, which exits in the region "A" and is matched with the central pixel of the mask of 5 × 5. The second determination unit 35 applies the above process to
- 10 all the pixels existing in the region "A" and determines pixels forming the boundary of the target object with pixels having over a predetermined pixel value among the newly determined pixel values. In accordance with the preferred embodiment of the present invention, pixels corresponding to 20 % of a high rank among the newly determined pixel values are determined as reference pixels representing the edge pixels.
- [0039] After the pixels forming the boundary are detected, the second determination unit 35 calculates a ratio of the number of the boundary pixels to all the number of pixels configuring the region "A" and variance of pixels forming the boundary. Thereafter, the ratio and the variance are applied to the following equation 1 so that the second determination unit 35 determines the capability of the contour detection of the target object.

$$D_{i} = \alpha R_{1i} + \beta R_{2i}$$
 Eq. 1

[0040] Wherein, i is the number of 2D ultrasound images for the same types of target object inputted to the second determination unit 35, D_i is a determination numerical value representing the capability of the contour detection, R_{1i} is a ratio of the number of the boundary pixels to all the number of pixels configuring the region "A", and R_{2i} is a variance of pixels forming the boundary. α and β , which are coefficients of an equation such as equation 1, are obtained through a contour detection experiment of various 2D ultrasound images for the same type of target object.

[0041] In particular, the second determination unit 35 selects various 2D ultrasound images for the same type of target object from the memory built in the 3D ultrasound diagnostic system. The second determination unit 35 gives a determination value "1" to 2D ultrasound images whose contour can be detected. It further gives a determination value "0" to 2D ultrasound images whose contour cannot be detected among the selected 2D ultrasound images. The second determination unit 35 calculates α and β when the square mean error value, which is defined in the following equation 2, is minimized by using the determination values of each 2D ultrasound image selected from the memory, as well as

35

the ratios R_{1i} and R_{2i}.

30

20

25

$$\epsilon = \sum_{i} (D_i - (\alpha R_{1i} + \beta R_{2i}))^2$$
 Eq. 2

- ⁴⁰ **[0042]** In order to minimize the square mean error value, the second determination unit 35 performs the partial differentiation for equation 2 for α and β . In case the resulting values of the equation performing the partial differentiation become "0", the α and β are determined as the coefficients of equation 1. The second determination unit 35 applies the R_{1i}, R_{2i}, α and β to equation 1 so that the determination values of the target object image in the ROI box, which are inputted through the first determination unit 33, can be calculated. Finally, the second determination unit 35 determines that the contour of the target object image in the ROI box can be detected when the calculated determination
- ⁴⁵ mines that the contour of the target object image in the ROI box can be detected when the calculated determination value is greater than a predetermined set value. Then, the inputted 2D ultrasound image is outputted to the contour detection unit 37 (illustrated in Fig. 2). On the other hand, when the calculated determination value is less than the predetermined set value, the second determination unit 35 determines that it is impossible to detect the contour of the target object image in the ROI box and stops the operation of the 3D ultrasound diagnostic system. Thereafter, the
- ⁵⁰ second determination unit 35 notifies that it is impossible to detect the contour of the target object image to the user. [0043] For the sake of convenience, a process for determining the capability of the contour detection for the top surface of the fetal face image by detecting the boundary pixels existing between the top bound of the ROI box and the fetal face image is described. However, boundary pixels existing between the bottom, left and right bounds of the ROI box and the fetal face image should be detected in order to determine the capability of the contour detection of
- ⁵⁵ fetal face image according to the above process. For such process, masks transforming the pixel values of the mask of 5×5 of Fig. 6A should be used. That is, for the bottom bound of the ROI box, a mask should be used in which the pixel values positioned at the first row 42 of the mask of 5×5 (illustrated in Fig. 6A) are "-1" and the pixel values

positioned at the fifth row 44 are "1". For the left bound of the ROI box, a mask should be used in which the pixel values positioned at the first column 46 of the mask of 5×5 (illustrated in Fig. 6A) are "-1" and the pixel values positioned at the fifth column 48 are "1". Also, for the right bound of the ROI box, a mask should be used in which the pixel values positioned at the first column 46 of the mask of 5×5 (illustrated in Fig. 6A) are "1" and the pixel values positioned at the first column 46 of the mask of 5×5 (illustrated in Fig. 6A) are "1" and the pixel values positioned at the first column 46 of the mask of 5×5 (illustrated in Fig. 6A) are "1" and the pixel values positioned at the fifth column 48 are "-1".

[0044] Since the processes for determining the capability of the contour detection for the bottom, left and right surfaces of the fetal face image are performed upon the above process, a detailed description will be omitted herein. Also, it is apparent that the determination process for determining the capability of the contour detection, which is described above, can be applied to an arbitrary target object.

- 10 [0045] The contour detection unit 37 (illustrated in Fig. 2) detects the contour of the target object image exiting in the ROI box, which is inputted from the second determination unit 35. First, in order to compensate the deficiency of the target object not having a sufficient brightness contrast, the contour detection unit 37 performs the smoothing of the target object by removing the speckle noise existing in the target object in the ROI box through the use of the Lee filter or the like. Thereafter, brightness contrast stretching for the target object is carried out by providing pixel values,
- ¹⁵ which are defined in equation 3, to the pixels configuring the target object image in the ROI box.

20

5

 $F(x) = \begin{cases} 0 & \text{for } x \le low \\ \frac{255 \times (x - low)}{(high - low)} & \text{for } low < x < high \\ 255 & \text{for } high \le x \end{cases}$ Eq. 3

25

[0046] Wherein, F(x) is a new pixel value provided by the contour detection unit 37, x is an old pixel value of the pixel configuring the target object image inputted from the second determination unit 35, 'low' is a critical value in a low range of the histogram for the target object inputted from the second determination unit 35, and 'high' is a critical value in a high range of the histogram for the target object inputted from the second determination unit 35.

³⁰ **[0047]** Figs. 7A and 7B represent examples showing the results of performing the smoothing and brightness contrast stretching for the target object in the ROI box, which is previously set, in the contour detection unit 37. Hereinafter, a process for detecting the contour of the fetal image will be described.

[0048] The contour detection unit 37 sets a binarization threshold value for the fetal image in the ROI box and performs the binarization for the fetal image by the set binarization threshold value as a reference. Next, the contour detection

- ³⁵ unit 37 detects regions, which are determined as a portion of the fetal image, by removing the noise regions from the binarization regions of the fetal image. That is, the contour detection unit 37 examines the brightness values for the binarization regions and calculates the mean thereof. In the binarization regions, if the brightness value of the binarization region is less than the mean brightness value, the binarization region is considered as a noise region and thereby removed. Also, even if the brightness value is greater than the mean brightness value, since the noise region may
- 40 exist, the contour detection unit 37 sets a threshold value for the size of the binarization region. Therefore, if the size of the binarization region is less than the threshold value, the binarization region is considered as a noise region and thereby removed. Figs. 8A and 8B represent examples showing the results of performing the binarization for the fetal image in the ROI box and removing the noise regions from the binarization regions in the contour detection unit 37. [0049] Hereinafter, a contour detection process performed in the detection unit 37 will be described in view of Fig.
- 9. If the binarization regions for the fetal image are generated as shown in Fig. 8B, the contour detection unit 37 moves the top bound (not shown) of the ROI box toward the bottom bound (not shown) in order to detect the contour of the fetal image. Next, the contour detection unit 37 extracts pixels corresponding to top surfaces of the binarization regions, which met with the top bound of the ROI box, and connects the extracted pixels with the pixels neighboring each other. Since a deep valley between the head and the body of the fetal image exists, the contour for the top surfaces of the
- ⁵⁰ fetal image are obtained as illustrated in Fig. 9A. If the separated contours as illustrated in Fig. 9A are detected, the contour detection unit 37 connects the contours to thereby obtain a final contour. An example, which applies the detected contour to the 2D ultrasound image displayed on the display device, is illustrated in Fig. 9B.
 [0050] For the sake of convenience, the contour detection process for only the top surface of the fetal image is described. Also, the contour detection for the bottom, left and right surfaces can be detected by applying the above
- ⁵⁵ process to the bottom, left and right bounds. Finally, the contour detection unit 37 outputs the 2D ultrasound image in which the contour detected from the fetal image is displayed to the 3D image processing unit 39 (illustrated in Fig. 2). [0051] The 3D image processing unit 39 forms the 3D ultrasound image by rendering volume data only for the target object. For such process, the 3D image processing unit 39 selects the volume data corresponding to the target object.

which is enclosed by the contour, inputted from the contour detection unit 37 from the voltage data stored in the voltage data acquisition unit (not shown). Next, the 3D image processing unit 39 performs the scan conversion for the selected volume data and then a typical volume rendering process is applied so that the 3D ultrasound image can be more accurately displayed.

- 5 [0052] As described above, since the ROI box is automatically generated and the contour of the target object image existing in the ROI box is automatically detected, the time consumption for the ROI box generation and the contour detection of the target object image can be reduced. Also, as the suitability of the size of the automatically generated ROI box and the capability of the contour detection from the target object image in the ROI box are determined, errors generated from the volume data rendering and the contour detection of the target object image can be reduced. As
- 10 such, a more accurate 3D ultrasound image for the target object can be provided to the user of the 3D ultrasound diagnostic system.

[0053] While the present invention has been described and illustrated with respect to a preferred embodiment of the invention, it will be apparent to those skilled in the art that variations and modifications are possible without deviating from the broad principles and teachings of the present invention which should be limited solely by the scope of the

15 claims appended hereto.

Claims

- 20 1. An apparatus for forming a 3-dimensioanl (3D) ultrasound image, comprising:
 - a first means for generating a region of interest (ROI) box on a 2D ultrasound image;
 - a second means for detecting a contour of a target object in the ROI box; and
 - a third means for forming a 3D ultrasound image by rendering volume data existing in the detected contour.

25

30

35

45

50

- 2. The apparatus as recited in claim 1, further comprising a first determination means for determining whether a size of the ROI box is suitable for a size of the 2D ultrasound image of the target object.
- 3. The apparatus as recited in claim 2, further comprising a second determination means for determining whether the contour of the 2D ultrasound image of the target object is detectable.
 - 4. The apparatus as recited in claim 1, wherein the first means comprises:

a first unit for segmenting the 2D ultra sound image into an image of the target object and an image of a background;

a second unit for setting the ROI box with a predetermined size according to types of the target object on the 2D ultra sound image; and

a third unit for adjusting a size of the ROI box to be suitable for a size of the target object.

40 5. The apparatus as recited in claim 4, wherein the third unit performs the following functions:

> after performing binarization for the target object image in the ROI box, removing a noise region in the binarization region of the target object image;

generating a binarization image for the target object by providing predetermined pixel values to the binarization region and a surface curved line for the binarization image; and

- adjusting bounds of the ROI box by moving each bound to pixels closest to the surface curved line.
- 6. The apparatus as recited in claim 2, wherein the first determination means determines whether a size of the ROI box is suitable by calculating square mean error values of predetermined standard histograms for each type of 2D target object images of target object and a histogram for the target object image in the ROI box.
- 7. The apparatus as recited in claim 3, wherein the second determination means performs the following functions:
- extracting pixels forming boundary of target object by matching pixels of a mask of a predetermined type with 55 pixels of the target object image in the ROI box one to one; determining a capability of contour detection of the target object by applying a ratio of the number of the boundary pixels to all the number of pixels matched with the mask one to one and variance of pixels forming
 - the boundary according to following equation:

$$D_i = \alpha R_{1i} + \beta R_{2i}$$

wherein, i is the number of the 2D ultrasound image for the same type of target objects previously stored in a memory, D_i is a determination numerical value representing the capability of the contour detection, R_{1i} is a ratio of the number of the boundary pixels to all the number of pixels matched with the mask one to one, and R_{2i} is a variance of pixels forming the boundary, and wherein α and β , which are coefficients obtained through a contour detection experiment of various 2D ultrasound images for the same type of target objects, are values when a square mean error value defined in following equation is minimized:

10

5

$$\epsilon = \sum_{i} (D_i - (\alpha R_{1i} + \beta R_{2i}))^2$$

15

wherein, to minimize the square mean error value, a partial differentiation of the square mean error value is performed for each α and β , and wherein the α and β are determined as the coefficients when resulting values of the partial differentiation are "0".

20

30

45

8. The apparatus as recited in claim 1, wherein the second means performs the following functions:

performing brightness contrast stretching for the target object image for sufficient brightness contrast of the target object in the ROI box;

²⁵ after performing binarization of the target object in the ROI box, removing a noise region from a binarization region of the target object;

extracting pixels existing in a surface of the binarization region which meets the bounds of the ROI box by adjusting the bounds of the ROI box; and

- detecting the contour of the target object by connecting the extracted pixels.
- 9. A method for forming a 3-dimensional (3D) ultrasound image, comprising the steps of:
 - a) generating a region of interest (ROI) box on a 2D ultrasound image;
 - b) detecting a contour of a target object in the ROI box; and
- 35 c) forming a 3D ultrasound image by rendering volume data existing in the detected contour.
 - **10.** The method as recited in claim 9, further comprising the step of determining whether a size of the ROI box is suitable for a size of the 2D ultrasound image of the target object.
- 40 **11.** The method as recited in claim 10, further comprising the step of determining whether the contour of the 2D ultrasound image of the target object is detectable.
 - 12. The method as recited in claim 9, wherein the step a) comprises the steps of:
 - a1) segmenting the 2D ultrasound image into an image of the target object and an image of a background neighboring the target object;
 - a2) setting a ROI box with a predetermined size on the 2D ultrasound image; and
 - a3) adjusting the size of the ROI box to be suitable for a size of the 2D ultrasound image of the target object.
- ⁵⁰ **13.** The method as recited in claim 12, wherein the step a3) comprises the steps of:

a3-1) after performing binarization for the target object image in the ROI box, removing a noise region in the binarization region of the target object image;

- a3-2) generating a binarization image for the target object by providing predetermined pixel values to the binarization region and a surface curved line for the binarization image; and
 - a3-3) adjusting bounds of the ROI box by moving each bound to pixels closest to the surface curved line.
 - 14. The method as recited in claim 10, wherein the step d) comprises the step of determining a size of the ROI box

by calculating square mean error values of predetermined standard histograms for each type of 2D target object images of target object and a histogram for the target object image in the ROI box.

15. The method as recited in claim 11, wherein the step e) comprises the steps of:

e1) extracting pixels forming boundary of target object by matching pixels of a mask of a predetermined type with pixels of the target object image in the ROI box one to one; and
e2) determining a capability of contour detection of the target object by applying a ratio of the number of the boundary pixels to all the number of pixels matched with the mask one to one and a variance of pixels forming

the boundary according to following equation:

$$D_i = \alpha R_{1i} + \beta R_{2i}$$

wherein, i is the number of the 2D ultrasound image for the same type of target objects previously stored in a memory, D; is a determination numerical value representing the capability of the contour detection, R_{1i} is a ratio of the number of the boundary pixels to all the number of pixels matched with the mask one to one, and R_{2i} is a variance of pixels forming the boundary, and wherein α and β , which are coefficients obtained through a contour detection experiment of various 2D ultrasound images for the same type of target objects, are values when a square mean error value defined in following equation is minimized:

$$\epsilon = \sum_{i} (D_{i} - (\alpha R_{1i} + \beta R_{2i}))^{2}$$

wherein, to minimize the square mean error value, a partial differentiation of the square mean error value is performed for each α and β , and wherien the α and β are determined as the coefficients when resulting values of the partial differentiation are "0".

30

35

5

15

20

25

16. The method as recited in claim 9, wherein the step b) comprises the steps of:

b1) performing brightness contrast stretching for the target object image for sufficient brightness contrast of the target object in the ROI box;

b2) after performing binarization of the target object in the ROI box, removing a noise region from a binarization region of the target object;

b3) extracting pixels existing in a surface of the binarization region which meets the bounds of the ROI box by adjusting the bounds of the ROI box; and

b4) detecting the contour of the target object by connecting the extracted pixels.

40

45

50

Fig. 1A



Fig. 1B





Fig. 2

Fig. 3



Fig. 4A



Fig. 4B





Fig. 4C

Fig. 5A



Fig. 5B



Fig. 5C







Fig. 5E



Fig. 5F



Fig. 6A



Fig. 6B



Fig. 7A



Fig. 7B



Fig. 8A



Fig. 8B



Fig. 9A



Fig. 9B





European Patent Office

EUROPEAN SEARCH REPORT

Application Number EP 05 00 0206

-

	DOCUMENTS CONSID			
Category	Citation of document with ir of relevant passa	idication, where appropriate, ges	Relevant to claim	CLASSIFICATION OF THE APPLICATION (Int.CI.7)
X A	EP 1 083 443 A (MED 14 March 2001 (2001 * abstract * * paragraph [0014]	UISON CO., LTD) -03-14) - paragraph [0027];	1-3,9-12 4-8, 13-16	A61B8/00 G01S15/89 G06T15/00
x	EP 1 323 380 A (MED 2 July 2003 (2003-C * paragraph [0011] figures 2,3a,3b *	DISON CO., LTD) 17-02) - paragraph [0013];	1,9	
A	US 6 464 642 B1 (KA 15 October 2002 (20 * abstract * * column 6, line 35 figure 2 *	WAGISHI TETSUYA) 02-10-15) 5 - column 13, line 20;	1,9	
A	US 6 193 660 B1 (JA 27 February 2001 (2 * abstract * * column 2, line 16 figures 1,2A-D,3 *	CKSON JOHN I ET AL) 001-02-27) 5 - column 10, line 47;	1,9	TECHNICAL FIELDS SEARCHED (Int.Cl.7) A61B C01S
A	WO 00/32106 A (WAKE VINING, DAVID, J; H DAVID,) 8 June 2000 * page 3, line 10 - * page 7, line 28 - figures 1,3 *	FOREST UNIVERSITY; UNT, GORDON, W; AHN, (2000-06-08) page 5, line 7 * page 15, line 9;	1,9	GOGT
	The present search report has I Place of search Munich	Deen drawn up for all claims Date of completion of the search 18 July 2005	Art	Examiner ikis, T
CA X : parti docu A : tech O : non P : inter	CATEGORY OF CITED DOCUMENTS T : theory or principle particularly relevant if taken alone after the filing data particularly relevant if combined with another D : document of the same category L : document cited f cohnological background active of the same category the stremediate document the stremediate document the same category t		I a underlying the ir ument, but publis e the application or other reasons ume patent family,	vention hed on, or corresponding

ANNEX TO THE EUROPEAN SEARCH REPORT ON EUROPEAN PATENT APPLICATION NO.

This annex lists the patent family members relating to the patent documents cited in the above-mentioned European search report. The members are as contained in the European Patent Office EDP file on The European Patent Office is in no way liable for these particulars which are merely given for the purpose of information.

18-07-2005

	F cite	^p atent document ed in search report		Publication date		Patent family member(s)		Publication date
	EP	1083443	A	14-03-2001	KR EP JP US	2001026857 1083443 2001137241 2004213445	A A2 A A1	06-04-2001 14-03-2001 22-05-2001 28-10-2004
	EP	1323380	A	02-07-2003	KR EP JP US	2003058423 1323380 2003284717 2003135119	A A2 A A1	07-07-2003 02-07-2003 07-10-2003 17-07-2003
	US	6464642	B1	15-10-2002	JP	2001128975	A	15-05-2001
	US	6193660	B1	27-02-2001	NONE			
	WO	0032106	A	08-06-2000	US US AU EP WO US AU CA EP WO	6366800 6119574 1526300 1133257 0032106 2002193687 1307300 2335788 1117881 0003102	B1 A A1 A1 A1 A1 A1 A2 A2 A2	02-04-2002 19-09-2000 19-06-2000 19-09-2001 08-06-2000 19-12-2002 01-02-2000 20-01-2000 25-07-2001 20-01-2000
- ORM P0459								

For more details about this annex : see Official Journal of the European Patent Office, No. 12/82

patsnap

专利名称(译)	用于形成3D超声图像的设备和方法					
公开(公告)号	EP1582150A1	公开(公告)日	2005-10-05			
申请号	EP2005000206	申请日	2005-01-07			
申请(专利权)人(译)	MEDISON CO. , LTD.					
当前申请(专利权)人(译)	MEDISON CO., LTD.					
[标]发明人	SHIM JAEYOON KWAK BYOUNG JOO KIM NAM CHUL KIM SANG HYUN YANGSAN CAMPUS					
发明人	SHIM, JAEYOON KWAK, BYOUNG JOO KIM, NAM CHUL KIM, SANG HYUN YANGSAN CAI	MPUS				
IPC分类号	A61B8/00 G01S7/539 G01S15/89	G06T15/08 G06T15/00				
CPC分类号	A61B8/00 A61B8/461 A61B8/467 A61B8/469 A61B8/483 Y10S128/916 E02D29/0275 E02D2250/0007 E02D2250/0023					
代理机构(译)	SCHMID , WOLFGANG					
优先权	1020040022567 2004-04-01 KR					
其他公开文献	EP1582150B1					
外部链接	<u>Espacenet</u>					
摘要(译) 本发明涉及一种3D超声ì	诊断 其诵讨自动检测目标对象的轮	廓仅形成 具	Fig. 2			

本发明涉及一种3D超声诊断,其通过自动检测目标对象的轮廓仅形成具 有在轮廓内退出的体数据的3D超声图像,包括:用于在2D超声上生成感 兴趣区域(ROI)盒的第一单元图片;第二单元,用于检测ROI盒中目标 对象的轮廓;第三单元,用于通过渲染检测到的轮廓中存在的体数据来形 成3D超声图像。

